

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2841</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>08/19/2025 09:37</u>
Lab File ID:	<u>VX047400.D</u>	Init. Calib. Date(s):	<u>08/15/2025 08/15/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:40 12:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.720	0.658		-8.61	20
1,2-Dichloroethane-d4	0.700	0.805		15	20
Dibromofluoromethane	0.325	0.344		5.85	20
Toluene-d8	1.160	1.173		1.12	20
4-Bromofluorobenzene	0.428	0.455		6.31	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.