

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2841

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc.

Datafile : VX047403.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0819WBS01	Vinyl chloride	20	16.6	ug/L	83			70 (65)	130 (117)	