

DATA PACKAGE VOLATILE ORGANICS

PROJECT NAME : FORMER SCHLUMBERGER STC PTC SITE D3868221

JACOBS ENGINEERING GROUP, INC.

412 Mt. Kemble Ave

Downtown Building

Morristown, NJ - 07960

Phone No: 9732670555

ORDER ID : Q2842

ATTENTION : John Ynfante



Laboratory Certification ID # 20012



1) VOLATILES DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For VOCMS Group3	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	16
9.3) LCS/LCSD Summary	18
9.4) Method Blank Summary	19
9.5) GS/MS Tune Summary	20
9.6) Internal Standard Area and RT Summary	22
10) Sample Data	24
10.1) RMW-03B-90.4-081225	25
10.2) RMW-02B-66.3-081225	34
10.3) MW-17B-55.5-081225	44
10.4) TB01-081225	53
11) Calibration Data Summary	61
11.1) Initial Calibration Data	62
11.1.1) VN071625	62
11.2) Continued Calibration Data	431
11.2.1) VN087552.D	431
12) QC Sample Data	490
12.1) Tune Raw Data	491
12.2) Method Blank Data	493
12.3) LCS Data	500
12.4) MS Data	553
12.5) MSD Data	606
13) Manual Integration	659
14) Analytical Runlogs	662
15) Standard Prep Logs	667
16) Shipping Document	756

Table Of Contents for Q2842

16.1) Chain Of Custody	757
16.2) ROC	758
16.3) Lab Certificate	762
16.4) Internal COC	763

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

Cover Page

Order ID : Q2842

Project ID : Former Schlumberger STC PTC Site D3868221

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

Q2842-01
Q2842-02
Q2842-03
Q2842-04
Q2842-05
Q2842-06
Q2842-08
Q2842-09
Q2842-10
Q2842-11

Client Sample Number

RMW-03B-90.4-081225
RMW-02B-66.3-081225
MW-17B-55.5-081225
MW-17B-55.5-081225MS
MW-17B-55.5-081225MSD
MW-06-6.5-081225
TB01-081225
MW-17B-55.5-081225
MW-17B-55.5-081225MS
MW-17B-55.5-081225MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:14 am, Oct 28, 2025

Date: 10/28/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger STC PTC Site D3868221

Project # N/A

Order ID # Q2842

Test Name: VOCMS Group3

A. Number of Samples and Date of Receipt:

6 Water samples were received on 08/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group3. This data package contains results for VOCMS Group3.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group3 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples RMW-03B-90.4-081225, RMW-02B-66.3-081225 and MW-17B-55.5-081225, All these samples were run at straight dilution after checking past history of these samples containing high amounts of compounds cis-1,2-Dichloroethene and Trichloroethene.

E. Additional Comments:

This Data Package has been revised due to Data Package type changed.

The SIM analysis is not required for the sample MW-17B-55.5-081225-SIM as all the SIM target analytes are detected at or above the sample adjusted CRQLs in the full scan analysis, a SIM analysis is not to be performed for that sample."

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 9:16 am, Oct 28, 2025

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2842

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds. The MSD recoveries met the requirements for all compounds. The RPD were met for all analysis. The Blank Spike met requirements for all compounds			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

This Data Package has been revised due to Data Package type changed

The SIM analysis is not required for the sample MW-17B-55.5-081225-SIM as all the SIM target analytes are detected at or above the sample adjusted CRQLs in the full scan analysis, a SIM analysis is not to be performed for that sample."

Samples RMW-03B-90.4-08122025, RMW-02B-66.3-08122025 and MW-17B-55.5-08122025, All these samples were run at straight dilution after checking past history of these samples containing high amounts of compounds cis-1,2-Dichloroethene and Trichloroethene.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

REVIEWED

QA REVIEW

By Sohil Jodhani, QA/QC Director at 8:51 am, Oct 28, 2025

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2842

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/28/2025

LAB CHRONICLE

OrderID:	Q2842	OrderDate:	8/13/2025 8:34:00 AM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	Location:	J23,J32,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2842-01	RMW-03B-90.4-08122 5	Water	VOCMS Group3	8260-Low	08/12/25		08/14/25	08/12/25
Q2842-02	RMW-02B-66.3-08122 5	Water	VOCMS Group3	8260-Low	08/12/25		08/14/25	08/12/25
Q2842-03	MW-17B-55.5-081225	Water	VOCMS Group3	8260-Low	08/12/25		08/14/25	08/12/25
Q2842-08	TB01-081225	Water	VOCMS Group3	8260-Low	08/12/25		08/14/25	08/12/25

Hit Summary Sheet
SW-846

SDG No.: Q2842
Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	RMW-03B-90.4-081225							
Q2842-01	RMW-03B-90.4-08	Water	cis-1,2-Dichloroethene	2800		7.60	40.0	ug/L
Q2842-01	RMW-03B-90.4-08	Water	Trichloroethene	810		3.70	40.0	ug/L
			Total Voc :	3610				
			Total Concentration:	3610				
Client ID:	RMW-02B-66.3-081225							
Q2842-02	RMW-02B-66.3-08	Water	Vinyl Chloride	240		10.4	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	1,1-Dichloroethene	130		9.20	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	cis-1,2-Dichloroethene	1600		7.60	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	Trichloroethene	1500		3.70	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	Tetrachloroethene	24.7	J	9.20	40.0	ug/L
			Total Voc :	3490				
			Total Concentration:	3490				
Client ID:	MW-17B-55.5-081225							
Q2842-03	MW-17B-55.5-0812	Water	cis-1,2-Dichloroethene	3000		19.0	100	ug/L
Q2842-03	MW-17B-55.5-0812	Water	Trichloroethene	3600		9.30	100	ug/L
Q2842-03	MW-17B-55.5-0812	Water	Tetrachloroethene	56.3	J	23.0	100	ug/L
			Total Voc :	6660				
			Total Concentration:	6660				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Surrogate Summary

SDG No.: Q2842

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2842-01	RMW-03B-90.4-081225	1,2-Dichloroethane-d4	50	53.7	107		70 (74)	130 (125)
		Dibromofluoromethane	50	45.5	91		70 (75)	130 (124)
		Toluene-d8	50	45.7	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.9	88		70 (77)	130 (121)
Q2842-02	RMW-02B-66.3-081225	1,2-Dichloroethane-d4	50	57.0	114		70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90		70 (75)	130 (124)
		Toluene-d8	50	46.0	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.0	90		70 (77)	130 (121)
Q2842-03	MW-17B-55.5-081225	1,2-Dichloroethane-d4	50	56.6	113		70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90		70 (75)	130 (124)
		Toluene-d8	50	46.5	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.5	91		70 (77)	130 (121)
Q2842-04MS	MW-17B-55.5-081225MS	1,2-Dichloroethane-d4	50	53.8	108		70 (74)	130 (125)
		Dibromofluoromethane	50	47.9	96		70 (75)	130 (124)
		Toluene-d8	50	47.9	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.7	103		70 (77)	130 (121)
Q2842-05MSD	MW-17B-55.5-081225MSD	1,2-Dichloroethane-d4	50	54.6	109		70 (74)	130 (125)
		Dibromofluoromethane	50	46.7	93		70 (75)	130 (124)
		Toluene-d8	50	48.2	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101		70 (77)	130 (121)
Q2842-08	TB01-081225	1,2-Dichloroethane-d4	50	55.1	110		70 (74)	130 (125)
		Dibromofluoromethane	50	44.9	90		70 (75)	130 (124)
		Toluene-d8	50	45.6	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.2	88		70 (77)	130 (121)
VN0814WBL01	VN0814WBL01	1,2-Dichloroethane-d4	50	52.9	106		70 (74)	130 (125)
		Dibromofluoromethane	50	44.2	88		70 (75)	130 (124)
		Toluene-d8	50	44.2	88		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.7	87		70 (77)	130 (121)
VN0814WBS01	VN0814WBS01	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	50.7	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106		70 (77)	130 (121)

() = LABORATORY INHOUSE LIMIT

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

() = LABORATORY INHOUSE LIMIT

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VN087569.D

										Limits	
Parameter	Spike	Sample	Result	Units	Rec		RPD	Qual	Low	High	RPD
		Result			Rec	Qual					
Lab Sample ID : Q2842-04MS		Client Sample ID : MW-17B-55.5-081225MS									
Vinyl chloride	5000	0	5400	ug/L	108				70 (69)	130 (125)	
1,1-Dichloroethene	5000	0	5200	ug/L	104				70 (53)	130 (162)	
1,1-Dichloroethane	5000	0	5300	ug/L	106				70 (78)	130 (122)	
cis-1,2-Dichloroethene	5000	3000	8100	ug/L	102				70 (82)	130 (124)	
1,1,1-Trichloroethane	5000	0	5400	ug/L	108				70 (83)	130 (117)	
Benzene	5000	0	5100	ug/L	102				70 (81)	130 (128)	
1,2-Dichloroethane	5000	0	5500	ug/L	110				70 (76)	130 (120)	
Trichloroethene	5000	3600	10000	ug/L	128				70 (28)	130 (175)	
1,1,2-Trichloroethane	5000	0	5300	ug/L	106				70 (27)	130 (175)	
Tetrachloroethene	5000	56.3	4300	ug/L	85				70 (48)	130 (153)	

() = LABORATORY INHOUSE LIMIT

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VN087570.D

Limits											
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Low	High	RPD
		Result			Qual	RPD	Qual				
Lab Sample ID :	Q2842-05MSD	Client Sample ID :	MW-17B-55.5-081225MSD								
Vinyl chloride	5000	0	5400	ug/L	108		0		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	5000	0	5300	ug/L	106		2		70 (53)	130 (162)	20 (20)
1,1-Dichloroethane	5000	0	5400	ug/L	108		2		70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	5000	3000	8000	ug/L	100		2		70 (82)	130 (124)	20 (20)
1,1,1-Trichloroethane	5000	0	5500	ug/L	110		2		70 (83)	130 (117)	20 (20)
Benzene	5000	0	5100	ug/L	102		0		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	5000	0	5300	ug/L	106		4		70 (76)	130 (120)	20 (20)
Trichloroethene	5000	3600	9500	ug/L	118		8		70 (28)	130 (175)	20 (20)
1,1,2-Trichloroethane	5000	0	5300	ug/L	106		0		70 (27)	130 (175)	20 (20)
Tetrachloroethene	5000	56.3	4300	ug/L	85		0		70 (48)	130 (153)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2842</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>JACOBS Engineering Group, Inc.</u>	Datafile :	<u>VN087555.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0814WBS01	Vinyl chloride	20	21.4	ug/L	107			70 (65)	130 (117)	
	1,1-Dichloroethene	20	20.6	ug/L	103			70 (74)	130 (110)	
	1,1-Dichloroethane	20	21.6	ug/L	108			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	21.2	ug/L	106			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	21.4	ug/L	107			70 (80)	130 (108)	
	Benzene	20	20.3	ug/L	102			70 (82)	130 (109)	
	1,2-Dichloroethane	20	21.5	ug/L	108			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			70 (83)	130 (112)	
	Tetrachloroethene	20	18.1	ug/L	91			70 (67)	130 (123)	

() = LABORATORY INHOUSE LIMIT

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0814WBL01

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q2842

Lab File ID: VN087554.D

Lab Sample ID: VN0814WBL01

Date Analyzed: 08/14/2025

Time Analyzed: 10:32

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0814WBS01	VN0814WBS01	VN087555.D	08/14/2025
TB01-081225	Q2842-08	VN087565.D	08/14/2025
RMW-03B-90.4-081225	Q2842-01	VN087566.D	08/14/2025
RMW-02B-66.3-081225	Q2842-02	VN087567.D	08/14/2025
MW-17B-55.5-081225	Q2842-03	VN087568.D	08/14/2025
MW-17B-55.5-081225MS	Q2842-04MS	VN087569.D	08/14/2025
MW-17B-55.5-081225MSD	Q2842-05MSD	VN087570.D	08/14/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
 BROMOFLUOROBENZENE (BFB)
Lab Name: AllianceContract: JACO05Lab Code: ACESDG NO.: Q2842Lab File ID: VN087327.DBFB Injection Date: 07/16/2025Instrument ID: MSVOA_NBFB Injection Time: 16:10GC Column: RXI-624 ID: 0.25 (mm)Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.8
75	30.0 - 60.0% of mass 95	50.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.1
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	70.9
175	5.0 - 9.0% of mass 174	3.6 (5.1) 1
176	95.0 - 101.0% of mass 174	68.7 (96.9) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN087328.D	07/16/2025	17:05
VSTDIC005	VSTDIC005	VN087329.D	07/16/2025	17:27
VSTDIC020	VSTDIC020	VN087330.D	07/16/2025	17:49
VSTDIC050	VSTDIC050	VN087331.D	07/16/2025	18:11
VSTDIC100	VSTDIC100	VN087332.D	07/16/2025	18:32
VSTDIC150	VSTDIC150	VN087333.D	07/16/2025	18:54



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VN087551.D
Instrument ID: MSVOA_N
GC Column: RXI-624 ID: 0.25 (mm)

Contract: JACO05
SDG NO.: Q2842
BFB Injection Date: 08/14/2025
BFB Injection Time: 09:01
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	5.6
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	65.2
175	5.0 - 9.0% of mass 174	4.4 (6.8) 1
176	95.0 - 101.0% of mass 174	64.4 (98.8) 1
177	5.0 - 9.0% of mass 176	4 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087552.D	08/14/2025	09:34
VN0814WBL01	VN0814WBL01	VN087554.D	08/14/2025	10:32
VN0814WBS01	VN0814WBS01	VN087555.D	08/14/2025	11:07
TB01-081225	Q2842-08	VN087565.D	08/14/2025	14:58
RMW-03B-90.4-081225	Q2842-01	VN087566.D	08/14/2025	15:20
RMW-02B-66.3-081225	Q2842-02	VN087567.D	08/14/2025	15:42
MW-17B-55.5-081225	Q2842-03	VN087568.D	08/14/2025	16:04
MW-17B-55.5-081225MS	Q2842-04MS	VN087569.D	08/14/2025	16:27
MW-17B-55.5-081225MSD	Q2842-05MSD	VN087570.D	08/14/2025	16:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q2842
Lab File ID: VN087552.D Date Analyzed: 08/14/2025
Instrument ID: MSVOA_N Time Analyzed: 09:34
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	271352	8.21	543098	9.08	500619	11.85
UPPER LIMIT	542704	8.712	1086200	9.582	1001240	12.347
LOWER LIMIT	135676	7.712	271549	8.582	250310	11.347
EPA SAMPLE NO.						
RMW-03B-90.4-081225	225616	8.21	494400	9.09	459245	11.85
RMW-02B-66.3-081225	218605	8.21	486591	9.09	438933	11.85
MW-17B-55.5-081225	219937	8.21	487780	9.09	455235	11.85
MW-17B-55.5-081225MS	238067	8.21	461160	9.08	435202	11.85
MW-17B-55.5-081225MSD	251785	8.21	498703	9.08	460182	11.85
TB01-081225	224412	8.21	498664	9.09	458661	11.85
VN0814WBL01	284398	8.21	560321	9.08	503771	11.85
VN0814WBS01	250453	8.21	485218	9.08	446518	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q2842
 Lab File ID: VN087552.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	255173	13.77				
UPPER LIMIT	510346	14.27				
LOWER LIMIT	127587	13.27				
EPA SAMPLE NO.						
RMW-03B-90.4-081225	213548	13.77				
RMW-02B-66.3-081225	211146	13.77				
MW-17B-55.5-081225	209426	13.77				
MW-17B-55.5-081225MS	231778	13.77				
MW-17B-55.5-081225MSD	240017	13.77				
TB01-081225	214662	13.77				
VN0814WBL01	234765	13.77				
VN0814WBS01	221983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	RMW-03B-90.4-081225	SDG No.:	Q2842
Lab Sample ID:	Q2842-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087566.D	40	08/14/25 15:20	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	10.4	U	10.4	40.0	ug/L
75-35-4	1,1-Dichloroethene	9.20	U	9.20	40.0	ug/L
75-34-3	1,1-Dichloroethane	9.20	U	9.20	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	2800		7.60	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	8.00	U	8.00	40.0	ug/L
71-43-2	Benzene	6.00	U	6.00	40.0	ug/L
107-06-2	1,2-Dichloroethane	8.80	U	8.80	40.0	ug/L
79-01-6	Trichloroethene	810		3.70	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	8.40	U	8.40	40.0	ug/L
127-18-4	Tetrachloroethene	9.20	U	9.20	40.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.7		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	45.5		70 (75) - 130 (124)	91%	SPK: 50
2037-26-5	Toluene-d8	45.7		70 (86) - 130 (113)	91%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.9		70 (77) - 130 (121)	88%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	226000	8.206			
540-36-3	1,4-Difluorobenzene	494000	9.088			
3114-55-4	Chlorobenzene-d5	459000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	214000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087566.D
Acq On : 14 Aug 2025 15:20
Operator : JC\MD
Sample : Q2842-01 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-081225

Quant Time: Aug 14 15:39:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.206	168	225616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	494400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	459245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	213548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	205410	53.657	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.320%		
35) Dibromofluoromethane	8.153	113	155257	45.525	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 91.040%		
50) Toluene-d8	10.547	98	555821	45.689	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 91.380%		
62) 4-Bromofluorobenzene	12.829	95	197243	43.886	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 87.780%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.371	101	188722	83.020	ug/l	95
27) cis-1,2-Dichloroethene	7.477	96	234562	70.144	ug/l	91
44) Trichloroethene	9.329	130	69531	20.207	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087566.D
Acq On : 14 Aug 2025 15:20
Operator : JC\MD
Sample : Q2842-01 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-081225

Quant Time: Aug 14 15:39:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. -0.006 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-081225


Tgt Ion:168 Resp: 225616
Ion Ratio Lower Upper
168 100
99 66.9 47.9 71.9



#9
1,1,2-Trichlorotrifluoroethane
Concen: 83.020 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.012 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Tgt Ion:101 Resp: 188722
Ion Ratio Lower Upper
101 100
85 49.1 37.3 55.9
151 68.6 58.9 88.3





#27

cis-1,2-Dichloroethene

Concen: 70.144 ug/l

RT: 7.477 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087566.D


Acq: 14 Aug 2025 15:20

Instrument :

MSVOA_N

ClientSampleId :

RMW-03B-90.4-081225






Tgt Ion: 96 Resp: 234562

Ion Ratio Lower Upper

96 100

61 135.3 0.0 297.8

98 62.2 0.0 132.4



#33

1,2-Dichloroethane-d4

Concen: 53.657 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087566.D




Acq: 14 Aug 2025 15:20


Tgt Ion: 65 Resp: 205410

Ion Ratio Lower Upper

65 100

67 51.7 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087566.D

Acq: 14 Aug 2025 15:20

Instrument :

MSVOA_N

ClientSampleId :

RMW-03B-90.4-081225





Tgt Ion:114 Resp: 494400

Ion Ratio Lower Upper

114 100

63 22.9 0.0 44.6

88 15.5 0.0 32.8



#35

Dibromofluoromethane

Concen: 45.525 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087566.D

Acq: 14 Aug 2025 15:20




Tgt Ion:113 Resp: 155257


Ion Ratio Lower Upper

113 100

111 102.1 82.5 123.7

192 15.1 13.7 20.5





#44

Trichloroethene

Concen: 20.207 ug/l

RT: 9.329 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN087566.D

Acq: 14 Aug 2025 15:20

Instrument :

MSVOA_N

ClientSampleId :





RMW-03B-90.4-081225

Tgt Ion:130 Resp: 69531

Ion Ratio Lower Upper

130 100

95 96.4 0.0 195.2



#50

Toluene-d8

Concen: 45.689 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. -0.000 min

Lab File: VN087566.D




Acq: 14 Aug 2025 15:20


Tgt Ion: 98 Resp: 555821

Ion Ratio Lower Upper

98 100

100 64.4 52.1 78.1









#62
4-Bromofluorobenzene
Concen: 43.886 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-081225


Tgt Ion: 95 Resp: 197243
Ion Ratio Lower Upper
95 100
174 69.3 0.0 149.4
176 64.4 0.0 141.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20


Tgt Ion: 117 Resp: 459245
Ion Ratio Lower Upper
117 100
82 58.2 47.4 71.2
119 32.0 23.8 35.8







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-081225



Tgt Ion:	152	Resp:	213548
Ion Ratio	Lower	Upper	
152	100		
115	63.6	31.1	93.5
150	154.0	0.0	349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	RMW-02B-66.3-081225	SDG No.:	Q2842
Lab Sample ID:	Q2842-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087567.D	40	08/14/25 15:42	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	240		10.4	40.0	ug/L
75-35-4	1,1-Dichloroethene	130		9.20	40.0	ug/L
75-34-3	1,1-Dichloroethane	9.20	U	9.20	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	1600		7.60	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	8.00	U	8.00	40.0	ug/L
71-43-2	Benzene	6.00	U	6.00	40.0	ug/L
107-06-2	1,2-Dichloroethane	8.80	U	8.80	40.0	ug/L
79-01-6	Trichloroethene	1500		3.70	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	8.40	U	8.40	40.0	ug/L
127-18-4	Tetrachloroethene	24.7	J	9.20	40.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.0		70 (74) - 130 (125)	114%	SPK: 50
1868-53-7	Dibromofluoromethane	45.1		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	46.0		70 (86) - 130 (113)	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.0		70 (77) - 130 (121)	90%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	219000	8.212			
540-36-3	1,4-Difluorobenzene	487000	9.088			
3114-55-4	Chlorobenzene-d5	439000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	211000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087567.D
 Acq On : 14 Aug 2025 15:42
 Operator : JC\MD
 Sample : Q2842-02 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-02B-66.3-081225

Quant Time: Aug 14 19:59:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	218605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	486591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	438933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	211146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	211465	57.010	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.020%			
35) Dibromofluoromethane	8.153	113	151404	45.108	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 90.220%			
50) Toluene-d8	10.547	98	550924	46.014	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.020%			
62) 4-Bromofluorobenzene	12.829	95	199247	45.043	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.080%			
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.542	62	17532	6.042	ug/l	97
12) 1,1-Dichloroethene	4.330	96	8400	3.365	ug/l	94
27) cis-1,2-Dichloroethene	7.471	96	128340	39.610	ug/l	94
44) Trichloroethene	9.335	130	129163	38.140	ug/l	94
64) Tetrachloroethene	11.076	164	1746	0.618	ug/l #	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087567.D
Acq On : 14 Aug 2025 15:42
Operator : JC\MD
Sample : Q2842-02 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-081225

Quant Time: Aug 14 19:59:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-081225


Tgt Ion:168 Resp: 218605
Ion Ratio Lower Upper
168 100
99 65.5 47.9 71.9



#4
Vinyl Chloride
Concen: 6.042 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42

Tgt Ion: 62 Resp: 17532
Ion Ratio Lower Upper
62 100
64 31.9 27.0 40.6





#12

1,1-Dichloroethene

Concen: 3.365 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-081225





Tgt Ion: 96 Resp: 8400

Ion Ratio Lower Upper

96 100

61 155.5 132.3 198.5

98 61.4 46.8 70.2



#27

cis-1,2-Dichloroethene

Concen: 39.610 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42




Tgt Ion: 96 Resp: 128340


Ion Ratio Lower Upper

96 100

61 139.9 0.0 297.8

98 64.5 0.0 132.4





#33

1,2-Dichloroethane-d4

Concen: 57.010 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :





RMW-02B-66.3-081225

Tgt Ion: 65 Resp: 211465

Ion Ratio Lower Upper

65 100

67 49.2 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42




Tgt Ion: 114 Resp: 486591


Ion Ratio Lower Upper

114 100

63 24.5 0.0 44.6

88 16.8 0.0 32.8





#35

Dibromofluoromethane

Concen: 45.108 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-081225





Tgt Ion:113 Resp: 151404

Ion Ratio Lower Upper

113 100

111 104.3 82.5 123.7

192 15.8 13.7 20.5



#44

Trichloroethene

Concen: 38.140 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087567.D




Acq: 14 Aug 2025 15:42


Tgt Ion:130 Resp: 129163

Ion Ratio Lower Upper

130 100

95 103.7 0.0 195.2









#50
Toluene-d8
Concen: 46.014 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-081225


Tgt Ion: 98 Resp: 550924
Ion Ratio Lower Upper
98 100
100 64.8 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 45.043 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42

Tgt Ion: 95 Resp: 199247
Ion Ratio Lower Upper
95 100
174 66.4 0.0 149.4
176 64.9 0.0 141.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.847 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-081225





Tgt Ion:117 Resp: 438933

Ion Ratio Lower Upper

117 100

82 61.8 47.4 71.2

119 31.0 23.8 35.8



#64

Tetrachloroethene

Concen: 0.618 ug/l

RT: 11.076 min Scan# 1551

Delta R.T. -0.012 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Tgt Ion:164 Resp: 1746




Ion Ratio Lower Upper


164 100

166 103.9 105.5 158.3#

129 82.5 77.4 116.2

131 70.4 77.3 115.9#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087567.D


Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-081225





Tgt Ion:152 Resp: 211146

Ion Ratio Lower Upper

152 100

115 63.3 31.1 93.5

150 156.0 0.0 349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225	SDG No.:	Q2842
Lab Sample ID:	Q2842-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087568.D	100	08/14/25 16:04	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	26.0	U	26.0	100	ug/L
75-35-4	1,1-Dichloroethene	23.0	U	23.0	100	ug/L
75-34-3	1,1-Dichloroethane	23.0	U	23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	3000		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	20.0	U	20.0	100	ug/L
71-43-2	Benzene	15.0	U	15.0	100	ug/L
107-06-2	1,2-Dichloroethane	22.0	U	22.0	100	ug/L
79-01-6	Trichloroethene	3600		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	21.0	U	21.0	100	ug/L
127-18-4	Tetrachloroethene	56.3	J	23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.6		70 (74) - 130 (125)	113%	SPK: 50
1868-53-7	Dibromofluoromethane	45.1		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	46.5		70 (86) - 130 (113)	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.5		70 (77) - 130 (121)	91%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	220000	8.212			
540-36-3	1,4-Difluorobenzene	488000	9.088			
3114-55-4	Chlorobenzene-d5	455000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	209000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087568.D
Acq On : 14 Aug 2025 16:04
Operator : JC\MD
Sample : Q2842-03 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225

Quant Time: Aug 14 19:59:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	219937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	487780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	455235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	209426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	211074	56.560	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.120%	
35) Dibromofluoromethane	8.153	113	151622	45.062	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	90.120%	
50) Toluene-d8	10.547	98	557660	46.463	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.920%	
62) 4-Bromofluorobenzene	12.829	95	201626	45.470	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.940%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.477	96	99130	30.410	ug/l	93
44) Trichloroethene	9.335	130	123655	36.424	ug/l	98
64) Tetrachloroethene	11.082	164	1650	0.563	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087568.D
Acq On : 14 Aug 2025 16:04
Operator : JC\MD
Sample : Q2842-03 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225

Quant Time: Aug 14 19:59:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225


Tgt Ion:168 Resp: 219937
Ion Ratio Lower Upper
168 100
99 65.5 47.9 71.9



#27
cis-1,2-Dichloroethene
Concen: 30.410 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Tgt Ion: 96 Resp: 99130
Ion Ratio Lower Upper
96 100
61 139.9 0.0 297.8
98 62.0 0.0 132.4





#33

1,2-Dichloroethane-d4

Concen: 56.560 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225

Tgt Ion: 65 Resp: 211074

Ion Ratio Lower Upper

65 100

67 49.4 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04




Tgt Ion: 114 Resp: 487780


Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.6

88 15.4 0.0 32.8





#35

Dibromofluoromethane

Concen: 45.062 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225





Tgt Ion:113 Resp: 151622

Ion Ratio Lower Upper

113 100

111 104.2 82.5 123.7

192 15.7 13.7 20.5



#44

Trichloroethene

Concen: 36.424 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087568.D




Acq: 14 Aug 2025 16:04


Tgt Ion:130 Resp: 123655

Ion Ratio Lower Upper

130 100

95 99.4 0.0 195.2









#50
Toluene-d8
Concen: 46.463 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225


Tgt Ion: 98 Resp: 557660
Ion Ratio Lower Upper
98 100
100 63.3 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 45.470 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Tgt Ion: 95 Resp: 201626
Ion Ratio Lower Upper
95 100
174 65.3 0.0 149.4
176 63.8 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225


Tgt Ion:117 Resp: 455235
Ion Ratio Lower Upper
117 100
82 59.8 47.4 71.2
119 31.3 23.8 35.8



#64
Tetrachloroethene
Concen: 0.563 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Tgt Ion:164 Resp: 1650
Ion Ratio Lower Upper
164 100
166 122.4 105.5 158.3
129 103.2 77.4 116.2
131 52.6 77.3 115.9#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087568.D


Acq: 14 Aug 2025 16:04

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225





Tgt Ion:152 Resp: 209426

Ion Ratio Lower Upper

152 100

115 64.3 31.1 93.5

150 157.1 0.0 349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	TB01-081225	SDG No.:	Q2842
Lab Sample ID:	Q2842-08	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087565.D	1	08/14/25 14:58	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.1		70 (74) - 130 (125)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	44.9		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	45.6		70 (86) - 130 (113)	91%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.2		70 (77) - 130 (121)	88%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	224000	8.212			
540-36-3	1,4-Difluorobenzene	499000	9.088			
3114-55-4	Chlorobenzene-d5	459000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	215000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087565.D
Acq On : 14 Aug 2025 14:58
Operator : JC\MD
Sample : Q2842-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-081225

Quant Time: Aug 14 15:25:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	224412	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.088	114	498664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	458661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	214662	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	209873	55.117	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.240%	
35) Dibromofluoromethane	8.153	113	154326	44.865	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	89.740%	
50) Toluene-d8	10.547	98	559384	45.589	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	91.180%	
62) 4-Bromofluorobenzene	12.829	95	200248	44.173	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.340%	
Target Compounds						
16) Acetone	4.424	43	8905	4.988	ug/l	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087565.D
Acq On : 14 Aug 2025 14:58
Operator : JC\MD
Sample : Q2842-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-081225

Quant Time: Aug 14 15:25:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58




Instrument :
MSVOA_N
ClientSampleId :
TB01-081225


Tgt Ion:168 Resp: 224412
Ion Ratio Lower Upper
168 100
99 72.0 47.9 71.9#



#16
Acetone
Concen: 4.988 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58

Tgt Ion: 43 Resp: 8905
Ion Ratio Lower Upper
43 100
58 21.4 22.3 33.5#





#33

1,2-Dichloroethane-d4

Concen: 55.117 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :





TB01-081225

Tgt Ion: 65 Resp: 209873

Ion Ratio Lower Upper

65 100

67 50.1 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58




Tgt Ion: 114 Resp: 498664


Ion Ratio Lower Upper

114 100

63 23.0 0.0 44.6

88 16.0 0.0 32.8





#35

Dibromofluoromethane

Concen: 44.865 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :

TB01-081225





Tgt Ion:113 Resp: 154326

Ion Ratio Lower Upper

113 100

111 103.2 82.5 123.7

192 15.9 13.7 20.5



#50

Toluene-d8

Concen: 45.589 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087565.D




Acq: 14 Aug 2025 14:58


Tgt Ion: 98 Resp: 559384

Ion Ratio Lower Upper

98 100

100 64.6 52.1 78.1









#62
4-Bromofluorobenzene
Concen: 44.173 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58




Instrument :
MSVOA_N
ClientSampleId :
TB01-081225


Tgt Ion: 95 Resp: 200248
Ion Ratio Lower Upper
95 100
174 68.6 0.0 149.4
176 63.7 0.0 141.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58

Tgt Ion: 117 Resp: 458661
Ion Ratio Lower Upper
117 100
82 63.8 47.4 71.2
119 32.1 23.8 35.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087565.D


Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :

TB01-081225





Tgt Ion:152 Resp: 214662

Ion Ratio Lower Upper

152 100

115 64.0 31.1 93.5

150 156.9 0.0 349.0





CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG No.: Q2842
 Instrument ID: MSVOA_N Calibration Date(s): 07/16/2025 07/16/2025
 Heated Purge: (Y/N) N Calibration Time(s): 17:05 18:54
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN087328.D		RRF005 = VN087329.D		RRF020 = VN087330.D			
		RRF050 = VN087331.D		RRF100 = VN087332.D		RRF150 = VN087333.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Vinyl Chloride	0.554	0.665	0.623	0.728	0.692	0.720	0.664	9.9	
1,1-Dichloroethene	0.635	0.641	0.553	0.545	0.514	0.537	0.571	9.4	
1,1-Dichloroethane	1.304	1.325	1.254	1.222	1.185	1.212	1.250	4.4	
cis-1,2-Dichloroethene	0.704	0.737	0.747	0.767	0.740	0.751	0.741	2.8	
1,1,1-Trichloroethane	1.043	1.146	1.096	1.084	1.049	1.085	1.084	3.4	
Benzene	1.370	1.430	1.502	1.553	1.483	1.499	1.473	4.3	
1,2-Dichloroethane	0.553	0.565	0.569	0.567	0.544	0.552	0.558	1.8	
Trichloroethene	0.373	0.330	0.337	0.356	0.339	0.352	0.348	4.5	
1,1,2-Trichloroethane	0.367	0.364	0.365	0.367	0.357	0.354	0.362	1.6	
Tetrachloroethene	0.329	0.338	0.317	0.320	0.310	0.317	0.322	3.2	
1,2-Dichloroethane-d4		0.939	0.820	0.802	0.818	0.863	0.848	6.6	
Dibromofluoromethane		0.347	0.341	0.332	0.348	0.356	0.345	2.6	
Toluene-d8		1.180	1.176	1.224	1.255	1.316	1.230	4.7	
4-Bromofluorobenzene		0.405	0.433	0.455	0.476	0.505	0.455	8.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N071625W.M
 Title : SW846 8260
 Last Update : Thu Jul 17 02:56:13 2025
 Response Via : Initial Calibration

Calibration Files

1 =VN087328.D 5 =VN087329.D 20 =VN087330.D 50 =VN087331.D 100 =VN087332.D 150 =VN087333.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.447	0.443	0.443	0.623	0.606	0.625	0.531	17.98
3) P	Chloromethane	0.714	0.659	0.588	0.690	0.659	0.698	0.668	6.70
4) C	Vinyl Chloride	0.554	0.665	0.623	0.728	0.692	0.720	0.664	9.93#
5) T	Bromomethane		0.328	0.308	0.355	0.356	0.370	0.344	7.22
6) T	Chloroethane	0.396	0.485	0.431	0.441	0.415	0.429	0.433	6.96
7) T	Trichlorofluor...	0.959	0.975	0.963	1.025	0.960	1.007	0.981	2.83
8) T	Diethyl Ether	0.335	0.391	0.399	0.395	0.371	0.393	0.381	6.40
9) T	1,1,2-Trichlor...	0.463	0.495	0.539	0.525	0.491	0.509	0.504	5.34
10) T	Methyl Iodide		0.288	0.360	0.494	0.554	0.564	0.452	27.17
11) T	Tert butyl alc...		0.164	0.155	0.161	0.161	0.164	0.161	2.41
12) CM	1,1-Dichloroet...	0.635	0.641	0.553	0.545	0.514	0.537	0.571	9.41#
13) T	Acrolein		0.140	0.112	0.121	0.130	0.143	0.129	9.85
14) T	Allyl chloride	0.949	0.978	1.134	1.002	0.995	1.141	1.033	8.02
15) T	Acrylonitrile	0.419	0.419	0.455	0.453	0.436	0.440	0.437	3.57
16) T	Acetone	0.455	0.426	0.393	0.387	0.361	0.366	0.398	9.15
17) T	Carbon Disulfide	1.686	1.685	1.669	1.733	1.643	1.739	1.693	2.20
18) T	Methyl Acetate	1.007	1.052	0.991	0.991	0.962	0.993	0.999	2.99
19) T	Methyl tert-bu...	1.911	2.099	2.106	2.167	2.129	2.213	2.104	4.93
20) T	Methylene Chlo...	1.107	0.788	0.700	0.672	0.655	0.672	0.766	22.70
21) T	trans-1,2-Dich...	0.667	0.669	0.643	0.653	0.618	0.613	0.644	3.69
22) T	Diisopropyl ether	1.797	2.199	2.353	2.286	2.165	2.201	2.167	8.94
23) T	Vinyl Acetate	1.421	1.685	2.101	2.114	2.007	2.044	1.895	14.81
24) P	1,1-Dichloroet...	1.304	1.325	1.254	1.222	1.185	1.212	1.250	4.38
25) T	2-Butanone	0.552	0.608	0.650	0.643	0.617	0.618	0.615	5.65
26) T	2,2-Dichloropr...	1.013	0.963	1.001	0.982	0.933	0.941	0.972	3.34
27) T	cis-1,2-Dichlo...	0.704	0.737	0.747	0.767	0.740	0.751	0.741	2.83
28) T	Bromochloromet...	0.596	0.640	0.578	0.595	0.597	0.584	0.598	3.64
29) T	Tetrahydrofuran	0.360	0.397	0.413	0.425	0.400	0.401	0.399	5.52
30) C	Chloroform	1.181	1.299	1.303	1.279	1.214	1.234	1.251	3.97#
31) T	Cyclohexane		1.148	1.039	1.046	0.981	1.002	1.043	6.16
32) T	1,1,1-Trichlor...	1.043	1.146	1.096	1.084	1.049	1.085	1.084	3.42
33) S	1,2-Dichloroet...		0.939	0.820	0.802	0.818	0.863	0.848	6.56
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.347	0.341	0.332	0.348	0.356	0.345	2.55
36) T	1,1-Dichloropr...	0.387	0.437	0.459	0.487	0.477	0.487	0.456	8.47
37) T	Ethyl Acetate	0.587	0.671	0.655	0.706	0.660	0.669	0.658	5.94
38) T	Carbon Tetrach...	0.453	0.523	0.498	0.517	0.504	0.518	0.502	5.10
39) T	Methylcyclohexane	0.447	0.442	0.482	0.529	0.519	0.541	0.493	8.62
40) TM	Benzene	1.370	1.430	1.502	1.553	1.483	1.499	1.473	4.34
41) T	Methacrylonitrile	0.289	0.330	0.362	0.368	0.353	0.363	0.344	8.81
42) TM	1,2-Dichloroet...	0.553	0.565	0.569	0.567	0.544	0.552	0.558	1.78
43) T	Isopropyl Acetate	0.938	0.999	1.029	1.071	1.038	1.055	1.022	4.64
44) TM	Trichloroethene	0.373	0.330	0.337	0.356	0.339	0.352	0.348	4.49
45) C	1,2-Dichloropr...	0.335	0.367	0.395	0.395	0.376	0.378	0.374	5.90#
46) T	Dibromomethane	0.265	0.296	0.279	0.289	0.274	0.278	0.280	3.94
47) T	Bromodichlorom...	0.568	0.572	0.559	0.569	0.553	0.565	0.564	1.21
48) T	Methyl methacr...	0.372	0.412	0.475	0.505	0.493	0.503	0.460	12.06
49) T	1,4-Dioxane	0.006	0.007	0.007	0.008	0.008	0.008	0.007	12.52
50) S	Toluene-d8		1.180	1.176	1.224	1.255	1.316	1.230	4.72
51) T	4-Methyl-2-Pen...	0.551	0.641	0.685	0.689	0.658	0.652	0.646	7.79
52) CM	Toluene	0.774	0.849	0.940	0.963	0.916	0.929	0.895	7.93#
53) T	t-1,3-Dichloro...	0.459	0.536	0.586	0.621	0.607	0.619	0.571	11.08
54) T	cis-1,3-Dichlo...	0.489	0.564	0.602	0.632	0.620	0.632	0.590	9.40
55) T	1,1,2-Trichlor...	0.367	0.364	0.365	0.367	0.357	0.354	0.362	1.57
56) T	Ethyl methacry...	0.348	0.486	0.567	0.627	0.626	0.656	0.552	21.20

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N071625W.M

57)	T	1,3-Dichloropr...	0.556	0.621	0.652	0.655	0.636	0.640	0.627	5.82
58)	T	2-Chloroethyl ...	0.220	0.273	0.303	0.346	0.310	0.332	0.297	15.20
59)	T	2-Hexanone	0.279	0.373	0.465	0.495	0.481	0.479	0.429	19.95
60)	T	Dibromochlorom...	0.352	0.416	0.425	0.430	0.424	0.433	0.413	7.39
61)	T	1,2-Dibromoethane	0.367	0.373	0.391	0.385	0.381	0.389	0.381	2.49
62)	S	4-Bromofluorob...	0.405	0.433	0.455	0.476	0.505	0.455		8.46
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.329	0.338	0.317	0.320	0.310	0.317	0.322	3.16
65)	PM	Chlorobenzene	1.139	1.131	1.133	1.119	1.092	1.122	1.123	1.48
66)	T	1,1,1,2-Tetrac...	0.324	0.406	0.393	0.393	0.381	0.394	0.382	7.64
67)	C	Ethyl Benzene	1.643	1.738	1.882	1.942	1.905	1.979	1.848	7.04#
68)	T	m/p-Xylenes	0.541	0.646	0.717	0.758	0.734	0.756	0.692	12.20
69)	T	o-Xylene	0.491	0.606	0.702	0.723	0.710	0.734	0.661	14.39
70)	T	Styrene	0.726	1.032	1.186	1.255	1.217	1.257	1.112	18.59
71)	P	Bromoform	0.246	0.302	0.314	0.328	0.322	0.337	0.308	10.69
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	2.524	2.889	3.242	3.396	3.302	3.529	3.147	11.86
74)	T	N-amyl acetate	1.309	1.256	1.249	1.138	1.378	1.515	1.307	9.83
75)	P	1,1,2,2-Tetrac...	1.159	1.181	1.228	1.207	1.156	1.174	1.184	2.37
76)	T	1,2,3-Trichlor...	1.208	1.238	1.101	1.191	0.969	1.019	1.121	9.79
77)	T	Bromobenzene	0.669	0.791	0.862	0.869	0.830	0.876	0.816	9.63
78)	T	n-propylbenzene	3.252	3.716	4.089	4.294	4.109	4.295	3.959	10.25
79)	T	2-Chlorotoluene	2.085	2.348	2.535	2.533	2.493	2.605	2.433	7.84
80)	T	1,3,5-Trimethy...	2.069	2.517	2.816	2.928	2.799	2.959	2.681	12.62
81)	T	trans-1,4-Dich...	0.327	0.448	0.404	0.417	0.453	0.410		12.30
82)	T	4-Chlorotoluene	2.225	2.460	2.618	2.647	2.552	2.699	2.533	6.79
83)	T	tert-Butylbenzene	1.781	2.068	2.303	2.423	2.376	2.485	2.239	11.92
84)	T	1,2,4-Trimethy...	2.201	2.411	2.906	3.003	2.872	3.034	2.738	12.64
85)	T	sec-Butylbenzene	3.120	3.136	3.460	3.565	3.387	3.570	3.373	5.98
86)	T	p-Isopropyltol...	2.081	2.417	2.806	2.991	2.893	3.032	2.703	13.90
87)	T	1,3-Dichlorobe...	1.448	1.592	1.651	1.657	1.588	1.674	1.602	5.19
88)	T	1,4-Dichlorobe...	1.807	1.709	1.742	1.703	1.627	1.677	1.711	3.56
89)	T	n-Butylbenzene	2.278	2.425	2.648	2.761	2.614	2.762	2.581	7.49
90)	T	Hexachloroethane	0.593	0.564	0.562	0.576	0.552	0.590	0.573	2.86
91)	T	1,2-Dichlorobe...	1.303	1.534	1.596	1.577	1.510	1.583	1.517	7.23
92)	T	1,2-Dibromo-3-...	0.339	0.325	0.304	0.302	0.290	0.305	0.311	5.76
93)	T	1,2,4-Trichlor...	0.761	0.860	0.875	0.937	0.921	0.994	0.891	8.94
94)	T	Hexachlorobuta...	0.351	0.337	0.324	0.329	0.316	0.331	0.331	3.57
95)	T	Naphthalene	2.486	2.658	3.113	3.441	3.451	3.797	3.158	16.01
96)	T	1,2,3-Trichlor...	0.803	0.844	0.887	0.922	0.917	0.992	0.894	7.37

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087328.D
 Acq On : 16 Jul 2025 17:05
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:16:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	166177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	312474	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	274838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	128808	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	2.142	85	1484	0.841	ug/l	98
3) Chloromethane	2.389	50	2372	1.069	ug/l	96
4) Vinyl Chloride	2.542	62	1841	0.835	ug/l #	68
6) Chloroethane	3.153	64	1316	0.915	ug/l	93
7) Trichlorofluoromethane	3.506	101	3187	0.977	ug/l	92
8) Diethyl Ether	3.965	74	1114	0.880	ug/l	79
9) 1,1,2-Trichlorotrifluo...	4.371	101	1539	0.919	ug/l #	31
12) 1,1-Dichloroethene	4.347	96	2112	1.113	ug/l #	53
14) Allyl chloride	5.012	41	3155m	0.919	ug/l	
15) Acrylonitrile	5.712	53	6965	4.794	ug/l #	80
16) Acetone	4.430	43	7566m	5.660	ug/l	
17) Carbon Disulfide	4.700	76	5602	0.996	ug/l #	71
18) Methyl Acetate	5.018	43	3347	1.008	ug/l #	74
19) Methyl tert-butyl Ether	5.789	73	6351m	0.908	ug/l	
20) Methylene Chloride	5.271	84	3679	1.019	ug/l #	74
21) trans-1,2-Dichloroethene	5.777	96	2216	1.036	ug/l #	53
22) Diisopropyl ether	6.659	45	5973	0.829	ug/l #	75
23) Vinyl Acetate	6.600	43	23617	3.749	ug/l #	93
24) 1,1-Dichloroethane	6.547	63	4333	1.043	ug/l #	83
25) 2-Butanone	7.477	43	9174	4.491	ug/l	99
26) 2,2-Dichloropropane	7.471	77	3368	1.043	ug/l #	67
27) cis-1,2-Dichloroethene	7.471	96	2340	0.950	ug/l	99
28) Bromochloromethane	7.794	49	1982	0.997	ug/l #	85
29) Tetrahydrofuran	7.835	42	5981	4.507	ug/l	94
30) Chloroform	7.953	83	3924	0.943	ug/l #	71
32) 1,1,1-Trichloroethane	8.165	97	3468	0.963	ug/l #	45
36) 1,1-Dichloropropene	8.353	75	2421	0.850	ug/l	96
37) Ethyl Acetate	7.547	43	3670	0.892	ug/l #	82
38) Carbon Tetrachloride	8.341	117	2833	0.903	ug/l #	83
39) Methylcyclohexane	9.588	83	2794	0.906	ug/l #	86
40) Benzene	8.588	78	8561	0.930	ug/l	97
41) Methacrylonitrile	7.771	41	1805	0.839	ug/l #	87
42) 1,2-Dichloroethane	8.653	62	3456	0.990	ug/l	89
43) Isopropyl Acetate	8.677	43	5865	0.919	ug/l #	95
44) Trichloroethene	9.341	130	2331	1.072	ug/l	70
45) 1,2-Dichloropropane	9.606	63	2094	0.895	ug/l #	86

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087328.D
 Acq On : 16 Jul 2025 17:05
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:16:50 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.694	93	1657	0.946	ug/l #	87
47) Bromodichloromethane	9.871	83	3549	1.006	ug/l #	91
48) Methyl methacrylate	9.671	41	2322	0.808	ug/l	94
49) 1,4-Dioxane	9.682	88	696	15.810	ug/l #	53
51) 4-Methyl-2-Pentanone	10.429	43	17216	4.263	ug/l	91
52) Toluene	10.612	92	4834	0.864	ug/l	91
53) t-1,3-Dichloropropene	10.818	75	2869	0.804	ug/l #	56
54) cis-1,3-Dichloropropene	10.294	75	3059	0.830	ug/l	94
55) 1,1,2-Trichloroethane	11.000	97	2294	1.013	ug/l #	83
56) Ethyl methacrylate	10.859	69	2173	1.217	ug/l #	90
57) 1,3-Dichloropropane	11.141	76	3477	0.888	ug/l	95
58) 2-Chloroethyl Vinyl ether	10.141	63	6887	3.707	ug/l	96
59) 2-Hexanone	11.182	43	8707	3.250	ug/l	87
60) Dibromochloromethane	11.347	129	2200	0.852	ug/l	82
61) 1,2-Dibromoethane	11.459	107	2291	0.962	ug/l	90
64) Tetrachloroethene	11.088	164	1807	1.022	ug/l #	75
65) Chlorobenzene	11.870	112	6259	1.014	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	1783	0.850	ug/l #	50
67) Ethyl Benzene	11.941	91	9031	0.889	ug/l	92
68) m/p-Xylenes	12.053	106	5952	1.565	ug/l	97
69) o-Xylene	12.388	106	2700	0.743	ug/l	91
70) Styrene	12.400	104	3989	0.653	ug/l	91
71) Bromoform	12.559	173	1351	0.797	ug/l #	94
73) Isopropylbenzene	12.682	105	6501	0.802	ug/l	95
74) N-amyl acetate	12.841	43	3372m	1.065	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	2986	0.979	ug/l #	78
76) 1,2,3-Trichloropropane	12.976	75	3113m	1.089	ug/l	
77) Bromobenzene	12.953	156	1724	0.820	ug/l	79
78) n-propylbenzene	13.017	91	8378	0.821	ug/l	95
79) 2-Chlorotoluene	13.100	91	5372	0.857	ug/l	93
80) 1,3,5-Trimethylbenzene	13.147	105	5329	0.772	ug/l	93
82) 4-Chlorotoluene	13.200	91	5733	0.878	ug/l	96
83) tert-Butylbenzene	13.412	119	4588	0.795	ug/l	90
84) 1,2,4-Trimethylbenzene	13.464	105	5671	0.804	ug/l	87
85) sec-Butylbenzene	13.594	105	8037	0.925	ug/l	97
86) p-Isopropyltoluene	13.712	119	5360	0.770	ug/l	92
87) 1,3-Dichlorobenzene	13.712	146	3731	0.904	ug/l	89
88) 1,4-Dichlorobenzene	13.794	146	4655m	1.056	ug/l	
89) n-Butylbenzene	14.035	91	5869	0.883	ug/l	91
90) Hexachloroethane	14.317	117	1527	1.035	ug/l	87
91) 1,2-Dichlorobenzene	14.088	146	3358	0.859	ug/l	85
92) 1,2-Dibromo-3-Chloropr...	14.711	75	874	1.091	ug/l #	56
93) 1,2,4-Trichlorobenzene	15.376	180	1960	0.854	ug/l	78
94) Hexachlorobutadiene	15.476	225	903	1.058	ug/l	76
95) Naphthalene	15.611	128	6404	0.787	ug/l #	95
96) 1,2,3-Trichlorobenzene	15.806	180	2068	0.898	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:168 Resp: 16617
Ion Ratio Lower Upper
168 100
99 63.3 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 0.841 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 85 Resp: 1484
Ion Ratio Lower Upper
85 100
87 36.7 17.8 53.3









#3
Chloromethane
Concen: 1.069 ug/l
RT: 2.389 min Scan# 74
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 2372
Ion Ratio Lower Upper
50 100
52 35.1 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 0.835 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 62 Resp: 1841
Ion Ratio Lower Upper
62 100
64 51.9 27.0 40.6#





#6
Chloroethane
Concen: 0.915 ug/l
RT: 3.153 min Scan# 204
Delta R.T. 0.024 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion: 64 Resp: 1310
Ion Ratio Lower Upper
64 100
66 34.8 24.6 36.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#7
Trichlorofluoromethane
Concen: 0.977 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05



Tgt Ion:101 Resp: 3187
Ion Ratio Lower Upper
101 100
103 68.4 49.8 74.6



#8
Diethyl Ether

Concen: 0.880 ug/l

RT: 3.965 min Scan# 341

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 74 Resp: 1114

Ion Ratio Lower Upper

74 100





45 122.6 50.8 152.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.919 ug/l

RT: 4.371 min Scan# 411

Delta R.T. 0.012 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05




Tgt Ion:101 Resp: 1539


Ion Ratio Lower Upper

101 100

85 20.5 37.3 55.9#

151 0.0 58.9 88.3#





#12

1,1-Dichloroethene

Concen: 1.113 ug/l

RT: 4.347 min Scan# 407

Delta R.T. 0.018 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	96	Resp:	211
Ion	Ratio	Lower	Upper
96	100		
61	102.7	132.3	198.5
98	21.8	46.8	70.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 0.919 ug/l m




RT: 5.012 min Scan# 520


Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	41	Resp:	3155
Ion	Ratio	Lower	Upper
41	100		
39	46.1	59.0	88.6
76	18.6	28.7	43.1





#15

Acrylonitrile

Concen: 4.794 ug/l

RT: 5.712 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	53	Resp:	696
Ion	Ratio	Lower	Upper
53	100		
52	75.2	65.5	98.3
51	7.2	28.7	43.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#16

Acetone

Concen: 5.660 ug/l m




RT: 4.430 min Scan# 421


Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	43	Resp:	7566
Ion	Ratio	Lower	Upper
43	100		
58	27.2	22.3	33.5





#17

Carbon Disulfide

Concen: 0.996 ug/l

RT: 4.700 min Scan# 4

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 76 Resp: 560

Ion Ratio Lower Upper

76 100





78 0.0 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#18

Methyl Acetate

Concen: 1.008 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.006 min

Lab File: VN087328.D




Acq: 16 Jul 2025 17:05


Tgt Ion: 43 Resp: 3347

Ion Ratio Lower Upper

43 100

74 9.5 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 0.908 ug/l m

RT: 5.789 min Scan# 6

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 635

Ion Ratio Lower Upper

73 100





57 13.5 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 1.019 ug/l

RT: 5.271 min Scan# 564

Delta R.T. 0.012 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion: 84 Resp: 3679




Ion Ratio Lower Upper


84 100

49 93.3 107.5 161.3#

51 29.8 30.2 45.2#

86 50.0 49.3 73.9





#21

trans-1,2-Dichloroethene

Concen: 1.036 ug/l

RT: 5.777 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 96 Resp: 2210

Ion Ratio Lower Upper

96 100

61 85.5 122.0 183.0#





98 37.1 50.0 75.0#

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#22

Diisopropyl ether

Concen: 0.829 ug/l

RT: 6.659 min Scan# 800

Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion: 45 Resp: 5973




Ion Ratio Lower Upper


45 100

43 77.6 42.8 64.2#

87 20.3 19.8 29.6

59 17.1 8.3 12.5#









#23
Vinyl Acetate
Concen: 3.749 ug/l
RT: 6.600 min Scan# 790
Delta R.T. 0.012 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
Client Sample Id :
VSTDICC001

Tgt Ion: 43 Resp: 2361
Ion Ratio Lower Upper
43 100
86 7.1 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#24
1,1-Dichloroethane
Concen: 1.043 ug/l
RT: 6.547 min Scan# 781
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 63 Resp: 4333
Ion Ratio Lower Upper
63 100
98 0.0 3.3 9.9#
100 0.0 2.5 7.4#





#25

2-Butanone

Concen: 4.491 ug/l

RT: 7.477 min Scan# 917

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 9174

Ion Ratio Lower Upper

43 100





72 24.9 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#26

2,2-Dichloropropane

Concen: 1.043 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087328.D




Acq: 16 Jul 2025 17:05


Tgt Ion: 77 Resp: 3368

Ion Ratio Lower Upper

77 100

97 6.5 11.1 33.1#





#27

cis-1,2-Dichloroethene

Concen: 0.950 ug/l

RT: 7.471 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	96	Resp:	2340
Ion	Ratio	Lower	Upper
96	100		
61	147.9	0.0	297.8
98	65.5	0.0	132.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 0.997 ug/l




RT: 7.794 min Scan# 993


Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	49	Resp:	1982
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.2
130	58.7	57.3	85.9





#29

Tetrahydrofuran

Concen: 4.507 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	42	Resp:	598
Ion	Ratio	Lower	Upper
42	100		
72	37.1	33.4	50.0
71	36.3	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 0.943 ug/l




RT: 7.953 min Scan# 1020


Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	83	Resp:	3924
Ion	Ratio	Lower	Upper
83	100		
85	42.9	52.7	79.1#





#32

1,1,1-Trichloroethane

Concen: 0.963 ug/l

RT: 8.165 min Scan# 1054

Delta R.T. 0.012 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 346

Ion Ratio Lower Upper

97 100

99 0.0 51.8 77.8





61 36.3 38.7 58.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 8.56 min

Lab File: VN087328.D


Acq: 16 Jul 2025 17:05


Tgt Ion: 65

Sig Exp Ratio

65 100

67 52.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:114 Resp: 312474

Ion Ratio Lower Upper

114 100

63 22.7 0.0 44.6





88 16.7 0.0 32.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.15 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.1

192 17.1





#36

1,1-Dichloropropene

Concen: 0.850 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt Ion:	75	Resp:	242
Ion	Ratio	Lower	Upper
75	100		
110	29.6	16.7	50.1
77	30.6	25.2	37.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#37

Ethyl Acetate

Concen: 0.892 ug/l




RT: 7.547 min Scan# 951


Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	43	Resp:	3670
Ion	Ratio	Lower	Upper
43	100		
61	3.7	10.9	16.3#
70	6.5	7.4	11.0#





#38

Carbon Tetrachloride

Concen: 0.903 ug/l

RT: 8.341 min Scan# 1086

Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	117	Resp:	283
Ion	Ratio	Lower	Upper
117	100		
119	83.1	80.2	120.2
121	21.7	25.4	38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#39

Methylcyclohexane

Concen: 0.906 ug/l




RT: 9.588 min Scan# 1298


Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	83	Resp:	2794
Ion	Ratio	Lower	Upper
83	100		
55	65.7	61.3	91.9
98	36.9	37.9	56.9





#40

Benzene

Concen: 0.930 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 78 Resp: 856

Ion Ratio Lower Upper

78 100





77 24.2 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#41

Methacrylonitrile

Concen: 0.839 ug/l

RT: 7.771 min Scan# 989

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion: 41 Resp: 1805




Ion Ratio Lower Upper


41 100

39 60.9 43.4 65.0

67 76.1 55.1 82.7

52 15.8 24.9 37.3#





#42

1,2-Dichloroethane

Concen: 0.990 ug/l

RT: 8.653 min Scan# 1139

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion: 62 Resp: 3450
Ion Ratio Lower Upper
62 100
98 5.1 0.0 18.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#43

Isopropyl Acetate

Concen: 0.919 ug/l




RT: 8.677 min Scan# 1143


Delta R.T. -0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion: 43 Resp: 5865
Ion Ratio Lower Upper
43 100
61 24.8 19.8 29.8
87 6.6 9.8 14.6









#44
Trichloroethene
Concen: 1.072 ug/l
RT: 9.341 min Scan# 1256
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:130 Resp: 233
Ion Ratio Lower Upper
130 100
95 68.2 0.0 195.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#45
1,2-Dichloropropane
Concen: 0.895 ug/l
RT: 9.606 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 63 Resp: 2094
Ion Ratio Lower Upper
63 100
65 41.9 27.0 40.4#





#46

Dibromomethane

Concen: 0.946 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	93	Resp:	165
Ion	Ratio	Lower	Upper
93	100		
95	62.2	65.8	98.8
174	84.2	69.9	104.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#47

Bromodichloromethane

Concen: 1.006 ug/l




RT: 9.871 min Scan# 1346


Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	83	Resp:	3549
Ion	Ratio	Lower	Upper
83	100		
85	68.8	50.4	75.6
127	0.0	5.8	8.8





#48

Methyl methacrylate

Concen: 0.808 ug/l

RT: 9.671 min Scan# 1311

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05





Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 41	Resp: 232
Ion Ratio	Lower Upper
41	100
69	87.9 64.1 96.1
39	56.1 45.5 68.3

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#49

1,4-Dioxane

Concen: 15.810 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion: 88	Resp: 696
Ion Ratio	Lower Upper
88	100
43	0.0 0.0 0.0
58	35.9 61.1 91.7





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 10.55 min



Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 98
Sig Exp Ratio
98 100
100 65.1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#51
4-Methyl-2-Pentanone
Concen: 4.263 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 43 Resp: 17216
Ion Ratio Lower Upper
43 100
58 33.0 30.8 46.2





#52
Toluene
Concen: 0.864 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 92 Resp: 4834
Ion Ratio Lower Upper
92 100
91 157.1 135.1 202.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#53
t-1,3-Dichloropropene
Concen: 0.804 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 75 Resp: 2869
Ion Ratio Lower Upper
75 100
77 53.0 23.6 35.4#





#54
cis-1,3-Dichloropropene
Concen: 0.830 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 75 Resp: 3059
Ion Ratio Lower Upper
75 100
77 31.6 24.2 36.2
39 53.2 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#55
1,1,2-Trichloroethane
Concen: 1.013 ug/l
RT: 11.000 min Scan# 1538
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 97 Resp: 2294
Ion Ratio Lower Upper
97 100
83 56.4 72.6 109.0#
85 58.9 46.7 70.1
99 62.8 54.1 81.1





#56
Ethyl methacrylate
Concen: 1.217 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 69 Resp: 217
Ion Ratio Lower Upper
69 100
41 75.4 55.1 82.7
39 42.6 27.9 41.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#57
1,3-Dichloropropane
Concen: 0.888 ug/l
RT: 11.141 min Scan# 1562
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 76 Resp: 3477
Ion Ratio Lower Upper
76 100
78 29.5 26.0 39.0









#58
2-Chloroethyl Vinyl ether
Concen: 3.707 ug/l
RT: 10.141 min Scan# 1392
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 63 Resp: 688
Ion Ratio Lower Upper
63 100
106 24.9 21.7 32.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#59
2-Hexanone
Concen: 3.250 ug/l
RT: 11.182 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 43 Resp: 8707
Ion Ratio Lower Upper
43 100
58 43.8 26.7 80.0





#60
Dibromochloromethane
Concen: 0.852 ug/l
RT: 11.347 min Scan# 1597
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion:129 Resp: 2200
Ion Ratio Lower Upper
129 100
127 93.6 39.1 117.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#61
1,2-Dibromoethane
Concen: 0.962 ug/l
RT: 11.459 min Scan# 1616
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05



Tgt Ion:107 Resp: 2291
Ion Ratio Lower Upper
107 100
109 87.3 77.5 116.3





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.83 min

Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 95
Sig Exp Ratio
95 100
174 74.7
176 70.6

Instrument :

MSVOA_N



ClientSampleId :

VSTDICC001

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion:117 Resp: 274838
Ion Ratio Lower Upper
117 100
82 54.2 47.4 71.2
119 31.2 23.8 35.8





#64
Tetrachloroethene
Concen: 1.022 ug/l
RT: 11.088 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion:164 Resp: 180
Ion Ratio Lower Upper
164 100
166 108.3 105.5 158.3
129 73.2 77.4 116.2
131 65.6 77.3 115.9#




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#65
Chlorobenzene
Concen: 1.014 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion:112 Resp: 6259
Ion Ratio Lower Upper
112 100
114 34.3 26.5 39.7









#66
1,1,1,2-Tetrachloroethane
Concen: 0.850 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:131 Resp: 178
Ion Ratio Lower Upper
131 100
133 120.9 47.4 142.3
119 0.0 33.1 99.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#67
Ethyl Benzene
Concen: 0.889 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 91 Resp: 9031
Ion Ratio Lower Upper
91 100
106 34.9 24.3 36.5





#68

m/p-Xylenes

Concen: 1.565 ug/l

RT: 12.053 min Scan# 1717

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:106 Resp: 595

Ion Ratio Lower Upper

106 100





91 198.2 162.0 243.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#69

o-Xylene

Concen: 0.743 ug/l

RT: 12.388 min Scan# 1774

Delta R.T. 0.012 min

Lab File: VN087328.D




Acq: 16 Jul 2025 17:05


Tgt Ion:106 Resp: 2700

Ion Ratio Lower Upper

106 100

91 229.8 107.7 323.3









#70
Styrene
Concen: 0.653 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:104 Resp: 3989
Ion Ratio Lower Upper
104 100
78 48.8 41.0 61.6
103 64.6 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#71
Bromoform
Concen: 0.797 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion:173 Resp: 1351
Ion Ratio Lower Upper
173 100
175 52.5 24.1 72.3
254 0.0 0.0 0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001




Tgt Ion:152 Resp: 128800
Ion Ratio Lower Upper
152 100
115 61.0 31.1 93.5
150 160.6 0.0 349.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#73
Isopropylbenzene
Concen: 0.802 ug/l
RT: 12.682 min Scan# 1824
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Tgt Ion:105 Resp: 6501
Ion Ratio Lower Upper
105 100
120 24.0 13.4 40.1





#74
N-amy1 acetate
Concen: 1.065 ug/l m
RT: 12.841 min Scan# 1851
Delta R.T. 0.347 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 43 Resp: 3372
Ion Ratio Lower Upper
43 100
70 0.0 37.6 56.4#
55 0.0 19.6 29.4#
61 0.0 20.6 31.0#




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#75
1,1,2,2-Tetrachloroethane
Concen: 0.979 ug/l
RT: 12.917 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 83 Resp: 2986
Ion Ratio Lower Upper
83 100
131 1.8 5.1 15.3#
85 82.1 32.5 97.4









#76
1,2,3-Trichloropropane
Concen: 1.089 ug/l m
RT: 12.976 min Scan# 1873
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 75 Resp: 311
Ion Ratio Lower Upper
75 100
77 153.5 94.5 283.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#77
Bromobenzene
Concen: 0.820 ug/l
RT: 12.953 min Scan# 1870
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion:156 Resp: 1724
Ion Ratio Lower Upper
156 100
77 277.2 114.9 344.6
158 99.6 48.5 145.5









#78
n-propylbenzene
Concen: 0.821 ug/l
RT: 13.017 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion: 91 Resp: 8378
Ion Ratio Lower Upper
91 100
120 19.9 11.3 33.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#79
2-Chlorotoluene
Concen: 0.857 ug/l
RT: 13.100 min Scan# 1895
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion: 91 Resp: 5372
Ion Ratio Lower Upper
91 100
126 29.6 16.9 50.6









#80
1,3,5-Trimethylbenzene
Concen: 0.772 ug/l
RT: 13.147 min Scan# 1904
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:105 Resp: 5329
Ion Ratio Lower Upper
105 100
120 52.9 24.3 72.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 0.878 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05


Tgt Ion: 91 Resp: 5733
Ion Ratio Lower Upper
91 100
126 31.1 16.6 49.7





#83
tert-Butylbenzene
Concen: 0.795 ug/l
RT: 13.412 min Scan# 1948
Delta R.T. -0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion:119 Resp: 4588
Ion Ratio Lower Upper
119 100
91 79.4 35.0 105.1
134 29.0 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 0.804 ug/l
RT: 13.464 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion:105 Resp: 5671
Ion Ratio Lower Upper
105 100
120 36.8 22.8 68.3









#85
sec-Butylbenzene
Concen: 0.925 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion:105 Resp: 803
Ion Ratio Lower Upper
105 100
134 18.1 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#86
p-Isopropyltoluene
Concen: 0.770 ug/l
RT: 13.712 min Scan# 1999
Delta R.T. 0.006 min
Lab File: VN087328.D
Acq: 16 Jul 2025 17:05

Tgt Ion:119 Resp: 5360
Ion Ratio Lower Upper
119 100
134 34.2 13.5 40.5
91 25.4 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 0.904 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	146	Resp:	373
Ion Ratio	Lower	Upper	
146	100		
111	36.5	21.4	64.3
148	72.6	31.6	95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 1.056 ug/l m




RT: 13.794 min Scan# 2013


Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	146	Resp:	4655
Ion Ratio	Lower	Upper	
146	100		
111	29.2	19.6	58.7
148	58.2	31.4	94.0





#89

n-Butylbenzene

Concen: 0.883 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	91	Resp:	586
Ion	Ratio	Lower	Upper
91	100		
92	44.0	26.2	78.6
134	25.9	12.4	37.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 1.035 ug/l




RT: 14.317 min Scan# 2102


Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	117	Resp:	1527
Ion	Ratio	Lower	Upper
117	100		
201	55.1	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 0.859 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	146	Resp:	3358
Ion	Ratio	Lower	Upper
146	100		
111	47.6	22.0	66.0
148	82.7	32.7	98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 1.091 ug/l




RT: 14.711 min Scan# 2169


Delta R.T. 0.012 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:	75	Resp:	874
Ion	Ratio	Lower	Upper
75	100		
155	37.2	37.3	111.8#
157	50.2	46.2	138.6





#93

1,2,4-Trichlorobenzene

Concen: 0.854 ug/l

RT: 15.376 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:180 Resp: 1960

Ion Ratio Lower Upper

180 100

182 121.8 47.3 141.8





145 32.6 15.4 46.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 1.058 ug/l

RT: 15.476 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05




Tgt Ion:225 Resp: 903


Ion Ratio Lower Upper

225 100

223 42.9 32.1 96.3

227 46.0 31.3 93.9





#95

Naphthalene

Concen: 0.787 ug/l

RT: 15.611 min Scan# 2323

Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:128	Resp:	640
Ion Ratio	Lower	Upper
128	100	
127	13.7	10.5 15.7
129	7.0	8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96

1,2,3-Trichlorobenzene

Concen: 0.898 ug/l




RT: 15.806 min Scan# 2355

Delta R.T. -0.006 min

Lab File: VN087328.D

Acq: 16 Jul 2025 17:05

Tgt Ion:180	Resp:	2068
Ion Ratio	Lower	Upper
180	100	
182	93.6	47.1 141.4
145	34.1	16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087329.D
 Acq On : 16 Jul 2025 17:27
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.206	168	172217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	319926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	285583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	146494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	16177	5.536	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	11.080%#		
35) Dibromofluoromethane	8.153	113	11111	5.035	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.060%#		
50) Toluene-d8	10.547	98	37751	4.796	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	9.600%#		
62) 4-Bromofluorobenzene	12.829	95	12943	4.450	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	8.900%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	7633	4.173	ug/l	90
3) Chloromethane	2.383	50	11350	4.934	ug/l	96
4) Vinyl Chloride	2.536	62	11459	5.013	ug/l	98
5) Bromomethane	2.971	94	5657	4.779	ug/l	87
6) Chloroethane	3.124	64	8358	5.606	ug/l #	81
7) Trichlorofluoromethane	3.506	101	16796	4.969	ug/l #	83
8) Diethyl Ether	3.959	74	6731	5.133	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.371	101	8532	4.917	ug/l #	19
10) Methyl Iodide	4.577	142	4952m	8.093	ug/l	
11) Tert butyl alcohol	5.536	59	14147	25.497	ug/l	98
12) 1,1-Dichloroethene	4.330	96	11037	5.613	ug/l #	83
13) Acrolein	4.177	56	12045	27.050	ug/l	96
14) Allyl chloride	5.006	41	16836	4.731	ug/l	97
15) Acrylonitrile	5.718	53	36114	23.986	ug/l	92
16) Acetone	4.430	43	36648	26.455	ug/l	96
17) Carbon Disulfide	4.695	76	29024	4.979	ug/l	94
18) Methyl Acetate	5.006	43	18125	5.265	ug/l	99
19) Methyl tert-butyl Ether	5.800	73	36155	4.989	ug/l	95
20) Methylene Chloride	5.259	84	13570	5.286	ug/l #	86
21) trans-1,2-Dichloroethene	5.765	96	11514	5.193	ug/l	81
22) Diisopropyl ether	6.671	45	37879	5.075	ug/l #	97
23) Vinyl Acetate	6.594	43	145089	22.225	ug/l #	94
24) 1,1-Dichloroethane	6.547	63	22819	5.299	ug/l	95
25) 2-Butanone	7.477	43	52336	24.722	ug/l	95
26) 2,2-Dichloropropane	7.483	77	16583	4.953	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	12690	4.972	ug/l	99
28) Bromochloromethane	7.789	49	11022	5.348	ug/l #	16
29) Tetrahydrofuran	7.830	42	34146	24.829	ug/l	99
30) Chloroform	7.959	83	22369	5.190	ug/l	84
31) Cyclohexane	8.236	56	19764	5.502	ug/l	87
32) 1,1,1-Trichloroethane	8.153	97	19739	5.287	ug/l #	51
36) 1,1-Dichloropropene	8.347	75	13965	4.790	ug/l	98
37) Ethyl Acetate	7.553	43	21482	5.102	ug/l	97
38) Carbon Tetrachloride	8.341	117	16718	5.205	ug/l #	89
39) Methylcyclohexane	9.582	83	14154	4.484	ug/l	86
40) Benzene	8.583	78	45748	4.855	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087329.D
 Acq On : 16 Jul 2025 17:27
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	10551	4.792	ug/l	97
42) 1,2-Dichloroethane	8.653	62	18086	5.061	ug/l	96
43) Isopropyl Acetate	8.677	43	31964	4.890	ug/l	100
44) Trichloroethene	9.335	130	10573	4.748	ug/l	79
45) 1,2-Dichloropropane	9.606	63	11728	4.898	ug/l	95
46) Dibromomethane	9.694	93	9473	5.284	ug/l	96
47) Bromodichloromethane	9.865	83	18300	5.068	ug/l #	82
48) Methyl methacrylate	9.665	41	13169	4.475	ug/l	95
49) 1,4-Dioxane	9.682	88	4176	92.653	ug/l #	87
51) 4-Methyl-2-Pentanone	10.430	43	102519	24.797	ug/l	97
52) Toluene	10.612	92	27158	4.742	ug/l	96
53) t-1,3-Dichloropropene	10.824	75	17133	4.688	ug/l #	87
54) cis-1,3-Dichloropropene	10.294	75	18046	4.780	ug/l	97
55) 1,1,2-Trichloroethane	11.000	97	11643	5.021	ug/l #	88
56) Ethyl methacrylate	10.859	69	15548	4.693	ug/l	95
57) 1,3-Dichloropropane	11.141	76	19868	4.956	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.147	63	43690	22.968	ug/l	94
59) 2-Hexanone	11.182	43	59736	21.778	ug/l	99
60) Dibromochloromethane	11.335	129	13321	5.037	ug/l	100
61) 1,2-Dibromoethane	11.453	107	11942	4.898	ug/l	89
64) Tetrachloroethene	11.082	164	9660	5.256	ug/l	95
65) Chlorobenzene	11.871	112	32304	5.038	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.941	131	11581	5.312	ug/l	96
67) Ethyl Benzene	11.947	91	49622	4.701	ug/l	92
68) m/p-Xylenes	12.059	106	36883	9.332	ug/l	98
69) o-Xylene	12.376	106	17293	4.580	ug/l	97
70) Styrene	12.394	104	29462	4.639	ug/l	99
71) Bromoform	12.559	173	8628	4.899	ug/l #	96
73) Isopropylbenzene	12.676	105	42320	4.590	ug/l	98
74) N-amyl acetate	12.647	43	18405m	5.113	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	17305	4.988	ug/l	95
76) 1,2,3-Trichloropropane	12.976	75	18135m	5.580	ug/l	
77) Bromobenzene	12.965	156	11589	4.847	ug/l	84
78) n-propylbenzene	13.018	91	54443	4.693	ug/l	99
79) 2-Chlorotoluene	13.106	91	34400	4.825	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	36874	4.694	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	4795	3.994	ug/l #	80
82) 4-Chlorotoluene	13.200	91	36031	4.854	ug/l	98
83) tert-Butylbenzene	13.418	119	30293	4.617	ug/l	96
84) 1,2,4-Trimethylbenzene	13.465	105	35326	4.403	ug/l	99
85) sec-Butylbenzene	13.594	105	45947	4.649	ug/l	98
86) p-Isopropyltoluene	13.706	119	35413	4.471	ug/l	97
87) 1,3-Dichlorobenzene	13.718	146	23326	4.970	ug/l	95
88) 1,4-Dichlorobenzene	13.794	146	25031	4.994	ug/l	90
89) n-Butylbenzene	14.035	91	35522	4.697	ug/l	95
90) Hexachloroethane	14.312	117	8261	4.923	ug/l	98
91) 1,2-Dichlorobenzene	14.088	146	22474	5.055	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	4757	5.223	ug/l	87
93) 1,2,4-Trichlorobenzene	15.364	180	12595	4.823	ug/l	97
94) Hexachlorobutadiene	15.476	225	4936	5.087	ug/l	98
95) Naphthalene	15.617	128	38936	4.208	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	12368	4.721	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087329.D
Acq On : 16 Jul 2025 17:27
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087329.D
Acq On : 16 Jul 2025 17:27
Operator : JC\MD
Sample : VSTDIC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Jul 17 02:17:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Instrument :

MSVOA_N

ClientSampleId :


VSTDIC005


Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1064
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:168 Resp: 17221
Ion Ratio Lower Upper
168 100
99 57.7 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 4.173 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 85 Resp: 7633
Ion Ratio Lower Upper
85 100
87 29.4 17.8 53.3









#3
Chloromethane
Concen: 4.934 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 11350
Ion Ratio Lower Upper
50 100
52 30.6 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 5.013 ug/l
RT: 2.536 min Scan# 99
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 62 Resp: 11459
Ion Ratio Lower Upper
62 100
64 34.7 27.0 40.6





#5

Bromomethane

Concen: 4.779 ug/l

RT: 2.971 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 94 Resp: 565

Ion Ratio Lower Upper

94 100





96 79.0 73.4 110.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#6

Chloroethane

Concen: 5.606 ug/l

RT: 3.124 min Scan# 199

Delta R.T. -0.006 min

Lab File: VN087329.D




Acq: 16 Jul 2025 17:27


Tgt Ion: 64 Resp: 8358

Ion Ratio Lower Upper

64 100

66 20.2 24.6 36.8#









#7
Trichlorofluoromethane
Concen: 4.969 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 101 Resp: 16790
Ion Ratio Lower Upper
101 100
103 48.8 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 5.133 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 74 Resp: 6731
Ion Ratio Lower Upper
74 100
45 111.0 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.917 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.012 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 853
Ion Ratio Lower Upper
101 100
85 0.0 37.3 55.9
151 0.0 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#10
Methyl Iodide
Concen: 8.093 ug/l m
RT: 4.577 min Scan# 446
Delta R.T. -0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion:142 Resp: 4952
Ion Ratio Lower Upper
142 100
127 44.4 35.7 53.5
141 15.6 10.4 15.6









#11
Tert butyl alcohol
Concen: 25.497 ug/l
RT: 5.536 min Scan# 609
Delta R.T. 0.012 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 59 Resp: 1414
Ion Ratio Lower Upper
59 100
57 9.7 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#12
1,1-Dichloroethene
Concen: 5.613 ug/l
RT: 4.330 min Scan# 404
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 96 Resp: 11037
Ion Ratio Lower Upper
96 100
61 142.2 132.3 198.5
98 46.1 46.8 70.2#





#13

Acrolein

Concen: 27.050 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 1204

Ion Ratio Lower Upper

56 100





55 66.7 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 4.731 ug/l

RT: 5.006 min Scan# 519

Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion: 41 Resp: 16836


Ion Ratio Lower Upper

41 100

39 71.9 59.0 88.6

76 38.3 28.7 43.1





#15

Acrylonitrile

Concen: 23.986 ug/l

RT: 5.718 min Scan# 640

Delta R.T. 0.006 min

Lab File: VN087329.D


Acq: 16 Jul 2025 17:27

Instrument :




MSVOA_N

ClientSampleId :

VSTDIC005



Tgt Ion:	53	Resp:	36114
Ion	Ratio	Lower	Upper
53	100		
52	90.3	65.5	98.3
51	37.7	28.7	43.1

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#16

Acetone


Concen: 26.455 ug/l

RT: 4.430 min Scan# 421



Delta R.T. 0.006 min


Lab File: VN087329.D

Acq: 16 Jul 2025 17:27



Tgt Ion:	43	Resp:	36648
Ion	Ratio	Lower	Upper
43	100		
58	30.0	22.3	33.5





#17

Carbon Disulfide

Concen: 4.979 ug/l

RT: 4.695 min Scan# 41

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27





Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:	76	Resp:	29024
Ion	Ratio	Lower	Upper
76	100		
78	8.6	8.6	13.0

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#18

Methyl Acetate

Concen: 5.265 ug/l




RT: 5.006 min Scan# 519


Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	43	Resp:	18125
Ion	Ratio	Lower	Upper
43	100		
74	22.8	17.8	26.6





#19

Methyl tert-butyl Ether

Concen: 4.989 ug/l

RT: 5.800 min Scan# 61

Delta R.T. 0.012 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 3615

Ion Ratio Lower Upper

73 100





57 20.7 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 5.286 ug/l

RT: 5.259 min Scan# 562

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion: 84 Resp: 13570




Ion Ratio Lower Upper


84 100

49 113.7 107.5 161.3

51 29.3 30.2 45.2#

86 55.5 49.3 73.9





#21

trans-1,2-Dichloroethene

Concen: 5.193 ug/l

RT: 5.765 min Scan# 649

Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:	96	Resp:	11514
Ion	Ratio	Lower	Upper
96	100		
61	124.3	122.0	183.0
98	53.3	50.0	75.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22

Diisopropyl ether

Concen: 5.075 ug/l




RT: 6.671 min Scan# 802


Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	45	Resp:	37879
Ion	Ratio	Lower	Upper
45	100		
43	55.4	42.8	64.2
87	23.4	19.8	29.6
59	13.5	8.3	12.5









#23
Vinyl Acetate
Concen: 22.225 ug/l
RT: 6.594 min Scan# 789
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 145089
Ion Ratio Lower Upper
43 100
86 11.8 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#24
1,1-Dichloroethane
Concen: 5.299 ug/l
RT: 6.547 min Scan# 781
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 63 Resp: 22819
Ion Ratio Lower Upper
63 100
98 5.1 3.3 9.9
100 6.7 2.5 7.4









#25
2-Butanone
Concen: 24.722 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 5233
Ion Ratio Lower Upper
43 100
72 27.0 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#26
2,2-Dichloropropane
Concen: 4.953 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.012 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 77 Resp: 16583
Ion Ratio Lower Upper
77 100
97 22.2 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 4.972 ug/l

RT: 7.471 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:	96	Resp:	12690
Ion	Ratio	Lower	Upper
96	100		
61	150.3	0.0	297.8
98	66.8	0.0	132.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 5.348 ug/l




RT: 7.789 min Scan# 992


Delta R.T. -0.012 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	49	Resp:	11022
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.2
130	0.0	57.3	85.9#





#29

Tetrahydrofuran

Concen: 24.829 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	42	Resp:	3414
Ion	Ratio	Lower	Upper
42	100		
72	42.1	33.4	50.0
71	37.6	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 5.190 ug/l




RT: 7.959 min Scan# 1021


Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	83	Resp:	22369
Ion	Ratio	Lower	Upper
83	100		
85	53.5	52.7	79.1





#31

Cyclohexane

Concen: 5.502 ug/l

RT: 8.236 min Scan# 1068

Delta R.T. -0.006 min

Lab File: VN087329.D


Acq: 16 Jul 2025 17:27

Instrument :




MSVOA_N

ClientSampleId :

VSTDIC005



Tgt Ion:	56	Resp:	1976
Ion Ratio	100	Lower	Upper
56	100		
69	32.8	22.7	34.1
84	69.3	65.8	98.6

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#32

1,1,1-Trichloroethane


Concen: 5.287 ug/l

RT: 8.153 min Scan# 1054



Delta R.T. 0.000 min


Lab File: VN087329.D

Acq: 16 Jul 2025 17:27



Tgt Ion:	97	Resp:	19739
Ion Ratio	100	Lower	Upper
97	100		
99	0.0	51.8	77.8
61	45.3	38.7	58.1





#33

1,2-Dichloroethane-d4

Concen: 5.536 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 1617

Ion Ratio Lower Upper

65 100





67 48.6 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion:114 Resp: 319926


Ion Ratio Lower Upper

114 100

63 21.3 0.0 44.6


88 15.3 0.0 32.8





#35
Dibromofluoromethane
Concen: 5.035 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005



Tgt Ion: 113 Resp: 1111
Ion Ratio Lower Upper
113 100
111 104.4 82.5 123.7
192 17.7 13.7 20.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#36
1,1-Dichloropropene
Concen: 4.790 ug/l
RT: 8.347 min Scan# 1087
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 75 Resp: 13965
Ion Ratio Lower Upper
75 100
110 35.8 16.7 50.1
77 31.2 25.2 37.8





#37

Ethyl Acetate

Concen: 5.102 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	43	Resp:	2148
Ion	Ratio	Lower	Upper
43	100		
61	12.2	10.9	16.3
70	8.5	7.4	11.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 5.205 ug/l




RT: 8.341 min Scan# 1086


Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	117	Resp:	16718
Ion	Ratio	Lower	Upper
117	100		
119	109.3	80.2	120.2
121	41.8	25.4	38.2#





#39

Methylcyclohexane

Concen: 4.484 ug/l

RT: 9.582 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion: 83	Resp: 14154
Ion Ratio	Lower Upper
83	100
55	91.6 61.3 91.9
98	53.8 37.9 56.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#40

Benzene

Concen: 4.855 ug/l




RT: 8.583 min Scan# 1127


Delta R.T. -0.012 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27


Tgt Ion: 78	Resp: 45748
Ion Ratio	Lower Upper
78	100
77	25.8 18.2 27.2





#41
Methacrylonitrile
Concen: 4.792 ug/l
RT: 7.765 min Scan# 911
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005




Tgt Ion: 41 Resp: 1055
Ion Ratio Lower Upper
41 100
39 52.1 43.4 65.0
67 65.8 55.1 82.7
52 31.6 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#42
1,2-Dichloroethane
Concen: 5.061 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27



Tgt Ion: 62 Resp: 18086
Ion Ratio Lower Upper
62 100
98 7.7 0.0 18.0





#43

Isopropyl Acetate

Concen: 4.890 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 43	Resp: 31964
Ion Ratio	Lower Upper
43 100	
61 24.7	19.8 29.8
87 12.2	9.8 14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene

Concen: 4.748 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:130	Resp: 10573
Ion Ratio	Lower Upper
130 100	
95 118.0	0.0 195.2





#45

1,2-Dichloropropane

Concen: 4.898 ug/l

RT: 9.606 min Scan# 11728

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 11728

Ion Ratio Lower Upper

63 100





65 30.7 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#46

Dibromomethane

Concen: 5.284 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion: 93 Resp: 9473


Ion Ratio Lower Upper

93 100

95 79.6 65.8 98.8

174 92.1 69.9 104.9





#47

Bromodichloromethane

Concen: 5.068 ug/l

RT: 9.865 min Scan# 1346

Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	83	Resp:	18300
Ion	Ratio	Lower	Upper
83	100		
85	47.9	50.4	75.6
127	9.7	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#48

Methyl methacrylate

Concen: 4.475 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	41	Resp:	13169
Ion	Ratio	Lower	Upper
41	100		
69	84.9	64.1	96.1
39	54.4	45.5	68.3





#49

1,4-Dioxane

Concen: 92.653 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 88 Resp: 4170

Ion Ratio Lower Upper

88 100

43 25.3 0.0 0.0





58 65.4 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 4.796 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087329.D




Acq: 16 Jul 2025 17:27


Tgt Ion: 98 Resp: 37751

Ion Ratio Lower Upper

98 100


100 66.6 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 24.797 ug/l
RT: 10.430 min Scan# 1441
Delta R.T. 0.001 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005




Tgt Ion: 43 Resp: 102519
Ion Ratio Lower Upper
43 100
58 36.6 30.8 46.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#52
Toluene
Concen: 4.742 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27



Tgt Ion: 92 Resp: 27158
Ion Ratio Lower Upper
92 100
91 174.6 135.1 202.7









#53
t-1,3-Dichloropropene
Concen: 4.688 ug/l
RT: 10.824 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 75 Resp: 1713
Ion Ratio Lower Upper
75 100
77 36.3 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 4.780 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 75 Resp: 18046
Ion Ratio Lower Upper
75 100
77 34.0 24.2 36.2
39 47.8 38.4 57.6





#55

1,1,2-Trichloroethane

Concen: 5.021 ug/l

RT: 11.000 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 1164

Ion Ratio Lower Upper

97 100

83 72.6 72.6 109.0

85 59.0 46.7 70.1





99 58.7 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#56

Ethyl methacrylate

Concen: 4.693 ug/l

RT: 10.859 min Scan# 1514

Delta R.T. -0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion: 69 Resp: 15548


Ion Ratio Lower Upper

69 100

41 67.5 55.1 82.7

39 28.8 27.9 41.9





#57

1,3-Dichloropropane

Concen: 4.956 ug/l

RT: 11.141 min Scan# 1562

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 1986

Ion Ratio Lower Upper

76 100





78 32.1 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 22.968 ug/l

RT: 10.147 min Scan# 1393

Delta R.T. 0.006 min

Lab File: VN087329.D




Acq: 16 Jul 2025 17:27


Tgt Ion: 63 Resp: 43690

Ion Ratio Lower Upper

63 100

106 30.2 21.7 32.5





#59

2-Hexanone

Concen: 21.778 ug/l

RT: 11.182 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 59730

Ion Ratio Lower Upper

43 100





58 52.6 26.7 80.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#60

Dibromochloromethane

Concen: 5.037 ug/l

RT: 11.335 min Scan# 1595

Delta R.T. -0.006 min

Lab File: VN087329.D




Acq: 16 Jul 2025 17:27


Tgt Ion:129 Resp: 13321

Ion Ratio Lower Upper

129 100


127 77.9 39.1 117.5





#61
1,2-Dibromoethane
Concen: 4.898 ug/l
RT: 11.453 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005




Tgt Ion:107 Resp: 1194
Ion Ratio Lower Upper
107 100
109 86.3 77.5 116.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#62
4-Bromofluorobenzene
Concen: 4.450 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Tgt Ion: 95 Resp: 12943
Ion Ratio Lower Upper
95 100
174 76.6 0.0 149.4
176 76.1 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005




Tgt Ion:117 Resp: 28558
Ion Ratio Lower Upper
117 100
82 57.9 47.4 71.2
119 30.8 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#64
Tetrachloroethene
Concen: 5.256 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27



Tgt Ion:164 Resp: 9660
Ion Ratio Lower Upper
164 100
166 119.2 105.5 158.3
129 97.5 77.4 116.2
131 95.3 77.3 115.9









#65
Chlorobenzene
Concen: 5.038 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:112 Resp: 32304
Ion Ratio Lower Upper
112 100
114 35.4 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 5.312 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion:131 Resp: 11581
Ion Ratio Lower Upper
131 100
133 92.2 47.4 142.3
119 70.7 33.1 99.2





#67

Ethyl Benzene

Concen: 4.701 ug/l

RT: 11.947 min Scan# 1699

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 4962

Ion Ratio Lower Upper

91 100





106 34.6 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#68

m/p-Xylenes

Concen: 9.332 ug/l

RT: 12.059 min Scan# 1718

Delta R.T. 0.006 min

Lab File: VN087329.D




Acq: 16 Jul 2025 17:27


Tgt Ion:106 Resp: 36883

Ion Ratio Lower Upper

106 100

91 205.1 162.0 243.0





#69

o-Xylene

Concen: 4.580 ug/l

RT: 12.376 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 1729

Ion Ratio Lower Upper

106 100





91 219.7 107.7 323.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#70

Styrene

Concen: 4.639 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion:104 Resp: 29462


Ion Ratio Lower Upper

104 100

78 50.2 41.0 61.6


103 55.1 43.9 65.9





#71
Bromoform
Concen: 4.899 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 8628
Ion Ratio Lower Upper
173 100
175 45.5 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27



Tgt Ion:152 Resp: 146494
Ion Ratio Lower Upper
152 100
115 60.1 31.1 93.5
150 152.1 0.0 349.0









#73
Isopropylbenzene
Concen: 4.590 ug/l
RT: 12.676 min Scan# 1818
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:105 Resp: 42320
Ion Ratio Lower Upper
105 100
120 27.7 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#74
N-ethyl acetate
Concen: 5.113 ug/l m
RT: 12.647 min Scan# 1818
Delta R.T. 0.153 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27


Tgt Ion: 43 Resp: 18405
Ion Ratio Lower Upper
43 100
70 0.0 37.6 56.4#
55 0.0 19.6 29.4#
61 0.0 20.6 31.0#





#75
1,1,2-Tetrachloroethane
Concen: 4.988 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005




Tgt Ion: 83 Resp: 1730
Ion Ratio Lower Upper
83 100
131 5.3 5.1 15.3
85 67.4 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#76
1,2,3-Trichloropropane
Concen: 5.580 ug/l m
RT: 12.976 min Scan# 1874
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27



Tgt Ion: 75 Resp: 18135
Ion Ratio Lower Upper
75 100
77 170.2 94.5 283.6









#77
Bromobenzene
Concen: 4.847 ug/l
RT: 12.965 min Scan# 1158
Delta R.T. 0.006 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion:156 Resp: 1158
Ion Ratio Lower Upper
156 100
77 266.3 114.9 344.6
158 94.4 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 4.693 ug/l
RT: 13.018 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 91 Resp: 54443
Ion Ratio Lower Upper
91 100
120 22.2 11.3 33.8









#79
2-Chlorotoluene
Concen: 4.825 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 91 Resp: 34400
Ion Ratio Lower Upper
91 100
126 33.2 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#80
1,3,5-Trimethylbenzene
Concen: 4.694 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion:105 Resp: 36874
Ion Ratio Lower Upper
105 100
120 48.3 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 3.994 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion: 75 Resp: 479
Ion Ratio Lower Upper
75 100
53 83.4 83.5 125.3
89 34.7 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 4.854 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion: 91 Resp: 36031
Ion Ratio Lower Upper
91 100
126 34.5 16.6 49.7









#83
tert-Butylbenzene
Concen: 4.617 ug/l
RT: 13.418 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion:119 Resp: 3029
Ion Ratio Lower Upper
119 100
91 67.0 35.0 105.1
134 27.7 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 4.403 ug/l
RT: 13.465 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion:105 Resp: 35326
Ion Ratio Lower Upper
105 100
120 45.9 22.8 68.3





#85

sec-Butylbenzene

Concen: 4.649 ug/l

RT: 13.594 min Scan# 1979

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:105 Resp: 4594

Ion Ratio Lower Upper

105 100





134 18.9 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#86

p-Isopropyltoluene

Concen: 4.471 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27




Tgt Ion:119 Resp: 35413


Ion Ratio Lower Upper

119 100

134 27.6 13.5 40.5

91 26.8 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 4.970 ug/l

RT: 13.718 min Scan# 20

Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:146	Resp:	23320
Ion Ratio	Lower	Upper
146	100	
111	38.8	21.4 64.3
148	66.1	31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 4.994 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:146	Resp:	25031
Ion Ratio	Lower	Upper
146	100	
111	48.2	19.6 58.7
148	68.2	31.4 94.0





#89

n-Butylbenzene

Concen: 4.697 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	91	Resp:	3552
Ion Ratio	Lower	Upper	
91	100		
92	48.5	26.2	78.6
134	23.0	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 4.923 ug/l




RT: 14.312 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	117	Resp:	8261
Ion Ratio	Lower	Upper	
117	100		
201	67.1	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 5.055 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion:	146	Resp:	2247
Ion Ratio	Lower	Upper	
146	100		
111	43.7	22.0	66.0
148	67.1	32.7	98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 5.223 ug/l




RT: 14.700 min Scan# 2167


Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	75	Resp:	4757
Ion Ratio	Lower	Upper	
75	100		
155	63.8	37.3	111.8
157	79.5	46.2	138.6





#93

1,2,4-Trichlorobenzene

Concen: 4.823 ug/l

RT: 15.364 min Scan# 21

Delta R.T. -0.006 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	180	Resp:	1259
Ion Ratio	Lower	Upper	
180	100		
182	93.5	47.3	141.8
145	35.6	15.4	46.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 5.087 ug/l




RT: 15.476 min Scan# 2299


Delta R.T. 0.000 min

Lab File: VN087329.D

Acq: 16 Jul 2025 17:27

Tgt Ion:	225	Resp:	4936
Ion Ratio	Lower	Upper	
225	100		
223	60.9	32.1	96.3
227	63.0	31.3	93.9









#95
Naphthalene
Concen: 4.208 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005

Tgt Ion:128 Resp: 38930
Ion Ratio Lower Upper
128 100
127 13.0 10.5 15.7
129 11.1 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 4.721 ug/l
RT: 15.812 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087329.D
Acq: 16 Jul 2025 17:27

Tgt Ion:180 Resp: 12368
Ion Ratio Lower Upper
180 100
182 92.4 47.1 141.4
145 33.5 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087330.D
 Acq On : 16 Jul 2025 17:49
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025
 Supervised By : Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	178514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	328159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	297202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	156249	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.565	65	58584	19.341	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	38.680%#
35) Dibromofluoromethane	8.153	113	44726	19.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	39.520%#
50) Toluene-d8	10.547	98	154381	19.119	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	38.240%#
62) 4-Bromofluorobenzene	12.829	95	56832	19.051	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	38.100%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	31604	16.669	ug/l	86
3) Chloromethane	2.383	50	41982	17.608	ug/l	94
4) Vinyl Chloride	2.542	62	44508	18.784	ug/l	99
5) Bromomethane	2.971	94	22025	17.950	ug/l	98
6) Chloroethane	3.130	64	30785	19.922	ug/l	100
7) Trichlorofluoromethane	3.506	101	68764	19.626	ug/l	97
8) Diethyl Ether	3.959	74	28520	20.984	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	38516	21.414	ug/l	95
10) Methyl Iodide	4.577	142	25698	17.955	ug/l	97
11) Tert butyl alcohol	5.530	59	55298	96.147	ug/l	99
12) 1,1-Dichloroethene	4.330	96	39480	19.370	ug/l	96
13) Acrolein	4.177	56	40135	86.954	ug/l	98
14) Allyl chloride	5.006	41	80983	21.955	ug/l	88
15) Acrylonitrile	5.712	53	162325	104.007	ug/l	99
16) Acetone	4.424	43	140161	97.607	ug/l	93
17) Carbon Disulfide	4.700	76	119191	19.725	ug/l #	93
18) Methyl Acetate	5.012	43	70735	19.824	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	150385	20.018	ug/l	97
20) Methylene Chloride	5.265	84	49951	20.427	ug/l	87
21) trans-1,2-Dichloroethene	5.771	96	45917	19.980	ug/l	90
22) Diisopropyl ether	6.659	45	168043	21.719	ug/l	97
23) Vinyl Acetate	6.594	43	750202	110.863	ug/l	98
24) 1,1-Dichloroethane	6.559	63	89552	20.062	ug/l	95
25) 2-Butanone	7.477	43	231983	105.718	ug/l	96
26) 2,2-Dichloropropane	7.483	77	71478	20.596	ug/l	97
27) cis-1,2-Dichloroethene	7.471	96	53347	20.162	ug/l	98
28) Bromochloromethane	7.794	49	41285	19.325	ug/l	100
29) Tetrahydrofuran	7.830	42	147436	103.426	ug/l	99
30) Chloroform	7.953	83	93018	20.819	ug/l	100
31) Cyclohexane	8.241	56	74161	19.915	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	78235	20.217	ug/l	97
36) 1,1-Dichloropropene	8.359	75	60302	20.163	ug/l	98
37) Ethyl Acetate	7.553	43	85921	19.893	ug/l	98
38) Carbon Tetrachloride	8.347	117	65336	19.832	ug/l	95
39) Methylcyclohexane	9.588	83	63264	19.539	ug/l	95
40) Benzene	8.588	78	197187	20.400	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087330.D
 Acq On : 16 Jul 2025 17:49
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	47550	21.054	ug/l	97
42) 1,2-Dichloroethane	8.659	62	74680	20.374	ug/l	99
43) Isopropyl Acetate	8.677	43	135010	20.136	ug/l	99
44) Trichloroethene	9.335	130	44202	19.354	ug/l	98
45) 1,2-Dichloropropane	9.606	63	51786	21.086	ug/l	97
46) Dibromomethane	9.688	93	36673	19.943	ug/l	96
47) Bromodichloromethane	9.871	83	73409	19.819	ug/l	93
48) Methyl methacrylate	9.665	41	62316	20.645	ug/l	99
49) 1,4-Dioxane	9.682	88	18509	400.357	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	449710	106.046	ug/l	99
52) Toluene	10.612	92	123391	21.002	ug/l	98
53) t-1,3-Dichloropropene	10.818	75	76959	20.530	ug/l	92
54) cis-1,3-Dichloropropene	10.294	75	79019	20.407	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	47960	20.164	ug/l	94
56) Ethyl methacrylate	10.859	69	74408	19.384	ug/l	99
57) 1,3-Dichloropropane	11.147	76	85537	20.800	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	198704	101.839	ug/l	100
59) 2-Hexanone	11.182	43	305342	108.526	ug/l	99
60) Dibromochloromethane	11.341	129	55783	20.564	ug/l	98
61) 1,2-Dibromoethane	11.447	107	51292	20.509	ug/l	97
64) Tetrachloroethene	11.082	164	37655	19.686	ug/l	93
65) Chlorobenzene	11.871	112	134668	20.183	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.941	131	46671	20.570	ug/l	98
67) Ethyl Benzene	11.947	91	223715	20.367	ug/l	100
68) m/p-Xylenes	12.053	106	170590	41.473	ug/l	98
69) o-Xylene	12.376	106	83415	21.230	ug/l	98
70) Styrene	12.388	104	140937	21.323	ug/l	100
71) Bromoform	12.559	173	37362	20.383	ug/l #	99
73) Isopropylbenzene	12.676	105	202595	20.601	ug/l	100
74) N-amyl acetate	12.512	43	78074m	20.334	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	76720	20.733	ug/l	98
76) 1,2,3-Trichloropropane	12.976	75	68839m	19.858	ug/l	
77) Bromobenzene	12.959	156	53881	21.127	ug/l	97
78) n-propylbenzene	13.017	91	255577	20.656	ug/l	99
79) 2-Chlorotoluene	13.106	91	158436	20.836	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	176006	21.006	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	27975	21.846	ug/l	90
82) 4-Chlorotoluene	13.200	91	163615	20.667	ug/l	99
83) tert-Butylbenzene	13.417	119	143954	20.571	ug/l	98
84) 1,2,4-Trimethylbenzene	13.465	105	181623	21.226	ug/l	100
85) sec-Butylbenzene	13.594	105	216256	20.516	ug/l	99
86) p-Isopropyltoluene	13.706	119	175352	20.758	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	103157	20.609	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	108898	20.370	ug/l	98
89) n-Butylbenzene	14.035	91	165484	20.516	ug/l	99
90) Hexachloroethane	14.312	117	35108	19.615	ug/l	96
91) 1,2-Dichlorobenzene	14.088	146	99750	21.036	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.700	75	19015	19.573	ug/l	98
93) 1,2,4-Trichlorobenzene	15.370	180	54714	19.643	ug/l	99
94) Hexachlorobutadiene	15.476	225	20248	19.563	ug/l	96
95) Naphthalene	15.617	128	194574	19.718	ug/l	100
96) 1,2,3-Trichlorobenzene	15.811	180	55463	19.850	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087330.D
Acq On : 16 Jul 2025 17:49
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:168 Resp: 178514
Ion Ratio Lower Upper
168 100
99 61.7 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 16.669 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 85 Resp: 31604
Ion Ratio Lower Upper
85 100
87 27.2 17.8 53.3









#3
Chloromethane
Concen: 17.608 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 4198
Ion Ratio Lower Upper
50 100
52 29.3 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 18.784 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 62 Resp: 44508
Ion Ratio Lower Upper
62 100
64 34.1 27.0 40.6









#5
Bromomethane
Concen: 17.950 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 94 Resp: 22025
Ion Ratio Lower Upper
94 100
96 93.6 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#6
Chloroethane
Concen: 19.922 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 64 Resp: 30785
Ion Ratio Lower Upper
64 100
66 30.6 24.6 36.8









#7
Trichlorofluoromethane
Concen: 19.626 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 68764
Ion Ratio Lower Upper
101 100
103 64.4 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 20.984 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 74 Resp: 28520
Ion Ratio Lower Upper
74 100
45 98.3 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.414 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 38510
Ion Ratio Lower Upper
101 100
85 44.0 37.3 55.9
151 69.3 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#10
Methyl Iodide
Concen: 17.955 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion:142 Resp: 25698
Ion Ratio Lower Upper
142 100
127 46.4 35.7 53.5
141 13.7 10.4 15.6





#11

Tert butyl alcohol

Concen: 96.147 ug/l

RT: 5.530 min Scan# 607

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 59 Resp: 55298

Ion Ratio Lower Upper

59 100





57 10.1 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#12

1,1-Dichloroethene

Concen: 19.370 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 96 Resp: 39480


Ion Ratio Lower Upper

96 100

61 170.8 132.3 198.5

98 55.6 46.8 70.2





#13

Acrolein

Concen: 86.954 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 4013

Ion Ratio Lower Upper

56 100





55 72.2 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 21.955 ug/l

RT: 5.006 min Scan# 519

Delta R.T. -0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 41 Resp: 80983


Ion Ratio Lower Upper

41 100

39 62.3 59.0 88.6

76 31.0 28.7 43.1





#15

Acrylonitrile

Concen: 104.007 ug/l

RT: 5.712 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 53 Resp: 16232

Ion Ratio Lower Upper

53 100

52 81.8 65.5 98.3





51 36.7 28.7 43.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#16

Acetone

Concen: 97.607 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion: 43 Resp: 140161

Ion Ratio Lower Upper

43 100

58 31.8 22.3 33.5





#17

Carbon Disulfide

Concen: 19.725 ug/l

RT: 4.700 min Scan# 467

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 11919

Ion Ratio Lower Upper

76 100





78 8.2 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#18

Methyl Acetate

Concen: 19.824 ug/l

RT: 5.012 min Scan# 520

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion: 43 Resp: 70735

Ion Ratio Lower Upper

43 100

74 21.6 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 20.018 ug/l

RT: 5.794 min Scan# 653

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 73 Resp: 15038

Ion Ratio Lower Upper

73 100





57 21.5 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 20.427 ug/l

RT: 5.265 min Scan# 563

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion: 84 Resp: 49951




Ion Ratio Lower Upper


84 100

49 112.6 107.5 161.3

51 33.4 30.2 45.2

86 57.5 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 19.980 ug/l
RT: 5.771 min Scan# 649
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 96 Resp: 4591
Ion Ratio Lower Upper
96 100
61 135.4 122.0 183.0
98 62.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22
Diisopropyl ether
Concen: 21.719 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 45 Resp: 168043
Ion Ratio Lower Upper
45 100
43 56.8 42.8 64.2
87 24.3 19.8 29.6
59 11.5 8.3 12.5





#23

Vinyl Acetate

Concen: 110.863 ug/l

RT: 6.594 min Scan# 71

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 75020

Ion Ratio Lower Upper

43 100





86 10.4 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#24

1,1-Dichloroethane

Concen: 20.062 ug/l

RT: 6.559 min Scan# 783

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 63 Resp: 89552


Ion Ratio Lower Upper

63 100

98 8.6 3.3 9.9

100 3.4 2.5 7.4









#25
2-Butanone
Concen: 105.718 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 23198
Ion Ratio Lower Upper
43 100
72 26.6 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#26
2,2-Dichloropropane
Concen: 20.596 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.012 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 77 Resp: 71478
Ion Ratio Lower Upper
77 100
97 20.5 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 20.162 ug/l

RT: 7.471 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC020

Tgt Ion:	96	Resp:	5334
Ion	Ratio	Lower	Upper
96	100		
61	151.5	0.0	297.8
98	67.4	0.0	132.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 19.325 ug/l




RT: 7.794 min Scan# 993


Delta R.T. -0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	49	Resp:	41285
Ion	Ratio	Lower	Upper
49	100		
129	2.7	0.0	4.2
130	71.2	57.3	85.9





#29

Tetrahydrofuran

Concen: 103.426 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 42 Resp: 147430

Ion Ratio Lower Upper

42 100

72 43.1 33.4 50.0





71 39.0 31.2 46.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 20.819 ug/l

RT: 7.953 min Scan# 1020

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion: 83 Resp: 93018

Ion Ratio Lower Upper

83 100

85 66.0 52.7 79.1





#31

Cyclohexane

Concen: 19.915 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 7416

Ion Ratio Lower Upper

56 100

69 29.6 22.7 34.1





84 83.3 65.8 98.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#32

1,1,1-Trichloroethane

Concen: 20.217 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 97 Resp: 78235


Ion Ratio Lower Upper

97 100

99 62.3 51.8 77.8

61 49.8 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 19.341 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 65 Resp: 58584

Ion Ratio Lower Upper

65 100





67 51.7 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion:114 Resp: 328159


Ion Ratio Lower Upper

114 100

63 21.5 0.0 44.6

88 14.5 0.0 32.8





#35

Dibromofluoromethane

Concen: 19.758 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:113 Resp: 44720

Ion Ratio Lower Upper

113 100

111 97.7 82.5 123.7





192 17.8 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#36

1,1-Dichloropropene

Concen: 20.163 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 75 Resp: 60302


Ion Ratio Lower Upper

75 100

110 34.9 16.7 50.1

77 31.9 25.2 37.8





#37

Ethyl Acetate

Concen: 19.893 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC020

Tgt Ion:	43	Resp:	8592
Ion	Ratio	Lower	Upper
43	100		
61	14.0	10.9	16.3
70	10.1	7.4	11.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 19.832 ug/l




RT: 8.347 min Scan# 1087


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	117	Resp:	65336
Ion	Ratio	Lower	Upper
117	100		
119	95.3	80.2	120.2
121	34.1	25.4	38.2





#39

Methylcyclohexane

Concen: 19.539 ug/l

RT: 9.588 min Scan# 1128

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49





Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 83	Resp: 63264
Ion Ratio	Lower Upper
83	100
55	71.5 61.3 91.9
98	50.0 37.9 56.9

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#40

Benzene

Concen: 20.400 ug/l




RT: 8.588 min Scan# 1128


Delta R.T. -0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion: 78	Resp: 197187
Ion Ratio	Lower Upper
78	100
77	23.9 18.2 27.2









#41
Methacrylonitrile
Concen: 21.054 ug/l
RT: 7.771 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
Client Sample Id :
VSTDICC020

Tgt Ion: 41 Resp: 47550
Ion Ratio Lower Upper
41 100
39 51.7 43.4 65.0
67 67.2 55.1 82.7
52 30.1 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#42
1,2-Dichloroethane
Concen: 20.374 ug/l
RT: 8.659 min Scan# 1140
Delta R.T. 0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 62 Resp: 74680
Ion Ratio Lower Upper
62 100
98 8.7 0.0 18.0





#43

Isopropyl Acetate

Concen: 20.136 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 135010

Ion Ratio Lower Upper

43 100

61 25.0 19.8 29.8





87 13.2 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene

Concen: 19.354 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion:130 Resp: 44202

Ion Ratio Lower Upper

130 100

95 95.4 0.0 195.2





#45

1,2-Dichloropropane

Concen: 21.086 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 5178

Ion Ratio Lower Upper

63 100





65 32.3 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#46

Dibromomethane

Concen: 19.943 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 93 Resp: 36673


Ion Ratio Lower Upper

93 100

95 83.3 65.8 98.8

174 93.5 69.9 104.9





#47

Bromodichloromethane

Concen: 19.819 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	83	Resp:	73409
Ion	Ratio	Lower	Upper
83	100		
85	69.3	50.4	75.6
127	8.0	5.8	8.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#48

Methyl methacrylate

Concen: 20.645 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	41	Resp:	62316
Ion	Ratio	Lower	Upper
41	100		
69	78.3	64.1	96.1
39	57.4	45.5	68.3





#49

1,4-Dioxane

Concen: 400.357 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 88 Resp: 18509

Ion Ratio Lower Upper

88 100

43 28.0 0.0 0.0





58 76.8 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 19.119 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion: 98 Resp: 154381

Ion Ratio Lower Upper

98 100


100 66.1 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 106.046 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 449710
Ion Ratio Lower Upper
43 100
58 37.6 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#52
Toluene
Concen: 21.002 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 92 Resp: 123391
Ion Ratio Lower Upper
92 100
91 166.7 135.1 202.7









#53
t-1,3-Dichloropropene
Concen: 20.530 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 76959
Ion Ratio Lower Upper
75 100
77 33.6 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 20.407 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 75 Resp: 79019
Ion Ratio Lower Upper
75 100
77 33.3 24.2 36.2
39 50.3 38.4 57.6





#55

1,1,2-Trichloroethane

Concen: 20.164 ug/l

RT: 11.000 min Scan# 1538

Delta R.T. -0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 97 Resp: 47960

Ion Ratio Lower Upper

97 100

83 83.6 72.6 109.0

85 53.1 46.7 70.1





99 64.0 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#56

Ethyl methacrylate

Concen: 19.384 ug/l

RT: 10.859 min Scan# 1514

Delta R.T. -0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion: 69 Resp: 74408


Ion Ratio Lower Upper

69 100

41 68.1 55.1 82.7

39 36.2 27.9 41.9





#57

1,3-Dichloropropane

Concen: 20.800 ug/l

RT: 11.147 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 8553

Ion Ratio Lower Upper

76 100





78 31.7 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 101.839 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion: 63 Resp: 198704

Ion Ratio Lower Upper

63 100

106 27.1 21.7 32.5





#59

2-Hexanone

Concen: 108.526 ug/l

RT: 11.182 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 30534

Ion Ratio Lower Upper

43 100





58 52.8 26.7 80.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#60

Dibromochloromethane

Concen: 20.564 ug/l

RT: 11.341 min Scan# 1596

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion:129 Resp: 55783

Ion Ratio Lower Upper

129 100

127 76.4 39.1 117.5









#61
1,2-Dibromoethane
Concen: 20.509 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:107 Resp: 51292
Ion Ratio Lower Upper
107 100
109 94.2 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#62
4-Bromofluorobenzene
Concen: 19.051 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 95 Resp: 56832
Ion Ratio Lower Upper
95 100
174 75.9 0.0 149.4
176 70.0 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:117 Resp: 29720
Ion Ratio Lower Upper
117 100
82 58.0 47.4 71.2
119 32.7 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#64
Tetrachloroethene
Concen: 19.686 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49


Tgt Ion:164 Resp: 37655
Ion Ratio Lower Upper
164 100
166 127.2 105.5 158.3
129 112.0 77.4 116.2
131 94.0 77.3 115.9





#65
Chlorobenzene
Concen: 20.183 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion:112 Resp: 134668
Ion Ratio Lower Upper
112 100
114 30.4 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.570 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion:131 Resp: 46671
Ion Ratio Lower Upper
131 100
133 94.6 47.4 142.3
119 68.8 33.1 99.2





#67

Ethyl Benzene

Concen: 20.367 ug/l

RT: 11.947 min Scan# 1699

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 22371

Ion Ratio Lower Upper

91 100





106 30.4 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#68

m/p-Xylenes

Concen: 41.473 ug/l

RT: 12.053 min Scan# 1717

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion:106 Resp: 170590

Ion Ratio Lower Upper

106 100

91 205.1 162.0 243.0





#69

o-Xylene

Concen: 21.230 ug/l

RT: 12.376 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:106 Resp: 8341

Ion Ratio Lower Upper

106 100





91 212.7 107.7 323.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#70

Styrene

Concen: 21.323 ug/l

RT: 12.388 min Scan# 1774

Delta R.T. -0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion:104 Resp: 140937


Ion Ratio Lower Upper

104 100

78 51.6 41.0 61.6


103 54.8 43.9 65.9





#71
Bromoform
Concen: 20.383 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:173 Resp: 37362
Ion Ratio Lower Upper
173 100
175 47.3 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49



Tgt Ion:152 Resp: 156249
Ion Ratio Lower Upper
152 100
115 61.9 31.1 93.5
150 162.0 0.0 349.0









#73
Isopropylbenzene
Concen: 20.601 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 105 Resp: 20259
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#74
N-ethyl acetate
Concen: 20.334 ug/l m
RT: 12.512 min Scan# 1795
Delta R.T. 0.018 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49


Tgt Ion: 43 Resp: 78074
Ion Ratio Lower Upper
43 100
70 17.4 37.6 56.4#
55 12.3 19.6 29.4#
61 7.9 20.6 31.0#





#75
1,1,2,2-Tetrachloroethane
Concen: 20.733 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion: 83 Resp: 76720
Ion Ratio Lower Upper
83 100
131 9.4 5.1 15.3
85 63.6 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#76
1,2,3-Trichloropropane
Concen: 19.858 ug/l m
RT: 12.976 min Scan# 1874
Delta R.T. 0.006 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49



Tgt Ion: 75 Resp: 68839
Ion Ratio Lower Upper
75 100
77 176.6 94.5 283.6









#77
Bromobenzene
Concen: 21.127 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:156 Resp: 5388
Ion Ratio Lower Upper
156 100
77 225.6 114.9 344.6
158 93.0 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 20.656 ug/l
RT: 13.017 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion: 91 Resp: 255577
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8





#79

2-Chlorotoluene

Concen: 20.836 ug/l

RT: 13.106 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 158430

Ion Ratio Lower Upper

91 100





126 31.5 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#80

1,3,5-Trimethylbenzene

Concen: 21.006 ug/l

RT: 13.153 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VN087330.D




Acq: 16 Jul 2025 17:49


Tgt Ion:105 Resp: 176006

Ion Ratio Lower Upper

105 100

120 48.1 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 21.846 ug/l
RT: 12.717 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 75 Resp: 27975
Ion Ratio Lower Upper
75 100
53 93.5 83.5 125.3
89 42.7 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 20.667 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49


Tgt Ion: 91 Resp: 163615
Ion Ratio Lower Upper
91 100
126 33.7 16.6 49.7





#83
tert-Butylbenzene
Concen: 20.571 ug/l
RT: 13.417 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020




Tgt Ion:119 Resp: 143954
Ion Ratio Lower Upper
119 100
91 70.2 35.0 105.1
134 27.9 12.6 37.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#84
1,2,4-Trimethylbenzene
Concen: 21.226 ug/l
RT: 13.465 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49



Tgt Ion:105 Resp: 181623
Ion Ratio Lower Upper
105 100
120 45.6 22.8 68.3





#85

sec-Butylbenzene

Concen: 20.516 ug/l

RT: 13.594 min Scan# 1979

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion:105 Resp: 216250

Ion Ratio Lower Upper

105 100





134 19.9 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#86

p-Isopropyltoluene

Concen: 20.758 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49




Tgt Ion:119 Resp: 175352


Ion Ratio Lower Upper

119 100

134 26.2 13.5 40.5

91 25.0 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 20.609 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:	146	Resp:	10315
Ion Ratio	Lower	Upper	
146	100		
111	42.3	21.4	64.3
148	63.9	31.6	95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 20.370 ug/l




RT: 13.788 min Scan# 1212


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	146	Resp:	108898
Ion Ratio	Lower	Upper	
146	100		
111	40.3	19.6	58.7
148	63.9	31.4	94.0





#89

n-Butylbenzene

Concen: 20.516 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC020

Tgt Ion:	91	Resp:	16548
Ion Ratio	Lower	Upper	
91	100		
92	51.8	26.2	78.6
134	24.5	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 19.615 ug/l




RT: 14.312 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	117	Resp:	35108
Ion Ratio	Lower	Upper	
117	100		
201	68.6	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 21.036 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:	146	Resp:	99750
Ion	Ratio	Lower	Upper
146	100		
111	42.6	22.0	66.0
148	62.0	32.7	98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.573 ug/l




RT: 14.700 min Scan# 2167


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	75	Resp:	19015
Ion	Ratio	Lower	Upper
75	100		
155	74.2	37.3	111.8
157	89.0	46.2	138.6





#93

1,2,4-Trichlorobenzene

Concen: 19.643 ug/l

RT: 15.370 min Scan# 2281

Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC020

Tgt Ion:	180	Resp:	54714
Ion Ratio	Lower	Upper	
180	100		
182	95.4	47.3	141.8
145	31.3	15.4	46.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 19.563 ug/l




RT: 15.476 min Scan# 2299


Delta R.T. 0.000 min

Lab File: VN087330.D

Acq: 16 Jul 2025 17:49

Tgt Ion:	225	Resp:	20248
Ion Ratio	Lower	Upper	
225	100		
223	62.9	32.1	96.3
227	68.0	31.3	93.9









#95
Naphthalene
Concen: 19.718 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:128 Resp: 194574
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 10.5 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.850 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087330.D
Acq: 16 Jul 2025 17:49

Tgt Ion:180 Resp: 55463
Ion Ratio Lower Upper
180 100
182 96.4 47.1 141.4
145 34.8 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025
 Supervised By : Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.212	168	182792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	325711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	300150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	158477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	146559	47.253	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.500%		
35) Dibromofluoromethane	8.147	113	108271	48.190	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.380%		
50) Toluene-d8	10.547	98	398776	49.757	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.520%		
62) 4-Bromofluorobenzene	12.829	95	148158	50.037	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.080%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	113853	58.643	ug/l	100
3) Chloromethane	2.383	50	126054	51.631	ug/l	100
4) Vinyl Chloride	2.542	62	132985	54.810	ug/l	100
5) Bromomethane	2.971	94	64982	51.719	ug/l	100
6) Chloroethane	3.130	64	80606	50.942	ug/l	100
7) Trichlorofluoromethane	3.506	101	187302	52.206	ug/l	100
8) Diethyl Ether	3.959	74	72151	51.843	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.359	101	95895	52.068	ug/l	100
10) Methyl Iodide	4.577	142	90277	47.912	ug/l	100
11) Tert butyl alcohol	5.524	59	146939	249.504	ug/l	100
12) 1,1-Dichloroethene	4.330	96	99569	47.708	ug/l	100
13) Acrolein	4.177	56	110711	234.247	ug/l	100
14) Allyl chloride	5.012	41	183089	48.475	ug/l	100
15) Acrylonitrile	5.712	53	414273	259.227	ug/l	100
16) Acetone	4.424	43	353373	240.328	ug/l	100
17) Carbon Disulfide	4.694	76	316833	51.205	ug/l	100
18) Methyl Acetate	5.012	43	181199	49.594	ug/l	100
19) Methyl tert-butyl Ether	5.788	73	396133	51.495	ug/l	100
20) Methylene Chloride	5.259	84	122921	50.001	ug/l	100
21) trans-1,2-Dichloroethene	5.771	96	119278	50.687	ug/l	100
22) Diisopropyl ether	6.665	45	417843	52.740	ug/l	100
23) Vinyl Acetate	6.588	43	1931809	278.796	ug/l	100
24) 1,1-Dichloroethane	6.553	63	223298	48.853	ug/l	100
25) 2-Butanone	7.471	43	588032	261.702	ug/l	100
26) 2,2-Dichloropropane	7.471	77	179431	50.492	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	140194	51.746	ug/l	100
28) Bromochloromethane	7.800	49	108748	49.713	ug/l	100
29) Tetrahydrofuran	7.829	42	388704	266.294	ug/l	100
30) Chloroform	7.953	83	233747	51.092	ug/l	100
31) Cyclohexane	8.241	56	191116	50.122	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	198135	50.002	ug/l	100
36) 1,1-Dichloropropene	8.353	75	158563	53.418	ug/l	100
37) Ethyl Acetate	7.547	43	230005	53.652	ug/l	100
38) Carbon Tetrachloride	8.347	117	168403	51.501	ug/l	100
39) Methylcyclohexane	9.582	83	172297	53.614	ug/l	100
40) Benzene	8.594	78	505723	52.714	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	119895	53.486	ug/l	100
42) 1,2-Dichloroethane	8.653	62	184670	50.759	ug/l	100
43) Isopropyl Acetate	8.677	43	348709	52.399	ug/l	100
44) Trichloroethene	9.335	130	115974	51.160	ug/l	100
45) 1,2-Dichloropropane	9.606	63	128530	52.726	ug/l	100
46) Dibromomethane	9.688	93	94179	51.601	ug/l	100
47) Bromodichloromethane	9.871	83	185222	50.382	ug/l	100
48) Methyl methacrylate	9.665	41	164476	54.899	ug/l	100
49) 1,4-Dioxane	9.682	88	52072	1134.802	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	1122857	266.771	ug/l	100
52) Toluene	10.612	92	313806	53.814	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	202129	54.327	ug/l	100
54) cis-1,3-Dichloropropene	10.294	75	205720	53.529	ug/l	100
55) 1,1,2-Trichloroethane	11.000	97	119612	50.666	ug/l	100
56) Ethyl methacrylate	10.859	69	204376	51.427	ug/l	100
57) 1,3-Dichloropropane	11.141	76	213343	52.267	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	562790	290.607	ug/l	100
59) 2-Hexanone	11.182	43	806824	288.920	ug/l	100
60) Dibromochloromethane	11.341	129	139955	51.982	ug/l	100
61) 1,2-Dibromoethane	11.453	107	125416	50.525	ug/l	100
64) Tetrachloroethene	11.088	164	96176	49.786	ug/l	100
65) Chlorobenzene	11.870	112	335833	49.837	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.941	131	117833	51.425	ug/l	100
67) Ethyl Benzene	11.947	91	582851	52.540	ug/l	100
68) m/p-Xylenes	12.053	106	455017	109.536	ug/l	100
69) o-Xylene	12.376	106	217121	54.717	ug/l	100
70) Styrene	12.394	104	376745	56.440	ug/l	100
71) Bromoform	12.559	173	98549	53.237	ug/l #	100
73) Isopropylbenzene	12.676	105	538245	53.964	ug/l	100
74) N-amyl acetate	12.494	43	180343m	46.310	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	191273	50.964	ug/l	100
76) 1,2,3-Trichloropropane	12.970	75	188824m	53.705	ug/l	
77) Bromobenzene	12.959	156	137669	53.221	ug/l	100
78) n-propylbenzene	13.017	91	680506	54.227	ug/l	100
79) 2-Chlorotoluene	13.106	91	401444	52.051	ug/l	100
80) 1,3,5-Trimethylbenzene	13.153	105	464063	54.607	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.717	75	63993	49.271	ug/l	100
82) 4-Chlorotoluene	13.200	91	419474	52.240	ug/l	100
83) tert-Butylbenzene	13.417	119	383986	54.099	ug/l	100
84) 1,2,4-Trimethylbenzene	13.464	105	475982	54.846	ug/l	100
85) sec-Butylbenzene	13.594	105	565038	52.851	ug/l	100
86) p-Isopropyltoluene	13.706	119	473942	55.316	ug/l	100
87) 1,3-Dichlorobenzene	13.711	146	262675	51.740	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	269881	49.773	ug/l	100
89) n-Butylbenzene	14.035	91	437514	53.478	ug/l	100
90) Hexachloroethane	14.311	117	91313	50.301	ug/l	100
91) 1,2-Dichlorobenzene	14.082	146	249988	51.977	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.700	75	47931	48.643	ug/l	100
93) 1,2,4-Trichlorobenzene	15.370	180	148472	52.553	ug/l	100
94) Hexachlorobutadiene	15.476	225	52130	49.659	ug/l	100
95) Naphthalene	15.617	128	545367	54.490	ug/l	100
96) 1,2,3-Trichlorobenzene	15.811	180	146080	51.547	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087331.D
Acq On : 16 Jul 2025 18:11
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

Client Sample Id :

VSTDICCC050

Manual Integrations

APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025

Supervised By : Semsettin Yesilyurt 07/17/2025


Quant Time: Jul 17 02:19:33 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025


Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
Client Sample Id :
VSTDICCC050



Tgt Ion: 168 Resp: 182792
Ion Ratio Lower Upper
168 100
99 59.9 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 58.643 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 85 Resp: 113853
Ion Ratio Lower Upper
85 100
87 35.5 17.8 53.3









#3
Chloromethane
Concen: 51.631 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 126054
Ion Ratio Lower Upper
50 100
52 32.9 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 54.810 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11


Tgt Ion: 62 Resp: 132985
Ion Ratio Lower Upper
62 100
64 33.8 27.0 40.6





#5
Bromomethane
Concen: 51.719 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 6498
Ion Ratio Lower Upper
94 100
96 91.8 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#6
Chloroethane
Concen: 50.942 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11


Tgt Ion: 64 Resp: 80606
Ion Ratio Lower Upper
64 100
66 30.7 24.6 36.8





#7
Trichlorofluoromethane
Concen: 52.206 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 101 Resp: 187302
Ion Ratio Lower Upper
101 100
103 62.2 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 51.843 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 74 Resp: 72151
Ion Ratio Lower Upper
74 100
45 101.7 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.068 ug/l
RT: 4.359 min Scan# 409
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 9589
Ion Ratio Lower Upper
101 100
85 46.6 37.3 55.9
151 73.6 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#10
Methyl Iodide
Concen: 47.912 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:142 Resp: 90277
Ion Ratio Lower Upper
142 100
127 44.6 35.7 53.5
141 13.0 10.4 15.6





#11

Tert butyl alcohol

Concen: 249.504 ug/l

RT: 5.524 min Scan# 607

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 146939

Ion Ratio Lower Upper

59 100





57 10.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#12

1,1-Dichloroethene

Concen: 47.708 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 96 Resp: 99569


Ion Ratio Lower Upper

96 100

61 165.4 132.3 198.5

98 58.5 46.8 70.2





#13

Acrolein

Concen: 234.247 ug/l

RT: 4.177 min Scan# 378

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 11071

Ion Ratio Lower Upper

56 100





55 70.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 48.475 ug/l

RT: 5.012 min Scan# 520

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 41 Resp: 183089


Ion Ratio Lower Upper

41 100

39 73.8 59.0 88.6

76 35.9 28.7 43.1





#15

Acrylonitrile

Concen: 259.227 ug/l

RT: 5.712 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 53 Resp: 41427

Ion Ratio Lower Upper

53 100

52 81.9 65.5 98.3





51 35.9 28.7 43.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#16

Acetone

Concen: 240.328 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion: 43 Resp: 353373

Ion Ratio Lower Upper

43 100

58 27.9 22.3 33.5









#17
Carbon Disulfide
Concen: 51.205 ug/l
RT: 4.694 min Scan# 466
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 76 Resp: 31683
Ion Ratio Lower Upper
76 100
78 10.8 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#18
Methyl Acetate
Concen: 49.594 ug/l
RT: 5.012 min Scan# 520
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 43 Resp: 181199
Ion Ratio Lower Upper
43 100
74 22.2 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 51.495 ug/l

RT: 5.788 min Scan# 652

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 39613

Ion Ratio Lower Upper

73 100





57 23.2 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 50.001 ug/l

RT: 5.259 min Scan# 562

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion: 84 Resp: 122921




Ion Ratio Lower Upper


84 100

49 134.4 107.5 161.3

51 37.7 30.2 45.2

86 61.6 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 50.687 ug/l
RT: 5.771 min Scan# 649
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
Client Sample Id :
VSTDICCC050

Tgt Ion: 96 Resp: 119278
Ion Ratio Lower Upper
96 100
61 152.5 122.0 183.0
98 62.5 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22
Diisopropyl ether
Concen: 52.740 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 45 Resp: 417843
Ion Ratio Lower Upper
45 100
43 53.5 42.8 64.2
87 24.7 19.8 29.6
59 10.4 8.3 12.5









#23
Vinyl Acetate
Concen: 278.796 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 1931809
Ion Ratio Lower Upper
43 100
86 9.6 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#24
1,1-Dichloroethane
Concen: 48.853 ug/l
RT: 6.553 min Scan# 782
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 63 Resp: 223298
Ion Ratio Lower Upper
63 100
98 6.6 3.3 9.9
100 4.9 2.5 7.4





#25

2-Butanone

Concen: 261.702 ug/l

RT: 7.471 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 58803

Ion Ratio Lower Upper

43 100





72 24.5 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#26

2,2-Dichloropropane

Concen: 50.492 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion: 77 Resp: 179431

Ion Ratio Lower Upper

77 100

97 22.1 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 51.746 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	96	Resp:	140194
Ion	Ratio	Lower	Upper
96	100		
61	148.9	0.0	297.8
98	66.2	0.0	132.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 49.713 ug/l




RT: 7.800 min Scan# 994


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	49	Resp:	108748
Ion	Ratio	Lower	Upper
49	100		
129	2.1	0.0	4.2
130	71.6	57.3	85.9





#29

Tetrahydrofuran

Concen: 266.294 ug/l

RT: 7.829 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	42	Resp:	388704
Ion	Ratio	Lower	Upper
42	100		
72	41.7	33.4	50.0
71	39.0	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 51.092 ug/l




RT: 7.953 min Scan# 1020


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	83	Resp:	233747
Ion	Ratio	Lower	Upper
83	100		
85	65.9	52.7	79.1





#31

Cyclohexane

Concen: 50.122 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 191110

Ion Ratio Lower Upper

56 100

69 28.4 22.7 34.1





84 82.2 65.8 98.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#32

1,1,1-Trichloroethane

Concen: 50.002 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 97 Resp: 198135


Ion Ratio Lower Upper

97 100

99 64.8 51.8 77.8

61 48.4 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 47.253 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 14655

Ion Ratio Lower Upper

65 100





67 52.0 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion:114 Resp: 325711


Ion Ratio Lower Upper

114 100

63 22.3 0.0 44.6

88 16.4 0.0 32.8





#35

Dibromofluoromethane

Concen: 48.190 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 113 Resp: 10827

Ion Ratio Lower Upper

113 100

111 103.1 82.5 123.7





192 17.1 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#36

1,1-Dichloropropene

Concen: 53.418 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 75 Resp: 158563


Ion Ratio Lower Upper

75 100

110 33.4 16.7 50.1

77 31.5 25.2 37.8





#37

Ethyl Acetate

Concen: 53.652 ug/l

RT: 7.547 min Scan# 951

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	43	Resp:	23000
Ion	Ratio	Lower	Upper
43	100		
61	13.6	10.9	16.3
70	9.2	7.4	11.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 51.501 ug/l




RT: 8.347 min Scan# 1087


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	117	Resp:	168403
Ion	Ratio	Lower	Upper
117	100		
119	100.2	80.2	120.2
121	31.8	25.4	38.2





#39

Methylcyclohexane

Concen: 53.614 ug/l

RT: 9.582 min Scan# 1297

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 17229

Ion Ratio Lower Upper

83 100

55 76.6 61.3 91.9





98 47.4 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#40

Benzene

Concen: 52.714 ug/l

RT: 8.594 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion: 78 Resp: 505723

Ion Ratio Lower Upper

78 100


77 22.7 18.2 27.2





#41
Methacrylonitrile
Concen: 53.486 ug/l
RT: 7.771 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 41 Resp: 11989
Ion Ratio Lower Upper
41 100
39 54.2 43.4 65.0
67 68.9 55.1 82.7
52 31.1 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#42
1,2-Dichloroethane
Concen: 50.759 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 62 Resp: 184670
Ion Ratio Lower Upper
62 100
98 9.0 0.0 18.0





#43

Isopropyl Acetate

Concen: 52.399 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	43	Resp:	348709
Ion	Ratio	Lower	Upper
43	100		
61	24.8	19.8	29.8
87	12.2	9.8	14.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene

Concen: 51.160 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	130	Resp:	115974
Ion	Ratio	Lower	Upper
130	100		
95	97.6	0.0	195.2









#45
1,2-Dichloropropane
Concen: 52.726 ug/l
RT: 9.606 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 63 Resp: 128530
Ion Ratio Lower Upper
63 100
65 33.7 27.0 40.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#46
Dibromomethane
Concen: 51.601 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 93 Resp: 94179
Ion Ratio Lower Upper
93 100
95 82.3 65.8 98.8
174 87.4 69.9 104.9





#47

Bromodichloromethane

Concen: 50.382 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11





Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:	83	Resp:	18522
Ion	Ratio	Lower	Upper
83	100		
85	63.0	50.4	75.6
127	7.3	5.8	8.8

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#48

Methyl methacrylate

Concen: 54.899 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	41	Resp:	164476
Ion	Ratio	Lower	Upper
41	100		
69	80.1	64.1	96.1
39	56.9	45.5	68.3





#49

1,4-Dioxane

Concen: 1134.802 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 52072

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 76.4 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 49.757 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. -0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion: 98 Resp: 398776

Ion Ratio Lower Upper

98 100


100 65.1 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 266.771 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 112285
Ion Ratio Lower Upper
43 100
58 38.5 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#52
Toluene
Concen: 53.814 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 92 Resp: 313806
Ion Ratio Lower Upper
92 100
91 168.9 135.1 202.7









#53
t-1,3-Dichloropropene
Concen: 54.327 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 202129
Ion Ratio Lower Upper
75 100
77 29.5 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 53.529 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 75 Resp: 205720
Ion Ratio Lower Upper
75 100
77 30.2 24.2 36.2
39 48.0 38.4 57.6





#55

1,1,2-Trichloroethane

Concen: 50.666 ug/l

RT: 11.000 min Scan# 1538

Delta R.T. -0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 11961

Ion Ratio Lower Upper

97 100

83 90.8 72.6 109.0

85 58.4 46.7 70.1





99 67.6 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#56

Ethyl methacrylate

Concen: 51.427 ug/l

RT: 10.859 min Scan# 1514

Delta R.T. -0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 69 Resp: 204376


Ion Ratio Lower Upper

69 100

41 68.9 55.1 82.7

39 34.9 27.9 41.9





#57

1,3-Dichloropropane

Concen: 52.267 ug/l

RT: 11.141 min Scan# 1562

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 21334

Ion Ratio Lower Upper

76 100





78 32.5 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 290.607 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion: 63 Resp: 562790

Ion Ratio Lower Upper

63 100


106 27.1 21.7 32.5





#59
2-Hexanone
Concen: 288.920 ug/l
RT: 11.182 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 806824
Ion Ratio Lower Upper
43 100
58 53.3 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#60
Dibromochloromethane
Concen: 51.982 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:129 Resp: 139955
Ion Ratio Lower Upper
129 100
127 78.3 39.1 117.5









#61
1,2-Dibromoethane
Concen: 50.525 ug/l
RT: 11.453 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 125410
Ion Ratio Lower Upper
107 100
109 96.9 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#62
4-Bromofluorobenzene
Concen: 50.037 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 95 Resp: 148158
Ion Ratio Lower Upper
95 100
174 74.7 0.0 149.4
176 70.6 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:117 Resp: 300150
Ion Ratio Lower Upper
117 100
82 59.3 47.4 71.2
119 29.8 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#64
Tetrachloroethene
Concen: 49.786 ug/l
RT: 11.088 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:164 Resp: 96176
Ion Ratio Lower Upper
164 100
166 131.9 105.5 158.3
129 96.8 77.4 116.2
131 96.6 77.3 115.9









#65
Chlorobenzene
Concen: 49.837 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 335833
Ion Ratio Lower Upper
112 100
114 33.1 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 51.425 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:131 Resp: 117833
Ion Ratio Lower Upper
131 100
133 94.9 47.4 142.3
119 66.1 33.1 99.2





#67

Ethyl Benzene

Concen: 52.540 ug/l

RT: 11.947 min Scan# 1699

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 58285

Ion Ratio Lower Upper

91 100





106 30.4 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#68

m/p-Xylenes

Concen: 109.536 ug/l

RT: 12.053 min Scan# 1717

Delta R.T. 0.000 min

Lab File: VN087331.D




Acq: 16 Jul 2025 18:11


Tgt Ion:106 Resp: 455017

Ion Ratio Lower Upper

106 100

91 202.5 162.0 243.0









#69
o-Xylene
Concen: 54.717 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 21712
Ion Ratio Lower Upper
106 100
91 215.5 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#70
Styrene
Concen: 56.440 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11


Tgt Ion:104 Resp: 376745
Ion Ratio Lower Upper
104 100
78 51.3 41.0 61.6
103 54.9 43.9 65.9





#71
Bromoform
Concen: 53.237 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:173 Resp: 98549
Ion Ratio Lower Upper
173 100
175 48.2 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11



Tgt Ion:152 Resp: 158477
Ion Ratio Lower Upper
152 100
115 62.3 31.1 93.5
150 174.5 0.0 349.0









#73
Isopropylbenzene
Concen: 53.964 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 53824
Ion Ratio Lower Upper
105 100
120 26.7 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#74
N-ethyl acetate
Concen: 46.310 ug/l m
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 43 Resp: 180343
Ion Ratio Lower Upper
43 100
70 42.5 37.6 56.4
55 22.2 19.6 29.4
61 23.4 20.6 31.0









#75
1,1,2,2-Tetrachloroethane
Concen: 50.964 ug/l
RT: 12.917 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 19127
Ion Ratio Lower Upper
83 100
131 10.2 5.1 15.3
85 64.9 32.5 97.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#76
1,2,3-Trichloropropane
Concen: 53.705 ug/l m
RT: 12.970 min Scan# 1873
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11


Tgt Ion: 75 Resp: 188824
Ion Ratio Lower Upper
75 100
77 167.4 94.5 283.6





#77
Bromobenzene
Concen: 53.221 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:156 Resp: 137669
Ion Ratio Lower Upper
156 100
77 229.7 114.9 344.6
158 97.0 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 54.227 ug/l
RT: 13.017 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 91 Resp: 680506
Ion Ratio Lower Upper
91 100
120 22.5 11.3 33.8









#79
2-Chlorotoluene
Concen: 52.051 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 401444
Ion Ratio Lower Upper
91 100
126 33.7 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#80
1,3,5-Trimethylbenzene
Concen: 54.607 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:105 Resp: 464063
Ion Ratio Lower Upper
105 100
120 48.5 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 49.271 ug/l
RT: 12.717 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 6399
Ion Ratio Lower Upper
75 100
53 104.5 83.5 125.3
89 48.0 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 52.240 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion: 91 Resp: 419474
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.7









#83
tert-Butylbenzene
Concen: 54.099 ug/l
RT: 13.417 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:119 Resp: 383980
Ion Ratio Lower Upper
119 100
91 70.1 35.0 105.1
134 25.1 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 54.846 ug/l
RT: 13.464 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:105 Resp: 475982
Ion Ratio Lower Upper
105 100
120 45.5 22.8 68.3









#85
sec-Butylbenzene
Concen: 52.851 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 565038
Ion Ratio Lower Upper
105 100
134 19.6 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#86
p-Isopropyltoluene
Concen: 55.316 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11

Tgt Ion:119 Resp: 473942
Ion Ratio Lower Upper
119 100
134 27.0 13.5 40.5
91 24.4 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 51.740 ug/l

RT: 13.711 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 262679

Ion Ratio Lower Upper

146 100

111 42.9 21.4 64.3





148 63.3 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 49.773 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion:146 Resp: 269881


Ion Ratio Lower Upper

146 100

111 39.1 19.6 58.7

148 62.7 31.4 94.0





#89

n-Butylbenzene

Concen: 53.478 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	91	Resp:	437514
Ion	Ratio	Lower	Upper
91	100		
92	52.4	26.2	78.6
134	24.8	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 50.301 ug/l




RT: 14.311 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Tgt Ion:	117	Resp:	91313
Ion	Ratio	Lower	Upper
117	100		
201	65.6	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 51.977 ug/l

RT: 14.082 min Scan# 2062

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 24998

Ion Ratio Lower Upper

146 100

111 44.0 22.0 66.0





148 65.4 32.7 98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 48.643 ug/l

RT: 14.700 min Scan# 2167

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion: 75 Resp: 47931


Ion Ratio Lower Upper

75 100

155 74.5 37.3 111.8

157 92.4 46.2 138.6





#93

1,2,4-Trichlorobenzene

Concen: 52.553 ug/l

RT: 15.370 min Scan# 2281

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:180 Resp: 148472

Ion Ratio Lower Upper

180 100

182 94.5 47.3 141.8





145 30.9 15.4 46.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 49.659 ug/l

RT: 15.476 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VN087331.D

Acq: 16 Jul 2025 18:11




Tgt Ion:225 Resp: 52130


Ion Ratio Lower Upper

225 100

223 64.2 32.1 96.3


227 62.6 31.3 93.9





#95
Naphthalene
Concen: 54.490 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:128 Resp: 54536
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 10.5 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 51.547 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087331.D
Acq: 16 Jul 2025 18:11



Tgt Ion:180 Resp: 146080
Ion Ratio Lower Upper
180 100
182 94.3 47.1 141.4
145 33.7 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087332.D
 Acq On : 16 Jul 2025 18:32
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jul 17 02:20:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	192275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	339591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	314849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	168489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	314416	96.373	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 192.740%#			
35) Dibromofluoromethane	8.153	113	236468	100.947	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 201.900%#			
50) Toluene-d8	10.547	98	852301	101.999	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 204.000%#			
62) 4-Bromofluorobenzene	12.829	95	323070	104.650	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 209.300%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	233197	114.190	ug/l	92
3) Chloromethane	2.383	50	253369	98.660	ug/l	98
4) Vinyl Chloride	2.542	62	266221	104.312	ug/l	94
5) Bromomethane	2.959	94	136793	103.503	ug/l	96
6) Chloroethane	3.124	64	159422	95.783	ug/l	92
7) Trichlorofluoromethane	3.501	101	369032	97.786	ug/l	98
8) Diethyl Ether	3.959	74	142652	97.445	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.353	101	188841	97.477	ug/l	100
10) Methyl Iodide	4.571	142	213127	100.524	ug/l	99
11) Tert butyl alcohol	5.536	59	309615	499.800	ug/l	99
12) 1,1-Dichloroethene	4.336	96	197760	90.083	ug/l	95
13) Acrolein	4.177	56	250115	503.102	ug/l	99
14) Allyl chloride	5.006	41	382812	96.355	ug/l	100
15) Acrylonitrile	5.706	53	839080	499.150	ug/l	100
16) Acetone	4.424	43	693485	448.375	ug/l	96
17) Carbon Disulfide	4.695	76	631695	97.056	ug/l	95
18) Methyl Acetate	5.012	43	369894	96.247	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	818763	101.186	ug/l	98
20) Methylene Chloride	5.259	84	251830	98.001	ug/l	97
21) trans-1,2-Dichloroethene	5.771	96	237704	96.030	ug/l	94
22) Diisopropyl ether	6.659	45	832709	99.921	ug/l	96
23) Vinyl Acetate	6.594	43	3859420	529.516	ug/l	97
24) 1,1-Dichloroethane	6.553	63	455820	94.807	ug/l	100
25) 2-Butanone	7.471	43	1185700	501.667	ug/l	98
26) 2,2-Dichloropropane	7.477	77	358596	95.932	ug/l	99
27) cis-1,2-Dichloroethene	7.477	96	284578	99.858	ug/l	98
28) Bromochloromethane	7.800	49	229544	99.757	ug/l	100
29) Tetrahydrofuran	7.830	42	768521	500.532	ug/l	98
30) Chloroform	7.953	83	466838	97.008	ug/l	99
31) Cyclohexane	8.241	56	377159	94.035	ug/l	95
32) 1,1,1-Trichloroethane	8.153	97	403445	96.794	ug/l	97
36) 1,1-Dichloropropene	8.353	75	324047	104.705	ug/l	99
37) Ethyl Acetate	7.547	43	448428	100.326	ug/l	99
38) Carbon Tetrachloride	8.347	117	342043	100.328	ug/l	99
39) Methylcyclohexane	9.582	83	352211	105.118	ug/l	96
40) Benzene	8.588	78	1007257	100.700	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087332.D
 Acq On : 16 Jul 2025 18:32
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:20:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	239659	102.544	ug/l	99
42) 1,2-Dichloroethane	8.653	62	369731	97.472	ug/l	99
43) Isopropyl Acetate	8.677	43	704936	101.598	ug/l	99
44) Trichloroethene	9.335	130	230373	97.472	ug/l	93
45) 1,2-Dichloropropane	9.606	63	255369	100.478	ug/l	96
46) Dibromomethane	9.694	93	185908	97.696	ug/l	98
47) Bromodichloromethane	9.871	83	375910	98.072	ug/l	97
48) Methyl methacrylate	9.665	41	334941	107.228	ug/l	97
49) 1,4-Dioxane	9.688	88	102159	2135.348	ug/l #	95
51) 4-Methyl-2-Pentanone	10.429	43	2234274	509.127	ug/l	100
52) Toluene	10.612	92	622259	102.349	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	412142	106.245	ug/l	100
54) cis-1,3-Dichloropropene	10.294	75	421386	105.164	ug/l	98
55) 1,1,2-Trichloroethane	11.000	97	242683	98.595	ug/l	97
56) Ethyl methacrylate	10.859	69	425458	98.980	ug/l	98
57) 1,3-Dichloropropane	11.147	76	431753	101.453	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	1052628	521.327	ug/l	99
59) 2-Hexanone	11.176	43	1631949	560.508	ug/l	99
60) Dibromochloromethane	11.341	129	287682	102.484	ug/l	99
61) 1,2-Dibromoethane	11.453	107	258936	100.051	ug/l	98
64) Tetrachloroethene	11.082	164	195044	96.252	ug/l	97
65) Chlorobenzene	11.871	112	687696	97.289	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	239722	99.736	ug/l	99
67) Ethyl Benzene	11.947	91	1199510	103.080	ug/l	99
68) m/p-Xylenes	12.053	106	924069	212.065	ug/l	99
69) o-Xylene	12.376	106	447092	107.413	ug/l	100
70) Styrene	12.394	104	766471	109.464	ug/l	98
71) Bromoform	12.559	173	202915	104.498	ug/l #	99
73) Isopropylbenzene	12.676	105	1112821	104.940	ug/l	100
74) N-amyl acetate	12.482	43	464246	112.129	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.918	83	389392	97.587	ug/l	98
76) 1,2,3-Trichloropropane	12.970	75	326434m	87.327	ug/l	
77) Bromobenzene	12.959	156	279589	101.662	ug/l	100
78) n-propylbenzene	13.018	91	1384517	103.771	ug/l	99
79) 2-Chlorotoluene	13.106	91	840073	102.452	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	943119	104.383	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	140673	101.874	ug/l	96
82) 4-Chlorotoluene	13.200	91	859965	100.734	ug/l	99
83) tert-Butylbenzene	13.418	119	800700	106.106	ug/l	99
84) 1,2,4-Trimethylbenzene	13.459	105	967878	104.898	ug/l	100
85) sec-Butylbenzene	13.594	105	1141320	100.410	ug/l	98
86) p-Isopropyltoluene	13.706	119	975014	107.037	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	535172	99.151	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	548212	95.097	ug/l	99
89) n-Butylbenzene	14.035	91	880763	101.260	ug/l	99
90) Hexachloroethane	14.312	117	185973	96.358	ug/l	98
91) 1,2-Dichlorobenzene	14.082	146	508902	99.523	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	97572	93.137	ug/l	94
93) 1,2,4-Trichlorobenzene	15.364	180	310448	103.357	ug/l	99
94) Hexachlorobutadiene	15.476	225	106429	95.360	ug/l	98
95) Naphthalene	15.611	128	1162930	109.289	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	308888	102.519	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087332.D
Acq On : 16 Jul 2025 18:32
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:20:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:168 Resp: 192275
Ion Ratio Lower Upper
168 100
99 66.5 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 114.190 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion: 85 Resp: 233197
Ion Ratio Lower Upper
85 100
87 30.7 17.8 53.3





#3
Chloromethane
Concen: 98.660 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 50 Resp: 253369
Ion Ratio Lower Upper
50 100
52 31.8 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#4
Vinyl Chloride
Concen: 104.312 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Tgt Ion: 62 Resp: 266221
Ion Ratio Lower Upper
62 100
64 30.2 27.0 40.6





#5
Bromomethane
Concen: 103.503 ug/l
RT: 2.959 min Scan# 11
Delta R.T. -0.012 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 13679
Ion Ratio Lower Upper
94 100
96 95.4 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#6
Chloroethane
Concen: 95.783 ug/l
RT: 3.124 min Scan# 199
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 64 Resp: 159422
Ion Ratio Lower Upper
64 100
66 35.0 24.6 36.8









#7
Trichlorofluoromethane
Concen: 97.786 ug/l
RT: 3.501 min Scan# 263
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:101 Resp: 36903
Ion Ratio Lower Upper
101 100
103 63.6 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 97.445 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 74 Resp: 142652
Ion Ratio Lower Upper
74 100
45 104.3 50.8 152.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 97.477 ug/l

RT: 4.353 min Scan# 408

Delta R.T. -0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32





Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:	101	Resp:	18884
Ion	Ratio	Lower	Upper
101	100		
85	47.3	37.3	55.9
151	73.6	58.9	88.3

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#10

Methyl Iodide

Concen: 100.524 ug/l




RT: 4.571 min Scan# 445


Delta R.T. -0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Tgt Ion:	142	Resp:	213127
Ion	Ratio	Lower	Upper
142	100		
127	45.8	35.7	53.5
141	13.1	10.4	15.6





#11

Tert butyl alcohol

Concen: 499.800 ug/l

RT: 5.536 min Scan# 607

Delta R.T. 0.012 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 59 Resp: 30961

Ion Ratio Lower Upper

59 100





57 10.7 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#12

1,1-Dichloroethene

Concen: 90.083 ug/l

RT: 4.336 min Scan# 405

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion: 96 Resp: 197760


Ion Ratio Lower Upper

96 100

61 170.8 132.3 198.5

98 63.3 46.8 70.2





#13

Acrolein

Concen: 503.102 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 25011

Ion Ratio Lower Upper

56 100





55 71.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 96.355 ug/l

RT: 5.006 min Scan# 519

Delta R.T. -0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion: 41 Resp: 382812


Ion Ratio Lower Upper

41 100

39 73.3 59.0 88.6


76 35.9 28.7 43.1





#15
Acrylonitrile
Concen: 499.150 ug/l
RT: 5.706 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 839080
Ion Ratio Lower Upper
53 100
52 82.1 65.5 98.3
51 35.6 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#16
Acetone
Concen: 448.375 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 43 Resp: 693485
Ion Ratio Lower Upper
43 100
58 30.0 22.3 33.5





#17

Carbon Disulfide

Concen: 97.056 ug/l

RT: 4.695 min Scan# 4

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 63169

Ion Ratio Lower Upper

76 100





78 8.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#18

Methyl Acetate

Concen: 96.247 ug/l

RT: 5.012 min Scan# 520

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion: 43 Resp: 369894

Ion Ratio Lower Upper

43 100

74 22.5 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 101.186 ug/l

RT: 5.789 min Scan# 652

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 73 Resp: 81876

Ion Ratio Lower Upper

73 100





57 22.1 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 98.001 ug/l

RT: 5.259 min Scan# 562

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Tgt Ion: 84 Resp: 251830




Ion Ratio Lower Upper


84 100

49 131.0 107.5 161.3

51 41.0 30.2 45.2

86 62.7 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 96.030 ug/l
RT: 5.771 min Scan# 649
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 96 Resp: 237704
Ion Ratio Lower Upper
96 100
61 143.5 122.0 183.0
98 64.7 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22
Diisopropyl ether
Concen: 99.921 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 45 Resp: 832709
Ion Ratio Lower Upper
45 100
43 56.9 42.8 64.2
87 25.7 19.8 29.6
59 11.2 8.3 12.5









#23
Vinyl Acetate
Concen: 529.516 ug/l
RT: 6.594 min Scan# 789
Delta R.T. 0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 3859420
Ion Ratio Lower Upper
43 100
86 10.7 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#24
1,1-Dichloroethane
Concen: 94.807 ug/l
RT: 6.553 min Scan# 782
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 63 Resp: 455820
Ion Ratio Lower Upper
63 100
98 6.5 3.3 9.9
100 4.7 2.5 7.4









#25
2-Butanone
Concen: 501.667 ug/l
RT: 7.471 min Scan# 911
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1185700
Ion Ratio Lower Upper
43 100
72 25.3 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#26
2,2-Dichloropropane
Concen: 95.932 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 77 Resp: 358596
Ion Ratio Lower Upper
77 100
97 21.9 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 99.858 ug/l

RT: 7.477 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087332.D


Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100



Tgt Ion: 96 Resp: 284578

Ion Ratio Lower Upper

96 100

61 147.5 0.0 297.8




98 63.9 0.0 132.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane


Concen: 99.757 ug/l

RT: 7.800 min Scan# 994

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32





Tgt Ion: 49 Resp: 229544


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.2

130 71.8 57.3 85.9





#29

Tetrahydrofuran

Concen: 500.532 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 42 Resp: 76852

Ion Ratio Lower Upper

42 100

72 42.4 33.4 50.0





71 40.2 31.2 46.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 97.008 ug/l

RT: 7.953 min Scan# 1020

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion: 83 Resp: 466838

Ion Ratio Lower Upper

83 100

85 65.3 52.7 79.1





#31

Cyclohexane

Concen: 94.035 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 377159

Ion Ratio Lower Upper

56 100

69 31.4 22.7 34.1





84 86.8 65.8 98.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#32

1,1,1-Trichloroethane

Concen: 96.794 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion: 97 Resp: 403445


Ion Ratio Lower Upper

97 100

99 62.9 51.8 77.8

61 46.4 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 96.373 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. -0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 65 Resp: 314410

Ion Ratio Lower Upper

65 100





67 51.7 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion:114 Resp: 339591


Ion Ratio Lower Upper

114 100

63 22.2 0.0 44.6

88 16.1 0.0 32.8





#35

Dibromofluoromethane

Concen: 100.947 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:113 Resp: 23646

Ion Ratio Lower Upper

113 100

111 100.2 82.5 123.7





192 17.5 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#36

1,1-Dichloropropene

Concen: 104.705 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion: 75 Resp: 324047


Ion Ratio Lower Upper

75 100

110 32.9 16.7 50.1

77 30.9 25.2 37.8





#37

Ethyl Acetate

Concen: 100.326 ug/l

RT: 7.547 min Scan# 951

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 448428

Ion Ratio Lower Upper

43 100

61 13.3 10.9 16.3





70 10.1 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 100.328 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion:117 Resp: 342043


Ion Ratio Lower Upper

117 100

119 99.6 80.2 120.2

121 31.0 25.4 38.2





#39

Methylcyclohexane

Concen: 105.118 ug/l

RT: 9.582 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 35221

Ion Ratio Lower Upper

83 100

55 80.6 61.3 91.9





98 50.1 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#40

Benzene

Concen: 100.700 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion: 78 Resp: 1007257

Ion Ratio Lower Upper

78 100


77 23.2 18.2 27.2





#41
Methacrylonitrile
Concen: 102.544 ug/l
RT: 7.765 min Scan# 988
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 41 Resp: 239659
Ion Ratio Lower Upper
41 100
39 54.5 43.4 65.0
67 69.8 55.1 82.7
52 30.2 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#42
1,2-Dichloroethane
Concen: 97.472 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 62 Resp: 369731
Ion Ratio Lower Upper
62 100
98 8.6 0.0 18.0





#43

Isopropyl Acetate

Concen: 101.598 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion:	43	Resp:	704930
Ion	Ratio	Lower	Upper
43	100		
61	24.8	19.8	29.8
87	12.8	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene

Concen: 97.472 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Tgt Ion:	130	Resp:	230373
Ion	Ratio	Lower	Upper
130	100		
95	104.4	0.0	195.2





#45

1,2-Dichloropropane

Concen: 100.478 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 255369

Ion Ratio Lower Upper

63 100





65 31.4 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#46

Dibromomethane

Concen: 97.696 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion: 93 Resp: 185908


Ion Ratio Lower Upper

93 100

95 83.5 65.8 98.8

174 90.3 69.9 104.9





#47

Bromodichloromethane

Concen: 98.072 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 375910
Ion Ratio Lower Upper
83 100
85 65.7 50.4 75.6
127 8.1 5.8 8.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#48

Methyl methacrylate

Concen: 107.228 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Tgt Ion: 41 Resp: 334941
Ion Ratio Lower Upper
41 100
69 83.2 64.1 96.1
39 58.3 45.5 68.3





#49

1,4-Dioxane

Concen: 2135.348 ug/l

RT: 9.688 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 88 Resp: 102159

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 72.4 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 101.999 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion: 98 Resp: 852301

Ion Ratio Lower Upper

98 100


100 66.6 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 509.127 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 43 Resp: 2234274
Ion Ratio Lower Upper
43 100
58 38.7 30.8 46.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#52
Toluene
Concen: 102.349 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Tgt Ion: 92 Resp: 622259
Ion Ratio Lower Upper
92 100
91 172.5 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 106.245 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 41214
Ion Ratio Lower Upper
75 100
77 29.7 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 105.164 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion: 75 Resp: 421386
Ion Ratio Lower Upper
75 100
77 30.9 24.2 36.2
39 49.1 38.4 57.6





#55
1,1,2-Trichloroethane
Concen: 98.595 ug/l
RT: 11.000 min Scan# 1538
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 97 Resp: 242683
Ion Ratio Lower Upper
97 100
83 89.9 72.6 109.0
85 55.4 46.7 70.1
99 62.7 54.1 81.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#56
Ethyl methacrylate
Concen: 98.980 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32



Tgt Ion: 69 Resp: 425458
Ion Ratio Lower Upper
69 100
41 69.0 55.1 82.7
39 38.9 27.9 41.9





#57

1,3-Dichloropropane

Concen: 101.453 ug/l

RT: 11.147 min Scan# 1563

Delta R.T. 0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 43175

Ion Ratio Lower Upper





76 100

78 32.3 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 521.327 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion: 63 Resp: 1052628

Ion Ratio Lower Upper

63 100

106 27.5 21.7 32.5









#59
2-Hexanone
Concen: 560.508 ug/l
RT: 11.176 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1631949
Ion Ratio Lower Upper
43 100
58 54.1 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#60
Dibromochloromethane
Concen: 102.484 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion:129 Resp: 287682
Ion Ratio Lower Upper
129 100
127 77.3 39.1 117.5





#61
1,2-Dibromoethane
Concen: 100.051 ug/l
RT: 11.453 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:107 Resp: 258930
Ion Ratio Lower Upper
107 100
109 94.9 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#62
4-Bromofluorobenzene
Concen: 104.650 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion: 95 Resp: 323070
Ion Ratio Lower Upper
95 100
174 75.5 0.0 149.4
176 72.2 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:117 Resp: 314849
Ion Ratio Lower Upper
117 100
82 55.1 47.4 71.2
119 31.4 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#64
Tetrachloroethene
Concen: 96.252 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion:164 Resp: 195044
Ion Ratio Lower Upper
164 100
166 123.8 105.5 158.3
129 97.4 77.4 116.2
131 95.2 77.3 115.9









#65
Chlorobenzene
Concen: 97.289 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:112 Resp: 687690
Ion Ratio Lower Upper
112 100
114 31.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 99.736 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion:131 Resp: 239722
Ion Ratio Lower Upper
131 100
133 95.7 47.4 142.3
119 66.2 33.1 99.2





#67

Ethyl Benzene

Concen: 103.080 ug/l

RT: 11.947 min Scan# 1699

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 1199510

Ion Ratio Lower Upper

91 100





106 30.9 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#68

m/p-Xylenes

Concen: 212.065 ug/l

RT: 12.053 min Scan# 1717

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion:106 Resp: 924069

Ion Ratio Lower Upper

106 100

91 204.6 162.0 243.0









#69
o-Xylene
Concen: 107.413 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 447092
Ion Ratio Lower Upper
106 100
91 216.0 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#70
Styrene
Concen: 109.464 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion:104 Resp: 766471
Ion Ratio Lower Upper
104 100
78 53.0 41.0 61.6
103 54.0 43.9 65.9





#71
Bromoform
Concen: 104.498 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion:173 Resp: 202915
Ion Ratio Lower Upper
173 100
175 47.7 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32



Tgt Ion:152 Resp: 168489
Ion Ratio Lower Upper
152 100
115 59.2 31.1 93.5
150 186.3 0.0 349.0









#73
Isopropylbenzene
Concen: 104.940 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 105 Resp: 111282
Ion Ratio Lower Upper
105 100
120 26.5 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#74
N-ethyl acetate
Concen: 112.129 ug/l
RT: 12.482 min Scan# 1790
Delta R.T. -0.012 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32


Tgt Ion: 43 Resp: 464246
Ion Ratio Lower Upper
43 100
70 41.1 37.6 56.4
55 23.5 19.6 29.4
61 22.9 20.6 31.0





#75
1,1,2,2-Tetrachloroethane
Concen: 97.587 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100




Tgt Ion: 83 Resp: 389392
Ion Ratio Lower Upper
83 100
131 10.2 5.1 15.3
85 63.5 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#76
1,2,3-Trichloropropane
Concen: 87.327 ug/l m
RT: 12.970 min Scan# 1873
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Tgt Ion: 75 Resp: 326434
Ion Ratio Lower Upper
75 100
77 196.5 94.5 283.6





#77
Bromobenzene
Concen: 101.662 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:156 Resp: 279589
Ion Ratio Lower Upper
156 100
77 229.4 114.9 344.6
158 97.6 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 103.771 ug/l
RT: 13.018 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 91 Resp: 1384517
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8





#79

2-Chlorotoluene

Concen: 102.452 ug/l

RT: 13.106 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 84007

Ion Ratio Lower Upper

91 100





126 32.8 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#80

1,3,5-Trimethylbenzene

Concen: 104.383 ug/l

RT: 13.153 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion:105 Resp: 943119

Ion Ratio Lower Upper

105 100

120 48.7 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 101.874 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 75 Resp: 14067
Ion Ratio Lower Upper
75 100
53 110.2 83.5 125.3
89 49.1 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 100.734 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 91 Resp: 859965
Ion Ratio Lower Upper
91 100
126 33.8 16.6 49.7









#83
tert-Butylbenzene
Concen: 106.106 ug/l
RT: 13.418 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:119 Resp: 800700
Ion Ratio Lower Upper
119 100
91 69.4 35.0 105.1
134 24.9 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 104.898 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion:105 Resp: 967878
Ion Ratio Lower Upper
105 100
120 45.5 22.8 68.3





#85

sec-Butylbenzene

Concen: 100.410 ug/l

RT: 13.594 min Scan# 1979

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:105 Resp: 1141320

Ion Ratio Lower Upper

105 100





134 20.3 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#86

p-Isopropyltoluene

Concen: 107.037 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion:119 Resp: 975014


Ion Ratio Lower Upper

119 100

134 26.0 13.5 40.5

91 24.5 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 99.151 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 53517

Ion Ratio Lower Upper

146 100

111 42.4 21.4 64.3





148 63.2 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 95.097 ug/l

RT: 13.788 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion:146 Resp: 548212


Ion Ratio Lower Upper

146 100

111 40.6 19.6 58.7

148 62.7 31.4 94.0





#89

n-Butylbenzene

Concen: 101.260 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 88076

Ion Ratio Lower Upper

91 100

92 52.7 26.2 78.6





134 25.7 12.4 37.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 96.358 ug/l

RT: 14.312 min Scan# 2101

Delta R.T. 0.000 min

Lab File: VN087332.D




Acq: 16 Jul 2025 18:32


Tgt Ion:117 Resp: 185973

Ion Ratio Lower Upper

117 100

201 67.5 32.8 98.4









#91
1,2-Dichlorobenzene
Concen: 99.523 ug/l
RT: 14.082 min Scan# 2062
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:146 Resp: 508902
Ion Ratio Lower Upper
146 100
111 43.8 22.0 66.0
148 63.6 32.7 98.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 93.137 ug/l
RT: 14.700 min Scan# 2167
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion: 75 Resp: 97572
Ion Ratio Lower Upper
75 100
155 78.0 37.3 111.8
157 99.6 46.2 138.6





#93

1,2,4-Trichlorobenzene

Concen: 103.357 ug/l

RT: 15.364 min Scan# 21

Delta R.T. -0.006 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:180 Resp: 31044

Ion Ratio Lower Upper

180 100

182 93.8 47.3 141.8





145 32.7 15.4 46.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 95.360 ug/l

RT: 15.476 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VN087332.D

Acq: 16 Jul 2025 18:32




Tgt Ion:225 Resp: 106429


Ion Ratio Lower Upper

225 100

223 61.0 32.1 96.3


227 62.4 31.3 93.9





#95
Naphthalene
Concen: 109.289 ug/l
RT: 15.611 min Scan# 2322
Delta R.T. -0.006 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 1162930
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 11.0 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 102.519 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087332.D
Acq: 16 Jul 2025 18:32

Tgt Ion:180 Resp: 308888
Ion Ratio Lower Upper
180 100
182 96.2 47.1 141.4
145 34.2 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jul 17 02:21:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	193853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	344966	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	317026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	165077	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.565	65	501760	152.545	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	305.080%#
35) Dibromofluoromethane	8.153	113	368306	154.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	309.560%#
50) Toluene-d8	10.547	98	1362098	160.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	320.940%#
62) 4-Bromofluorobenzene	12.829	95	522221	166.525	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	333.040%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	363327	176.463	ug/l	95
3) Chloromethane	2.383	50	405800	156.728	ug/l	99
4) Vinyl Chloride	2.542	62	418507	162.646	ug/l	94
5) Bromomethane	2.942	94	215335	161.604	ug/l	95
6) Chloroethane	3.118	64	249466	148.663	ug/l	100
7) Trichlorofluoromethane	3.501	101	585457	153.872	ug/l	97
8) Diethyl Ether	3.959	74	228544	154.847	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.353	101	296076	151.587	ug/l	99
10) Methyl Iodide	4.571	142	328079	150.516	ug/l	99
11) Tert butyl alcohol	5.536	59	478279	765.784	ug/l	100
12) 1,1-Dichloroethene	4.330	96	312325	141.111	ug/l	97
13) Acrolein	4.177	56	415488	828.945	ug/l	100
14) Allyl chloride	5.006	41	663403	165.621	ug/l	91
15) Acrylonitrile	5.706	53	1279363	754.869	ug/l	100
16) Acetone	4.424	43	1064061	682.373	ug/l	96
17) Carbon Disulfide	4.695	76	1011315	154.118	ug/l	95
18) Methyl Acetate	5.012	43	577536	149.052	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	1286728	157.724	ug/l	95
20) Methylene Chloride	5.259	84	390988	151.266	ug/l	94
21) trans-1,2-Dichloroethene	5.765	96	356549	142.870	ug/l	96
22) Diisopropyl ether	6.659	45	1280267	152.375	ug/l	99
23) Vinyl Acetate	6.589	43	5943020	808.750	ug/l	98
24) 1,1-Dichloroethane	6.553	63	704727	145.384	ug/l	98
25) 2-Butanone	7.477	43	1797146	754.179	ug/l	98
26) 2,2-Dichloropropane	7.477	77	547190	145.193	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	437021	152.102	ug/l	99
28) Bromochloromethane	7.800	49	339529	146.355	ug/l	100
29) Tetrahydrofuran	7.830	42	1166702	753.679	ug/l	98
30) Chloroform	7.953	83	717424	147.865	ug/l	98
31) Cyclohexane	8.241	56	582988	144.170	ug/l	94
32) 1,1,1-Trichloroethane	8.153	97	630986	150.152	ug/l	98
36) 1,1-Dichloropropene	8.353	75	503793	160.248	ug/l	99
37) Ethyl Acetate	7.553	43	692256	152.464	ug/l	98
38) Carbon Tetrachloride	8.347	117	535575	154.647	ug/l	98
39) Methylcyclohexane	9.583	83	559851	164.485	ug/l	98
40) Benzene	8.588	78	1550984	152.642	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	375482	158.156	ug/l	99
42) 1,2-Dichloroethane	8.653	62	571634	148.352	ug/l	100
43) Isopropyl Acetate	8.677	43	1091736	154.893	ug/l	99
44) Trichloroethene	9.335	130	364760	151.927	ug/l	96
45) 1,2-Dichloropropane	9.606	63	391676	151.708	ug/l	97
46) Dibromomethane	9.694	93	287252	148.602	ug/l	98
47) Bromodichloromethane	9.871	83	584546	150.128	ug/l	99
48) Methyl methacrylate	9.665	41	520990	164.191	ug/l	99
49) 1,4-Dioxane	9.688	88	157400	3238.745	ug/l #	93
51) 4-Methyl-2-Pentanone	10.430	43	3375607	757.219	ug/l	99
52) Toluene	10.612	92	961314	155.654	ug/l	99
53) t-1,3-Dichloropropene	10.818	75	640188	162.462	ug/l	94
54) cis-1,3-Dichloropropene	10.294	75	654301	160.747	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	365858	146.322	ug/l	99
56) Ethyl methacrylate	10.859	69	679202	150.296	ug/l	96
57) 1,3-Dichloropropane	11.147	76	662154	153.168	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	1717427	837.324	ug/l	100
59) 2-Hexanone	11.177	43	2477578	837.689	ug/l	98
60) Dibromochloromethane	11.341	129	448307	157.217	ug/l	99
61) 1,2-Dibromoethane	11.453	107	402978	153.282	ug/l	97
64) Tetrachloroethene	11.082	164	301404	147.718	ug/l	97
65) Chlorobenzene	11.871	112	1066757	149.878	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	375162	155.013	ug/l	98
67) Ethyl Benzene	11.941	91	1881920	160.613	ug/l	99
68) m/p-Xylenes	12.053	106	1437290	327.579	ug/l	98
69) o-Xylene	12.376	106	698363	166.628	ug/l	98
70) Styrene	12.394	104	1195068	169.502	ug/l	99
71) Bromoform	12.559	173	320960	164.154	ug/l #	99
73) Isopropylbenzene	12.676	105	1747567	168.203	ug/l	100
74) N-amyl acetate	12.482	43	750104	184.917	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.918	83	581568	148.761	ug/l	100
76) 1,2,3-Trichloropropane	12.971	75	504791m	137.832	ug/l	
77) Bromobenzene	12.959	156	433814	161.001	ug/l	99
78) n-propylbenzene	13.018	91	2127215	162.733	ug/l	100
79) 2-Chlorotoluene	13.106	91	1290205	160.600	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	1465159	165.514	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	224191	165.713	ug/l	95
82) 4-Chlorotoluene	13.200	91	1336462	159.785	ug/l	100
83) tert-Butylbenzene	13.418	119	1230683	166.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	1502651	166.222	ug/l	100
85) sec-Butylbenzene	13.594	105	1767949	158.754	ug/l	99
86) p-Isopropyltoluene	13.706	119	1501303	168.219	ug/l	98
87) 1,3-Dichlorobenzene	13.712	146	828885	156.742	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	830255	147.000	ug/l	99
89) n-Butylbenzene	14.035	91	1367813	160.505	ug/l	98
90) Hexachloroethane	14.312	117	292183	154.518	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	784106	156.513	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.694	75	151090	147.203	ug/l	93
93) 1,2,4-Trichlorobenzene	15.370	180	492237	167.266	ug/l	99
94) Hexachlorobutadiene	15.476	225	163922	149.909	ug/l	99
95) Naphthalene	15.612	128	1880423	180.370	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	491195	166.396	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087333.D
Acq On : 16 Jul 2025 18:54
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Quant Time: Jul 17 02:21:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:168 Resp: 19385
Ion Ratio Lower Upper
168 100
99 61.8 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 176.463 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 85 Resp: 363327
Ion Ratio Lower Upper
85 100
87 32.9 17.8 53.3









#3
Chloromethane
Concen: 156.728 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 50 Resp: 405800
Ion Ratio Lower Upper
50 100
52 32.3 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 162.646 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 62 Resp: 418507
Ion Ratio Lower Upper
62 100
64 30.6 27.0 40.6





#5

Bromomethane

Concen: 161.604 ug/l

RT: 2.942 min Scan# 1

Delta R.T. -0.029 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 94 Resp: 21533

Ion Ratio Lower Upper

94 100





96 96.7 73.4 110.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#6

Chloroethane

Concen: 148.663 ug/l

RT: 3.118 min Scan# 198

Delta R.T. -0.012 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 64 Resp: 249466

Ion Ratio Lower Upper

64 100

66 30.7 24.6 36.8









#7
Trichlorofluoromethane
Concen: 153.872 ug/l
RT: 3.501 min Scan# 263
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 58545
Ion Ratio Lower Upper
101 100
103 64.5 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 154.847 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 74 Resp: 228544
Ion Ratio Lower Upper
74 100
45 100.7 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 151.587 ug/l
RT: 4.353 min Scan# 409
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:101 Resp: 296070
Ion Ratio Lower Upper
101 100
85 46.1 37.3 55.9
151 73.1 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#10
Methyl Iodide
Concen: 150.516 ug/l
RT: 4.571 min Scan# 445
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion:142 Resp: 328079
Ion Ratio Lower Upper
142 100
127 44.0 35.7 53.5
141 14.1 10.4 15.6





#11

Tert butyl alcohol

Concen: 765.784 ug/l

RT: 5.536 min Scan# 607

Delta R.T. 0.012 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 478279

Ion Ratio Lower Upper

59 100





57 10.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#12

1,1-Dichloroethene

Concen: 141.111 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 96 Resp: 312325


Ion Ratio Lower Upper

96 100

61 168.2 132.3 198.5

98 63.0 46.8 70.2





#13

Acrolein

Concen: 828.945 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 415488

Ion Ratio Lower Upper

56 100





55 70.4 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 165.621 ug/l

RT: 5.006 min Scan# 519

Delta R.T. -0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 41 Resp: 663403


Ion Ratio Lower Upper

41 100

39 64.9 59.0 88.6

76 32.4 28.7 43.1





#15

Acrylonitrile

Concen: 754.869 ug/l

RT: 5.706 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 53 Resp: 127936

Ion Ratio Lower Upper

53 100

52 81.7 65.5 98.3





51 35.5 28.7 43.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#16

Acetone

Concen: 682.373 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 43 Resp: 1064061

Ion Ratio Lower Upper

43 100

58 29.9 22.3 33.5









#17
Carbon Disulfide
Concen: 154.118 ug/l
RT: 4.695 min Scan# 40
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 76 Resp: 101131
Ion Ratio Lower Upper
76 100
78 9.1 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#18
Methyl Acetate
Concen: 149.052 ug/l
RT: 5.012 min Scan# 520
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 43 Resp: 577536
Ion Ratio Lower Upper
43 100
74 22.7 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 157.724 ug/l

RT: 5.789 min Scan# 652

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 73 Resp: 1286728

Ion Ratio Lower Upper





73 100

57 20.8 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 151.266 ug/l

RT: 5.259 min Scan# 562

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Tgt Ion: 84 Resp: 390988




Ion Ratio Lower Upper


84 100

49 122.7 107.5 161.3

51 37.7 30.2 45.2


86 62.5 49.3 73.9





#21
trans-1,2-Dichloroethene
Concen: 142.870 ug/l
RT: 5.765 min Scan# 649
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 356549
Ion Ratio Lower Upper
96 100
61 145.5 122.0 183.0
98 62.9 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22
Diisopropyl ether
Concen: 152.375 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 45 Resp: 1280267
Ion Ratio Lower Upper
45 100
43 53.9 42.8 64.2
87 25.9 19.8 29.6
59 11.2 8.3 12.5





#23

Vinyl Acetate

Concen: 808.750 ug/l

RT: 6.589 min Scan# 788

Delta R.T. 0.001 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 5943020

Ion Ratio Lower Upper

43 100





86 10.2 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#24

1,1-Dichloroethane

Concen: 145.384 ug/l

RT: 6.553 min Scan# 782

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 63 Resp: 704727


Ion Ratio Lower Upper

63 100

98 7.0 3.3 9.9

100 4.2 2.5 7.4





#25

2-Butanone

Concen: 754.179 ug/l

RT: 7.477 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 179714

Ion Ratio Lower Upper

43 100





72 25.6 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#26

2,2-Dichloropropane

Concen: 145.193 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.006 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 77 Resp: 547190

Ion Ratio Lower Upper

77 100

97 22.2 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 152.102 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 96 Resp: 43702

Ion Ratio Lower Upper

96 100

61 149.1 0.0 297.8





98 63.7 0.0 132.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 146.355 ug/l

RT: 7.800 min Scan# 994

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 49 Resp: 339529


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.2

130 71.9 57.3 85.9





#29

Tetrahydrofuran

Concen: 753.679 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 116670

Ion Ratio Lower Upper

42 100

72 43.0 33.4 50.0





71 40.3 31.2 46.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 147.865 ug/l

RT: 7.953 min Scan# 1020

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 83 Resp: 717424

Ion Ratio Lower Upper

83 100

85 64.4 52.7 79.1





#31

Cyclohexane

Concen: 144.170 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 58298

Ion Ratio Lower Upper

56 100

69 31.4 22.7 34.1





84 87.7 65.8 98.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#32

1,1,1-Trichloroethane

Concen: 150.152 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 97 Resp: 630986


Ion Ratio Lower Upper

97 100

99 64.9 51.8 77.8

61 45.0 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 152.545 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 65 Resp: 501760

Ion Ratio Lower Upper

65 100





67 51.9 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion:114 Resp: 344966


Ion Ratio Lower Upper

114 100

63 20.5 0.0 44.6

88 15.2 0.0 32.8





#35

Dibromofluoromethane

Concen: 154.778 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion:113 Resp: 368300

Ion Ratio Lower Upper

113 100

111 103.8 82.5 123.7





192 17.8 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#36

1,1-Dichloropropene

Concen: 160.248 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 75 Resp: 503793


Ion Ratio Lower Upper

75 100

110 33.1 16.7 50.1

77 30.9 25.2 37.8





#37

Ethyl Acetate

Concen: 152.464 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 692250

Ion Ratio Lower Upper

43 100

61 13.2 10.9 16.3





70 10.1 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 154.647 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion:117 Resp: 535575


Ion Ratio Lower Upper

117 100

119 98.0 80.2 120.2

121 30.7 25.4 38.2





#39

Methylcyclohexane

Concen: 164.485 ug/l

RT: 9.583 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 83 Resp: 55985

Ion Ratio Lower Upper

83 100

55 77.8 61.3 91.9





98 48.6 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#40

Benzene

Concen: 152.642 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 78 Resp: 1550984

Ion Ratio Lower Upper

78 100


77 23.5 18.2 27.2





#41
Methacrylonitrile
Concen: 158.156 ug/l
RT: 7.765 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
Client Sample Id :
VSTDICC150



Tgt Ion: 41 Resp: 37548
Ion Ratio Lower Upper
41 100
39 53.7 43.4 65.0
67 68.5 55.1 82.7
52 29.7 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#42
1,2-Dichloroethane
Concen: 148.352 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 62 Resp: 571634
Ion Ratio Lower Upper
62 100
98 9.0 0.0 18.0





#43

Isopropyl Acetate

Concen: 154.893 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087333.D


Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150



Tgt Ion: 43 Resp: 1091730

Ion Ratio Lower Upper

43 100

61 24.8 19.8 29.8




87 12.8 9.8 14.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene

Concen: 151.927 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion:130 Resp: 364760

Ion Ratio Lower Upper

130 100

95 101.7 0.0 195.2





#45

1,2-Dichloropropane

Concen: 151.708 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 391670

Ion Ratio Lower Upper

63 100





65 31.9 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#46

Dibromomethane

Concen: 148.602 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 93 Resp: 287252


Ion Ratio Lower Upper

93 100

95 83.3 65.8 98.8

174 89.7 69.9 104.9





#47

Bromodichloromethane

Concen: 150.128 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 584540
Ion Ratio Lower Upper
83 100
85 63.8 50.4 75.6
127 8.4 5.8 8.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#48

Methyl methacrylate

Concen: 164.191 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Tgt Ion: 41 Resp: 520990
Ion Ratio Lower Upper
41 100
69 81.7 64.1 96.1
39 57.0 45.5 68.3





#49

1,4-Dioxane

Concen: 3238.745 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 88 Resp: 157400

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 70.5 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 160.469 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 98 Resp: 1362098

Ion Ratio Lower Upper

98 100


100 66.2 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 757.219 ug/l
RT: 10.430 min Scan# 1441
Delta R.T. 0.001 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 43 Resp: 337560
Ion Ratio Lower Upper
43 100
58 38.8 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#52
Toluene
Concen: 155.654 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 92 Resp: 961314
Ion Ratio Lower Upper
92 100
91 170.6 135.1 202.7









#53
t-1,3-Dichloropropene
Concen: 162.462 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 75 Resp: 640188
Ion Ratio Lower Upper
75 100
77 33.0 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 160.747 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 75 Resp: 654301
Ion Ratio Lower Upper
75 100
77 32.0 24.2 36.2
39 51.0 38.4 57.6





#55

1,1,2-Trichloroethane

Concen: 146.322 ug/l

RT: 11.000 min Scan# 1538

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 97 Resp: 36585

Ion Ratio Lower Upper

97 100

83 90.5 72.6 109.0

85 57.3 46.7 70.1





99 66.0 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#56

Ethyl methacrylate

Concen: 150.296 ug/l

RT: 10.859 min Scan# 1514

Delta R.T. -0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 69 Resp: 679202


Ion Ratio Lower Upper

69 100

41 66.4 55.1 82.7

39 37.8 27.9 41.9





#57

1,3-Dichloropropane

Concen: 153.168 ug/l

RT: 11.147 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 76 Resp: 662154

Ion Ratio Lower Upper

76 100





78 32.2 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 837.324 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion: 63 Resp: 1717427

Ion Ratio Lower Upper

63 100

106 27.4 21.7 32.5









#59
2-Hexanone
Concen: 837.689 ug/l
RT: 11.177 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion: 43 Resp: 2477578
Ion Ratio Lower Upper
43 100
58 54.4 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#60
Dibromochloromethane
Concen: 157.217 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54


Tgt Ion:129 Resp: 448307
Ion Ratio Lower Upper
129 100
127 77.8 39.1 117.5





#61
1,2-Dibromoethane
Concen: 153.282 ug/l
RT: 11.453 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150




Tgt Ion: 107 Resp: 402978
Ion Ratio Lower Upper
107 100
109 94.2 77.5 116.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#62
4-Bromofluorobenzene
Concen: 166.525 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54



Tgt Ion: 95 Resp: 522221
Ion Ratio Lower Upper
95 100
174 74.8 0.0 149.4
176 71.2 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:117 Resp: 317020
Ion Ratio Lower Upper
117 100
82 57.3 47.4 71.2
119 31.6 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#64
Tetrachloroethene
Concen: 147.718 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54


Tgt Ion:164 Resp: 301404
Ion Ratio Lower Upper
164 100
166 124.4 105.5 158.3
129 99.0 77.4 116.2
131 95.9 77.3 115.9





#65
Chlorobenzene
Concen: 149.878 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150




Tgt Ion:112 Resp: 106675
Ion Ratio Lower Upper
112 100
114 31.6 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 155.013 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Tgt Ion:131 Resp: 375162
Ion Ratio Lower Upper
131 100
133 93.5 47.4 142.3
119 64.4 33.1 99.2





#67
Ethyl Benzene
Concen: 160.613 ug/l
RT: 11.941 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 91 Resp: 1881920
Ion Ratio Lower Upper
91 100
106 29.9 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#68
m/p-Xylenes
Concen: 327.579 ug/l
RT: 12.053 min Scan# 1717
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion:106 Resp: 1437290
Ion Ratio Lower Upper
106 100
91 206.0 162.0 243.0





#69

o-Xylene

Concen: 166.628 ug/l

RT: 12.376 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion:106 Resp: 69836

Ion Ratio Lower Upper

106 100





91 218.8 107.7 323.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#70

Styrene

Concen: 169.502 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion:104 Resp: 1195068


Ion Ratio Lower Upper

104 100

78 52.6 41.0 61.6


103 54.2 43.9 65.9





#71
Bromoform
Concen: 164.154 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150




Tgt Ion:173 Resp: 320960
Ion Ratio Lower Upper
173 100
175 49.0 24.1 72.3
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.771 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54



Tgt Ion:152 Resp: 165077
Ion Ratio Lower Upper
152 100
115 62.1 31.1 93.5
150 205.0 0.0 349.0









#73
Isopropylbenzene
Concen: 168.203 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 174756
Ion Ratio Lower Upper
105 100
120 26.7 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#74
N-ethyl acetate
Concen: 184.917 ug/l
RT: 12.482 min Scan# 1790
Delta R.T. -0.012 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54


Tgt Ion: 43 Resp: 750104
Ion Ratio Lower Upper
43 100
70 38.7 37.6 56.4
55 21.4 19.6 29.4
61 23.2 20.6 31.0





#75
1,1,2,2-Tetrachloroethane
Concen: 148.761 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150




Tgt Ion: 83 Resp: 58156
Ion Ratio Lower Upper
83 100
131 10.6 5.1 15.3
85 65.0 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#76
1,2,3-Trichloropropane
Concen: 137.832 ug/l m
RT: 12.971 min Scan# 1873
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Tgt Ion: 75 Resp: 504791
Ion Ratio Lower Upper
75 100
77 195.6 94.5 283.6





#77
Bromobenzene
Concen: 161.001 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:156 Resp: 433814
Ion Ratio Lower Upper
156 100
77 227.6 114.9 344.6
158 95.5 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 162.733 ug/l
RT: 13.018 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 91 Resp: 2127215
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8





#79

2-Chlorotoluene

Concen: 160.600 ug/l

RT: 13.106 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion: 91 Resp: 129020

Ion Ratio Lower Upper

91 100





126 32.6 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#80

1,3,5-Trimethylbenzene

Concen: 165.514 ug/l

RT: 13.153 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion:105 Resp: 1465159

Ion Ratio Lower Upper

105 100


120 48.8 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 165.713 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion: 75 Resp: 22419
Ion Ratio Lower Upper
75 100
53 97.9 83.5 125.3
89 47.3 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 159.785 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion: 91 Resp: 1336462
Ion Ratio Lower Upper
91 100
126 33.1 16.6 49.7









#83
tert-Butylbenzene
Concen: 166.457 ug/l
RT: 13.418 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:119 Resp: 123068
Ion Ratio Lower Upper
119 100
91 70.4 35.0 105.1
134 27.7 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 166.222 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion:105 Resp: 1502651
Ion Ratio Lower Upper
105 100
120 45.2 22.8 68.3









#85
sec-Butylbenzene
Concen: 158.754 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Tgt Ion:105 Resp: 1767949
Ion Ratio Lower Upper
105 100
134 19.9 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#86
p-Isopropyltoluene
Concen: 168.219 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion:119 Resp: 1501303
Ion Ratio Lower Upper
119 100
134 25.8 13.5 40.5
91 25.0 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 156.742 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 82888

Ion Ratio Lower Upper

146 100

111 42.0 21.4 64.3





148 62.5 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 147.000 ug/l

RT: 13.788 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion:146 Resp: 830255


Ion Ratio Lower Upper

146 100

111 40.5 19.6 58.7

148 63.1 31.4 94.0





#89

n-Butylbenzene

Concen: 160.505 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 136781

Ion Ratio Lower Upper

91 100

92 53.7 26.2 78.6





134 25.8 12.4 37.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 154.518 ug/l

RT: 14.312 min Scan# 2101

Delta R.T. 0.000 min

Lab File: VN087333.D




Acq: 16 Jul 2025 18:54


Tgt Ion:117 Resp: 292183

Ion Ratio Lower Upper

117 100

201 68.0 32.8 98.4





#91

1,2-Dichlorobenzene

Concen: 156.513 ug/l

RT: 14.082 min Scan# 2062

Delta R.T. 0.000 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC150

Tgt Ion:146 Resp: 784100

Ion Ratio Lower Upper

146 100

111 43.9 22.0 66.0





148 64.1 32.7 98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 147.203 ug/l

RT: 14.694 min Scan# 2166

Delta R.T. -0.006 min

Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion: 75 Resp: 151090


Ion Ratio Lower Upper

75 100

155 78.4 37.3 111.8

157 101.0 46.2 138.6





#93

1,2,4-Trichlorobenzene

Concen: 167.266 ug/l

RT: 15.370 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN087333.D


Acq: 16 Jul 2025 18:54

Instrument :




MSVOA_N

ClientSampleId :

VSTDICC150



Tgt Ion:	180	Resp:	49223
Ion Ratio	Lower	Upper	
180	100		
182	94.3	47.3	141.8
145	32.1	15.4	46.4

Manual Integrations
APPROVEDReviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

#94

Hexachlorobutadiene


Concen: 149.909 ug/l

RT: 15.476 min Scan# 2299



Delta R.T. 0.000 min


Lab File: VN087333.D

Acq: 16 Jul 2025 18:54




Tgt Ion:	225	Resp:	163922
Ion Ratio	Lower	Upper	
225	100		
223	65.1	32.1	96.3
227	63.1	31.3	93.9





#95
Naphthalene
Concen: 180.370 ug/l
RT: 15.612 min Scan# 2322
Delta R.T. -0.006 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150



Tgt Ion:128 Resp: 188042
Ion Ratio Lower Upper
128 100
127 12.9 10.5 15.7
129 10.8 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 166.396 ug/l
RT: 15.812 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087333.D
Acq: 16 Jul 2025 18:54

Tgt Ion:180 Resp: 491195
Ion Ratio Lower Upper
180 100
182 95.2 47.1 141.4
145 34.2 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.212	168	198299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	343165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	319692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	166959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	162194	48.204	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.400%		
35) Dibromofluoromethane	8.153	113	118201	49.934	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.860%		
50) Toluene-d8	10.547	98	435474	51.573	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.140%		
62) 4-Bromofluorobenzene	12.829	95	165160	52.942	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	105.880%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	122218	58.029	ug/l	94
3) Chloromethane	2.383	50	135040	50.986	ug/l	99
4) Vinyl Chloride	2.542	62	141054	53.589	ug/l	93
5) Bromomethane	2.971	94	79496	58.322	ug/l	100
6) Chloroethane	3.130	64	84406	49.172	ug/l	97
7) Trichlorofluoromethane	3.506	101	192022	49.336	ug/l	93
8) Diethyl Ether	3.959	74	76789	50.861	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	98940	49.520	ug/l	99
10) Methyl Iodide	4.577	142	103627	50.369	ug/l	98
11) Tert butyl alcohol	5.530	59	167953	262.884	ug/l	99
12) 1,1-Dichloroethene	4.330	96	102861	45.432	ug/l	97
13) Acrolein	4.171	56	142098	277.145	ug/l	97
14) Allyl chloride	5.006	41	200790	49.004	ug/l	99
15) Acrylonitrile	5.706	53	437956	252.616	ug/l	99
16) Acetone	4.424	43	375335	237.913	ug/l	100
17) Carbon Disulfide	4.700	76	335264	49.947	ug/l	96
18) Methyl Acetate	5.018	43	194657	49.111	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	425099	50.939	ug/l	94
20) Methylene Chloride	5.265	84	131885	49.445	ug/l	93
21) trans-1,2-Dichloroethene	5.771	96	120707	47.283	ug/l	97
22) Diisopropyl ether	6.659	45	439978	51.191	ug/l	97
23) Vinyl Acetate	6.594	43	2052885	273.101	ug/l	98
24) 1,1-Dichloroethane	6.553	63	238854	48.170	ug/l	99
25) 2-Butanone	7.477	43	622581	255.411	ug/l	98
26) 2,2-Dichloropropane	7.477	77	176180	45.700	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	148552	50.543	ug/l	96
28) Bromochloromethane	7.800	49	116494	49.089	ug/l	98
29) Tetrahydrofuran	7.830	42	412353	260.404	ug/l	100
30) Chloroform	7.947	83	239752	48.306	ug/l	100
31) Cyclohexane	8.241	56	202863	49.042	ug/l	96
32) 1,1,1-Trichloroethane	8.153	97	212405	49.412	ug/l	99
36) 1,1-Dichloropropene	8.353	75	167912	53.690	ug/l	99
37) Ethyl Acetate	7.553	43	241759	53.525	ug/l	99
38) Carbon Tetrachloride	8.347	117	181974	52.821	ug/l	90
39) Methylcyclohexane	9.582	83	179820	53.109	ug/l	96
40) Benzene	8.588	78	529446	52.380	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:00:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:56:13 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	127286	53.895	ug/l	98
42) 1,2-Dichloroethane	8.653	62	194804	50.821	ug/l	98
43) Isopropyl Acetate	8.677	43	374957	53.477	ug/l	100
44) Trichloroethene	9.335	130	118775	49.731	ug/l	95
45) 1,2-Dichloropropane	9.606	63	135864	52.900	ug/l	99
46) Dibromomethane	9.688	93	98534	51.241	ug/l	98
47) Bromodichloromethane	9.871	83	200236	51.696	ug/l	97
48) Methyl methacrylate	9.665	41	177068	56.096	ug/l	99
49) 1,4-Dioxane	9.682	88	56917	1177.300	ug/l #	97
51) 4-Methyl-2-Pentanone	10.429	43	1192765	268.966	ug/l	99
52) Toluene	10.612	92	320922	52.236	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	214141	54.628	ug/l	97
54) cis-1,3-Dichloropropene	10.294	75	216023	53.351	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	124090	49.889	ug/l	97
56) Ethyl methacrylate	10.859	69	222410	53.042	ug/l	98
57) 1,3-Dichloropropane	11.147	76	224493	52.202	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	581689	285.088	ug/l	99
59) 2-Hexanone	11.176	43	849057	288.579	ug/l	100
60) Dibromochloromethane	11.341	129	149449	52.685	ug/l	100
61) 1,2-Dibromoethane	11.447	107	135330	51.746	ug/l	94
64) Tetrachloroethene	11.082	164	98391	47.819	ug/l	97
65) Chlorobenzene	11.870	112	347934	48.477	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	123321	50.530	ug/l	99
67) Ethyl Benzene	11.947	91	608779	51.523	ug/l	99
68) m/p-Xylenes	12.053	106	465762	105.269	ug/l	97
69) o-Xylene	12.376	106	224934	53.221	ug/l	99
70) Styrene	12.394	104	387319	54.477	ug/l	98
71) Bromoform	12.559	173	102502	51.987	ug/l #	99
73) Isopropylbenzene	12.676	105	573938	54.619	ug/l	100
74) N-amyl acetate	12.494	43	218170m	49.972	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	203947	51.580	ug/l	98
76) 1,2,3-Trichloropropane	12.976	75	175797m	46.955	ug/l	
77) Bromobenzene	12.959	156	142714	52.368	ug/l	99
78) n-propylbenzene	13.012	91	709437	53.661	ug/l	99
79) 2-Chlorotoluene	13.106	91	426373	52.475	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	484100	54.071	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	69894	51.080	ug/l	97
82) 4-Chlorotoluene	13.200	91	443817	52.464	ug/l	99
83) tert-Butylbenzene	13.417	119	413882	55.349	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	500792	54.773	ug/l	99
85) sec-Butylbenzene	13.594	105	594712	52.800	ug/l	98
86) p-Isopropyltoluene	13.706	119	496291	54.982	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	277102	51.809	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	282084	49.381	ug/l	99
89) n-Butylbenzene	14.035	91	458964	53.250	ug/l	99
90) Hexachloroethane	14.312	117	93464	48.870	ug/l	95
91) 1,2-Dichlorobenzene	14.088	146	260874	51.485	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	50904	49.035	ug/l	97
93) 1,2,4-Trichlorobenzene	15.370	180	155818	52.351	ug/l	99
94) Hexachlorobutadiene	15.476	225	55599	50.273	ug/l	98
95) Naphthalene	15.617	128	586097	55.585	ug/l	98
96) 1,2,3-Trichlorobenzene	15.811	180	157909	52.890	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Manual Integrations APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:168 Resp: 198299
Ion Ratio Lower Upper
168 100
99 57.5 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#2
Dichlorodifluoromethane
Concen: 58.029 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 85 Resp: 122218
Ion Ratio Lower Upper
85 100
87 32.0 17.8 53.3









#3
Chloromethane
Concen: 50.986 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Tgt Ion: 50 Resp: 135040
Ion Ratio Lower Upper
50 100
52 33.3 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#4
Vinyl Chloride
Concen: 53.589 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 62 Resp: 141054
Ion Ratio Lower Upper
62 100
64 30.0 27.0 40.6





#5

Bromomethane

Concen: 58.322 ug/l

RT: 2.971 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 94 Resp: 79490

Ion Ratio Lower Upper

94 100





96 92.2 73.4 110.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#6

Chloroethane

Concen: 49.172 ug/l

RT: 3.130 min Scan# 200

Delta R.T. 0.000 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 64 Resp: 84406

Ion Ratio Lower Upper

64 100

66 32.5 24.6 36.8









#7
Trichlorofluoromethane
Concen: 49.336 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Tgt Ion:101 Resp: 19202
Ion Ratio Lower Upper
101 100
103 67.6 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#8
Diethyl Ether
Concen: 50.861 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 74 Resp: 76789
Ion Ratio Lower Upper
74 100
45 99.6 50.8 152.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.520 ug/l

RT: 4.365 min Scan# 410

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :





ICVVN071625

Tgt Ion:	101	Resp:	98940
Ion	Ratio	Lower	Upper
101	100		
85	46.6	37.3	55.9
151	72.5	58.9	88.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#10

Methyl Iodide

Concen: 50.369 ug/l




RT: 4.577 min Scan# 446


Delta R.T. -0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Tgt Ion:	142	Resp:	103627
Ion	Ratio	Lower	Upper
142	100		
127	46.5	35.7	53.5
141	12.6	10.4	15.6





#11

Tert butyl alcohol

Concen: 262.884 ug/l

RT: 5.530 min Scan# 607

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 59 Resp: 16795

Ion Ratio Lower Upper

59 100





57 10.0 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#12

1,1-Dichloroethene

Concen: 45.432 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 96 Resp: 102861


Ion Ratio Lower Upper

96 100

61 165.3 132.3 198.5

98 66.1 46.8 70.2





#13

Acrolein

Concen: 277.145 ug/l

RT: 4.171 min Scan# 31

Delta R.T. -0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 56 Resp: 142098

Ion Ratio Lower Upper

56 100





55 72.9 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#14

Allyl chloride

Concen: 49.004 ug/l

RT: 5.006 min Scan# 519

Delta R.T. -0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 41 Resp: 200790


Ion Ratio Lower Upper

41 100

39 72.4 59.0 88.6

76 35.6 28.7 43.1





#15

Acrylonitrile

Concen: 252.616 ug/l

RT: 5.706 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN087335.D


Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625



Tgt Ion: 53 Resp: 437950

Ion Ratio Lower Upper

53 100

52 82.7 65.5 98.3




51 36.6 28.7 43.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#16

Acetone

Concen: 237.913 ug/l

RT: 4.424 min Scan# 420

Delta R.T. 0.000 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 43 Resp: 375335

Ion Ratio Lower Upper

43 100

58 28.1 22.3 33.5





#17

Carbon Disulfide

Concen: 49.947 ug/l

RT: 4.700 min Scan# 467

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 76 Resp: 335264

Ion Ratio Lower Upper

76 100





78 9.3 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#18

Methyl Acetate

Concen: 49.111 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.006 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 43 Resp: 194657

Ion Ratio Lower Upper

43 100

74 22.5 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 50.939 ug/l

RT: 5.789 min Scan# 652

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 73 Resp: 425099

Ion Ratio Lower Upper

73 100





57 20.2 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#20

Methylene Chloride

Concen: 49.445 ug/l

RT: 5.265 min Scan# 563

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Tgt Ion: 84 Resp: 131885




Ion Ratio Lower Upper


84 100

49 124.5 107.5 161.3

51 41.3 30.2 45.2

86 59.3 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 47.283 ug/l
RT: 5.771 min Scan# 649
Delta R.T. -0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Tgt Ion: 96 Resp: 12070
Ion Ratio Lower Upper
96 100
61 148.5 122.0 183.0
98 61.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#22
Diisopropyl ether
Concen: 51.191 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 45 Resp: 439978
Ion Ratio Lower Upper
45 100
43 55.2 42.8 64.2
87 26.3 19.8 29.6
59 12.3 8.3 12.5





#23

Vinyl Acetate

Concen: 273.101 ug/l

RT: 6.594 min Scan# 789

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 43 Resp: 205288

Ion Ratio Lower Upper

43 100





86 10.2 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#24

1,1-Dichloroethane

Concen: 48.170 ug/l

RT: 6.553 min Scan# 782

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 63 Resp: 238854


Ion Ratio Lower Upper

63 100

98 6.4 3.3 9.9

100 4.6 2.5 7.4





#25

2-Butanone

Concen: 255.411 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 43 Resp: 62258

Ion Ratio Lower Upper

43 100





72 25.4 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#26

2,2-Dichloropropane

Concen: 45.700 ug/l

RT: 7.477 min Scan# 939

Delta R.T. 0.006 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 77 Resp: 176180

Ion Ratio Lower Upper

77 100

97 22.2 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 50.543 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 96 Resp: 148552

Ion Ratio Lower Upper

96 100

61 144.0 0.0 297.8





98 63.7 0.0 132.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#28

Bromochloromethane

Concen: 49.089 ug/l

RT: 7.800 min Scan# 994

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 49 Resp: 116494


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.2

130 69.6 57.3 85.9





#29

Tetrahydrofuran

Concen: 260.404 ug/l

RT: 7.830 min Scan# 94

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N





ClientSampleId :

ICVVN071625

Tgt Ion:	42	Resp:	41235
Ion	Ratio	Lower	Upper
42	100		
72	41.3	33.4	50.0
71	38.9	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#30

Chloroform

Concen: 48.306 ug/l




RT: 7.947 min Scan# 1019


Delta R.T. -0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Tgt Ion:	83	Resp:	239752
Ion	Ratio	Lower	Upper
83	100		
85	65.9	52.7	79.1





#31

Cyclohexane

Concen: 49.042 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 56 Resp: 20286

Ion Ratio Lower Upper

56 100

69 30.0 22.7 34.1





84 85.9 65.8 98.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#32

1,1,1-Trichloroethane

Concen: 49.412 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 97 Resp: 212405


Ion Ratio Lower Upper

97 100

99 64.8 51.8 77.8

61 46.8 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 48.204 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 65 Resp: 162194

Ion Ratio Lower Upper

65 100





67 51.3 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion:114 Resp: 343165


Ion Ratio Lower Upper

114 100

63 22.6 0.0 44.6

88 15.4 0.0 32.8





#35

Dibromofluoromethane

Concen: 49.934 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion:113 Resp: 11820

Ion Ratio Lower Upper

113 100

111 107.2 82.5 123.7





192 17.6 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#36

1,1-Dichloropropene

Concen: 53.690 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 75 Resp: 167912


Ion Ratio Lower Upper

75 100

110 32.7 16.7 50.1

77 30.7 25.2 37.8





#37

Ethyl Acetate

Concen: 53.525 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 43 Resp: 241759

Ion Ratio Lower Upper

43 100

61 12.9 10.9 16.3





70 9.6 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#38

Carbon Tetrachloride

Concen: 52.821 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion:117 Resp: 181974


Ion Ratio Lower Upper

117 100

119 90.1 80.2 120.2

121 27.5 25.4 38.2





#39

Methylcyclohexane

Concen: 53.109 ug/l

RT: 9.582 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 83 Resp: 179820

Ion Ratio Lower Upper

83 100

55 79.9 61.3 91.9





98 50.2 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#40

Benzene

Concen: 52.380 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 78 Resp: 529446

Ion Ratio Lower Upper

78 100


77 24.4 18.2 27.2





#41
Methacrylonitrile
Concen: 53.895 ug/l
RT: 7.765 min Scan# 989
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion: 41 Resp: 127280
Ion Ratio Lower Upper
41 100
39 53.9 43.4 65.0
67 71.0 55.1 82.7
52 30.0 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#42
1,2-Dichloroethane
Concen: 50.821 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 62 Resp: 194804
Ion Ratio Lower Upper
62 100
98 8.2 0.0 18.0





#43

Isopropyl Acetate

Concen: 53.477 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087335.D


Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :




ICVVN071625



Tgt Ion:	43	Resp:	37495
Ion	Ratio	Lower	Upper
43	100		
61	24.9	19.8	29.8
87	12.1	9.8	14.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#44

Trichloroethene


Concen: 49.731 ug/l

RT: 9.335 min Scan# 1255



Delta R.T. 0.000 min


Lab File: VN087335.D

Acq: 16 Jul 2025 19:59



Tgt Ion:	130	Resp:	118775
Ion	Ratio	Lower	Upper
130	100		
95	102.8	0.0	195.2





#45

1,2-Dichloropropane

Concen: 52.900 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 63 Resp: 135864

Ion Ratio Lower Upper

63 100





65 34.0 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#46

Dibromomethane

Concen: 51.241 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 93 Resp: 98534


Ion Ratio Lower Upper

93 100

95 79.2 65.8 98.8

174 86.5 69.9 104.9





#47

Bromodichloromethane

Concen: 51.696 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N





ClientSampleId :

ICVVN071625

Tgt Ion:	83	Resp:	200230
Ion	Ratio	Lower	Upper
83	100		
85	60.5	50.4	75.6
127	8.2	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#48

Methyl methacrylate

Concen: 56.096 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Tgt Ion:	41	Resp:	177068
Ion	Ratio	Lower	Upper
41	100		
69	79.9	64.1	96.1
39	56.0	45.5	68.3





#49

1,4-Dioxane

Concen: 1177.300 ug/l

RT: 9.682 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N





ClientSampleId :

ICVVN071625

Tgt Ion:	88	Resp:	5691
Ion	Ratio	Lower	Upper
88	100		
43	0.0	0.0	0.0
58	73.7	61.1	91.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#50

Toluene-d8

Concen: 51.573 ug/l




RT: 10.547 min Scan# 1461


Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59


Tgt Ion:	98	Resp:	435474
Ion	Ratio	Lower	Upper
98	100		
100	68.2	52.1	78.1





#51
4-Methyl-2-Pentanone
Concen: 268.966 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion: 43 Resp: 119276
Ion Ratio Lower Upper
43 100
58 38.0 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#52
Toluene
Concen: 52.236 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59


Tgt Ion: 92 Resp: 320922
Ion Ratio Lower Upper
92 100
91 173.1 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 54.628 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion: 75 Resp: 21414
Ion Ratio Lower Upper
75 100
77 30.9 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#54
cis-1,3-Dichloropropene
Concen: 53.351 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 75 Resp: 216023
Ion Ratio Lower Upper
75 100
77 33.0 24.2 36.2
39 50.8 38.4 57.6





#55

1,1,2-Trichloroethane

Concen: 49.889 ug/l

RT: 11.000 min Scan# 1538

Delta R.T. -0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 97 Resp: 124090

Ion Ratio Lower Upper

97 100

83 88.7 72.6 109.0

85 57.4 46.7 70.1





99 61.9 54.1 81.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#56

Ethyl methacrylate

Concen: 53.042 ug/l

RT: 10.859 min Scan# 1514

Delta R.T. -0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 69 Resp: 222410


Ion Ratio Lower Upper

69 100

41 67.6 55.1 82.7

39 34.0 27.9 41.9





#57

1,3-Dichloropropane

Concen: 52.202 ug/l

RT: 11.147 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 76 Resp: 22449

Ion Ratio Lower Upper

76 100





78 31.9 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#58

2-Chloroethyl Vinyl ether

Concen: 285.088 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion: 63 Resp: 581689

Ion Ratio Lower Upper

63 100


106 27.5 21.7 32.5





#59
2-Hexanone
Concen: 288.579 ug/l
RT: 11.176 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion: 43 Resp: 84905
Ion Ratio Lower Upper
43 100
58 53.4 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#60
Dibromochloromethane
Concen: 52.685 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59


Tgt Ion:129 Resp: 149449
Ion Ratio Lower Upper
129 100
127 78.0 39.1 117.5





#61
1,2-Dibromoethane
Concen: 51.746 ug/l
RT: 11.447 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:107 Resp: 135330
Ion Ratio Lower Upper
107 100
109 91.4 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#62
4-Bromofluorobenzene
Concen: 52.942 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59


Tgt Ion: 95 Resp: 165160
Ion Ratio Lower Upper
95 100
174 71.8 0.0 149.4
176 70.6 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1182
Delta R.T. -0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:117 Resp: 319692
Ion Ratio Lower Upper
117 100
82 57.6 47.4 71.2
119 31.0 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#64
Tetrachloroethene
Concen: 47.819 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion:164 Resp: 98391
Ion Ratio Lower Upper
164 100
166 127.0 105.5 158.3
129 99.0 77.4 116.2
131 93.7 77.3 115.9









#65
Chlorobenzene
Concen: 48.477 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Tgt Ion:112 Resp: 347934
Ion Ratio Lower Upper
112 100
114 31.8 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.530 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion:131 Resp: 123321
Ion Ratio Lower Upper
131 100
133 95.5 47.4 142.3
119 65.0 33.1 99.2





#67

Ethyl Benzene

Concen: 51.523 ug/l

RT: 11.947 min Scan# 1699

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 91 Resp: 608779

Ion Ratio Lower Upper

91 100





106 30.9 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#68

m/p-Xylenes

Concen: 105.269 ug/l

RT: 12.053 min Scan# 1717

Delta R.T. 0.000 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion:106 Resp: 465762

Ion Ratio Lower Upper

106 100


91 206.7 162.0 243.0





#69
o-Xylene
Concen: 53.221 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:106 Resp: 224934
Ion Ratio Lower Upper
106 100
91 217.3 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#70
Styrene
Concen: 54.477 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59


Tgt Ion:104 Resp: 387319
Ion Ratio Lower Upper
104 100
78 53.1 41.0 61.6
103 54.5 43.9 65.9





#71
Bromoform
Concen: 51.987 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625




Tgt Ion:173 Resp: 10250
Ion Ratio Lower Upper
173 100
175 49.0 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Tgt Ion:152 Resp: 166959
Ion Ratio Lower Upper
152 100
115 60.2 31.1 93.5
150 174.4 0.0 349.0





#73
Isopropylbenzene
Concen: 54.619 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625




Tgt Ion:105 Resp: 573938
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#74
N-ethyl acetate
Concen: 49.972 ug/l m
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Tgt Ion: 43 Resp: 218170
Ion Ratio Lower Upper
43 100
70 37.9 37.6 56.4
55 19.4 19.6 29.4#
61 21.6 20.6 31.0





#75
1,1,2,2-Tetrachloroethane
Concen: 51.580 ug/l
RT: 12.917 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625




Tgt Ion: 83 Resp: 20394
Ion Ratio Lower Upper
83 100
131 10.2 5.1 15.3
85 63.4 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025




#76
1,2,3-Trichloropropane
Concen: 46.955 ug/l m
RT: 12.976 min Scan# 1874
Delta R.T. 0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Tgt Ion: 75 Resp: 175797
Ion Ratio Lower Upper
75 100
77 187.3 94.5 283.6





#77
Bromobenzene
Concen: 52.368 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:156 Resp: 142714
Ion Ratio Lower Upper
156 100
77 230.7 114.9 344.6
158 98.1 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#78
n-propylbenzene
Concen: 53.661 ug/l
RT: 13.012 min Scan# 1880
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 91 Resp: 709437
Ion Ratio Lower Upper
91 100
120 21.9 11.3 33.8





#79

2-Chlorotoluene

Concen: 52.475 ug/l

RT: 13.106 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion: 91 Resp: 42637

Ion Ratio Lower Upper

91 100





126 33.1 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#80

1,3,5-Trimethylbenzene

Concen: 54.071 ug/l

RT: 13.153 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VN087335.D




Acq: 16 Jul 2025 19:59


Tgt Ion:105 Resp: 484100

Ion Ratio Lower Upper

105 100


120 49.5 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 51.080 ug/l
RT: 12.717 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion: 75 Resp: 69894
Ion Ratio Lower Upper
75 100
53 100.2 83.5 125.3
89 48.6 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#82
4-Chlorotoluene
Concen: 52.464 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion: 91 Resp: 443817
Ion Ratio Lower Upper
91 100
126 32.4 16.6 49.7









#83
tert-Butylbenzene
Concen: 55.349 ug/l
RT: 13.417 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Tgt Ion:119 Resp: 41388
Ion Ratio Lower Upper
119 100
91 69.7 35.0 105.1
134 27.6 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#84
1,2,4-Trimethylbenzene
Concen: 54.773 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion:105 Resp: 500792
Ion Ratio Lower Upper
105 100
120 46.0 22.8 68.3





#85

sec-Butylbenzene

Concen: 52.800 ug/l

RT: 13.594 min Scan# 1979

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion:105 Resp: 594712

Ion Ratio Lower Upper

105 100





134 20.4 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#86

p-Isopropyltoluene

Concen: 54.982 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion:119 Resp: 496291


Ion Ratio Lower Upper

119 100

134 26.5 13.5 40.5

91 24.6 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 51.809 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion:146 Resp: 277102

Ion Ratio Lower Upper

146 100

111 42.0 21.4 64.3





148 62.5 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#88

1,4-Dichlorobenzene

Concen: 49.381 ug/l

RT: 13.788 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion:146 Resp: 282084


Ion Ratio Lower Upper

146 100

111 40.5 19.6 58.7

148 63.2 31.4 94.0





#89

n-Butylbenzene

Concen: 53.250 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N





ClientSampleId :

ICVVN071625

Tgt Ion:	91	Resp:	458964
Ion	Ratio	Lower	Upper
91	100		
92	53.0	26.2	78.6
134	25.7	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#90

Hexachloroethane

Concen: 48.870 ug/l




RT: 14.312 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Tgt Ion:	117	Resp:	93464
Ion	Ratio	Lower	Upper
117	100		
201	69.6	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 51.485 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion:146 Resp: 260874

Ion Ratio Lower Upper

146 100

111 43.4 22.0 66.0





148 64.0 32.7 98.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 49.035 ug/l

RT: 14.700 min Scan# 2167

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion: 75 Resp: 50904


Ion Ratio Lower Upper

75 100

155 72.3 37.3 111.8

157 95.1 46.2 138.6





#93

1,2,4-Trichlorobenzene

Concen: 52.351 ug/l

RT: 15.370 min Scan# 2281

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Tgt Ion:180 Resp: 155818

Ion Ratio Lower Upper

180 100

182 94.5 47.3 141.8





145 33.3 15.4 46.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025



#94

Hexachlorobutadiene

Concen: 50.273 ug/l

RT: 15.476 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VN087335.D

Acq: 16 Jul 2025 19:59




Tgt Ion:225 Resp: 55599


Ion Ratio Lower Upper

225 100

223 62.7 32.1 96.3


227 64.6 31.3 93.9





#95
Naphthalene
Concen: 55.585 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59




Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625



Tgt Ion:128 Resp: 58609
Ion Ratio Lower Upper
128 100
127 12.5 10.5 15.7
129 11.1 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



#96
1,2,3-Trichlorobenzene
Concen: 52.890 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087335.D
Acq: 16 Jul 2025 19:59

Tgt Ion:180 Resp: 157909
Ion Ratio Lower Upper
180 100
182 94.8 47.1 141.4
145 34.2 16.9 50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.531	0.616	-16.0	107	0.00
3 P	Chloromethane	0.668	0.681	-1.9	107	0.00
4 C	Vinyl Chloride	0.664	0.711	-7.1#	106	0.00
5 T	Bromomethane	0.344	0.401	-16.6	122	0.00
6 T	Chloroethane	0.433	0.426	1.6	105	0.00
7 T	Trichlorofluoromethane	0.981	0.968	1.3	103	0.00
8 T	Diethyl Ether	0.381	0.387	-1.6	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.499	1.0	103	0.00
10 T	Methyl Iodide	0.452	0.523	-15.7	115	0.00
11 T	Tert butyl alcohol	0.161	0.169	-5.0	114	0.00
12 CM	1,1-Dichloroethene	0.571	0.519	9.1#	103	0.00
13 T	Acrolein	0.129	0.143	-10.9	128	0.00
14 T	Allyl chloride	1.033	1.013	1.9	110	0.00
15 T	Acrylonitrile	0.437	0.442	-1.1	106	0.00
16 T	Acetone	0.398	0.379	4.8	106	0.00
17 T	Carbon Disulfide	1.693	1.691	0.1	106	0.00
18 T	Methyl Acetate	0.999	0.982	1.7	107	0.00
19 T	Methyl tert-butyl Ether	2.104	2.144	-1.9	107	0.00
20 T	Methylene Chloride	0.766	0.665	13.2	107	0.00
21 T	trans-1,2-Dichloroethene	0.644	0.609	5.4	101	0.00
22 T	Diisopropyl ether	2.167	2.219	-2.4	105	0.00
23 T	Vinyl Acetate	1.895	2.070	-9.2	106	0.00
24 P	1,1-Dichloroethane	1.250	1.205	3.6	107	0.00
25 T	2-Butanone	0.615	0.628	-2.1	106	0.00
26 T	2,2-Dichloropropane	0.972	0.888	8.6	98	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.749	-1.1	106	0.00
28 T	Bromochloromethane	0.598	0.587	1.8	107	0.00
29 T	Tetrahydrofuran	0.399	0.416	-4.3	106	0.00
30 C	Chloroform	1.251	1.209	3.4#	103	0.00
31 T	Cyclohexane	1.043	1.023	1.9	106	0.00
32 T	1,1,1-Trichloroethane	1.084	1.071	1.2	107	0.00
33 S	1,2-Dichloroethane-d4	0.848	0.818	3.5	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.345	0.344	0.3	109	0.00
36 T	1,1-Dichloropropene	0.456	0.489	-7.2	106	0.00
37 T	Ethyl Acetate	0.658	0.704	-7.0	105	0.00
38 T	Carbon Tetrachloride	0.502	0.530	-5.6	108	0.00
39 T	Methylcyclohexane	0.493	0.524	-6.3	104	0.00
40 TM	Benzene	1.473	1.543	-4.8	105	0.00
41 T	Methacrylonitrile	0.344	0.371	-7.8	106	0.00
42 TM	1,2-Dichloroethane	0.558	0.568	-1.8	105	0.00
43 T	Isopropyl Acetate	1.022	1.093	-6.9	108	0.00
44 TM	Trichloroethene	0.348	0.346	0.6	102	0.00
45 C	1,2-Dichloropropane	0.374	0.396	-5.9#	106	0.00
46 T	Dibromomethane	0.280	0.287	-2.5	105	0.00
47 T	Bromodichloromethane	0.564	0.583	-3.4	108	0.00
48 T	Methyl methacrylate	0.460	0.516	-12.2	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	109	0.00
50 S	Toluene-d8	1.230	1.269	-3.2	109	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.695	-7.6	106	0.00
52 CM	Toluene	0.895	0.935	-4.5#	102	0.00
53 T	t-1,3-Dichloropropene	0.571	0.624	-9.3	106	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.630	-6.8	105	0.00
55 T	1,1,2-Trichloroethane	0.362	0.362	0.0	104	0.00
56 T	Ethyl methacrylate	0.552	0.648	-17.4	109	0.00
57 T	1,3-Dichloropropane	0.627	0.654	-4.3	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.339	-14.1	103	0.00
59 T	2-Hexanone	0.429	0.495	-15.4	105	0.00
60 T	Dibromochloromethane	0.413	0.436	-5.6	107	0.00
61 T	1,2-Dibromoethane	0.381	0.394	-3.4	108	0.00
62 S	4-Bromofluorobenzene	0.455	0.481	-5.7	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.322	0.308	4.3	102	0.00
65 PM	Chlorobenzene	1.123	1.088	3.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.386	-1.0	105	0.00
67 C	Ethyl Benzene	1.848	1.904	-3.0#	104	0.00
68 T	m/p-Xylenes	0.692	0.728	-5.2	102	0.00
69 T	o-Xylene	0.661	0.704	-6.5	104	0.00
70 T	Styrene	1.112	1.212	-9.0	103	0.00
71 P	Bromoform	0.308	0.321	-4.2	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.147	3.438	-9.2	107	0.00
74 T	N-amyl acetate	1.307	1.307	0.0	121	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.222	-3.2	107	0.00
76 T	1,2,3-Trichloropropane	1.121	1.053	6.1	93	0.00
77 T	Bromobenzene	0.816	0.855	-4.8	104	0.00
78 T	n-propylbenzene	3.959	4.249	-7.3	104	0.00
79 T	2-Chlorotoluene	2.433	2.554	-5.0	106	0.00
80 T	1,3,5-Trimethylbenzene	2.681	2.900	-8.2	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.410	0.419	-2.2	109	0.00
82 T	4-Chlorotoluene	2.533	2.658	-4.9	106	0.00
83 T	tert-Butylbenzene	2.239	2.479	-10.7	108	0.00
84 T	1,2,4-Trimethylbenzene	2.738	2.999	-9.5	105	0.00
85 T	sec-Butylbenzene	3.373	3.562	-5.6	105	0.00
86 T	p-Isopropyltoluene	2.703	2.973	-10.0	105	0.00
87 T	1,3-Dichlorobenzene	1.602	1.660	-3.6	105	0.00
88 T	1,4-Dichlorobenzene	1.711	1.690	1.2	105	0.00
89 T	n-Butylbenzene	2.581	2.749	-6.5	105	0.00
90 T	Hexachloroethane	0.573	0.560	2.3	102	0.00
91 T	1,2-Dichlorobenzene	1.517	1.563	-3.0	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.311	0.305	1.9	106	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.933	-4.7	105	0.00
94 T	Hexachlorobutadiene	0.331	0.333	-0.6	107	0.00
95 T	Naphthalene	3.158	3.510	-11.1	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICV VN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.894	0.946	-5.8	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	58.029	-16.1	107	0.00
3 P	Chloromethane	50.000	50.986	-2.0	107	0.00
4 C	Vinyl Chloride	50.000	53.589	-7.2#	106	0.00
5 T	Bromomethane	50.000	58.322	-16.6	122	0.00
6 T	Chloroethane	50.000	49.172	1.7	105	0.00
7 T	Trichlorofluoromethane	50.000	49.336	1.3	103	0.00
8 T	Diethyl Ether	50.000	50.861	-1.7	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.520	1.0	103	0.00
10 T	Methyl Iodide	50.000	50.369	-0.7	115	0.00
11 T	Tert butyl alcohol	250.000	262.884	-5.2	114	0.00
12 CM	1,1-Dichloroethene	50.000	45.432	9.1#	103	0.00
13 T	Acrolein	250.000	277.145	-10.9	128	0.00
14 T	Allyl chloride	50.000	49.004	2.0	110	0.00
15 T	Acrylonitrile	250.000	252.616	-1.0	106	0.00
16 T	Acetone	250.000	237.913	4.8	106	0.00
17 T	Carbon Disulfide	50.000	49.947	0.1	106	0.00
18 T	Methyl Acetate	50.000	49.111	1.8	107	0.00
19 T	Methyl tert-butyl Ether	50.000	50.939	-1.9	107	0.00
20 T	Methylene Chloride	50.000	49.445	1.1	107	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.283	5.4	101	0.00
22 T	Diisopropyl ether	50.000	51.191	-2.4	105	0.00
23 T	Vinyl Acetate	250.000	273.101	-9.2	106	0.00
24 P	1,1-Dichloroethane	50.000	48.170	3.7	107	0.00
25 T	2-Butanone	250.000	255.411	-2.2	106	0.00
26 T	2,2-Dichloropropane	50.000	45.700	8.6	98	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.543	-1.1	106	0.00
28 T	Bromochloromethane	50.000	49.089	1.8	107	0.00
29 T	Tetrahydrofuran	250.000	260.404	-4.2	106	0.00
30 C	Chloroform	50.000	48.306	3.4#	103	0.00
31 T	Cyclohexane	50.000	49.042	1.9	106	0.00
32 T	1,1,1-Trichloroethane	50.000	49.412	1.2	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.204	3.6	111	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35 S	Dibromofluoromethane	50.000	49.934	0.1	109	0.00
36 T	1,1-Dichloropropene	50.000	53.690	-7.4	106	0.00
37 T	Ethyl Acetate	50.000	53.525	-7.0	105	0.00
38 T	Carbon Tetrachloride	50.000	52.821	-5.6	108	0.00
39 T	Methylcyclohexane	50.000	53.109	-6.2	104	0.00
40 TM	Benzene	50.000	52.380	-4.8	105	0.00
41 T	Methacrylonitrile	50.000	53.895	-7.8	106	0.00
42 TM	1,2-Dichloroethane	50.000	50.821	-1.6	105	0.00
43 T	Isopropyl Acetate	50.000	53.477	-7.0	108	0.00
44 TM	Trichloroethene	50.000	49.731	0.5	102	0.00
45 C	1,2-Dichloropropane	50.000	52.900	-5.8#	106	0.00
46 T	Dibromomethane	50.000	51.241	-2.5	105	0.00
47 T	Bromodichloromethane	50.000	51.696	-3.4	108	0.00
48 T	Methyl methacrylate	50.000	56.096	-12.2	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1177.300	-17.7	109	0.00
50 S	Toluene-d8	50.000	51.573	-3.1	109	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.966	-7.6	106	0.00
52 CM	Toluene	50.000	52.236	-4.5#	102	0.00
53 T	t-1,3-Dichloropropene	50.000	54.628	-9.3	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.351	-6.7	105	0.00
55 T	1,1,2-Trichloroethane	50.000	49.889	0.2	104	0.00
56 T	Ethyl methacrylate	50.000	53.042	-6.1	109	0.00
57 T	1,3-Dichloropropane	50.000	52.202	-4.4	105	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	285.088	-14.0	103	0.00
59 T	2-Hexanone	250.000	288.579	-15.4	105	0.00
60 T	Dibromochloromethane	50.000	52.685	-5.4	107	0.00
61 T	1,2-Dibromoethane	50.000	51.746	-3.5	108	0.00
62 S	4-Bromofluorobenzene	50.000	52.942	-5.9	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	47.819	4.4	102	0.00
65 PM	Chlorobenzene	50.000	48.477	3.0	104	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.530	-1.1	105	0.00
67 C	Ethyl Benzene	50.000	51.523	-3.0#	104	0.00
68 T	m/p-Xylenes	100.000	105.269	-5.3	102	0.00
69 T	o-Xylene	50.000	53.221	-6.4	104	0.00
70 T	Styrene	50.000	54.477	-9.0	103	0.00
71 P	Bromoform	50.000	51.987	-4.0	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	54.619	-9.2	107	0.00
74 T	N-ethyl acetate	50.000	49.972	0.1	121	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.580	-3.2	107	0.00
76 T	1,2,3-Trichloropropane	50.000	46.955	6.1	93	0.00
77 T	Bromobenzene	50.000	52.368	-4.7	104	0.00
78 T	n-propylbenzene	50.000	53.661	-7.3	104	0.00
79 T	2-Chlorotoluene	50.000	52.475	-5.0	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.071	-8.1	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.080	-2.2	109	0.00
82 T	4-Chlorotoluene	50.000	52.464	-4.9	106	0.00
83 T	tert-Butylbenzene	50.000	55.349	-10.7	108	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.773	-9.5	105	0.00
85 T	sec-Butylbenzene	50.000	52.800	-5.6	105	0.00
86 T	p-Isopropyltoluene	50.000	54.982	-10.0	105	0.00
87 T	1,3-Dichlorobenzene	50.000	51.809	-3.6	105	0.00
88 T	1,4-Dichlorobenzene	50.000	49.381	1.2	105	0.00
89 T	n-Butylbenzene	50.000	53.250	-6.5	105	0.00
90 T	Hexachloroethane	50.000	48.870	2.3	102	0.00
91 T	1,2-Dichlorobenzene	50.000	51.485	-3.0	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.035	1.9	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.351	-4.7	105	0.00
94 T	Hexachlorobutadiene	50.000	50.273	-0.5	107	0.00
95 T	Naphthalene	50.000	55.585	-11.2	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.890	-5.8	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2842</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/14/2025</u> <u>09:34</u>
Lab File ID:	<u>VN087552.D</u>	Init. Calib. Date(s):	<u>07/16/2025</u> <u>07/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>17:05</u> <u>18:54</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.678		2.11	20
1,1-Dichloroethene	0.571	0.562		-1.58	20
1,1-Dichloroethane	1.250	1.280	0.1	2.4	20
cis-1,2-Dichloroethene	0.741	0.782		5.53	20
1,1,1-Trichloroethane	1.084	1.139		5.07	20
Benzene	1.473	1.404		-4.68	20
1,2-Dichloroethane	0.558	0.573		2.69	20
Trichloroethene	0.348	0.305		-12.36	20
1,1,2-Trichloroethane	0.362	0.350		-3.32	20
Tetrachloroethene	0.322	0.260		-19.25	20
1,2-Dichloroethane-d4	0.848	0.912		7.55	20
Dibromofluoromethane	0.345	0.310		-10.15	20
Toluene-d8	1.230	1.125		-8.54	20
4-Bromofluorobenzene	0.455	0.433		-4.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 08/14/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.212	168	271352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	543098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	500619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	255173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	247424	53.738	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.480%			
35) Dibromofluoromethane	8.153	113	168221	44.903	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 89.800%			
50) Toluene-d8	10.547	98	611058	45.726	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 91.460%			
62) 4-Bromofluorobenzene	12.829	95	235280	47.655	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.300%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	160120	55.557	ug/l	96
3) Chloromethane	2.383	50	169400	46.740	ug/l	99
4) Vinyl Chloride	2.542	62	183878	51.052	ug/l	96
5) Bromomethane	2.965	94	102629	55.023	ug/l	98
6) Chloroethane	3.130	64	126750	53.961	ug/l	97
7) Trichlorofluoromethane	3.506	101	267426	50.212	ug/l	97
8) Diethyl Ether	3.965	74	117073	56.667	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	145784	53.322	ug/l	96
10) Methyl Iodide	4.577	142	103333	38.232	ug/l #	89
11) Tert butyl alcohol	5.536	59	241449	276.178	ug/l	100
12) 1,1-Dichloroethene	4.336	96	152389	49.187	ug/l	99
13) Acrolein	4.177	56	160450	228.690	ug/l	99
14) Allyl chloride	5.012	41	286033	51.014	ug/l	90
15) Acrylonitrile	5.712	53	598909	252.452	ug/l	98
16) Acetone	4.424	43	657966	304.782	ug/l	97
17) Carbon Disulfide	4.700	76	405663	44.164	ug/l	98
18) Methyl Acetate	5.018	43	317943	58.620	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	659220	57.727	ug/l	95
20) Methylene Chloride	5.259	84	192849	52.881	ug/l	97
21) trans-1,2-Dichloroethene	5.777	96	170971	48.942	ug/l	88
22) Diisopropyl ether	6.659	45	674122	57.318	ug/l	95
23) Vinyl Acetate	6.589	43	3071299	298.585	ug/l	100
24) 1,1-Dichloroethane	6.553	63	347405	51.200	ug/l	98
25) 2-Butanone	7.471	43	892036	267.432	ug/l	97
26) 2,2-Dichloropropane	7.471	77	298522	56.588	ug/l	98
27) cis-1,2-Dichloroethene	7.477	96	212209	52.764	ug/l	95
28) Bromochloromethane	7.794	49	184324	56.761	ug/l	92
29) Tetrahydrofuran	7.824	42	572583	264.243	ug/l	100
30) Chloroform	7.947	83	372180	54.800	ug/l	98
31) Cyclohexane	8.235	56	278777	49.251	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	309142	52.554	ug/l	98
36) 1,1-Dichloropropene	8.353	75	237939	48.073	ug/l	97
37) Ethyl Acetate	7.547	43	340534	47.639	ug/l	98
38) Carbon Tetrachloride	8.347	117	246721	45.251	ug/l	97
39) Methylcyclohexane	9.582	83	254444	47.484	ug/l	96
40) Benzene	8.588	78	762513	47.666	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/14/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	178825	47.844	ug/l	96
42) 1,2-Dichloroethane	8.653	62	311061	51.277	ug/l	97
43) Isopropyl Acetate	8.671	43	575331	51.848	ug/l	98
44) Trichloroethene	9.335	130	165523	43.791	ug/l	95
45) 1,2-Dichloropropane	9.606	63	200161	49.245	ug/l	92
46) Dibromomethane	9.694	93	147474	48.459	ug/l	96
47) Bromodichloromethane	9.871	83	298595	48.711	ug/l	100
48) Methyl methacrylate	9.665	41	263753	52.798	ug/l	93
49) 1,4-Dioxane	9.682	88	77554	1013.618	ug/l #	99
51) 4-Methyl-2-Pentanone	10.429	43	1771617	252.428	ug/l	98
52) Toluene	10.612	92	472630	48.609	ug/l	95
53) t-1,3-Dichloropropene	10.818	75	316497	51.017	ug/l	96
54) cis-1,3-Dichloropropene	10.294	75	326996	51.028	ug/l #	86
55) 1,1,2-Trichloroethane	11.000	97	190088	48.289	ug/l	95
56) Ethyl methacrylate	10.859	69	331191	50.043	ug/l #	92
57) 1,3-Dichloropropane	11.147	76	336112	49.384	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	934209	289.306	ug/l	99
59) 2-Hexanone	11.176	43	1221722	262.377	ug/l	98
60) Dibromochloromethane	11.341	129	217527	48.455	ug/l	100
61) 1,2-Dibromoethane	11.453	107	202290	48.875	ug/l	97
64) Tetrachloroethene	11.082	164	130358	40.459	ug/l	93
65) Chlorobenzene	11.871	112	508857	45.275	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	183045	47.896	ug/l	99
67) Ethyl Benzene	11.947	91	914813	49.442	ug/l	100
68) m/p-Xylenes	12.053	106	683841	98.699	ug/l	95
69) o-Xylene	12.376	106	337782	51.038	ug/l	95
70) Styrene	12.394	104	583897	52.445	ug/l	97
71) Bromoform	12.559	173	143351	46.429	ug/l #	99
73) Isopropylbenzene	12.676	105	854755	53.222	ug/l	98
74) N-amyl acetate	12.500	43	259021	38.819	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.918	83	304854	50.447	ug/l	100
76) 1,2,3-Trichloropropane	12.976	75	257264m	44.960	ug/l	
77) Bromobenzene	12.959	156	203849	48.942	ug/l	93
78) n-propylbenzene	13.012	91	1079956	53.447	ug/l	97
79) 2-Chlorotoluene	13.106	91	652303	52.528	ug/l	95
80) 1,3,5-Trimethylbenzene	13.153	105	735691	53.765	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.718	75	101369	48.472	ug/l	95
82) 4-Chlorotoluene	13.200	91	684119	52.913	ug/l	96
83) tert-Butylbenzene	13.417	119	613860	53.713	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	762982	54.601	ug/l	95
85) sec-Butylbenzene	13.594	105	915066	53.157	ug/l	98
86) p-Isopropyltoluene	13.706	119	762503	55.271	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	399195	48.835	ug/l	99
88) 1,4-Dichlorobenzene	13.794	146	400669	45.893	ug/l	97
89) n-Butylbenzene	14.035	91	736497	55.909	ug/l	98
90) Hexachloroethane	14.312	117	139358	47.677	ug/l	96
91) 1,2-Dichlorobenzene	14.082	146	381913	49.316	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	70382	44.360	ug/l	99
93) 1,2,4-Trichlorobenzene	15.370	180	226838	49.866	ug/l	98
94) Hexachlorobutadiene	15.476	225	82198	48.630	ug/l	99
95) Naphthalene	15.617	128	838952	52.059	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	220386	48.298	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

MSVOA_N

Client SampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By : John Carlone 08/14/2025

Supervised By : Mahesh Dadoda 08/19/2025


Quant Time: Aug 14 15:23:00 2025


Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:56:13 2025


Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 168 Resp: 27135
Ion Ratio Lower Upper
168 100
99 70.1 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#2
Dichlorodifluoromethane
Concen: 55.557 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 85 Resp: 160120
Ion Ratio Lower Upper
85 100
87 33.2 17.8 53.3





#3
Chloromethane
Concen: 46.740 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 50 Resp: 169400
Ion Ratio Lower Upper
50 100
52 32.5 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#4
Vinyl Chloride
Concen: 51.052 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion: 62 Resp: 183878
Ion Ratio Lower Upper
62 100
64 31.6 27.0 40.6





#5
Bromomethane
Concen: 55.023 ug/l
RT: 2.965 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDC050



Tgt Ion: 94 Resp: 102629
Ion Ratio Lower Upper
94 100
96 93.4 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#6
Chloroethane
Concen: 53.961 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 64 Resp: 126750
Ion Ratio Lower Upper
64 100
66 29.0 24.6 36.8









#7
Trichlorofluoromethane
Concen: 50.212 ug/l
RT: 3.506 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 101 Resp: 267420
Ion Ratio Lower Upper
101 100
103 64.5 49.8 74.6




Manual Integrations
APPROVED


Reviewed By : John Carlone 08/14/2025
Supervised By : Mahesh Dadoda 08/19/2025



#8
Diethyl Ether
Concen: 56.667 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 74 Resp: 117073
Ion Ratio Lower Upper
74 100
45 103.9 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.322 ug/l
RT: 4.359 min Scan# 409
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 145784
Ion Ratio Lower Upper
101 100
85 46.7 37.3 55.9
151 67.5 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#10
Methyl Iodide
Concen: 38.232 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion:142 Resp: 103333
Ion Ratio Lower Upper
142 100
127 53.7 35.7 53.5#
141 13.4 10.4 15.6





#11

Tert butyl alcohol

Concen: 276.178 ug/l

RT: 5.536 min Scan# 607

Delta R.T. 0.012 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 241449

Ion Ratio Lower Upper

59 100





57 10.5 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#12

1,1-Dichloroethene

Concen: 49.187 ug/l

RT: 4.336 min Scan# 405

Delta R.T. 0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34




Tgt Ion: 96 Resp: 152389


Ion Ratio Lower Upper

96 100

61 166.8 132.3 198.5


98 59.3 46.8 70.2





#13
Acrolein
Concen: 228.690 ug/l
RT: 4.177 min Scan# 378
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 160450
Ion Ratio Lower Upper
56 100
55 69.7 56.2 84.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#14
Allyl chloride
Concen: 51.014 ug/l
RT: 5.012 min Scan# 520
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion: 41 Resp: 286033
Ion Ratio Lower Upper
41 100
39 85.7 59.0 88.6
76 34.9 28.7 43.1





#15
Acrylonitrile
Concen: 252.452 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 598909
Ion Ratio Lower Upper
53 100
52 84.2 65.5 98.3
51 35.6 28.7 43.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#16
Acetone
Concen: 304.782 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion: 43 Resp: 657966
Ion Ratio Lower Upper
43 100
58 29.4 22.3 33.5









#17
Carbon Disulfide
Concen: 44.164 ug/l
RT: 4.700 min Scan# 40566
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 40566
Ion Ratio Lower Upper
76 100
78 9.9 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#18
Methyl Acetate
Concen: 58.620 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 43 Resp: 317943
Ion Ratio Lower Upper
43 100
74 21.8 17.8 26.6









#19
Methyl tert-butyl Ether
Concen: 57.727 ug/l
RT: 5.789 min Scan# 652
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 659220
Ion Ratio Lower Upper
73 100
57 20.8 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#20
Methylene Chloride
Concen: 52.881 ug/l
RT: 5.259 min Scan# 562
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 84 Resp: 192849
Ion Ratio Lower Upper
84 100
49 130.5 107.5 161.3
51 38.0 30.2 45.2
86 59.2 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 48.942 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 96 Resp: 17097
Ion Ratio Lower Upper
96 100
61 133.9 122.0 183.0
98 59.0 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#22
Diisopropyl ether
Concen: 57.318 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 45 Resp: 674122
Ion Ratio Lower Upper
45 100
43 58.4 42.8 64.2
87 25.3 19.8 29.6
59 10.9 8.3 12.5





#23

Vinyl Acetate

Concen: 298.585 ug/l

RT: 6.589 min Scan# 788

Delta R.T. 0.001 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 307129

Ion Ratio Lower Upper

43 100





86 9.6 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#24

1,1-Dichloroethane

Concen: 51.200 ug/l

RT: 6.553 min Scan# 782

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34




Tgt Ion: 63 Resp: 347405


Ion Ratio Lower Upper

63 100

98 7.1 3.3 9.9


100 3.9 2.5 7.4





#25
2-Butanone
Concen: 267.432 ug/l
RT: 7.471 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 892030
Ion Ratio Lower Upper
43 100
72 23.0 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#26
2,2-Dichloropropane
Concen: 56.588 ug/l
RT: 7.471 min Scan# 938
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 77 Resp: 298522
Ion Ratio Lower Upper
77 100
97 21.0 11.1 33.1





#27
cis-1,2-Dichloroethene
Concen: 52.764 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 212209
Ion Ratio Lower Upper
96 100
61 155.5 0.0 297.8
98 63.2 0.0 132.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#28
Bromochloromethane
Concen: 56.761 ug/l
RT: 7.794 min Scan# 993
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 49 Resp: 184324
Ion Ratio Lower Upper
49 100
129 1.8 0.0 4.2
130 64.7 57.3 85.9





#29

Tetrahydrofuran

Concen: 264.243 ug/l

RT: 7.824 min Scan# 94

Delta R.T. -0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 57258

Ion Ratio Lower Upper

42 100

72 41.8 33.4 50.0





71 39.2 31.2 46.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#30

Chloroform

Concen: 54.800 ug/l

RT: 7.947 min Scan# 1019

Delta R.T. -0.006 min

Lab File: VN087552.D




Acq: 14 Aug 2025 09:34


Tgt Ion: 83 Resp: 372180

Ion Ratio Lower Upper

83 100


85 64.1 52.7 79.1





#31
Cyclohexane
Concen: 49.251 ug/l
RT: 8.235 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 27877
Ion Ratio Lower Upper
56 100
69 31.0 22.7 34.1
84 80.7 65.8 98.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#32
1,1,1-Trichloroethane
Concen: 52.554 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion: 97 Resp: 309142
Ion Ratio Lower Upper
97 100
99 65.7 51.8 77.8
61 51.1 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 53.738 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 247424

Ion Ratio Lower Upper

65 100





67 49.4 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34




Tgt Ion:114 Resp: 543098


Ion Ratio Lower Upper

114 100

63 21.6 0.0 44.6

88 16.7 0.0 32.8





#35

Dibromofluoromethane

Concen: 44.903 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 16822

Ion Ratio Lower Upper

113 100

111 103.3 82.5 123.7





192 15.6 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#36

1,1-Dichloropropene

Concen: 48.073 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34




Tgt Ion: 75 Resp: 237939


Ion Ratio Lower Upper

75 100

110 31.4 16.7 50.1

77 30.7 25.2 37.8





#37

Ethyl Acetate

Concen: 47.639 ug/l

RT: 7.547 min Scan# 951

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N





ClientSampleId :

VSTDC050

Tgt Ion:	43	Resp:	340534
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.9	16.3
70	9.4	7.4	11.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#38

Carbon Tetrachloride

Concen: 45.251 ug/l




RT: 8.347 min Scan# 1087


Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Tgt Ion:	117	Resp:	246721
Ion	Ratio	Lower	Upper
117	100		
119	97.5	80.2	120.2
121	30.1	25.4	38.2









#39
Methylcyclohexane
Concen: 47.484 ug/l
RT: 9.582 min Scan# 1129
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 25444
Ion Ratio Lower Upper
83 100
55 80.6 61.3 91.9
98 48.5 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#40
Benzene
Concen: 47.666 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 78 Resp: 762513
Ion Ratio Lower Upper
78 100
77 24.7 18.2 27.2





#41
Methacrylonitrile
Concen: 47.844 ug/l
RT: 7.765 min Scan# 911
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 178825
Ion Ratio Lower Upper
41 100
39 58.4 43.4 65.0
67 71.7 55.1 82.7
52 32.6 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#42
1,2-Dichloroethane
Concen: 51.277 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 62 Resp: 311061
Ion Ratio Lower Upper
62 100
98 8.0 0.0 18.0





#43

Isopropyl Acetate

Concen: 51.848 ug/l

RT: 8.671 min Scan# 1142

Delta R.T. -0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 57533

Ion Ratio Lower Upper

43

61

87

100

23.2

11.9

19.8

9.8

29.8





14.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#44

Trichloroethene

Concen: 43.791 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Tgt Ion:130 Resp: 165523

Ion Ratio Lower Upper

130




95


100

102.4

0.0

195.2





#45

1,2-Dichloropropane

Concen: 49.245 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 20016

Ion Ratio Lower Upper

63 100





65 28.9 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#46

Dibromomethane

Concen: 48.459 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34




Tgt Ion: 93 Resp: 147474


Ion Ratio Lower Upper

93 100

95 81.6 65.8 98.8

174 81.6 69.9 104.9





#47

Bromodichloromethane

Concen: 48.711 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:	83	Resp:	29859
Ion	Ratio	Lower	Upper
83	100		
85	62.8	50.4	75.6
127	8.0	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#48

Methyl methacrylate

Concen: 52.798 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Tgt Ion:	41	Resp:	263753
Ion	Ratio	Lower	Upper
41	100		
69	78.3	64.1	96.1
39	67.0	45.5	68.3





#49

1,4-Dioxane

Concen: 1013.618 ug/l

RT: 9.682 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 77554

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 75.2 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#50

Toluene-d8

Concen: 45.726 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087552.D




Acq: 14 Aug 2025 09:34


Tgt Ion: 98 Resp: 611058

Ion Ratio Lower Upper

98 100


100 65.0 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 252.428 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 177161
Ion Ratio Lower Upper
43 100
58 37.2 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#52
Toluene
Concen: 48.609 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 92 Resp: 472630
Ion Ratio Lower Upper
92 100
91 176.0 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 51.017 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 31649
Ion Ratio Lower Upper
75 100
77 31.8 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#54
cis-1,3-Dichloropropene
Concen: 51.028 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 75 Resp: 326996
Ion Ratio Lower Upper
75 100
77 30.5 24.2 36.2
39 63.5 38.4 57.6#





#55
1,1,2-Trichloroethane
Concen: 48.289 ug/l
RT: 11.000 min Scan# 1538
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 19008
Ion Ratio Lower Upper
97 100
83 87.7 72.6 109.0
85 55.4 46.7 70.1
99 61.2 54.1 81.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#56
Ethyl methacrylate
Concen: 50.043 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion: 69 Resp: 331191
Ion Ratio Lower Upper
69 100
41 70.4 55.1 82.7
39 46.2 27.9 41.9#





#57

1,3-Dichloropropane

Concen: 49.384 ug/l

RT: 11.147 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 33611

Ion Ratio Lower Upper

76 100





78 32.2 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#58

2-Chloroethyl Vinyl ether

Concen: 289.306 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087552.D




Acq: 14 Aug 2025 09:34


Tgt Ion: 63 Resp: 934209

Ion Ratio Lower Upper

63 100

106 26.6 21.7 32.5









#59
2-Hexanone
Concen: 262.377 ug/l
RT: 11.176 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 122172
Ion Ratio Lower Upper
43 100
58 51.9 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#60
Dibromochloromethane
Concen: 48.455 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion:129 Resp: 217527
Ion Ratio Lower Upper
129 100
127 78.5 39.1 117.5





#61
1,2-Dibromoethane
Concen: 48.875 ug/l
RT: 11.453 min Scan# 1615
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 107 Resp: 202290
Ion Ratio Lower Upper
107 100
109 94.1 77.5 116.3

Manual Integrations
APPROVED



Reviewed By : John Carlone 08/14/2025
Supervised By : Mahesh Dadoda 08/19/2025




#62
4-Bromofluorobenzene
Concen: 47.655 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion: 95 Resp: 235280
Ion Ratio Lower Upper
95 100
174 70.2 0.0 149.4
176 66.7 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 11847
Delta R.T. -0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 500619
Ion Ratio Lower Upper
117 100
82 59.9 47.4 71.2
119 34.0 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#64
Tetrachloroethene
Concen: 40.459 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion:164 Resp: 130358
Ion Ratio Lower Upper
164 100
166 125.7 105.5 158.3
129 108.2 77.4 116.2
131 101.7 77.3 115.9









#65
Chlorobenzene
Concen: 45.275 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 50885
Ion Ratio Lower Upper
112 100
114 31.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 47.896 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion:131 Resp: 183045
Ion Ratio Lower Upper
131 100
133 93.8 47.4 142.3
119 65.7 33.1 99.2





#67
Ethyl Benzene
Concen: 49.442 ug/l
RT: 11.947 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 91481
Ion Ratio Lower Upper
91 100
106 30.5 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#68
m/p-Xylenes
Concen: 98.699 ug/l
RT: 12.053 min Scan# 1717
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion:106 Resp: 683841
Ion Ratio Lower Upper
106 100
91 210.7 162.0 243.0









#69
o-Xylene
Concen: 51.038 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 33778
Ion Ratio Lower Upper
106 100
91 223.0 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#70
Styrene
Concen: 52.445 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion:104 Resp: 583897
Ion Ratio Lower Upper
104 100
78 53.9 41.0 61.6
103 56.4 43.9 65.9





#71
Bromoform
Concen: 46.429 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 14335
Ion Ratio Lower Upper
173 100
175 47.4 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion:152 Resp: 255173
Ion Ratio Lower Upper
152 100
115 64.3 31.1 93.5
150 173.5 0.0 349.0





#73
Isopropylbenzene
Concen: 53.222 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 105 Resp: 85475
Ion Ratio Lower Upper
105 100
120 25.4 13.4 40.1

Manual Integrations
APPROVED



Reviewed By : John Carlone 08/14/2025
Supervised By : Mahesh Dadoda 08/19/2025




#74
N-ethyl acetate
Concen: 38.819 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion: 43 Resp: 259021
Ion Ratio Lower Upper
43 100
70 36.1 37.6 56.4#
55 20.5 19.6 29.4
61 21.2 20.6 31.0





#75
1,1,2,2-Tetrachloroethane
Concen: 50.447 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 304854
Ion Ratio Lower Upper
83 100
131 9.5 5.1 15.3
85 64.8 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#76
1,2,3-Trichloropropane
Concen: 44.960 ug/l m
RT: 12.976 min Scan# 1874
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Tgt Ion: 75 Resp: 257264
Ion Ratio Lower Upper
75 100
77 193.5 94.5 283.6





#77
Bromobenzene
Concen: 48.942 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 203849
Ion Ratio Lower Upper
156 100
77 244.2 114.9 344.6
158 98.6 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#78
n-propylbenzene
Concen: 53.447 ug/l
RT: 13.012 min Scan# 1880
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 91 Resp: 1079956
Ion Ratio Lower Upper
91 100
120 20.9 11.3 33.8





#79

2-Chlorotoluene

Concen: 52.528 ug/l

RT: 13.106 min Scan# 1896

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 65230

Ion Ratio Lower Upper

91 100





126 30.8 16.9 50.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#80

1,3,5-Trimethylbenzene

Concen: 53.765 ug/l

RT: 13.153 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VN087552.D




Acq: 14 Aug 2025 09:34


Tgt Ion:105 Resp: 735691

Ion Ratio Lower Upper

105 100


120 46.2 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 48.472 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 101369
Ion Ratio Lower Upper
75 100
53 109.0 83.5 125.3
89 43.7 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#82
4-Chlorotoluene
Concen: 52.913 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion: 91 Resp: 684119
Ion Ratio Lower Upper
91 100
126 31.1 16.6 49.7









#83
tert-Butylbenzene
Concen: 53.713 ug/l
RT: 13.417 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:119 Resp: 613860
Ion Ratio Lower Upper
119 100
91 72.3 35.0 105.1
134 26.5 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#84
1,2,4-Trimethylbenzene
Concen: 54.601 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion:105 Resp: 762982
Ion Ratio Lower Upper
105 100
120 42.4 22.8 68.3









#85
sec-Butylbenzene
Concen: 53.157 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 915060
Ion Ratio Lower Upper
105 100
134 18.9 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#86
p-Isopropyltoluene
Concen: 55.271 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion:119 Resp: 762503
Ion Ratio Lower Upper
119 100
134 25.0 13.5 40.5
91 25.5 12.2 36.6









#87
1,3-Dichlorobenzene
Concen: 48.835 ug/l
RT: 13.712 min Scan# 1999
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 39919
Ion Ratio Lower Upper
146 100
111 42.6 21.4 64.3
148 63.9 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#88
1,4-Dichlorobenzene
Concen: 45.893 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Tgt Ion:146 Resp: 400669
Ion Ratio Lower Upper
146 100
111 42.1 19.6 58.7
148 63.8 31.4 94.0





#89

n-Butylbenzene

Concen: 55.909 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:	91	Resp:	73649
Ion	Ratio	Lower	Upper
91	100		
92	53.2	26.2	78.6
134	23.5	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#90

Hexachloroethane

Concen: 47.677 ug/l




RT: 14.312 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Tgt Ion:	117	Resp:	139358
Ion	Ratio	Lower	Upper
117	100		
201	62.5	32.8	98.4









#91
1,2-Dichlorobenzene
Concen: 49.316 ug/l
RT: 14.082 min Scan# 2062
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 38191
Ion Ratio Lower Upper
146 100
111 45.7 22.0 66.0
148 64.3 32.7 98.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 44.360 ug/l
RT: 14.700 min Scan# 2167
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34


Tgt Ion: 75 Resp: 70382
Ion Ratio Lower Upper
75 100
155 73.3 37.3 111.8
157 92.7 46.2 138.6





#93
1,2,4-Trichlorobenzene
Concen: 49.866 ug/l
RT: 15.370 min Scan# 2281
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:180 Resp: 226838
Ion Ratio Lower Upper
180 100
182 95.3 47.3 141.8
145 33.8 15.4 46.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025




#94
Hexachlorobutadiene
Concen: 48.630 ug/l
RT: 15.476 min Scan# 2299
Delta R.T. 0.000 min
Lab File: VN087552.D
Acq: 14 Aug 2025 09:34



Tgt Ion:225 Resp: 82198
Ion Ratio Lower Upper
225 100
223 63.6 32.1 96.3
227 64.0 31.3 93.9





#95

Naphthalene

Concen: 52.059 ug/l

RT: 15.617 min Scan# 2323

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:	128	Resp:	83895
Ion Ratio	Lower	Upper	
128	100		
127	12.9	10.5	15.7
129	10.8	8.4	12.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025

Supervised By :Mahesh Dadoda 08/19/2025



#96

1,2,3-Trichlorobenzene

Concen: 48.298 ug/l




RT: 15.811 min Scan# 2356

Delta R.T. 0.000 min

Lab File: VN087552.D

Acq: 14 Aug 2025 09:34

Tgt Ion:	180	Resp:	220386
Ion Ratio	Lower	Upper	
180	100		
182	96.4	47.1	141.4
145	36.0	16.9	50.6



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	148	0.00
2 T	Dichlorodifluoromethane	0.531	0.590	-11.1	141	0.00
3 P	Chloromethane	0.668	0.624	6.6	134	0.00
4 C	Vinyl Chloride	0.664	0.678	-2.1#	138	0.00
5 T	Bromomethane	0.344	0.378	-9.9	158#	0.00
6 T	Chloroethane	0.433	0.467	-7.9	157#	0.00
7 T	Trichlorofluoromethane	0.981	0.986	-0.5	143	0.00
8 T	Diethyl Ether	0.381	0.431	-13.1	162#	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.537	-6.5	152#	0.00
10 T	Methyl Iodide	0.452	0.381	15.7	114	0.00
11 T	Tert butyl alcohol	0.161	0.178	-10.6	164#	0.01
12 CM	1,1-Dichloroethene	0.571	0.562	1.6#	153#	0.00
13 T	Acrolein	0.129	0.118	8.5	145	0.00
14 T	Allyl chloride	1.033	1.054	-2.0	156#	0.00
15 T	Acrylonitrile	0.437	0.441	-0.9	145	0.00
16 T	Acetone	0.398	0.485	-21.9	186#	0.00
17 T	Carbon Disulfide	1.693	1.495	11.7	128	0.00
18 T	Methyl Acetate	0.999	1.172	-17.3	175#	0.00
19 T	Methyl tert-butyl Ether	2.104	2.429	-15.4	166#	0.00
20 T	Methylene Chloride	0.766	0.711	7.2	157#	0.00
21 T	trans-1,2-Dichloroethene	0.644	0.630	2.2	143	0.00
22 T	Diisopropyl ether	2.167	2.484	-14.6	161#	0.00
23 T	Vinyl Acetate	1.895	2.264	-19.5	159#	0.00
24 P	1,1-Dichloroethane	1.250	1.280	-2.4	156#	0.00
25 T	2-Butanone	0.615	0.657	-6.8	152#	0.00
26 T	2,2-Dichloropropane	0.972	1.100	-13.2	166#	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.782	-5.5	151#	0.00
28 T	Bromochloromethane	0.598	0.679	-13.5	169#	0.00
29 T	Tetrahydrofuran	0.399	0.422	-5.8	147	0.00
30 C	Chloroform	1.251	1.372	-9.7#	159#	0.00
31 T	Cyclohexane	1.043	1.027	1.5	146	0.00
32 T	1,1,1-Trichloroethane	1.084	1.139	-5.1	156#	0.00
33 S	1,2-Dichloroethane-d4	0.848	0.912	-7.5	169#	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	167#	0.00
35 S	Dibromofluoromethane	0.345	0.310	10.1	155#	0.00
36 T	1,1-Dichloropropene	0.456	0.438	3.9	150#	0.00
37 T	Ethyl Acetate	0.658	0.627	4.7	148	0.00
38 T	Carbon Tetrachloride	0.502	0.454	9.6	147	0.00
39 T	Methylcyclohexane	0.493	0.469	4.9	148	0.00
40 TM	Benzene	1.473	1.404	4.7	151#	0.00
41 T	Methacrylonitrile	0.344	0.329	4.4	149	0.00
42 TM	1,2-Dichloroethane	0.558	0.573	-2.7	168#	0.00
43 T	Isopropyl Acetate	1.022	1.059	-3.6	165#	0.00
44 TM	Trichloroethene	0.348	0.305	12.4	143	0.00
45 C	1,2-Dichloropropane	0.374	0.369	1.3#	156#	0.00
46 T	Dibromomethane	0.280	0.272	2.9	157#	0.00
47 T	Bromodichloromethane	0.564	0.550	2.5	161#	0.00
48 T	Methyl methacrylate	0.460	0.486	-5.7	160#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	149	0.00
50 S	Toluene-d8	1.230	1.125	8.5	153#	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.652	-0.9	158#	0.00
52 CM	Toluene	0.895	0.870	2.8#	151#	0.00
53 T	t-1,3-Dichloropropene	0.571	0.583	-2.1	157#	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.602	-2.0	159#	0.00
55 T	1,1,2-Trichloroethane	0.362	0.350	3.3	159#	0.00
56 T	Ethyl methacrylate	0.552	0.610	-10.5	162#	0.00
57 T	1,3-Dichloropropane	0.627	0.619	1.3	158#	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.344	-15.8	166#	0.00
59 T	2-Hexanone	0.429	0.450	-4.9	151#	0.00
60 T	Dibromochloromethane	0.413	0.401	2.9	155#	0.00
61 T	1,2-Dibromoethane	0.381	0.372	2.4	161#	0.00
62 S	4-Bromofluorobenzene	0.455	0.433	4.8	159#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	167#	0.00
64 T	Tetrachloroethene	0.322	0.260	19.3	136	0.00
65 PM	Chlorobenzene	1.123	1.016	9.5	152#	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.366	4.2	155#	0.00
67 C	Ethyl Benzene	1.848	1.827	1.1#	157#	0.00
68 T	m/p-Xylenes	0.692	0.683	1.3	150#	0.00
69 T	o-Xylene	0.661	0.675	-2.1	156#	0.00
70 T	Styrene	1.112	1.166	-4.9	155#	0.00
71 P	Bromoform	0.308	0.286	7.1	145	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	161#	0.00
73 T	Isopropylbenzene	3.147	3.350	-6.5	159#	0.00
74 T	N-amyl acetate	1.307	1.015	22.3	144	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.195	-0.9	159#	0.00
76 T	1,2,3-Trichloropropane	1.121	1.008	10.1	136	0.00
77 T	Bromobenzene	0.816	0.799	2.1	148	0.00
78 T	n-propylbenzene	3.959	4.232	-6.9	159#	0.00
79 T	2-Chlorotoluene	2.433	2.556	-5.1	162#	0.00
80 T	1,3,5-Trimethylbenzene	2.681	2.883	-7.5	159#	0.00
81 T	trans-1,4-Dichloro-2-butene	0.410	0.397	3.2	158#	0.00
82 T	4-Chlorotoluene	2.533	2.681	-5.8	163#	0.00
83 T	tert-Butylbenzene	2.239	2.406	-7.5	160#	0.00
84 T	1,2,4-Trimethylbenzene	2.738	2.990	-9.2	160#	0.00
85 T	sec-Butylbenzene	3.373	3.586	-6.3	162#	0.00
86 T	p-Isopropyltoluene	2.703	2.988	-10.5	161#	0.00
87 T	1,3-Dichlorobenzene	1.602	1.564	2.4	152#	0.00
88 T	1,4-Dichlorobenzene	1.711	1.570	8.2	148	0.00
89 T	n-Butylbenzene	2.581	2.886	-11.8	168#	0.00
90 T	Hexachloroethane	0.573	0.546	4.7	153#	0.00
91 T	1,2-Dichlorobenzene	1.517	1.497	1.3	153#	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.311	0.276	11.3	147	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.889	0.2	153#	0.00
94 T	Hexachlorobutadiene	0.331	0.322	2.7	158#	0.00
95 T	Naphthalene	3.158	3.288	-4.1	154#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.894	0.864	3.4	151#	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	148	0.00
2 T	Dichlorodifluoromethane	50.000	55.557	-11.1	141	0.00
3 P	Chloromethane	50.000	46.740	6.5	134	0.00
4 C	Vinyl Chloride	50.000	51.052	-2.1#	138	0.00
5 T	Bromomethane	50.000	55.023	-10.0	158	0.00
6 T	Chloroethane	50.000	53.961	-7.9	157	0.00
7 T	Trichlorofluoromethane	50.000	50.212	-0.4	143	0.00
8 T	Diethyl Ether	50.000	56.667	-13.3	162	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.322	-6.6	152	0.00
10 T	Methyl Iodide	50.000	38.232	23.5	114	0.00
11 T	Tert butyl alcohol	250.000	276.178	-10.5	164	0.01
12 CM	1,1-Dichloroethene	50.000	49.187	1.6#	153	0.00
13 T	Acrolein	250.000	228.690	8.5	145	0.00
14 T	Allyl chloride	50.000	51.014	-2.0	156	0.00
15 T	Acrylonitrile	250.000	252.452	-1.0	145	0.00
16 T	Acetone	250.000	304.782	-21.9	186	0.00
17 T	Carbon Disulfide	50.000	44.164	11.7	128	0.00
18 T	Methyl Acetate	50.000	58.620	-17.2	175	0.00
19 T	Methyl tert-butyl Ether	50.000	57.727	-15.5	166	0.00
20 T	Methylene Chloride	50.000	52.881	-5.8	157	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.942	2.1	143	0.00
22 T	Diisopropyl ether	50.000	57.318	-14.6	161	0.00
23 T	Vinyl Acetate	250.000	298.585	-19.4	159	0.00
24 P	1,1-Dichloroethane	50.000	51.200	-2.4	156	0.00
25 T	2-Butanone	250.000	267.432	-7.0	152	0.00
26 T	2,2-Dichloropropane	50.000	56.588	-13.2	166	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.764	-5.5	151	0.00
28 T	Bromochloromethane	50.000	56.761	-13.5	169	0.00
29 T	Tetrahydrofuran	250.000	264.243	-5.7	147	0.00
30 C	Chloroform	50.000	54.800	-9.6#	159	0.00
31 T	Cyclohexane	50.000	49.251	1.5	146	0.00
32 T	1,1,1-Trichloroethane	50.000	52.554	-5.1	156	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.738	-7.5	169	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	167	0.00
35 S	Dibromofluoromethane	50.000	44.903	10.2	155	0.00
36 T	1,1-Dichloropropene	50.000	48.073	3.9	150	0.00
37 T	Ethyl Acetate	50.000	47.639	4.7	148	0.00
38 T	Carbon Tetrachloride	50.000	45.251	9.5	147	0.00
39 T	Methylcyclohexane	50.000	47.484	5.0	148	0.00
40 TM	Benzene	50.000	47.666	4.7	151	0.00
41 T	Methacrylonitrile	50.000	47.844	4.3	149	0.00
42 TM	1,2-Dichloroethane	50.000	51.277	-2.6	168	0.00
43 T	Isopropyl Acetate	50.000	51.848	-3.7	165	0.00
44 TM	Trichloroethene	50.000	43.791	12.4	143	0.00
45 C	1,2-Dichloropropane	50.000	49.245	1.5#	156	0.00
46 T	Dibromomethane	50.000	48.459	3.1	157	0.00
47 T	Bromodichloromethane	50.000	48.711	2.6	161	0.00
48 T	Methyl methacrylate	50.000	52.798	-5.6	160	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1013.618	-1.4	149	0.00
50 S	Toluene-d8	50.000	45.726	8.5	153	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.428	-1.0	158	0.00
52 CM	Toluene	50.000	48.609	2.8#	151	0.00
53 T	t-1,3-Dichloropropene	50.000	51.017	-2.0	157	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.028	-2.1	159	0.00
55 T	1,1,2-Trichloroethane	50.000	48.289	3.4	159	0.00
56 T	Ethyl methacrylate	50.000	50.043	-0.1	162	0.00
57 T	1,3-Dichloropropane	50.000	49.384	1.2	158	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	289.306	-15.7	166	0.00
59 T	2-Hexanone	250.000	262.377	-5.0	151	0.00
60 T	Dibromochloromethane	50.000	48.455	3.1	155	0.00
61 T	1,2-Dibromoethane	50.000	48.875	2.3	161	0.00
62 S	4-Bromofluorobenzene	50.000	47.655	4.7	159	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	167	0.00
64 T	Tetrachloroethene	50.000	40.459	19.1	136	0.00
65 PM	Chlorobenzene	50.000	45.275	9.5	152	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.896	4.2	155	0.00
67 C	Ethyl Benzene	50.000	49.442	1.1#	157	0.00
68 T	m/p-Xylenes	100.000	98.699	1.3	150	0.00
69 T	o-Xylene	50.000	51.038	-2.1	156	0.00
70 T	Styrene	50.000	52.445	-4.9	155	0.00
71 P	Bromoform	50.000	46.429	7.1	145	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	161	0.00
73 T	Isopropylbenzene	50.000	53.222	-6.4	159	0.00
74 T	N-ethyl acetate	50.000	38.819	22.4	144	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.447	-0.9	159	0.00
76 T	1,2,3-Trichloropropane	50.000	44.960	10.1	136	0.00
77 T	Bromobenzene	50.000	48.942	2.1	148	0.00
78 T	n-propylbenzene	50.000	53.447	-6.9	159	0.00
79 T	2-Chlorotoluene	50.000	52.528	-5.1	162	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.765	-7.5	159	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.472	3.1	158	0.00
82 T	4-Chlorotoluene	50.000	52.913	-5.8	163	0.00
83 T	tert-Butylbenzene	50.000	53.713	-7.4	160	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.601	-9.2	160	0.00
85 T	sec-Butylbenzene	50.000	53.157	-6.3	162	0.00
86 T	p-Isopropyltoluene	50.000	55.271	-10.5	161	0.00
87 T	1,3-Dichlorobenzene	50.000	48.835	2.3	152	0.00
88 T	1,4-Dichlorobenzene	50.000	45.893	8.2	148	0.00
89 T	n-Butylbenzene	50.000	55.909	-11.8	168	0.00
90 T	Hexachloroethane	50.000	47.677	4.6	153	0.00
91 T	1,2-Dichlorobenzene	50.000	49.316	1.4	153	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.360	11.3	147	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.866	0.3	153	0.00
94 T	Hexachlorobutadiene	50.000	48.630	2.7	158	0.00
95 T	Naphthalene	50.000	52.059	-4.1	154	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	50.000	48.298	3.4	151	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087327.D
Acq On : 16 Jul 2025 16:10
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Title : SW846 8260
Last Update : Thu Jul 17 02:56:13 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1839



Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.8	8426	PASS
75	95	30	60	50.8	21643	PASS
95	95	100	100	100.0	42640	PASS
96	95	5	9	6.1	2612	PASS
173	174	0.00	2	0.8	255	PASS
174	95	50	100	70.9	30229	PASS
175	174	5	9	5.1	1538	PASS
176	174	95	101	96.9	29285	PASS
177	176	5	9	7.0	2046	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087551.D
Acq On : 14 Aug 2025 09:01
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Title : SW846 8260
Last Update : Thu Jul 17 02:56:13 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1840

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.8	14484	PASS
75	95	30	60	57.2	34829	PASS
95	95	100	100	100.0	60920	PASS
96	95	5	9	5.6	3407	PASS
173	174	0.00	2	1.3	536	PASS
174	95	50	100	65.2	39723	PASS
175	174	5	9	6.8	2685	PASS
176	174	95	101	98.8	39240	PASS
177	176	5	9	6.2	2451	PASS

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VN0814WBL01	SDG No.:	Q2842
Lab Sample ID:	VN0814WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087554.D	1	08/14/25 10:32	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	44.2		70 (75) - 130 (124)	88%	SPK: 50
2037-26-5	Toluene-d8	44.2		70 (86) - 130 (113)	88%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.7		70 (77) - 130 (121)	87%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	284000	8.206			
540-36-3	1,4-Difluorobenzene	560000	9.083			
3114-55-4	Chlorobenzene-d5	504000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	235000	13.771			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087554.D
Acq On : 14 Aug 2025 10:32
Operator : JC\MD
Sample : VN0814WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.206	168	284398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	560321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	503771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	234765	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	255137	52.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.740%
35) Dibromofluoromethane	8.153	113	170840	44.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.400%
50) Toluene-d8	10.547	98	608969	44.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.340%
62) 4-Bromofluorobenzene	12.829	95	222642	43.709	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.420%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087554.D
Acq On : 14 Aug 2025 10:32
Operator : JC\MD
Sample : VN0814WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1106
Delta R.T. -0.006 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01


Tgt Ion:168 Resp: 284398
Ion Ratio Lower Upper
168 100
99 67.0 47.9 71.9



#33
1,2-Dichloroethane-d4
Concen: 52.871 ug/l
RT: 8.565 min Scan# 1124
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion: 65 Resp: 255137
Ion Ratio Lower Upper
65 100
67 50.2 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 1

Delta R.T. 0.001 min

Lab File: VN087554.D

Acq: 14 Aug 2025 10:32

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBL01





Tgt Ion:114 Resp: 560321

Ion Ratio Lower Upper

114 100

63 23.2 0.0 44.6

88 16.3 0.0 32.8



#35

Dibromofluoromethane

Concen: 44.201 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087554.D

Acq: 14 Aug 2025 10:32




Tgt Ion:113 Resp: 170840


Ion Ratio Lower Upper

113 100

111 103.1 82.5 123.7

192 16.3 13.7 20.5









#50
Toluene-d8
Concen: 44.169 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01


Tgt Ion: 98 Resp: 608969
Ion Ratio Lower Upper
98 100
100 64.2 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 43.709 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion: 95 Resp: 222642
Ion Ratio Lower Upper
95 100
174 66.5 0.0 149.4
176 64.9 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Tgt Ion	Ratio	Lower	Upper
117	100		
82	60.7	47.4	71.2
119	31.6	23.8	35.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.771 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion	Ratio	Lower	Upper
152	100		
115	64.5	31.1	93.5
150	156.1	0.0	349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VN0814WBS01	SDG No.:	Q2842
Lab Sample ID:	VN0814WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087555.D	1	08/14/25 11:07	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	21.4		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.6		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.6		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	21.2		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.4		0.20	1.00	ug/L
71-43-2	Benzene	20.3		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.7		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	18.1		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.6		70 (74) - 130 (125)	115%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	250000	8.206			
540-36-3	1,4-Difluorobenzene	485000	9.082			
3114-55-4	Chlorobenzene-d5	447000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	222000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087555.D
 Acq On : 14 Aug 2025 11:07
 Operator : JC\MD
 Sample : VN0814WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 08/18/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	250453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	485218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	446518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	221983	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.565	65	244982	57.648	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.300%	
35) Dibromofluoromethane	8.153	113	169668	50.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.380%	
50) Toluene-d8	10.547	98	605258	50.695	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.380%	
62) 4-Bromofluorobenzene	12.829	95	233458	52.926	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.860%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	60972	22.921	ug/l	92
3) Chloromethane	2.383	50	64803	19.372	ug/l	100
4) Vinyl Chloride	2.542	62	71080	21.381	ug/l	99
5) Bromomethane	2.983	94	38871	22.579	ug/l	96
6) Chloroethane	3.136	64	47553	21.934	ug/l	92
7) Trichlorofluoromethane	3.506	101	103028	20.959	ug/l	98
8) Diethyl Ether	3.965	74	45341	23.778	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.359	101	56020	22.200	ug/l	96
10) Methyl Iodide	4.571	142	33911	17.223	ug/l	89
11) Tert butyl alcohol	5.530	59	92129	114.174	ug/l	97
12) 1,1-Dichloroethene	4.336	96	58932	20.609	ug/l	91
13) Acrolein	4.177	56	52972	81.801	ug/l	95
14) Allyl chloride	5.012	41	106153	20.512	ug/l	91
15) Acrylonitrile	5.706	53	228245	104.238	ug/l	97
16) Acetone	4.430	43	234225	117.551	ug/l	98
17) Carbon Disulfide	4.700	76	152230	17.956	ug/l	97
18) Methyl Acetate	5.024	43	120344	24.040	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	247283	23.461	ug/l	98
20) Methylene Chloride	5.259	84	73232	21.375	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	62532	19.394	ug/l	92
22) Diisopropyl ether	6.659	45	247597	22.809	ug/l	95
23) Vinyl Acetate	6.588	43	1129441	118.964	ug/l	100
24) 1,1-Dichloroethane	6.559	63	135050	21.564	ug/l	97
25) 2-Butanone	7.477	43	332813	108.103	ug/l	95
26) 2,2-Dichloropropane	7.477	77	115630	23.748	ug/l	97
27) cis-1,2-Dichloroethene	7.477	96	78759	21.217	ug/l	94
28) Bromochloromethane	7.800	49	75409	25.159	ug/l	95
29) Tetrahydrofuran	7.829	42	208941	104.471	ug/l	98
30) Chloroform	7.953	83	140698	22.445	ug/l	99
31) Cyclohexane	8.241	56	111696	21.380	ug/l	97
32) 1,1,1-Trichloroethane	8.153	97	115982	21.362	ug/l	96
36) 1,1-Dichloropropene	8.353	75	90752	20.523	ug/l	97
37) Ethyl Acetate	7.547	43	131078	20.524	ug/l	98
38) Carbon Tetrachloride	8.347	117	93434	19.181	ug/l	96
39) Methylcyclohexane	9.582	83	96372	20.130	ug/l	93
40) Benzene	8.588	78	289566	20.261	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087555.D
 Acq On : 14 Aug 2025 11:07
 Operator : JC\MD
 Sample : VN0814WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBS01

Quant Time: Aug 14 15:23:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/18/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	66759	19.992	ug/l	95
42) 1,2-Dichloroethane	8.659	62	116694	21.531	ug/l	97
43) Isopropyl Acetate	8.676	43	215090	21.696	ug/l	98
44) Trichloroethene	9.335	130	64743	19.172	ug/l	82
45) 1,2-Dichloropropane	9.606	63	74233	20.442	ug/l	94
46) Dibromomethane	9.688	93	56540	20.795	ug/l	94
47) Bromodichloromethane	9.871	83	112437	20.530	ug/l	99
48) Methyl methacrylate	9.665	41	98225	22.008	ug/l	95
49) 1,4-Dioxane	9.688	88	30568	447.176	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	661207	105.450	ug/l	98
52) Toluene	10.612	92	177586	20.443	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	116338	20.990	ug/l	99
54) cis-1,3-Dichloropropene	10.294	75	116909	20.420	ug/l #	82
55) 1,1,2-Trichloroethane	10.994	97	72631	20.652	ug/l	92
56) Ethyl methacrylate	10.859	69	120174	21.091	ug/l #	93
57) 1,3-Dichloropropane	11.141	76	128302	21.100	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	328582	113.893	ug/l	98
59) 2-Hexanone	11.182	43	431325	103.681	ug/l	98
60) Dibromochloromethane	11.335	129	76256	19.012	ug/l	98
61) 1,2-Dibromoethane	11.447	107	75775	20.492	ug/l	95
64) Tetrachloroethene	11.088	164	51988	18.090	ug/l	94
65) Chlorobenzene	11.870	112	192917	19.244	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.941	131	65609	19.247	ug/l	97
67) Ethyl Benzene	11.947	91	336124	20.367	ug/l	98
68) m/p-Xylenes	12.053	106	249595	40.389	ug/l	94
69) o-Xylene	12.376	106	119950	20.320	ug/l	93
70) Styrene	12.394	104	205053	20.649	ug/l	96
71) Bromoform	12.559	173	48296	17.538	ug/l #	99
73) Isopropylbenzene	12.676	105	314281	22.495	ug/l	97
74) N-amyl acetate	12.535	43	108512m	18.694	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.923	83	115038	21.882	ug/l	99
76) 1,2,3-Trichloropropane	12.976	75	100267m	20.143	ug/l	
77) Bromobenzene	12.959	156	73544	20.297	ug/l	91
78) n-propylbenzene	13.017	91	389520	22.160	ug/l	99
79) 2-Chlorotoluene	13.106	91	241476	22.353	ug/l	94
80) 1,3,5-Trimethylbenzene	13.153	105	262222	22.029	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.723	75	32197	17.698	ug/l	91
82) 4-Chlorotoluene	13.200	91	250722	22.292	ug/l	95
83) tert-Butylbenzene	13.417	119	222918	22.422	ug/l	96
84) 1,2,4-Trimethylbenzene	13.464	105	269488	22.169	ug/l	99
85) sec-Butylbenzene	13.594	105	329434	21.998	ug/l	99
86) p-Isopropyltoluene	13.706	119	272080	22.671	ug/l	97
87) 1,3-Dichlorobenzene	13.711	146	145480	20.458	ug/l	98
88) 1,4-Dichlorobenzene	13.794	146	150121	19.766	ug/l	97
89) n-Butylbenzene	14.035	91	270370	23.593	ug/l	98
90) Hexachloroethane	14.311	117	49037	19.285	ug/l	96
91) 1,2-Dichlorobenzene	14.088	146	138916	20.620	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	27318	19.792	ug/l	92
93) 1,2,4-Trichlorobenzene	15.364	180	83245	21.036	ug/l	99
94) Hexachlorobutadiene	15.476	225	31065	21.127	ug/l	99
95) Naphthalene	15.617	128	298554	21.296	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	78959	19.891	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087555.D
Acq On : 14 Aug 2025 11:07
Operator : JC\MD
Sample : VN0814WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed						

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Manual Integrations APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1064
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 168 Resp: 25045
Ion Ratio Lower Upper
168 100
99 68.7 47.9 71.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 08/18/2025
Supervised By : Mahesh Dadoda 08/19/2025



#2
Dichlorodifluoromethane
Concen: 22.921 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 85 Resp: 60972
Ion Ratio Lower Upper
85 100
87 31.1 17.8 53.3









#3
Chloromethane
Concen: 19.372 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 50 Resp: 6480
Ion Ratio Lower Upper
50 100
52 33.1 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#4
Vinyl Chloride
Concen: 21.381 ug/l
RT: 2.542 min Scan# 100
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 62 Resp: 71080
Ion Ratio Lower Upper
62 100
64 33.0 27.0 40.6









#5
Bromomethane
Concen: 22.579 ug/l
RT: 2.983 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 94 Resp: 3887
Ion Ratio Lower Upper
94 100
96 95.3 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#6
Chloroethane
Concen: 21.934 ug/l
RT: 3.136 min Scan# 201
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 64 Resp: 47553
Ion Ratio Lower Upper
64 100
66 26.5 24.6 36.8









#7
Trichlorofluoromethane
Concen: 20.959 ug/l
RT: 3.506 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion:101 Resp: 103028
Ion Ratio Lower Upper
101 100
103 63.5 49.8 74.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#8
Diethyl Ether
Concen: 23.778 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion: 74 Resp: 45341
Ion Ratio Lower Upper
74 100
45 104.4 50.8 152.5





#9
1,1,2-Trichlorotrifluoroethane
Concen: 22.200 ug/l
RT: 4.359 min Scan# 409
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion:101 Resp: 56020
Ion Ratio Lower Upper
101 100
85 50.2 37.3 55.9
151 70.4 58.9 88.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#10
Methyl Iodide
Concen: 17.223 ug/l
RT: 4.571 min Scan# 445
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07



Tgt Ion:142 Resp: 33911
Ion Ratio Lower Upper
142 100
127 52.5 35.7 53.5
141 15.3 10.4 15.6





#11

Tert butyl alcohol

Concen: 114.174 ug/l

RT: 5.530 min Scan# 607

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 59 Resp: 92129

Ion Ratio Lower Upper

59 100





57 11.4 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#12

1,1-Dichloroethene

Concen: 20.609 ug/l

RT: 4.336 min Scan# 405

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion: 96 Resp: 58932


Ion Ratio Lower Upper

96 100

61 175.2 132.3 198.5

98 68.3 46.8 70.2





#13

Acrolein

Concen: 81.801 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 56 Resp: 5297

Ion Ratio Lower Upper

56 100





55 74.1 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#14

Allyl chloride

Concen: 20.512 ug/l

RT: 5.012 min Scan# 520

Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion: 41 Resp: 106153


Ion Ratio Lower Upper

41 100

39 84.4 59.0 88.6


76 35.6 28.7 43.1





#15
Acrylonitrile
Concen: 104.238 ug/l
RT: 5.706 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 53 Resp: 22824
Ion Ratio Lower Upper
53 100
52 85.4 65.5 98.3
51 37.0 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#16
Acetone
Concen: 117.551 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 43 Resp: 234225
Ion Ratio Lower Upper
43 100
58 29.0 22.3 33.5









#17
Carbon Disulfide
Concen: 17.956 ug/l
RT: 4.700 min Scan# 467
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 76 Resp: 152230
Ion Ratio Lower Upper
76 100
78 9.8 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#18
Methyl Acetate
Concen: 24.040 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.012 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 43 Resp: 120344
Ion Ratio Lower Upper
43 100
74 23.2 17.8 26.6









#19
Methyl tert-butyl Ether
Concen: 23.461 ug/l
RT: 5.794 min Scan# 653
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 73 Resp: 24728
Ion Ratio Lower Upper
73 100
57 22.2 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#20
Methylene Chloride
Concen: 21.375 ug/l
RT: 5.259 min Scan# 562
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 84 Resp: 73232
Ion Ratio Lower Upper
84 100
49 128.1 107.5 161.3
51 38.2 30.2 45.2
86 62.8 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 19.394 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 96 Resp: 6253
Ion Ratio Lower Upper
96 100
61 138.4 122.0 183.0
98 61.8 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#22
Diisopropyl ether
Concen: 22.809 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 45 Resp: 247597
Ion Ratio Lower Upper
45 100
43 58.6 42.8 64.2
87 25.3 19.8 29.6
59 12.1 8.3 12.5









#23
Vinyl Acetate
Concen: 118.964 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 43 Resp: 112944
Ion Ratio Lower Upper
43 100
86 9.7 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#24
1,1-Dichloroethane
Concen: 21.564 ug/l
RT: 6.559 min Scan# 783
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion: 63 Resp: 135050
Ion Ratio Lower Upper
63 100
98 7.1 3.3 9.9
100 3.5 2.5 7.4





#25
2-Butanone
Concen: 108.103 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 43 Resp: 33281
Ion Ratio Lower Upper
43 100
72 22.1 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#26
2,2-Dichloropropane
Concen: 23.748 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 77 Resp: 115630
Ion Ratio Lower Upper
77 100
97 20.9 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 21.217 ug/l

RT: 7.477 min Scan# 938

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01





Tgt Ion:	96	Resp:	7875
Ion	Ratio	Lower	Upper
96	100		
61	159.0	0.0	297.8
98	65.0	0.0	132.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#28

Bromochloromethane

Concen: 25.159 ug/l




RT: 7.800 min Scan# 994


Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Tgt Ion:	49	Resp:	75409
Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.2
130	67.2	57.3	85.9





#29

Tetrahydrofuran

Concen: 104.471 ug/l

RT: 7.829 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07





Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion:	42	Resp:	20894
Ion Ratio	Lower	Upper	
42	100		
72	42.6	33.4	50.0
71	41.0	31.2	46.8

Manual Integrations
APPROVEDReviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025

#30

Chloroform

Concen: 22.445 ug/l




RT: 7.953 min Scan# 1020


Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07


Tgt Ion:	83	Resp:	140698
Ion Ratio	Lower	Upper	
83	100		
85	65.5	52.7	79.1





#31
Cyclohexane
Concen: 21.380 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 56 Resp: 111690
Ion Ratio Lower Upper
56 100
69 29.0 22.7 34.1
84 78.7 65.8 98.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#32
1,1,1-Trichloroethane
Concen: 21.362 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 97 Resp: 115982
Ion Ratio Lower Upper
97 100
99 60.1 51.8 77.8
61 49.7 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 57.648 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 65 Resp: 24498

Ion Ratio Lower Upper

65 100





67 50.5 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion:114 Resp: 485218


Ion Ratio Lower Upper

114 100

63 22.7 0.0 44.6

88 15.8 0.0 32.8





#35

Dibromofluoromethane

Concen: 50.692 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion:113 Resp: 16966

Ion Ratio Lower Upper

113 100

111 102.4 82.5 123.7





192 16.1 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#36

1,1-Dichloropropene

Concen: 20.523 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion: 75 Resp: 90752


Ion Ratio Lower Upper

75 100

110 30.8 16.7 50.1

77 31.0 25.2 37.8





#37

Ethyl Acetate

Concen: 20.524 ug/l

RT: 7.547 min Scan# 951

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 43 Resp: 131078

Ion Ratio Lower Upper

43 100

61 12.5 10.9 16.3





70 9.3 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#38

Carbon Tetrachloride

Concen: 19.181 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion:117 Resp: 93434


Ion Ratio Lower Upper

117 100

119 98.4 80.2 120.2

121 26.6 25.4 38.2





#39

Methylcyclohexane

Concen: 20.130 ug/l

RT: 9.582 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :





VN0814WBS01

Tgt Ion:	83	Resp:	96372
Ion Ratio	Lower	Upper	
83	100		
55	83.1	61.3	91.9
98	51.6	37.9	56.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#40

Benzene

Concen: 20.261 ug/l




RT: 8.588 min Scan# 1128


Delta R.T. -0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07


Tgt Ion:	78	Resp:	289566
Ion Ratio	Lower	Upper	
78	100		
77	23.5	18.2	27.2





#41
Methacrylonitrile
Concen: 19.992 ug/l
RT: 7.765 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 41 Resp: 6675
Ion Ratio Lower Upper
41 100
39 61.1 43.4 65.0
67 70.9 55.1 82.7
52 33.6 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#42
1,2-Dichloroethane
Concen: 21.531 ug/l
RT: 8.659 min Scan# 1140
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 62 Resp: 116694
Ion Ratio Lower Upper
62 100
98 7.9 0.0 18.0





#43

Isopropyl Acetate

Concen: 21.696 ug/l

RT: 8.676 min Scan# 1143

Delta R.T. -0.001 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :





VN0814WBS01

Tgt Ion:	43	Resp:	215090
Ion	Ratio	Lower	Upper
43	100		
61	23.3	19.8	29.8
87	12.2	9.8	14.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#44

Trichloroethene

Concen: 19.172 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Tgt Ion:	130	Resp:	64743
Ion	Ratio	Lower	Upper
130	100		
95	115.4	0.0	195.2





#45

1,2-Dichloropropane

Concen: 20.442 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 63 Resp: 7423

Ion Ratio Lower Upper

63 100





65 30.1 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#46

Dibromomethane

Concen: 20.795 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion: 93 Resp: 56540


Ion Ratio Lower Upper

93 100

95 81.5 65.8 98.8

174 76.8 69.9 104.9





#47

Bromodichloromethane

Concen: 20.530 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :





VN0814WBS01

Tgt Ion:	83	Resp:	11243
Ion	Ratio	Lower	Upper
83	100		
85	63.2	50.4	75.6
127	8.3	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#48

Methyl methacrylate

Concen: 22.008 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Tgt Ion:	41	Resp:	98225
Ion	Ratio	Lower	Upper
41	100		
69	78.9	64.1	96.1
39	65.2	45.5	68.3





#49

1,4-Dioxane

Concen: 447.176 ug/l

RT: 9.688 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 88 Resp: 3056

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 76.6 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#50

Toluene-d8

Concen: 50.695 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. -0.000 min

Lab File: VN087555.D




Acq: 14 Aug 2025 11:07


Tgt Ion: 98 Resp: 605258

Ion Ratio Lower Upper

98 100


100 64.5 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 105.450 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 43 Resp: 66120
Ion Ratio Lower Upper
43 100
58 37.3 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#52
Toluene
Concen: 20.443 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion: 92 Resp: 177586
Ion Ratio Lower Upper
92 100
91 172.6 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 20.990 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 75 Resp: 116338
Ion Ratio Lower Upper
75 100
77 30.0 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#54
cis-1,3-Dichloropropene
Concen: 20.420 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion: 75 Resp: 116909
Ion Ratio Lower Upper
75 100
77 33.7 24.2 36.2
39 64.8 38.4 57.6#





#55
1,1,2-Trichloroethane
Concen: 20.652 ug/l
RT: 10.994 min Scan# 1537
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion: 97 Resp: 7263
Ion Ratio Lower Upper
97 100
83 82.8 72.6 109.0
85 57.9 46.7 70.1
99 57.5 54.1 81.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#56
Ethyl methacrylate
Concen: 21.091 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Tgt Ion: 69 Resp: 120174
Ion Ratio Lower Upper
69 100
41 70.4 55.1 82.7
39 44.5 27.9 41.9#





#57
1,3-Dichloropropane
Concen: 21.100 ug/l
RT: 11.141 min Scan# 1562
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion: 76 Resp: 128302
Ion Ratio Lower Upper
76 100
78 32.4 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#58
2-Chloroethyl Vinyl ether
Concen: 113.893 ug/l
RT: 10.141 min Scan# 1392
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Tgt Ion: 63 Resp: 328582
Ion Ratio Lower Upper
63 100
106 25.9 21.7 32.5





#59
2-Hexanone
Concen: 103.681 ug/l
RT: 11.182 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion: 43 Resp: 43132
Ion Ratio Lower Upper
43 100
58 51.5 26.7 80.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#60
Dibromochloromethane
Concen: 19.012 ug/l
RT: 11.335 min Scan# 1595
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Tgt Ion:129 Resp: 76256
Ion Ratio Lower Upper
129 100
127 76.6 39.1 117.5





#61
1,2-Dibromoethane
Concen: 20.492 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion:107 Resp: 75775
Ion Ratio Lower Upper
107 100
109 91.5 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#62
4-Bromofluorobenzene
Concen: 52.926 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion: 95 Resp: 233458
Ion Ratio Lower Upper
95 100
174 65.4 0.0 149.4
176 64.8 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion:117 Resp: 446518
Ion Ratio Lower Upper
117 100
82 58.5 47.4 71.2
119 31.2 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#64
Tetrachloroethene
Concen: 18.090 ug/l
RT: 11.088 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion:164 Resp: 51988
Ion Ratio Lower Upper
164 100
166 126.8 105.5 158.3
129 106.6 77.4 116.2
131 92.4 77.3 115.9









#65
Chlorobenzene
Concen: 19.244 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion:112 Resp: 19291
Ion Ratio Lower Upper
112 100
114 29.1 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 19.247 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion:131 Resp: 65609
Ion Ratio Lower Upper
131 100
133 98.9 47.4 142.3
119 65.8 33.1 99.2





#67
Ethyl Benzene
Concen: 20.367 ug/l
RT: 11.947 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion: 91 Resp: 336124
Ion Ratio Lower Upper
91 100
106 29.4 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#68
m/p-Xylenes
Concen: 40.389 ug/l
RT: 12.053 min Scan# 1717
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07



Tgt Ion:106 Resp: 249595
Ion Ratio Lower Upper
106 100
91 211.9 162.0 243.0









#69
o-Xylene
Concen: 20.320 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion:106 Resp: 119950
Ion Ratio Lower Upper
106 100
91 227.3 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#70
Styrene
Concen: 20.649 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion:104 Resp: 205053
Ion Ratio Lower Upper
104 100
78 55.7 41.0 61.6
103 56.6 43.9 65.9





#71
Bromoform
Concen: 17.538 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion:173 Resp: 48290
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Tgt Ion:152 Resp: 221983
Ion Ratio Lower Upper
152 100
115 65.8 31.1 93.5
150 165.1 0.0 349.0





#73
Isopropylbenzene
Concen: 22.495 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 105 Resp: 31428
Ion Ratio Lower Upper
105 100
120 25.2 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#74
N-ethyl acetate
Concen: 18.694 ug/l m
RT: 12.535 min Scan# 1799
Delta R.T. 0.041 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 43 Resp: 108512
Ion Ratio Lower Upper
43 100
70 16.4 37.6 56.4#
55 3.6 19.6 29.4#
61 3.5 20.6 31.0#





#75

1,1,2,2-Tetrachloroethane

Concen: 21.882 ug/l

RT: 12.923 min Scan# 1865

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion: 83 Resp: 115038

Ion Ratio Lower Upper

83 100

131 9.5 5.1 15.3





85 64.1 32.5 97.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#76

1,2,3-Trichloropropane

Concen: 20.143 ug/l m

RT: 12.976 min Scan# 1874

Delta R.T. 0.006 min

Lab File: VN087555.D




Acq: 14 Aug 2025 11:07


Tgt Ion: 75 Resp: 100267

Ion Ratio Lower Upper

75 100


77 182.8 94.5 283.6





#77
Bromobenzene
Concen: 20.297 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 156 Resp: 73544
Ion Ratio Lower Upper
156 100
77 249.2 114.9 344.6
158 99.8 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#78
n-propylbenzene
Concen: 22.160 ug/l
RT: 13.017 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 91 Resp: 389520
Ion Ratio Lower Upper
91 100
120 21.8 11.3 33.8









#79
2-Chlorotoluene
Concen: 22.353 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion: 91 Resp: 241470
Ion Ratio Lower Upper
91 100
126 30.3 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#80
1,3,5-Trimethylbenzene
Concen: 22.029 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07


Tgt Ion:105 Resp: 262222
Ion Ratio Lower Upper
105 100
120 47.6 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 17.698 ug/l
RT: 12.723 min Scan# 1831
Delta R.T. 0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion: 75 Resp: 3219
Ion Ratio Lower Upper
75 100
53 117.4 83.5 125.3
89 49.5 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#82
4-Chlorotoluene
Concen: 22.292 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion: 91 Resp: 250722
Ion Ratio Lower Upper
91 100
126 30.5 16.6 49.7





#83

tert-Butylbenzene

Concen: 22.422 ug/l

RT: 13.417 min Scan# 1949

Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion:119 Resp: 222918

Ion Ratio Lower Upper

119 100

91 73.3 35.0 105.1





134 27.2 12.6 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#84

1,2,4-Trimethylbenzene

Concen: 22.169 ug/l

RT: 13.464 min Scan# 1957

Delta R.T. 0.000 min

Lab File: VN087555.D




Acq: 14 Aug 2025 11:07


Tgt Ion:105 Resp: 269488

Ion Ratio Lower Upper

105 100

120 44.7 22.8 68.3









#85
sec-Butylbenzene
Concen: 21.998 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Tgt Ion:105 Resp: 329434
Ion Ratio Lower Upper
105 100
134 19.3 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#86
p-Isopropyltoluene
Concen: 22.671 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. -0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion:119 Resp: 272080
Ion Ratio Lower Upper
119 100
134 25.6 13.5 40.5
91 26.4 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 20.458 ug/l

RT: 13.711 min Scan# 1999

Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion:146 Resp: 145480

Ion Ratio Lower Upper

146 100

111 44.2 21.4 64.3





148 64.7 31.6 95.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#88

1,4-Dichlorobenzene

Concen: 19.766 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion:146 Resp: 150121


Ion Ratio Lower Upper

146 100

111 42.6 19.6 58.7

148 64.2 31.4 94.0





#89

n-Butylbenzene

Concen: 23.593 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01





Tgt Ion:	91	Resp:	270370
Ion	Ratio	Lower	Upper
91	100		
92	51.6	26.2	78.6
134	23.3	12.4	37.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#90

Hexachloroethane

Concen: 19.285 ug/l




RT: 14.311 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Tgt Ion:	117	Resp:	49037
Ion	Ratio	Lower	Upper
117	100		
201	62.3	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 20.620 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBS01

Tgt Ion:146 Resp: 138910

Ion Ratio Lower Upper

146 100

111 46.0 22.0 66.0





148 64.4 32.7 98.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/18/2025

Supervised By :Mahesh Dadoda 08/19/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.792 ug/l

RT: 14.700 min Scan# 2167

Delta R.T. -0.000 min

Lab File: VN087555.D

Acq: 14 Aug 2025 11:07




Tgt Ion: 75 Resp: 27318


Ion Ratio Lower Upper

75 100

155 69.0 37.3 111.8


157 83.3 46.2 138.6





#93
1,2,4-Trichlorobenzene
Concen: 21.036 ug/l
RT: 15.364 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01




Tgt Ion:180 Resp: 8324
Ion Ratio Lower Upper
180 100
182 94.2 47.3 141.8
145 31.6 15.4 46.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025




#94
Hexachlorobutadiene
Concen: 21.127 ug/l
RT: 15.476 min Scan# 2299
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Tgt Ion:225 Resp: 31065
Ion Ratio Lower Upper
225 100
223 64.6 32.1 96.3
227 63.2 31.3 93.9





#95
Naphthalene
Concen: 21.296 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01



Tgt Ion:128 Resp: 298554
Ion Ratio Lower Upper
128 100
127 12.4 10.5 15.7
129 10.7 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.891 ug/l
RT: 15.811 min Scan# 2356
Delta R.T. 0.000 min
Lab File: VN087555.D
Acq: 14 Aug 2025 11:07

Tgt Ion:180 Resp: 78959
Ion Ratio Lower Upper
180 100
182 100.0 47.1 141.4
145 35.6 16.9 50.6



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MS	SDG No.:	Q2842
Lab Sample ID:	Q2842-04MS	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087569.D	100	08/14/25 16:27	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5200		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5300		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8100		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5400		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5500		22.0	100	ug/L
79-01-6	Trichloroethene	10000		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.8		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	47.9		70 (75) - 130 (124)	96%	SPK: 50
2037-26-5	Toluene-d8	47.9		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	238000	8.206			
540-36-3	1,4-Difluorobenzene	461000	9.083			
3114-55-4	Chlorobenzene-d5	435000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	232000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087569.D
 Acq On : 14 Aug 2025 16:27
 Operator : JC\MD
 Sample : Q2842-04MS 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MS

Quant Time: Aug 14 19:59:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.206	168	238067	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.083	114	461160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	435202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	231778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	217494	53.842	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.680%			
35) Dibromofluoromethane	8.153	113	152403	47.909	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.820%			
50) Toluene-d8	10.547	98	543343	47.883	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.760%			
62) 4-Bromofluorobenzene	12.829	95	216897	51.737	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.480%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.142	85	143419	56.720	ug/l	100
3) Chloromethane	2.383	50	152078	47.827	ug/l	100
4) Vinyl Chloride	2.542	62	169468	53.629	ug/l	95
5) Bromomethane	2.971	94	89339	54.595	ug/l	96
6) Chloroethane	3.136	64	114275	55.452	ug/l	97
7) Trichlorofluoromethane	3.501	101	253745	54.304	ug/l	97
8) Diethyl Ether	3.965	74	112077	61.834	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	138606	57.785	ug/l	95
10) Methyl Iodide	4.577	142	91710	38.610	ug/l	91
11) Tert butyl alcohol	5.536	59	219902	286.700	ug/l	100
12) 1,1-Dichloroethene	4.330	96	140041	51.521	ug/l	99
13) Acrolein	4.177	56	135355	219.895	ug/l	99
14) Allyl chloride	5.012	41	254609	51.759	ug/l	94
15) Acrylonitrile	5.712	53	562090	270.058	ug/l	98
16) Acetone	4.430	43	530238	279.957	ug/l	98
17) Carbon Disulfide	4.700	76	356757	44.270	ug/l #	93
18) Methyl Acetate	5.018	43	302650	63.602	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	601237	60.011	ug/l	96
20) Methylene Chloride	5.265	84	169920	53.111	ug/l	99
21) trans-1,2-Dichloroethene	5.771	96	153170	49.977	ug/l	93
22) Diisopropyl ether	6.659	45	612749	59.384	ug/l	97
23) Vinyl Acetate	6.594	43	2775483	307.552	ug/l	99
24) 1,1-Dichloroethane	6.553	63	318408	53.488	ug/l	98
25) 2-Butanone	7.471	43	850355	290.579	ug/l	98
26) 2,2-Dichloropropane	7.471	77	249304	53.865	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	284580	80.651	ug/l	97
28) Bromochloromethane	7.800	49	143451	50.351	ug/l	93
29) Tetrahydrofuran	7.830	42	541901	285.049	ug/l	99
30) Chloroform	7.953	83	341066	57.240	ug/l	99
31) Cyclohexane	8.241	56	253936	51.134	ug/l	96
32) 1,1,1-Trichloroethane	8.153	97	279470	54.153	ug/l	97
36) 1,1-Dichloropropene	8.353	75	214559	51.052	ug/l	98
37) Ethyl Acetate	7.553	43	307770	50.705	ug/l	98
38) Carbon Tetrachloride	8.347	117	228035	49.255	ug/l	93
39) Methylcyclohexane	9.583	83	237505	52.198	ug/l	96
40) Benzene	8.588	78	696316	51.262	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087569.D
 Acq On : 14 Aug 2025 16:27
 Operator : JC\MD
 Sample : Q2842-04MS 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MS

Quant Time: Aug 14 19:59:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	175869	55.413	ug/l	98
42) 1,2-Dichloroethane	8.653	62	283996	55.133	ug/l	97
43) Isopropyl Acetate	8.677	43	532619	56.527	ug/l	97
44) Trichloroethene	9.335	130	321363	100.127	ug/l	88
45) 1,2-Dichloropropane	9.606	63	182390	52.845	ug/l	94
46) Dibromomethane	9.694	93	137378	53.162	ug/l	95
47) Bromodichloromethane	9.871	83	277559	53.324	ug/l	96
48) Methyl methacrylate	9.665	41	245497	57.875	ug/l	93
49) 1,4-Dioxane	9.688	88	69859	1075.274	ug/l #	97
51) 4-Methyl-2-Pentanone	10.430	43	1677266	281.447	ug/l	98
52) Toluene	10.612	92	435684	52.770	ug/l	96
53) t-1,3-Dichloropropene	10.818	75	283224	53.765	ug/l	95
54) cis-1,3-Dichloropropene	10.294	75	296099	54.416	ug/l #	88
55) 1,1,2-Trichloroethane	11.000	97	176388	52.770	ug/l	94
56) Ethyl methacrylate	10.859	69	305796	54.213	ug/l	93
57) 1,3-Dichloropropane	11.147	76	313150	54.186	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	584151	213.042	ug/l	99
59) 2-Hexanone	11.182	43	1146887	290.068	ug/l	100
60) Dibromochloromethane	11.341	129	195208	51.209	ug/l	99
61) 1,2-Dibromoethane	11.447	107	185145	52.680	ug/l	98
64) Tetrachloroethene	11.088	164	120472	43.011	ug/l	96
65) Chlorobenzene	11.871	112	472723	48.382	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.941	131	165027	49.672	ug/l	100
67) Ethyl Benzene	11.941	91	844688	52.514	ug/l	99
68) m/p-Xylenes	12.053	106	627919	104.251	ug/l	93
69) o-Xylene	12.376	106	307518	53.449	ug/l	94
70) Styrene	12.394	104	539497	55.741	ug/l	98
71) Bromoform	12.559	173	129055	48.082	ug/l #	99
73) Isopropylbenzene	12.676	105	797605	54.677	ug/l	98
74) N-amyl acetate	12.500	43	204124	33.679	ug/l #	93
75) 1,1,2,2-Tetrachloroethane	12.918	83	285137	51.946	ug/l	99
76) 1,2,3-Trichloropropane	12.976	75	244697m	47.080	ug/l	
77) Bromobenzene	12.959	156	183370	48.469	ug/l	91
78) n-propylbenzene	13.018	91	990768	53.982	ug/l	97
79) 2-Chlorotoluene	13.106	91	600287	53.218	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	686579	55.240	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.718	75	87284	45.950	ug/l	93
82) 4-Chlorotoluene	13.200	91	625581	53.269	ug/l	97
83) tert-Butylbenzene	13.418	119	576103	55.497	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	715753	56.391	ug/l	95
85) sec-Butylbenzene	13.594	105	866181	55.396	ug/l	99
86) p-Isopropyltoluene	13.706	119	703516	56.143	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	370720	49.929	ug/l	99
88) 1,4-Dichlorobenzene	13.794	146	374267	47.195	ug/l	97
89) n-Butylbenzene	14.035	91	697153	58.265	ug/l	98
90) Hexachloroethane	14.312	117	126427	47.619	ug/l	93
91) 1,2-Dichlorobenzene	14.082	146	361117	51.338	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	69760	48.406	ug/l	95
93) 1,2,4-Trichlorobenzene	15.370	180	212082	51.328	ug/l	99
94) Hexachlorobutadiene	15.476	225	79123	51.536	ug/l	97
95) Naphthalene	15.617	128	789585	53.941	ug/l	99
96) 1,2,3-Trichlorobenzene	15.817	180	206242	49.760	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087569.D
Acq On : 14 Aug 2025 16:27
Operator : JC\MD
Sample : Q2842-04MS 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Quant Time: Aug 14 19:59:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Manual Integrations


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1064
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 168 Resp: 23806
Ion Ratio Lower Upper
168 100
99 72.2 47.9 71.9




Manual Integrations
APPROVED


Reviewed By : John Carlone 08/15/2025
Supervised By : Mahesh Dadoda 08/19/2025



#2
Dichlorodifluoromethane
Concen: 56.720 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 85 Resp: 143419
Ion Ratio Lower Upper
85 100
87 35.3 17.8 53.3





#3
Chloromethane
Concen: 47.827 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 50 Resp: 152078
Ion Ratio Lower Upper
50 100
52 33.1 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#4
Vinyl Chloride
Concen: 53.629 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 62 Resp: 169468
Ion Ratio Lower Upper
62 100
64 31.1 27.0 40.6





#5
Bromomethane
Concen: 54.595 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion: 94 Resp: 89339
Ion Ratio Lower Upper
94 100
96 95.4 73.4 110.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#6
Chloroethane
Concen: 55.452 ug/l
RT: 3.136 min Scan# 201
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Tgt Ion: 64 Resp: 114275
Ion Ratio Lower Upper
64 100
66 32.3 24.6 36.8





#7
Trichlorofluoromethane
Concen: 54.304 ug/l
RT: 3.501 min Scan# 20
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 101 Resp: 25374
Ion Ratio Lower Upper
101 100
103 60.0 49.8 74.6




Manual Integrations
APPROVED


Reviewed By : John Carlone 08/15/2025
Supervised By : Mahesh Dadoda 08/19/2025



#8
Diethyl Ether
Concen: 61.834 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 74 Resp: 112077
Ion Ratio Lower Upper
74 100
45 99.5 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 57.785 ug/l
RT: 4.359 min Scan# 409
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Tgt Ion:101 Resp: 138600
Ion Ratio Lower Upper
101 100
85 48.3 37.3 55.9
151 68.6 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#10
Methyl Iodide
Concen: 38.610 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion:142 Resp: 91710
Ion Ratio Lower Upper
142 100
127 51.6 35.7 53.5
141 12.4 10.4 15.6





#11

Tert butyl alcohol

Concen: 286.700 ug/l

RT: 5.536 min Scan# 607

Delta R.T. 0.012 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 59 Resp: 21990

Ion Ratio Lower Upper

59 100





57 10.3 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#12

1,1-Dichloroethene

Concen: 51.521 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion: 96 Resp: 140041


Ion Ratio Lower Upper

96 100

61 165.7 132.3 198.5

98 56.0 46.8 70.2





#13

Acrolein

Concen: 219.895 ug/l

RT: 4.177 min Scan# 378

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 56 Resp: 13535

Ion Ratio Lower Upper

56 100





55 69.8 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#14

Allyl chloride

Concen: 51.759 ug/l

RT: 5.012 min Scan# 520

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion: 41 Resp: 254609


Ion Ratio Lower Upper

41 100

39 80.7 59.0 88.6

76 34.6 28.7 43.1





#15

Acrylonitrile

Concen: 270.058 ug/l

RT: 5.712 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN087569.D


Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS






Tgt Ion	Ratio	Lower	Upper
53	100		
52	82.8	65.5	98.3
51	37.5	28.7	43.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#16

Acetone


Concen: 279.957 ug/l

RT: 4.430 min Scan# 421



Delta R.T. 0.006 min


Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion	Ratio	Lower	Upper
43	100		
58	28.7	22.3	33.5





#17
Carbon Disulfide
Concen: 44.270 ug/l
RT: 4.700 min Scan# 467
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 76 Resp: 35675
Ion Ratio Lower Upper
76 100
78 8.2 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#18
Methyl Acetate
Concen: 63.602 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 43 Resp: 302650
Ion Ratio Lower Upper
43 100
74 22.3 17.8 26.6





#19
Methyl tert-butyl Ether
Concen: 60.011 ug/l
RT: 5.789 min Scan# 601
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion: 73 Resp: 60123
Ion Ratio Lower Upper
73 100
57 21.1 18.6 27.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#20
Methylene Chloride
Concen: 53.111 ug/l
RT: 5.265 min Scan# 563
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27



Tgt Ion: 84 Resp: 169920
Ion Ratio Lower Upper
84 100
49 134.1 107.5 161.3
51 40.2 30.2 45.2
86 62.0 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 49.977 ug/l
RT: 5.771 min Scan# 649
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Tgt Ion: 96 Resp: 153170
Ion Ratio Lower Upper
96 100
61 141.9 122.0 183.0
98 66.8 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#22
Diisopropyl ether
Concen: 59.384 ug/l
RT: 6.659 min Scan# 800
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 45 Resp: 612749
Ion Ratio Lower Upper
45 100
43 57.1 42.8 64.2
87 25.2 19.8 29.6
59 10.5 8.3 12.5





#23

Vinyl Acetate

Concen: 307.552 ug/l

RT: 6.594 min Scan# 789

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 43 Resp: 277548

Ion Ratio Lower Upper

43 100





86 10.1 7.7 11.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#24

1,1-Dichloroethane

Concen: 53.488 ug/l

RT: 6.553 min Scan# 782

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion: 63 Resp: 318408


Ion Ratio Lower Upper

63 100

98 6.9 3.3 9.9


100 3.5 2.5 7.4





#25
2-Butanone
Concen: 290.579 ug/l
RT: 7.471 min Scan# 938
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 43 Resp: 85035
Ion Ratio Lower Upper
43 100
72 23.5 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#26
2,2-Dichloropropane
Concen: 53.865 ug/l
RT: 7.471 min Scan# 938
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 77 Resp: 249304
Ion Ratio Lower Upper
77 100
97 22.4 11.1 33.1





#27

cis-1,2-Dichloroethene

Concen: 80.651 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N





ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion:	96	Resp:	284580
Ion	Ratio	Lower	Upper
96	100		
61	146.1	0.0	297.8
98	63.4	0.0	132.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#28

Bromochloromethane

Concen: 50.351 ug/l




RT: 7.800 min Scan# 994


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:	49	Resp:	143451
Ion	Ratio	Lower	Upper
49	100		
129	1.0	0.0	4.2
130	65.8	57.3	85.9





#29

Tetrahydrofuran

Concen: 285.049 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N





ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion:	42	Resp:	54190
Ion	Ratio	Lower	Upper
42	100		
72	41.0	33.4	50.0
71	38.9	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#30

Chloroform

Concen: 57.240 ug/l




RT: 7.953 min Scan# 1020


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27


Tgt Ion:	83	Resp:	341066
Ion	Ratio	Lower	Upper
83	100		
85	64.9	52.7	79.1





#31
Cyclohexane
Concen: 51.134 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 56 Resp: 253930
Ion Ratio Lower Upper
56 100
69 29.3 22.7 34.1
84 86.7 65.8 98.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#32
1,1,1-Trichloroethane
Concen: 54.153 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 97 Resp: 279470
Ion Ratio Lower Upper
97 100
99 65.8 51.8 77.8
61 51.4 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 53.842 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 65 Resp: 217494

Ion Ratio Lower Upper

65 100





67 50.2 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 1212

Delta R.T. 0.001 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion:114 Resp: 461160


Ion Ratio Lower Upper

114 100

63 24.0 0.0 44.6

88 15.1 0.0 32.8





#35

Dibromofluoromethane

Concen: 47.909 ug/l

RT: 8.153 min Scan# 1053

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion:113 Resp: 15240

Ion Ratio Lower Upper

113 100

111 102.2 82.5 123.7





192 16.1 13.7 20.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#36

1,1-Dichloropropene

Concen: 51.052 ug/l

RT: 8.353 min Scan# 1088

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion: 75 Resp: 214559


Ion Ratio Lower Upper

75 100

110 31.6 16.7 50.1

77 31.4 25.2 37.8





#37

Ethyl Acetate

Concen: 50.705 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 43 Resp: 307770

Ion Ratio Lower Upper

43 100

61 14.1 10.9 16.3





70 10.2 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#38

Carbon Tetrachloride

Concen: 49.255 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion:117 Resp: 228035


Ion Ratio Lower Upper

117 100

119 92.1 80.2 120.2

121 30.9 25.4 38.2





#39

Methylcyclohexane

Concen: 52.198 ug/l

RT: 9.583 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087569.D


Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS






Tgt Ion:	83	Resp:	23750
Ion Ratio	Lower	Upper	
83	100		
55	80.4	61.3	91.9
98	49.6	37.9	56.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#40

Benzene


Concen: 51.262 ug/l

RT: 8.588 min Scan# 1128



Delta R.T. -0.006 min


Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion:	78	Resp:	696316
Ion Ratio	Lower	Upper	
78	100		
77	23.6	18.2	27.2





#41
Methacrylonitrile
Concen: 55.413 ug/l
RT: 7.765 min Scan# 911
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 41 Resp: 175869
Ion Ratio Lower Upper
41 100
39 56.6 43.4 65.0
67 68.3 55.1 82.7
52 31.3 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#42
1,2-Dichloroethane
Concen: 55.133 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 62 Resp: 283996
Ion Ratio Lower Upper
62 100
98 8.0 0.0 18.0





#43

Isopropyl Acetate

Concen: 56.527 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:	43	Resp:	532619
Ion	Ratio	Lower	Upper
43	100		
61	23.0	19.8	29.8
87	11.7	9.8	14.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#44

Trichloroethene

Concen: 100.127 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:	130	Resp:	321363
Ion	Ratio	Lower	Upper
130	100		
95	109.1	0.0	195.2





#45

1,2-Dichloropropane

Concen: 52.845 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 63 Resp: 182390

Ion Ratio Lower Upper

63 100





65 30.5 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#46

Dibromomethane

Concen: 53.162 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion: 93 Resp: 137378


Ion Ratio Lower Upper

93 100

95 79.9 65.8 98.8

174 79.9 69.9 104.9





#47

Bromodichloromethane

Concen: 53.324 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:	83	Resp:	27755
Ion	Ratio	Lower	Upper
83	100		
85	65.9	50.4	75.6
127	8.3	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#48

Methyl methacrylate

Concen: 57.875 ug/l




RT: 9.665 min Scan# 1311


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:	41	Resp:	245497
Ion	Ratio	Lower	Upper
41	100		
69	77.6	64.1	96.1
39	65.8	45.5	68.3





#49

1,4-Dioxane

Concen: 1075.274 ug/l

RT: 9.688 min Scan# 1314

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 88 Resp: 69859

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 73.7 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#50

Toluene-d8

Concen: 47.883 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087569.D




Acq: 14 Aug 2025 16:27


Tgt Ion: 98 Resp: 543343

Ion Ratio Lower Upper

98 100


100 64.5 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 281.447 ug/l
RT: 10.430 min Scan# 1441
Delta R.T. 0.001 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 43 Resp: 167726
Ion Ratio Lower Upper
43 100
58 37.1 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#52
Toluene
Concen: 52.770 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 92 Resp: 435684
Ion Ratio Lower Upper
92 100
91 174.1 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 53.765 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 75 Resp: 283224
Ion Ratio Lower Upper
75 100
77 32.3 23.6 35.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#54
cis-1,3-Dichloropropene
Concen: 54.416 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 75 Resp: 296099
Ion Ratio Lower Upper
75 100
77 32.2 24.2 36.2
39 59.6 38.4 57.6#





#55
1,1,2-Trichloroethane
Concen: 52.770 ug/l
RT: 11.000 min Scan# 1538
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 97 Resp: 17638
Ion Ratio Lower Upper
97 100
83 84.2 72.6 109.0
85 56.5 46.7 70.1
99 62.7 54.1 81.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#56
Ethyl methacrylate
Concen: 54.213 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 69 Resp: 305796
Ion Ratio Lower Upper
69 100
41 73.1 55.1 82.7
39 41.9 27.9 41.9





#57

1,3-Dichloropropane

Concen: 54.186 ug/l

RT: 11.147 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion: 76 Resp: 313150

Ion Ratio Lower Upper

76 100





78 32.4 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#58

2-Chloroethyl Vinyl ether

Concen: 213.042 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087569.D




Acq: 14 Aug 2025 16:27


Tgt Ion: 63 Resp: 584151

Ion Ratio Lower Upper

63 100


106 26.7 21.7 32.5





#59
2-Hexanone
Concen: 290.068 ug/l
RT: 11.182 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion: 43 Resp: 114688
Ion Ratio Lower Upper
43 100
58 53.1 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#60
Dibromochloromethane
Concen: 51.209 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion:129 Resp: 195208
Ion Ratio Lower Upper
129 100
127 78.9 39.1 117.5





#61
1,2-Dibromoethane
Concen: 52.680 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:107 Resp: 18514
Ion Ratio Lower Upper
107 100
109 95.1 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#62
4-Bromofluorobenzene
Concen: 51.737 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion: 95 Resp: 216897
Ion Ratio Lower Upper
95 100
174 67.8 0.0 149.4
176 66.2 0.0 141.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:117 Resp: 43520
Ion Ratio Lower Upper
117 100
82 61.7 47.4 71.2
119 30.9 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#64
Tetrachloroethene
Concen: 43.011 ug/l
RT: 11.088 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion:164 Resp: 120472
Ion Ratio Lower Upper
164 100
166 123.4 105.5 158.3
129 96.3 77.4 116.2
131 93.6 77.3 115.9





#65
Chlorobenzene
Concen: 48.382 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:112 Resp: 47272
Ion Ratio Lower Upper
112 100
114 32.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.672 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion:131 Resp: 165027
Ion Ratio Lower Upper
131 100
133 95.2 47.4 142.3
119 66.5 33.1 99.2





#67
Ethyl Benzene
Concen: 52.514 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion: 91 Resp: 844688
Ion Ratio Lower Upper
91 100
106 30.8 24.3 36.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#68
m/p-Xylenes
Concen: 104.251 ug/l
RT: 12.053 min Scan# 1717
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Tgt Ion:106 Resp: 627919
Ion Ratio Lower Upper
106 100
91 212.8 162.0 243.0





#69
o-Xylene
Concen: 53.449 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:106 Resp: 307518
Ion Ratio Lower Upper
106 100
91 224.3 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#70
Styrene
Concen: 55.741 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion:104 Resp: 539497
Ion Ratio Lower Upper
104 100
78 53.4 41.0 61.6
103 55.2 43.9 65.9





#71
Bromoform
Concen: 48.082 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion:173 Resp: 12905
Ion Ratio Lower Upper
173 100
175 47.3 24.1 72.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Tgt Ion:152 Resp: 231778
Ion Ratio Lower Upper
152 100
115 64.6 31.1 93.5
150 170.8 0.0 349.0





#73
Isopropylbenzene
Concen: 54.677 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion: 105 Resp: 797609
Ion Ratio Lower Upper
105 100
120 25.7 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#74
N-ethyl acetate
Concen: 33.679 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Tgt Ion: 43 Resp: 204124
Ion Ratio Lower Upper
43 100
70 42.1 37.6 56.4
55 22.5 19.6 29.4
61 20.2 20.6 31.0#





#75
1,1,2,2-Tetrachloroethane
Concen: 51.946 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion: 83 Resp: 28513
Ion Ratio Lower Upper
83 100
131 9.2 5.1 15.3
85 65.5 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#76
1,2,3-Trichloropropane
Concen: 47.080 ug/l m
RT: 12.976 min Scan# 1874
Delta R.T. 0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Tgt Ion: 75 Resp: 244697
Ion Ratio Lower Upper
75 100
77 188.3 94.5 283.6





#77
Bromobenzene
Concen: 48.469 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:156 Resp: 183370
Ion Ratio Lower Upper
156 100
77 251.3 114.9 344.6
158 98.0 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#78
n-propylbenzene
Concen: 53.982 ug/l
RT: 13.018 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion: 91 Resp: 990768
Ion Ratio Lower Upper
91 100
120 21.3 11.3 33.8









#79
2-Chlorotoluene
Concen: 53.218 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Tgt Ion: 91 Resp: 60028
Ion Ratio Lower Upper
91 100
126 31.6 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#80
1,3,5-Trimethylbenzene
Concen: 55.240 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion:105 Resp: 686579
Ion Ratio Lower Upper
105 100
120 45.6 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 45.950 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion	Ratio	Lower	Upper
75	100		
53	113.7	83.5	125.3
89	45.2	38.4	57.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#82
4-Chlorotoluene
Concen: 53.269 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27


Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.3	16.6	49.7





#83
tert-Butylbenzene
Concen: 55.497 ug/l
RT: 13.418 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS




Tgt Ion:119 Resp: 57610
Ion Ratio Lower Upper
119 100
91 71.9 35.0 105.1
134 26.6 12.6 37.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#84
1,2,4-Trimethylbenzene
Concen: 56.391 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27



Tgt Ion:105 Resp: 715753
Ion Ratio Lower Upper
105 100
120 42.2 22.8 68.3





#85

sec-Butylbenzene

Concen: 55.396 ug/l

RT: 13.594 min Scan# 1979

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MS

Tgt Ion:105 Resp: 86618

Ion Ratio Lower Upper

105 100





134 19.2 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#86

p-Isopropyltoluene

Concen: 56.143 ug/l

RT: 13.706 min Scan# 1998

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27




Tgt Ion:119 Resp: 703516


Ion Ratio Lower Upper

119 100

134 25.7 13.5 40.5

91 26.3 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 49.929 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:	146	Resp:	370720
Ion	Ratio	Lower	Upper
146	100		
111	43.8	21.4	64.3
148	64.4	31.6	95.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#88

1,4-Dichlorobenzene

Concen: 47.195 ug/l




RT: 13.794 min Scan# 2013


Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:	146	Resp:	374267
Ion	Ratio	Lower	Upper
146	100		
111	43.1	19.6	58.7
148	63.7	31.4	94.0





#89

n-Butylbenzene

Concen: 58.265 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:	91	Resp:	69715
Ion Ratio	Lower	Upper	
91	100		
92	51.5	26.2	78.6
134	23.1	12.4	37.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#90

Hexachloroethane

Concen: 47.619 ug/l




RT: 14.312 min Scan# 2101


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:	117	Resp:	126427
Ion Ratio	Lower	Upper	
117	100		
201	60.0	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 51.338 ug/l

RT: 14.082 min Scan# 2062

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:	146	Resp:	36111
Ion Ratio	Lower	Upper	
146	100		
111	43.8	22.0	66.0
148	63.4	32.7	98.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 48.406 ug/l




RT: 14.700 min Scan# 2167


Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27


Tgt Ion:	75	Resp:	69760
Ion Ratio	Lower	Upper	
75	100		
155	68.3	37.3	111.8
157	89.9	46.2	138.6





#93
1,2,4-Trichlorobenzene
Concen: 51.328 ug/l
RT: 15.370 min Scan# 21208
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS



Tgt Ion:180 Resp: 21208
Ion Ratio Lower Upper
180 100
182 94.3 47.3 141.8
145 33.1 15.4 46.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#94
Hexachlorobutadiene
Concen: 51.536 ug/l
RT: 15.476 min Scan# 2299
Delta R.T. 0.000 min
Lab File: VN087569.D
Acq: 14 Aug 2025 16:27

Tgt Ion:225 Resp: 79123
Ion Ratio Lower Upper
225 100
223 60.9 32.1 96.3
227 60.9 31.3 93.9





#95

Naphthalene

Concen: 53.941 ug/l

RT: 15.617 min Scan# 2323

Delta R.T. 0.000 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MS

Tgt Ion:128	Resp: 78958
Ion Ratio	Lower Upper
128 100	
127 13.4	10.5 15.7
129 10.9	8.4 12.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.760 ug/l




RT: 15.817 min Scan# 2357

Delta R.T. 0.006 min

Lab File: VN087569.D

Acq: 14 Aug 2025 16:27

Tgt Ion:180	Resp: 206242
Ion Ratio	Lower Upper
180 100	
182 94.9	47.1 141.4
145 35.0	16.9 50.6



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MSD	SDG No.:	Q2842
Lab Sample ID:	Q2842-05MSD	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087570.D	100	08/14/25 16:49	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5300		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5400		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8000		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5500		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5300		22.0	100	ug/L
79-01-6	Trichloroethene	9500		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	46.7		70 (75) - 130 (124)	93%	SPK: 50
2037-26-5	Toluene-d8	48.2		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	252000	8.212			
540-36-3	1,4-Difluorobenzene	499000	9.082			
3114-55-4	Chlorobenzene-d5	460000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	240000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087570.D
 Acq On : 14 Aug 2025 16:49
 Operator : JC\MD
 Sample : Q2842-05MSD 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MSD

Quant Time: Aug 14 20:00:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 08/15/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	251785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	498703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	460182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	240017	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.565	65	233131	54.569	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.140%	
35) Dibromofluoromethane	8.153	113	160785	46.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.480%	
50) Toluene-d8	10.547	98	591485	48.202	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.400%	
62) 4-Bromofluorobenzene	12.829	95	228797	50.467	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.940%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	160380	59.972	ug/l	95
3) Chloromethane	2.383	50	164845	49.018	ug/l	95
4) Vinyl Chloride	2.542	62	181197	54.217	ug/l	94
5) Bromomethane	2.971	94	111058	64.170	ug/l	100
6) Chloroethane	3.130	64	124775	57.248	ug/l	97
7) Trichlorofluoromethane	3.506	101	269630	54.560	ug/l	92
8) Diethyl Ether	3.965	74	116948	61.006	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	146884	57.900	ug/l	97
10) Methyl Iodide	4.583	142	119433	46.239	ug/l	95
11) Tert butyl alcohol	5.536	59	241582	297.805	ug/l	99
12) 1,1-Dichloroethene	4.342	96	151233	52.607	ug/l	98
13) Acrolein	4.177	56	121388	186.460	ug/l	99
14) Allyl chloride	5.018	41	274012	52.668	ug/l	93
15) Acrylonitrile	5.706	53	596577	271.011	ug/l	97
16) Acetone	4.424	43	565138	282.126	ug/l	97
17) Carbon Disulfide	4.700	76	383063	44.945	ug/l	97
18) Methyl Acetate	5.024	43	323417	64.264	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	630203	59.475	ug/l	96
20) Methylene Chloride	5.265	84	180635	53.387	ug/l #	89
21) trans-1,2-Dichloroethene	5.777	96	165046	50.918	ug/l	96
22) Diisopropyl ether	6.665	45	653071	59.843	ug/l	98
23) Vinyl Acetate	6.594	43	2963851	310.532	ug/l	99
24) 1,1-Dichloroethane	6.559	63	337459	53.599	ug/l	99
25) 2-Butanone	7.477	43	880575	284.512	ug/l	97
26) 2,2-Dichloropropane	7.477	77	251846	51.450	ug/l	98
27) cis-1,2-Dichloroethene	7.471	96	296951	79.572	ug/l	98
28) Bromochloromethane	7.794	49	161291	53.528	ug/l	92
29) Tetrahydrofuran	7.830	42	579208	288.074	ug/l	99
30) Chloroform	7.953	83	365342	57.974	ug/l	98
31) Cyclohexane	8.241	56	269452	51.302	ug/l	99
32) 1,1,1-Trichloroethane	8.153	97	301283	55.199	ug/l	97
36) 1,1-Dichloropropene	8.359	75	232570	51.172	ug/l	98
37) Ethyl Acetate	7.547	43	344525	52.488	ug/l	98
38) Carbon Tetrachloride	8.347	117	240818	48.100	ug/l	96
39) Methylcyclohexane	9.582	83	251368	51.086	ug/l	95
40) Benzene	8.588	78	748859	50.980	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087570.D
 Acq On : 14 Aug 2025 16:49
 Operator : JC\MD
 Sample : Q2842-05MSD 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MSD

Quant Time: Aug 14 20:00:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	188160	54.823	ug/l	97
42) 1,2-Dichloroethane	8.659	62	295609	53.067	ug/l	97
43) Isopropyl Acetate	8.677	43	557773	54.740	ug/l	99
44) Trichloroethene	9.335	130	330475	95.214	ug/l	96
45) 1,2-Dichloropropane	9.606	63	191212	51.231	ug/l	98
46) Dibromomethane	9.694	93	141089	50.488	ug/l	95
47) Bromodichloromethane	9.871	83	288056	51.174	ug/l	98
48) Methyl methacrylate	9.665	41	268246	58.477	ug/l	94
49) 1,4-Dioxane	9.682	88	74858	1065.478	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	1774583	275.360	ug/l	97
52) Toluene	10.612	92	460377	51.563	ug/l	96
53) t-1,3-Dichloropropene	10.818	75	308221	54.105	ug/l	96
54) cis-1,3-Dichloropropene	10.294	75	308333	52.399	ug/l #	87
55) 1,1,2-Trichloroethane	11.000	97	190288	52.643	ug/l	92
56) Ethyl methacrylate	10.859	69	326690	53.587	ug/l #	95
57) 1,3-Dichloropropane	11.147	76	329546	52.730	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	412625	139.157	ug/l	99
59) 2-Hexanone	11.182	43	1237547	289.435	ug/l	98
60) Dibromochloromethane	11.341	129	205718	49.903	ug/l	98
61) 1,2-Dibromoethane	11.447	107	193708	50.967	ug/l	98
64) Tetrachloroethene	11.082	164	127664	43.104	ug/l	94
65) Chlorobenzene	11.871	112	500916	48.485	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.941	131	176284	50.180	ug/l	99
67) Ethyl Benzene	11.941	91	898707	52.840	ug/l	99
68) m/p-Xylenes	12.053	106	662341	103.996	ug/l	93
69) o-Xylene	12.376	106	328068	53.926	ug/l	96
70) Styrene	12.394	104	567805	55.481	ug/l	98
71) Bromoform	12.559	173	133281	46.961	ug/l #	95
73) Isopropylbenzene	12.676	105	839673	55.585	ug/l	99
74) N-amyl acetate	12.500	43	234757	37.404	ug/l #	94
75) 1,1,2,2-Tetrachloroethane	12.918	83	304955	53.650	ug/l	99
76) 1,2,3-Trichloropropane	12.970	75	297389m	55.254	ug/l	
77) Bromobenzene	12.959	156	197712	50.466	ug/l	94
78) n-propylbenzene	13.018	91	1061543	55.853	ug/l	98
79) 2-Chlorotoluene	13.106	91	637513	54.578	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	725164	56.342	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.718	75	93157	47.359	ug/l	94
82) 4-Chlorotoluene	13.200	91	661408	54.387	ug/l	97
83) tert-Butylbenzene	13.418	119	601617	55.966	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	747848	56.897	ug/l	96
85) sec-Butylbenzene	13.594	105	905664	55.933	ug/l	98
86) p-Isopropyltoluene	13.706	119	748208	57.660	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	393523	51.181	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	395731	48.189	ug/l	98
89) n-Butylbenzene	14.035	91	730169	58.929	ug/l	98
90) Hexachloroethane	14.312	117	136983	49.824	ug/l	91
91) 1,2-Dichlorobenzene	14.088	146	372655	51.160	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	74799	50.121	ug/l	97
93) 1,2,4-Trichlorobenzene	15.370	180	220758	51.594	ug/l	98
94) Hexachlorobutadiene	15.476	225	77907	49.002	ug/l	98
95) Naphthalene	15.617	128	847533	55.913	ug/l	100
96) 1,2,3-Trichlorobenzene	15.817	180	217612	50.701	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087570.D
Acq On : 14 Aug 2025 16:49
Operator : JC\MD
Sample : Q2842-05MSD 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Aug 14 20:00:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Manual Integrations

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:168 Resp: 25178
Ion Ratio Lower Upper
168 100
99 66.6 47.9 71.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#2
Dichlorodifluoromethane
Concen: 59.972 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 85 Resp: 160380
Ion Ratio Lower Upper
85 100
87 32.8 17.8 53.3





#3
Chloromethane
Concen: 49.018 ug/l
RT: 2.383 min Scan# 71
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 50 Resp: 16484
Ion Ratio Lower Upper
50 100
52 35.6 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#4
Vinyl Chloride
Concen: 54.217 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 62 Resp: 181197
Ion Ratio Lower Upper
62 100
64 30.4 27.0 40.6









#5
Bromomethane
Concen: 64.170 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion: 94 Resp: 111058
Ion Ratio Lower Upper
94 100
96 91.8 73.4 110.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#6
Chloroethane
Concen: 57.248 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 64 Resp: 124775
Ion Ratio Lower Upper
64 100
66 32.4 24.6 36.8









#7
Trichlorofluoromethane
Concen: 54.560 ug/l
RT: 3.506 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion: 101 Resp: 269630
Ion Ratio Lower Upper
101 100
103 68.3 49.8 74.6




Manual Integrations
APPROVED


Reviewed By : John Carlone 08/15/2025
Supervised By : Mahesh Dadoda 08/19/2025



#8
Diethyl Ether
Concen: 61.006 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 74 Resp: 116948
Ion Ratio Lower Upper
74 100
45 102.1 50.8 152.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 57.900 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:101 Resp: 146884
Ion Ratio Lower Upper
101 100
85 48.3 37.3 55.9
151 71.3 58.9 88.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#10
Methyl Iodide
Concen: 46.239 ug/l
RT: 4.583 min Scan# 447
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion:142 Resp: 119433
Ion Ratio Lower Upper
142 100
127 48.0 35.7 53.5
141 15.3 10.4 15.6





#11

Tert butyl alcohol

Concen: 297.805 ug/l

RT: 5.536 min Scan# 607

Delta R.T. 0.012 min

Lab File: VN087570.D


Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD



Tgt Ion: 59 Resp: 24158

Ion Ratio Lower Upper

59 100




57 10.9 8.3 12.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#12

1,1-Dichloroethene


Concen: 52.607 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.012 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49





Tgt Ion: 96 Resp: 151233


Ion Ratio Lower Upper

96 100

61 164.0 132.3 198.5

98 54.4 46.8 70.2





#13

Acrolein

Concen: 186.460 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 56 Resp: 121388

Ion Ratio Lower Upper

56 100





55 69.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#14

Allyl chloride

Concen: 52.668 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49




Tgt Ion: 41 Resp: 274012


Ion Ratio Lower Upper

41 100

39 82.5 59.0 88.6


76 35.0 28.7 43.1





#15
Acrylonitrile
Concen: 271.011 ug/l
RT: 5.706 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 53 Resp: 59657
Ion Ratio Lower Upper
53 100
52 84.4 65.5 98.3
51 37.8 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#16
Acetone
Concen: 282.126 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 43 Resp: 565138
Ion Ratio Lower Upper
43 100
58 29.4 22.3 33.5









#17
Carbon Disulfide
Concen: 44.945 ug/l
RT: 4.700 min Scan# 467
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion: 76 Resp: 38306
Ion Ratio Lower Upper
76 100
78 9.6 8.6 13.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#18
Methyl Acetate
Concen: 64.264 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.012 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 43 Resp: 323417
Ion Ratio Lower Upper
43 100
74 22.3 17.8 26.6





#19

Methyl tert-butyl Ether

Concen: 59.475 ug/l

RT: 5.794 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 73 Resp: 63020

Ion Ratio Lower Upper

73 100





57 21.4 18.6 27.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#20

Methylene Chloride

Concen: 53.387 ug/l

RT: 5.265 min Scan# 563

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Tgt Ion: 84 Resp: 180635




Ion Ratio Lower Upper


84 100

49 145.8 107.5 161.3

51 46.8 30.2 45.2#

86 68.9 49.3 73.9









#21
trans-1,2-Dichloroethene
Concen: 50.918 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion: 96 Resp: 165040
Ion Ratio Lower Upper
96 100
61 156.9 122.0 183.0
98 59.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#22
Diisopropyl ether
Concen: 59.843 ug/l
RT: 6.665 min Scan# 801
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 45 Resp: 653071
Ion Ratio Lower Upper
45 100
43 55.6 42.8 64.2
87 25.1 19.8 29.6
59 11.5 8.3 12.5





#23
Vinyl Acetate
Concen: 310.532 ug/l
RT: 6.594 min Scan# 789
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 43 Resp: 296385
Ion Ratio Lower Upper
43 100
86 9.9 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#24
1,1-Dichloroethane
Concen: 53.599 ug/l
RT: 6.559 min Scan# 783
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 63 Resp: 337459
Ion Ratio Lower Upper
63 100
98 6.0 3.3 9.9
100 4.8 2.5 7.4





#25
2-Butanone
Concen: 284.512 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 43 Resp: 88057
Ion Ratio Lower Upper
43 100
72 25.9 19.6 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#26
2,2-Dichloropropane
Concen: 51.450 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 77 Resp: 251846
Ion Ratio Lower Upper
77 100
97 23.0 11.1 33.1





#27
cis-1,2-Dichloroethene
Concen: 79.572 ug/l
RT: 7.471 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 96 Resp: 29695
Ion Ratio Lower Upper
96 100
61 150.5 0.0 297.8
98 63.9 0.0 132.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#28
Bromochloromethane
Concen: 53.528 ug/l
RT: 7.794 min Scan# 993
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 49 Resp: 161291
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.2
130 64.8 57.3 85.9





#29

Tetrahydrofuran

Concen: 288.074 ug/l

RT: 7.830 min Scan# 999

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N





ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion:	42	Resp:	579208
Ion	Ratio	Lower	Upper
42	100		
72	41.2	33.4	50.0
71	39.4	31.2	46.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#30

Chloroform

Concen: 57.974 ug/l




RT: 7.953 min Scan# 1020


Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49


Tgt Ion:	83	Resp:	365342
Ion	Ratio	Lower	Upper
83	100		
85	64.4	52.7	79.1





#31
Cyclohexane
Concen: 51.302 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 56 Resp: 26945
Ion Ratio Lower Upper
56 100
69 28.5 22.7 34.1
84 83.6 65.8 98.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#32
1,1,1-Trichloroethane
Concen: 55.199 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 97 Resp: 301283
Ion Ratio Lower Upper
97 100
99 63.5 51.8 77.8
61 50.9 38.7 58.1





#33

1,2-Dichloroethane-d4

Concen: 54.569 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 65 Resp: 23313

Ion Ratio Lower Upper

65 100





67 51.1 0.0 104.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49




Tgt Ion:114 Resp: 498703


Ion Ratio Lower Upper

114 100

63 23.3 0.0 44.6

88 14.7 0.0 32.8





#35

Dibromofluoromethane

Concen: 46.739 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MSD

Tgt Ion:	113	Resp:	16078
Ion	Ratio	Lower	Upper
113	100		
111	104.8	82.5	123.7
192	16.5	13.7	20.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#36

1,1-Dichloropropene

Concen: 51.172 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Tgt Ion:	75	Resp:	232570
Ion	Ratio	Lower	Upper
75	100		
110	32.3	16.7	50.1
77	30.6	25.2	37.8





#37

Ethyl Acetate

Concen: 52.488 ug/l

RT: 7.547 min Scan# 951

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 43 Resp: 34452

Ion Ratio Lower Upper

43 100

61 12.5 10.9 16.3





70 9.3 7.4 11.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#38

Carbon Tetrachloride

Concen: 48.100 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49




Tgt Ion:117 Resp: 240818


Ion Ratio Lower Upper

117 100

119 95.8 80.2 120.2

121 29.6 25.4 38.2





#39

Methylcyclohexane

Concen: 51.086 ug/l

RT: 9.582 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 83 Resp: 251368

Ion Ratio Lower Upper

83 100

55 83.5 61.3 91.9





98 48.1 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#40

Benzene

Concen: 50.980 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VN087570.D




Acq: 14 Aug 2025 16:49


Tgt Ion: 78 Resp: 748859

Ion Ratio Lower Upper

78 100


77 24.4 18.2 27.2





#41
Methacrylonitrile
Concen: 54.823 ug/l
RT: 7.765 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 41 Resp: 188160
Ion Ratio Lower Upper
41 100
39 57.2 43.4 65.0
67 67.2 55.1 82.7
52 30.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#42
1,2-Dichloroethane
Concen: 53.067 ug/l
RT: 8.659 min Scan# 1140
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 62 Resp: 295609
Ion Ratio Lower Upper
62 100
98 7.9 0.0 18.0





#43

Isopropyl Acetate

Concen: 54.740 ug/l

RT: 8.677 min Scan# 1143

Delta R.T. -0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MSD

Tgt Ion:	43	Resp:	55777
Ion	Ratio	Lower	Upper
43	100		
61	24.2	19.8	29.8
87	12.1	9.8	14.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#44

Trichloroethene

Concen: 95.214 ug/l




RT: 9.335 min Scan# 1255


Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Tgt Ion:	130	Resp:	330475
Ion	Ratio	Lower	Upper
130	100		
95	101.4	0.0	195.2





#45

1,2-Dichloropropane

Concen: 51.231 ug/l

RT: 9.606 min Scan# 1301

Delta R.T. -0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 63 Resp: 19121

Ion Ratio Lower Upper

63 100





65 32.4 27.0 40.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#46

Dibromomethane

Concen: 50.488 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49




Tgt Ion: 93 Resp: 141089


Ion Ratio Lower Upper

93 100

95 84.5 65.8 98.8

174 80.4 69.9 104.9





#47

Bromodichloromethane

Concen: 51.174 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. 0.000 min

Lab File: VN087570.D


Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD






Tgt Ion:	83	Resp:	28805
Ion	Ratio	Lower	Upper
83	100		
85	64.4	50.4	75.6
127	7.4	5.8	8.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#48

Methyl methacrylate


Concen: 58.477 ug/l

RT: 9.665 min Scan# 1311



Delta R.T. 0.000 min


Lab File: VN087570.D

Acq: 14 Aug 2025 16:49



Tgt Ion:	41	Resp:	268246
Ion	Ratio	Lower	Upper
41	100		
69	77.2	64.1	96.1
39	64.3	45.5	68.3





#49

1,4-Dioxane

Concen: 1065.478 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 88 Resp: 7485

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0





58 76.0 61.1 91.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#50

Toluene-d8

Concen: 48.202 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087570.D




Acq: 14 Aug 2025 16:49


Tgt Ion: 98 Resp: 591485

Ion Ratio Lower Upper

98 100


100 63.5 52.1 78.1





#51
4-Methyl-2-Pentanone
Concen: 275.360 ug/l
RT: 10.429 min Scan# 1441
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 43 Resp: 177458
Ion Ratio Lower Upper
43 100
58 36.9 30.8 46.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#52
Toluene
Concen: 51.563 ug/l
RT: 10.612 min Scan# 1472
Delta R.T. -0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 92 Resp: 460377
Ion Ratio Lower Upper
92 100
91 175.0 135.1 202.7





#53
t-1,3-Dichloropropene
Concen: 54.105 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion: 75 Resp: 30822
Ion Ratio Lower Upper
75 100
77 31.5 23.6 35.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#54
cis-1,3-Dichloropropene
Concen: 52.399 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Tgt Ion: 75 Resp: 308333
Ion Ratio Lower Upper
75 100
77 33.1 24.2 36.2
39 60.3 38.4 57.6#





#55
1,1,2-Trichloroethane
Concen: 52.643 ug/l
RT: 11.000 min Scan# 1538
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion: 97 Resp: 190288
Ion Ratio Lower Upper
97 100
83 84.6 72.6 109.0
85 51.8 46.7 70.1
99 59.5 54.1 81.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#56
Ethyl methacrylate
Concen: 53.587 ug/l
RT: 10.859 min Scan# 1514
Delta R.T. -0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49



Tgt Ion: 69 Resp: 326690
Ion Ratio Lower Upper
69 100
41 69.8 55.1 82.7
39 42.5 27.9 41.9#





#57

1,3-Dichloropropane

Concen: 52.730 ug/l

RT: 11.147 min Scan# 1563

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD

Tgt Ion: 76 Resp: 329540

Ion Ratio Lower Upper

76 100





78 31.9 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#58

2-Chloroethyl Vinyl ether

Concen: 139.157 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. 0.000 min

Lab File: VN087570.D




Acq: 14 Aug 2025 16:49


Tgt Ion: 63 Resp: 412625

Ion Ratio Lower Upper

63 100


106 26.7 21.7 32.5





#59
2-Hexanone
Concen: 289.435 ug/l
RT: 11.182 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 43 Resp: 123754
Ion Ratio Lower Upper
43 100
58 51.6 26.7 80.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#60
Dibromochloromethane
Concen: 49.903 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion:129 Resp: 205718
Ion Ratio Lower Upper
129 100
127 80.0 39.1 117.5





#61
1,2-Dibromoethane
Concen: 50.967 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion:107 Resp: 193708
Ion Ratio Lower Upper
107 100
109 95.3 77.5 116.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#62
4-Bromofluorobenzene
Concen: 50.467 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 95 Resp: 228797
Ion Ratio Lower Upper
95 100
174 69.5 0.0 149.4
176 65.8 0.0 141.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.847 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN087570.D


Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD



Tgt Ion:117 Resp: 46018

Ion Ratio Lower Upper

117 100

82 59.6 47.4 71.2




119 32.0 23.8 35.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#64

Tetrachloroethene


Concen: 43.104 ug/l

RT: 11.082 min Scan# 1552

Delta R.T. -0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49



Tgt Ion:164 Resp: 127664



Ion Ratio Lower Upper


164 100

166 125.4 105.5 158.3

129 102.7 77.4 116.2

131 101.9 77.3 115.9









#65
Chlorobenzene
Concen: 48.485 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:112 Resp: 500910
Ion Ratio Lower Upper
112 100
114 32.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 50.180 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion:131 Resp: 176284
Ion Ratio Lower Upper
131 100
133 93.4 47.4 142.3
119 65.7 33.1 99.2





#67
Ethyl Benzene
Concen: 52.840 ug/l
RT: 11.941 min Scan# 1699
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 91 Resp: 89870
Ion Ratio Lower Upper
91 100
106 29.8 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#68
m/p-Xylenes
Concen: 103.996 ug/l
RT: 12.053 min Scan# 1717
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion:106 Resp: 662341
Ion Ratio Lower Upper
106 100
91 213.4 162.0 243.0





#69
o-Xylene
Concen: 53.926 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion:106 Resp: 328068
Ion Ratio Lower Upper
106 100
91 222.1 107.7 323.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#70
Styrene
Concen: 55.481 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion:104 Resp: 567805
Ion Ratio Lower Upper
104 100
78 53.5 41.0 61.6
103 55.8 43.9 65.9





#71
Bromoform
Concen: 46.961 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion:173 Resp: 13328
Ion Ratio Lower Upper
173 100
175 51.5 24.1 72.3
254 0.1 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Tgt Ion:152 Resp: 240017
Ion Ratio Lower Upper
152 100
115 64.3 31.1 93.5
150 173.6 0.0 349.0





#73
Isopropylbenzene
Concen: 55.585 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion: 105 Resp: 83967
Ion Ratio Lower Upper
105 100
120 26.0 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#74
N-ethyl acetate
Concen: 37.404 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Tgt Ion: 43 Resp: 234757
Ion Ratio Lower Upper
43 100
70 46.9 37.6 56.4
55 18.2 19.6 29.4#
61 20.4 20.6 31.0#





#75
1,1,2,2-Tetrachloroethane
Concen: 53.650 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion: 83 Resp: 304958
Ion Ratio Lower Upper
83 100
131 9.3 5.1 15.3
85 63.9 32.5 97.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#76
1,2,3-Trichloropropane
Concen: 55.254 ug/l m
RT: 12.970 min Scan# 1873
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Tgt Ion: 75 Resp: 297389
Ion Ratio Lower Upper
75 100
77 162.0 94.5 283.6





#77
Bromobenzene
Concen: 50.466 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 156 Resp: 197712
Ion Ratio Lower Upper
156 100
77 243.7 114.9 344.6
158 98.0 48.5 145.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#78
n-propylbenzene
Concen: 55.853 ug/l
RT: 13.018 min Scan# 1881
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion: 91 Resp: 1061543
Ion Ratio Lower Upper
91 100
120 21.4 11.3 33.8









#79
2-Chlorotoluene
Concen: 54.578 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion: 91 Resp: 63751
Ion Ratio Lower Upper
91 100
126 31.5 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#80
1,3,5-Trimethylbenzene
Concen: 56.342 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion:105 Resp: 725164
Ion Ratio Lower Upper
105 100
120 46.0 24.3 72.8





#81
trans-1,4-Dichloro-2-butene
Concen: 47.359 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD



Tgt Ion: 75 Resp: 9315
Ion Ratio Lower Upper
75 100
53 112.1 83.5 125.3
89 45.5 38.4 57.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#82
4-Chlorotoluene
Concen: 54.387 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49


Tgt Ion: 91 Resp: 661408
Ion Ratio Lower Upper
91 100
126 31.7 16.6 49.7





#83
tert-Butylbenzene
Concen: 55.966 ug/l
RT: 13.418 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD




Tgt Ion:119 Resp: 60161
Ion Ratio Lower Upper
119 100
91 72.4 35.0 105.1
134 26.6 12.6 37.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025




#84
1,2,4-Trimethylbenzene
Concen: 56.897 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49



Tgt Ion:105 Resp: 747848
Ion Ratio Lower Upper
105 100
120 42.8 22.8 68.3









#85
sec-Butylbenzene
Concen: 55.933 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:105 Resp: 905664
Ion Ratio Lower Upper
105 100
134 18.7 9.8 29.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#86
p-Isopropyltoluene
Concen: 57.660 ug/l
RT: 13.706 min Scan# 1998
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion:119 Resp: 748208
Ion Ratio Lower Upper
119 100
134 25.1 13.5 40.5
91 26.0 12.2 36.6





#87

1,3-Dichlorobenzene

Concen: 51.181 ug/l

RT: 13.712 min Scan# 1999

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MSD

Tgt Ion:	146	Resp:	39352
Ion Ratio	Lower	Upper	
146	100		
111	42.6	21.4	64.3
148	62.9	31.6	95.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#88

1,4-Dichlorobenzene

Concen: 48.189 ug/l




RT: 13.788 min Scan# 1212


Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Tgt Ion:	146	Resp:	395731
Ion Ratio	Lower	Upper	
146	100		
111	42.0	19.6	58.7
148	62.7	31.4	94.0









#89
n-Butylbenzene
Concen: 58.929 ug/l
RT: 14.035 min Scan# 2054
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:	91	Resp:	730169
Ion Ratio	Lower	Upper	
91	100		
92	53.1	26.2	78.6
134	23.0	12.4	37.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#90
Hexachloroethane
Concen: 49.824 ug/l
RT: 14.312 min Scan# 2101
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion:	117	Resp:	136983
Ion Ratio	Lower	Upper	
117	100		
201	58.6	32.8	98.4





#91

1,2-Dichlorobenzene

Concen: 51.160 ug/l

RT: 14.088 min Scan# 2062

Delta R.T. 0.006 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-081225MSD

Tgt Ion:	146	Resp:	37265
Ion Ratio	Lower	Upper	
146	100		
111	46.0	22.0	66.0
148	64.6	32.7	98.1

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#92

1,2-Dibromo-3-Chloropropane

Concen: 50.121 ug/l




RT: 14.700 min Scan# 2167


Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49

Tgt Ion:	75	Resp:	74799
Ion Ratio	Lower	Upper	
75	100		
155	71.2	37.3	111.8
157	90.8	46.2	138.6





#93

1,2,4-Trichlorobenzene

Concen: 51.594 ug/l

RT: 15.370 min Scan# 2281

Delta R.T. 0.000 min

Lab File: VN087570.D


Acq: 14 Aug 2025 16:49

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-081225MSD



Tgt Ion:180 Resp: 220758

Ion Ratio Lower Upper

180 100

182 95.2 47.3 141.8




145 33.7 15.4 46.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 08/15/2025

Supervised By :Mahesh Dadoda 08/19/2025



#94

Hexachlorobutadiene


Concen: 49.002 ug/l

RT: 15.476 min Scan# 2299

Delta R.T. 0.000 min

Lab File: VN087570.D

Acq: 14 Aug 2025 16:49





Tgt Ion:225 Resp: 77907


Ion Ratio Lower Upper

225 100

223 66.5 32.1 96.3

227 63.9 31.3 93.9









#95
Naphthalene
Concen: 55.913 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Tgt Ion:128 Resp: 84753
Ion Ratio Lower Upper
128 100
127 13.2 10.5 15.7
129 10.4 8.4 12.6




Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025



#96
1,2,3-Trichlorobenzene
Concen: 50.701 ug/l
RT: 15.817 min Scan# 2357
Delta R.T. 0.006 min
Lab File: VN087570.D
Acq: 14 Aug 2025 16:49

Tgt Ion:180 Resp: 217612
Ion Ratio Lower Upper
180 100
182 93.8 47.1 141.4
145 36.1 16.9 50.6



Manual Integration Report

Sequence:	VN071625	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN087328.D	1,2,3-Trichloropropane	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	1,4-Dichlorobenzene	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Acetone	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Allyl chloride	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Methyl tert-butyl Ether	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	N-amyl acetate	MMDadod a	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	1,2,3-Trichloropropane	MMDadod a	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	Methyl Iodide	MMDadod a	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	N-amyl acetate	MMDadod a	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC020	VN087330.D	1,2,3-Trichloropropane	MMDadod a	7/17/2025 8:19:54 AM	Sam	7/17/2025 8:24:56 AM	Peak Integrated by Software
VSTDICC020	VN087330.D	N-amyl acetate	MMDadod a	7/17/2025 8:19:54 AM	Sam	7/17/2025 8:24:56 AM	Peak Integrated by Software
VSTDICCC050	VN087331.D	1,2,3-Trichloropropane	MMDadod a	7/17/2025 8:19:56 AM	Sam	7/17/2025 8:25:01 AM	Peak Integrated by Software
VSTDICCC050	VN087331.D	N-amyl acetate	MMDadod a	7/17/2025 8:19:56 AM	Sam	7/17/2025 8:25:01 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VN071625	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN087332.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:58 AM	Sam	7/17/2025 8:25:00 AM	Peak Integrated by Software
VSTDICC150	VN087333.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:59 AM	Sam	7/17/2025 8:25:02 AM	Peak Integrated by Software
VSTDICV050	VN087335.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:20:01 AM	Sam	7/17/2025 8:25:03 AM	Peak Integrated by Software
VSTDICV050	VN087335.D	N-amyl acetate	MMDadoda	7/17/2025 8:20:01 AM	Sam	7/17/2025 8:25:03 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VN081425	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN087552.D	1,2,3-Trichloropropane	JOHN	8/14/2025 4:35:55 PM	MMDadoda	8/19/2025 1:44:04 AM	Peak Integrated by Software
VN0814WBS01	VN087555.D	1,2,3-Trichloropropane	JOHN	8/18/2025 8:11:28 AM	MMDadoda	8/19/2025 1:44:07 AM	Peak Integrated by Software
VN0814WBS01	VN087555.D	N-amyl acetate	JOHN	8/18/2025 8:11:28 AM	MMDadoda	8/19/2025 1:44:07 AM	Peak Integrated by Software
Q2842-04MS	VN087569.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:03 AM	MMDadoda	8/19/2025 1:44:16 AM	Peak Integrated by Software
Q2842-05MSD	VN087570.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:08 AM	MMDadoda	8/19/2025 1:44:19 AM	Peak Integrated by Software
VSTDCCC050	VN087571.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:12 AM	MMDadoda	8/19/2025 1:44:23 AM	Peak Integrated by Software

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN071625

Review By	Mahesh Dadoda	Review On	7/17/2025 8:20:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:25:10 AM
SubDirectory	VN071625	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134794		
Initial Calibration Stds	VP134795,VP134796,VP134797,VP134798,VP134799,VP134800		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP134801		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087327.D	16 Jul 2025 16:10	JC\MD	Ok
2	VSTDIC001	VN087328.D	16 Jul 2025 17:05	JC\MD	Ok,M
3	VSTDIC005	VN087329.D	16 Jul 2025 17:27	JC\MD	Ok,M
4	VSTDIC020	VN087330.D	16 Jul 2025 17:49	JC\MD	Ok,M
5	VSTDIC050	VN087331.D	16 Jul 2025 18:11	JC\MD	Ok,M
6	VSTDIC100	VN087332.D	16 Jul 2025 18:32	JC\MD	Ok,M
7	VSTDIC150	VN087333.D	16 Jul 2025 18:54	JC\MD	Ok,M
8	IBLK	VN087334.D	16 Jul 2025 19:16	JC\MD	Ok
9	VSTDICV050	VN087335.D	16 Jul 2025 19:59	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135139		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087551.D	14 Aug 2025 09:01	JC\MD	Ok
2	VSTDCCC050	VN087552.D	14 Aug 2025 09:34	JC\MD	Ok,M
3	VN0814MBL01	VN087553.D	14 Aug 2025 10:10	JC\MD	Ok
4	VN0814WBL01	VN087554.D	14 Aug 2025 10:32	JC\MD	Ok
5	VN0814WBS01	VN087555.D	14 Aug 2025 11:07	JC\MD	Ok,M
6	VN0814WBSD01	VN087556.D	14 Aug 2025 11:41	JC\MD	Ok,M
7	Q2832-02	VN087557.D	14 Aug 2025 12:03	JC\MD	Ok
8	Q2832-04	VN087558.D	14 Aug 2025 12:24	JC\MD	Ok,M
9	Q2836-01	VN087559.D	14 Aug 2025 12:46	JC\MD	Ok
10	Q2836-05	VN087560.D	14 Aug 2025 13:08	JC\MD	Ok
11	Q2836-09	VN087561.D	14 Aug 2025 13:30	JC\MD	Ok
12	Q2836-13	VN087562.D	14 Aug 2025 13:52	JC\MD	Ok
13	Q2838-04	VN087563.D	14 Aug 2025 14:14	JC\MD	Ok
14	Q2838-08	VN087564.D	14 Aug 2025 14:36	JC\MD	Ok
15	Q2842-08	VN087565.D	14 Aug 2025 14:58	JC\MD	Ok
16	Q2842-01	VN087566.D	14 Aug 2025 15:20	JC\MD	Ok
17	Q2842-02	VN087567.D	14 Aug 2025 15:42	JC\MD	Ok
18	Q2842-03	VN087568.D	14 Aug 2025 16:04	JC\MD	Ok
19	Q2842-04MS	VN087569.D	14 Aug 2025 16:27	JC\MD	Ok,M
20	Q2842-05MSD	VN087570.D	14 Aug 2025 16:49	JC\MD	Ok,M
21	VSTDCCC050	VN087571.D	14 Aug 2025 17:11	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN071625

Review By	Maresh Dadoda	Review On	7/17/2025 8:20:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:25:10 AM
SubDirectory	VN071625	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134794
Initial Calibration Stds	VP134795,VP134796,VP134797,VP134798,VP134799,VP134800
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134801
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087327.D	16 Jul 2025 16:10		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN087328.D	16 Jul 2025 17:05		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN087329.D	16 Jul 2025 17:27	% d fail for com.#10 in 5 ppb	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN087330.D	16 Jul 2025 17:49	LR- 10,20	JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN087331.D	16 Jul 2025 18:11	QR- 56	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN087332.D	16 Jul 2025 18:32		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN087333.D	16 Jul 2025 18:54		JC\MD	Ok,M
8	IBLK	IBLK	VN087334.D	16 Jul 2025 19:16		JC\MD	Ok
9	VSTDICV050	ICVVN071625	VN087335.D	16 Jul 2025 19:59		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135139
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087551.D	14 Aug 2025 09:01		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087552.D	14 Aug 2025 09:34	pH#Lot#V12668	JC\MD	Ok,M
3	VN0814MBL01	VN0814MBL01	VN087553.D	14 Aug 2025 10:10		JC\MD	Ok
4	VN0814WBL01	VN0814WBL01	VN087554.D	14 Aug 2025 10:32		JC\MD	Ok
5	VN0814WBS01	VN0814WBS01	VN087555.D	14 Aug 2025 11:07		JC\MD	Ok,M
6	VN0814WBSD01	VN0814WBSD01	VN087556.D	14 Aug 2025 11:41		JC\MD	Ok,M
7	Q2832-02	TG-S01	VN087557.D	14 Aug 2025 12:03	vial B pH#5.0	JC\MD	Ok
8	Q2832-04	TG-S02	VN087558.D	14 Aug 2025 12:24	vial B pH#5.0	JC\MD	Ok,M
9	Q2836-01	WC-A2-15-G	VN087559.D	14 Aug 2025 12:46	vial B pH#5.0	JC\MD	Ok
10	Q2836-05	WC-A2-16-G	VN087560.D	14 Aug 2025 13:08	vial B pH#5.0	JC\MD	Ok
11	Q2836-09	WC-A2-17-G	VN087561.D	14 Aug 2025 13:30	vial B pH#5.0	JC\MD	Ok
12	Q2836-13	WC-A5-02-G	VN087562.D	14 Aug 2025 13:52	vial B pH#5.0	JC\MD	Ok
13	Q2838-04	TP-11	VN087563.D	14 Aug 2025 14:14	vial B pH#5.0	JC\MD	Ok
14	Q2838-08	TP-10	VN087564.D	14 Aug 2025 14:36	vial B pH#5.0	JC\MD	Ok
15	Q2842-08	TB01-081225	VN087565.D	14 Aug 2025 14:58	vial A pH<2 TB	JC\MD	Ok
16	Q2842-01	RMW-03B-90.4-081225	VN087566.D	14 Aug 2025 15:20	vial A pH<2	JC\MD	Ok
17	Q2842-02	RMW-02B-66.3-081225	VN087567.D	14 Aug 2025 15:42	vial A pH<2	JC\MD	Ok
18	Q2842-03	MW-17B-55.5-081225	VN087568.D	14 Aug 2025 16:04	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135139		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141		
19	Q2842-04MS	MW-17B-55.5-081225M	VN087569.D 14 Aug 2025 16:27 vial A pH<2 JC\MD Ok,M
20	Q2842-05MSD	MW-17B-55.5-081225M	VN087570.D 14 Aug 2025 16:49 vial A pH<2 JC\MD Ok,M
21	VSTDCCC050	VSTDCCC050EC	VN087571.D 14 Aug 2025 17:11 JC\MD Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2842

Test : VOCMS Group3

Prepbatch ID :

Sequence ID/Qc Batch ID: VN081425,

Standard ID :

VP134142,VP134149,VP134933,VP134956,VP134957,VP135059,VP135129,VP135139,VP135140,VP135141,

Chemical ID :

V13391,V14290,V14444,V14446,V14507,V14508,V14529,V14530,V14625,V14626,V14636,V14637,V14638,V14639,V14668,V14671,V14673,V14675,V14702,V14705,V14716,V14745,V14751,V14806,V14807,V14843,V14906,V14929,V15057,V15058,V15059,W3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP134142	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 1.00000ml of V14668 + 1.00000ml of V14671 + 1.00000ml of V14673 + 1.00000ml of V14675 + 16.00000ml of V14929 = Final
Quantity: 20.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP134149	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								06/10/2025

FROM 1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final
Quantity: 50.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP134933	07/29/2025	01/29/2026	Semsettin Yesilyurt	None	None	Maresh Dadoda
08/19/2025								

FROM 0.40000ml of V14906 + 24.60000ml of V14625 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP134956	08/01/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
08/06/2025								

FROM 0.25000ml of V14290 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP134957	08/01/2025	11/22/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/06/2025

FROM 0.50000ml of V13391 + 49.50000ml of V14625 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP135059	08/11/2025	09/19/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/19/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14444 + 1.00000ml of V14446 + 1.00000ml of V14507 + 1.00000ml of V14508 + 1.00000ml of V14529 + 1.00000ml of V14530 + 1.00000ml of V14705 + 1.00000ml of V14745 + 1.00000ml of V14751 + 1.00000ml of V14806 + 1.00000ml of V14807 + 1.50000ml of V14702 + 1.50000ml of V14716 + 10.60000ml of V14625 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP135129	08/14/2025	09/13/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/19/2025

FROM 1.00000ml of V15059 + 1.50000ml of V15057 + 1.50000ml of V15058 + 21.00000ml of V14625 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP135139	08/14/2025	08/15/2025	John Carlone	None	None	Maresh Dadoda
								08/19/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP134957 = Final Quantity: 40.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP135140	08/14/2025	08/15/2025	John Carlone	None	None	Maresh Dadoda
FROM 39.94450ml of W3112 + 0.00500ml of VP134142 + 0.00500ml of VP134933 + 0.00800ml of VP134956 + 0.01250ml of VP134149 + 0.01250ml of VP135059 + 0.01250ml of VP135129 = Final Quantity: 40.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP135141	08/14/2025	08/15/2025	John Carlone	None	None	Maresh Dadoda
FROM 39.94450ml of W3112 + 0.00500ml of VP134142 + 0.00500ml of VP134933 + 0.00800ml of VP134956 + 0.01250ml of VP134149 + 0.01250ml of VP135059 + 0.01250ml of VP135129 = Final Quantity: 40.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	08/08/2025 / SAM	08/15/2024 / SAM	V14444

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	08/08/2025 / SAM	08/15/2024 / SAM	V14446

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	02/08/2026	08/08/2025 / SAM	09/17/2024 / SAM	V14507

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	02/08/2026	08/08/2025 / SAM	09/17/2024 / SAM	V14508

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	02/08/2026	08/08/2025 / SAM	09/18/2024 / SAM	V14529

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	02/08/2026	08/08/2025 / SAM	09/18/2024 / SAM	V14530

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	2310762004	01/29/2026	07/29/2025 / SAM	11/26/2024 / SAM	V14625

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	2310762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14671

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14675

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14705

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14716

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14745

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14751

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	02/08/2026	08/08/2025 / SAM	01/08/2025 / SAM	V14806

LOTS

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	02/08/2026	08/08/2025 / SAM	01/08/2025 / SAM	V14807

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0223904	07/29/2026	07/29/2025 / SAM	03/24/2025 / SAM	V14906

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15057

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15058

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15059

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance


Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC



Ken Koehnlein
Sr. Manager, Quality Assurance



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12

Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Flask Uncertainty

5E-05 Balance Uncertainty

Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(KME)	Lot	DR	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information		
														(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3460mg/kg
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg
11. Methyl acrylate	(0075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 787mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG02JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Pentachloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	N/A	or-rat 848mg/kg
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	158-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	35171	101623	0.05	5.00	40003.2	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	or-rat 725mg/kg
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1800mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	108-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloropropane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloroethane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10081-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10081-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	or-rat 2699mg/kg
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1060mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-rat




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetone	12.00
5	Isodimethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methoxyvinylbenzene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethane	28.73
23	Toluene	29.34
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.24
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.85
28	1,2-Dibromomethane	33.85
29	Chlorobenzene	35.33
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	35.33
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.33
33	Styrene	35.33
34	Isopropylbenzene/Bromofluorane	35.33
35	cis-1,4-Dichloro-3-butene	36.48
36	1,1,2,2-Tetrachloroethane	36.80
37	1,2,2-Trichloropropane	37.27
38	n-Propylbenzene	37.92
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.62
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropylbenzene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.26
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		
Methanol	METHYL ALCOHOL	CAS#: 67-56-1
		% (optional) > 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		


Compound	(K049)	Lot	Dir.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS Information			
	Part Number	Number	Factor	Vol. (mL)	Conc. (µg/mL)	Conc. (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc. (µg/mL)	Uncertainty	(+/-) (µg/mL)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2450mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	or-rat 14800mg/kg	
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm(28mg/m3/8H)(skin)	or-rat 75mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H)(skin)	or-rat 120mg/kg	
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm(35mg/m3/8H)(skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKBW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 787mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H)(skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	158-59-2	N/A	N/A	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	158-59-2	N/A	N/A	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg	
22. 1,1-Dichloroethane	32261	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 933mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	78-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	N/A	or-rat 106mg/kg	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	127-18-4	25 ppm (170mg/m3/8H)(final)	or-rat 2629mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40018.5	2000	NA	NA	0.017	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-12-8	0.001 ppm	or-rat 170mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-83-4	20 ppm (8H)	or-rat 108mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	1999.8	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg	
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	1999.8	23.0	10081-01-5	N/A	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	10081-02-6	N/A	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	1999.8	23.0	10081-02-6	N/A	N/A	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	1999.8	23.0	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.8	22.9	830-20-6	N/A	or-rat 670mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H)(skin)	or-rat 800mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H)(skin)	or-rat 936mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 240mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-1	N/A	or-rat 2699mg/kg	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
52. Toluene	35162	050823	0.05	5.00	40008.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-88-3	200 ppm	or-rat 5000mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	N/A	or-rat 1390mg/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-6	N/A	or-rat 5000mg/kg	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	135-98-8	N/A	or-rat 2240mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40000.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg	
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A		



Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1,1-Trichloroethane	11.10
4	Acetone	12.00
5	Isodimethane	12.21
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methyl methacrylate/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethane	28.73
23	Toluene	29.34
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloropropane	31.88
28	1,2-Dibromopropane	33.88
29	Chlorobenzene	35.53
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	35.53
31	m-Xylene/p-Xylene	35.53
32	o-Xylene	35.53
33	Styrene	35.53
34	Isopropylbenzene/Bromobenzene	35.53
35	cis-1,4-Dichloro-3-butene	36.40
36	1,1,2-Trichloropropane	37.27
37	n-Propylbenzene	37.27
38	trans-1,4-Dichloro-3-butene	38.05
39	Bromobenzene	38.14
40	1,3,5-Trimethylbenzene	38.50
41	2-Chlorobenzene	38.63
42	4-Chlorobenzene	38.77
43	tert-Butylbenzene	39.76
44	1,2,4-Trimethylbenzene	39.91
45	Perchloroethane	40.17
46	sec-Butylbenzene	40.52
47	p-Isopropylbenzene	41.02
48	1,3-Trichlorobenzene	41.42
49	1,4-Dichlorobenzene	41.83
50	n-Butylbenzene	42.52
51	1,2-Dichlorobenzene	42.18
52	1,2-Dichloropropane	46.17
53	1,2-Dichloro-3-methylpropane	46.40
54	Nitrobenzene	49.26
55	1,2,4-Trifluorobenzene	49.72
56	Hexachlorocyclopentadiene	50.56
57	Naphthalene	51.16
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water

			Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM

Dec 09/17/24

2 vlog



ANAB ISO 17034 Accredited
AR-1539 Certificate Number
<https://Absolutestandards.com>

CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 091424
Description: Acrolein

Solvent(s): Lot#
Water 072324Q


Expiration Date:	101424
Recommended Storage:	Refrigerate (4 °C)
Concentration (µg/mL):	5000
NIST Test ID#:	6UTB

Weight(s) shown below were combined and diluted to (mL):	10.0	0.001	Flask Uncertainty

		091424
Formulated By:	Justin Dippold	DATE
		091424
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05186	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.). Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM

Dec 09/17/24

2 vials

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT


Part Number: **91980**
Lot Number: **091424**
Description: **Acrolein**Solvent(s):
Water
Lot#
072324Q

Formulated By:	Justin Dippold
DATE: 091424	
Reviewed By:	Pedro L. Rentas
DATE: 091424	

Expiration Date: 101424
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2 = 200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.





Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97
	CAS#: 67-56-1	

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

See 1216124
20 vial
✓ 14630 to
✓ 14649



Weight(s) shown below were combined and diluted to (mL): **50.0**

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

<i>Prashant Chauhan</i>		120524
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		120524
Reviewed By:	Pedro L. Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	TWA 200 ppm
Skin notation		TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.		
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **95318**
Lot Number: **120524**
Description: **2-Chloroethyl vinyl ether**

Solvent(s):
Methanol

Lot#
EJ143-US

Expiration Date: **120527**
Recommended Storage: **Refrigerate (4 °C)**
Nominal Concentration (µg/mL): **10000**
NIST Test ID#: **6UTB**

Weight(s) shown below were combined and diluted to (mL): **50.0**



5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Prashant Chauhan 120524
Formulated By: Prashant Chauhan DATE
Pedro L. Rentas 120524
Reviewed By: Pedro L. Rentas DATE

✓ 14630 to
✓ 14649

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	ori-rat 250mg/kg

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp. = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024 Balance Serial # B707717271

Dylan Murphy
Dylan Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555581 **Lot No.:** A0210184

Description : Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 1127510105

REVISED
By Restek America, Inc. 11-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL
30 ml

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105


Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pabon at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105


Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial
Dec 03/31/25
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
gravimetric

V14904 to V14913



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.


This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0223904
Description : Custom 8260A/B Surrogate Mix
Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-33313	99%	25,108.0 µg/mL	+/- 1,421.9957
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	0000268853	99%	25,108.0 µg/mL	+/- 1,421.9957
3	Dibromofluoromethane	1868-53-7	VENKAT02	99%	25,232.0 µg/mL	+/- 1,429.0184
4	Toluene-d8	2037-26-5	PR-34141	99%	25,156.0 µg/mL	+/- 1,424.7141

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%


Brittany Federlino - Operations Tech I

Date Mixed: 27-Mar-2025

Balance: B251644995



Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rev 05/09/25
Avantor™



V 14921 to
V 14938

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM

Rec 08/14/25 5 Jia



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 081325
Description: Acrolein

Solvent(s): Water
Lot#: 041725Q

<i>Eli Allaga</i>		081325
Formulated By:	Eli Allaga	DATE
<i>Pedro Rentas</i>		081325
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 091325
Recommended Storage: Refrigerate (2°C to 8°C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

V15057 +0
V15061

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05176	5009.9	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
• Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
Rev 1.0, 2/25/2025



Certified Reference Material CRM

Rec 08/14/25 5 Jia



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 081325
Description: Acrolein

Solvent(s): Water
Lot#: 041725Q

<i>Eli Allaga</i>		081325
Formulated By:	Eli Allaga	DATE
<i>Pedro Rentas</i>		081325
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 091325
Recommended Storage: Refrigerate (2°C to 8°C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

V15057 + 0
V15061

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05176	5009.9	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
• Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
Rev 1.0, 2/25/2025



Certified Reference Material CRM

Rec 08/14/25 5 Jia



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 081325
Description: Acrolein

Solvent(s):
Water
Lot#: 041725Q

<i>Eli Allaga</i>		081325
Formulated By:	Eli Allaga	DATE
<i>Pedro Rentas</i>		081325
Reviewed By:	Pedro L. Rentas	DATE



Expiration Date: 091325
Recommended Storage: Refrigerate (2°C to 8°C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB

V15057 + 0
V15061

Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05176	5009.9	52.6	107-02-8	0.1 ppm	ori-rat 46mg/kg

Method: GC6MSD-1. Detector: Mass Selective Detector (Scan mode). Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas. NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately
Long term storage is not recommended. Please contact our technical department if further information is required.



• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
• Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).
Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Jacobs
ADDRESS: 412 Mt Kemble Ave Suite 100
CITY Morrisstown STATE: NJ ZIP: 07960
ATTENTION: John Ynfante John.Ynfante@jacobs.com
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: STC PTC
PROJECT NO.: D3868221 LOCATION: Princeton Junction
PROJECT MANAGER: Mary Murphy
e-mail: Mary.Murphy@Jacobs.com
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Mary Murphy PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard TAT DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. STC SPECIFIC VOL'S
2. 1.4 DIOL'S
3. TOTAL METALS
4. COBALT
5. ALUMINUM
6. TDS (5M250C)
7. AMMONIUM
8. TRACE METALS
9. STANDARD

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		A/E	E	B/E	E	E	E	E	A/E		← Specify Preservatives A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
								1	2	3	4	5	6	7	8	9		
1.	RMW-03B-90.4-08122025	GW		✓	8/12/25	1230	6	✓	✓					✓				
2.	RMW-02B-66.3-08122025	GW		✓	8/12/25	1530	5	✓	✓					✓				
3.	MW-17B-55.5-08122025	GW		✓	8/14/25	1154	27	✓	✓	✓	✓	✓	✓	✓			MS/MSD	
4.	MW-06-6.5-08122025	GW		✓	8/14/25	1530	2		✓									
5.	MW-17B-55.5-081225-SIM	GW		✓	8/12/25	1154	3								✓			
6.	TB01-081225	DI		✓	8/12/25	1600	2	✓										
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>8/12/25 1700</u>	RECEIVED BY: 1. <u>[Signature]</u>	DATE/TIME: <u>8-12-25 1700</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.0</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME:	RECEIVED BY: 2. <u>[Signature]</u>	DATE/TIME:	Comments: <u>See work order for list of site specific VOGs</u>
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>8-12-25</u>	RECEIVED BY: 3. <u>[Signature]</u>	DATE/TIME:	Comments: <u>Dissolved Iron is Picked Filtered</u>

From: Yazmeen Gomez
Sent: Thursday, August 14, 2025 10:30 AM
To: Ynfante, John; Sohil Jodhani
Cc: Mohammad Ahmed
Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.
Attachments: quant2841.pdf; quant2842.pdf

John,

The lab informed me –

Q2841- 42 sample screened with 8260 method as they are usually highly contaminated samples for Trichloroethene

Q2841-01 (MW-06-6.5-08122025) at 200x

Q2842-07(MW-17B-55.5-081225-SIM) 200x

Sample can be analyzed with 8260 under low dilution

Sim analysis is not possible as it may contaminate instrument

Please see attached.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Ynfante, John <John.Ynfante@jacobs.com>
Sent: Wednesday, August 13, 2025 3:23 PM
To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Cc: Yazmeen Gomez <Yazmeen.Gomez@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Thanks Sohil

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Sent: Wednesday, August 13, 2025 1:23 PM
To: Ynfante, John <John.Ynfante@jacobs.com>
Cc: Yazmeen Gomez <Yazmeen.Gomez@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>;
 Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,

Sample IDs are updated as per below request. Lab will notify you if in case VOC-SIM (for Vinyl chloride) is not required for the samples based on low level VOA analysis.

Thanks & Regards,



Sohil Jodhani (he/him/his)
 QA/QC Director
 Alliance Technical Group, LLC-Newark
 Main: 908-789-8900
 Direct: 908-728-3152
 Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com    

From: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Sent: Wednesday, August 13, 2025 2:10 PM
To: Ynfante, John <John.Ynfante@jacobs.com>; Data-EWR <Data-EWR@alliancetg.com>; Yazmeen Gomez <yazmeen.gomez@alliancetg.com>
Cc: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Subject: Re: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,

yaz is out of the office today, but i will coordinate with lab to make sure we communicate with you and run these analysis accordingly



Mohammad Ahmed
 Laboratory Director
 An Alliance Technical Group Company
 Main: 908-789-8900
 Direct: 908-728-3151
 Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Ynfante, John <John.Ynfante@jacobs.com>

Sent: Wednesday, August 13, 2025 2:01 PM

To: Data-EWR <Data-EWR@alliancetg.com>; Yazmeen Gomez <yazmeen.gomez@alliancetg.com>

Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Yazmeen,

A couple of comments on this Princeton login for Q2842:

1. I assume your SFAM-SIM method for VOCs can still only report vinyl chloride out of our project list of VOCs – if that is the case then note that sample Q2842-07 (MW-17B-55.5-081225-SIM) and any other SFAM-SIM samples only need to be analyzed if their corresponding 8260D-Low analysis shows the vinyl chloride to be non-detect and of course if the lab is able to analyze them by SIM. I know up to this point the lab has said the concentrations have been too high in the SIM samples to even run the SFAM-SIM on so let me know how it looks on those after your analysts run the regular VOCs run on them.
2. The date part of the sample IDs should have used a MMDDYY format, not MMDDYYYY as is listed on the chain. Please change the “2025” bit to just “25” on all these sample IDs. For example, “RMW-03B-90.4-08122025” should be changed to “RMW-03B-90.4-081225”, etc.

Thanks

- John Y.


From: Data-EWR@alliancetg.com <Data-EWR@alliancetg.com>

Sent: Wednesday, August 13, 2025 9:38 AM

To: Ongjoco, Alec <Alec.Ongjoco@jacobs.com>; Dillon, Alexa <Alexa.Dillon@jacobs.com>; Lader, Chelsea <Chelsea.Lader@jacobs.com>; Holmes, Daniel <Daniel.Holmes@jacobs.com>; Reamer, David <David.Reamer@jacobs.com>; Ynfante, John <John.Ynfante@jacobs.com>; Murphy, Mary <Mary.Murphy@jacobs.com>; Warren, Melissa <Melissa.Warren@jacobs.com>; Asher, Sarah <Sarah.Asher@jacobs.com>

Cc: yazmeen.gomez@alliancetg.com

Subject: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.



To John Ynfante;

Please see the attached Login Summary for the following project, or download the file using your login credentials from the link below.

Order ID : Q2842
Project ID : Former Schlumberger STC PTC Site D3868221
Download File : <https://chemtech.net/secureLogin.aspx>
Order Date : 8/13/2025 8:34:00 AM

Alliance's Project Manager : YAZMEEN GOMEZ , yazmeen.gomez@alliancetg.com , 908-728-3147
Alliance's Sales Executive : Jordan Hedvat , jordan.hedvat@alliancetg.com , 908-728-3144

Thank you for the opportunity to provide you with our services. For any questions please feel free to contact your project manager.

Click Here for our short online customer Survey <http://chemtech.net/ClientSurvey.aspx>.

Thank you,

Alliance Technical Group LLC.

Notice: The information transmitted in this e-mail message and in any attachments is intended Solely for the attention of the named addressee(s) and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon, this information by persons or entities other than the intended recipient is strictly prohibited and may be unlawful. If you have received this transmission in error, please notify us immediately by return e-mail, and permanently delete this transmission, including attachments if any, from any computer.

■

NOTICE - This communication may contain confidential and privileged information that is for the sole use of the intended recipient. Any viewing, copying or distribution of, or reliance on this message by unintended recipients is strictly prohibited. If you have received this message in error, please notify us immediately by replying to the message and deleting it from your computer.

NOTICE - This communication may contain confidential and privileged information that is for the sole use of the intended recipient. Any viewing, copying or distribution of, or reliance on this message by unintended recipients is strictly prohibited. If you have received this message in error, please notify us immediately by replying to the message and deleting it from your computer.

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2842	JACO05	Order Date : 8/13/2025 8:34:00 AM	Project Mgr : Deepak
Client Name : JACOBS Engineering Grou		Project Name : Former Schlumberger STC	Report Type : Level 34
Client Contact : John Ynfante		Receive DateTime : 8/12/2025 6:22:00 PM	EDD Type : CH2MHILL
Invoice Name : JACOBS Engineering Grou		Purchase Order :	Hard Copy Date :
Invoice Contact : John Ynfante			Date Signoff : 8/13/2025 10:38:03 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2842-01	RMW-03B-90.4-081225	Water	08/12/2025	12:30					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-02	RMW-02B-66.3-081225	Water	08/12/2025	15:30					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-03	MW-17B-55.5-081225	Water	08/12/2025	11:54					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-04	Q2842-03MS	Water	08/12/2025	11:54					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-05	Q2842-03MSD	Water	08/12/2025	11:54					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-08	TB01-081225	Water	08/12/2025	16:00					
					VOCMS Group3		8260-Low	10 Bus. Days	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2842	JACO05	Order Date : 8/13/2025 8:34:00 AM	Project Mgr : Deepak
Client Name : JACOBS Engineering Grou		Project Name : Former Schlumberger STC	Report Type : Level ~~3~~ 4
Client Contact : John Ynfante		Receive DateTime : 8/12/2025 6:22:00 PM	EDD Type : CH2MHILL
Invoice Name : JACOBS Engineering Grou		Purchase Order :	Hard Copy Date :
Invoice Contact : John Ynfante			Date Signoff : 8/13/2025 10:38:03 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinquished By :

Date / Time :

df
8/13/25 09:15

Received By :

Date / Time :

Jim
08/13/25 09:15

Storage Area : VOA Refridgerator Room

Rg Ay