

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2842

Project: Former Schlumberger STC PTC Site D3868221

RunNo.: LB136803

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.3	10	103	90-110	08/06/2025
Chloride	mg/L	3.2	3	107	90-110	08/06/2025
Fluoride	mg/L	2	2	100	90-110	08/06/2025
Nitrite	mg/L	3.1	3	103	90-110	08/06/2025
Nitrate	mg/L	2.6	2.5	104	90-110	08/06/2025
Sulfate	mg/L	15.4	15	103	90-110	08/06/2025
Orthophosphate as P	mg/L	5	5	100	90-110	08/06/2025
Sample ID: CCV1						
Bromide	mg/L	10	10	100	90-110	08/13/2025
Chloride	mg/L	3	3	100	90-110	08/13/2025
Fluoride	mg/L	2	2	100	90-110	08/13/2025
Nitrite	mg/L	3	3	100	90-110	08/13/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/13/2025
Sulfate	mg/L	14.9	15	99	90-110	08/13/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	08/13/2025
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	08/13/2025
Chloride	mg/L	3	3	100	90-110	08/13/2025
Fluoride	mg/L	2	2	100	90-110	08/13/2025
Nitrite	mg/L	3	3	100	90-110	08/13/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/13/2025
Sulfate	mg/L	14.7	15	98	90-110	08/13/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	08/13/2025