

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q2842
Project: Former Schlumberger STC PTC Site D3868221	Sample ID: Q2842-03
Client ID: MW-17B-55.5-081225DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	197		196		1	0.51		08/14/2025
Alkalinity	mg/L	+/-20	88.2		87.6		1	1		08/15/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q2842
Project: Former Schlumberger STC PTC Site D3868221	Sample ID: Q2842-03
Client ID: MW-17B-55.5-081225MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Nitrate	mg/L	+/-15	2.60		2.60		1	0		08/13/2025
Chloride	mg/L	+/-15	23.5	OR	23.3	OR	1	1		08/13/2025
Sulfate	mg/L	+/-15	55.1	OR	55.4	OR	1	1		08/13/2025
Bromide	mg/L	+/-15	10.3		10.5		1	2		08/13/2025
Nitrite	mg/L	+/-15	3.00		3.10		1	3		08/13/2025
Fluoride	mg/L	+/-15	2.30		2.40		1	4		08/13/2025
Orthophosphate as P	mg/L	+/-15	3.10		3.30		1	6		08/13/2025