

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	RMW-03B-90.4-08122025DL	SDG No.:	Q2842
Lab Sample ID:	Q2842-01DL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	910 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037620.D	2	08/18/25 08:41	08/20/25 05:27	PB169286

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
123-91-1	1,4-Dioxane	6.60	D	0.15	0.44	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.25		30 (20) - 150 (139)	62%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 (54) - 150 (157)	87%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		30 (27) - 130 (154)	74%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		30 (30) - 130 (155)	85%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.48		30 (54) - 130 (175)	120%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2330	7.71			
1146-65-2	Naphthalene-d8	5460	10.498			
15067-26-2	Acenaphthene-d10	2690	14.345			
1517-22-2	Phenanthrene-d10	5770	17.087			
1719-03-5	Chrysene-d12	4770	21.268			
1520-96-3	Perylene-d12	4470	23.496			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products