

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081225

Review By	Rahul	Review On	8/13/2025 11:25:23 AM
Supervise By	Jagrut	Supervise On	8/13/2025 11:26:50 AM
SubDirectory	BN081225	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081225

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037576.D	12 Aug 2025 15:05		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037577.D	12 Aug 2025 15:49	A Fresh Calibration is required	RC/JU	Not Ok
3	SSTDICC0.1	SSTDICC0.1	BN037578.D	12 Aug 2025 16:26		RC/JU	Ok
4	SSTDICC0.2	SSTDICC0.2	BN037579.D	12 Aug 2025 17:03		RC/JU	Ok
5	SSTDICCC0.4	SSTDICCC0.4	BN037580.D	12 Aug 2025 17:39	Compound #20 kept on LR	RC/JU	Ok
6	SSTDICC0.8	SSTDICC0.8	BN037581.D	12 Aug 2025 18:16		RC/JU	Ok
7	SSTDICC1.6	SSTDICC1.6	BN037582.D	12 Aug 2025 18:52		RC/JU	Ok
8	SSTDICC3.2	SSTDICC3.2	BN037583.D	12 Aug 2025 19:29		RC/JU	Ok
9	SSTDICC5.0	SSTDICC5.0	BN037584.D	12 Aug 2025 20:05	Comp #20,23,24 removed from 5ppm	RC/JU	Ok
10	SSTDICV0.4	ICVBN081225	BN037585.D	12 Aug 2025 20:42		RC/JU	Ok
11	PB169094BL	PB169094BL	BN037586.D	12 Aug 2025 21:55		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081925

Review By	Rahul	Review On	8/21/2025 3:02:23 PM
Supervise By	Jagrut	Supervise On	8/21/2025 5:56:51 PM
SubDirectory	BN081925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081225

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6846 SP6830, 1ul/100ul sample SP6854

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037599.D	19 Aug 2025 10:00		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037600.D	19 Aug 2025 10:39		RC/JU	Ok
3	PB169286BL	PB169286BL	BN037601.D	19 Aug 2025 11:16	Not Used	RC/JU	Not Ok
4	Q2842-01	RMW-03B-90.4-081225	BN037602.D	19 Aug 2025 11:52	Need 2X.	RC/JU	Dilution
5	Q2842-02	RMW-02B-66.3-081225	BN037603.D	19 Aug 2025 12:28	Need 10X.	RC/JU	Dilution
6	Q2842-03	MW-17B-55.5-081225	BN037604.D	19 Aug 2025 13:04		RC/JU	Ok
7	Q2842-04MS	MW-17B-55.5-081225	BN037605.D	19 Aug 2025 13:40	Recovery fail very high , need noncon.	RC/JU	Ok,M
8	Q2842-05MSD	MW-17B-55.5-081225	BN037606.D	19 Aug 2025 14:17	Recovery fail very high , need noncon.	RC/JU	Ok,M
9	Q2842-06	MW-06-6.5-081225	BN037607.D	19 Aug 2025 14:53		RC/JU	Ok
10	Q2869-01	MW-19B-71.5-081325	BN037608.D	19 Aug 2025 15:29	Need 2X.	RC/JU	Dilution
11	Q2869-03	MW-01-6.5-081325	BN037609.D	19 Aug 2025 16:05		RC/JU	Ok
12	Q2869-04	MW-11A-13.5-081325	BN037610.D	19 Aug 2025 16:42		RC/JU	Ok
13	Q2869-05	MW-19B-71.5-081325	BN037611.D	19 Aug 2025 17:18		RC/JU	Ok
14	Q2869-07	EB01-081325	BN037612.D	19 Aug 2025 17:54		RC/JU	Ok
15	Q2869-10	EB02-081325	BN037613.D	19 Aug 2025 18:30		RC/JU	Ok
16	Q2878-01	MW-18B-57.5-081325	BN037614.D	19 Aug 2025 19:07		RC/JU	Ok
17	Q2878-05	MW-11A-37.5-081425	BN037615.D	19 Aug 2025 19:43	Need 20X.	RC/JU	Dilution

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081925

Review By	Rahul	Review On	8/21/2025 3:02:23 PM
Supervise By	Jagrut	Supervise On	8/21/2025 5:56:51 PM
SubDirectory	BN081925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081225
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	SP6757 SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6846 SP6830,1ul/100ul sample SP6854		

18	SSTDCCC0.4	SSTDCCC0.4EC	BN037616.D	19 Aug 2025 20:19		RC/JU	Ok
19	DFTPP	DFTPP	BN037617.D	20 Aug 2025 03:35		RC/JU	Ok
20	SSTDCCC0.4	SSTDCCC0.4	BN037618.D	20 Aug 2025 04:14		RC/JU	Ok
21	PB169286BL	PB169286BL	BN037619.D	20 Aug 2025 04:50		RC/JU	Ok
22	Q2842-01DL	RMW-03B-90.4-081225	BN037620.D	20 Aug 2025 05:27		RC/JU	Ok
23	Q2842-02DL	RMW-02B-66.3-081225	BN037621.D	20 Aug 2025 06:03		RC/JU	Ok
24	Q2869-01DL	MW-19B-71.5-081325D	BN037622.D	20 Aug 2025 06:38		RC/JU	Ok
25	Q2881-01	RW8-SP100-20250814	BN037623.D	20 Aug 2025 07:13		RC/JU	Ok
26	Q2881-02	RW8-SP303-20250814	BN037624.D	20 Aug 2025 07:49		RC/JU	Ok
27	Q2882-01	RW7-SP100-20250814	BN037625.D	20 Aug 2025 08:25		RC/JU	Ok
28	Q2882-02	RW7-SP201-20250814	BN037626.D	20 Aug 2025 09:00		RC/JU	Ok
29	Q2882-03	RW7-SP302-20250814	BN037627.D	20 Aug 2025 09:37		RC/JU	Ok
30	Q2882-04	RW7-SP303-20250814	BN037628.D	20 Aug 2025 10:13		RC/JU	Ok
31	Q2878-05DL	MW-11A-37.5-081425D	BN037629.D	20 Aug 2025 10:49		RC/JU	Ok
32	PB169286BS	PB169286BS	BN037630.D	20 Aug 2025 11:25		RC/JU	Ok
33	SSTDCCC0.4	SSTDCCC0.4EC	BN037631.D	20 Aug 2025 14:56		RC/JU	Ok

M : Manual Integration