

Surrogate Summary

SW-846

SDG No.: Q2842

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169286BL	PB169286BL	2-Methylnaphthalene-d10	0.4	0.32	81		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.33	83		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.36	90		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	86		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.37	93		30 (54)	130 (175)
PB169286BS	PB169286BS	2-Methylnaphthalene-d10	0.4	0.33	82		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.31	76		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.35	87		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.36	90		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.35	88		30 (54)	130 (175)
Q2842-01	RMW-03B-90.4-08122025	2-Methylnaphthalene-d10	0.4	0.26	66		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.38	94		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.30	74		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	87		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.49	123		30 (54)	130 (175)
Q2842-01DL	RMW-03B-90.4-08122025DL	2-Methylnaphthalene-d10	0.4	0.25	62		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	87		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.29	74		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.34	85		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.48	120		30 (54)	130 (175)
Q2842-02	RMW-02B-66.3-08122025	2-Methylnaphthalene-d10	0.4	0.30	76		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.39	97		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.33	83		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.33	83		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.49	122		30 (54)	130 (175)
Q2842-02DL	RMW-02B-66.3-08122025DL	2-Methylnaphthalene-d10	0.4	0.32	80		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.45	113		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.36	90		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.36	90		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.55	138	*	30 (54)	130 (175)
Q2842-03	MW-17B-55.5-08122025	2-Methylnaphthalene-d10	0.4	0.29	72		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.39	97		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.31	77		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.31	76		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.53	132	*	30 (54)	130 (175)
Q2842-04MS	MW-17B-55.5-08122025MS	2-Methylnaphthalene-d10	0.4	0.31	77		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.38	94		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.33	83		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.32	81		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.47	117		30 (54)	130 (175)
Q2842-05MSD	MW-17B-55.5-08122025MSD	2-Methylnaphthalene-d10	0.4	0.30	76		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	93		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.32	79		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.32	81		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.47	117		30 (54)	130 (175)
Q2842-06	MW-06-6.5-08122025	2-Methylnaphthalene-d10	0.4	0.26	66		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.36	89		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.30	75		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	87		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.43	108		30 (54)	130 (175)