

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2842

Client ID : SSTDCCC0.4

Date Analyzed: 08/20/2025

Lab File ID: BN037618.D

Time Analyzed: 04:14

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2622	7.71	6166	10.50	3094	14.35
UPPER LIMIT	5244	8.21	12332	10.998	6188	14.845
LOWER LIMIT	1311	7.21	3083	9.998	1547	13.845
EPA SAMPLE NO.						
01 PB169286BL	2323	7.71	5229	10.50	2344	14.35
02 PB169286BS	2389	7.71	5616	10.49	2629	14.35
03 RMW-03B-90.4-08122025DL	2332	7.71	5464	10.50	2692	14.35
04 RMW-02B-66.3-08122025DL	1938	7.71	4467	10.50	2275	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.