

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2842

Client ID: SSTDCCC0.4

Date Analyzed: 08/20/2025

Lab File ID: BN037618.D

Time Analyzed: 04:14

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6113	17.086	5494	21.268	5322	23.499
UPPER LIMIT	12226	17.586	10988	21.768	10644	23.999
LOWER LIMIT	3056.5	16.586	2747	20.768	2661	22.999
EPA SAMPLE NO.						
01 PB169286BL	4466	17.09	3733	21.27	3552	23.50
02 PB169286BS	5040	17.09	4175	21.27	3780	23.50
03 RMW-03B-90.4-08122025DL	5770	17.09	4774	21.27	4470	23.50
04 RMW-02B-66.3-08122025DL	4809	17.09	3968	21.27	3640	23.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.