

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jul 17 02:21:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	193853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	344966	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	317026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	165077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	501760	152.545	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 305.080%#			
35) Dibromofluoromethane	8.153	113	368306	154.778	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 309.560%#			
50) Toluene-d8	10.547	98	1362098	160.469	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 320.940%#			
62) 4-Bromofluorobenzene	12.829	95	522221	166.525	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 333.040%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	363327	176.463	ug/l	95
3) Chloromethane	2.383	50	405800	156.728	ug/l	99
4) Vinyl Chloride	2.542	62	418507	162.646	ug/l	94
5) Bromomethane	2.942	94	215335	161.604	ug/l	95
6) Chloroethane	3.118	64	249466	148.663	ug/l	100
7) Trichlorofluoromethane	3.501	101	585457	153.872	ug/l	97
8) Diethyl Ether	3.959	74	228544	154.847	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.353	101	296076	151.587	ug/l	99
10) Methyl Iodide	4.571	142	328079	150.516	ug/l	99
11) Tert butyl alcohol	5.536	59	478279	765.784	ug/l	100
12) 1,1-Dichloroethene	4.330	96	312325	141.111	ug/l	97
13) Acrolein	4.177	56	415488	828.945	ug/l	100
14) Allyl chloride	5.006	41	663403	165.621	ug/l	91
15) Acrylonitrile	5.706	53	1279363	754.869	ug/l	100
16) Acetone	4.424	43	1064061	682.373	ug/l	96
17) Carbon Disulfide	4.695	76	1011315	154.118	ug/l	95
18) Methyl Acetate	5.012	43	577536	149.052	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	1286728	157.724	ug/l	95
20) Methylene Chloride	5.259	84	390988	151.266	ug/l	94
21) trans-1,2-Dichloroethene	5.765	96	356549	142.870	ug/l	96
22) Diisopropyl ether	6.659	45	1280267	152.375	ug/l	99
23) Vinyl Acetate	6.589	43	5943020	808.750	ug/l	98
24) 1,1-Dichloroethane	6.553	63	704727	145.384	ug/l	98
25) 2-Butanone	7.477	43	1797146	754.179	ug/l	98
26) 2,2-Dichloropropane	7.477	77	547190	145.193	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	437021	152.102	ug/l	99
28) Bromochloromethane	7.800	49	339529	146.355	ug/l	100
29) Tetrahydrofuran	7.830	42	1166702	753.679	ug/l	98
30) Chloroform	7.953	83	717424	147.865	ug/l	98
31) Cyclohexane	8.241	56	582988	144.170	ug/l	94
32) 1,1,1-Trichloroethane	8.153	97	630986	150.152	ug/l	98
36) 1,1-Dichloropropene	8.353	75	503793	160.248	ug/l	99
37) Ethyl Acetate	7.553	43	692256	152.464	ug/l	98
38) Carbon Tetrachloride	8.347	117	535575	154.647	ug/l	98
39) Methylcyclohexane	9.583	83	559851	164.485	ug/l	98
40) Benzene	8.588	78	1550984	152.642	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jul 17 02:21:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	375482	158.156	ug/l	99
42) 1,2-Dichloroethane	8.653	62	571634	148.352	ug/l	100
43) Isopropyl Acetate	8.677	43	1091736	154.893	ug/l	99
44) Trichloroethene	9.335	130	364760	151.927	ug/l	96
45) 1,2-Dichloropropane	9.606	63	391676	151.708	ug/l	97
46) Dibromomethane	9.694	93	287252	148.602	ug/l	98
47) Bromodichloromethane	9.871	83	584546	150.128	ug/l	99
48) Methyl methacrylate	9.665	41	520990	164.191	ug/l	99
49) 1,4-Dioxane	9.688	88	157400	3238.745	ug/l #	93
51) 4-Methyl-2-Pentanone	10.430	43	3375607	757.219	ug/l	99
52) Toluene	10.612	92	961314	155.654	ug/l	99
53) t-1,3-Dichloropropene	10.818	75	640188	162.462	ug/l	94
54) cis-1,3-Dichloropropene	10.294	75	654301	160.747	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	365858	146.322	ug/l	99
56) Ethyl methacrylate	10.859	69	679202	150.296	ug/l	96
57) 1,3-Dichloropropane	11.147	76	662154	153.168	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	1717427	837.324	ug/l	100
59) 2-Hexanone	11.177	43	2477578	837.689	ug/l	98
60) Dibromochloromethane	11.341	129	448307	157.217	ug/l	99
61) 1,2-Dibromoethane	11.453	107	402978	153.282	ug/l	97
64) Tetrachloroethene	11.082	164	301404	147.718	ug/l	97
65) Chlorobenzene	11.871	112	1066757	149.878	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	375162	155.013	ug/l	98
67) Ethyl Benzene	11.941	91	1881920	160.613	ug/l	99
68) m/p-Xylenes	12.053	106	1437290	327.579	ug/l	98
69) o-Xylene	12.376	106	698363	166.628	ug/l	98
70) Styrene	12.394	104	1195068	169.502	ug/l	99
71) Bromoform	12.559	173	320960	164.154	ug/l #	99
73) Isopropylbenzene	12.676	105	1747567	168.203	ug/l	100
74) N-amyl acetate	12.482	43	750104	184.917	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.918	83	581568	148.761	ug/l	100
76) 1,2,3-Trichloropropane	12.971	75	504791m	137.832	ug/l	
77) Bromobenzene	12.959	156	433814	161.001	ug/l	99
78) n-propylbenzene	13.018	91	2127215	162.733	ug/l	100
79) 2-Chlorotoluene	13.106	91	1290205	160.600	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	1465159	165.514	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	224191	165.713	ug/l	95
82) 4-Chlorotoluene	13.200	91	1336462	159.785	ug/l	100
83) tert-Butylbenzene	13.418	119	1230683	166.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	1502651	166.222	ug/l	100
85) sec-Butylbenzene	13.594	105	1767949	158.754	ug/l	99
86) p-Isopropyltoluene	13.706	119	1501303	168.219	ug/l	98
87) 1,3-Dichlorobenzene	13.712	146	828885	156.742	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	830255	147.000	ug/l	99
89) n-Butylbenzene	14.035	91	1367813	160.505	ug/l	98
90) Hexachloroethane	14.312	117	292183	154.518	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	784106	156.513	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.694	75	151090	147.203	ug/l	93
93) 1,2,4-Trichlorobenzene	15.370	180	492237	167.266	ug/l	99
94) Hexachlorobutadiene	15.476	225	163922	149.909	ug/l	99
95) Naphthalene	15.612	128	1880423	180.370	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	491195	166.396	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087333.D
Acq On : 16 Jul 2025 18:54
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087333.D
Acq On : 16 Jul 2025 18:54
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_N

Client Sample Id :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025
Supervised By : Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

