

Cover Page

Order ID : Q2842

Project ID : Former Schlumberger STC PTC Site D3868221

Client : JACOBS Engineering Group, Inc.

Lab Sample Number

Q2842-01
Q2842-02
Q2842-03
Q2842-04
Q2842-05
Q2842-06
Q2842-08
Q2842-09
Q2842-10
Q2842-11

Client Sample Number

RMW-03B-90.4-08122025
RMW-02B-66.3-08122025
MW-17B-55.5-08122025
Q2842-03MS
Q2842-03MSD
MW-06-6.5-08122025
TB01-081225
MW-17B-55.5-08122025
Q2842-09MS
Q2842-09MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 8/20/2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2842

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 08/20/2025

LAB CHRONICLE

OrderID:	Q2842	OrderDate:	8/13/2025 8:34:00 AM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger STC PTC Site D3868221
Contact:	John Ynfante	Location:	J23,J32,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2842-01	RMW-03B-90.4-08122 025	Water			08/12/25			08/12/25
			VOCMS Group3	8260-Low			08/14/25	
Q2842-02	RMW-02B-66.3-08122 025	Water			08/12/25			08/12/25
			VOCMS Group3	8260-Low			08/14/25	
Q2842-03	MW-17B-55.5-081220 25	Water			08/12/25			08/12/25
			VOCMS Group3	8260-Low			08/14/25	
Q2842-08	TB01-081225	Water			08/12/25			08/12/25
			VOCMS Group3	8260-Low			08/14/25	

Hit Summary Sheet
SW-846

SDG No.: Q2842

Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	RMW-03B-90.4-08122025							
Q2842-01	RMW-03B-90.4-08	Water	cis-1,2-Dichloroethene	2800		7.60	40.0	ug/L
Q2842-01	RMW-03B-90.4-08	Water	Trichloroethene	810		3.70	40.0	ug/L
			Total Voc :	3610				
			Total Concentration:	3610				
Client ID:	RMW-02B-66.3-08122025							
Q2842-02	RMW-02B-66.3-08	Water	Vinyl Chloride	240		10.4	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	1,1-Dichloroethene	130		9.20	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	cis-1,2-Dichloroethene	1600		7.60	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	Trichloroethene	1500		3.70	40.0	ug/L
Q2842-02	RMW-02B-66.3-08	Water	Tetrachloroethene	24.7	J	9.20	40.0	ug/L
			Total Voc :	3490				
			Total Concentration:	3490				
Client ID:	MW-17B-55.5-08122025							
Q2842-03	MW-17B-55.5-0812	Water	cis-1,2-Dichloroethene	3000		19.0	100	ug/L
Q2842-03	MW-17B-55.5-0812	Water	Trichloroethene	3600		9.30	100	ug/L
Q2842-03	MW-17B-55.5-0812	Water	Tetrachloroethene	56.3	J	23.0	100	ug/L
			Total Voc :	6660				
			Total Concentration:	6660				



QC SUMMARY

Surrogate Summary

SDG No.: **Q2842**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2842-01	RMW-03B-90.4-08122025	1,2-Dichloroethane-d4	50	53.7	107		70 (74)	130 (125)
		Dibromofluoromethane	50	45.5	91		70 (75)	130 (124)
		Toluene-d8	50	45.7	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.9	88		70 (77)	130 (121)
Q2842-02	RMW-02B-66.3-08122025	1,2-Dichloroethane-d4	50	57.0	114		70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90		70 (75)	130 (124)
		Toluene-d8	50	46.0	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.0	90		70 (77)	130 (121)
Q2842-03	MW-17B-55.5-08122025	1,2-Dichloroethane-d4	50	56.6	113		70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90		70 (75)	130 (124)
		Toluene-d8	50	46.5	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.5	91		70 (77)	130 (121)
Q2842-04MS	MW-17B-55.5-081225MS	1,2-Dichloroethane-d4	50	53.8	108		70 (74)	130 (125)
		Dibromofluoromethane	50	47.9	96		70 (75)	130 (124)
		Toluene-d8	50	47.9	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.7	103		70 (77)	130 (121)
Q2842-05MSD	MW-17B-55.5-081225MSD	1,2-Dichloroethane-d4	50	54.6	109		70 (74)	130 (125)
		Dibromofluoromethane	50	46.7	93		70 (75)	130 (124)
		Toluene-d8	50	48.2	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101		70 (77)	130 (121)
Q2842-08	TB01-081225	1,2-Dichloroethane-d4	50	55.1	110		70 (74)	130 (125)
		Dibromofluoromethane	50	44.9	90		70 (75)	130 (124)
		Toluene-d8	50	45.6	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.2	88		70 (77)	130 (121)
VN0814WBL01	VN0814WBL01	1,2-Dichloroethane-d4	50	52.9	106		70 (74)	130 (125)
		Dibromofluoromethane	50	44.2	88		70 (75)	130 (124)
		Toluene-d8	50	44.2	88		70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.7	87		70 (77)	130 (121)
VN0814WBS01	VN0814WBS01	1,2-Dichloroethane-d4	50	57.6	115		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	50.7	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.9	106		70 (77)	130 (121)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VN087569.D

Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Limits		
		Result			Qual	RPD		Qual	Low	High	RPD
Lab Sample ID : Q2842-04MS		Client Sample ID :		MW-17B-55.5-081225MS							
Vinyl chloride	5000	0	5400	ug/L	108				70 (69)	130 (125)	
1,1-Dichloroethene	5000	0	5200	ug/L	104				70 (53)	130 (162)	
1,1-Dichloroethane	5000	0	5300	ug/L	106				70 (78)	130 (122)	
cis-1,2-Dichloroethene	5000	3000	8100	ug/L	102				70 (82)	130 (124)	
1,1,1-Trichloroethane	5000	0	5400	ug/L	108				70 (83)	130 (117)	
Benzene	5000	0	5100	ug/L	102				70 (81)	130 (128)	
1,2-Dichloroethane	5000	0	5500	ug/L	110				70 (76)	130 (120)	
Trichloroethene	5000	3600	10000	ug/L	128				70 (28)	130 (175)	
1,1,2-Trichloroethane	5000	0	5300	ug/L	106				70 (27)	130 (175)	
Tetrachloroethene	5000	56.3	4300	ug/L	85				70 (48)	130 (153)	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VN087570.D

Limits										
Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD	RPD Qual	Low	High	RPD
Lab Sample ID :	Q2842-05MSD	Client Sample ID :	MW-17B-55.5-081225MSD							
Vinyl chloride	5000	0	5400	ug/L	108		0	70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	5000	0	5300	ug/L	106		2	70 (53)	130 (162)	20 (20)
1,1-Dichloroethane	5000	0	5400	ug/L	108		2	70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	5000	3000	8000	ug/L	100		2	70 (82)	130 (124)	20 (20)
1,1,1-Trichloroethane	5000	0	5500	ug/L	110		2	70 (83)	130 (117)	20 (20)
Benzene	5000	0	5100	ug/L	102		0	70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	5000	0	5300	ug/L	106		4	70 (76)	130 (120)	20 (20)
Trichloroethene	5000	3600	9500	ug/L	118		8	70 (28)	130 (175)	20 (20)
1,1,2-Trichloroethane	5000	0	5300	ug/L	106		0	70 (27)	130 (175)	20 (20)
Tetrachloroethene	5000	56.3	4300	ug/L	85		0	70 (48)	130 (153)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2842</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>JACOBS Engineering Group, Inc.</u>	Datafile :	<u>VN087555.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0814WBS01	Vinyl chloride	20	21.4	ug/L	107			70 (65)	130 (117)	
	1,1-Dichloroethene	20	20.6	ug/L	103			70 (74)	130 (110)	
	1,1-Dichloroethane	20	21.6	ug/L	108			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	21.2	ug/L	106			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	21.4	ug/L	107			70 (80)	130 (108)	
	Benzene	20	20.3	ug/L	102			70 (82)	130 (109)	
	1,2-Dichloroethane	20	21.5	ug/L	108			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			70 (83)	130 (112)	
	Tetrachloroethene	20	18.1	ug/L	91			70 (67)	130 (123)	

VOLATILE METHOD BLANK SUMMARY

Client ID

VN0814WBL01

Lab Name: Alliance

Contract: JAC005

Lab Code: ACE

SDG NO.: Q2842

Lab File ID: VN087554.D

Lab Sample ID: VN0814WBL01

Date Analyzed: 08/14/2025

Time Analyzed: 10:32

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0814WBS01	VN0814WBS01	VN087555.D	08/14/2025
TB01-081225	Q2842-08	VN087565.D	08/14/2025
RMW-03B-90.4-08122025	Q2842-01	VN087566.D	08/14/2025
RMW-02B-66.3-08122025	Q2842-02	VN087567.D	08/14/2025
MW-17B-55.5-08122025	Q2842-03	VN087568.D	08/14/2025
MW-17B-55.5-081225MS	Q2842-04MS	VN087569.D	08/14/2025
MW-17B-55.5-081225MSD	Q2842-05MSD	VN087570.D	08/14/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q2842

Lab File ID: VN087327.D

BFB Injection Date: 07/16/2025

Instrument ID: MSVOA_N

BFB Injection Time: 16:10

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.8
75	30.0 - 60.0% of mass 95	50.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.1
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	70.9
175	5.0 - 9.0% of mass 174	3.6 (5.1) 1
176	95.0 - 101.0% of mass 174	68.7 (96.9) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN087328.D	07/16/2025	17:05
VSTDIC005	VSTDIC005	VN087329.D	07/16/2025	17:27
VSTDIC020	VSTDIC020	VN087330.D	07/16/2025	17:49
VSTDIC050	VSTDIC050	VN087331.D	07/16/2025	18:11
VSTDIC100	VSTDIC100	VN087332.D	07/16/2025	18:32
VSTDIC150	VSTDIC150	VN087333.D	07/16/2025	18:54



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q2842

Lab File ID: VN087551.D

BFB Injection Date: 08/14/2025

Instrument ID: MSVOA_N

BFB Injection Time: 09:01

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	5.6
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	65.2
175	5.0 - 9.0% of mass 174	4.4 (6.8) 1
176	95.0 - 101.0% of mass 174	64.4 (98.8) 1
177	5.0 - 9.0% of mass 176	4 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087552.D	08/14/2025	09:34
VN0814WBL01	VN0814WBL01	VN087554.D	08/14/2025	10:32
VN0814WBS01	VN0814WBS01	VN087555.D	08/14/2025	11:07
TB01-081225	Q2842-08	VN087565.D	08/14/2025	14:58
RMW-03B-90.4-08122025	Q2842-01	VN087566.D	08/14/2025	15:20
RMW-02B-66.3-08122025	Q2842-02	VN087567.D	08/14/2025	15:42
MW-17B-55.5-08122025	Q2842-03	VN087568.D	08/14/2025	16:04
MW-17B-55.5-081225MS	Q2842-04MS	VN087569.D	08/14/2025	16:27
MW-17B-55.5-081225MSD	Q2842-05MSD	VN087570.D	08/14/2025	16:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q2842
 Lab File ID: VN087552.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	271352	8.21	543098	9.08	500619	11.85
UPPER LIMIT	542704	8.712	1086200	9.582	1001240	12.347
LOWER LIMIT	135676	7.712	271549	8.582	250310	11.347
EPA SAMPLE NO.						
RMW-03B-90.4-08122025	225616	8.21	494400	9.09	459245	11.85
RMW-02B-66.3-08122025	218605	8.21	486591	9.09	438933	11.85
MW-17B-55.5-08122025	219937	8.21	487780	9.09	455235	11.85
MW-17B-55.5-081225MS	238067	8.21	461160	9.08	435202	11.85
MW-17B-55.5-081225MSD	251785	8.21	498703	9.08	460182	11.85
TB01-081225	224412	8.21	498664	9.09	458661	11.85
VN0814WBL01	284398	8.21	560321	9.08	503771	11.85
VN0814WBS01	250453	8.21	485218	9.08	446518	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q2842
 Lab File ID: VN087552.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:34
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	255173	13.77				
UPPER LIMIT	510346	14.27				
LOWER LIMIT	127587	13.27				
EPA SAMPLE NO.						
RMW-03B-90.4-08122025	213548	13.77				
RMW-02B-66.3-08122025	211146	13.77				
MW-17B-55.5-08122025	209426	13.77				
MW-17B-55.5-081225MS	231778	13.77				
MW-17B-55.5-081225MSD	240017	13.77				
TB01-081225	214662	13.77				
VN0814WBL01	234765	13.77				
VN0814WBS01	221983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	RMW-03B-90.4-08122025	SDG No.:	Q2842
Lab Sample ID:	Q2842-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087566.D	40	08/14/25 15:20	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	10.4	U	10.4	40.0	ug/L
75-35-4	1,1-Dichloroethene	9.20	U	9.20	40.0	ug/L
75-34-3	1,1-Dichloroethane	9.20	U	9.20	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	2800		7.60	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	8.00	U	8.00	40.0	ug/L
71-43-2	Benzene	6.00	U	6.00	40.0	ug/L
107-06-2	1,2-Dichloroethane	8.80	U	8.80	40.0	ug/L
79-01-6	Trichloroethene	810		3.70	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	8.40	U	8.40	40.0	ug/L
127-18-4	Tetrachloroethene	9.20	U	9.20	40.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.7		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	45.5		70 (75) - 130 (124)	91%	SPK: 50
2037-26-5	Toluene-d8	45.7		70 (86) - 130 (113)	91%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.9		70 (77) - 130 (121)	88%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	226000	8.206			
540-36-3	1,4-Difluorobenzene	494000	9.088			
3114-55-4	Chlorobenzene-d5	459000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	214000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087566.D
 Acq On : 14 Aug 2025 15:20
 Operator : JC\MD
 Sample : Q2842-01 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-03B-90.4-08122025

Quant Time: Aug 14 15:39:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.206	168	225616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	494400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	459245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	213548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	205410	53.657	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.320%			
35) Dibromofluoromethane	8.153	113	155257	45.525	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.040%			
50) Toluene-d8	10.547	98	555821	45.689	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 91.380%			
62) 4-Bromofluorobenzene	12.829	95	197243	43.886	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 87.780%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.371	101	188722	83.020	ug/l	95
27) cis-1,2-Dichloroethene	7.477	96	234562	70.144	ug/l	91
44) Trichloroethene	9.329	130	69531	20.207	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087566.D
Acq On : 14 Aug 2025 15:20
Operator : JC\MD
Sample : Q2842-01 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-08122025

Quant Time: Aug 14 15:39:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-08122025


Tgt Ion:168 Resp: 225616
Ion Ratio Lower Upper
168 100
99 66.9 47.9 71.9



#9
1,1,2-Trichlorotrifluoroethane
Concen: 83.020 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.012 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Tgt Ion:101 Resp: 188722
Ion Ratio Lower Upper
101 100
85 49.1 37.3 55.9
151 68.6 58.9 88.3









#27
cis-1,2-Dichloroethene
Concen: 70.144 ug/l
RT: 7.477 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-08122025


Tgt Ion:	96	Resp:	234562
Ion	Ratio	Lower	Upper
96	100		
61	135.3	0.0	297.8
98	62.2	0.0	132.4



#33
1,2-Dichloroethane-d4
Concen: 53.657 ug/l
RT: 8.565 min Scan# 1124
Delta R.T. 0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Tgt Ion:	65	Resp:	205410
Ion	Ratio	Lower	Upper
65	100		
67	51.7	0.0	104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN087566.D

Acq: 14 Aug 2025 15:20

Instrument :

MSVOA_N

ClientSampleId :

RMW-03B-90.4-08122025





Tgt Ion:114 Resp: 494400

Ion Ratio Lower Upper

114 100

63 22.9 0.0 44.6

88 15.5 0.0 32.8



#35

Dibromofluoromethane

Concen: 45.525 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087566.D

Acq: 14 Aug 2025 15:20




Tgt Ion:113 Resp: 155257


Ion Ratio Lower Upper

113 100

111 102.1 82.5 123.7

192 15.1 13.7 20.5









#44
Trichloroethene
Concen: 20.207 ug/l
RT: 9.329 min Scan# 1254
Delta R.T. -0.006 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-08122025


Tgt Ion:130 Resp: 69531
Ion Ratio Lower Upper
130 100
95 96.4 0.0 195.2



#50
Toluene-d8
Concen: 45.689 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Tgt Ion: 98 Resp: 555821
Ion Ratio Lower Upper
98 100
100 64.4 52.1 78.1









#62
4-Bromofluorobenzene
Concen: 43.886 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20




Instrument :
MSVOA_N
ClientSampleId :
RMW-03B-90.4-08122025


Tgt Ion	Resp	Ion Ratio	Lower	Upper
95	197243	100		
174		69.3	0.0	149.4
176		64.4	0.0	141.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087566.D
Acq: 14 Aug 2025 15:20

Tgt Ion	Resp	Ion Ratio	Lower	Upper
117	459245	100		
82		58.2	47.4	71.2
119		32.0	23.8	35.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087566.D


Acq: 14 Aug 2025 15:20

Instrument :

MSVOA_N

ClientSampleId :

RMW-03B-90.4-08122025





Tgt Ion:152 Resp: 213548

Ion Ratio Lower Upper

152 100

115 63.6 31.1 93.5

150 154.0 0.0 349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	RMW-02B-66.3-08122025	SDG No.:	Q2842
Lab Sample ID:	Q2842-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087567.D	40	08/14/25 15:42	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	240		10.4	40.0	ug/L
75-35-4	1,1-Dichloroethene	130		9.20	40.0	ug/L
75-34-3	1,1-Dichloroethane	9.20	U	9.20	40.0	ug/L
156-59-2	cis-1,2-Dichloroethene	1600		7.60	40.0	ug/L
71-55-6	1,1,1-Trichloroethane	8.00	U	8.00	40.0	ug/L
71-43-2	Benzene	6.00	U	6.00	40.0	ug/L
107-06-2	1,2-Dichloroethane	8.80	U	8.80	40.0	ug/L
79-01-6	Trichloroethene	1500		3.70	40.0	ug/L
79-00-5	1,1,2-Trichloroethane	8.40	U	8.40	40.0	ug/L
127-18-4	Tetrachloroethene	24.7	J	9.20	40.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.0		70 (74) - 130 (125)	114%	SPK: 50
1868-53-7	Dibromofluoromethane	45.1		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	46.0		70 (86) - 130 (113)	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.0		70 (77) - 130 (121)	90%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	219000	8.212			
540-36-3	1,4-Difluorobenzene	487000	9.088			
3114-55-4	Chlorobenzene-d5	439000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	211000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087567.D
 Acq On : 14 Aug 2025 15:42
 Operator : JC\MD
 Sample : Q2842-02 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-02B-66.3-08122025

Quant Time: Aug 14 19:59:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	218605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	486591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	438933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	211146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	211465	57.010	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.020%
35) Dibromofluoromethane	8.153	113	151404	45.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.220%
50) Toluene-d8	10.547	98	550924	46.014	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.829	95	199247	45.043	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.080%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.542	62	17532	6.042	ug/l	97
12) 1,1-Dichloroethene	4.330	96	8400	3.365	ug/l	94
27) cis-1,2-Dichloroethene	7.471	96	128340	39.610	ug/l	94
44) Trichloroethene	9.335	130	129163	38.140	ug/l	94
64) Tetrachloroethene	11.076	164	1746	0.618	ug/l #	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087567.D
Acq On : 14 Aug 2025 15:42
Operator : JC\MD
Sample : Q2842-02 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-08122025

Quant Time: Aug 14 19:59:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-08122025


Tgt Ion:168 Resp: 218605
Ion Ratio Lower Upper
168 100
99 65.5 47.9 71.9



#4
Vinyl Chloride
Concen: 6.042 ug/l
RT: 2.542 min Scan# 100
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42

Tgt Ion: 62 Resp: 17532
Ion Ratio Lower Upper
62 100
64 31.9 27.0 40.6





#12

1,1-Dichloroethene

Concen: 3.365 ug/l

RT: 4.330 min Scan# 404

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-08122025





Tgt Ion: 96 Resp: 8400

Ion Ratio Lower Upper

96 100

61 155.5 132.3 198.5

98 61.4 46.8 70.2



#27

cis-1,2-Dichloroethene

Concen: 39.610 ug/l

RT: 7.471 min Scan# 938

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42




Tgt Ion: 96 Resp: 128340


Ion Ratio Lower Upper

96 100

61 139.9 0.0 297.8

98 64.5 0.0 132.4





#33

1,2-Dichloroethane-d4

Concen: 57.010 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :





RMW-02B-66.3-08122025

Tgt Ion: 65 Resp: 211465

Ion Ratio Lower Upper

65 100

67 49.2 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42




Tgt Ion: 114 Resp: 486591


Ion Ratio Lower Upper

114 100

63 24.5 0.0 44.6

88 16.8 0.0 32.8





#35

Dibromofluoromethane

Concen: 45.108 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087567.D

Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-08122025





Tgt Ion:113 Resp: 151404

Ion Ratio Lower Upper

113 100

111 104.3 82.5 123.7

192 15.8 13.7 20.5



#44

Trichloroethene

Concen: 38.140 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087567.D




Acq: 14 Aug 2025 15:42


Tgt Ion:130 Resp: 129163

Ion Ratio Lower Upper

130 100

95 103.7 0.0 195.2









#50
Toluene-d8
Concen: 46.014 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-08122025


Tgt Ion: 98 Resp: 550924
Ion Ratio Lower Upper
98 100
100 64.8 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 45.043 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42

Tgt Ion: 95 Resp: 199247
Ion Ratio Lower Upper
95 100
174 66.4 0.0 149.4
176 64.9 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1170
Delta R.T. 0.000 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
RMW-02B-66.3-08122025


Tgt Ion:117 Resp: 438933
Ion Ratio Lower Upper
117 100
82 61.8 47.4 71.2
119 31.0 23.8 35.8



#64
Tetrachloroethene
Concen: 0.618 ug/l
RT: 11.076 min Scan# 1551
Delta R.T. -0.012 min
Lab File: VN087567.D
Acq: 14 Aug 2025 15:42

Tgt Ion:164 Resp: 1746
Ion Ratio Lower Upper
164 100
166 103.9 105.5 158.3#
129 82.5 77.4 116.2
131 70.4 77.3 115.9#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087567.D


Acq: 14 Aug 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

RMW-02B-66.3-08122025





Tgt Ion:152 Resp: 211146

Ion Ratio Lower Upper

152 100

115 63.3 31.1 93.5

150 156.0 0.0 349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-08122025	SDG No.:	Q2842
Lab Sample ID:	Q2842-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087568.D	100	08/14/25 16:04	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	26.0	U	26.0	100	ug/L
75-35-4	1,1-Dichloroethene	23.0	U	23.0	100	ug/L
75-34-3	1,1-Dichloroethane	23.0	U	23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	3000		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	20.0	U	20.0	100	ug/L
71-43-2	Benzene	15.0	U	15.0	100	ug/L
107-06-2	1,2-Dichloroethane	22.0	U	22.0	100	ug/L
79-01-6	Trichloroethene	3600		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	21.0	U	21.0	100	ug/L
127-18-4	Tetrachloroethene	56.3	J	23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.6		70 (74) - 130 (125)	113%	SPK: 50
1868-53-7	Dibromofluoromethane	45.1		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	46.5		70 (86) - 130 (113)	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.5		70 (77) - 130 (121)	91%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	220000	8.212			
540-36-3	1,4-Difluorobenzene	488000	9.088			
3114-55-4	Chlorobenzene-d5	455000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	209000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087568.D
 Acq On : 14 Aug 2025 16:04
 Operator : JC\MD
 Sample : Q2842-03 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-08122025

Quant Time: Aug 14 19:59:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	219937	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	487780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	455235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	209426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	211074	56.560	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.120%
35) Dibromofluoromethane	8.153	113	151622	45.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	10.547	98	557660	46.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.920%
62) 4-Bromofluorobenzene	12.829	95	201626	45.470	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.940%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	7.477	96	99130	30.410	ug/l	93
44) Trichloroethene	9.335	130	123655	36.424	ug/l	98
64) Tetrachloroethene	11.082	164	1650	0.563	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087568.D
Acq On : 14 Aug 2025 16:04
Operator : JC\MD
Sample : Q2842-03 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-08122025

Quant Time: Aug 14 19:59:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-08122025


Tgt Ion:168 Resp: 219937
Ion Ratio Lower Upper
168 100
99 65.5 47.9 71.9



#27
cis-1,2-Dichloroethene
Concen: 30.410 ug/l
RT: 7.477 min Scan# 939
Delta R.T. 0.006 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Tgt Ion: 96 Resp: 99130
Ion Ratio Lower Upper
96 100
61 139.9 0.0 297.8
98 62.0 0.0 132.4





#33

1,2-Dichloroethane-d4

Concen: 56.560 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04

Instrument :

MSVOA_N

ClientSampleId :





MW-17B-55.5-08122025

Tgt Ion: 65 Resp: 211074

Ion Ratio Lower Upper

65 100

67 49.4 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04




Tgt Ion: 114 Resp: 487780


Ion Ratio Lower Upper

114 100

63 22.1 0.0 44.6

88 15.4 0.0 32.8





#35

Dibromofluoromethane

Concen: 45.062 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087568.D

Acq: 14 Aug 2025 16:04

Instrument :

MSVOA_N

ClientSampleId :

MW-17B-55.5-08122025





Tgt Ion:113 Resp: 151622

Ion Ratio Lower Upper

113 100

111 104.2 82.5 123.7

192 15.7 13.7 20.5



#44

Trichloroethene

Concen: 36.424 ug/l

RT: 9.335 min Scan# 1255

Delta R.T. 0.000 min

Lab File: VN087568.D




Acq: 14 Aug 2025 16:04


Tgt Ion:130 Resp: 123655

Ion Ratio Lower Upper

130 100

95 99.4 0.0 195.2









#50
Toluene-d8
Concen: 46.463 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-08122025


Tgt Ion: 98 Resp: 557660
Ion Ratio Lower Upper
98 100
100 63.3 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 45.470 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Tgt Ion: 95 Resp: 201626
Ion Ratio Lower Upper
95 100
174 65.3 0.0 149.4
176 63.8 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04




Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-08122025


Tgt Ion:117 Resp: 455235
Ion Ratio Lower Upper
117 100
82 59.8 47.4 71.2
119 31.3 23.8 35.8



#64
Tetrachloroethene
Concen: 0.563 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.006 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04


Tgt Ion:164 Resp: 1650
Ion Ratio Lower Upper
164 100
166 122.4 105.5 158.3
129 103.2 77.4 116.2
131 52.6 77.3 115.9#







#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN087568.D
Acq: 14 Aug 2025 16:04

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-08122025



Tgt Ion	Ratio	Lower	Upper
152	100		
115	64.3	31.1	93.5
150	157.1	0.0	349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MS	SDG No.:	Q2842
Lab Sample ID:	Q2842-04MS	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087569.D	100	08/14/25 16:27	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5200		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5300		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8100		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5400		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5500		22.0	100	ug/L
79-01-6	Trichloroethene	10000		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.8		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	47.9		70 (75) - 130 (124)	96%	SPK: 50
2037-26-5	Toluene-d8	47.9		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	238000	8.206			
540-36-3	1,4-Difluorobenzene	461000	9.083			
3114-55-4	Chlorobenzene-d5	435000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	232000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MSD	SDG No.:	Q2842
Lab Sample ID:	Q2842-05MSD	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087570.D	100	08/14/25 16:49	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5300		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5400		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8000		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5500		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5300		22.0	100	ug/L
79-01-6	Trichloroethene	9500		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	46.7		70 (75) - 130 (124)	93%	SPK: 50
2037-26-5	Toluene-d8	48.2		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	252000	8.212			
540-36-3	1,4-Difluorobenzene	499000	9.082			
3114-55-4	Chlorobenzene-d5	460000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	240000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	TB01-081225	SDG No.:	Q2842
Lab Sample ID:	Q2842-08	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087565.D	1	08/14/25 14:58	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.1		70 (74) - 130 (125)	110%	SPK: 50
1868-53-7	Dibromofluoromethane	44.9		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	45.6		70 (86) - 130 (113)	91%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.2		70 (77) - 130 (121)	88%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	224000	8.212			
540-36-3	1,4-Difluorobenzene	499000	9.088			
3114-55-4	Chlorobenzene-d5	459000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	215000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087565.D
 Acq On : 14 Aug 2025 14:58
 Operator : JC\MD
 Sample : Q2842-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-081225

Quant Time: Aug 14 15:25:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.212	168	224412	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.088	114	498664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	458661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	214662	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	209873	55.117	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.240%
35) Dibromofluoromethane	8.153	113	154326	44.865	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.740%
50) Toluene-d8	10.547	98	559384	45.589	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.180%
62) 4-Bromofluorobenzene	12.829	95	200248	44.173	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.340%
Target Compounds						
16) Acetone	4.424	43	8905	4.988	ug/l	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087565.D
Acq On : 14 Aug 2025 14:58
Operator : JC\MD
Sample : Q2842-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-081225

Quant Time: Aug 14 15:25:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. -0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58




Instrument :
MSVOA_N
ClientSampleId :
TB01-081225


Tgt Ion:168 Resp: 224412
Ion Ratio Lower Upper
168 100
99 72.0 47.9 71.9#



#16
Acetone
Concen: 4.988 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58

Tgt Ion: 43 Resp: 8905
Ion Ratio Lower Upper
43 100
58 21.4 22.3 33.5#





#33

1,2-Dichloroethane-d4

Concen: 55.117 ug/l

RT: 8.565 min Scan# 1124

Delta R.T. 0.000 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :





TB01-081225

Tgt Ion: 65 Resp: 209873

Ion Ratio Lower Upper

65 100

67 50.1 0.0 104.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.088 min Scan# 1213

Delta R.T. 0.006 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58




Tgt Ion: 114 Resp: 498664


Ion Ratio Lower Upper

114 100

63 23.0 0.0 44.6

88 16.0 0.0 32.8





#35

Dibromofluoromethane

Concen: 44.865 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087565.D

Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :

TB01-081225





Tgt Ion:113 Resp: 154326

Ion Ratio Lower Upper

113 100

111 103.2 82.5 123.7

192 15.9 13.7 20.5



#50

Toluene-d8

Concen: 45.589 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. 0.000 min

Lab File: VN087565.D




Acq: 14 Aug 2025 14:58


Tgt Ion: 98 Resp: 559384

Ion Ratio Lower Upper

98 100

100 64.6 52.1 78.1









#62
4-Bromofluorobenzene
Concen: 44.173 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58




Instrument :
MSVOA_N
ClientSampleId :
TB01-081225


Tgt Ion	Resp	Lower	Upper
95	100		
174	68.6	0.0	149.4
176	63.7	0.0	141.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. -0.000 min
Lab File: VN087565.D
Acq: 14 Aug 2025 14:58

Tgt Ion	Ratio	Lower	Upper
117	100		
82	63.8	47.4	71.2
119	32.1	23.8	35.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.770 min Scan# 20

Delta R.T. 0.000 min

Lab File: VN087565.D


Acq: 14 Aug 2025 14:58

Instrument :

MSVOA_N

ClientSampleId :

TB01-081225





Tgt Ion:152 Resp: 214662

Ion Ratio Lower Upper

152 100

115 64.0 31.1 93.5

150 156.9 0.0 349.0





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG No.: Q2842
 Instrument ID: MSVOA_N Calibration Date(s): 07/16/2025 07/16/2025
 Heated Purge: (Y/N) N Calibration Time(s): 17:05 18:54
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN087328.D		RRF005 = VN087329.D		RRF020 = VN087330.D			
		RRF050 = VN087331.D		RRF100 = VN087332.D		RRF150 = VN087333.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Vinyl Chloride	0.554	0.665	0.623	0.728	0.692	0.720	0.664	9.9	
1,1-Dichloroethene	0.635	0.641	0.553	0.545	0.514	0.537	0.571	9.4	
1,1-Dichloroethane	1.304	1.325	1.254	1.222	1.185	1.212	1.250	4.4	
cis-1,2-Dichloroethene	0.704	0.737	0.747	0.767	0.740	0.751	0.741	2.8	
1,1,1-Trichloroethane	1.043	1.146	1.096	1.084	1.049	1.085	1.084	3.4	
Benzene	1.370	1.430	1.502	1.553	1.483	1.499	1.473	4.3	
1,2-Dichloroethane	0.553	0.565	0.569	0.567	0.544	0.552	0.558	1.8	
Trichloroethene	0.373	0.330	0.337	0.356	0.339	0.352	0.348	4.5	
1,1,2-Trichloroethane	0.367	0.364	0.365	0.367	0.357	0.354	0.362	1.6	
Tetrachloroethene	0.329	0.338	0.317	0.320	0.310	0.317	0.322	3.2	
1,2-Dichloroethane-d4		0.939	0.820	0.802	0.818	0.863	0.848	6.6	
Dibromofluoromethane		0.347	0.341	0.332	0.348	0.356	0.345	2.6	
Toluene-d8		1.180	1.176	1.224	1.255	1.316	1.230	4.7	
4-Bromofluorobenzene		0.405	0.433	0.455	0.476	0.505	0.455	8.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N071625W.M
 Title : SW846 8260
 Last Update : Thu Jul 17 02:56:13 2025
 Response Via : Initial Calibration

Calibration Files

1 =VN087328.D 5 =VN087329.D 20 =VN087330.D 50 =VN087331.D 100 =VN087332.D 150 =VN087333.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.447	0.443	0.443	0.623	0.606	0.625	0.531	17.98
3) P	Chloromethane	0.714	0.659	0.588	0.690	0.659	0.698	0.668	6.70
4) C	Vinyl Chloride	0.554	0.665	0.623	0.728	0.692	0.720	0.664	9.93#
5) T	Bromomethane		0.328	0.308	0.355	0.356	0.370	0.344	7.22
6) T	Chloroethane	0.396	0.485	0.431	0.441	0.415	0.429	0.433	6.96
7) T	Trichlorofluor...	0.959	0.975	0.963	1.025	0.960	1.007	0.981	2.83
8) T	Diethyl Ether	0.335	0.391	0.399	0.395	0.371	0.393	0.381	6.40
9) T	1,1,2-Trichlor...	0.463	0.495	0.539	0.525	0.491	0.509	0.504	5.34
10) T	Methyl Iodide		0.288	0.360	0.494	0.554	0.564	0.452	27.17
11) T	Tert butyl alc...		0.164	0.155	0.161	0.161	0.164	0.161	2.41
12) CM	1,1-Dichloroet...	0.635	0.641	0.553	0.545	0.514	0.537	0.571	9.41#
13) T	Acrolein		0.140	0.112	0.121	0.130	0.143	0.129	9.85
14) T	Allyl chloride	0.949	0.978	1.134	1.002	0.995	1.141	1.033	8.02
15) T	Acrylonitrile	0.419	0.419	0.455	0.453	0.436	0.440	0.437	3.57
16) T	Acetone	0.455	0.426	0.393	0.387	0.361	0.366	0.398	9.15
17) T	Carbon Disulfide	1.686	1.685	1.669	1.733	1.643	1.739	1.693	2.20
18) T	Methyl Acetate	1.007	1.052	0.991	0.991	0.962	0.993	0.999	2.99
19) T	Methyl tert-bu...	1.911	2.099	2.106	2.167	2.129	2.213	2.104	4.93
20) T	Methylene Chlo...	1.107	0.788	0.700	0.672	0.655	0.672	0.766	22.70
21) T	trans-1,2-Dich...	0.667	0.669	0.643	0.653	0.618	0.613	0.644	3.69
22) T	Diisopropyl ether	1.797	2.199	2.353	2.286	2.165	2.201	2.167	8.94
23) T	Vinyl Acetate	1.421	1.685	2.101	2.114	2.007	2.044	1.895	14.81
24) P	1,1-Dichloroet...	1.304	1.325	1.254	1.222	1.185	1.212	1.250	4.38
25) T	2-Butanone	0.552	0.608	0.650	0.643	0.617	0.618	0.615	5.65
26) T	2,2-Dichloropr...	1.013	0.963	1.001	0.982	0.933	0.941	0.972	3.34
27) T	cis-1,2-Dichlo...	0.704	0.737	0.747	0.767	0.740	0.751	0.741	2.83
28) T	Bromochloromet...	0.596	0.640	0.578	0.595	0.597	0.584	0.598	3.64
29) T	Tetrahydrofuran	0.360	0.397	0.413	0.425	0.400	0.401	0.399	5.52
30) C	Chloroform	1.181	1.299	1.303	1.279	1.214	1.234	1.251	3.97#
31) T	Cyclohexane		1.148	1.039	1.046	0.981	1.002	1.043	6.16
32) T	1,1,1-Trichlor...	1.043	1.146	1.096	1.084	1.049	1.085	1.084	3.42
33) S	1,2-Dichloroet...		0.939	0.820	0.802	0.818	0.863	0.848	6.56
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.347	0.341	0.332	0.348	0.356	0.345	2.55
36) T	1,1-Dichloropr...	0.387	0.437	0.459	0.487	0.477	0.487	0.456	8.47
37) T	Ethyl Acetate	0.587	0.671	0.655	0.706	0.660	0.669	0.658	5.94
38) T	Carbon Tetrach...	0.453	0.523	0.498	0.517	0.504	0.518	0.502	5.10
39) T	Methylcyclohexane	0.447	0.442	0.482	0.529	0.519	0.541	0.493	8.62
40) TM	Benzene	1.370	1.430	1.502	1.553	1.483	1.499	1.473	4.34
41) T	Methacrylonitrile	0.289	0.330	0.362	0.368	0.353	0.363	0.344	8.81
42) TM	1,2-Dichloroet...	0.553	0.565	0.569	0.567	0.544	0.552	0.558	1.78
43) T	Isopropyl Acetate	0.938	0.999	1.029	1.071	1.038	1.055	1.022	4.64
44) TM	Trichloroethene	0.373	0.330	0.337	0.356	0.339	0.352	0.348	4.49
45) C	1,2-Dichloropr...	0.335	0.367	0.395	0.395	0.376	0.378	0.374	5.90#
46) T	Dibromomethane	0.265	0.296	0.279	0.289	0.274	0.278	0.280	3.94
47) T	Bromodichlorom...	0.568	0.572	0.559	0.569	0.553	0.565	0.564	1.21
48) T	Methyl methacr...	0.372	0.412	0.475	0.505	0.493	0.503	0.460	12.06
49) T	1,4-Dioxane	0.006	0.007	0.007	0.008	0.008	0.008	0.007	12.52
50) S	Toluene-d8		1.180	1.176	1.224	1.255	1.316	1.230	4.72
51) T	4-Methyl-2-Pen...	0.551	0.641	0.685	0.689	0.658	0.652	0.646	7.79
52) CM	Toluene	0.774	0.849	0.940	0.963	0.916	0.929	0.895	7.93#
53) T	t-1,3-Dichloro...	0.459	0.536	0.586	0.621	0.607	0.619	0.571	11.08
54) T	cis-1,3-Dichlo...	0.489	0.564	0.602	0.632	0.620	0.632	0.590	9.40
55) T	1,1,2-Trichlor...	0.367	0.364	0.365	0.367	0.357	0.354	0.362	1.57
56) T	Ethyl methacry...	0.348	0.486	0.567	0.627	0.626	0.656	0.552	21.20


Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N071625W.M

57)	T	1,3-Dichloropr...	0.556	0.621	0.652	0.655	0.636	0.640	0.627	5.82
58)	T	2-Chloroethyl ...	0.220	0.273	0.303	0.346	0.310	0.332	0.297	15.20
59)	T	2-Hexanone	0.279	0.373	0.465	0.495	0.481	0.479	0.429	19.95
60)	T	Dibromochlorom...	0.352	0.416	0.425	0.430	0.424	0.433	0.413	7.39
61)	T	1,2-Dibromoethane	0.367	0.373	0.391	0.385	0.381	0.389	0.381	2.49
62)	S	4-Bromofluorob...	0.405	0.433	0.455	0.476	0.505	0.455		8.46
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.329	0.338	0.317	0.320	0.310	0.317	0.322	3.16
65)	PM	Chlorobenzene	1.139	1.131	1.133	1.119	1.092	1.122	1.123	1.48
66)	T	1,1,1,2-Tetrac...	0.324	0.406	0.393	0.393	0.381	0.394	0.382	7.64
67)	C	Ethyl Benzene	1.643	1.738	1.882	1.942	1.905	1.979	1.848	7.04#
68)	T	m/p-Xylenes	0.541	0.646	0.717	0.758	0.734	0.756	0.692	12.20
69)	T	o-Xylene	0.491	0.606	0.702	0.723	0.710	0.734	0.661	14.39
70)	T	Styrene	0.726	1.032	1.186	1.255	1.217	1.257	1.112	18.59
71)	P	Bromoform	0.246	0.302	0.314	0.328	0.322	0.337	0.308	10.69
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	2.524	2.889	3.242	3.396	3.302	3.529	3.147	11.86
74)	T	N-amyl acetate	1.309	1.256	1.249	1.138	1.378	1.515	1.307	9.83
75)	P	1,1,2,2-Tetrac...	1.159	1.181	1.228	1.207	1.156	1.174	1.184	2.37
76)	T	1,2,3-Trichlor...	1.208	1.238	1.101	1.191	0.969	1.019	1.121	9.79
77)	T	Bromobenzene	0.669	0.791	0.862	0.869	0.830	0.876	0.816	9.63
78)	T	n-propylbenzene	3.252	3.716	4.089	4.294	4.109	4.295	3.959	10.25
79)	T	2-Chlorotoluene	2.085	2.348	2.535	2.533	2.493	2.605	2.433	7.84
80)	T	1,3,5-Trimethy...	2.069	2.517	2.816	2.928	2.799	2.959	2.681	12.62
81)	T	trans-1,4-Dich...	0.327	0.448	0.404	0.417	0.453	0.410		12.30
82)	T	4-Chlorotoluene	2.225	2.460	2.618	2.647	2.552	2.699	2.533	6.79
83)	T	tert-Butylbenzene	1.781	2.068	2.303	2.423	2.376	2.485	2.239	11.92
84)	T	1,2,4-Trimethy...	2.201	2.411	2.906	3.003	2.872	3.034	2.738	12.64
85)	T	sec-Butylbenzene	3.120	3.136	3.460	3.565	3.387	3.570	3.373	5.98
86)	T	p-Isopropyltol...	2.081	2.417	2.806	2.991	2.893	3.032	2.703	13.90
87)	T	1,3-Dichlorobe...	1.448	1.592	1.651	1.657	1.588	1.674	1.602	5.19
88)	T	1,4-Dichlorobe...	1.807	1.709	1.742	1.703	1.627	1.677	1.711	3.56
89)	T	n-Butylbenzene	2.278	2.425	2.648	2.761	2.614	2.762	2.581	7.49
90)	T	Hexachloroethane	0.593	0.564	0.562	0.576	0.552	0.590	0.573	2.86
91)	T	1,2-Dichlorobe...	1.303	1.534	1.596	1.577	1.510	1.583	1.517	7.23
92)	T	1,2-Dibromo-3-...	0.339	0.325	0.304	0.302	0.290	0.305	0.311	5.76
93)	T	1,2,4-Trichlor...	0.761	0.860	0.875	0.937	0.921	0.994	0.891	8.94
94)	T	Hexachlorobuta...	0.351	0.337	0.324	0.329	0.316	0.331	0.331	3.57
95)	T	Naphthalene	2.486	2.658	3.113	3.441	3.451	3.797	3.158	16.01
96)	T	1,2,3-Trichlor...	0.803	0.844	0.887	0.922	0.917	0.992	0.894	7.37

(#) = Out of Range

Ethyl methacrylate


Response Ratio



$R = 2.028e-002 A^2 + 5.966e-001 A - 7.578e-003$
Coef of Det (r^2) = 0.999797 Curve Fit: Quadratic
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N071625W.M
Calibration Table Last Updated: Thu Jul 17 02:56:13 2025

Methylene Chloride

Response Ratio



$$\text{Response} = 6.638\text{e-}001 * \text{Amt} + 8.611\text{e-}003$$


Coef of Det (r^2) = 0.999669 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N071625W.M

Calibration Table Last Updated: Thu Jul 17 02:56:13 2025

Methyl Iodide

Response Ratio



$$\text{Response} = 5.841\text{e-}001 * \text{Amt} - 6.578\text{e-}002$$

Coef of Det (r^2) = 0.998697 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N071625W.M

Calibration Table Last Updated: Thu Jul 17 02:56:13 2025

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087328.D
 Acq On : 16 Jul 2025 17:05
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:16:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	166177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	312474	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	274838	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	128808	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

					Qvalue	
2) Dichlorodifluoromethane	2.142	85	1484	0.841	ug/l	98
3) Chloromethane	2.389	50	2372	1.069	ug/l	96
4) Vinyl Chloride	2.542	62	1841	0.835	ug/l #	68
6) Chloroethane	3.153	64	1316	0.915	ug/l	93
7) Trichlorofluoromethane	3.506	101	3187	0.977	ug/l	92
8) Diethyl Ether	3.965	74	1114	0.880	ug/l	79
9) 1,1,2-Trichlorotrifluo...	4.371	101	1539	0.919	ug/l #	31
12) 1,1-Dichloroethene	4.347	96	2112	1.113	ug/l #	53
14) Allyl chloride	5.012	41	3155m	0.919	ug/l	
15) Acrylonitrile	5.712	53	6965	4.794	ug/l #	80
16) Acetone	4.430	43	7566m	5.660	ug/l	
17) Carbon Disulfide	4.700	76	5602	0.996	ug/l #	71
18) Methyl Acetate	5.018	43	3347	1.008	ug/l #	74
19) Methyl tert-butyl Ether	5.789	73	6351m	0.908	ug/l	
20) Methylene Chloride	5.271	84	3679	1.019	ug/l #	74
21) trans-1,2-Dichloroethene	5.777	96	2216	1.036	ug/l #	53
22) Diisopropyl ether	6.659	45	5973	0.829	ug/l #	75
23) Vinyl Acetate	6.600	43	23617	3.749	ug/l #	93
24) 1,1-Dichloroethane	6.547	63	4333	1.043	ug/l #	83
25) 2-Butanone	7.477	43	9174	4.491	ug/l	99
26) 2,2-Dichloropropane	7.471	77	3368	1.043	ug/l #	67
27) cis-1,2-Dichloroethene	7.471	96	2340	0.950	ug/l	99
28) Bromochloromethane	7.794	49	1982	0.997	ug/l #	85
29) Tetrahydrofuran	7.835	42	5981	4.507	ug/l	94
30) Chloroform	7.953	83	3924	0.943	ug/l #	71
32) 1,1,1-Trichloroethane	8.165	97	3468	0.963	ug/l #	45
36) 1,1-Dichloropropene	8.353	75	2421	0.850	ug/l	96
37) Ethyl Acetate	7.547	43	3670	0.892	ug/l #	82
38) Carbon Tetrachloride	8.341	117	2833	0.903	ug/l #	83
39) Methylcyclohexane	9.588	83	2794	0.906	ug/l #	86
40) Benzene	8.588	78	8561	0.930	ug/l	97
41) Methacrylonitrile	7.771	41	1805	0.839	ug/l #	87
42) 1,2-Dichloroethane	8.653	62	3456	0.990	ug/l	89
43) Isopropyl Acetate	8.677	43	5865	0.919	ug/l #	95
44) Trichloroethene	9.341	130	2331	1.072	ug/l	70
45) 1,2-Dichloropropane	9.606	63	2094	0.895	ug/l #	86

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087328.D
 Acq On : 16 Jul 2025 17:05
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:16:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.694	93	1657	0.946	ug/l #	87
47) Bromodichloromethane	9.871	83	3549	1.006	ug/l #	91
48) Methyl methacrylate	9.671	41	2322	0.808	ug/l	94
49) 1,4-Dioxane	9.682	88	696	15.810	ug/l #	53
51) 4-Methyl-2-Pentanone	10.429	43	17216	4.263	ug/l	91
52) Toluene	10.612	92	4834	0.864	ug/l	91
53) t-1,3-Dichloropropene	10.818	75	2869	0.804	ug/l #	56
54) cis-1,3-Dichloropropene	10.294	75	3059	0.830	ug/l	94
55) 1,1,2-Trichloroethane	11.000	97	2294	1.013	ug/l #	83
56) Ethyl methacrylate	10.859	69	2173	1.217	ug/l #	90
57) 1,3-Dichloropropane	11.141	76	3477	0.888	ug/l	95
58) 2-Chloroethyl Vinyl ether	10.141	63	6887	3.707	ug/l	96
59) 2-Hexanone	11.182	43	8707	3.250	ug/l	87
60) Dibromochloromethane	11.347	129	2200	0.852	ug/l	82
61) 1,2-Dibromoethane	11.459	107	2291	0.962	ug/l	90
64) Tetrachloroethene	11.088	164	1807	1.022	ug/l #	75
65) Chlorobenzene	11.870	112	6259	1.014	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	1783	0.850	ug/l #	50
67) Ethyl Benzene	11.941	91	9031	0.889	ug/l	92
68) m/p-Xylenes	12.053	106	5952	1.565	ug/l	97
69) o-Xylene	12.388	106	2700	0.743	ug/l	91
70) Styrene	12.400	104	3989	0.653	ug/l	91
71) Bromoform	12.559	173	1351	0.797	ug/l #	94
73) Isopropylbenzene	12.682	105	6501	0.802	ug/l	95
74) N-amyl acetate	12.841	43	3372m	1.065	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	2986	0.979	ug/l #	78
76) 1,2,3-Trichloropropane	12.976	75	3113m	1.089	ug/l	
77) Bromobenzene	12.953	156	1724	0.820	ug/l	79
78) n-propylbenzene	13.017	91	8378	0.821	ug/l	95
79) 2-Chlorotoluene	13.100	91	5372	0.857	ug/l	93
80) 1,3,5-Trimethylbenzene	13.147	105	5329	0.772	ug/l	93
82) 4-Chlorotoluene	13.200	91	5733	0.878	ug/l	96
83) tert-Butylbenzene	13.412	119	4588	0.795	ug/l	90
84) 1,2,4-Trimethylbenzene	13.464	105	5671	0.804	ug/l	87
85) sec-Butylbenzene	13.594	105	8037	0.925	ug/l	97
86) p-Isopropyltoluene	13.712	119	5360	0.770	ug/l	92
87) 1,3-Dichlorobenzene	13.712	146	3731	0.904	ug/l	89
88) 1,4-Dichlorobenzene	13.794	146	4655m	1.056	ug/l	
89) n-Butylbenzene	14.035	91	5869	0.883	ug/l	91
90) Hexachloroethane	14.317	117	1527	1.035	ug/l	87
91) 1,2-Dichlorobenzene	14.088	146	3358	0.859	ug/l	85
92) 1,2-Dibromo-3-Chloropr...	14.711	75	874	1.091	ug/l #	56
93) 1,2,4-Trichlorobenzene	15.376	180	1960	0.854	ug/l	78
94) Hexachlorobutadiene	15.476	225	903	1.058	ug/l	76
95) Naphthalene	15.611	128	6404	0.787	ug/l #	95
96) 1,2,3-Trichlorobenzene	15.806	180	2068	0.898	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\  
Data File : VN087328.D  
Acq On    : 16 Jul 2025  17:05  
Operator  : JC\MD  
Sample    : VSTDIC001  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 3      Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:16:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087329.D
 Acq On : 16 Jul 2025 17:27
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	172217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	319926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	285583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	146494	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	16177	5.536	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	11.080%#		
35) Dibromofluoromethane	8.153	113	11111	5.035	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.060%#		
50) Toluene-d8	10.547	98	37751	4.796	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	9.600%#		
62) 4-Bromofluorobenzene	12.829	95	12943	4.450	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	8.900%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	7633	4.173	ug/l	90
3) Chloromethane	2.383	50	11350	4.934	ug/l	96
4) Vinyl Chloride	2.536	62	11459	5.013	ug/l	98
5) Bromomethane	2.971	94	5657	4.779	ug/l	87
6) Chloroethane	3.124	64	8358	5.606	ug/l #	81
7) Trichlorofluoromethane	3.506	101	16796	4.969	ug/l #	83
8) Diethyl Ether	3.959	74	6731	5.133	ug/l	91
9) 1,1,2-Trichlorotrifluo...	4.371	101	8532	4.917	ug/l #	19
10) Methyl Iodide	4.577	142	4952m	8.093	ug/l	
11) Tert butyl alcohol	5.536	59	14147	25.497	ug/l	98
12) 1,1-Dichloroethene	4.330	96	11037	5.613	ug/l #	83
13) Acrolein	4.177	56	12045	27.050	ug/l	96
14) Allyl chloride	5.006	41	16836	4.731	ug/l	97
15) Acrylonitrile	5.718	53	36114	23.986	ug/l	92
16) Acetone	4.430	43	36648	26.455	ug/l	96
17) Carbon Disulfide	4.695	76	29024	4.979	ug/l	94
18) Methyl Acetate	5.006	43	18125	5.265	ug/l	99
19) Methyl tert-butyl Ether	5.800	73	36155	4.989	ug/l	95
20) Methylene Chloride	5.259	84	13570	5.286	ug/l #	86
21) trans-1,2-Dichloroethene	5.765	96	11514	5.193	ug/l	81
22) Diisopropyl ether	6.671	45	37879	5.075	ug/l #	97
23) Vinyl Acetate	6.594	43	145089	22.225	ug/l #	94
24) 1,1-Dichloroethane	6.547	63	22819	5.299	ug/l	95
25) 2-Butanone	7.477	43	52336	24.722	ug/l	95
26) 2,2-Dichloropropane	7.483	77	16583	4.953	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	12690	4.972	ug/l	99
28) Bromochloromethane	7.789	49	11022	5.348	ug/l #	16
29) Tetrahydrofuran	7.830	42	34146	24.829	ug/l	99
30) Chloroform	7.959	83	22369	5.190	ug/l	84
31) Cyclohexane	8.236	56	19764	5.502	ug/l	87
32) 1,1,1-Trichloroethane	8.153	97	19739	5.287	ug/l #	51
36) 1,1-Dichloropropene	8.347	75	13965	4.790	ug/l	98
37) Ethyl Acetate	7.553	43	21482	5.102	ug/l	97
38) Carbon Tetrachloride	8.341	117	16718	5.205	ug/l #	89
39) Methylcyclohexane	9.582	83	14154	4.484	ug/l	86
40) Benzene	8.583	78	45748	4.855	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087329.D
 Acq On : 16 Jul 2025 17:27
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	10551	4.792	ug/l	97
42) 1,2-Dichloroethane	8.653	62	18086	5.061	ug/l	96
43) Isopropyl Acetate	8.677	43	31964	4.890	ug/l	100
44) Trichloroethene	9.335	130	10573	4.748	ug/l	79
45) 1,2-Dichloropropane	9.606	63	11728	4.898	ug/l	95
46) Dibromomethane	9.694	93	9473	5.284	ug/l	96
47) Bromodichloromethane	9.865	83	18300	5.068	ug/l #	82
48) Methyl methacrylate	9.665	41	13169	4.475	ug/l	95
49) 1,4-Dioxane	9.682	88	4176	92.653	ug/l #	87
51) 4-Methyl-2-Pentanone	10.430	43	102519	24.797	ug/l	97
52) Toluene	10.612	92	27158	4.742	ug/l	96
53) t-1,3-Dichloropropene	10.824	75	17133	4.688	ug/l #	87
54) cis-1,3-Dichloropropene	10.294	75	18046	4.780	ug/l	97
55) 1,1,2-Trichloroethane	11.000	97	11643	5.021	ug/l #	88
56) Ethyl methacrylate	10.859	69	15548	4.693	ug/l	95
57) 1,3-Dichloropropane	11.141	76	19868	4.956	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.147	63	43690	22.968	ug/l	94
59) 2-Hexanone	11.182	43	59736	21.778	ug/l	99
60) Dibromochloromethane	11.335	129	13321	5.037	ug/l	100
61) 1,2-Dibromoethane	11.453	107	11942	4.898	ug/l	89
64) Tetrachloroethene	11.082	164	9660	5.256	ug/l	95
65) Chlorobenzene	11.871	112	32304	5.038	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.941	131	11581	5.312	ug/l	96
67) Ethyl Benzene	11.947	91	49622	4.701	ug/l	92
68) m/p-Xylenes	12.059	106	36883	9.332	ug/l	98
69) o-Xylene	12.376	106	17293	4.580	ug/l	97
70) Styrene	12.394	104	29462	4.639	ug/l	99
71) Bromoform	12.559	173	8628	4.899	ug/l #	96
73) Isopropylbenzene	12.676	105	42320	4.590	ug/l	98
74) N-amyl acetate	12.647	43	18405m	5.113	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	17305	4.988	ug/l	95
76) 1,2,3-Trichloropropane	12.976	75	18135m	5.580	ug/l	
77) Bromobenzene	12.965	156	11589	4.847	ug/l	84
78) n-propylbenzene	13.018	91	54443	4.693	ug/l	99
79) 2-Chlorotoluene	13.106	91	34400	4.825	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	36874	4.694	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	4795	3.994	ug/l #	80
82) 4-Chlorotoluene	13.200	91	36031	4.854	ug/l	98
83) tert-Butylbenzene	13.418	119	30293	4.617	ug/l	96
84) 1,2,4-Trimethylbenzene	13.465	105	35326	4.403	ug/l	99
85) sec-Butylbenzene	13.594	105	45947	4.649	ug/l	98
86) p-Isopropyltoluene	13.706	119	35413	4.471	ug/l	97
87) 1,3-Dichlorobenzene	13.718	146	23326	4.970	ug/l	95
88) 1,4-Dichlorobenzene	13.794	146	25031	4.994	ug/l	90
89) n-Butylbenzene	14.035	91	35522	4.697	ug/l	95
90) Hexachloroethane	14.312	117	8261	4.923	ug/l	98
91) 1,2-Dichlorobenzene	14.088	146	22474	5.055	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	4757	5.223	ug/l	87
93) 1,2,4-Trichlorobenzene	15.364	180	12595	4.823	ug/l	97
94) Hexachlorobutadiene	15.476	225	4936	5.087	ug/l	98
95) Naphthalene	15.617	128	38936	4.208	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	12368	4.721	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087329.D
Acq On : 16 Jul 2025 17:27
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:17:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087330.D
 Acq On : 16 Jul 2025 17:49
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	178514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	328159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	297202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	156249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	58584	19.341	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	38.680%#		
35) Dibromofluoromethane	8.153	113	44726	19.758	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	39.520%#		
50) Toluene-d8	10.547	98	154381	19.119	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	38.240%#		
62) 4-Bromofluorobenzene	12.829	95	56832	19.051	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	38.100%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	31604	16.669	ug/l	86
3) Chloromethane	2.383	50	41982	17.608	ug/l	94
4) Vinyl Chloride	2.542	62	44508	18.784	ug/l	99
5) Bromomethane	2.971	94	22025	17.950	ug/l	98
6) Chloroethane	3.130	64	30785	19.922	ug/l	100
7) Trichlorofluoromethane	3.506	101	68764	19.626	ug/l	97
8) Diethyl Ether	3.959	74	28520	20.984	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	38516	21.414	ug/l	95
10) Methyl Iodide	4.577	142	25698	17.955	ug/l	97
11) Tert butyl alcohol	5.530	59	55298	96.147	ug/l	99
12) 1,1-Dichloroethene	4.330	96	39480	19.370	ug/l	96
13) Acrolein	4.177	56	40135	86.954	ug/l	98
14) Allyl chloride	5.006	41	80983	21.955	ug/l	88
15) Acrylonitrile	5.712	53	162325	104.007	ug/l	99
16) Acetone	4.424	43	140161	97.607	ug/l	93
17) Carbon Disulfide	4.700	76	119191	19.725	ug/l #	93
18) Methyl Acetate	5.012	43	70735	19.824	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	150385	20.018	ug/l	97
20) Methylene Chloride	5.265	84	49951	20.427	ug/l	87
21) trans-1,2-Dichloroethene	5.771	96	45917	19.980	ug/l	90
22) Diisopropyl ether	6.659	45	168043	21.719	ug/l	97
23) Vinyl Acetate	6.594	43	750202	110.863	ug/l	98
24) 1,1-Dichloroethane	6.559	63	89552	20.062	ug/l	95
25) 2-Butanone	7.477	43	231983	105.718	ug/l	96
26) 2,2-Dichloropropane	7.483	77	71478	20.596	ug/l	97
27) cis-1,2-Dichloroethene	7.471	96	53347	20.162	ug/l	98
28) Bromochloromethane	7.794	49	41285	19.325	ug/l	100
29) Tetrahydrofuran	7.830	42	147436	103.426	ug/l	99
30) Chloroform	7.953	83	93018	20.819	ug/l	100
31) Cyclohexane	8.241	56	74161	19.915	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	78235	20.217	ug/l	97
36) 1,1-Dichloropropene	8.359	75	60302	20.163	ug/l	98
37) Ethyl Acetate	7.553	43	85921	19.893	ug/l	98
38) Carbon Tetrachloride	8.347	117	65336	19.832	ug/l	95
39) Methylcyclohexane	9.588	83	63264	19.539	ug/l	95
40) Benzene	8.588	78	197187	20.400	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087330.D
 Acq On : 16 Jul 2025 17:49
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	47550	21.054	ug/l	97
42) 1,2-Dichloroethane	8.659	62	74680	20.374	ug/l	99
43) Isopropyl Acetate	8.677	43	135010	20.136	ug/l	99
44) Trichloroethene	9.335	130	44202	19.354	ug/l	98
45) 1,2-Dichloropropane	9.606	63	51786	21.086	ug/l	97
46) Dibromomethane	9.688	93	36673	19.943	ug/l	96
47) Bromodichloromethane	9.871	83	73409	19.819	ug/l	93
48) Methyl methacrylate	9.665	41	62316	20.645	ug/l	99
49) 1,4-Dioxane	9.682	88	18509	400.357	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	449710	106.046	ug/l	99
52) Toluene	10.612	92	123391	21.002	ug/l	98
53) t-1,3-Dichloropropene	10.818	75	76959	20.530	ug/l	92
54) cis-1,3-Dichloropropene	10.294	75	79019	20.407	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	47960	20.164	ug/l	94
56) Ethyl methacrylate	10.859	69	74408	19.384	ug/l	99
57) 1,3-Dichloropropane	11.147	76	85537	20.800	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	198704	101.839	ug/l	100
59) 2-Hexanone	11.182	43	305342	108.526	ug/l	99
60) Dibromochloromethane	11.341	129	55783	20.564	ug/l	98
61) 1,2-Dibromoethane	11.447	107	51292	20.509	ug/l	97
64) Tetrachloroethene	11.082	164	37655	19.686	ug/l	93
65) Chlorobenzene	11.871	112	134668	20.183	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.941	131	46671	20.570	ug/l	98
67) Ethyl Benzene	11.947	91	223715	20.367	ug/l	100
68) m/p-Xylenes	12.053	106	170590	41.473	ug/l	98
69) o-Xylene	12.376	106	83415	21.230	ug/l	98
70) Styrene	12.388	104	140937	21.323	ug/l	100
71) Bromoform	12.559	173	37362	20.383	ug/l #	99
73) Isopropylbenzene	12.676	105	202595	20.601	ug/l	100
74) N-amyl acetate	12.512	43	78074m	20.334	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.918	83	76720	20.733	ug/l	98
76) 1,2,3-Trichloropropane	12.976	75	68839m	19.858	ug/l	
77) Bromobenzene	12.959	156	53881	21.127	ug/l	97
78) n-propylbenzene	13.017	91	255577	20.656	ug/l	99
79) 2-Chlorotoluene	13.106	91	158436	20.836	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	176006	21.006	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	27975	21.846	ug/l	90
82) 4-Chlorotoluene	13.200	91	163615	20.667	ug/l	99
83) tert-Butylbenzene	13.417	119	143954	20.571	ug/l	98
84) 1,2,4-Trimethylbenzene	13.465	105	181623	21.226	ug/l	100
85) sec-Butylbenzene	13.594	105	216256	20.516	ug/l	99
86) p-Isopropyltoluene	13.706	119	175352	20.758	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	103157	20.609	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	108898	20.370	ug/l	98
89) n-Butylbenzene	14.035	91	165484	20.516	ug/l	99
90) Hexachloroethane	14.312	117	35108	19.615	ug/l	96
91) 1,2-Dichlorobenzene	14.088	146	99750	21.036	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.700	75	19015	19.573	ug/l	98
93) 1,2,4-Trichlorobenzene	15.370	180	54714	19.643	ug/l	99
94) Hexachlorobutadiene	15.476	225	20248	19.563	ug/l	96
95) Naphthalene	15.617	128	194574	19.718	ug/l	100
96) 1,2,3-Trichlorobenzene	15.811	180	55463	19.850	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087330.D
Acq On : 16 Jul 2025 17:49
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:18:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	182792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	325711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	300150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	158477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	146559	47.253	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.500%		
35) Dibromofluoromethane	8.147	113	108271	48.190	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.380%		
50) Toluene-d8	10.547	98	398776	49.757	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.520%		
62) 4-Bromofluorobenzene	12.829	95	148158	50.037	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.080%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	113853	58.643	ug/l	100
3) Chloromethane	2.383	50	126054	51.631	ug/l	100
4) Vinyl Chloride	2.542	62	132985	54.810	ug/l	100
5) Bromomethane	2.971	94	64982	51.719	ug/l	100
6) Chloroethane	3.130	64	80606	50.942	ug/l	100
7) Trichlorofluoromethane	3.506	101	187302	52.206	ug/l	100
8) Diethyl Ether	3.959	74	72151	51.843	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.359	101	95895	52.068	ug/l	100
10) Methyl Iodide	4.577	142	90277	47.912	ug/l	100
11) Tert butyl alcohol	5.524	59	146939	249.504	ug/l	100
12) 1,1-Dichloroethene	4.330	96	99569	47.708	ug/l	100
13) Acrolein	4.177	56	110711	234.247	ug/l	100
14) Allyl chloride	5.012	41	183089	48.475	ug/l	100
15) Acrylonitrile	5.712	53	414273	259.227	ug/l	100
16) Acetone	4.424	43	353373	240.328	ug/l	100
17) Carbon Disulfide	4.694	76	316833	51.205	ug/l	100
18) Methyl Acetate	5.012	43	181199	49.594	ug/l	100
19) Methyl tert-butyl Ether	5.788	73	396133	51.495	ug/l	100
20) Methylene Chloride	5.259	84	122921	50.001	ug/l	100
21) trans-1,2-Dichloroethene	5.771	96	119278	50.687	ug/l	100
22) Diisopropyl ether	6.665	45	417843	52.740	ug/l	100
23) Vinyl Acetate	6.588	43	1931809	278.796	ug/l	100
24) 1,1-Dichloroethane	6.553	63	223298	48.853	ug/l	100
25) 2-Butanone	7.471	43	588032	261.702	ug/l	100
26) 2,2-Dichloropropane	7.471	77	179431	50.492	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	140194	51.746	ug/l	100
28) Bromochloromethane	7.800	49	108748	49.713	ug/l	100
29) Tetrahydrofuran	7.829	42	388704	266.294	ug/l	100
30) Chloroform	7.953	83	233747	51.092	ug/l	100
31) Cyclohexane	8.241	56	191116	50.122	ug/l	100
32) 1,1,1-Trichloroethane	8.153	97	198135	50.002	ug/l	100
36) 1,1-Dichloropropene	8.353	75	158563	53.418	ug/l	100
37) Ethyl Acetate	7.547	43	230005	53.652	ug/l	100
38) Carbon Tetrachloride	8.347	117	168403	51.501	ug/l	100
39) Methylcyclohexane	9.582	83	172297	53.614	ug/l	100
40) Benzene	8.594	78	505723	52.714	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.771	41	119895	53.486	ug/l	100
42) 1,2-Dichloroethane	8.653	62	184670	50.759	ug/l	100
43) Isopropyl Acetate	8.677	43	348709	52.399	ug/l	100
44) Trichloroethene	9.335	130	115974	51.160	ug/l	100
45) 1,2-Dichloropropane	9.606	63	128530	52.726	ug/l	100
46) Dibromomethane	9.688	93	94179	51.601	ug/l	100
47) Bromodichloromethane	9.871	83	185222	50.382	ug/l	100
48) Methyl methacrylate	9.665	41	164476	54.899	ug/l	100
49) 1,4-Dioxane	9.682	88	52072	1134.802	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	1122857	266.771	ug/l	100
52) Toluene	10.612	92	313806	53.814	ug/l	100
53) t-1,3-Dichloropropene	10.818	75	202129	54.327	ug/l	100
54) cis-1,3-Dichloropropene	10.294	75	205720	53.529	ug/l	100
55) 1,1,2-Trichloroethane	11.000	97	119612	50.666	ug/l	100
56) Ethyl methacrylate	10.859	69	204376	51.427	ug/l	100
57) 1,3-Dichloropropane	11.141	76	213343	52.267	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	562790	290.607	ug/l	100
59) 2-Hexanone	11.182	43	806824	288.920	ug/l	100
60) Dibromochloromethane	11.341	129	139955	51.982	ug/l	100
61) 1,2-Dibromoethane	11.453	107	125416	50.525	ug/l	100
64) Tetrachloroethene	11.088	164	96176	49.786	ug/l	100
65) Chlorobenzene	11.870	112	335833	49.837	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.941	131	117833	51.425	ug/l	100
67) Ethyl Benzene	11.947	91	582851	52.540	ug/l	100
68) m/p-Xylenes	12.053	106	455017	109.536	ug/l	100
69) o-Xylene	12.376	106	217121	54.717	ug/l	100
70) Styrene	12.394	104	376745	56.440	ug/l	100
71) Bromoform	12.559	173	98549	53.237	ug/l #	100
73) Isopropylbenzene	12.676	105	538245	53.964	ug/l	100
74) N-amyl acetate	12.494	43	180343m	46.310	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	191273	50.964	ug/l	100
76) 1,2,3-Trichloropropane	12.970	75	188824m	53.705	ug/l	
77) Bromobenzene	12.959	156	137669	53.221	ug/l	100
78) n-propylbenzene	13.017	91	680506	54.227	ug/l	100
79) 2-Chlorotoluene	13.106	91	401444	52.051	ug/l	100
80) 1,3,5-Trimethylbenzene	13.153	105	464063	54.607	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.717	75	63993	49.271	ug/l	100
82) 4-Chlorotoluene	13.200	91	419474	52.240	ug/l	100
83) tert-Butylbenzene	13.417	119	383986	54.099	ug/l	100
84) 1,2,4-Trimethylbenzene	13.464	105	475982	54.846	ug/l	100
85) sec-Butylbenzene	13.594	105	565038	52.851	ug/l	100
86) p-Isopropyltoluene	13.706	119	473942	55.316	ug/l	100
87) 1,3-Dichlorobenzene	13.711	146	262675	51.740	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	269881	49.773	ug/l	100
89) n-Butylbenzene	14.035	91	437514	53.478	ug/l	100
90) Hexachloroethane	14.311	117	91313	50.301	ug/l	100
91) 1,2-Dichlorobenzene	14.082	146	249988	51.977	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.700	75	47931	48.643	ug/l	100
93) 1,2,4-Trichlorobenzene	15.370	180	148472	52.553	ug/l	100
94) Hexachlorobutadiene	15.476	225	52130	49.659	ug/l	100
95) Naphthalene	15.617	128	545367	54.490	ug/l	100
96) 1,2,3-Trichlorobenzene	15.811	180	146080	51.547	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087331.D
Acq On : 16 Jul 2025 18:11
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:19:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087331.D
 Acq On : 16 Jul 2025 18:11
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

Client Sample Id :

VSTDICCC050

Manual Integrations

APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025

Supervised By : Semsettin Yesilyurt 07/17/2025


Quant Time: Jul 17 02:19:33 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087332.D
 Acq On : 16 Jul 2025 18:32
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Jul 17 02:20:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	192275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	339591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	314849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	168489	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.559	65	314416	96.373	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 192.740%#	
35) Dibromofluoromethane	8.153	113	236468	100.947	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 201.900%#	
50) Toluene-d8	10.547	98	852301	101.999	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 204.000%#	
62) 4-Bromofluorobenzene	12.829	95	323070	104.650	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 209.300%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	233197	114.190	ug/l	92
3) Chloromethane	2.383	50	253369	98.660	ug/l	98
4) Vinyl Chloride	2.542	62	266221	104.312	ug/l	94
5) Bromomethane	2.959	94	136793	103.503	ug/l	96
6) Chloroethane	3.124	64	159422	95.783	ug/l	92
7) Trichlorofluoromethane	3.501	101	369032	97.786	ug/l	98
8) Diethyl Ether	3.959	74	142652	97.445	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.353	101	188841	97.477	ug/l	100
10) Methyl Iodide	4.571	142	213127	100.524	ug/l	99
11) Tert butyl alcohol	5.536	59	309615	499.800	ug/l	99
12) 1,1-Dichloroethene	4.336	96	197760	90.083	ug/l	95
13) Acrolein	4.177	56	250115	503.102	ug/l	99
14) Allyl chloride	5.006	41	382812	96.355	ug/l	100
15) Acrylonitrile	5.706	53	839080	499.150	ug/l	100
16) Acetone	4.424	43	693485	448.375	ug/l	96
17) Carbon Disulfide	4.695	76	631695	97.056	ug/l	95
18) Methyl Acetate	5.012	43	369894	96.247	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	818763	101.186	ug/l	98
20) Methylene Chloride	5.259	84	251830	98.001	ug/l	97
21) trans-1,2-Dichloroethene	5.771	96	237704	96.030	ug/l	94
22) Diisopropyl ether	6.659	45	832709	99.921	ug/l	96
23) Vinyl Acetate	6.594	43	3859420	529.516	ug/l	97
24) 1,1-Dichloroethane	6.553	63	455820	94.807	ug/l	100
25) 2-Butanone	7.471	43	1185700	501.667	ug/l	98
26) 2,2-Dichloropropane	7.477	77	358596	95.932	ug/l	99
27) cis-1,2-Dichloroethene	7.477	96	284578	99.858	ug/l	98
28) Bromochloromethane	7.800	49	229544	99.757	ug/l	100
29) Tetrahydrofuran	7.830	42	768521	500.532	ug/l	98
30) Chloroform	7.953	83	466838	97.008	ug/l	99
31) Cyclohexane	8.241	56	377159	94.035	ug/l	95
32) 1,1,1-Trichloroethane	8.153	97	403445	96.794	ug/l	97
36) 1,1-Dichloropropene	8.353	75	324047	104.705	ug/l	99
37) Ethyl Acetate	7.547	43	448428	100.326	ug/l	99
38) Carbon Tetrachloride	8.347	117	342043	100.328	ug/l	99
39) Methylcyclohexane	9.582	83	352211	105.118	ug/l	96
40) Benzene	8.588	78	1007257	100.700	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087332.D
 Acq On : 16 Jul 2025 18:32
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:20:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	239659	102.544	ug/l	99
42) 1,2-Dichloroethane	8.653	62	369731	97.472	ug/l	99
43) Isopropyl Acetate	8.677	43	704936	101.598	ug/l	99
44) Trichloroethene	9.335	130	230373	97.472	ug/l	93
45) 1,2-Dichloropropane	9.606	63	255369	100.478	ug/l	96
46) Dibromomethane	9.694	93	185908	97.696	ug/l	98
47) Bromodichloromethane	9.871	83	375910	98.072	ug/l	97
48) Methyl methacrylate	9.665	41	334941	107.228	ug/l	97
49) 1,4-Dioxane	9.688	88	102159	2135.348	ug/l #	95
51) 4-Methyl-2-Pentanone	10.429	43	2234274	509.127	ug/l	100
52) Toluene	10.612	92	622259	102.349	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	412142	106.245	ug/l	100
54) cis-1,3-Dichloropropene	10.294	75	421386	105.164	ug/l	98
55) 1,1,2-Trichloroethane	11.000	97	242683	98.595	ug/l	97
56) Ethyl methacrylate	10.859	69	425458	98.980	ug/l	98
57) 1,3-Dichloropropane	11.147	76	431753	101.453	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	1052628	521.327	ug/l	99
59) 2-Hexanone	11.176	43	1631949	560.508	ug/l	99
60) Dibromochloromethane	11.341	129	287682	102.484	ug/l	99
61) 1,2-Dibromoethane	11.453	107	258936	100.051	ug/l	98
64) Tetrachloroethene	11.082	164	195044	96.252	ug/l	97
65) Chlorobenzene	11.871	112	687696	97.289	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	239722	99.736	ug/l	99
67) Ethyl Benzene	11.947	91	1199510	103.080	ug/l	99
68) m/p-Xylenes	12.053	106	924069	212.065	ug/l	99
69) o-Xylene	12.376	106	447092	107.413	ug/l	100
70) Styrene	12.394	104	766471	109.464	ug/l	98
71) Bromoform	12.559	173	202915	104.498	ug/l #	99
73) Isopropylbenzene	12.676	105	1112821	104.940	ug/l	100
74) N-amyl acetate	12.482	43	464246	112.129	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.918	83	389392	97.587	ug/l	98
76) 1,2,3-Trichloropropane	12.970	75	326434m	87.327	ug/l	
77) Bromobenzene	12.959	156	279589	101.662	ug/l	100
78) n-propylbenzene	13.018	91	1384517	103.771	ug/l	99
79) 2-Chlorotoluene	13.106	91	840073	102.452	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	943119	104.383	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	140673	101.874	ug/l	96
82) 4-Chlorotoluene	13.200	91	859965	100.734	ug/l	99
83) tert-Butylbenzene	13.418	119	800700	106.106	ug/l	99
84) 1,2,4-Trimethylbenzene	13.459	105	967878	104.898	ug/l	100
85) sec-Butylbenzene	13.594	105	1141320	100.410	ug/l	98
86) p-Isopropyltoluene	13.706	119	975014	107.037	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	535172	99.151	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	548212	95.097	ug/l	99
89) n-Butylbenzene	14.035	91	880763	101.260	ug/l	99
90) Hexachloroethane	14.312	117	185973	96.358	ug/l	98
91) 1,2-Dichlorobenzene	14.082	146	508902	99.523	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	97572	93.137	ug/l	94
93) 1,2,4-Trichlorobenzene	15.364	180	310448	103.357	ug/l	99
94) Hexachlorobutadiene	15.476	225	106429	95.360	ug/l	98
95) Naphthalene	15.611	128	1162930	109.289	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	308888	102.519	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087332.D
Acq On : 16 Jul 2025 18:32
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:20:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087332.D
 Acq On : 16 Jul 2025 18:32
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025


Quant Time: Jul 17 02:20:26 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Jul 17 02:21:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	193853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	344966	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	317026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	165077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	501760	152.545	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 305.080%#			
35) Dibromofluoromethane	8.153	113	368306	154.778	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 309.560%#			
50) Toluene-d8	10.547	98	1362098	160.469	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 320.940%#			
62) 4-Bromofluorobenzene	12.829	95	522221	166.525	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 333.040%#			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	363327	176.463	ug/l	95
3) Chloromethane	2.383	50	405800	156.728	ug/l	99
4) Vinyl Chloride	2.542	62	418507	162.646	ug/l	94
5) Bromomethane	2.942	94	215335	161.604	ug/l	95
6) Chloroethane	3.118	64	249466	148.663	ug/l	100
7) Trichlorofluoromethane	3.501	101	585457	153.872	ug/l	97
8) Diethyl Ether	3.959	74	228544	154.847	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.353	101	296076	151.587	ug/l	99
10) Methyl Iodide	4.571	142	328079	150.516	ug/l	99
11) Tert butyl alcohol	5.536	59	478279	765.784	ug/l	100
12) 1,1-Dichloroethene	4.330	96	312325	141.111	ug/l	97
13) Acrolein	4.177	56	415488	828.945	ug/l	100
14) Allyl chloride	5.006	41	663403	165.621	ug/l	91
15) Acrylonitrile	5.706	53	1279363	754.869	ug/l	100
16) Acetone	4.424	43	1064061	682.373	ug/l	96
17) Carbon Disulfide	4.695	76	1011315	154.118	ug/l	95
18) Methyl Acetate	5.012	43	577536	149.052	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	1286728	157.724	ug/l	95
20) Methylene Chloride	5.259	84	390988	151.266	ug/l	94
21) trans-1,2-Dichloroethene	5.765	96	356549	142.870	ug/l	96
22) Diisopropyl ether	6.659	45	1280267	152.375	ug/l	99
23) Vinyl Acetate	6.589	43	5943020	808.750	ug/l	98
24) 1,1-Dichloroethane	6.553	63	704727	145.384	ug/l	98
25) 2-Butanone	7.477	43	1797146	754.179	ug/l	98
26) 2,2-Dichloropropane	7.477	77	547190	145.193	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	437021	152.102	ug/l	99
28) Bromochloromethane	7.800	49	339529	146.355	ug/l	100
29) Tetrahydrofuran	7.830	42	1166702	753.679	ug/l	98
30) Chloroform	7.953	83	717424	147.865	ug/l	98
31) Cyclohexane	8.241	56	582988	144.170	ug/l	94
32) 1,1,1-Trichloroethane	8.153	97	630986	150.152	ug/l	98
36) 1,1-Dichloropropene	8.353	75	503793	160.248	ug/l	99
37) Ethyl Acetate	7.553	43	692256	152.464	ug/l	98
38) Carbon Tetrachloride	8.347	117	535575	154.647	ug/l	98
39) Methylcyclohexane	9.583	83	559851	164.485	ug/l	98
40) Benzene	8.588	78	1550984	152.642	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:09:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	375482	158.156	ug/l	99
42) 1,2-Dichloroethane	8.653	62	571634	148.352	ug/l	100
43) Isopropyl Acetate	8.677	43	1091736	154.893	ug/l	99
44) Trichloroethene	9.335	130	364760	151.927	ug/l	96
45) 1,2-Dichloropropane	9.606	63	391676	151.708	ug/l	97
46) Dibromomethane	9.694	93	287252	148.602	ug/l	98
47) Bromodichloromethane	9.871	83	584546	150.128	ug/l	99
48) Methyl methacrylate	9.665	41	520990	164.191	ug/l	99
49) 1,4-Dioxane	9.688	88	157400	3238.745	ug/l #	93
51) 4-Methyl-2-Pentanone	10.430	43	3375607	757.219	ug/l	99
52) Toluene	10.612	92	961314	155.654	ug/l	99
53) t-1,3-Dichloropropene	10.818	75	640188	162.462	ug/l	94
54) cis-1,3-Dichloropropene	10.294	75	654301	160.747	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	365858	146.322	ug/l	99
56) Ethyl methacrylate	10.859	69	679202	150.296	ug/l	96
57) 1,3-Dichloropropane	11.147	76	662154	153.168	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	1717427	837.324	ug/l	100
59) 2-Hexanone	11.177	43	2477578	837.689	ug/l	98
60) Dibromochloromethane	11.341	129	448307	157.217	ug/l	99
61) 1,2-Dibromoethane	11.453	107	402978	153.282	ug/l	97
64) Tetrachloroethene	11.082	164	301404	147.718	ug/l	97
65) Chlorobenzene	11.871	112	1066757	149.878	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	375162	155.013	ug/l	98
67) Ethyl Benzene	11.941	91	1881920	160.613	ug/l	99
68) m/p-Xylenes	12.053	106	1437290	327.579	ug/l	98
69) o-Xylene	12.376	106	698363	166.628	ug/l	98
70) Styrene	12.394	104	1195068	169.502	ug/l	99
71) Bromoform	12.559	173	320960	164.154	ug/l #	99
73) Isopropylbenzene	12.676	105	1747567	168.203	ug/l	100
74) N-amyl acetate	12.482	43	750104	184.917	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.918	83	581568	148.761	ug/l	100
76) 1,2,3-Trichloropropane	12.971	75	504791m	137.832	ug/l	
77) Bromobenzene	12.959	156	433814	161.001	ug/l	99
78) n-propylbenzene	13.018	91	2127215	162.733	ug/l	100
79) 2-Chlorotoluene	13.106	91	1290205	160.600	ug/l	98
80) 1,3,5-Trimethylbenzene	13.153	105	1465159	165.514	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.718	75	224191	165.713	ug/l	95
82) 4-Chlorotoluene	13.200	91	1336462	159.785	ug/l	100
83) tert-Butylbenzene	13.418	119	1230683	166.457	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	1502651	166.222	ug/l	100
85) sec-Butylbenzene	13.594	105	1767949	158.754	ug/l	99
86) p-Isopropyltoluene	13.706	119	1501303	168.219	ug/l	98
87) 1,3-Dichlorobenzene	13.712	146	828885	156.742	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	830255	147.000	ug/l	99
89) n-Butylbenzene	14.035	91	1367813	160.505	ug/l	98
90) Hexachloroethane	14.312	117	292183	154.518	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	784106	156.513	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.694	75	151090	147.203	ug/l	93
93) 1,2,4-Trichlorobenzene	15.370	180	492237	167.266	ug/l	99
94) Hexachlorobutadiene	15.476	225	163922	149.909	ug/l	99
95) Naphthalene	15.612	128	1880423	180.370	ug/l	99
96) 1,2,3-Trichlorobenzene	15.812	180	491195	166.396	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087333.D
Acq On : 16 Jul 2025 18:54
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 02:21:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:09:29 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087333.D
 Acq On : 16 Jul 2025 18:54
 Operator : JC\MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_N

Client Sample Id :

VSTDIC150

Manual Integrations

APPROVED

Reviewed By : Mahesh Dadoda 07/17/2025

Supervised By : Semsettin Yesilyurt 07/17/2025


Quant Time: Jul 17 02:21:25 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:09:29 2025

Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
 Supervised By :Semsettin Yesilyurt 07/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	198299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	343165	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	319692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	166959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	162194	48.204	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.400%		
35) Dibromofluoromethane	8.153	113	118201	49.934	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.860%		
50) Toluene-d8	10.547	98	435474	51.573	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.140%		
62) 4-Bromofluorobenzene	12.829	95	165160	52.942	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	105.880%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	122218	58.029	ug/l	94
3) Chloromethane	2.383	50	135040	50.986	ug/l	99
4) Vinyl Chloride	2.542	62	141054	53.589	ug/l	93
5) Bromomethane	2.971	94	79496	58.322	ug/l	100
6) Chloroethane	3.130	64	84406	49.172	ug/l	97
7) Trichlorofluoromethane	3.506	101	192022	49.336	ug/l	93
8) Diethyl Ether	3.959	74	76789	50.861	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	98940	49.520	ug/l	99
10) Methyl Iodide	4.577	142	103627	50.369	ug/l	98
11) Tert butyl alcohol	5.530	59	167953	262.884	ug/l	99
12) 1,1-Dichloroethene	4.330	96	102861	45.432	ug/l	97
13) Acrolein	4.171	56	142098	277.145	ug/l	97
14) Allyl chloride	5.006	41	200790	49.004	ug/l	99
15) Acrylonitrile	5.706	53	437956	252.616	ug/l	99
16) Acetone	4.424	43	375335	237.913	ug/l	100
17) Carbon Disulfide	4.700	76	335264	49.947	ug/l	96
18) Methyl Acetate	5.018	43	194657	49.111	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	425099	50.939	ug/l	94
20) Methylene Chloride	5.265	84	131885	49.445	ug/l	93
21) trans-1,2-Dichloroethene	5.771	96	120707	47.283	ug/l	97
22) Diisopropyl ether	6.659	45	439978	51.191	ug/l	97
23) Vinyl Acetate	6.594	43	2052885	273.101	ug/l	98
24) 1,1-Dichloroethane	6.553	63	238854	48.170	ug/l	99
25) 2-Butanone	7.477	43	622581	255.411	ug/l	98
26) 2,2-Dichloropropane	7.477	77	176180	45.700	ug/l	100
27) cis-1,2-Dichloroethene	7.471	96	148552	50.543	ug/l	96
28) Bromochloromethane	7.800	49	116494	49.089	ug/l	98
29) Tetrahydrofuran	7.830	42	412353	260.404	ug/l	100
30) Chloroform	7.947	83	239752	48.306	ug/l	100
31) Cyclohexane	8.241	56	202863	49.042	ug/l	96
32) 1,1,1-Trichloroethane	8.153	97	212405	49.412	ug/l	99
36) 1,1-Dichloropropene	8.353	75	167912	53.690	ug/l	99
37) Ethyl Acetate	7.553	43	241759	53.525	ug/l	99
38) Carbon Tetrachloride	8.347	117	181974	52.821	ug/l	90
39) Methylcyclohexane	9.582	83	179820	53.109	ug/l	96
40) Benzene	8.588	78	529446	52.380	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

ICVVN071625

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025

Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:00:24 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M

Quant Title : SW846 8260

QLast Update : Thu Jul 17 02:56:13 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	127286	53.895	ug/l	98
42) 1,2-Dichloroethane	8.653	62	194804	50.821	ug/l	98
43) Isopropyl Acetate	8.677	43	374957	53.477	ug/l	100
44) Trichloroethene	9.335	130	118775	49.731	ug/l	95
45) 1,2-Dichloropropane	9.606	63	135864	52.900	ug/l	99
46) Dibromomethane	9.688	93	98534	51.241	ug/l	98
47) Bromodichloromethane	9.871	83	200236	51.696	ug/l	97
48) Methyl methacrylate	9.665	41	177068	56.096	ug/l	99
49) 1,4-Dioxane	9.682	88	56917	1177.300	ug/l #	97
51) 4-Methyl-2-Pentanone	10.429	43	1192765	268.966	ug/l	99
52) Toluene	10.612	92	320922	52.236	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	214141	54.628	ug/l	97
54) cis-1,3-Dichloropropene	10.294	75	216023	53.351	ug/l	96
55) 1,1,2-Trichloroethane	11.000	97	124090	49.889	ug/l	97
56) Ethyl methacrylate	10.859	69	222410	53.042	ug/l	98
57) 1,3-Dichloropropane	11.147	76	224493	52.202	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	581689	285.088	ug/l	99
59) 2-Hexanone	11.176	43	849057	288.579	ug/l	100
60) Dibromochloromethane	11.341	129	149449	52.685	ug/l	100
61) 1,2-Dibromoethane	11.447	107	135330	51.746	ug/l	94
64) Tetrachloroethene	11.082	164	98391	47.819	ug/l	97
65) Chlorobenzene	11.870	112	347934	48.477	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.941	131	123321	50.530	ug/l	99
67) Ethyl Benzene	11.947	91	608779	51.523	ug/l	99
68) m/p-Xylenes	12.053	106	465762	105.269	ug/l	97
69) o-Xylene	12.376	106	224934	53.221	ug/l	99
70) Styrene	12.394	104	387319	54.477	ug/l	98
71) Bromoform	12.559	173	102502	51.987	ug/l #	99
73) Isopropylbenzene	12.676	105	573938	54.619	ug/l	100
74) N-amyl acetate	12.494	43	218170m	49.972	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	203947	51.580	ug/l	98
76) 1,2,3-Trichloropropane	12.976	75	175797m	46.955	ug/l	
77) Bromobenzene	12.959	156	142714	52.368	ug/l	99
78) n-propylbenzene	13.012	91	709437	53.661	ug/l	99
79) 2-Chlorotoluene	13.106	91	426373	52.475	ug/l	99
80) 1,3,5-Trimethylbenzene	13.153	105	484100	54.071	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.717	75	69894	51.080	ug/l	97
82) 4-Chlorotoluene	13.200	91	443817	52.464	ug/l	99
83) tert-Butylbenzene	13.417	119	413882	55.349	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	500792	54.773	ug/l	99
85) sec-Butylbenzene	13.594	105	594712	52.800	ug/l	98
86) p-Isopropyltoluene	13.706	119	496291	54.982	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	277102	51.809	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	282084	49.381	ug/l	99
89) n-Butylbenzene	14.035	91	458964	53.250	ug/l	99
90) Hexachloroethane	14.312	117	93464	48.870	ug/l	95
91) 1,2-Dichlorobenzene	14.088	146	260874	51.485	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	50904	49.035	ug/l	97
93) 1,2,4-Trichlorobenzene	15.370	180	155818	52.351	ug/l	99
94) Hexachlorobutadiene	15.476	225	55599	50.273	ug/l	98
95) Naphthalene	15.617	128	586097	55.585	ug/l	98
96) 1,2,3-Trichlorobenzene	15.811	180	157909	52.890	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Manual Integrations

Reviewed By :Mahesh Dadoda 07/17/2025
Supervised By :Semsettin Yesilyurt 07/17/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.531	0.616	-16.0	107	0.00
3 P	Chloromethane	0.668	0.681	-1.9	107	0.00
4 C	Vinyl Chloride	0.664	0.711	-7.1#	106	0.00
5 T	Bromomethane	0.344	0.401	-16.6	122	0.00
6 T	Chloroethane	0.433	0.426	1.6	105	0.00
7 T	Trichlorofluoromethane	0.981	0.968	1.3	103	0.00
8 T	Diethyl Ether	0.381	0.387	-1.6	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.499	1.0	103	0.00
10 T	Methyl Iodide	0.452	0.523	-15.7	115	0.00
11 T	Tert butyl alcohol	0.161	0.169	-5.0	114	0.00
12 CM	1,1-Dichloroethene	0.571	0.519	9.1#	103	0.00
13 T	Acrolein	0.129	0.143	-10.9	128	0.00
14 T	Allyl chloride	1.033	1.013	1.9	110	0.00
15 T	Acrylonitrile	0.437	0.442	-1.1	106	0.00
16 T	Acetone	0.398	0.379	4.8	106	0.00
17 T	Carbon Disulfide	1.693	1.691	0.1	106	0.00
18 T	Methyl Acetate	0.999	0.982	1.7	107	0.00
19 T	Methyl tert-butyl Ether	2.104	2.144	-1.9	107	0.00
20 T	Methylene Chloride	0.766	0.665	13.2	107	0.00
21 T	trans-1,2-Dichloroethene	0.644	0.609	5.4	101	0.00
22 T	Diisopropyl ether	2.167	2.219	-2.4	105	0.00
23 T	Vinyl Acetate	1.895	2.070	-9.2	106	0.00
24 P	1,1-Dichloroethane	1.250	1.205	3.6	107	0.00
25 T	2-Butanone	0.615	0.628	-2.1	106	0.00
26 T	2,2-Dichloropropane	0.972	0.888	8.6	98	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.749	-1.1	106	0.00
28 T	Bromochloromethane	0.598	0.587	1.8	107	0.00
29 T	Tetrahydrofuran	0.399	0.416	-4.3	106	0.00
30 C	Chloroform	1.251	1.209	3.4#	103	0.00
31 T	Cyclohexane	1.043	1.023	1.9	106	0.00
32 T	1,1,1-Trichloroethane	1.084	1.071	1.2	107	0.00
33 S	1,2-Dichloroethane-d4	0.848	0.818	3.5	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.345	0.344	0.3	109	0.00
36 T	1,1-Dichloropropene	0.456	0.489	-7.2	106	0.00
37 T	Ethyl Acetate	0.658	0.704	-7.0	105	0.00
38 T	Carbon Tetrachloride	0.502	0.530	-5.6	108	0.00
39 T	Methylcyclohexane	0.493	0.524	-6.3	104	0.00
40 TM	Benzene	1.473	1.543	-4.8	105	0.00
41 T	Methacrylonitrile	0.344	0.371	-7.8	106	0.00
42 TM	1,2-Dichloroethane	0.558	0.568	-1.8	105	0.00
43 T	Isopropyl Acetate	1.022	1.093	-6.9	108	0.00
44 TM	Trichloroethene	0.348	0.346	0.6	102	0.00
45 C	1,2-Dichloropropane	0.374	0.396	-5.9#	106	0.00
46 T	Dibromomethane	0.280	0.287	-2.5	105	0.00
47 T	Bromodichloromethane	0.564	0.583	-3.4	108	0.00
48 T	Methyl methacrylate	0.460	0.516	-12.2	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	109	0.00
50 S	Toluene-d8	1.230	1.269	-3.2	109	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.695	-7.6	106	0.00
52 CM	Toluene	0.895	0.935	-4.5#	102	0.00
53 T	t-1,3-Dichloropropene	0.571	0.624	-9.3	106	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.630	-6.8	105	0.00
55 T	1,1,2-Trichloroethane	0.362	0.362	0.0	104	0.00
56 T	Ethyl methacrylate	0.552	0.648	-17.4	109	0.00
57 T	1,3-Dichloropropane	0.627	0.654	-4.3	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.339	-14.1	103	0.00
59 T	2-Hexanone	0.429	0.495	-15.4	105	0.00
60 T	Dibromochloromethane	0.413	0.436	-5.6	107	0.00
61 T	1,2-Dibromoethane	0.381	0.394	-3.4	108	0.00
62 S	4-Bromofluorobenzene	0.455	0.481	-5.7	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.322	0.308	4.3	102	0.00
65 PM	Chlorobenzene	1.123	1.088	3.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.386	-1.0	105	0.00
67 C	Ethyl Benzene	1.848	1.904	-3.0#	104	0.00
68 T	m/p-Xylenes	0.692	0.728	-5.2	102	0.00
69 T	o-Xylene	0.661	0.704	-6.5	104	0.00
70 T	Styrene	1.112	1.212	-9.0	103	0.00
71 P	Bromoform	0.308	0.321	-4.2	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.147	3.438	-9.2	107	0.00
74 T	N-amyl acetate	1.307	1.307	0.0	121	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.222	-3.2	107	0.00
76 T	1,2,3-Trichloropropane	1.121	1.053	6.1	93	0.00
77 T	Bromobenzene	0.816	0.855	-4.8	104	0.00
78 T	n-propylbenzene	3.959	4.249	-7.3	104	0.00
79 T	2-Chlorotoluene	2.433	2.554	-5.0	106	0.00
80 T	1,3,5-Trimethylbenzene	2.681	2.900	-8.2	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.410	0.419	-2.2	109	0.00
82 T	4-Chlorotoluene	2.533	2.658	-4.9	106	0.00
83 T	tert-Butylbenzene	2.239	2.479	-10.7	108	0.00
84 T	1,2,4-Trimethylbenzene	2.738	2.999	-9.5	105	0.00
85 T	sec-Butylbenzene	3.373	3.562	-5.6	105	0.00
86 T	p-Isopropyltoluene	2.703	2.973	-10.0	105	0.00
87 T	1,3-Dichlorobenzene	1.602	1.660	-3.6	105	0.00
88 T	1,4-Dichlorobenzene	1.711	1.690	1.2	105	0.00
89 T	n-Butylbenzene	2.581	2.749	-6.5	105	0.00
90 T	Hexachloroethane	0.573	0.560	2.3	102	0.00
91 T	1,2-Dichlorobenzene	1.517	1.563	-3.0	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.311	0.305	1.9	106	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.933	-4.7	105	0.00
94 T	Hexachlorobutadiene	0.331	0.333	-0.6	107	0.00
95 T	Naphthalene	3.158	3.510	-11.1	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.894	0.946	-5.8	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	58.029	-16.1	107	0.00
3 P	Chloromethane	50.000	50.986	-2.0	107	0.00
4 C	Vinyl Chloride	50.000	53.589	-7.2#	106	0.00
5 T	Bromomethane	50.000	58.322	-16.6	122	0.00
6 T	Chloroethane	50.000	49.172	1.7	105	0.00
7 T	Trichlorofluoromethane	50.000	49.336	1.3	103	0.00
8 T	Diethyl Ether	50.000	50.861	-1.7	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.520	1.0	103	0.00
10 T	Methyl Iodide	50.000	50.369	-0.7	115	0.00
11 T	Tert butyl alcohol	250.000	262.884	-5.2	114	0.00
12 CM	1,1-Dichloroethene	50.000	45.432	9.1#	103	0.00
13 T	Acrolein	250.000	277.145	-10.9	128	0.00
14 T	Allyl chloride	50.000	49.004	2.0	110	0.00
15 T	Acrylonitrile	250.000	252.616	-1.0	106	0.00
16 T	Acetone	250.000	237.913	4.8	106	0.00
17 T	Carbon Disulfide	50.000	49.947	0.1	106	0.00
18 T	Methyl Acetate	50.000	49.111	1.8	107	0.00
19 T	Methyl tert-butyl Ether	50.000	50.939	-1.9	107	0.00
20 T	Methylene Chloride	50.000	49.445	1.1	107	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.283	5.4	101	0.00
22 T	Diisopropyl ether	50.000	51.191	-2.4	105	0.00
23 T	Vinyl Acetate	250.000	273.101	-9.2	106	0.00
24 P	1,1-Dichloroethane	50.000	48.170	3.7	107	0.00
25 T	2-Butanone	250.000	255.411	-2.2	106	0.00
26 T	2,2-Dichloropropane	50.000	45.700	8.6	98	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.543	-1.1	106	0.00
28 T	Bromochloromethane	50.000	49.089	1.8	107	0.00
29 T	Tetrahydrofuran	250.000	260.404	-4.2	106	0.00
30 C	Chloroform	50.000	48.306	3.4#	103	0.00
31 T	Cyclohexane	50.000	49.042	1.9	106	0.00
32 T	1,1,1-Trichloroethane	50.000	49.412	1.2	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.204	3.6	111	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35 S	Dibromofluoromethane	50.000	49.934	0.1	109	0.00
36 T	1,1-Dichloropropene	50.000	53.690	-7.4	106	0.00
37 T	Ethyl Acetate	50.000	53.525	-7.0	105	0.00
38 T	Carbon Tetrachloride	50.000	52.821	-5.6	108	0.00
39 T	Methylcyclohexane	50.000	53.109	-6.2	104	0.00
40 TM	Benzene	50.000	52.380	-4.8	105	0.00
41 T	Methacrylonitrile	50.000	53.895	-7.8	106	0.00
42 TM	1,2-Dichloroethane	50.000	50.821	-1.6	105	0.00
43 T	Isopropyl Acetate	50.000	53.477	-7.0	108	0.00
44 TM	Trichloroethene	50.000	49.731	0.5	102	0.00
45 C	1,2-Dichloropropane	50.000	52.900	-5.8#	106	0.00
46 T	Dibromomethane	50.000	51.241	-2.5	105	0.00
47 T	Bromodichloromethane	50.000	51.696	-3.4	108	0.00
48 T	Methyl methacrylate	50.000	56.096	-12.2	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
 Data File : VN087335.D
 Acq On : 16 Jul 2025 19:59
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN071625

Quant Time: Jul 17 03:00:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1177.300	-17.7	109	0.00
50 S	Toluene-d8	50.000	51.573	-3.1	109	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.966	-7.6	106	0.00
52 CM	Toluene	50.000	52.236	-4.5#	102	0.00
53 T	t-1,3-Dichloropropene	50.000	54.628	-9.3	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.351	-6.7	105	0.00
55 T	1,1,2-Trichloroethane	50.000	49.889	0.2	104	0.00
56 T	Ethyl methacrylate	50.000	53.042	-6.1	109	0.00
57 T	1,3-Dichloropropane	50.000	52.202	-4.4	105	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	285.088	-14.0	103	0.00
59 T	2-Hexanone	250.000	288.579	-15.4	105	0.00
60 T	Dibromochloromethane	50.000	52.685	-5.4	107	0.00
61 T	1,2-Dibromoethane	50.000	51.746	-3.5	108	0.00
62 S	4-Bromofluorobenzene	50.000	52.942	-5.9	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	47.819	4.4	102	0.00
65 PM	Chlorobenzene	50.000	48.477	3.0	104	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.530	-1.1	105	0.00
67 C	Ethyl Benzene	50.000	51.523	-3.0#	104	0.00
68 T	m/p-Xylenes	100.000	105.269	-5.3	102	0.00
69 T	o-Xylene	50.000	53.221	-6.4	104	0.00
70 T	Styrene	50.000	54.477	-9.0	103	0.00
71 P	Bromoform	50.000	51.987	-4.0	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	54.619	-9.2	107	0.00
74 T	N-ethyl acetate	50.000	49.972	0.1	121	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.580	-3.2	107	0.00
76 T	1,2,3-Trichloropropane	50.000	46.955	6.1	93	0.00
77 T	Bromobenzene	50.000	52.368	-4.7	104	0.00
78 T	n-propylbenzene	50.000	53.661	-7.3	104	0.00
79 T	2-Chlorotoluene	50.000	52.475	-5.0	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.071	-8.1	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.080	-2.2	109	0.00
82 T	4-Chlorotoluene	50.000	52.464	-4.9	106	0.00
83 T	tert-Butylbenzene	50.000	55.349	-10.7	108	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.773	-9.5	105	0.00
85 T	sec-Butylbenzene	50.000	52.800	-5.6	105	0.00
86 T	p-Isopropyltoluene	50.000	54.982	-10.0	105	0.00
87 T	1,3-Dichlorobenzene	50.000	51.809	-3.6	105	0.00
88 T	1,4-Dichlorobenzene	50.000	49.381	1.2	105	0.00
89 T	n-Butylbenzene	50.000	53.250	-6.5	105	0.00
90 T	Hexachloroethane	50.000	48.870	2.3	102	0.00
91 T	1,2-Dichlorobenzene	50.000	51.485	-3.0	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.035	1.9	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.351	-4.7	105	0.00
94 T	Hexachlorobutadiene	50.000	50.273	-0.5	107	0.00
95 T	Naphthalene	50.000	55.585	-11.2	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087335.D
Acq On : 16 Jul 2025 19:59
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071625

Quant Time: Jul 17 03:00:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.890	-5.8	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2842</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/14/2025</u> <u>09:34</u>
Lab File ID:	<u>VN087552.D</u>	Init. Calib. Date(s):	<u>07/16/2025</u> <u>07/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>17:05</u> <u>18:54</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.678		2.11	20
1,1-Dichloroethene	0.571	0.562		-1.58	20
1,1-Dichloroethane	1.250	1.280	0.1	2.4	20
cis-1,2-Dichloroethene	0.741	0.782		5.53	20
1,1,1-Trichloroethane	1.084	1.139		5.07	20
Benzene	1.473	1.404		-4.68	20
1,2-Dichloroethane	0.558	0.573		2.69	20
Trichloroethene	0.348	0.305		-12.36	20
1,1,2-Trichloroethane	0.362	0.350		-3.32	20
Tetrachloroethene	0.322	0.260		-19.25	20
1,2-Dichloroethane-d4	0.848	0.912		7.55	20
Dibromofluoromethane	0.345	0.310		-10.15	20
Toluene-d8	1.230	1.125		-8.54	20
4-Bromofluorobenzene	0.455	0.433		-4.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/14/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	271352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	543098	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	500619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	255173	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.565	65	247424	53.738	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.480%	
35) Dibromofluoromethane	8.153	113	168221	44.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 89.800%	
50) Toluene-d8	10.547	98	611058	45.726	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 91.460%	
62) 4-Bromofluorobenzene	12.829	95	235280	47.655	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.300%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.142	85	160120	55.557	ug/l	96
3) Chloromethane	2.383	50	169400	46.740	ug/l	99
4) Vinyl Chloride	2.542	62	183878	51.052	ug/l	96
5) Bromomethane	2.965	94	102629	55.023	ug/l	98
6) Chloroethane	3.130	64	126750	53.961	ug/l	97
7) Trichlorofluoromethane	3.506	101	267426	50.212	ug/l	97
8) Diethyl Ether	3.965	74	117073	56.667	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	145784	53.322	ug/l	96
10) Methyl Iodide	4.577	142	103333	38.232	ug/l #	89
11) Tert butyl alcohol	5.536	59	241449	276.178	ug/l	100
12) 1,1-Dichloroethene	4.336	96	152389	49.187	ug/l	99
13) Acrolein	4.177	56	160450	228.690	ug/l	99
14) Allyl chloride	5.012	41	286033	51.014	ug/l	90
15) Acrylonitrile	5.712	53	598909	252.452	ug/l	98
16) Acetone	4.424	43	657966	304.782	ug/l	97
17) Carbon Disulfide	4.700	76	405663	44.164	ug/l	98
18) Methyl Acetate	5.018	43	317943	58.620	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	659220	57.727	ug/l	95
20) Methylene Chloride	5.259	84	192849	52.881	ug/l	97
21) trans-1,2-Dichloroethene	5.777	96	170971	48.942	ug/l	88
22) Diisopropyl ether	6.659	45	674122	57.318	ug/l	95
23) Vinyl Acetate	6.589	43	3071299	298.585	ug/l	100
24) 1,1-Dichloroethane	6.553	63	347405	51.200	ug/l	98
25) 2-Butanone	7.471	43	892036	267.432	ug/l	97
26) 2,2-Dichloropropane	7.471	77	298522	56.588	ug/l	98
27) cis-1,2-Dichloroethene	7.477	96	212209	52.764	ug/l	95
28) Bromochloromethane	7.794	49	184324	56.761	ug/l	92
29) Tetrahydrofuran	7.824	42	572583	264.243	ug/l	100
30) Chloroform	7.947	83	372180	54.800	ug/l	98
31) Cyclohexane	8.235	56	278777	49.251	ug/l	98
32) 1,1,1-Trichloroethane	8.153	97	309142	52.554	ug/l	98
36) 1,1-Dichloropropene	8.353	75	237939	48.073	ug/l	97
37) Ethyl Acetate	7.547	43	340534	47.639	ug/l	98
38) Carbon Tetrachloride	8.347	117	246721	45.251	ug/l	97
39) Methylcyclohexane	9.582	83	254444	47.484	ug/l	96
40) Benzene	8.588	78	762513	47.666	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/14/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	178825	47.844	ug/l	96
42) 1,2-Dichloroethane	8.653	62	311061	51.277	ug/l	97
43) Isopropyl Acetate	8.671	43	575331	51.848	ug/l	98
44) Trichloroethene	9.335	130	165523	43.791	ug/l	95
45) 1,2-Dichloropropane	9.606	63	200161	49.245	ug/l	92
46) Dibromomethane	9.694	93	147474	48.459	ug/l	96
47) Bromodichloromethane	9.871	83	298595	48.711	ug/l	100
48) Methyl methacrylate	9.665	41	263753	52.798	ug/l	93
49) 1,4-Dioxane	9.682	88	77554	1013.618	ug/l #	99
51) 4-Methyl-2-Pentanone	10.429	43	1771617	252.428	ug/l	98
52) Toluene	10.612	92	472630	48.609	ug/l	95
53) t-1,3-Dichloropropene	10.818	75	316497	51.017	ug/l	96
54) cis-1,3-Dichloropropene	10.294	75	326996	51.028	ug/l #	86
55) 1,1,2-Trichloroethane	11.000	97	190088	48.289	ug/l	95
56) Ethyl methacrylate	10.859	69	331191	50.043	ug/l #	92
57) 1,3-Dichloropropane	11.147	76	336112	49.384	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	934209	289.306	ug/l	99
59) 2-Hexanone	11.176	43	1221722	262.377	ug/l	98
60) Dibromochloromethane	11.341	129	217527	48.455	ug/l	100
61) 1,2-Dibromoethane	11.453	107	202290	48.875	ug/l	97
64) Tetrachloroethene	11.082	164	130358	40.459	ug/l	93
65) Chlorobenzene	11.871	112	508857	45.275	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.941	131	183045	47.896	ug/l	99
67) Ethyl Benzene	11.947	91	914813	49.442	ug/l	100
68) m/p-Xylenes	12.053	106	683841	98.699	ug/l	95
69) o-Xylene	12.376	106	337782	51.038	ug/l	95
70) Styrene	12.394	104	583897	52.445	ug/l	97
71) Bromoform	12.559	173	143351	46.429	ug/l #	99
73) Isopropylbenzene	12.676	105	854755	53.222	ug/l	98
74) N-amyl acetate	12.500	43	259021	38.819	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.918	83	304854	50.447	ug/l	100
76) 1,2,3-Trichloropropane	12.976	75	257264m	44.960	ug/l	
77) Bromobenzene	12.959	156	203849	48.942	ug/l	93
78) n-propylbenzene	13.012	91	1079956	53.447	ug/l	97
79) 2-Chlorotoluene	13.106	91	652303	52.528	ug/l	95
80) 1,3,5-Trimethylbenzene	13.153	105	735691	53.765	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.718	75	101369	48.472	ug/l	95
82) 4-Chlorotoluene	13.200	91	684119	52.913	ug/l	96
83) tert-Butylbenzene	13.417	119	613860	53.713	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	762982	54.601	ug/l	95
85) sec-Butylbenzene	13.594	105	915066	53.157	ug/l	98
86) p-Isopropyltoluene	13.706	119	762503	55.271	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	399195	48.835	ug/l	99
88) 1,4-Dichlorobenzene	13.794	146	400669	45.893	ug/l	97
89) n-Butylbenzene	14.035	91	736497	55.909	ug/l	98
90) Hexachloroethane	14.312	117	139358	47.677	ug/l	96
91) 1,2-Dichlorobenzene	14.082	146	381913	49.316	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	70382	44.360	ug/l	99
93) 1,2,4-Trichlorobenzene	15.370	180	226838	49.866	ug/l	98
94) Hexachlorobutadiene	15.476	225	82198	48.630	ug/l	99
95) Naphthalene	15.617	128	838952	52.059	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	220386	48.298	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 08/14/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	148	0.00
2 T	Dichlorodifluoromethane	0.531	0.590	-11.1	141	0.00
3 P	Chloromethane	0.668	0.624	6.6	134	0.00
4 C	Vinyl Chloride	0.664	0.678	-2.1#	138	0.00
5 T	Bromomethane	0.344	0.378	-9.9	158#	0.00
6 T	Chloroethane	0.433	0.467	-7.9	157#	0.00
7 T	Trichlorofluoromethane	0.981	0.986	-0.5	143	0.00
8 T	Diethyl Ether	0.381	0.431	-13.1	162#	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.504	0.537	-6.5	152#	0.00
10 T	Methyl Iodide	0.452	0.381	15.7	114	0.00
11 T	Tert butyl alcohol	0.161	0.178	-10.6	164#	0.01
12 CM	1,1-Dichloroethene	0.571	0.562	1.6#	153#	0.00
13 T	Acrolein	0.129	0.118	8.5	145	0.00
14 T	Allyl chloride	1.033	1.054	-2.0	156#	0.00
15 T	Acrylonitrile	0.437	0.441	-0.9	145	0.00
16 T	Acetone	0.398	0.485	-21.9	186#	0.00
17 T	Carbon Disulfide	1.693	1.495	11.7	128	0.00
18 T	Methyl Acetate	0.999	1.172	-17.3	175#	0.00
19 T	Methyl tert-butyl Ether	2.104	2.429	-15.4	166#	0.00
20 T	Methylene Chloride	0.766	0.711	7.2	157#	0.00
21 T	trans-1,2-Dichloroethene	0.644	0.630	2.2	143	0.00
22 T	Diisopropyl ether	2.167	2.484	-14.6	161#	0.00
23 T	Vinyl Acetate	1.895	2.264	-19.5	159#	0.00
24 P	1,1-Dichloroethane	1.250	1.280	-2.4	156#	0.00
25 T	2-Butanone	0.615	0.657	-6.8	152#	0.00
26 T	2,2-Dichloropropane	0.972	1.100	-13.2	166#	0.00
27 T	cis-1,2-Dichloroethene	0.741	0.782	-5.5	151#	0.00
28 T	Bromochloromethane	0.598	0.679	-13.5	169#	0.00
29 T	Tetrahydrofuran	0.399	0.422	-5.8	147	0.00
30 C	Chloroform	1.251	1.372	-9.7#	159#	0.00
31 T	Cyclohexane	1.043	1.027	1.5	146	0.00
32 T	1,1,1-Trichloroethane	1.084	1.139	-5.1	156#	0.00
33 S	1,2-Dichloroethane-d4	0.848	0.912	-7.5	169#	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	167#	0.00
35 S	Dibromofluoromethane	0.345	0.310	10.1	155#	0.00
36 T	1,1-Dichloropropene	0.456	0.438	3.9	150#	0.00
37 T	Ethyl Acetate	0.658	0.627	4.7	148	0.00
38 T	Carbon Tetrachloride	0.502	0.454	9.6	147	0.00
39 T	Methylcyclohexane	0.493	0.469	4.9	148	0.00
40 TM	Benzene	1.473	1.404	4.7	151#	0.00
41 T	Methacrylonitrile	0.344	0.329	4.4	149	0.00
42 TM	1,2-Dichloroethane	0.558	0.573	-2.7	168#	0.00
43 T	Isopropyl Acetate	1.022	1.059	-3.6	165#	0.00
44 TM	Trichloroethene	0.348	0.305	12.4	143	0.00
45 C	1,2-Dichloropropane	0.374	0.369	1.3#	156#	0.00
46 T	Dibromomethane	0.280	0.272	2.9	157#	0.00
47 T	Bromodichloromethane	0.564	0.550	2.5	161#	0.00
48 T	Methyl methacrylate	0.460	0.486	-5.7	160#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	149	0.00
50 S	Toluene-d8	1.230	1.125	8.5	153#	0.00
51 T	4-Methyl-2-Pentanone	0.646	0.652	-0.9	158#	0.00
52 CM	Toluene	0.895	0.870	2.8#	151#	0.00
53 T	t-1,3-Dichloropropene	0.571	0.583	-2.1	157#	0.00
54 T	cis-1,3-Dichloropropene	0.590	0.602	-2.0	159#	0.00
55 T	1,1,2-Trichloroethane	0.362	0.350	3.3	159#	0.00
56 T	Ethyl methacrylate	0.552	0.610	-10.5	162#	0.00
57 T	1,3-Dichloropropane	0.627	0.619	1.3	158#	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.344	-15.8	166#	0.00
59 T	2-Hexanone	0.429	0.450	-4.9	151#	0.00
60 T	Dibromochloromethane	0.413	0.401	2.9	155#	0.00
61 T	1,2-Dibromoethane	0.381	0.372	2.4	161#	0.00
62 S	4-Bromofluorobenzene	0.455	0.433	4.8	159#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	167#	0.00
64 T	Tetrachloroethene	0.322	0.260	19.3	136	0.00
65 PM	Chlorobenzene	1.123	1.016	9.5	152#	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.366	4.2	155#	0.00
67 C	Ethyl Benzene	1.848	1.827	1.1#	157#	0.00
68 T	m/p-Xylenes	0.692	0.683	1.3	150#	0.00
69 T	o-Xylene	0.661	0.675	-2.1	156#	0.00
70 T	Styrene	1.112	1.166	-4.9	155#	0.00
71 P	Bromoform	0.308	0.286	7.1	145	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	161#	0.00
73 T	Isopropylbenzene	3.147	3.350	-6.5	159#	0.00
74 T	N-amyl acetate	1.307	1.015	22.3	144	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.195	-0.9	159#	0.00
76 T	1,2,3-Trichloropropane	1.121	1.008	10.1	136	0.00
77 T	Bromobenzene	0.816	0.799	2.1	148	0.00
78 T	n-propylbenzene	3.959	4.232	-6.9	159#	0.00
79 T	2-Chlorotoluene	2.433	2.556	-5.1	162#	0.00
80 T	1,3,5-Trimethylbenzene	2.681	2.883	-7.5	159#	0.00
81 T	trans-1,4-Dichloro-2-butene	0.410	0.397	3.2	158#	0.00
82 T	4-Chlorotoluene	2.533	2.681	-5.8	163#	0.00
83 T	tert-Butylbenzene	2.239	2.406	-7.5	160#	0.00
84 T	1,2,4-Trimethylbenzene	2.738	2.990	-9.2	160#	0.00
85 T	sec-Butylbenzene	3.373	3.586	-6.3	162#	0.00
86 T	p-Isopropyltoluene	2.703	2.988	-10.5	161#	0.00
87 T	1,3-Dichlorobenzene	1.602	1.564	2.4	152#	0.00
88 T	1,4-Dichlorobenzene	1.711	1.570	8.2	148	0.00
89 T	n-Butylbenzene	2.581	2.886	-11.8	168#	0.00
90 T	Hexachloroethane	0.573	0.546	4.7	153#	0.00
91 T	1,2-Dichlorobenzene	1.517	1.497	1.3	153#	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.311	0.276	11.3	147	0.00
93 T	1,2,4-Trichlorobenzene	0.891	0.889	0.2	153#	0.00
94 T	Hexachlorobutadiene	0.331	0.322	2.7	158#	0.00
95 T	Naphthalene	3.158	3.288	-4.1	154#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.894	0.864	3.4	151#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	148	0.00
2 T	Dichlorodifluoromethane	50.000	55.557	-11.1	141	0.00
3 P	Chloromethane	50.000	46.740	6.5	134	0.00
4 C	Vinyl Chloride	50.000	51.052	-2.1#	138	0.00
5 T	Bromomethane	50.000	55.023	-10.0	158	0.00
6 T	Chloroethane	50.000	53.961	-7.9	157	0.00
7 T	Trichlorofluoromethane	50.000	50.212	-0.4	143	0.00
8 T	Diethyl Ether	50.000	56.667	-13.3	162	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.322	-6.6	152	0.00
10 T	Methyl Iodide	50.000	38.232	23.5	114	0.00
11 T	Tert butyl alcohol	250.000	276.178	-10.5	164	0.01
12 CM	1,1-Dichloroethene	50.000	49.187	1.6#	153	0.00
13 T	Acrolein	250.000	228.690	8.5	145	0.00
14 T	Allyl chloride	50.000	51.014	-2.0	156	0.00
15 T	Acrylonitrile	250.000	252.452	-1.0	145	0.00
16 T	Acetone	250.000	304.782	-21.9	186	0.00
17 T	Carbon Disulfide	50.000	44.164	11.7	128	0.00
18 T	Methyl Acetate	50.000	58.620	-17.2	175	0.00
19 T	Methyl tert-butyl Ether	50.000	57.727	-15.5	166	0.00
20 T	Methylene Chloride	50.000	52.881	-5.8	157	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.942	2.1	143	0.00
22 T	Diisopropyl ether	50.000	57.318	-14.6	161	0.00
23 T	Vinyl Acetate	250.000	298.585	-19.4	159	0.00
24 P	1,1-Dichloroethane	50.000	51.200	-2.4	156	0.00
25 T	2-Butanone	250.000	267.432	-7.0	152	0.00
26 T	2,2-Dichloropropane	50.000	56.588	-13.2	166	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.764	-5.5	151	0.00
28 T	Bromochloromethane	50.000	56.761	-13.5	169	0.00
29 T	Tetrahydrofuran	250.000	264.243	-5.7	147	0.00
30 C	Chloroform	50.000	54.800	-9.6#	159	0.00
31 T	Cyclohexane	50.000	49.251	1.5	146	0.00
32 T	1,1,1-Trichloroethane	50.000	52.554	-5.1	156	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.738	-7.5	169	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	167	0.00
35 S	Dibromofluoromethane	50.000	44.903	10.2	155	0.00
36 T	1,1-Dichloropropene	50.000	48.073	3.9	150	0.00
37 T	Ethyl Acetate	50.000	47.639	4.7	148	0.00
38 T	Carbon Tetrachloride	50.000	45.251	9.5	147	0.00
39 T	Methylcyclohexane	50.000	47.484	5.0	148	0.00
40 TM	Benzene	50.000	47.666	4.7	151	0.00
41 T	Methacrylonitrile	50.000	47.844	4.3	149	0.00
42 TM	1,2-Dichloroethane	50.000	51.277	-2.6	168	0.00
43 T	Isopropyl Acetate	50.000	51.848	-3.7	165	0.00
44 TM	Trichloroethene	50.000	43.791	12.4	143	0.00
45 C	1,2-Dichloropropane	50.000	49.245	1.5#	156	0.00
46 T	Dibromomethane	50.000	48.459	3.1	157	0.00
47 T	Bromodichloromethane	50.000	48.711	2.6	161	0.00
48 T	Methyl methacrylate	50.000	52.798	-5.6	160	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087552.D
 Acq On : 14 Aug 2025 09:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1013.618	-1.4	149	0.00
50 S	Toluene-d8	50.000	45.726	8.5	153	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.428	-1.0	158	0.00
52 CM	Toluene	50.000	48.609	2.8#	151	0.00
53 T	t-1,3-Dichloropropene	50.000	51.017	-2.0	157	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.028	-2.1	159	0.00
55 T	1,1,2-Trichloroethane	50.000	48.289	3.4	159	0.00
56 T	Ethyl methacrylate	50.000	50.043	-0.1	162	0.00
57 T	1,3-Dichloropropane	50.000	49.384	1.2	158	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	289.306	-15.7	166	0.00
59 T	2-Hexanone	250.000	262.377	-5.0	151	0.00
60 T	Dibromochloromethane	50.000	48.455	3.1	155	0.00
61 T	1,2-Dibromoethane	50.000	48.875	2.3	161	0.00
62 S	4-Bromofluorobenzene	50.000	47.655	4.7	159	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	167	0.00
64 T	Tetrachloroethene	50.000	40.459	19.1	136	0.00
65 PM	Chlorobenzene	50.000	45.275	9.5	152	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	47.896	4.2	155	0.00
67 C	Ethyl Benzene	50.000	49.442	1.1#	157	0.00
68 T	m/p-Xylenes	100.000	98.699	1.3	150	0.00
69 T	o-Xylene	50.000	51.038	-2.1	156	0.00
70 T	Styrene	50.000	52.445	-4.9	155	0.00
71 P	Bromoform	50.000	46.429	7.1	145	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	161	0.00
73 T	Isopropylbenzene	50.000	53.222	-6.4	159	0.00
74 T	N-amyl acetate	50.000	38.819	22.4	144	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.447	-0.9	159	0.00
76 T	1,2,3-Trichloropropane	50.000	44.960	10.1	136	0.00
77 T	Bromobenzene	50.000	48.942	2.1	148	0.00
78 T	n-propylbenzene	50.000	53.447	-6.9	159	0.00
79 T	2-Chlorotoluene	50.000	52.528	-5.1	162	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.765	-7.5	159	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.472	3.1	158	0.00
82 T	4-Chlorotoluene	50.000	52.913	-5.8	163	0.00
83 T	tert-Butylbenzene	50.000	53.713	-7.4	160	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.601	-9.2	160	0.00
85 T	sec-Butylbenzene	50.000	53.157	-6.3	162	0.00
86 T	p-Isopropyltoluene	50.000	55.271	-10.5	161	0.00
87 T	1,3-Dichlorobenzene	50.000	48.835	2.3	152	0.00
88 T	1,4-Dichlorobenzene	50.000	45.893	8.2	148	0.00
89 T	n-Butylbenzene	50.000	55.909	-11.8	168	0.00
90 T	Hexachloroethane	50.000	47.677	4.6	153	0.00
91 T	1,2-Dichlorobenzene	50.000	49.316	1.4	153	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.360	11.3	147	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.866	0.3	153	0.00
94 T	Hexachlorobutadiene	50.000	48.630	2.7	158	0.00
95 T	Naphthalene	50.000	52.059	-4.1	154	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087552.D
Acq On : 14 Aug 2025 09:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 15:23:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.298	3.4	151	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071625\
Data File : VN087327.D
Acq On : 16 Jul 2025 16:10
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Title : SW846 8260
Last Update : Thu Jul 17 02:56:13 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1839

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	8426	PASS
75	95	30	60	50.8	21643	PASS
95	95	100	100	100.0	42640	PASS
96	95	5	9	6.1	2612	PASS
173	174	0.00	2	0.8	255	PASS
174	95	50	100	70.9	30229	PASS
175	174	5	9	5.1	1538	PASS
176	174	95	101	96.9	29285	PASS
177	176	5	9	7.0	2046	PASS

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.95	354	47.80	181	61.95	2016	74.00	742
37.00	2290	48.10	60	63.05	2010	75.00	2164
37.95	1679	49.00	1995	63.85	261	76.00	181
38.95	954	50.00	8426	64.10	115	76.75	14
39.90	107	51.00	2608	64.95	542	77.00	122
40.95	135	52.95	119	66.85	179	78.10	60
42.90	65	54.90	138	68.00	4227	78.85	1870
43.85	725	55.95	722	68.95	4698	79.85	287
44.95	593	56.95	1299	69.95	494	80.10	209
46.80	115	60.00	486	71.95	381	80.85	1827
47.00	608	60.95	1996	72.95	1978	81.90	375

Average of 12.823 to 12.835 min.: VN087327.D\data.ms

BFB

Modified:subtracted



m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
86.00	53	103.70	156	127.90	178	173.95	30229
86.95	1927	104.00	90	130.00	72	174.90	709
87.90	1486	104.80	81	136.80	89	175.05	1538
90.85	118	105.90	261	140.85	580	175.90	29285
91.10	107	112.80	55	142.90	707	176.90	2046
91.90	1340	115.90	151	144.90	80	281.10	60
93.00	2153	116.60	61	146.90	74		
94.00	5304	117.00	209	147.70	57		
95.00	42640	117.20	66	171.85	576		
95.95	2612	119.00	301	172.70	163		
96.80	56	124.10	52	173.00	255		

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087551.D
Acq On : 14 Aug 2025 09:01
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Title : SW846 8260
Last Update : Thu Jul 17 02:56:13 2025



AutoFind: Scans 1848, 1849, 1850; Background Corrected with Scan 1840

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.8	14484	PASS
75	95	30	60	57.2	34829	PASS
95	95	100	100	100.0	60920	PASS
96	95	5	9	5.6	3407	PASS
173	174	0.00	2	1.3	536	PASS
174	95	50	100	65.2	39723	PASS
175	174	5	9	6.8	2685	PASS
176	174	95	101	98.8	39240	PASS
177	176	5	9	6.2	2451	PASS

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.00	805	49.95	14484	63.00	2963	74.00	1241
36.95	4361	50.95	5265	64.00	607	75.00	3482
37.95	3826	52.10	400	64.80	64	76.00	321
38.95	1459	53.80	50	65.10	174	76.80	37
40.00	321	55.00	367	67.05	225	77.10	310
40.90	158	55.95	1213	67.95	7067	77.80	254
43.95	478	56.95	2372	68.90	7296	78.20	87
44.95	699	58.80	53	69.90	564	78.90	3182
47.00	1337	59.90	752	71.75	448	79.95	646
47.85	509	60.95	3931	72.00	227	80.85	2939
48.95	3305	61.90	3852	72.90	2741	81.85	729

Average of 12.823 to 12.835 min.: VN087551.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
84.90	53	103.60	56	118.80	492	143.70	53
86.95	2577	103.85	295	122.40	52	146.80	70
87.95	2550	104.80	79	127.70	126	147.80	109
90.70	164	105.70	404	128.10	73	148.00	129
90.90	275	106.90	186	129.85	282	155.00	182
91.85	2248	114.60	54	133.00	81	156.90	53
92.95	3098	115.85	328	134.85	244	160.80	136
94.00	8132	116.80	285	136.85	122	171.70	369
95.00	60920	116.95	325	140.90	863	172.00	97
95.95	3407	117.80	270	142.85	826	172.50	283
96.80	81	118.10	99	143.10	183	172.95	536

Average of 12.823 to 12.835 min.: VN087551.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
173.90	39723						
174.85	2685						
175.90	39240						
176.90	2451						
191.00	51						
206.75	119						

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VN0814WBL01	SDG No.:	Q2842
Lab Sample ID:	VN0814WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087554.D	1	08/14/25 10:32	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	44.2		70 (75) - 130 (124)	88%	SPK: 50
2037-26-5	Toluene-d8	44.2		70 (86) - 130 (113)	88%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.7		70 (77) - 130 (121)	87%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	284000	8.206			
540-36-3	1,4-Difluorobenzene	560000	9.083			
3114-55-4	Chlorobenzene-d5	504000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	235000	13.771			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087554.D
Acq On : 14 Aug 2025 10:32
Operator : JC\MD
Sample : VN0814WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	284398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	560321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	503771	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	234765	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	255137	52.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.740%
35) Dibromofluoromethane	8.153	113	170840	44.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.400%
50) Toluene-d8	10.547	98	608969	44.169	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.340%
62) 4-Bromofluorobenzene	12.829	95	222642	43.709	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.420%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087554.D
Acq On : 14 Aug 2025 10:32
Operator : JC\MD
Sample : VN0814WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 15:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1106
Delta R.T. -0.006 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01


Tgt Ion:168 Resp: 284398
Ion Ratio Lower Upper
168 100
99 67.0 47.9 71.9



#33
1,2-Dichloroethane-d4
Concen: 52.871 ug/l
RT: 8.565 min Scan# 1124
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion: 65 Resp: 255137
Ion Ratio Lower Upper
65 100
67 50.2 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 11

Delta R.T. 0.001 min

Lab File: VN087554.D

Acq: 14 Aug 2025 10:32

Instrument :

MSVOA_N

ClientSampleId :

VN0814WBL01





Tgt Ion:114 Resp: 560321

Ion Ratio Lower Upper

114 100

63 23.2 0.0 44.6

88 16.3 0.0 32.8



#35

Dibromofluoromethane

Concen: 44.201 ug/l

RT: 8.153 min Scan# 1054

Delta R.T. 0.006 min

Lab File: VN087554.D

Acq: 14 Aug 2025 10:32




Tgt Ion:113 Resp: 170840


Ion Ratio Lower Upper

113 100

111 103.1 82.5 123.7

192 16.3 13.7 20.5









#50
Toluene-d8
Concen: 44.169 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01


Tgt Ion: 98 Resp: 608969
Ion Ratio Lower Upper
98 100
100 64.2 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 43.709 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion: 95 Resp: 222642
Ion Ratio Lower Upper
95 100
174 66.5 0.0 149.4
176 64.9 0.0 141.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32




Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Tgt Ion	Ratio	Lower	Upper
117	100		
82	60.7	47.4	71.2
119	31.6	23.8	35.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.771 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN087554.D
Acq: 14 Aug 2025 10:32

Tgt Ion	Ratio	Lower	Upper
152	100		
115	64.5	31.1	93.5
150	156.1	0.0	349.0



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	
Client Sample ID:	VN0814WBS01	SDG No.:	Q2842
Lab Sample ID:	VN0814WBS01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087555.D	1	08/14/25 11:07	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	21.4		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	20.6		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.6		0.23	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	21.2		0.19	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.4		0.20	1.00	ug/L
71-43-2	Benzene	20.3		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.090	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.7		0.21	1.00	ug/L
127-18-4	Tetrachloroethene	18.1		0.23	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.6		70 (74) - 130 (125)	115%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.9		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	250000	8.206			
540-36-3	1,4-Difluorobenzene	485000	9.082			
3114-55-4	Chlorobenzene-d5	447000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	222000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087555.D
 Acq On : 14 Aug 2025 11:07
 Operator : JC\MD
 Sample : VN0814WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/18/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	250453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	485218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	446518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	221983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	244982	57.648	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.300%			
35) Dibromofluoromethane	8.153	113	169668	50.692	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.380%			
50) Toluene-d8	10.547	98	605258	50.695	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.380%			
62) 4-Bromofluorobenzene	12.829	95	233458	52.926	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.860%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	60972	22.921	ug/l	92
3) Chloromethane	2.383	50	64803	19.372	ug/l	100
4) Vinyl Chloride	2.542	62	71080	21.381	ug/l	99
5) Bromomethane	2.983	94	38871	22.579	ug/l	96
6) Chloroethane	3.136	64	47553	21.934	ug/l	92
7) Trichlorofluoromethane	3.506	101	103028	20.959	ug/l	98
8) Diethyl Ether	3.965	74	45341	23.778	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.359	101	56020	22.200	ug/l	96
10) Methyl Iodide	4.571	142	33911	17.223	ug/l	89
11) Tert butyl alcohol	5.530	59	92129	114.174	ug/l	97
12) 1,1-Dichloroethene	4.336	96	58932	20.609	ug/l	91
13) Acrolein	4.177	56	52972	81.801	ug/l	95
14) Allyl chloride	5.012	41	106153	20.512	ug/l	91
15) Acrylonitrile	5.706	53	228245	104.238	ug/l	97
16) Acetone	4.430	43	234225	117.551	ug/l	98
17) Carbon Disulfide	4.700	76	152230	17.956	ug/l	97
18) Methyl Acetate	5.024	43	120344	24.040	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	247283	23.461	ug/l	98
20) Methylene Chloride	5.259	84	73232	21.375	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	62532	19.394	ug/l	92
22) Diisopropyl ether	6.659	45	247597	22.809	ug/l	95
23) Vinyl Acetate	6.588	43	1129441	118.964	ug/l	100
24) 1,1-Dichloroethane	6.559	63	135050	21.564	ug/l	97
25) 2-Butanone	7.477	43	332813	108.103	ug/l	95
26) 2,2-Dichloropropane	7.477	77	115630	23.748	ug/l	97
27) cis-1,2-Dichloroethene	7.477	96	78759	21.217	ug/l	94
28) Bromochloromethane	7.800	49	75409	25.159	ug/l	95
29) Tetrahydrofuran	7.829	42	208941	104.471	ug/l	98
30) Chloroform	7.953	83	140698	22.445	ug/l	99
31) Cyclohexane	8.241	56	111696	21.380	ug/l	97
32) 1,1,1-Trichloroethane	8.153	97	115982	21.362	ug/l	96
36) 1,1-Dichloropropene	8.353	75	90752	20.523	ug/l	97
37) Ethyl Acetate	7.547	43	131078	20.524	ug/l	98
38) Carbon Tetrachloride	8.347	117	93434	19.181	ug/l	96
39) Methylcyclohexane	9.582	83	96372	20.130	ug/l	93
40) Benzene	8.588	78	289566	20.261	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087555.D
 Acq On : 14 Aug 2025 11:07
 Operator : JC\MD
 Sample : VN0814WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/18/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	66759	19.992	ug/l	95
42) 1,2-Dichloroethane	8.659	62	116694	21.531	ug/l	97
43) Isopropyl Acetate	8.676	43	215090	21.696	ug/l	98
44) Trichloroethene	9.335	130	64743	19.172	ug/l	82
45) 1,2-Dichloropropane	9.606	63	74233	20.442	ug/l	94
46) Dibromomethane	9.688	93	56540	20.795	ug/l	94
47) Bromodichloromethane	9.871	83	112437	20.530	ug/l	99
48) Methyl methacrylate	9.665	41	98225	22.008	ug/l	95
49) 1,4-Dioxane	9.688	88	30568	447.176	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	661207	105.450	ug/l	98
52) Toluene	10.612	92	177586	20.443	ug/l	97
53) t-1,3-Dichloropropene	10.818	75	116338	20.990	ug/l	99
54) cis-1,3-Dichloropropene	10.294	75	116909	20.420	ug/l #	82
55) 1,1,2-Trichloroethane	10.994	97	72631	20.652	ug/l	92
56) Ethyl methacrylate	10.859	69	120174	21.091	ug/l #	93
57) 1,3-Dichloropropane	11.141	76	128302	21.100	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	328582	113.893	ug/l	98
59) 2-Hexanone	11.182	43	431325	103.681	ug/l	98
60) Dibromochloromethane	11.335	129	76256	19.012	ug/l	98
61) 1,2-Dibromoethane	11.447	107	75775	20.492	ug/l	95
64) Tetrachloroethene	11.088	164	51988	18.090	ug/l	94
65) Chlorobenzene	11.870	112	192917	19.244	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.941	131	65609	19.247	ug/l	97
67) Ethyl Benzene	11.947	91	336124	20.367	ug/l	98
68) m/p-Xylenes	12.053	106	249595	40.389	ug/l	94
69) o-Xylene	12.376	106	119950	20.320	ug/l	93
70) Styrene	12.394	104	205053	20.649	ug/l	96
71) Bromoform	12.559	173	48296	17.538	ug/l #	99
73) Isopropylbenzene	12.676	105	314281	22.495	ug/l	97
74) N-amyl acetate	12.535	43	108512m	18.694	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.923	83	115038	21.882	ug/l	99
76) 1,2,3-Trichloropropane	12.976	75	100267m	20.143	ug/l	
77) Bromobenzene	12.959	156	73544	20.297	ug/l	91
78) n-propylbenzene	13.017	91	389520	22.160	ug/l	99
79) 2-Chlorotoluene	13.106	91	241476	22.353	ug/l	94
80) 1,3,5-Trimethylbenzene	13.153	105	262222	22.029	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.723	75	32197	17.698	ug/l	91
82) 4-Chlorotoluene	13.200	91	250722	22.292	ug/l	95
83) tert-Butylbenzene	13.417	119	222918	22.422	ug/l	96
84) 1,2,4-Trimethylbenzene	13.464	105	269488	22.169	ug/l	99
85) sec-Butylbenzene	13.594	105	329434	21.998	ug/l	99
86) p-Isopropyltoluene	13.706	119	272080	22.671	ug/l	97
87) 1,3-Dichlorobenzene	13.711	146	145480	20.458	ug/l	98
88) 1,4-Dichlorobenzene	13.794	146	150121	19.766	ug/l	97
89) n-Butylbenzene	14.035	91	270370	23.593	ug/l	98
90) Hexachloroethane	14.311	117	49037	19.285	ug/l	96
91) 1,2-Dichlorobenzene	14.088	146	138916	20.620	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	27318	19.792	ug/l	92
93) 1,2,4-Trichlorobenzene	15.364	180	83245	21.036	ug/l	99
94) Hexachlorobutadiene	15.476	225	31065	21.127	ug/l	99
95) Naphthalene	15.617	128	298554	21.296	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	78959	19.891	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087555.D
Acq On : 14 Aug 2025 11:07
Operator : JC\MD
Sample : VN0814WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087555.D
Acq On : 14 Aug 2025 11:07
Operator : JC\MD
Sample : VN0814WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 15:23:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MS	SDG No.:	Q2842
Lab Sample ID:	Q2842-04MS	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087569.D	100	08/14/25 16:27	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5200		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5300		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8100		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5400		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5500		22.0	100	ug/L
79-01-6	Trichloroethene	10000		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.8		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	47.9		70 (75) - 130 (124)	96%	SPK: 50
2037-26-5	Toluene-d8	47.9		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	238000	8.206			
540-36-3	1,4-Difluorobenzene	461000	9.083			
3114-55-4	Chlorobenzene-d5	435000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	232000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087569.D
 Acq On : 14 Aug 2025 16:27
 Operator : JC\MD
 Sample : Q2842-04MS 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MS

Quant Time: Aug 14 19:59:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	238067	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.083	114	461160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	435202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	231778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	217494	53.842	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.680%			
35) Dibromofluoromethane	8.153	113	152403	47.909	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.820%			
50) Toluene-d8	10.547	98	543343	47.883	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.760%			
62) 4-Bromofluorobenzene	12.829	95	216897	51.737	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.480%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	143419	56.720	ug/l	100
3) Chloromethane	2.383	50	152078	47.827	ug/l	100
4) Vinyl Chloride	2.542	62	169468	53.629	ug/l	95
5) Bromomethane	2.971	94	89339	54.595	ug/l	96
6) Chloroethane	3.136	64	114275	55.452	ug/l	97
7) Trichlorofluoromethane	3.501	101	253745	54.304	ug/l	97
8) Diethyl Ether	3.965	74	112077	61.834	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.359	101	138606	57.785	ug/l	95
10) Methyl Iodide	4.577	142	91710	38.610	ug/l	91
11) Tert butyl alcohol	5.536	59	219902	286.700	ug/l	100
12) 1,1-Dichloroethene	4.330	96	140041	51.521	ug/l	99
13) Acrolein	4.177	56	135355	219.895	ug/l	99
14) Allyl chloride	5.012	41	254609	51.759	ug/l	94
15) Acrylonitrile	5.712	53	562090	270.058	ug/l	98
16) Acetone	4.430	43	530238	279.957	ug/l	98
17) Carbon Disulfide	4.700	76	356757	44.270	ug/l #	93
18) Methyl Acetate	5.018	43	302650	63.602	ug/l	100
19) Methyl tert-butyl Ether	5.789	73	601237	60.011	ug/l	96
20) Methylene Chloride	5.265	84	169920	53.111	ug/l	99
21) trans-1,2-Dichloroethene	5.771	96	153170	49.977	ug/l	93
22) Diisopropyl ether	6.659	45	612749	59.384	ug/l	97
23) Vinyl Acetate	6.594	43	2775483	307.552	ug/l	99
24) 1,1-Dichloroethane	6.553	63	318408	53.488	ug/l	98
25) 2-Butanone	7.471	43	850355	290.579	ug/l	98
26) 2,2-Dichloropropane	7.471	77	249304	53.865	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	284580	80.651	ug/l	97
28) Bromochloromethane	7.800	49	143451	50.351	ug/l	93
29) Tetrahydrofuran	7.830	42	541901	285.049	ug/l	99
30) Chloroform	7.953	83	341066	57.240	ug/l	99
31) Cyclohexane	8.241	56	253936	51.134	ug/l	96
32) 1,1,1-Trichloroethane	8.153	97	279470	54.153	ug/l	97
36) 1,1-Dichloropropene	8.353	75	214559	51.052	ug/l	98
37) Ethyl Acetate	7.553	43	307770	50.705	ug/l	98
38) Carbon Tetrachloride	8.347	117	228035	49.255	ug/l	93
39) Methylcyclohexane	9.583	83	237505	52.198	ug/l	96
40) Benzene	8.588	78	696316	51.262	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087569.D
 Acq On : 14 Aug 2025 16:27
 Operator : JC\MD
 Sample : Q2842-04MS 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MS

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/15/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 19:59:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	175869	55.413	ug/l	98
42) 1,2-Dichloroethane	8.653	62	283996	55.133	ug/l	97
43) Isopropyl Acetate	8.677	43	532619	56.527	ug/l	97
44) Trichloroethene	9.335	130	321363	100.127	ug/l	88
45) 1,2-Dichloropropane	9.606	63	182390	52.845	ug/l	94
46) Dibromomethane	9.694	93	137378	53.162	ug/l	95
47) Bromodichloromethane	9.871	83	277559	53.324	ug/l	96
48) Methyl methacrylate	9.665	41	245497	57.875	ug/l	93
49) 1,4-Dioxane	9.688	88	69859	1075.274	ug/l #	97
51) 4-Methyl-2-Pentanone	10.430	43	1677266	281.447	ug/l	98
52) Toluene	10.612	92	435684	52.770	ug/l	96
53) t-1,3-Dichloropropene	10.818	75	283224	53.765	ug/l	95
54) cis-1,3-Dichloropropene	10.294	75	296099	54.416	ug/l #	88
55) 1,1,2-Trichloroethane	11.000	97	176388	52.770	ug/l	94
56) Ethyl methacrylate	10.859	69	305796	54.213	ug/l	93
57) 1,3-Dichloropropane	11.147	76	313150	54.186	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	584151	213.042	ug/l	99
59) 2-Hexanone	11.182	43	1146887	290.068	ug/l	100
60) Dibromochloromethane	11.341	129	195208	51.209	ug/l	99
61) 1,2-Dibromoethane	11.447	107	185145	52.680	ug/l	98
64) Tetrachloroethene	11.088	164	120472	43.011	ug/l	96
65) Chlorobenzene	11.871	112	472723	48.382	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.941	131	165027	49.672	ug/l	100
67) Ethyl Benzene	11.941	91	844688	52.514	ug/l	99
68) m/p-Xylenes	12.053	106	627919	104.251	ug/l	93
69) o-Xylene	12.376	106	307518	53.449	ug/l	94
70) Styrene	12.394	104	539497	55.741	ug/l	98
71) Bromoform	12.559	173	129055	48.082	ug/l #	99
73) Isopropylbenzene	12.676	105	797605	54.677	ug/l	98
74) N-amyl acetate	12.500	43	204124	33.679	ug/l #	93
75) 1,1,2,2-Tetrachloroethane	12.918	83	285137	51.946	ug/l	99
76) 1,2,3-Trichloropropane	12.976	75	244697m	47.080	ug/l	
77) Bromobenzene	12.959	156	183370	48.469	ug/l	91
78) n-propylbenzene	13.018	91	990768	53.982	ug/l	97
79) 2-Chlorotoluene	13.106	91	600287	53.218	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	686579	55.240	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.718	75	87284	45.950	ug/l	93
82) 4-Chlorotoluene	13.200	91	625581	53.269	ug/l	97
83) tert-Butylbenzene	13.418	119	576103	55.497	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	715753	56.391	ug/l	95
85) sec-Butylbenzene	13.594	105	866181	55.396	ug/l	99
86) p-Isopropyltoluene	13.706	119	703516	56.143	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	370720	49.929	ug/l	99
88) 1,4-Dichlorobenzene	13.794	146	374267	47.195	ug/l	97
89) n-Butylbenzene	14.035	91	697153	58.265	ug/l	98
90) Hexachloroethane	14.312	117	126427	47.619	ug/l	93
91) 1,2-Dichlorobenzene	14.082	146	361117	51.338	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	69760	48.406	ug/l	95
93) 1,2,4-Trichlorobenzene	15.370	180	212082	51.328	ug/l	99
94) Hexachlorobutadiene	15.476	225	79123	51.536	ug/l	97
95) Naphthalene	15.617	128	789585	53.941	ug/l	99
96) 1,2,3-Trichlorobenzene	15.817	180	206242	49.760	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087569.D
Acq On : 14 Aug 2025 16:27
Operator : JC\MD
Sample : Q2842-04MS 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 19:59:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\  
Data File : VN087569.D  
Acq On    : 14 Aug 2025  16:27  
Operator  : JC\MD  
Sample    : Q2842-04MS 100X  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 19   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MS

Manual Integrations APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 19:59:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration



Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/12/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/12/25
Client Sample ID:	MW-17B-55.5-081225MSD	SDG No.:	Q2842
Lab Sample ID:	Q2842-05MSD	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087570.D	100	08/14/25 16:49	VN081425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5400		26.0	100	ug/L
75-35-4	1,1-Dichloroethene	5300		23.0	100	ug/L
75-34-3	1,1-Dichloroethane	5400		23.0	100	ug/L
156-59-2	cis-1,2-Dichloroethene	8000		19.0	100	ug/L
71-55-6	1,1,1-Trichloroethane	5500		20.0	100	ug/L
71-43-2	Benzene	5100		15.0	100	ug/L
107-06-2	1,2-Dichloroethane	5300		22.0	100	ug/L
79-01-6	Trichloroethene	9500		9.30	100	ug/L
79-00-5	1,1,2-Trichloroethane	5300		21.0	100	ug/L
127-18-4	Tetrachloroethene	4300		23.0	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	46.7		70 (75) - 130 (124)	93%	SPK: 50
2037-26-5	Toluene-d8	48.2		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	252000	8.212			
540-36-3	1,4-Difluorobenzene	499000	9.082			
3114-55-4	Chlorobenzene-d5	460000	11.847			
3855-82-1	1,4-Dichlorobenzene-d4	240000	13.77			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of OC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087570.D
 Acq On : 14 Aug 2025 16:49
 Operator : JC\MD
 Sample : Q2842-05MSD 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MSD

Quant Time: Aug 14 20:00:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	251785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	498703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	460182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	240017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	233131	54.569	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.140%			
35) Dibromofluoromethane	8.153	113	160785	46.739	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.480%			
50) Toluene-d8	10.547	98	591485	48.202	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.400%			
62) 4-Bromofluorobenzene	12.829	95	228797	50.467	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.940%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	160380	59.972	ug/l	95
3) Chloromethane	2.383	50	164845	49.018	ug/l	95
4) Vinyl Chloride	2.542	62	181197	54.217	ug/l	94
5) Bromomethane	2.971	94	111058	64.170	ug/l	100
6) Chloroethane	3.130	64	124775	57.248	ug/l	97
7) Trichlorofluoromethane	3.506	101	269630	54.560	ug/l	92
8) Diethyl Ether	3.965	74	116948	61.006	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	146884	57.900	ug/l	97
10) Methyl Iodide	4.583	142	119433	46.239	ug/l	95
11) Tert butyl alcohol	5.536	59	241582	297.805	ug/l	99
12) 1,1-Dichloroethene	4.342	96	151233	52.607	ug/l	98
13) Acrolein	4.177	56	121388	186.460	ug/l	99
14) Allyl chloride	5.018	41	274012	52.668	ug/l	93
15) Acrylonitrile	5.706	53	596577	271.011	ug/l	97
16) Acetone	4.424	43	565138	282.126	ug/l	97
17) Carbon Disulfide	4.700	76	383063	44.945	ug/l	97
18) Methyl Acetate	5.024	43	323417	64.264	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	630203	59.475	ug/l	96
20) Methylene Chloride	5.265	84	180635	53.387	ug/l #	89
21) trans-1,2-Dichloroethene	5.777	96	165046	50.918	ug/l	96
22) Diisopropyl ether	6.665	45	653071	59.843	ug/l	98
23) Vinyl Acetate	6.594	43	2963851	310.532	ug/l	99
24) 1,1-Dichloroethane	6.559	63	337459	53.599	ug/l	99
25) 2-Butanone	7.477	43	880575	284.512	ug/l	97
26) 2,2-Dichloropropane	7.477	77	251846	51.450	ug/l	98
27) cis-1,2-Dichloroethene	7.471	96	296951	79.572	ug/l	98
28) Bromochloromethane	7.794	49	161291	53.528	ug/l	92
29) Tetrahydrofuran	7.830	42	579208	288.074	ug/l	99
30) Chloroform	7.953	83	365342	57.974	ug/l	98
31) Cyclohexane	8.241	56	269452	51.302	ug/l	99
32) 1,1,1-Trichloroethane	8.153	97	301283	55.199	ug/l	97
36) 1,1-Dichloropropene	8.359	75	232570	51.172	ug/l	98
37) Ethyl Acetate	7.547	43	344525	52.488	ug/l	98
38) Carbon Tetrachloride	8.347	117	240818	48.100	ug/l	96
39) Methylcyclohexane	9.582	83	251368	51.086	ug/l	95
40) Benzene	8.588	78	748859	50.980	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
 Data File : VN087570.D
 Acq On : 14 Aug 2025 16:49
 Operator : JC\MD
 Sample : Q2842-05MSD 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-17B-55.5-081225MSD

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/15/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 20:00:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.765	41	188160	54.823	ug/l	97
42) 1,2-Dichloroethane	8.659	62	295609	53.067	ug/l	97
43) Isopropyl Acetate	8.677	43	557773	54.740	ug/l	99
44) Trichloroethene	9.335	130	330475	95.214	ug/l	96
45) 1,2-Dichloropropane	9.606	63	191212	51.231	ug/l	98
46) Dibromomethane	9.694	93	141089	50.488	ug/l	95
47) Bromodichloromethane	9.871	83	288056	51.174	ug/l	98
48) Methyl methacrylate	9.665	41	268246	58.477	ug/l	94
49) 1,4-Dioxane	9.682	88	74858	1065.478	ug/l #	100
51) 4-Methyl-2-Pentanone	10.429	43	1774583	275.360	ug/l	97
52) Toluene	10.612	92	460377	51.563	ug/l	96
53) t-1,3-Dichloropropene	10.818	75	308221	54.105	ug/l	96
54) cis-1,3-Dichloropropene	10.294	75	308333	52.399	ug/l #	87
55) 1,1,2-Trichloroethane	11.000	97	190288	52.643	ug/l	92
56) Ethyl methacrylate	10.859	69	326690	53.587	ug/l #	95
57) 1,3-Dichloropropane	11.147	76	329546	52.730	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	412625	139.157	ug/l	99
59) 2-Hexanone	11.182	43	1237547	289.435	ug/l	98
60) Dibromochloromethane	11.341	129	205718	49.903	ug/l	98
61) 1,2-Dibromoethane	11.447	107	193708	50.967	ug/l	98
64) Tetrachloroethene	11.082	164	127664	43.104	ug/l	94
65) Chlorobenzene	11.871	112	500916	48.485	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.941	131	176284	50.180	ug/l	99
67) Ethyl Benzene	11.941	91	898707	52.840	ug/l	99
68) m/p-Xylenes	12.053	106	662341	103.996	ug/l	93
69) o-Xylene	12.376	106	328068	53.926	ug/l	96
70) Styrene	12.394	104	567805	55.481	ug/l	98
71) Bromoform	12.559	173	133281	46.961	ug/l #	95
73) Isopropylbenzene	12.676	105	839673	55.585	ug/l	99
74) N-amyl acetate	12.500	43	234757	37.404	ug/l #	94
75) 1,1,2,2-Tetrachloroethane	12.918	83	304955	53.650	ug/l	99
76) 1,2,3-Trichloropropane	12.970	75	297389m	55.254	ug/l	
77) Bromobenzene	12.959	156	197712	50.466	ug/l	94
78) n-propylbenzene	13.018	91	1061543	55.853	ug/l	98
79) 2-Chlorotoluene	13.106	91	637513	54.578	ug/l	96
80) 1,3,5-Trimethylbenzene	13.153	105	725164	56.342	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.718	75	93157	47.359	ug/l	94
82) 4-Chlorotoluene	13.200	91	661408	54.387	ug/l	97
83) tert-Butylbenzene	13.418	119	601617	55.966	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	747848	56.897	ug/l	96
85) sec-Butylbenzene	13.594	105	905664	55.933	ug/l	98
86) p-Isopropyltoluene	13.706	119	748208	57.660	ug/l	97
87) 1,3-Dichlorobenzene	13.712	146	393523	51.181	ug/l	100
88) 1,4-Dichlorobenzene	13.788	146	395731	48.189	ug/l	98
89) n-Butylbenzene	14.035	91	730169	58.929	ug/l	98
90) Hexachloroethane	14.312	117	136983	49.824	ug/l	91
91) 1,2-Dichlorobenzene	14.088	146	372655	51.160	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	74799	50.121	ug/l	97
93) 1,2,4-Trichlorobenzene	15.370	180	220758	51.594	ug/l	98
94) Hexachlorobutadiene	15.476	225	77907	49.002	ug/l	98
95) Naphthalene	15.617	128	847533	55.913	ug/l	100
96) 1,2,3-Trichlorobenzene	15.817	180	217612	50.701	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087570.D
Acq On : 14 Aug 2025 16:49
Operator : JC\MD
Sample : Q2842-05MSD 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 20:00:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081425\
Data File : VN087570.D
Acq On    : 14 Aug 2025  16:49
Operator  : JC\MD
Sample    : Q2842-05MSD 100X
Misc      : 5.0mL/MSVOA_N/WATER
ALS Vial  : 20   Sample Multiplier: 1
```

Instrument :
MSVOA_N
ClientSampleId :
MW-17B-55.5-081225MSD

Manual Integrations

Reviewed By :John Carlone 08/15/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 14 20:00:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration



Manual Integration Report

Sequence:	VN071625	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VN087328.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	1,4-Dichlorobenzene	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Acetone	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Allyl chloride	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	Methyl tert-butyl Ether	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC001	VN087328.D	N-amyl acetate	MMDadoda	7/17/2025 8:19:52 AM	Sam	7/17/2025 8:24:59 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	Methyl Iodide	MMDadoda	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC005	VN087329.D	N-amyl acetate	MMDadoda	7/17/2025 8:19:53 AM	Sam	7/17/2025 8:24:55 AM	Peak Integrated by Software
VSTDICC020	VN087330.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:54 AM	Sam	7/17/2025 8:24:56 AM	Peak Integrated by Software
VSTDICC020	VN087330.D	N-amyl acetate	MMDadoda	7/17/2025 8:19:54 AM	Sam	7/17/2025 8:24:56 AM	Peak Integrated by Software
VSTDICCC050	VN087331.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:56 AM	Sam	7/17/2025 8:25:01 AM	Peak Integrated by Software
VSTDICCC050	VN087331.D	N-amyl acetate	MMDadoda	7/17/2025 8:19:56 AM	Sam	7/17/2025 8:25:01 AM	Peak Integrated by Software

Manual Integration Report

Sequence:

VN071625

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN087332.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:58 AM	Sam	7/17/2025 8:25:00 AM	Peak Integrated by Software
VSTDICC150	VN087333.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:19:59 AM	Sam	7/17/2025 8:25:02 AM	Peak Integrated by Software
VSTDICV050	VN087335.D	1,2,3-Trichloropropane	MMDadoda	7/17/2025 8:20:01 AM	Sam	7/17/2025 8:25:03 AM	Peak Integrated by Software
VSTDICV050	VN087335.D	N-amyl acetate	MMDadoda	7/17/2025 8:20:01 AM	Sam	7/17/2025 8:25:03 AM	Peak Integrated by Software

Manual Integration Report

Sequence:	VN081425	Instrument	MSVOA_n
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN087552.D	1,2,3-Trichloropropane	JOHN	8/14/2025 4:35:55 PM	MMDadoda	8/19/2025 1:44:04 AM	Peak Integrated by Software
VN0814WBS01	VN087555.D	1,2,3-Trichloropropane	JOHN	8/18/2025 8:11:28 AM	MMDadoda	8/19/2025 1:44:07 AM	Peak Integrated by Software
VN0814WBS01	VN087555.D	N-amyl acetate	JOHN	8/18/2025 8:11:28 AM	MMDadoda	8/19/2025 1:44:07 AM	Peak Integrated by Software
Q2842-04MS	VN087569.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:03 AM	MMDadoda	8/19/2025 1:44:16 AM	Peak Integrated by Software
Q2842-05MSD	VN087570.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:08 AM	MMDadoda	8/19/2025 1:44:19 AM	Peak Integrated by Software
VSTDCCC050	VN087571.D	1,2,3-Trichloropropane	JOHN	8/15/2025 8:40:12 AM	MMDadoda	8/19/2025 1:44:23 AM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN071625

Review By	Mahesh Dadoda	Review On	7/17/2025 8:20:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:25:10 AM
SubDirectory	VN071625	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134794		
Initial Calibration Stds	VP134795,VP134796,VP134797,VP134798,VP134799,VP134800		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP134801		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087327.D	16 Jul 2025 16:10	JC\MD	Ok
2	VSTDIC001	VN087328.D	16 Jul 2025 17:05	JC\MD	Ok,M
3	VSTDIC005	VN087329.D	16 Jul 2025 17:27	JC\MD	Ok,M
4	VSTDIC020	VN087330.D	16 Jul 2025 17:49	JC\MD	Ok,M
5	VSTDIC050	VN087331.D	16 Jul 2025 18:11	JC\MD	Ok,M
6	VSTDIC100	VN087332.D	16 Jul 2025 18:32	JC\MD	Ok,M
7	VSTDIC150	VN087333.D	16 Jul 2025 18:54	JC\MD	Ok,M
8	IBLK	VN087334.D	16 Jul 2025 19:16	JC\MD	Ok
9	VSTDICV050	VN087335.D	16 Jul 2025 19:59	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135139		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN087551.D	14 Aug 2025 09:01	JC\MD	Ok
2	VSTDCCC050	VN087552.D	14 Aug 2025 09:34	JC\MD	Ok,M
3	VN0814MBL01	VN087553.D	14 Aug 2025 10:10	JC\MD	Ok
4	VN0814WBL01	VN087554.D	14 Aug 2025 10:32	JC\MD	Ok
5	VN0814WBS01	VN087555.D	14 Aug 2025 11:07	JC\MD	Ok,M
6	VN0814WBSD01	VN087556.D	14 Aug 2025 11:41	JC\MD	Ok,M
7	Q2832-02	VN087557.D	14 Aug 2025 12:03	JC\MD	Ok
8	Q2832-04	VN087558.D	14 Aug 2025 12:24	JC\MD	Ok,M
9	Q2836-01	VN087559.D	14 Aug 2025 12:46	JC\MD	Ok
10	Q2836-05	VN087560.D	14 Aug 2025 13:08	JC\MD	Ok
11	Q2836-09	VN087561.D	14 Aug 2025 13:30	JC\MD	Ok
12	Q2836-13	VN087562.D	14 Aug 2025 13:52	JC\MD	Ok
13	Q2838-04	VN087563.D	14 Aug 2025 14:14	JC\MD	Ok
14	Q2838-08	VN087564.D	14 Aug 2025 14:36	JC\MD	Ok
15	Q2842-08	VN087565.D	14 Aug 2025 14:58	JC\MD	Ok
16	Q2842-01	VN087566.D	14 Aug 2025 15:20	JC\MD	Ok
17	Q2842-02	VN087567.D	14 Aug 2025 15:42	JC\MD	Ok
18	Q2842-03	VN087568.D	14 Aug 2025 16:04	JC\MD	Ok
19	Q2842-04MS	VN087569.D	14 Aug 2025 16:27	JC\MD	Ok,M
20	Q2842-05MSD	VN087570.D	14 Aug 2025 16:49	JC\MD	Ok,M
21	VSTDCCC050	VN087571.D	14 Aug 2025 17:11	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN071625

Review By	Maresh Dadoda	Review On	7/17/2025 8:20:05 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:25:10 AM
SubDirectory	VN071625	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134794 VP134795,VP134796,VP134797,VP134798,VP134799,VP134800 VP134801

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087327.D	16 Jul 2025 16:10		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN087328.D	16 Jul 2025 17:05		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN087329.D	16 Jul 2025 17:27	% d fail for com.#10 in 5 ppb	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN087330.D	16 Jul 2025 17:49	LR- 10,20	JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN087331.D	16 Jul 2025 18:11	QR- 56	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN087332.D	16 Jul 2025 18:32		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN087333.D	16 Jul 2025 18:54		JC\MD	Ok,M
8	IBLK	IBLK	VN087334.D	16 Jul 2025 19:16		JC\MD	Ok
9	VSTDICV050	ICVVN071625	VN087335.D	16 Jul 2025 19:59		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135139
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087551.D	14 Aug 2025 09:01		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087552.D	14 Aug 2025 09:34	pH#Lot#V12668	JC\MD	Ok,M
3	VN0814MBL01	VN0814MBL01	VN087553.D	14 Aug 2025 10:10		JC\MD	Ok
4	VN0814WBL01	VN0814WBL01	VN087554.D	14 Aug 2025 10:32		JC\MD	Ok
5	VN0814WBS01	VN0814WBS01	VN087555.D	14 Aug 2025 11:07		JC\MD	Ok,M
6	VN0814WBSD01	VN0814WBSD01	VN087556.D	14 Aug 2025 11:41		JC\MD	Ok,M
7	Q2832-02	TG-S01	VN087557.D	14 Aug 2025 12:03	vial B pH#5.0	JC\MD	Ok
8	Q2832-04	TG-S02	VN087558.D	14 Aug 2025 12:24	vial B pH#5.0	JC\MD	Ok,M
9	Q2836-01	WC-A2-15-G	VN087559.D	14 Aug 2025 12:46	vial B pH#5.0	JC\MD	Ok
10	Q2836-05	WC-A2-16-G	VN087560.D	14 Aug 2025 13:08	vial B pH#5.0	JC\MD	Ok
11	Q2836-09	WC-A2-17-G	VN087561.D	14 Aug 2025 13:30	vial B pH#5.0	JC\MD	Ok
12	Q2836-13	WC-A5-02-G	VN087562.D	14 Aug 2025 13:52	vial B pH#5.0	JC\MD	Ok
13	Q2838-04	TP-11	VN087563.D	14 Aug 2025 14:14	vial B pH#5.0	JC\MD	Ok
14	Q2838-08	TP-10	VN087564.D	14 Aug 2025 14:36	vial B pH#5.0	JC\MD	Ok
15	Q2842-08	TB01-081225	VN087565.D	14 Aug 2025 14:58	vial A pH<2 TB	JC\MD	Ok
16	Q2842-01	RMW-03B-90.4-081220	VN087566.D	14 Aug 2025 15:20	vial A pH<2	JC\MD	Ok
17	Q2842-02	RMW-02B-66.3-081220	VN087567.D	14 Aug 2025 15:42	vial A pH<2	JC\MD	Ok
18	Q2842-03	MW-17B-55.5-0812202	VN087568.D	14 Aug 2025 16:04	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN081425

Review By	John Carlone	Review On	8/15/2025 8:42:25 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:44:35 AM
SubDirectory	VN081425	HP Acquire Method	HP Processing Method 82N071625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135139		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135140,VP135141		

19	Q2842-04MS	MW-17B-55.5-081225M	VN087569.D	14 Aug 2025 16:27	vial A pH<2	JC\MD	Ok,M
20	Q2842-05MSD	MW-17B-55.5-081225M	VN087570.D	14 Aug 2025 16:49	vial A pH<2	JC\MD	Ok,M
21	VSTDCCC050	VSTDCCC050EC	VN087571.D	14 Aug 2025 17:11		JC\MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : Q2842

Test : VOCMS Group3

Prepbatch ID :

Sequence ID/Qc Batch ID: VN081425,

Standard ID :

VP134142,VP134149,VP134933,VP134956,VP134957,VP135059,VP135129,VP135139,VP135140,VP135141,

Chemical ID :

V13391,V14290,V14444,V14446,V14507,V14508,V14529,V14530,V14625,V14626,V14636,V14637,V14638,V14639,V14668,V14671,V14673,V14675,V14702,V14705,V14716,V14745,V14751,V14806,V14807,V14843,V14906,V14929,V15057,V15058,V15059,W3112,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP134142	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025
<u>FROM</u> 1.00000ml of V14668 + 1.00000ml of V14671 + 1.00000ml of V14673 + 1.00000ml of V14675 + 16.00000ml of V14929 = Final Quantity: 20.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP134149	06/06/2025	12/06/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 06/10/2025
<u>FROM</u> 1.00000ml of V14636 + 1.00000ml of V14637 + 1.00000ml of V14638 + 1.00000ml of V14639 + 46.00000ml of V14929 = Final Quantity: 50.000 ml								

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP134933	07/29/2025	01/29/2026	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/19/2025

FROM 0.40000ml of V14906 + 24.60000ml of V14625 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP134956	08/01/2025	11/09/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/06/2025

FROM 0.25000ml of V14290 + 24.75000ml of V14626 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP134957	08/01/2025	11/22/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/06/2025

FROM 0.50000ml of V13391 + 49.50000ml of V14625 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP135059	08/11/2025	09/19/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/19/2025

FROM 0.40000ml of V14843 + 1.00000ml of V14444 + 1.00000ml of V14446 + 1.00000ml of V14507 + 1.00000ml of V14508 + 1.00000ml of V14529 + 1.00000ml of V14530 + 1.00000ml of V14705 + 1.00000ml of V14745 + 1.00000ml of V14751 + 1.00000ml of V14806 + 1.00000ml of V14807 + 1.50000ml of V14702 + 1.50000ml of V14716 + 10.60000ml of V14625 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP135129	08/14/2025	09/13/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda
								08/19/2025

FROM 1.00000ml of V15059 + 1.50000ml of V15057 + 1.50000ml of V15058 + 21.00000ml of V14625 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP135139	08/14/2025	08/15/2025	John Carlone	None	None	Maresh Dadoda
								08/19/2025

FROM 39.98400ml of W3112 + 0.01600ml of VP134957 = Final Quantity: 40.000 ml



VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP135140	08/14/2025	08/15/2025	John Carlone	None	None	Mahesh Dadoda 08/19/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP134142 + 0.00500ml of VP134933 + 0.00800ml of VP134956 + 0.01250ml of VP134149 + 0.01250ml of VP135059 + 0.01250ml of VP135129 = Final Quantity: 40.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP135141	08/14/2025	08/15/2025	John Carlone	None	None	Mahesh Dadoda 08/19/2025
<u>FROM</u>	39.94450ml of W3112 + 0.00500ml of VP134142 + 0.00500ml of VP134933 + 0.00800ml of VP134956 + 0.01250ml of VP134149 + 0.01250ml of VP135059 + 0.01250ml of VP135129 = Final Quantity: 40.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	12/12/2025	12/12/2024 / SAM	04/15/2024 / SAM	V14290

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	08/08/2025 / SAM	08/15/2024 / SAM	V14444

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	08/08/2025 / SAM	08/15/2024 / SAM	V14446

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	02/08/2026	08/08/2025 / SAM	09/17/2024 / SAM	V14507

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	02/08/2026	08/08/2025 / SAM	09/17/2024 / SAM	V14508

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	02/08/2026	08/08/2025 / SAM	09/18/2024 / SAM	V14529

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	091724	02/08/2026	08/08/2025 / SAM	09/18/2024 / SAM	V14530

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	01/29/2026	07/29/2025 / SAM	11/26/2024 / SAM	V14625

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	11/09/2025	05/09/2025 / SAM	11/26/2024 / SAM	V14626

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14636

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14637

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14638

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ 2-Chloroethyl vinyl ether	120524	12/06/2025	06/06/2025 / SAM	12/06/2024 / SAM	V14639

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14668

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14671

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0214960	12/06/2025	06/06/2025 / SAM	12/09/2024 / SAM	V14675

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14705

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14716

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14745

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	02/08/2026	08/08/2025 / SAM	12/17/2024 / SAM	V14751

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	02/08/2026	08/08/2025 / SAM	01/08/2025 / SAM	V14806

LOTS

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0220471	02/08/2026	08/08/2025 / SAM	01/08/2025 / SAM	V14807

LOTS

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0217535	11/12/2025	05/12/2025 / SAM	01/21/2025 / SAM	V14843

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0223904	07/29/2026	07/29/2025 / SAM	03/24/2025 / SAM	V14906

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	24G0262002	12/06/2025	06/06/2025 / SAM	05/09/2025 / SAM	V14929

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15057

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15058

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	081325	09/13/2025	08/14/2025 / SAM	08/14/2025 / SAM	V15059

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 23I0762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Rentes	021624
DATE		

Compound	(KMe) Part Number	Lot Number	DR Factor	Initial Vol. (mL)	Initial Conc. (µg/mL)	Nominal Conc. (µg/mL)	Purity (%)	Purity Uncertainty	Uncertainty Pipette (mL)	Target Weight(g)	Actual Weight(g)	Actual Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
														CAS#	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-6	N/A	N/A
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP9041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3460mg/kg
10. Methacrylonitrile	(0472)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
14. 2-Nitropropane	(0461)	HG002JK	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg
15. Pentachloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18990	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 435g/kg
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 915mg/kg
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	or-rat 1235mg/kg
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg
22. 1,1-Dichloroethane	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 108mg/kg
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg
45. Benzene	35162	050823	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4894mg/kg
46. Bromobenzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	N/A	N/A
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
50. Naphthalene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-3	200 ppm	or-rat 5000mg/kg
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-63-6	N/A	or-rat 5g/kg
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-8	N/A	or-rat 5000mg/kg
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-8	N/A	or-rat 2240mg/kg
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73-1	N/A	or-mus 1062mg/kg
64. 1,3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	108-46-7	75 ppm (450mg/m3/8H)	or-rat




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Aryl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethane	18.00
12	Methoxyvinyltoluene/Methyl acrylate/Chloroform	18.49
13	Isobutanol/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethane	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2-Pentopropane	27.07
22	cis-1,2-Dichloroethane	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	31.84
28	1,2-Dibromomethane	33.06
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.28
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water
----------------------------------	---	--	---



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption	ingestion and inhalation.
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing.	Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Ree 03117/24

CERTIFIED WEIGHT REPORT

Part Number: 95317

Lot Number: 021624

Description: Universal VOA Megamix

69 components

Expiration Date: 021627

Recommended Storage: Freezer (0 °C)

Nominal Concentration (µg/mL): 2000

NIST Test ID#: BUTB

Solvent(s): Lot#
Methanol EG359-USQ12Weight(s) shown below were combined and diluted to (mL): 100.0 0.021 Balance Uncertainty
Flask Uncertainty

Formulated By:	Prashant Chauhan	021624
Reviewed By:	Pedro L. Renteria	021624
DATE		

Compound	(K049)	Lot	Dir.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Expanded	SDS Information				
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Actual Conc (µg/mL)	Uncertainty (µg/mL)	(Solvent Safety Info. On Attached pg.)	CASE	OSHA PEL (TWA)	LD50
1. Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2480mg/kg	
2. Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg	
3. Carbon disulfide	(0060)	MKCR8561	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg	
4. cis-1,4-Dichloro-2-butene	(1196)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21056	0.21069	2001.1	8.5	1478-11-5	N/A	N/A	
5. trans-1,4-Dichloro-2-butene	(0486)	MKBP6041V	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20746	2001.7	8.4	110-57-6	N/A	N/A	
6. Diethyl ether	(0153)	IK18CAS000C	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	N/A	N/A	
7. Ethyl methacrylate	(0381)	06128PX	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	N/A	N/A	
8. Iodomethane	(0489)	SHBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 14800mg/kg	
9. 2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 2460mg/kg	
10. Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	128-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 120mg/kg	
11. Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg	
12. Methyl methacrylate	(0404)	MKGW6137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg	
13. Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg	
14. 2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	78-46-9	10 ppm (35mg/m3/8H)	or-rat 720mg/kg	
15. Perchloroethane	(0460)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	78-01-7	N/A	N/A	
16. 1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	78-13-1	1000 ppm (7600mg/m3/8H)	or-rat 43g/kg	
17. Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	N/A	or-rat 915mg/kg	
18. Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-48-1	N/A	or-rat 848mg/kg	
19. cis-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	N/A	N/A	
20. trans-1,2-Dichloroethene	35171	101623	0.05	5.00	40003.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	N/A	N/A	
21. Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 820mg/kg	
22. 1,1-Dichloroethane	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg	
23. Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 933mg/kg	
24. Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg	
25. Chloroform	95321	020724	0.10	10.00	20004.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	60 ppm (240mg/m3) (CL)	or-rat 908mg/kg	
26. Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-95-3	N/A	or-rat 106mg/kg	
27. 1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	N/A	N/A	
28. 2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg	
29. Trichloroethene	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.8	20.8	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg	
30. 1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg	
31. 1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	86-12-8	0.001 ppm	or-rat 170mg/kg	
32. 1,2-Dibromoethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg	
33. 1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-06-2	50 ppm (8H)	or-rat 670mg/kg	
34. 1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2002.0	22.9	78-87-5	75 ppm (350mg/m3/8H)	or-rat 1847mg/kg	
35. 1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	N/A	or-rat 3600mg/kg	
36. 1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	28.7	563-58-6	N/A	N/A	
37. cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	N/A	N/A	
38. trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.6	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	N/A	N/A	
39. Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	29.7	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg	
40. 1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	830-20-6	N/A	or-rat 670mg/kg	
41. 1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg	
42. 1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (45mg/m3/8H) (skin)	or-rat 936mg/kg	
43. Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-mus 2402mg/kg	
44. 1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (60mg/m3/8H)	or-rat 149.8mg/kg	
45. Benzene	35162	050823	0.05	5.00	40006.0	2000	NA	NA	0.017	NA	NA	1999.9	22.9	71-43-2	1 ppm	or-rat 4894mg/kg	
46. Bromobenzene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	N/A	or-rat 2699mg/kg	
47. n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg	
48. Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	99-87-6	N/A	or-rat 4750mg/kg	
49. p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg	
50. Naphthalene	35162	050823	0.05	5.00	40005.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	100-42-5	100 ppm	or-rat 5000mg/kg	
51. Styrene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-98-3	200 ppm	or-rat 5000mg/kg	
52. Toluene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-81-6	N/A	or-rat 1390mg/kg	
53. 1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.8	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 756mg/kg	
54. 1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	N/A	or-rat 5g/kg	
55. 1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-87-6	N/A	or-rat 5000mg/kg	
56. 1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg	
57. m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	96-06-6	N/A	N/A	
58. tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	135-98-6	N/A	or-rat 2240mg/kg	
59. sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg	
60. Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg	
61. 2-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	N/A	or-rat 2100mg/kg	
62. 4-Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	95-50-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg	
63. 1,2-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	541-73			




Run 16, "P95317 L021624 (2000µg/mL in MeOH)"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments

GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.53
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.60
5	Iodomethane	12.91
6	Allyl chloride	12.96
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethene	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.74
11	cis-1,2-Dichloroethene	18.00
12	Methoxyvinyltoluene/Methyl acrylate/Chloroform	18.49
13	Isobutene/1,1,1-Trichloroethane	18.91
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.48
17	Trichloroethene	23.88
18	1,2-Dichloropropane	24.52
19	Methyl methacrylate	25.13
20	Bromochloropropane	25.46
21	Dibromomethane/2,2-Dichloropropane	27.07
22	cis-1,2-Dichloroethene	28.03
23	Toluene	28.73
24	Ethyl methacrylate/Trans-1,2-Dichloropropane	29.34
25	1,1,2-Trichloroethane	30.34
26	Tetrachloroethane/1,2-Dichloropropane	31.16
27	Dibromochloromethane	32.84
28	1,2-Dibromomethane	33.26
29	Chlorobenzene	33.40
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.85
31	m-Xylene/p-Xylene	35.33
32	o-Xylene	35.70
33	Styrene	35.70
34	Isopropylbenzene/Bromofarm	36.48
35	cis-1,4-Dichloro-3-butene	36.80
36	1,1,2-Trichloropropane	37.23
37	n-Propylbenzene	37.77
38	trans-1,4-Dichloro-3-butene	37.92
39	Bromobenzene	38.05
40	1,3,5-Trimethylbenzene	38.14
41	1,3,5-Trimethylbenzene	38.50
42	Chlorobenzene	38.63
43	4-Chlorobenzene	38.77
44	tert-Butylbenzene	39.76
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Trichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.52
52	1,2-Dichlorobenzene	42.58
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.40
55	1,2,4-Trifluorobenzene	49.26
56	Hexachlorocyclopentadiene	49.72
57	Naphthalene	50.56
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water
----------------------------------	---	--	---



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#	% (optional)
Methanol METHYL ALCOHOL	67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

IATA
UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



Certified Reference Material CRM



Rec 09/17/24

CERTIFIED WEIGHT REPORT

Part Number:

91980

Lot Number:

091424

Description:

Acrolein

Solvent(s):

Water

Lot#

072324Q

Expiration Date:

101424

Recommended Storage:

Refrigerate (4 °C)

Nominal Concentration (µg/mL):

5000

NIST Test ID#:

6UTB

5E-05 Balance Uncertainty

0.001 Flask Uncertainty

10.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (±) (µg/mL)	(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)	LD50
----------	-----	------------	----------------------	------------	-------------	------------------	------------------	---------------------	----------------------------------	--	------	----------------	------

1. Acrolein	5	103755V10F	5000	97	0.5	0.05186	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg
-------------	---	------------	------	----	-----	---------	---------	--------	------	----------	---------	-----------------

Method: GC/MSD-1, Detector: Mass Selective Detector (Scan mode). Columns: Vocol (60m X 0.25mm ID X 1.5µm film thickness), Oven Profile: Temp. 1 = 35°C (Time 1 = 0min.), Temp. 2 = 200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas, NOTE: Due to the instability of acrolein in solution, all solutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

TIC: [BSB2]79005.D

Scan 232 (8.927 min): [BSB2]79005.D

Abundance

27

250000

8.93

200000

C=CC=O

50000

56

150000

40000

30000

20000

50000

10000

37

Time-->

0

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00

m/z-->

0

20 30 40 50 60 70 80

65 75 85

119

158 169

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (±) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Certified Reference Material CRM



Dec 09/17/24

CERTIFIED WEIGHT REPORT

Part Number:

91980

Lot Number:

091424

Description:

Acrolein

Solvent(s):

Water

Lot#

072324Q

Expiration Date:

101424

Recommended Storage:

Refrigerate (4 °C)

Nominal Concentration (µg/mL):

5000

NIST Test ID#:

6UTB

5E-05 Balance Uncertainty

0.001 Flask Uncertainty

10.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)	LD50
----------	-----	------------	----------------------	------------	-------------	------------------	------------------	---------------------	------------------------------------	--	----------------	------

1. Acrolein	5	103755V10F	5000	97	0.5	0.05186	0.05175	5008.9	52.5	107-02-8	0.1 ppm	ori-rat 46mg/kg
-------------	---	------------	------	----	-----	---------	---------	--------	------	----------	---------	-----------------

Method: GC/MSD-1, Detector: Mass Selective Detector (Scan mode). Columns: Vocol (60m X 0.25mm ID X 1.5µm film thickness), Oven Profile: Temp. 1 = 35°C (Time 1 = 0min.), Temp. 2 = 200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas, NOTE: Due to the instability of acrolein in solution, all solutions thereof, should be used immediately. Long term storage is not recommended. Please contact our technical department if further information is required.

TIC: [BSB2]79005.D

Scan 232 (8.927 min): [BSB2]79005.D

Abundance

27

250000

8.93

200000

C=CC=O

50000

56

150000

40000

30000

20000

50000

10000

37

Time-->

10.00 15.00 20.00 25.00 30.00 35.00 40.00 45.00 50.00 55.00 60.00

m/z-->

0

20 30 40 50 60 70 80 90 100 110 120 130 140 150 160 170

44 65 75 85 119 158 169

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Dec 12/16/24
20 vial

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#
Methanol EJ143-US

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL):


5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Formulated By: Prashant Chauhan 120524
Reviewed By: Pedro L. Rentas 120524

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information	
										(Solvent Safety Info. On Attached pg.)	LD50

1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	or-rat 250mg/kg
------------------------------	----	----------	-------	----	-----	---------	---------	---------	------	----------	-----	-----------------

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp = 220°C. Analyst: Candice Warren.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



See 1216124
20 vial

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#
Methanol EJ143-US

Expiration Date:
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Formulated By: Prashant Chauhan 120524
Reviewed By: Pedro L. Rentas 120524


Weight(s) shown below were combined and diluted to (mL):

5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (±) (µg/mL)	SDS Information	
										(Solvent Safety Info. On Attached pg.)	LD50

1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	40.5	110-75-8	N/A	or-rat 250mg/kg
------------------------------	----	----------	-------	----	-----	---------	---------	---------	------	----------	-----	-----------------

Method: GC6MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., **Injector B Temp:** = 200°C, **Detector B Temp.** = 220°C. **Analyst:** Candice Warren.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (±) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

GHS/OSHA Compliant

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rosotto Dr. Hamden CT. 06514	Emergency Telephone International Date Prepared/Revised	1-352-323-3500 January 1, 2024

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302, 332	If on skin, wash with soap and water	P305, 351, 338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Components (Specific Chemical Identity; Common Name(s))	% (optional)
Methanol METHYL ALCOHOL CAS#: 67-56-1	> 97

INTENDED USE: REFERENCE MATERIAL

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Methanol	67-56-1 TWA 200 ppm		
Skin notation	TWA 200 ppm		
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



See 1216124
20 vial

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#
Methanol EJ143-US

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

Weight(s) shown below were combined and diluted to (mL):


5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Formulated By: Prashant Chauhan 120524
Reviewed By: Pedro L. Rentas 120524

Expanded Uncertainty (Solvent Safety Info. On Attached pg.)
Uncertainty (+/-) (µg/mL) CAS# OSHA PEL (TWA) LDSO

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	or-rat 250mg/kg
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	N/A

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



See 1216124
20 vial

CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 120524
Description: 2-Chloroethyl vinyl ether

Solvent(s): Lot#
Methanol EJ143-US

Expiration Date: 120527
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB

✓ 14630 to
✓ 14649

Weight(s) shown below were combined and diluted to (mL):


5E-05 Balance Uncertainty
0.001 Flask Uncertainty

Expanded Uncertainty (Solvent Safety Info. On Attached pg.)
Uncertainty (±) (µg/mL) CAS# OSHA PEL (TWA) LDSO

SDS Information

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	or-rat 250mg/kg
1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.50536	0.50550	10002.9	N/A

Method: GC6MSD-1.M. Detector: MSD. Column: (60m X 0.25mm X 1.5 µm). Oven Profile: Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., Injector B Temp = 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (±) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))		% (optional)
Methanol	METHYL ALCOHOL	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist. Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Storage Conditions	Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1	67-56-1	TWA 200 ppm
Skin notation			TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.			
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use.	Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.			

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix
8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555581 **Lot No.:** A0210184

Description: Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 11275.10105



Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus
V14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618
Description : VOA Calibration Mix #1
VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : July 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Dec 12/17/24
30 v. 4
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

✓ 14697-to-14726



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B707717271


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
CERTIFIED REFERENCE MATERIAL

Dec: 12/09/24
Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
Dec: 12/09/24
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

10 vial.
CERTIFIED REFERENCE MATERIAL

Dec: 12/09/24
Certificate of Analysis
chromatographic plus
V14667-5
V14676



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0214960
Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	SYN240416CTH	99%	2,012.0 µg/mL	+/- 113.0519

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Stacey Wanner - Operations Technician I

Date Mixed: 08-Aug-2024

Balance Serial # 1127510105

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 14-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
30 ml
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24
CERTIFIED REFERENCE MATERIAL

30 ml
Certificate of Analysis
chromatographic plus

V14727 to
V14756



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826
Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2031 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Tom Suckal - Mix Technician

Date Mixed: 23-Sep-2024

Balance Serial # B707717271



Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

✓ 14842 to 14846



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 **Lot No.:** A0217535
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : October 31, 2027 **Storage:** 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	SHBQ8002-1	99%	50,007.5 µg/mL	+/- 717.6137

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A. O. E.
Aaron Enyart - Operations Tech I

Date Mixed: 07-Oct-2024

Balance Serial # B251644995

Brittany Federinko
Brittany Federinko - Operations Tech I

Date Passed: 09-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

2014 Dec 01/08/21
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

V14803 - V14822



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0220471
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2026 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RD240423RSR	99%	8,066.0 µg/mL	+/- 278.7979

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 24-Dec-2024

Balance Serial # 1127510105

Dillan Murphy - Operations Technician I

Date Passed: 02-Jan-2025

REVIEWED
By Jennifer Pollock at 7:12 am, Jan 05, 2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



10 vial
See 03/31/25
V14904 to V14913

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555582 Lot No.: A0223904

Description: Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size: 2 mL Pkg Amt: > 1 mL

Expiration Date: March 31, 2028 Storage: 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-33313	99%	25,108.0 µg/mL	+/- 1,421.9957
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	0000268853	99%	25,108.0 µg/mL	+/- 1,421.9957
3	Dibromofluoromethane	1868-53-7	VENKAT02	99%	25,232.0 µg/mL	+/- 1,429.0184
4	Toluene-d8	2037-26-5	PR-34141	99%	25,156.0 µg/mL	+/- 1,424.7141

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Brittany Federlino

Brittany Federlino - Operations Tech I

Date Mixed: 27-Mar-2025

Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

Rec 05/09/25
Avantor™



*V14921 to
V14938*

Material No.: 9077-02
Batch No.: 24G0262002
Manufactured Date: 2024-05-14
Expiration Date: 2027-05-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μeq/g)	≤ 0.3	0.3
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 82608 CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Mark
Jamie Croak
Director Quality Operations, Bioscience Production



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **081325**
Description: **Acrolein**

Solvent(s): **Water**
Lot# **041725Q**

Formulated By:	<i>Eli Allaga</i>	081325
	EI Allaga	DATE
Reviewed By:	<i>Pedro L. Rentas</i>	081325
	Pedro L. Rentas	DATE


Expiration Date: **091325**
Recommended Storage: **Refrigerate (2°C to 8°C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**
Weight(s) shown below were combined and diluted to (mL): **10.0**
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

11505740
115061

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (±) (µg/mL)	SDS Information	
										(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)

1. Acrolein 5 103755V10F 5000 97 0.5 0.05166 0.05176 5009.9 52.6 107-02-8 0.1 ppm *or: rat 46mg/kg*

Method: GC/MSD-1, Detector: Mass Selective Detector (Scan mode), Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness), Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C, Analyst: Pedro Rentas, NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1996).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 081325
Description: Acrolein

Solvent(s): Lot#
Water 041725Q

Formulated By:	EI Allaga	081325
DATE		
Reviewed By:	Pedro L. Renteria	081325
DATE		


Expiration Date: 091325
Recommended Storage: Refrigerate (2°C to 8°C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 10.0
5E-05 Balance Uncertainty
0.001 Flask Uncertainty

11505740
115061

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (±) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
----------	-----	------------	----------------------	------------	--------------------	------------------	------------------	---------------------	----------------------------------	------	----------------	------

1. Acrolein 5 103755V10F 5000 97 0.5 0.05166 0.05176 5009.9 52.6 107-02-8 0.1 ppm or: rat 46mg/kg

Method: GC/MSD-1, Detector: Mass Selective Detector (Scan mode), Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness), Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C, Analyst: Pedro Renteria, NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1996).
- Rev 1.0, 2/25/2025



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number: **91980**
Lot Number: **081325**
Description: **Acrolein**
Expiration Date: **091325**
Recommended Storage: **Refrigerate (2°C to 8°C)**
Nominal Concentration (µg/mL): **5000**
NIST Test ID#: **6UTB**


Solvent(s): **Water** Lot# **041725Q**
Weight(s) shown below were combined and diluted to (mL): **10.0** **5E-05 Balance Uncertainty**
0.001 Flask Uncertainty

Handwritten: 11505740
115061

Formulated By:	<i>Elu Aliaga</i>	081325
Reviewed By:	<i>Pedro L. Renteria</i>	081325
DATE		

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Acrolein	5	103755V10F	5000	97	0.5	0.05166	0.05176	5009.9	52.6	107-02-8	0.1 ppm	or: rat 46mg/kg

Method: GC6MSD-1, Detector: Mass Selective Detector (Scan mode), Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness), Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C, Analyst: Pedro Renteria, NOTE: Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated by an ISO 17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1996).
- Rev 1.0, 2/25/2025



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Jacobs**
ADDRESS: **412 Mt Kemble Ave Suite 100**
CITY **Morrisstown** STATE: **NJ** ZIP: **07960**
ATTENTION: **John Ynfante John.Ynfante@jacobs.com**
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **STC PTC**
PROJECT NO.: **D3868221** LOCATION: **Princeton Junction**
PROJECT MANAGER: **Mary Murphy**
e-mail: **Mary.Murphy@Jacobs.com**
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: **Mary Murphy** PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) **Standard TAT** DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1. SPECIFIC VOL% (4.100-12.4)
2. I.H. DIOL (6.000-12.4)
3. TOTAL WAXES (6.000-12.4)
4. DIOL (6.000-12.4)
5. ALKYL (6.000-12.4)
6. TDS (5.000-12.4)
7. AMON (5.000-12.4)
8. TRAC (5.000-12.4)
9. STAM (5.000-12.4)

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		A/E	E	B/E	E	E	E	E	A/E		← Specify Preservatives A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
								1	2	3	4	5	6	7	8	9		
1.	RMW-03B-90.4-08122025	GW		✓	8/12/25	1230	6	✓	✓					✓				
2.	RMW-02B-66.3-08122025 RMW-02B-66.3-08122025	GW		✓	8/12/25	1530	5	✓	✓					✓				
3.	MW-17B-55.5-08122025	GW		✓	8/14/25	1154	27	✓	✓	✓	✓	✓	✓	✓			MS/MSD	
4.	MW-06-6.5-08122025	GW		✓	8/14/25	1530	2		✓									
5.	MW-17B-55.5-081225-SIM	GW		✓	8/12/25	1154	3								✓			
6.	TB01-081225	DI		✓	8/12/25	1600	2	✓										
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. 1700	DATE/TIME: 8/12/25	RECEIVED BY: 1. 8-12-25	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.0 °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2.	Comments: See work order for list of site specific VOBs
RELINQUISHED BY SAMPLER: 3. 8-12-25	DATE/TIME: 8-12-25	RECEIVED BY: 3.	Comments: Dissolved Iron is Picked Filtered

From: Sohil Jodhani
Sent: Wednesday, August 13, 2025 2:23 PM
To: Ynfante, John
Cc: Yazmeen Gomez; Mohammad Ahmed; Nimisha Pandya
Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,

Sample IDs are updated as per below request. Lab will notify you if in case VOC-SIM (for Vinyl chloride) is not required for the samples based on low level VOA analysis.

Thanks & Regards,



There's a better way.

Sohil Jodhani (he/him/his)
QA/QC Director
Alliance Technical Group, LLC-Newark
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Sent: Wednesday, August 13, 2025 2:10 PM
To: Ynfante, John <John.Ynfante@jacobs.com>; Data-EWR <Data-EWR@alliancetg.com>; Yazmeen Gomez <yazmeen.gomez@alliancetg.com>
Cc: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>
Subject: Re: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

Hi John,
yaz is out of the office today , but i will coordinate with lab to make sure we communicate with you and run these analysis accordingly



There's a better way.

Mohammad Ahmed
Laboratory Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3151
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

From: Ynfante, John <John.Ynfante@jacobs.com>

Sent: Wednesday, August 13, 2025 2:01 PM

To: Data-EWR <Data-EWR@alliancetg.com>; Yazmeen Gomez <yazmeen.gomez@alliancetg.com>

Subject: RE: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Yazmeen,

A couple of comments on this Princeton login for Q2842:

1. I assume your SFAM-SIM method for VOCs can still only report vinyl chloride out of our project list of VOCs – if that is the case then note that sample Q2842-07 (MW-17B-55.5-081225-SIM) and any other SFAM-SIM samples only need to be analyzed if their corresponding 8260D-Low analysis shows the vinyl chloride to be non-detect and of course if the lab is able to analyze them by SIM. I know up to this point the lab has said the concentrations have been too high in the SIM samples to even run the SFAM-SIM on so let me know how it looks on those after your analysts run the regular VOCs run on them.
2. The date part of the sample IDs should have used a MMDDYY format, not MMDDYYYY as is listed on the chain. Please change the “2025” bit to just “25” on all these sample IDs. For example, “RMW-03B-90.4-08122025” should be changed to “RMW-03B-90.4-081225”, etc.

Thanks

- John Y.


From: Data-EWR@alliancetg.com <Data-EWR@alliancetg.com>

Sent: Wednesday, August 13, 2025 9:38 AM

To: Ongjoco, Alec <Alec.Ongjoco@jacobs.com>; Dillon, Alexa <Alexa.Dillon@jacobs.com>; Lader, Chelsea <Chelsea.Lader@jacobs.com>; Holmes, Daniel <Daniel.Holmes@jacobs.com>; Reamer, David <David.Reamer@jacobs.com>; Ynfante, John <John.Ynfante@jacobs.com>; Murphy, Mary <Mary.Murphy@jacobs.com>; Warren, Melissa <Melissa.Warren@jacobs.com>; Asher, Sarah <Sarah.Asher@jacobs.com>

Cc: yazmeen.gomez@alliancetg.com

Subject: [EXTERNAL] Login Summary Details For Project Former Schlumberger STC PTC Site D3868221-Q2842.



To John Ynfante;

Please see the attached Login Summary for the following project, or download the file using your login credentials from the link below.

Order ID : Q2842
Project ID : Former Schlumberger STC PTC Site D3868221
Download File : <https://chemtech.net/secureLogin.aspx>
Order Date : 8/13/2025 8:34:00 AM

Alliance's Project Manager : YAZMEEN GOMEZ , yazmeen.gomez@alliancetg.com , 908-728-3147
Alliance's Sales Executive : Jordan Hedvat , jordan.hedvat@alliancetg.com , 908-728-3144

Thank you for the opportunity to provide you with our services. For any questions please feel free to contact your project manager.

Click Here for our short online customer Survey <http://chemtech.net/ClientSurvey.aspx>.

Thank you,

Alliance Technical Group LLC.

Notice: The information transmitted in this e-mail message and in any attachments is intended Solely for the attention of the named addressee(s) and may contain confidential and/or privileged material. Any review, retransmission, dissemination or other use of, or taking of any action in reliance upon, this information by persons or entities other than the intended recipient is strictly prohibited and may be unlawful. If you have received this transmission in error, please notify us immediately by return e-mail, and permanently delete this transmission, including attachments if any, from any computer.

■

NOTICE - This communication may contain confidential and privileged information that is for the sole use of the intended recipient. Any viewing, copying or distribution of, or reliance on this message by unintended recipients is strictly prohibited. If you have received this message in error, please notify us immediately by replying to the message and deleting it from your computer.

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DP 08/18/2025

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2842 JACO05

Client Name : JACOBS Engineering Grou

Client Contact : John Ynfante

Invoice Name : JACOBS Engineering Grou

Invoice Contact : John Ynfante

Order Date : 8/13/2025 8:34:00 AM

Project Name : Former Schlumberger STC

Receive DateTime : 8/12/2025 6:22:00 PM

Purchase Order :

Project Mgr :

Level 3

Report Type : ~~Level 4~~

EDD Type : CH2MHILL

Hard Copy Date :

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2842-01	RMW-03B-90.4-08122025	Water	08/12/2025	12:30					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-02	RMW-02B-66.3-08122025	Water	08/12/2025	15:30					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-03	MW-17B-55.5-08122025	Water	08/12/2025	11:54					
	Q2842-03MS				VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-04	Q2842-01MS	Water	08/12/2025	11:54					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-05	Q2842-03MSD Q2842-01MSD	Water	08/12/2025	11:54					
					VOCMS Group3		8260-Low	10 Bus. Days	
Q2842-07	MW-17B-55.5-081225-SIM MW-17B-55.5-08122025-SIM	Water	08/12/2025	11:54					
					VOC-SIM		SFAM_VOCSIM	10 Bus. Days	
Q2842-08	TB01-08122025	Water	08/12/2025	16:00					
					VOCMS Group3		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2842	JACO05	Order Date : 8/13/2025 8:34:00 AM	Project Mgr :
Client Name : JACOBS Engineering Grou		Project Name : Former Schlumberger STC	Report Type : Level 4
Client Contact : John Ynfante		Receive DateTime : 8/12/2025 6:22:00 PM	EDD Type : CH2MHILL
Invoice Name : JACOBS Engineering Grou		Purchase Order :	Hard Copy Date :
Invoice Contact : John Ynfante			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
--------	-----------	--------	----------------	----------------	------	------------	--------	----------	--------------

Relinquished By :

Date / Time :


8-13-25 0915

Received By :

Date / Time :


8/13/25 09:15 Jg # 4

Storage Area : VOA Refridgerator Room