

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc

Datafile : VN087569.D

Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Limits		
		Result			Qual	RPD		Qual	Low	High	RPD
Lab Sample ID :	Q2842-04MS	Client Sample ID :		MW-17B-55.5-081225MS							
Vinyl chloride	5000	0	5400	ug/L	108				70 (69)	130 (125)	
1,1-Dichloroethene	5000	0	5200	ug/L	104				70 (53)	130 (162)	
1,1-Dichloroethane	5000	0	5300	ug/L	106				70 (78)	130 (122)	
cis-1,2-Dichloroethene	5000	3000	8100	ug/L	102				70 (82)	130 (124)	
1,1,1-Trichloroethane	5000	0	5400	ug/L	108				70 (83)	130 (117)	
Benzene	5000	0	5100	ug/L	102				70 (81)	130 (128)	
1,2-Dichloroethane	5000	0	5500	ug/L	110				70 (76)	130 (120)	
Trichloroethene	5000	3600	10000	ug/L	128				70 (28)	130 (175)	
1,1,2-Trichloroethane	5000	0	5300	ug/L	106				70 (27)	130 (175)	
Tetrachloroethene	5000	56.3	4300	ug/L	85				70 (48)	130 (153)	