

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2842

Analytical Method: SW8260-Low

Client: JACOBS Engineering Group, Inc.

Datafile : VN087570.D

Parameter	Spike	Sample	Result	Units	Rec		RPD	Limits			
		Result			Qual	RPD	Qual	Low	High	RPD	
Lab Sample ID :	Q2842-05MSD	Client Sample ID :	MW-17B-55.5-081225MSD								
Vinyl chloride	5000	0	5400	ug/L	108		0		70 (69)	130 (125)	20 (20)
1,1-Dichloroethene	5000	0	5300	ug/L	106		2		70 (53)	130 (162)	20 (20)
1,1-Dichloroethane	5000	0	5400	ug/L	108		2		70 (78)	130 (122)	20 (20)
cis-1,2-Dichloroethene	5000	3000	8000	ug/L	100		2		70 (82)	130 (124)	20 (20)
1,1,1-Trichloroethane	5000	0	5500	ug/L	110		2		70 (83)	130 (117)	20 (20)
Benzene	5000	0	5100	ug/L	102		0		70 (81)	130 (128)	20 (20)
1,2-Dichloroethane	5000	0	5300	ug/L	106		4		70 (76)	130 (120)	20 (20)
Trichloroethene	5000	3600	9500	ug/L	118		8		70 (28)	130 (175)	20 (20)
1,1,2-Trichloroethane	5000	0	5300	ug/L	106		0		70 (27)	130 (175)	20 (20)
Tetrachloroethene	5000	56.3	4300	ug/L	85		0		70 (48)	130 (153)	20 (20)