

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>JACO05</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2842</u>
Lab File ID:	<u>VN087552.D</u>	Date Analyzed:	<u>08/14/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:34</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	255173	13.77				
UPPER LIMIT	510346	14.27				
LOWER LIMIT	127587	13.27				
EPA SAMPLE NO.						
RMW-03B-90.4-08122025	213548	13.77				
RMW-02B-66.3-08122025	211146	13.77				
MW-17B-55.5-08122025	209426	13.77				
MW-17B-55.5-081225MS	231778	13.77				
MW-17B-55.5-081225MSD	240017	13.77				
TB01-081225	214662	13.77				
VN0814WBL01	234765	13.77				
VN0814WBS01	221983	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.