

**Metals**

**- 2a -**

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc

**SDG No.:** Q2845

**Contract:** TULL01

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| ICV01     | Copper  | 1010           | 1000       | 102           | 95 - 105                  | P | 08/15/2025       | 11:50            | LB136851      |
|           | Iron    | 3940           | 4000       | 98            | 95 - 105                  | P | 08/15/2025       | 11:50            | LB136851      |
|           | Lead    | 3820           | 4000       | 96            | 95 - 105                  | P | 08/15/2025       | 11:50            | LB136851      |

---

**Metals**

**- 2a -**

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc

**SDG No.:** Q2845

**Contract:** TULL01

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| LLICV01   | Copper  | 21.8           | 20.0       | 109           | 80 - 120                  | P | 08/15/2025       | 11:56            | LB136851      |
|           | Iron    | 103            | 100        | 103           | 80 - 120                  | P | 08/15/2025       | 11:56            | LB136851      |
|           | Lead    | 11.7           | 12.0       | 98            | 80 - 120                  | P | 08/15/2025       | 11:56            | LB136851      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Tully Environmental, Inc

**SDG No.:** Q2845

**Contract:** TULL01

**Lab Code:** ACE

**Initial Calibration Source:** EPA

**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Copper  | 1240           | 1250       | 99            | 90 - 110                  | P | 08/15/2025       | 12:26            | LB136851      |
|           | Iron    | 4980           | 5000       | 100           | 90 - 110                  | P | 08/15/2025       | 12:26            | LB136851      |
|           | Lead    | 4870           | 5000       | 97            | 90 - 110                  | P | 08/15/2025       | 12:26            | LB136851      |
| CCV02     | Copper  | 1210           | 1250       | 96            | 90 - 110                  | P | 08/15/2025       | 13:40            | LB136851      |
|           | Iron    | 4840           | 5000       | 97            | 90 - 110                  | P | 08/15/2025       | 13:40            | LB136851      |
|           | Lead    | 4680           | 5000       | 94            | 90 - 110                  | P | 08/15/2025       | 13:40            | LB136851      |
| CCV03     | Copper  | 1220           | 1250       | 97            | 90 - 110                  | P | 08/15/2025       | 15:04            | LB136851      |
|           | Iron    | 5030           | 5000       | 101           | 90 - 110                  | P | 08/15/2025       | 15:04            | LB136851      |
|           | Lead    | 4720           | 5000       | 94            | 90 - 110                  | P | 08/15/2025       | 15:04            | LB136851      |
| CCV04     | Copper  | 1240           | 1250       | 99            | 90 - 110                  | P | 08/15/2025       | 15:55            | LB136851      |
|           | Iron    | 5040           | 5000       | 101           | 90 - 110                  | P | 08/15/2025       | 15:55            | LB136851      |
|           | Lead    | 4800           | 5000       | 96            | 90 - 110                  | P | 08/15/2025       | 15:55            | LB136851      |
| CCV05     | Copper  | 1250           | 1250       | 100           | 90 - 110                  | P | 08/15/2025       | 16:38            | LB136851      |
|           | Iron    | 5050           | 5000       | 101           | 90 - 110                  | P | 08/15/2025       | 16:38            | LB136851      |
|           | Lead    | 4830           | 5000       | 97            | 90 - 110                  | P | 08/15/2025       | 16:38            | LB136851      |