



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc

SDG No.: Q2845

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.0	20.0	110	65 - 135	P	08/15/2025	12:04	LB136851
	Iron	103	100	103	65 - 135	P	08/15/2025	12:04	LB136851
	Lead	12.6	12.0	105	65 - 135	P	08/15/2025	12:04	LB136851