

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q2845

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-6.62	2.0	331	-18	22	08/15/2025	12:09	LB136851
	Iron	102000	101000	101	85600	116500	08/15/2025	12:09	LB136851
	Lead	-4.50			-12	12	08/15/2025	12:09	LB136851
ICSAB01	Copper	483	511	94	434	588	08/15/2025	12:13	LB136851
	Iron	102000	99300	103	84400	114500	08/15/2025	12:13	LB136851
	Lead	42.4	49.0	86	37	61	08/15/2025	12:13	LB136851
ICSA	Copper	15.1	2.0	755	-18	22	08/15/2025	12:17	LB136851
	Iron	96400	101000	95	85600	116500	08/15/2025	12:17	LB136851
	Lead	-16.5			-12	12	08/15/2025	12:17	LB136851
ICSAB	Copper	507	511	99	434	588	08/15/2025	12:21	LB136851
	Iron	97900	99300	99	84400	114500	08/15/2025	12:21	LB136851
	Lead	53.0	49.0	108	37	61	08/15/2025	12:21	LB136851