

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2845</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2845-04</u>	client id: <u>002 35th Ave (july)MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2845-04MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	222		74.0		150	99		P
Iron	ug/L	75 - 125	7760		5070		1500	180	N	P
Lead	ug/L	75 - 125	445		11.4		500	87		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2845</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2845-04</u>	client id: <u>002 35th Ave (july)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2845-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	191		74.0		150	78		P
Iron	ug/L	75 - 125	5420		5070		1500	23	N	P
Lead	ug/L	75 - 125	438		11.4		500	85		P