

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2845

Contract: TULL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169267BS							
Copper	ug/L	150	155		103	85 - 115	P
Iron	ug/L	1500	1490		99	85 - 115	P
Lead	ug/L	500	470		94	85 - 115	P