

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2845</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tully Environmental, Inc</u>	Datafile :	<u>VN087596.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0818WBS01	Benzene	20	17.9	ug/L	90			65	135	
	Toluene	20	18.0	ug/L	90			70	130	
	Ethyl Benzene	20	17.7	ug/L	89			60	140	
	m/p-Xylenes	40	36.0	ug/L	90			87	111	
	o-Xylene	20	17.8	ug/L	89			87	111	