

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2845
 Lab File ID: VN087595.D Date Analyzed: 08/18/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:29
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	60212	7.80	341524	9.08	305141	11.85
UPPER LIMIT	120424	8.3	683048	9.582	610282	12.347
LOWER LIMIT	30106	7.3	170762	8.582	152571	11.347
EPA SAMPLE NO.						
001 willets Pt Blvd(july)	59234	7.80	334882	9.08	296304	11.85
002 35th Ave (july)	48558	7.80	265247	9.08	239168	11.85
002 35th Ave (july)DL	55447	7.79	303389	9.08	270310	11.85
VN0818WBL01	60369	7.80	336175	9.08	294352	11.85
VN0818WBS01	64865	7.80	356557	9.08	320424	11.85
VN0818WBSD01	61980	7.79	351353	9.08	310931	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.