

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2862

Project: Linden Generating Station - LRSA

Run No.: LB136821

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136821BS							
BOD5	mg/L	198	206		104	1	84.6-115.4	08/14/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2862

Project: Linden Generating Station - LRSA

Run No.: LB136825

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136825BS							
COD	mg/L	50	47.9		96	1	90-110	08/15/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2862

Project: Linden Generating Station - LRSA

Run No.: LB136829

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136829BS							
TSS	mg/L	550	532		97	1	90-110	08/15/2025