

## **CASE NARRATIVE**

**Sciacca General Contractors, LLC**

**Project Name: 507 Franklin Ave**

**Project # N/A**

**Order ID # Q2868**

**Test Name: EPH\_F2**

### **A. Number of Samples and Date of Receipt:**

5 Solid samples were received on 08/13/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: EPH\_F2. This data package contains results for EPH\_F2.

### **C. Analytical Techniques:**

The analysis were performed on instrument FID\_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis of EPH\_F2s was based on method NJEPH and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for 1DL [1-chlorooctadecane (SURR)-150%], 2DL [1-chlorooctadecane (SURR)-260% ], 3DL [1-chlorooctadecane (SURR)-149% ], 5DL [1-chlorooctadecane (SURR)152% ]. Due to high concentration of compounds, these samples required dilution. Therefore, samples were reanalyzed with dilution and reported .

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD {Q2868-07MSD} with File ID: FE055414.D recoveries met the requirements for all compounds except for Aliphatic C28-C40[14%], Aliphatic C9-C28[6%] due to matrix interference.

The RPD for {Q2868-07MSD} with File ID: FE055414.D met criteria except for Aliphatic C28-C40[132%], Aliphatic C9-C28[161.6%] due to difference in MS-MSD concentrations.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration met the requirements.

Sample 1,2,3,4,5 were diluted due to high concentration for Aliphatic compounds.

**E. Additional Comments:**

The soil samples results are based on a dry weight basis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_