

Continuing Calibration Report for SequenceID : FE082025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055404.D

Aliphatic C9-C12	6808628.000	60.000	3.321	6.956	113477.133	116782.597	2.830
Aliphatic C12-C16	4917613.000	40.000	6.957	10.409	122940.325	125227.070	1.826
Aliphatic C16-C21	7724195.000	60.000	10.410	13.790	128736.583	126404.887	-1.845
Aliphatic C21-C28	9827088.000	80.000	13.791	17.463	122838.600	120505.719	-1.936
Aliphatic C28-C40	14229532.000	120.000	17.464	22.490	118579.433	113987.195	-4.029
Aliphatic EPH	43507056.000	360.000	3.321	22.490	120852.933	119220.147	-1.370