

Continuing Calibration Report for SequenceID : FE082025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055416.D

Aliphatic C9-C12	6775879.000	60.000	3.321	6.956	112931.317	116782.597	3.298
Aliphatic C12-C16	4891200.000	40.000	6.957	10.409	122280.000	125227.070	2.353
Aliphatic C16-C21	7720477.000	60.000	10.410	13.790	128674.617	126404.887	-1.796
Aliphatic C21-C28	9966576.000	80.000	13.791	17.463	124582.200	120505.719	-3.383
Aliphatic C28-C40	15774740.000	120.000	17.464	22.490	131456.167	113987.195	-15.325
Aliphatic EPH	45128872.000	360.000	3.321	22.490	125357.978	119220.147	-5.148