

Continuing Calibration Report for SequenceID : FE082125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055419.D

Aliphatic C9-C12	7465718.000	60.000	3.319	6.954	124428.633	116782.597	-6.547
Aliphatic C12-C16	5344651.000	40.000	6.955	10.407	133616.275	125227.070	-6.699
Aliphatic C16-C21	8289208.000	60.000	10.408	13.787	138153.467	126404.887	-9.294
Aliphatic C21-C28	10521404.000	80.000	13.788	17.460	131517.550	120505.719	-9.138
Aliphatic C28-C40	14885380.000	120.000	17.461	22.485	124044.833	113987.195	-8.823
Aliphatic EPH	46506361.000	360.000	3.319	22.485	129184.336	119220.147	-8.358