

Continuing Calibration Report for SequenceID : FE082125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055426.D

Aliphatic C9-C12	7170278.000	60.000	3.319	6.954	119504.633	116782.597	-2.331
Aliphatic C12-C16	5104014.000	40.000	6.955	10.407	127600.350	125227.070	-1.895
Aliphatic C16-C21	7932486.000	60.000	10.408	13.787	132208.100	126404.887	-4.591
Aliphatic C21-C28	10243034.000	80.000	13.788	17.460	128037.925	120505.719	-6.250
Aliphatic C28-C40	14319200.000	120.000	17.461	22.485	119326.667	113987.195	-4.684
Aliphatic EPH	44769012.000	360.000	3.319	22.485	124358.367	119220.147	-4.310