

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

ProjectID: 507 Franklin Ave

Lab Code: ACE

SDG No.: Q2868

Calibration Sequence : FG080525		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	159549953	93853	FG016345.D
850	82027875	96503	FG016346.D
340	33685502	99075	FG016347.D
170	17686270	104037	FG016348.D
85	8678686	102102	FG016349.D
AVG RF : 99114		% RSD : 4.15	AVG RT : 15.2092