

**TPH GC CONTINUING CALIBRATION SUMMARY**

**50 PPM TRPH STD**

Lab Name: Alliane Contract: SCIA01  
ProjectID: 507 Franklin Ave  
Lab Code: ACE SDG No.: Q2868  
DataFile: FG016406.D Analyst Name: YP\AJ Analyst Date: 08-26-2025

| Conc. (PPM) | Area Count | RF     | Average RF | %D    |
|-------------|------------|--------|------------|-------|
| 500         | 54497296   | 108995 | 109253     | 0.236 |