

Initial and Continuing Calibration Verification

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Project: Former Schlumberger STC PTC Site D3868221

RunNo.: LB136833

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.3	10	103	90-110	08/06/2025
Chloride	mg/L	3.2	3	107	90-110	08/06/2025
Fluoride	mg/L	2	2	100	90-110	08/06/2025
Nitrite	mg/L	3.1	3	103	90-110	08/06/2025
Nitrate	mg/L	2.6	2.5	104	90-110	08/06/2025
Sulfate	mg/L	15.4	15	103	90-110	08/06/2025
Orthophosphate as P	mg/L	5	5	100	90-110	08/06/2025
Sample ID: CCV1						
Bromide	mg/L	10.2	10	102	90-110	08/15/2025
Chloride	mg/L	3.1	3	103	90-110	08/15/2025
Fluoride	mg/L	2	2	100	90-110	08/15/2025
Nitrite	mg/L	3	3	100	90-110	08/15/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/15/2025
Sulfate	mg/L	15.2	15	101	90-110	08/15/2025
Orthophosphate as P	mg/L	5	5	100	90-110	08/15/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	08/15/2025
Chloride	mg/L	3.1	3	103	90-110	08/15/2025
Fluoride	mg/L	2	2	100	90-110	08/15/2025
Nitrite	mg/L	3	3	100	90-110	08/15/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/15/2025
Sulfate	mg/L	15.2	15	101	90-110	08/15/2025
Orthophosphate as P	mg/L	4.6	5	92	90-110	08/15/2025
Sample ID: CCV3						
Bromide	mg/L	10.2	10	102	90-110	08/15/2025
Chloride	mg/L	3.1	3	103	90-110	08/15/2025
Fluoride	mg/L	2	2	100	90-110	08/15/2025
Nitrite	mg/L	3	3	100	90-110	08/15/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/15/2025
Sulfate	mg/L	15.1	15	101	90-110	08/15/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	08/15/2025