

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/13/25
Project:	Former Schlumberger STC PTC Site D3868221	Date Received:	08/13/25
Client Sample ID:	MW-19B-71.5-081325	SDG No.:	Q2869
Lab Sample ID:	Q2869-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	21300	1	7.81		50.0	ug/L	08/14/25 14:10	08/15/25 15:55	6020B	3010A

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group3			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits