


Batch Folder: D:\Agilent\ICPMH\1\DATA\P7081525-MS.b\
Analysis File: P7081525-MS.batch.bin
DA Date-Time: 2025-08-15 17:00:51
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	002CALB.d	S00	2025-08-15 12:34:51
2	006CALS.d	S02	2025-08-15 12:48:25
3	007CALS.d	S03	2025-08-15 12:51:42
4	008CALS.d	S04	2025-08-15 12:54:53
5	009CALS.d	S05	2025-08-15 12:57:53
6	010CALS.d	S06	2025-08-15 13:00:51
7	011CALS.d	S07	2025-08-15 13:03:39
8	012CALS.d	S08	2025-08-15 13:06:29

9 Be [No Gas] ISTD:6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	4.54	0.0000	P	44.4
2	☐	1.000	1.140	885.63	0.0011	P	6.3
3	☐	50.000	52.171	40799.64	0.0521	P	1.2
4	☐	125.000	128.416	101471.65	0.1283	P	1.0
5	☐	250.000	255.671	200675.74	0.2555	P	1.7
6	☐	500.000	513.581	397234.28	0.5132	P	2.5
7	☐	1000.000	991.256	778273.95	0.9904	P	1.1
8	☐			142.36	0.0002	P	10.8

$$y = 9.9915E-004 * x + 6.0956E-006$$

$$R = 0.9998$$


$$DL = 0.008127$$

$$BEC = 0.006101$$

Weight: <None>


Min Conc: <None>

12 C [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

12 C [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

23 Na [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	38978.10	0.0928	P	0.7
2	☐	500.000	549.030	278404.13	0.6620	P	1.2
3	☐	5000.000	5220.923	2363656.56	5.5057	A	1.0
4	☐	12500.000	13374.940	5956904.77	13.9596	A	1.4
5	☐	25000.000	26514.688	11671013.86	27.5826	A	1.1
6	☐	50000.000	51171.642	23189127.15	53.1464	A	1.4
7	☐	100000.00	105208.68	46655050.97	109.1708	A	0.9
8	☐	500000.00	498741.23	230713507.6	517.1767	A	0.3

$$y = 0.0010 * x + 0.0928$$


$$R = 0.9999$$

$$DL = 1.766$$

$$BEC = 89.48$$

Weight: <None>

Min Conc: <None>



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	3041.44	0.0072	P	4.4
2	☐	500.000	531.007	157820.76	0.3752	P	0.6
3	☐	5000.000	5093.941	1518754.58	3.5375	A	0.6
4	☐	12500.000	13152.131	3892621.99	9.1220	A	0.3
5	☐	25000.000	26036.390	7637702.60	18.0511	A	1.9
6	☐	50000.000	50050.511	15135563.38	34.6934	A	2.9
7	☐	100000.00	102886.31	30474068.71	71.3099	A	0.5
8	☐	500000.00	499348.59	154359308.8	346.0674	A	2.8

$$y = 6.9302E-004 * x + 0.0072$$


R = 1.0000

DL = 1.388

BEC = 10.45

Weight: <None>

Min Conc: <None>



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1902.36	0.0045	P	2.9
2	☐	20.000	19.419	5222.06	0.0124	P	0.8
3	☐	1000.000	996.613	175727.53	0.4093	P	0.5
4	☐	2500.000	2544.435	442913.76	1.0380	P	1.0
5	☐	5000.000	5042.997	868547.34	2.0528	P	0.4
6	☐	10000.000	9814.931	1741270.08	3.9910	A	1.7
7	☐	20000.000	20251.636	3516957.38	8.2299	A	0.7
8	☐	100000.00	99964.953	18114209.45	40.6062	A	1.3

$$y = 4.0616E-004 * x + 0.0045$$

R = 1.0000


DL = 0.9825

BEC = 11.15

Weight: <None>

Min Conc: <None>

28 Si [No Gas] ISTD : 45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	548289.54	0.1454	P	1.3
2	☐	10.000	0.502	558152.49	0.1473	P	0.8
3	☐	50.000	63.765	1468051.82	0.3793	A	1.1
4	☐	125.000	144.217	2588714.39	0.6744	A	1.7
5	☐	250.000	272.557	4321776.67	1.1451	A	0.8
6	☐	500.000	525.853	7816369.19	2.0741	A	4.3
7	☐	1000.000	978.439	14344445.20	3.7341	A	2.2
8	☐			632768.87	0.1621	P	1.5

$$y = 0.0037 * x + 0.1454$$

R = 0.9989


DL = 1.58

BEC = 39.65

Weight: <None>

Min Conc: <None>

31 P [He] ISTD : 45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	-16.202	951.15	0.0023	P	1.3
2	☐	0.000	16.202	1254.51	0.0030	P	4.1
3	☐	1000.000	967.592	10360.27	0.0241	P	1.4
4	☐	2500.000	2542.538	25235.39	0.0591	P	1.3
5	☐	5000.000	5073.840	48830.15	0.1154	P	2.0
6	☐	10000.000	9700.172	95217.92	0.2182	P	1.9
7	☐	20000.000	20127.757	192308.99	0.4500	P	0.2
8	☐			916.70	0.0021	P	5.0

$$y = 2.2227E-005 * x + 0.0026$$

R = 0.9998

DL = 10.27


BEC = 118

Weight: <None>

Min Conc: <None>

Calibration for 008CAL.S.d

34 S [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	-48.173	165885.82	0.0440	P	1.1
2	☐	0.000	48.173	168920.77	0.0446	P	0.9
3	☐	1000.000	927.210	192648.05	0.0498	P	0.4
4	☐	2500.000	2581.394	228669.29	0.0596	P	0.6
5	☐	5000.000	5126.382	281669.05	0.0746	P	0.2
6	☐	10000.000	10129.039	392893.12	0.1042	P	3.1
7	☐	20000.000	19897.350	622620.00	0.1621	P	0.4
8	☐			161101.96	0.0413	P	0.6

$$y = 5.9189\text{E-}006 * x + 0.0443$$

R = 0.9999


DL = 225.1

BEC = 7483

Weight: <None>


Min Conc: <None>

35 Cl [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

35 Cl [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

39 K [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	78335.57	0.1864	P	1.3
2	☐	500.000	572.751	271860.36	0.6464	P	1.4
3	☐	2500.000	2642.403	991096.82	2.3086	P	1.3
4	☐	6250.000	6533.877	2318582.10	5.4340	A	1.6
5	☐	12500.000	13002.508	4497263.37	10.6291	A	0.3
6	☐	25000.000	24899.927	8807345.77	20.1841	A	0.6
7	☐	50000.000	51475.641	17746903.48	41.5277	A	1.2
8	☐	250000.00	249681.08	89531349.78	200.7109	A	1.5

$$y = 8.0312E-004 * x + 0.1864$$

$$R = 1.0000$$


$$DL = 8.755$$

$$BEC = 232.2$$

Weight: <None>

Min Conc: <None>

43 Ca [No Gas] ISTD: 45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	644.46	0.0002	P	1.2
2	☐	500.000	577.378	9322.95	0.0025	P	1.0
3	☐	5000.000	5318.528	82272.87	0.0213	P	0.7
4	☐	12500.000	13676.275	208803.20	0.0544	P	0.7
5	☐	25000.000	27106.423	406249.20	0.1076	P	0.4
6	☐	50000.000	54380.856	813253.85	0.2158	P	3.2
7	☐	100000.00	103844.53	1582305.67	0.4119	A	1.5
8	☐	500000.00	498655.01	7717566.07	1.9772	A	0.6

$$y = 3.9647\text{E-}006 * x + 1.7095\text{E-}004$$

$$R = 0.9999$$


$$DL = 1.612$$

$$BEC = 43.12$$

Weight: <None>

Min Conc: <None>

44 Ca [No Gas] ISTD: 45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	28525.80	0.0076	P	2.9
2	☐	500.000	574.614	168752.79	0.0445	P	0.7
3	☐	5000.000	5038.901	1283667.52	0.3317	A	1.7
4	☐	12500.000	13001.700	3239420.96	0.8439	A	1.3
5	☐	25000.000	25960.252	6330893.24	1.6774	A	0.9
6	☐	50000.000	51640.899	12547688.98	3.3293	A	3.4
7	☐	100000.00	103837.12	25688150.45	6.6867	A	1.4
8	☐	500000.00	499007.46	125319719.2	32.1055	A	0.6

$$y = 6.4324\text{E-}005 * x + 0.0076$$

$$R = 1.0000$$


$$DL = 10.28$$

$$BEC = 117.6$$

Weight: <None>

Min Conc: <None>

47 Ti [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	104.44	0.0000	P	18.3
2	☐	5.000	5.262	4171.72	0.0011	P	2.1
3	☐	500.000	482.840	381211.85	0.0985	P	0.3
4	☐	1250.000	1238.391	969619.22	0.2526	P	0.7
5	☐	2500.000	2501.833	1925771.20	0.5102	A	1.6
6	☐	5000.000	4980.441	3828273.83	1.0157	A	3.3
7	☐	10000.000	10011.630	7843437.66	2.0418	A	2.1
8	☐			1124.50	0.0003	P	1.0

$$y = 2.0394E-004 * x + 2.7732E-005$$

$$R = 1.0000$$


$$DL = 0.07475$$

$$BEC = 0.136$$

Weight: <None>

Min Conc: <None>

51 V [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	23.34	0.0001	P	66.2
2	☐	5.000	5.522	14479.27	0.0344	P	0.4
3	☐	50.000	49.044	131078.12	0.3053	P	0.7
4	☐	125.000	124.716	331246.82	0.7763	P	1.2
5	☐	250.000	250.757	660382.63	1.5608	P	0.3
6	☐	500.000	486.395	1320867.41	3.0275	A	2.0
7	☐	1000.000	1006.694	2677677.41	6.2659	A	0.7
8	☐			1320.07	0.0030	P	5.5

$$y = 0.0062 * x + 5.5780E-005$$

$$R = 0.9999$$


$$DL = 0.0178$$

$$BEC = 0.008962$$

Weight: <None>

Min Conc: <None>

52 Cr [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2722.56	0.0065	P	7.1
2	☐	2.000	2.162	9312.99	0.0222	P	3.7
3	☐	50.000	50.019	158390.73	0.3689	P	1.4
4	☐	125.000	127.597	397296.11	0.9311	P	1.0
5	☐	250.000	252.678	777412.10	1.8375	P	1.2
6	☐	500.000	485.637	1538162.63	3.5256	A	2.2
7	☐	1000.000	1006.186	3118485.20	7.2976	A	2.1
8	☐			17335.62	0.0389	P	1.3

$$y = 0.0072 * x + 0.0065$$

$$R = 0.9998$$


$$DL = 0.1908$$

$$BEC = 0.8946$$

Weight: <None>

Min Conc: <None>

54 Fe [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	3137.02	0.0075	P	2.1
2	☐	50.000	58.132	11865.88	0.0282	P	1.3
3	☐	2500.000	2802.334	432635.63	1.0077	P	0.8
4	☐	6250.000	7082.675	1081960.49	2.5355	P	0.2
5	☐	12500.000	13533.433	2047025.75	4.8381	A	0.2
6	☐	25000.000	25970.976	4047814.11	9.2775	A	1.7
7	☐	50000.000	54099.156	8254974.25	19.3175	A	1.4
8	☐	250000.00	249007.55	39651350.24	88.8878	A	1.1

$$y = 3.5694E-004 * x + 0.0075$$


$$R = 0.9999$$

$$DL = 1.345$$

$$BEC = 20.91$$

Weight: <None>

Min Conc: <None>



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1170.39	0.0028	P	1.8
2	☐	1.000	1.010	3801.13	0.0090	P	1.1
3	☐	500.000	498.933	1326785.14	3.0905	P	1.0
4	☐	1250.000	1257.977	3322684.80	7.7878	A	3.1
5	☐	2500.000	2523.109	6607661.43	15.6171	A	0.7
6	☐	5000.000	4880.369	13179680.10	30.2051	A	0.7
7	☐	10000.000	10053.095	26587058.29	62.2167	A	1.3
8	☐			14682.25	0.0329	P	1.3

$$y = 0.0062 * x + 0.0028$$


$$R = 0.9999$$

$$DL = 0.02467$$

$$BEC = 0.4501$$

Weight: <None>

Min Conc: <None>



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	27505.39	0.0655	P	2.4
2	☐	50.000	57.127	183421.08	0.4362	P	2.0
3	☐	2500.000	2615.449	7314966.90	17.0387	A	0.9
4	☐	6250.000	6667.192	18489629.45	43.3330	A	1.4
5	☐	12500.000	13301.286	36549257.51	86.3857	A	0.9
6	☐	25000.000	25567.078	72421181.14	165.9859	A	1.5
7	☐	50000.000	52756.586	146341352.2	342.4354	A	1.7
8	☐	250000.00	249340.32	721783153.6	1,618.188	A	2.5

$$y = 0.0065 * x + 0.0655$$


$$R = 0.9999$$

$$DL = 0.7395$$

$$BEC = 10.09$$

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1947.93	0.0046	P	2.7
2	☐	50.000	58.689	5961.36	0.0142	P	4.5
3	☐	2500.000	2707.433	190744.67	0.4443	P	1.2
4	☐	6250.000	6885.520	479072.60	1.1228	P	2.0
5	☐	12500.000	13810.256	950839.92	2.2473	P	0.6
6	☐	25000.000	25586.111	1814804.12	4.1597	A	2.5
7	☐	50000.000	52280.616	3630094.49	8.4947	A	0.6
8	☐	250000.00	249401.78	18068910.70	40.5059	A	1.0

$$y = 1.6239E-004 * x + 0.0046$$


R = 1.0000

DL = 2.27

BEC = 28.54

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	183.34	0.0004	P	18.3
2	☐	1.000	1.151	5532.18	0.0132	P	1.7
3	☐	50.000	49.939	237144.83	0.5524	P	1.8
4	☐	125.000	127.178	599990.72	1.4061	P	0.8
5	☐	250.000	252.367	1180323.00	2.7898	P	0.8
6	☐	500.000	489.646	2361452.41	5.4124	A	1.7
7	☐	1000.000	1004.316	4743836.87	11.1009	A	0.8
8	☐			9370.76	0.0210	P	3.0

$$y = 0.0111 * x + 4.3606E-004$$

R = 0.9999


DL = 0.02168

BEC = 0.03945

Weight: <None>

Min Conc: <None>

60 Ni [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	621.13	0.0015	P	7.2
2	☐	1.000	1.008	1821.24	0.0043	P	2.0
3	☐	50.000	50.880	62398.31	0.1454	P	2.0
4	☐	125.000	129.814	157265.31	0.3686	P	1.1
5	☐	250.000	255.908	306799.63	0.7251	P	0.9
6	☐	500.000	491.655	607303.52	1.3918	P	1.1
7	☐	1000.000	1002.050	1211531.07	2.8350	P	0.4
8	☐			5374.36	0.0120	P	3.9

$$y = 0.0028 * x + 0.0015$$

$$R = 0.9999$$


$$DL = 0.1136$$

$$BEC = 0.5227$$

Weight: <None>

Min Conc: <None>

63 Cu [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1514.54	0.0036	P	3.5
2	☐	2.000	2.089	8979.42	0.0214	P	2.0
3	☐	500.000	504.209	1840708.62	4.2876	A	1.3
4	☐	1250.000	1277.106	4630979.72	10.8544	A	3.0
5	☐	2500.000	2554.035	9182976.46	21.7037	A	0.1
6	☐	5000.000	4916.159	18228076.67	41.7731	A	0.5
7	☐	10000.000	10024.813	36401408.35	85.1782	A	0.9
8	☐			8009.99	0.0180	P	1.2

$$y = 0.0085 * x + 0.0036$$

$$R = 0.9999$$


$$DL = 0.0446$$

$$BEC = 0.4243$$

Weight: <None>

Min Conc: <None>

66 Zn [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	4596.31	0.0109	P	2.2
2	☐	5.000	4.496	7539.75	0.0179	P	0.7
3	☐	500.000	514.923	348125.54	0.8109	P	1.4
4	☐	1250.000	1302.681	868244.26	2.0347	P	1.0
5	☐	2500.000	2607.539	1718729.63	4.0619	A	1.3
6	☐	5000.000	4985.014	3383717.07	7.7555	A	1.9
7	☐	10000.000	9973.277	6626229.42	15.5050	A	1.0
8	☐			8224.56	0.0184	P	2.5

$$y = 0.0016 * x + 0.0109$$

$$R = 0.9999$$


$$DL = 0.456$$

$$BEC = 7.042$$

Weight: <None>

Min Conc: <None>

75 As [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	14.44	0.0000	P	14.5
2	☐	1.000	1.082	430.01	0.0001	P	6.0
3	☐	50.000	50.765	19688.72	0.0061	P	1.4
4	☐	125.000	127.106	48794.63	0.0153	P	1.1
5	☐	250.000	256.337	98750.46	0.0309	P	0.2
6	☐	500.000	486.198	194030.45	0.0586	P	3.4
7	☐	1000.000	1005.015	393254.23	0.1211	P	1.3
8	☐			481.12	0.0001	P	4.4

$$y = 1.2053E-004 * x + 4.2753E-006$$

$$R = 0.9998$$


$$DL = 0.01546$$

$$BEC = 0.03547$$

Weight: <None>

Min Conc: <None>

77 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	0.00	0.0000	P	
2	☐	5.000	5.623	78.89	0.0000	P	35.8
3	☐	50.000	48.836	690.02	0.0002	P	5.8
4	☐	125.000	131.157	1836.80	0.0006	P	3.6
5	☐	250.000	253.991	3568.25	0.0011	P	5.4
6	☐	500.000	486.225	7080.65	0.0021	P	1.8
7	☐	1000.000	1005.175	14339.26	0.0044	P	3.4
8	☐			24.44	0.0000	P	46.9

$$y = 4.3956E-006 * x + 0.0000E+000$$

$$R = 0.9998$$


$$DL = 0$$

$$BEC = 0$$

Weight: <None>

Min Conc: <None>

78 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	450.01	0.0001	P	4.4
2	☐	5.000	7.157	758.91	0.0002	P	3.5
3	☐	50.000	51.852	2859.20	0.0009	P	4.4
4	☐	125.000	134.543	6669.34	0.0021	P	1.1
5	☐	250.000	261.406	12601.03	0.0039	P	2.8
6	☐	500.000	495.083	24336.81	0.0074	P	3.3
7	☐	1000.000	998.311	47675.71	0.0147	P	1.6
8	☐			584.46	0.0002	P	7.8

$$y = 1.4579E-005 * x + 1.3311E-004$$

$$R = 0.9999$$

$$DL = 1.214$$

$$BEC = 9.13$$

Weight: <None>

Min Conc: <None>

81 Br [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			9494.26		P	1.5
2	☐			9572.08		P	1.9
3	☐			9563.18		P	1.7
4	☐			9492.04		P	1.9
5	☐			9241.87		P	1.9
6	☐			9214.08		P	2.7
7	☐			9214.07		P	2.0
8	☐			9513.15		P	2.8

81 Br [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			234.45		P	18.0
2	☐			225.56		P	17.1
3	☐			193.34		P	19.9
4	☐			195.56		P	17.2
5	☐			234.45		P	18.5
6	☐			184.45		P	10.3
7	☐			120.00		P	14.4
8	☐			217.78		P	4.7

Calibration for 008CAL.S.d

82 Se [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	185.68	0.0000	P	16.6
2	☐	5.000	5.547	787.70	0.0001	P	10.6
3	☐	50.000	52.302	5966.63	0.0006	P	3.2
4	☐	125.000	129.509	14738.30	0.0013	P	1.9
5	☐	250.000	257.187	28635.33	0.0027	P	1.9
6	☐	500.000	513.402	57164.23	0.0053	P	4.4
7	☐	1000.000	990.821	114466.16	0.0102	P	1.3
8	☐			275.98	0.0000	P	39.9

$$y = 1.0295E-005 * x + 1.6232E-005$$

R = 0.9998


DL = 0.784

BEC = 1.577

Weight: <None>


Min Conc: <None>

83 Kr [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐			482.24		P	0.4
2	☐			444.46		P	9.7
3	☐			450.01		P	11.8
4	☐			500.01		P	8.1
5	☐			475.57		P	5.0
6	☐			478.91		P	7.0
7	☐			486.68		P	15.1
8	☐			502.24		P	14.3

86 Sr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	836.70	0.0001	P	3.5
2	☐	1.000	1.059	2769.18	0.0003	P	4.1
3	☐	50.000	50.086	94680.22	0.0088	P	1.8
4	☐	125.000	126.350	241258.24	0.0221	P	1.5
5	☐	250.000	253.419	475455.96	0.0442	P	2.1
6	☐	500.000	505.661	950851.46	0.0882	P	4.5
7	☐	1000.000	996.142	1945652.94	0.1737	A	1.1
8	☐			25285.14	0.0023	P	2.0

$$y = 1.7425E-004 * x + 7.3172E-005$$

$$R = 1.0000$$


$$DL = 0.04367$$

$$BEC = 0.4199$$

Weight: <None>

Min Conc: <None>

88 Sr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	763.36	0.0001	P	6.8
2	☐	1.000	1.031	17639.55	0.0016	P	1.1
3	☐	50.000	49.791	818653.13	0.0761	P	0.8
4	☐	125.000	125.630	2095739.87	0.1919	A	1.6
5	☐	250.000	251.923	4135395.29	0.3847	A	2.8
6	☐	500.000	505.432	8320970.22	0.7718	A	4.8
7	☐	1000.000	996.735	17052614.74	1.5219	A	2.0
8	☐			212726.82	0.0190	P	1.7

$$y = 0.0015 * x + 6.6778E-005$$

$$R = 1.0000$$


$$DL = 0.008871$$

$$BEC = 0.04374$$

Weight: <None>

Min Conc: <None>

90 Zr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	397.79	0.0000	P	8.5
2	☐	1.000	1.019	10586.14	0.0010	P	2.5
3	☐	50.000	49.080	492913.69	0.0458	P	1.3
4	☐	125.000	123.140	1254937.29	0.1149	P	0.7
5	☐	250.000	249.526	2502501.07	0.2328	A	1.1
6	☐	500.000	502.586	5055159.75	0.4688	A	4.5
7	☐	1000.000	999.104	10441880.40	0.9320	A	1.8
8	☐			9512.05	0.0009	P	3.2

$$y = 9.3278E-004 * x + 3.4801E-005$$

$$R = 1.0000$$


$$DL = 0.00956$$

$$BEC = 0.03731$$

Weight: <None>

Min Conc: <None>

91 Zr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	90.00	0.0000	P	29.3
2	☐	1.000	0.974	2232.41	0.0002	P	6.2
3	☐	50.000	49.740	109992.33	0.0102	P	0.8
4	☐	125.000	124.709	279819.87	0.0256	P	1.7
5	☐	250.000	251.263	554798.77	0.0516	P	2.0
6	☐	500.000	502.431	1112501.54	0.1032	P	5.2
7	☐	1000.000	998.518	2297826.18	0.2051	A	1.1
8	☐			2095.73	0.0002	P	1.1

$$y = 2.0538E-004 * x + 7.8613E-006$$

$$R = 1.0000$$


$$DL = 0.03369$$

$$BEC = 0.03828$$

Weight: <None>

Min Conc: <None>

94 Mo [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	295.57	0.0000	P	4.0
2	☐	5.000	6.292	12809.03	0.0012	P	0.6
3	☐	500.000	503.530	1004237.76	0.0933	P	0.5
4	☐	1250.000	1265.425	2561785.13	0.2345	A	1.3
5	☐	2500.000	2510.293	5000520.00	0.4652	A	2.9
6	☐	5000.000	5026.588	10044853.53	0.9315	A	4.2
7	☐	10000.000	9982.027	20727213.86	1.8499	A	1.1
8	☐			7433.06	0.0007	P	4.1

$$y = 1.8532E-004 * x + 2.5838E-005$$

$$R = 1.0000$$


$$DL = 0.01683$$

$$BEC = 0.1394$$

Weight: <None>

Min Conc: <None>

95 Mo [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	296.67	0.0000	P	3.9
2	☐	5.000	5.360	15962.10	0.0015	P	1.5
3	☐	500.000	497.312	1456880.43	0.1354	A	0.7
4	☐	1250.000	1273.772	3787843.66	0.3468	A	0.4
5	☐	2500.000	2534.171	7416081.49	0.6899	A	1.5
6	☐	5000.000	5046.128	14807720.19	1.3737	A	5.6
7	☐	10000.000	9965.556	30397139.82	2.7129	A	2.2
8	☐			6851.67	0.0006	P	1.8

$$y = 2.7223E-004 * x + 2.5936E-005$$

$$R = 1.0000$$


$$DL = 0.01121$$

$$BEC = 0.09527$$

Weight: <None>

Min Conc: <None>

96 Mo [No Gas] ISTD: 89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	314.45	0.0000	P	6.6
2	☐	5.000	5.354	17223.59	0.0016	P	0.9
3	☐	500.000	492.531	1559257.03	0.1449	A	1.7
4	☐	1250.000	1272.393	4088688.65	0.3744	A	1.1
5	☐	2500.000	2545.676	8051004.95	0.7489	A	1.2
6	☐	5000.000	5038.867	15976042.54	1.4824	A	6.3
7	☐	10000.000	9966.721	32852735.62	2.9321	A	2.0
8	☐			12045.08	0.0011	P	1.1

$$y = 2.9419\text{E-}004 * x + 2.7507\text{E-}005$$

R = 1.0000


DL = 0.01857

BEC = 0.0935

Weight: <None>

Min Conc: <None>

97 Mo [No Gas] ISTD: 89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	197.78	0.0000	P	9.7
2	☐	5.000	5.416	9840.08	0.0009	P	3.5
3	☐	500.000	502.397	896833.06	0.0834	P	1.0
4	☐	1250.000	1255.702	2275108.01	0.2083	A	2.4
5	☐	2500.000	2532.515	4516128.30	0.4201	A	1.9
6	☐	5000.000	5071.104	9069446.25	0.8413	A	5.0
7	☐	10000.000	9955.486	18505295.70	1.6515	A	0.9
8	☐			4312.89	0.0004	P	2.8

$$y = 1.6589\text{E-}004 * x + 1.7306\text{E-}005$$

R = 1.0000


DL = 0.03028

BEC = 0.1043

Weight: <None>

Min Conc: <None>

98 Mo [No Gas] ISTD: 89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	431.12	0.0000	P	8.6
2	☐	5.000	5.436	25635.95	0.0024	P	2.0
3	☐	500.000	489.671	2275993.94	0.2115	A	0.4
4	☐	1250.000	1259.469	5941789.28	0.5440	A	1.0
5	☐	2500.000	2532.628	11759742.88	1.0939	A	1.1
6	☐	5000.000	5063.414	23582450.20	2.1870	A	4.4
7	☐	10000.000	9959.468	48197010.39	4.3017	A	1.3
8	☐			10936.41	0.0010	P	1.0

$$y = 4.3192\text{E-}004 * x + 3.7697\text{E-}005$$

$$R = 1.0000$$


$$DL = 0.02249$$

$$BEC = 0.08728$$

Weight: <None>

Min Conc: <None>

106 Cd [No Gas] ISTD: 159 Tb [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2858.09	0.0004	P	5.6
2	☐	1.000	1.097	3080.36	0.0004	P	4.2
3	☐	50.000	52.068	10747.42	0.0014	P	0.6
4	☐	125.000	131.568	22583.24	0.0029	P	0.5
5	☐	250.000	256.378	41491.58	0.0054	P	2.4
6	☐	500.000	517.531	79611.49	0.0104	P	3.5
7	☐	1000.000	988.716	157062.06	0.0195	P	2.0
8	☐			3131.48	0.0004	P	3.4

$$y = 1.9381\text{E-}005 * x + 3.8224\text{E-}004$$

$$R = 0.9997$$


$$DL = 3.29$$

$$BEC = 19.72$$

Weight: <None>

Min Conc: <None>

107 Ag [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	132.23	0.0000	P	4.5
2	☐	1.000	1.083	8120.13	0.0011	P	1.1
3	☐	50.000	50.785	378933.29	0.0491	P	1.8
4	☐	125.000	130.218	968505.96	0.1258	P	1.9
5	☐	250.000	256.128	1917844.09	0.2474	A	3.3
6	☐	500.000	520.571	3844173.48	0.5028	A	3.1
7	☐	1000.000	987.491	7663528.50	0.9537	A	2.1
8	☐			2244.67	0.0003	P	38.5

$$y = 9.6574E-004 * x + 1.7651E-005$$

$$R = 0.9997$$


$$DL = 0.002448$$

$$BEC = 0.01828$$

Weight: <None>

Min Conc: <None>

108 Cd [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	8.89	0.0000	P	86.8
2	☐	1.000	1.016	118.89	0.0000	P	15.3
3	☐	50.000	50.282	5498.87	0.0007	P	2.8
4	☐	125.000	131.067	14274.91	0.0019	P	2.5
5	☐	250.000	254.516	27902.48	0.0036	P	3.3
6	☐	500.000	516.598	55845.04	0.0073	P	3.6
7	☐	1000.000	989.799	112455.22	0.0140	P	1.7
8	☐			100.00	0.0000	P	14.1

$$y = 1.4136E-005 * x + 1.1940E-006$$


$$R = 0.9998$$

$$DL = 0.2199$$

$$BEC = 0.08447$$

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	87.78	0.0000	P	22.1
2	☐	1.000	1.093	7875.56	0.0010	P	2.2
3	☐	50.000	50.829	366265.18	0.0474	P	2.7
4	☐	125.000	129.707	931739.15	0.1210	P	2.2
5	☐	250.000	254.865	1843718.93	0.2378	A	1.3
6	☐	500.000	518.533	3698944.46	0.4837	A	2.3
7	☐	1000.000	988.887	7412962.39	0.9225	A	1.9
8	☐			2104.63	0.0003	P	29.9

$$y = 9.3282E-004 * x + 1.1725E-005$$


$$R = 0.9997$$

$$DL = 0.00832$$

$$BEC = 0.01257$$

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2012.22	0.0003	P	5.9
2	☐	1.000	1.129	3749.81	0.0005	P	1.0
3	☐	50.000	51.211	79837.12	0.0103	P	2.3
4	☐	125.000	131.198	200681.52	0.0261	P	2.1
5	☐	250.000	255.352	391334.94	0.0505	P	2.4
6	☐	500.000	517.722	780305.76	0.1021	P	3.0
7	☐	1000.000	988.965	1564484.21	0.1947	P	2.3
8	☐			2996.52	0.0004	P	1.8

$$y = 1.9660E-004 * x + 2.6914E-004$$


$$R = 0.9997$$

$$DL = 0.2408$$

$$BEC = 1.369$$

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2036.84	0.0003	P	1.7
2	☐	5.000	5.261	26213.95	0.0034	P	1.4
3	☐	50.000	49.580	232130.80	0.0301	P	1.3
4	☐	125.000	127.376	591272.48	0.0768	P	2.0
5	☐	250.000	248.694	1160375.99	0.1497	P	2.6
6	☐	500.000	513.684	2361162.02	0.3088	A	3.8
7	☐	1000.000	993.207	4795976.77	0.5969	A	2.6
8	☐			2463.57	0.0003	P	5.7

$$y = 6.0068E-004 * x + 2.7207E-004$$


$$R = 0.9999$$

$$DL = 0.02254$$

$$BEC = 0.4529$$

Weight: <None>

Min Conc: <None>



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	44.44	0.0000	P	15.4
2	☐	2.000	2.070	11559.18	0.0015	P	2.0
3	☐	50.000	48.637	273686.29	0.0354	P	1.7
4	☐	125.000	124.225	696828.29	0.0905	P	2.2
5	☐	250.000	243.579	1375903.43	0.1774	P	1.9
6	☐	500.000	507.270	2825392.63	0.3695	A	3.3
7	☐	1000.000	998.135	5842781.79	0.7271	A	2.1
8	☐			4773.07	0.0006	P	2.1

$$y = 7.2844E-004 * x + 5.9169E-006$$

$$R = 0.9999$$


$$DL = 0.00376$$

$$BEC = 0.008123$$

Weight: <None>

Min Conc: <None>

135 Ba [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	145.56	0.0000	P	9.4
2	☐	10.000	10.336	13910.23	0.0018	P	1.4
3	☐	250.000	238.821	321774.15	0.0417	P	2.1
4	☐	625.000	615.013	825819.51	0.1073	P	2.8
5	☐	1250.000	1238.737	1675324.28	0.2160	A	1.0
6	☐	2500.000	2567.656	3423352.69	0.4477	A	3.3
7	☐	5000.000	4970.794	6964852.53	0.8668	A	2.4
8	☐			4305.14	0.0005	P	2.5

$$y = 1.7437E-004 * x + 1.9492E-005$$

R = 0.9999


DL = 0.03144

BEC = 0.1118

Weight: <None>

Min Conc: <None>

137 Ba [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	206.67	0.0000	P	21.7
2	☐	10.000	10.485	24087.17	0.0032	P	0.5
3	☐	250.000	244.566	563545.65	0.0730	P	0.9
4	☐	625.000	626.170	1438011.72	0.1868	P	2.4
5	☐	1250.000	1263.197	2921197.70	0.3767	A	2.2
6	☐	2500.000	2543.630	5800288.25	0.7586	A	3.0
7	☐	5000.000	4975.010	11921447.60	1.4836	A	2.5
8	☐			7659.91	0.0010	P	0.7

$$y = 2.9821E-004 * x + 2.7748E-005$$


R = 0.9999

DL = 0.06059


BEC = 0.09305

Weight: <None>

Min Conc: <None>

150 Nd [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			5.56		P	91.6
2	☐			6.67		P	50.0
3	☐			25.55		P	65.7
4	☐			32.22		P	6.0
5	☐			57.78		P	23.3
6	☐			113.33		P	20.6
7	☐			218.89		P	4.9
8	☐			134.45		P	5.2

150 Nd [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			11.11		P	17.3
2	☐			2.22		P	173.
3	☐			7.78		P	49.5
4	☐			13.33		P	43.3
5	☐			17.78		P	54.1
6	☐			30.00		P	57.7
7	☐			78.89		P	8.8
8	☐			77.78		P	29.2

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			10.00		P	66.7
2	☐			7.78		P	107.
3	☐			12.22		P	68.7
4	☐			28.89		P	6.7
5	☐			34.44		P	29.6
6	☐			60.00		P	29.4
7	☐			147.78		P	24.3
8	☐			284.45		P	3.6

156 Dy [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			2.22		P	173.
2	☐			7.78		P	49.5
3	☐			23.33		P	28.6
4	☐			61.11		P	27.5
5	☐			106.67		P	13.6
6	☐			197.78		P	4.9
7	☐			398.90		P	3.9
8	☐			266.67		P	2.5

160 Gd [No Gas] No available calibration poin...

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			45.56		P	48.7
2	☐			37.78		P	65.0
3	☐			54.44		P	15.4
4	☐			71.11		P	7.2
5	☐			61.11		P	22.0
6	☐			101.11		P	24.3
7	☐			133.34		P	24.6
8	☐			371.12		P	11.7

160 Gd [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			414.46		P	14.5
2	☐			405.57		P	3.7
3	☐			392.23		P	3.4
4	☐			431.12		P	14.0
5	☐			381.12		P	7.9
6	☐			391.12		P	11.0
7	☐			446.68		P	22.0
8	☐			593.35		P	11.4


164 Er [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			55.55		P	28.4
2	☐			67.78		P	10.2
3	☐			60.00		P	24.2
4	☐			77.78		P	25.8
5	☐			104.44		P	9.2
6	☐			128.89		P	29.8
7	☐			198.89		P	2.6
8	☐			643.36		P	7.7

164 Er [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			26.67		P	12.5
2	☐			33.33		P	36.1
3	☐			55.55		P	12.5
4	☐			62.22		P	37.6
5	☐			57.78		P	32.8
6	☐			91.11		P	11.2
7	☐			180.01		P	3.2
8	☐			475.57		P	11.1

203 TI [No Gas] ISTD :209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	491.12	0.0001	P	6.2
2	☐	1.000	1.008	5657.89	0.0012	P	3.8
3	☐	50.000	46.969	248332.78	0.0521	P	0.8
4	☐	125.000	118.727	629906.06	0.1316	P	0.9
5	☐	250.000	235.451	1240389.86	0.2610	P	0.5
6	☐	500.000	495.254	2643025.73	0.5488	A	5.3
7	☐	1000.000	1006.946	5406310.09	1.1157	A	1.4
8	☐			7237.54	0.0016	P	1.7

$$y = 0.0011 * x + 1.0539E-004$$

R = 0.9999


DL = 0.01778

BEC = 0.09513

Weight: <None>

Min Conc: <None>

205 TI [No Gas] ISTD :209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1058.94	0.0002	P	6.2
2	☐	1.000	1.000	13279.97	0.0029	P	0.9
3	☐	50.000	46.596	587453.93	0.1233	P	1.0
4	☐	125.000	118.236	1495926.35	0.3126	P	1.6
5	☐	250.000	247.639	3111475.20	0.6545	A	2.1
6	☐	500.000	498.926	6352941.15	1.3185	A	3.8
7	☐	1000.000	1002.143	12831207.17	2.6481	A	2.8
8	☐			17158.73	0.0038	P	1.3

$$y = 0.0026 * x + 2.2723E-004$$

R = 1.0000


DL = 0.01596

BEC = 0.086

Weight: <None>

Min Conc: <None>

206 [Pb] [No Gas] ISTD : 209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	730.03	0.0002	P	2.9
2	☐	1.000	1.010	4970.96	0.0011	P	2.1
3	☐	250.000	233.489	1010835.11	0.2122	P	1.2
4	☐	625.000	620.060	2695583.80	0.5634	A	1.9
5	☐	1250.000	1224.246	5286699.47	1.1122	A	1.4
6	☐	2500.000	2467.839	10799250.67	2.2418	A	4.6
7	☐	5000.000	5023.962	22113909.39	4.5636	A	1.4
8	☐			6834.01	0.0015	P	1.7

$$y = 9.0833E-004 * x + 1.5667E-004$$

$$R = 1.0000$$


$$DL = 0.01494$$

$$BEC = 0.1725$$

Weight: <None>

Min Conc: <None>

207 [Pb] [No Gas] ISTD : 209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	634.47	0.0001	P	4.4
2	☐	1.000	1.004	4306.29	0.0009	P	4.8
3	☐	250.000	231.820	873852.11	0.1835	P	1.0
4	☐	625.000	622.444	2356337.13	0.4924	A	1.0
5	☐	1250.000	1238.979	4658342.47	0.9800	A	0.4
6	☐	2500.000	2479.681	9443417.50	1.9613	A	5.9
7	☐	5000.000	5014.143	19216409.57	3.9657	A	1.8
8	☐			5946.91	0.0013	P	2.0

$$y = 7.9088E-004 * x + 1.3615E-004$$

$$R = 1.0000$$


$$DL = 0.02251$$

$$BEC = 0.1721$$

Weight: <None>

Min Conc: <None>

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2884.59	0.0006	P	2.0
2	☐	1.000	0.999	19641.32	0.0042	P	1.1
3	☐	250.000	237.635	4110585.89	0.8631	A	1.6
4	☐	625.000	623.673	10834128.41	2.2641	A	0.9
5	☐	1250.000	1239.547	21387228.29	4.4994	A	0.9
6	☐	2500.000	2479.440	43340352.55	8.9993	A	5.3
7	☐	5000.000	5013.677	88178067.16	18.1969	A	1.4
8	☐			27301.04	0.0060	P	2.4

$$y = 0.0036 * x + 6.1899E-004$$

$$R = 1.0000$$


$$DL = 0.01045$$

$$BEC = 0.1706$$

Weight: <None>

Min Conc: <None>

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	33.33	0.0000	P	35.4
2	☐	1.000	1.002	16681.71	0.0036	P	2.6
3	☐	50.000	45.627	780274.89	0.1638	P	0.9
4	☐	125.000	122.277	2100844.41	0.4390	A	0.8
5	☐	250.000	245.045	4182255.63	0.8798	A	0.8
6	☐	500.000	492.546	8515777.65	1.7685	A	5.7
7	☐	1000.000	1005.525	17493707.79	3.6103	A	2.3
8	☐			1526.77	0.0003	P	2.3

$$y = 0.0036 * x + 7.1369E-006$$


$$R = 0.9999$$

$$DL = 0.00211$$


$$BEC = 0.001988$$

Weight: <None>


Min Conc: <None>

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		748123.53		P	2.0
2	☐	1.000		773543.95		P	1.2
3	☐	1.000		782676.76		P	1.1
4	☐	1.000		790869.47		P	1.2
5	☐	1.000		785626.64		P	1.1
6	☐	1.000		774382.35		P	2.3
7	☐	1.000		785867.70		P	1.5
8	☐	1.000		798716.29		M	3.2

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3770144.35		A	1.0
2	☐	1.000		3789862.03		A	0.6
3	☐	1.000		3870330.43		A	0.6
4	☐	1.000		3838896.99		A	0.6
5	☐	1.000		3774175.71		A	0.2
6	☐	1.000		3770793.66		A	2.2
7	☐	1.000		3842040.85		A	1.2
8	☐	1.000		3903403.38		A	1.2

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		420191.01		P	1.2
2	☐	1.000		420611.39		P	1.7
3	☐	1.000		429337.95		P	1.0
4	☐	1.000		426723.40		P	0.9
5	☐	1.000		423106.50		P	0.8
6	☐	1.000		436359.71		P	1.1
7	☐	1.000		427346.40		P	0.5
8	☐	1.000		446108.90		P	0.9

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		11437602.61		A	1.2
2	☐	1.000		10747729.07		A	1.5
3	☐	1.000		10759308.94		A	1.0
4	☐	1.000		10922671.51		A	1.0
5	☐	1.000		10750853.59		A	1.2
6	☐	1.000		10794955.88		A	3.9
7	☐	1.000		11205016.50		A	1.0
8	☐	1.000		11185932.47		A	1.7

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3381893.35		A	1.1
2	☐	1.000		3195496.27		A	1.2
3	☐	1.000		3215911.83		A	1.5
4	☐	1.000		3184579.36		A	2.0
5	☐	1.000		3195784.23		A	1.4
6	☐	1.000		3313338.01		A	3.5
7	☐	1.000		3246749.78		A	1.9
8	☐	1.000		3366618.35		A	1.1

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6175629.35		A	1.7
2	☐	1.000		6174383.73		A	1.2
3	☐	1.000		6276530.60		A	0.8
4	☐	1.000		6218468.38		A	0.2
5	☐	1.000		6209017.62		A	1.3
6	☐	1.000		6178190.88		A	3.8
7	☐	1.000		6283700.74		A	1.2
8	☐	1.000		6030690.19		A	1.9

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3983387.47		A	0.4
2	☐	1.000		3963046.96		A	0.4
3	☐	1.000		4020495.22		A	1.5
4	☐	1.000		4001669.39		A	1.4
5	☐	1.000		3984416.78		A	1.0
6	☐	1.000		4072720.81		A	1.7
7	☐	1.000		3995791.82		A	1.3
8	☐	1.000		3898779.70		A	0.4

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6185882.93		A	1.3
2	☐	1.000		6228557.03		A	0.4
3	☐	1.000		6322550.16		A	0.5
4	☐	1.000		6256744.46		A	1.0
5	☐	1.000		6303319.91		A	0.9
6	☐	1.000		6202131.38		A	2.5
7	☐	1.000		6385297.32		A	0.9
8	☐	1.000		6292140.09		A	0.6

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2541458.48		A	0.6
2	☐	1.000		2531247.34		A	1.8
3	☐	1.000		2591594.99		A	0.6
4	☐	1.000		2570651.83		A	1.1
5	☐	1.000		2572539.55		A	0.6
6	☐	1.000		2630128.94		A	1.3
7	☐	1.000		2517083.90		A	1.2
8	☐	1.000		2606104.72		A	1.1

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7484825.09		A	3.7
2	☐	1.000		7636167.66		A	1.4
3	☐	1.000		7724501.00		A	1.1
4	☐	1.000		7702287.73		A	2.0
5	☐	1.000		7755921.06		A	1.8
6	☐	1.000		7649818.50		A	2.3
7	☐	1.000		8038383.56		A	2.3
8	☐	1.000		7975342.80		A	2.2

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5514585.68		A	1.3
2	☐	1.000		5519492.84		A	1.4
3	☐	1.000		5650831.72		A	1.0
4	☐	1.000		5587427.90		A	1.2
5	☐	1.000		5583165.61		A	2.2
6	☐	1.000		5702775.12		A	1.2
7	☐	1.000		5770509.98		A	1.5
8	☐	1.000		5855801.51		A	1.0

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7279091.97		A	0.8
2	☐	1.000		7341027.74		A	1.1
3	☐	1.000		7511921.21		A	1.0
4	☐	1.000		7443987.04		A	1.4
5	☐	1.000		7481026.41		A	0.7
6	☐	1.000		7490551.55		A	3.8
7	☐	1.000		7746536.00		A	0.9
8	☐	1.000		7768534.68		A	1.3

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5535984.05		A	0.6
2	☐	1.000		5448847.11		A	3.2
3	☐	1.000		5592308.81		A	1.5
4	☐	1.000		5608677.62		A	1.4
5	☐	1.000		5592626.58		A	1.2
6	☐	1.000		5773979.22		A	2.2
7	☐	1.000		5766551.09		A	1.8
8	☐	1.000		5870940.05		A	0.7

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4660528.26		A	0.9
2	☐	1.000		4628526.32		A	0.9
3	☐	1.000		4762895.48		A	0.6
4	☐	1.000		4785393.23		A	1.1
5	☐	1.000		4753384.41		A	0.5
6	☐	1.000		4824283.96		A	4.9
7	☐	1.000		4846296.11		A	1.2
8	☐	1.000		4540685.80		A	2.2

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4154512.13		A	2.1
2	☐	1.000		4118972.19		A	1.7
3	☐	1.000		4209074.31		A	1.8
4	☐	1.000		4180950.32		A	0.4
5	☐	1.000		4141088.48		A	0.5
6	☐	1.000		4272960.21		A	1.6
7	☐	1.000		4213701.50		A	1.4
8	☐	1.000		3991117.16		A	0.7

US EPA Tune Check Report

Operator Name Jaswal

Acq/Data Batch D:\Agilent\ICPMH\1\DATA\IP7081525-MS.b

Acq. Date-Time 2025-08-15 11:45:40

Report Comment ---

Instrument Name G8403A JP14410463

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		5335	53353.85			0.734	5.000
24		124321	1243211.78			1.282	5.000
25		15993	159929.38			1.017	5.000
26		19306	193058.70			1.231	5.000
59		76986	769862.87			0.578	5.000
113		9745	97451.74			0.726	5.000
115		113953	1139527.42			1.265	5.000
206		23801	238005.69			2.138	5.000
207		20468	204677.19			2.054	5.000
208		49920	499195.13			2.153	5.000
220		0	1.60			86.715	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	5327	5324	5308	5314	5404
24	121833	124144	125432	125965	124232
25	15988	15955	15822	15939	16261
26	18954	19432	19372	19571	19200
59	76818	76932	76414	77141	77626
113	9712	9653	9758	9756	9846
115	115484	111616	113900	114117	114646
206	23273	23331	24278	24355	23767
207	20068	19967	20901	20704	20699
208	48827	48798	51039	50864	50069

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	0	0	0	0	0

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	9248.66	9.10	8.90 - 9.10	
24	224233.18	24.05	23.90 - 24.10	
25	28496.85	25.00	24.90 - 25.10	
26	33453.05	26.00	25.90 - 26.10	
59	135463.21	59.00	58.90 - 59.10	
113	18775.26	113.00	112.90 - 113.10	
115	218342.94	115.00	114.90 - 115.10	
206	47970.86	206.00	205.90 - 206.10	
207	41239.38	207.00	206.90 - 207.10	
208	100630.97	208.00	207.90 - 208.10	
220			-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.60	0.735	0.900	
24	0.60	0.778	0.900	
25	0.60	0.733	0.900	
26	0.62	0.741	0.900	
59	0.60	0.739	0.900	
113	0.53	0.762	0.900	
115	0.52	0.735	0.900	
206	0.52	0.762	0.900	
207	0.51	0.774	0.900	
208	0.51	0.778	0.900	
220				

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.87 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	7.8 V	Deflect	16.0 V
Extract 2	-150.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Omega Bias	-65 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	170 V		

QP Parameters

Mass Gain	137	Axis Gain	0.9974	QP Bias	-3.0 V
Mass Offset	130	Axis Offset	0.16		

Hardware Settings

Torch

Torch H	-1.9 mm	Torch V	0.2 mm
---------	---------	---------	--------

EM

Discriminator	4.7 mV	Analog HV	2292 V	Pulse HV	1180 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		30146	301463.73			0.636	
89		80973	809733.29			1.449	
205		28684	286843.46			1.266	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	30159	30213	30407	30069	29884
89	79982	81496	82435	79577	81377
205	29221	28788	28718	28363	28332

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.87 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	9.9 V	Deflect	3.0 V
-----------	-------	------------	-------	---------	-------

US EPA Tune Check Report

Extract 2	-195.0 V	Cell Entrance	-40 V	Plate Bias	-60 V
Omega Bias	-95 V	Cell Exit	-60 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	3.9 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		
QP Parameters					
Mass Gain	137	Axis Gain	0.9974	QP Bias	-13.0 V
Mass Offset	130	Axis Offset	0.16		
Hardware Settings					
Torch					
Torch H	-1.9 mm	Torch V	0.2 mm		
EM					
Discriminator	4.7 mV	Analog HV	2292 V	Pulse HV	1180 V

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.27	0.36	-0.09	0.00	N/A	ppb
Antimony	121-1	0.00	0.00	0.00	0.00	N/A	ppb
Arsenic	75-2	0.00	0.01	0.00	0.00	N/A	ppb
Barium	135-1	-0.01	0.00	0.01	0.00	N/A	ppb
Barium	137-1	-0.02	0.00	0.02	0.00	N/A	ppb
Beryllium	9-1	0.00	0.00	0.00	0.00	N/A	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	-0.08	0.05	0.00	N/A	ppb
Cadmium	106-1	-0.14	-1.02	1.16	0.00	N/A	ppb
Cadmium	111-1	-0.02	-0.07	0.09	0.00	N/A	ppb
Calcium	43-1	-0.52	0.55	-0.04	0.00	N/A	ppb
Calcium	44-1	-3.63	0.45	3.18	0.00	N/A	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.07	0.01	0.06	0.00	N/A	ppb
Cobalt	59-2	0.00	0.01	0.00	0.00	N/A	ppb
Copper	63-2	0.00	-0.02	0.01	0.00	N/A	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				100		%
Indium	115-2				100		%
Iron	56-2	0.28	-0.14	-0.14	0.00	N/A	ppb
Iron	57-2	-0.43	0.87	-0.45	0.00	N/A	ppb
Iron	54-2	-0.34	0.51	-0.17	0.00	N/A	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.00	0.01	0.00	0.00	N/A	ppb
Lead	207-1	-0.01	0.00	0.01	0.00	N/A	ppb
Lead	208-1	0.00	0.00	0.00	0.00	N/A	ppb
Lithium	6-1				100		%
Magnesium	24-2	-0.18	-0.35	0.53	0.00	N/A	ppb
Manganese	55-2	-0.01	0.01	0.00	0.00	N/A	ppb
Molybdenum	94-1	0.00	0.01	-0.01	0.00	N/A	ppb
Molybdenum	95-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	96-1	-0.01	0.00	0.01	0.00	N/A	ppb
Molybdenum	97-1	-0.01	0.00	0.01	0.00	N/A	ppb
Molybdenum	98-1	0.00	0.01	0.00	0.00	N/A	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.03	0.04	-0.02	0.00	N/A	ppb
Phosphorus	31-2	-17.46	-14.84	-16.31	-16.20		ppb
Potassium	39-2	-2.15	-1.18	3.32	0.00	N/A	ppb
Rhodium	103-1				100		%
Rhodium	103-2				100		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	0.16	-0.30	0.14	0.00	N/A	ppb
Selenium	77-2	0.00	0.00	0.00	0.00	N/A	ppb
Selenium	78-2	-0.14	0.46	-0.31	0.00	N/A	ppb
Silicon	28-1	-0.59	0.16	0.43	0.00	N/A	ppb
Silver	107-1	0.00	0.00	0.00	0.00	N/A	ppb
Silver	109-1	0.00	0.00	0.00	0.00	N/A	ppb
Sodium	23-2	-0.63	0.09	0.54	0.00	N/A	ppb
Strontium	86-1	-0.01	-0.01	0.02	0.00	N/A	ppb
Strontium	88-1	0.00	0.00	0.00	0.00	N/A	ppb
Sulfur	34-1	-141.32	24.22	-27.41	-48.17		ppb
Terbium	159-1				100		%
Terbium	159-2				100		%
Thallium	203-1	0.01	0.00	-0.01	0.00	N/A	ppb
Thallium	205-1	-0.01	0.00	0.00	0.00	N/A	ppb
Tin	118-1	0.01	-0.01	0.00	0.00	N/A	ppb
Titanium	47-1	-0.03	0.01	0.02	0.00	N/A	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	N/A	ppb
Vanadium	51-2	-0.01	0.00	0.01	0.00	N/A	ppb
Yttrium	89-1				100		%
Yttrium	89-2				100		%
Zinc	66-2	-0.14	-0.01	0.16	0.00	N/A	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	N/A	ppb
Zirconium	91-1	0.00	0.01	-0.01	0.00	N/A	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	19.68	19.37	19.21	19.42	1.23	ppb
Antimony	121-1	2.06	2.03	2.12	2.07	2.03	ppb
Arsenic	75-2	1.16	1.04	1.05	1.08	6.15	ppb
Barium	135-1	10.23	10.28	10.50	10.34	1.37	ppb
Barium	137-1	10.43	10.52	10.51	10.48	0.46	ppb
Beryllium	9-1	1.13	1.22	1.07	1.14	6.35	ppb
Bismuth	209-1				99		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.13	1.09	0.82	1.02	16.58	ppb
Cadmium	106-1	0.67	0.53	2.09	1.10	79.00	ppb
Cadmium	111-1	1.16	1.10	1.13	1.13	2.32	ppb
Calcium	43-1	573.08	584.85	574.20	577.38	1.13	ppb
Calcium	44-1	569.28	579.15	575.41	574.61	0.87	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.05	2.28	2.15	2.16	5.26	ppb
Cobalt	59-2	1.13	1.17	1.15	1.15	1.71	ppb
Copper	63-2	2.04	2.10	2.13	2.09	2.37	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				98		%
Indium	115-1				101		%
Indium	115-2				100		%
Iron	56-2	55.54	57.94	57.91	57.13	2.41	ppb
Iron	57-2	62.77	55.00	58.30	58.69	6.65	ppb
Iron	54-2	57.09	59.10	58.21	58.13	1.73	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.00	1.04	0.99	1.01	2.41	ppb
Lead	207-1	0.97	0.97	1.07	1.00	5.62	ppb
Lead	208-1	1.01	0.99	1.00	1.00	1.34	ppb
Lithium	6-1				103		%
Magnesium	24-2	527.51	532.16	533.35	531.01	0.58	ppb
Manganese	55-2	1.01	1.03	0.99	1.01	1.60	ppb
Molybdenum	94-1	6.32	6.25	6.31	6.29	0.64	ppb
Molybdenum	95-1	5.43	5.28	5.37	5.36	1.48	ppb
Molybdenum	96-1	5.30	5.39	5.37	5.35	0.92	ppb
Molybdenum	97-1	5.20	5.57	5.48	5.42	3.52	ppb
Molybdenum	98-1	5.35	5.39	5.56	5.44	2.05	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.04	0.98	1.01	1.01	2.98	ppb
Phosphorus	31-2	10.95	21.98	15.67	16.20	34.16	ppb
Potassium	39-2	562.28	584.74	571.23	572.75	1.97	ppb
Rhodium	103-1				100		%
Rhodium	103-2				99		%
Scandium	45-1				101		%
Scandium	45-2				100		%
Selenium	82-1	5.61	4.76	6.27	5.55	13.65	ppb
Selenium	77-2	6.74	6.83	3.30	5.62	35.75	ppb
Selenium	78-2	6.69	6.99	7.80	7.16	8.01	ppb
Silicon	28-1	0.18	0.84	0.48	0.50	65.97	ppb
Silver	107-1	1.09	1.07	1.09	1.08	1.10	ppb
Silver	109-1	1.09	1.07	1.12	1.09	2.25	ppb
Sodium	23-2	541.37	557.09	548.64	549.03	1.43	ppb
Strontium	86-1	1.00	1.05	1.12	1.06	5.72	ppb
Strontium	88-1	1.02	1.04	1.04	1.03	1.12	ppb
Sulfur	34-1	13.85	123.57	7.10	48.17	135.73	ppb
Terbium	159-1				102		%
Terbium	159-2				100		%
Thallium	203-1	1.06	0.98	0.99	1.01	4.13	ppb
Thallium	205-1	1.00	0.99	1.01	1.00	0.97	ppb
Tin	118-1	5.34	5.25	5.19	5.26	1.50	ppb
Titanium	47-1	5.21	5.39	5.18	5.26	2.15	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.02	0.97	1.01	1.00	2.58	ppb
Vanadium	51-2	5.52	5.50	5.55	5.52	0.41	ppb
Yttrium	89-1				94		%
Yttrium	89-2				94		%
Zinc	66-2	4.55	4.41	4.53	4.50	1.73	ppb
Zirconium	90-1	1.01	1.00	1.05	1.02	2.63	ppb
Zirconium	91-1	0.92	0.96	1.04	0.97	6.39	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	996.12	991.36	1002.36	996.61	0.55	ppb
Antimony	121-1	47.94	49.56	48.41	48.64	1.72	ppb
Arsenic	75-2	51.55	50.29	50.46	50.77	1.35	ppb
Barium	135-1	234.76	244.44	237.27	238.82	2.10	ppb
Barium	137-1	242.19	246.44	245.06	244.57	0.89	ppb
Beryllium	9-1	51.98	51.66	52.87	52.17	1.21	ppb
Bismuth	209-1				102		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	49.71	51.87	49.26	50.28	2.78	ppb
Cadmium	106-1	51.63	52.08	52.49	52.07	0.82	ppb
Cadmium	111-1	50.28	52.60	50.75	51.21	2.39	ppb
Calcium	43-1	5362.11	5301.51	5291.97	5318.53	0.72	ppb
Calcium	44-1	4983.33	4992.59	5140.78	5038.90	1.75	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	49.19	50.44	50.43	50.02	1.43	ppb
Cobalt	59-2	49.94	49.03	50.86	49.94	1.83	ppb
Copper	63-2	505.44	497.13	510.05	504.21	1.30	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				101		%
Indium	115-1				102		%
Indium	115-2				102		%
Iron	56-2	2602.28	2602.85	2641.22	2615.45	0.85	ppb
Iron	57-2	2704.31	2675.82	2742.17	2707.43	1.23	ppb
Iron	54-2	2797.92	2782.50	2826.58	2802.33	0.80	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	235.87	234.20	230.40	233.49	1.20	ppb
Lead	207-1	232.64	233.64	229.18	231.82	1.01	ppb
Lead	208-1	237.16	241.55	234.20	237.63	1.56	ppb
Lithium	6-1				105		%
Magnesium	24-2	5064.74	5121.42	5095.66	5093.94	0.56	ppb
Manganese	55-2	499.45	493.56	503.79	498.93	1.03	ppb
Molybdenum	94-1	500.88	503.80	505.91	503.53	0.50	ppb
Molybdenum	95-1	493.75	500.14	498.05	497.31	0.66	ppb
Molybdenum	96-1	483.50	499.41	494.68	492.53	1.66	ppb
Molybdenum	97-1	497.43	502.21	507.56	502.40	1.01	ppb
Molybdenum	98-1	488.64	491.93	488.44	489.67	0.40	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	50.20	50.37	52.07	50.88	2.03	ppb
Phosphorus	31-2	982.24	951.79	968.74	967.59	1.58	ppb
Potassium	39-2	2624.45	2616.67	2686.09	2642.40	1.44	ppb
Rhodium	103-1				102		%
Rhodium	103-2				101		%
Scandium	45-1				103		%
Scandium	45-2				102		%
Selenium	82-1	50.87	51.80	54.24	52.30	3.33	ppb
Selenium	77-2	50.53	45.59	50.38	48.84	5.75	ppb
Selenium	78-2	52.83	53.93	48.80	51.85	5.21	ppb
Silicon	28-1	62.42	64.42	64.45	63.76	1.83	ppb
Silver	107-1	49.81	51.68	50.86	50.78	1.85	ppb
Silver	109-1	49.26	51.80	51.42	50.83	2.69	ppb
Sodium	23-2	5185.19	5195.98	5281.60	5220.92	1.01	ppb
Strontium	86-1	49.13	50.20	50.93	50.09	1.80	ppb
Strontium	88-1	49.33	49.91	50.14	49.79	0.83	ppb
Sulfur	34-1	907.40	906.79	967.45	927.21	3.76	ppb
Terbium	159-1				103		%
Terbium	159-2				102		%
Thallium	203-1	47.21	47.16	46.54	46.97	0.80	ppb
Thallium	205-1	46.89	46.85	46.05	46.60	1.02	ppb
Tin	118-1	48.89	50.21	49.64	49.58	1.33	ppb
Titanium	47-1	482.22	481.74	484.56	482.84	0.31	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	45.79	45.91	45.18	45.63	0.85	ppb
Vanadium	51-2	48.99	48.76	49.39	49.04	0.65	ppb
Yttrium	89-1				94		%
Yttrium	89-2				95		%
Zinc	66-2	514.14	507.92	522.71	514.92	1.44	ppb
Zirconium	90-1	48.37	49.27	49.60	49.08	1.30	ppb
Zirconium	91-1	49.49	49.50	50.23	49.74	0.85	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2517.06	2551.95	2564.30	2544.44	0.96	ppb
Antimony	121-1	121.75	127.08	123.85	124.22	2.16	ppb
Arsenic	75-2	126.77	125.85	128.70	127.11	1.15	ppb
Barium	135-1	600.31	633.87	610.86	615.01	2.79	ppb
Barium	137-1	610.49	640.08	627.94	626.17	2.38	ppb
Beryllium	9-1	128.56	129.56	127.12	128.42	0.95	ppb
Bismuth	209-1				103		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	127.43	133.52	132.25	131.07	2.45	ppb
Cadmium	106-1	130.64	132.11	131.96	131.57	0.61	ppb
Cadmium	111-1	128.11	133.67	131.80	131.20	2.16	ppb
Calcium	43-1	13582.66	13785.25	13660.92	13676.28	0.75	ppb
Calcium	44-1	12901.24	13200.65	12903.21	13001.70	1.33	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	126.66	129.00	127.14	127.60	0.97	ppb
Cobalt	59-2	126.78	128.32	126.43	127.18	0.79	ppb
Copper	63-2	1249.82	1320.73	1260.77	1277.11	2.99	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				101		%
Indium	115-1				101		%
Indium	115-2				101		%
Iron	56-2	6610.46	6777.37	6613.75	6667.19	1.43	ppb
Iron	57-2	6829.26	7040.73	6786.57	6885.52	1.98	ppb
Iron	54-2	7064.65	7099.11	7084.26	7082.67	0.24	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	606.57	628.50	625.11	620.06	1.90	ppb
Lead	207-1	620.44	617.76	629.13	622.44	0.95	ppb
Lead	208-1	617.67	628.77	624.58	623.67	0.90	ppb
Lithium	6-1				106		%
Magnesium	24-2	13165.97	13108.00	13182.43	13152.13	0.30	ppb
Manganese	55-2	1215.10	1291.07	1267.76	1257.98	3.09	ppb
Molybdenum	94-1	1246.89	1270.65	1278.74	1265.43	1.31	ppb
Molybdenum	95-1	1269.51	1272.35	1279.46	1273.77	0.40	ppb
Molybdenum	96-1	1281.08	1279.57	1256.53	1272.39	1.08	ppb
Molybdenum	97-1	1252.28	1287.01	1227.81	1255.70	2.37	ppb
Molybdenum	98-1	1257.52	1273.26	1247.62	1259.47	1.03	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	128.21	130.53	130.70	129.81	1.07	ppb
Phosphorus	31-2	2572.29	2551.29	2504.04	2542.54	1.37	ppb
Potassium	39-2	6487.45	6660.94	6453.24	6533.88	1.70	ppb
Rhodium	103-1				101		%
Rhodium	103-2				100		%
Scandium	45-1				102		%
Scandium	45-2				102		%
Selenium	82-1	129.23	132.14	127.15	129.51	1.94	ppb
Selenium	77-2	134.24	133.57	125.66	131.16	3.64	ppb
Selenium	78-2	133.77	133.50	136.35	134.54	1.17	ppb
Silicon	28-1	142.03	147.85	142.77	144.22	2.20	ppb
Silver	107-1	128.02	132.87	129.75	130.22	1.89	ppb
Silver	109-1	127.42	132.94	128.75	129.71	2.22	ppb
Sodium	23-2	13230.83	13311.10	13582.90	13374.94	1.38	ppb
Strontium	86-1	126.39	128.24	124.43	126.35	1.51	ppb
Strontium	88-1	124.75	127.88	124.25	125.63	1.57	ppb
Sulfur	34-1	2585.02	2644.03	2515.13	2581.39	2.50	ppb
Terbium	159-1				103		%
Terbium	159-2				101		%
Thallium	203-1	117.66	119.71	118.81	118.73	0.87	ppb
Thallium	205-1	116.27	120.06	118.38	118.24	1.61	ppb
Tin	118-1	125.70	130.36	126.07	127.38	2.03	ppb
Titanium	47-1	1237.40	1247.89	1229.87	1238.39	0.73	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	121.14	122.63	123.07	122.28	0.83	ppb
Vanadium	51-2	123.97	126.49	123.68	124.72	1.24	ppb
Yttrium	89-1				95		%
Yttrium	89-2				94		%
Zinc	66-2	1307.00	1313.11	1287.93	1302.68	1.01	ppb
Zirconium	90-1	123.09	124.00	122.33	123.14	0.68	ppb
Zirconium	91-1	125.68	126.19	122.26	124.71	1.71	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5037.90	5065.59	5025.50	5043.00	0.41	ppb
Antimony	121-1	240.03	248.75	241.95	243.58	1.88	ppb
Arsenic	75-2	255.69	256.97	256.35	256.34	0.25	ppb
Barium	135-1	1224.30	1242.50	1249.41	1238.74	1.05	ppb
Barium	137-1	1234.31	1291.05	1264.23	1263.20	2.25	ppb
Beryllium	9-1	250.99	259.55	256.47	255.67	1.70	ppb
Bismuth	209-1				102		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	249.70	264.07	249.78	254.52	3.25	ppb
Cadmium	106-1	252.85	264.08	252.21	256.38	2.60	ppb
Cadmium	111-1	251.10	262.40	252.56	255.35	2.41	ppb
Calcium	43-1	27082.31	27222.69	27014.27	27106.42	0.39	ppb
Calcium	44-1	25719.03	26191.97	25969.76	25960.25	0.91	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	250.98	250.91	256.13	252.68	1.18	ppb
Cobalt	59-2	250.32	252.36	254.42	252.37	0.81	ppb
Copper	63-2	2554.89	2554.82	2552.39	2554.04	0.06	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				101		%
Indium	115-1				102		%
Indium	115-2				101		%
Iron	56-2	13193.72	13438.49	13271.65	13301.29	0.94	ppb
Iron	57-2	13729.71	13904.60	13796.46	13810.26	0.64	ppb
Iron	54-2	13553.12	13494.82	13552.35	13533.43	0.25	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1205.98	1228.38	1238.38	1224.25	1.36	ppb
Lead	207-1	1237.27	1244.54	1235.13	1238.98	0.40	ppb
Lead	208-1	1226.87	1247.17	1244.60	1239.55	0.89	ppb
Lithium	6-1				105		%
Magnesium	24-2	26045.03	26517.86	25546.28	26036.39	1.87	ppb
Manganese	55-2	2514.84	2542.55	2511.94	2523.11	0.67	ppb
Molybdenum	94-1	2427.25	2550.99	2552.64	2510.29	2.86	ppb
Molybdenum	95-1	2489.79	2562.85	2549.87	2534.17	1.54	ppb
Molybdenum	96-1	2512.12	2575.27	2549.64	2545.68	1.25	ppb
Molybdenum	97-1	2479.44	2569.29	2548.82	2532.52	1.86	ppb
Molybdenum	98-1	2509.50	2565.28	2523.10	2532.63	1.15	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	254.66	258.67	254.40	255.91	0.93	ppb
Phosphorus	31-2	5123.71	5142.41	4955.39	5073.84	2.03	ppb
Potassium	39-2	13019.71	13032.95	12954.86	13002.51	0.32	ppb
Rhodium	103-1				101		%
Rhodium	103-2				100		%
Scandium	45-1				100		%
Scandium	45-2				101		%
Selenium	82-1	251.94	261.53	258.09	257.19	1.89	ppb
Selenium	77-2	256.08	266.43	239.46	253.99	5.36	ppb
Selenium	78-2	252.85	267.58	263.79	261.41	2.93	ppb
Silicon	28-1	270.52	275.28	271.87	272.56	0.90	ppb
Silver	107-1	250.20	265.83	252.36	256.13	3.31	ppb
Silver	109-1	251.92	258.57	254.10	254.86	1.33	ppb
Sodium	23-2	26785.64	26548.16	26210.26	26514.69	1.09	ppb
Strontium	86-1	247.57	257.86	254.83	253.42	2.09	ppb
Strontium	88-1	246.63	259.93	249.21	251.92	2.80	ppb
Sulfur	34-1	5152.02	5126.48	5100.65	5126.38	0.50	ppb
Terbium	159-1				104		%
Terbium	159-2				101		%
Thallium	203-1	234.90	236.87	234.59	235.45	0.52	ppb
Thallium	205-1	250.89	241.72	250.31	247.64	2.07	ppb
Tin	118-1	244.15	256.26	245.68	248.69	2.65	ppb
Titanium	47-1	2455.41	2527.68	2522.40	2501.83	1.61	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	243.27	244.63	247.24	245.04	0.82	ppb
Vanadium	51-2	249.99	250.97	251.32	250.76	0.27	ppb
Yttrium	89-1				94		%
Yttrium	89-2				94		%
Zinc	66-2	2645.68	2579.96	2596.97	2607.54	1.31	ppb
Zirconium	90-1	246.65	250.14	251.79	249.53	1.05	ppb
Zirconium	91-1	245.59	254.24	253.96	251.26	1.95	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9731.37	9708.12	10005.30	9814.93	1.68	ppb
Antimony	121-1	491.11	524.86	505.84	507.27	3.34	ppb
Arsenic	75-2	467.74	491.48	499.38	486.20	3.39	ppb
Barium	135-1	2481.16	2650.75	2571.06	2567.66	3.30	ppb
Barium	137-1	2465.74	2616.31	2548.84	2543.63	2.97	ppb
Beryllium	9-1	501.50	527.35	511.89	513.58	2.53	ppb
Bismuth	209-1				104		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	501.93	537.37	510.49	516.60	3.58	ppb
Cadmium	106-1	497.26	534.81	520.51	517.53	3.66	ppb
Cadmium	111-1	500.25	531.08	521.83	517.72	3.06	ppb
Calcium	43-1	52750.44	56190.16	54201.97	54380.86	3.18	ppb
Calcium	44-1	50040.11	53541.74	51340.84	51640.90	3.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	479.73	479.19	497.99	485.64	2.20	ppb
Cobalt	59-2	487.15	482.99	498.80	489.65	1.67	ppb
Copper	63-2	4938.43	4889.92	4920.13	4916.16	0.50	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				100		%
Indium	115-2				103		%
Iron	56-2	25365.60	25314.07	26021.56	25567.08	1.54	ppb
Iron	57-2	25507.62	24991.01	26259.70	25586.11	2.49	ppb
Iron	54-2	25775.09	25658.24	26479.61	25970.98	1.71	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2386.07	2598.40	2419.05	2467.84	4.63	ppb
Lead	207-1	2348.52	2639.23	2451.29	2479.68	5.95	ppb
Lead	208-1	2363.14	2623.17	2452.01	2479.44	5.33	ppb
Lithium	6-1				104		%
Magnesium	24-2	49392.92	49057.79	51700.82	50050.51	2.88	ppb
Manganese	55-2	4873.87	4850.77	4916.46	4880.37	0.68	ppb
Molybdenum	94-1	4795.06	5211.47	5073.23	5026.59	4.22	ppb
Molybdenum	95-1	4730.12	5267.41	5140.85	5046.13	5.57	ppb
Molybdenum	96-1	4694.75	5324.37	5097.47	5038.87	6.33	ppb
Molybdenum	97-1	4795.74	5291.32	5126.25	5071.10	4.98	ppb
Molybdenum	98-1	4852.22	5293.31	5044.71	5063.41	4.37	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	496.97	485.72	492.28	491.65	1.15	ppb
Phosphorus	31-2	9742.72	9495.21	9862.59	9700.17	1.93	ppb
Potassium	39-2	24968.01	24735.07	24996.70	24899.93	0.58	ppb
Rhodium	103-1				100		%
Rhodium	103-2				102		%
Scandium	45-1				100		%
Scandium	45-2				104		%
Selenium	82-1	489.08	534.16	516.96	513.40	4.43	ppb
Selenium	77-2	485.96	477.58	495.14	486.23	1.81	ppb
Selenium	78-2	475.90	502.34	507.01	495.08	3.39	ppb
Silicon	28-1	497.98	544.16	535.42	525.85	4.67	ppb
Silver	107-1	502.04	530.49	529.18	520.57	3.09	ppb
Silver	109-1	504.63	526.74	524.23	518.53	2.34	ppb
Sodium	23-2	51274.91	50415.20	51824.81	51171.64	1.39	ppb
Strontium	86-1	482.68	528.21	506.10	505.66	4.50	ppb
Strontium	88-1	479.29	527.56	509.45	505.43	4.82	ppb
Sulfur	34-1	9564.78	10637.73	10184.61	10129.04	5.32	ppb
Terbium	159-1				102		%
Terbium	159-2				103		%
Thallium	203-1	475.22	524.94	485.61	495.25	5.30	ppb
Thallium	205-1	481.03	519.02	496.73	498.93	3.83	ppb
Tin	118-1	493.73	532.26	515.06	513.68	3.76	ppb
Titanium	47-1	4856.90	5168.06	4916.37	4980.44	3.32	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	464.26	520.62	492.76	492.55	5.72	ppb
Vanadium	51-2	478.80	482.72	497.66	486.40	2.05	ppb
Yttrium	89-1				94		%
Yttrium	89-2				98		%
Zinc	66-2	4891.10	4980.82	5083.12	4985.01	1.93	ppb
Zirconium	90-1	478.91	523.84	505.00	502.59	4.49	ppb
Zirconium	91-1	473.49	524.45	509.35	502.43	5.21	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	20164.04	20167.90	20422.96	20251.64	0.73	ppb
Antimony	121-1	988.13	984.05	1022.22	998.14	2.10	ppb
Arsenic	75-2	997.32	997.22	1020.51	1005.02	1.34	ppb
Barium	135-1	4905.82	4900.62	5105.94	4970.79	2.36	ppb
Barium	137-1	4886.78	4919.77	5118.48	4975.01	2.52	ppb
Beryllium	9-1	980.66	1001.56	991.55	991.26	1.05	ppb
Bismuth	209-1				104		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	970.51	994.88	1004.00	989.80	1.75	ppb
Cadmium	106-1	978.75	975.68	1011.72	988.72	2.02	ppb
Cadmium	111-1	974.32	976.84	1015.74	988.97	2.35	ppb
Calcium	43-1	102111.58	104994.72	104427.29	103844.53	1.47	ppb
Calcium	44-1	102210.20	104695.36	104605.80	103837.12	1.36	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	983.98	1026.23	1008.35	1006.19	2.11	ppb
Cobalt	59-2	1001.52	998.09	1013.33	1004.32	0.80	ppb
Copper	63-2	10071.38	10083.45	9919.62	10024.81	0.91	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				104		%
Indium	115-1				103		%
Indium	115-2				99		%
Iron	56-2	52695.20	53685.48	51889.08	52756.59	1.71	ppb
Iron	57-2	51932.05	52317.29	52592.51	52280.62	0.63	ppb
Iron	54-2	53224.29	54693.13	54380.04	54099.16	1.43	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	5038.50	5086.17	4947.21	5023.96	1.41	ppb
Lead	207-1	4967.48	5120.27	4954.68	5014.14	1.84	ppb
Lead	208-1	5005.61	5086.52	4948.90	5013.68	1.38	ppb
Lithium	6-1				105		%
Magnesium	24-2	103192.15	102305.52	103161.27	102886.31	0.49	ppb
Manganese	55-2	9907.64	10102.79	10148.85	10053.09	1.27	ppb
Molybdenum	94-1	9951.59	10099.90	9894.59	9982.03	1.06	ppb
Molybdenum	95-1	9960.22	10185.64	9750.81	9965.56	2.18	ppb
Molybdenum	96-1	9915.05	10184.14	9800.97	9966.72	1.97	ppb
Molybdenum	97-1	9984.39	10025.60	9856.47	9955.49	0.89	ppb
Molybdenum	98-1	9859.16	10110.01	9909.24	9959.47	1.33	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	999.46	1006.24	1000.45	1002.05	0.37	ppb
Phosphorus	31-2	20139.88	20072.79	20170.60	20127.76	0.25	ppb
Potassium	39-2	51938.74	50798.00	51690.18	51475.64	1.17	ppb
Rhodium	103-1				102		%
Rhodium	103-2				100		%
Scandium	45-1				102		%
Scandium	45-2				102		%
Selenium	82-1	976.75	1002.68	993.03	990.82	1.32	ppb
Selenium	77-2	974.74	998.82	1041.97	1005.18	3.39	ppb
Selenium	78-2	979.83	1003.62	1011.48	998.31	1.65	ppb
Silicon	28-1	954.63	998.52	982.17	978.44	2.27	ppb
Silver	107-1	971.43	979.61	1011.43	987.49	2.14	ppb
Silver	109-1	975.21	981.30	1010.15	988.89	1.89	ppb
Sodium	23-2	106242.12	104958.56	104425.37	105208.69	0.89	ppb
Strontium	86-1	985.92	1007.58	994.93	996.14	1.09	ppb
Strontium	88-1	993.16	1017.85	979.19	996.73	1.96	ppb
Sulfur	34-1	19829.77	19825.69	20036.59	19897.35	0.61	ppb
Terbium	159-1				107		%
Terbium	159-2				105		%
Thallium	203-1	1014.52	1015.71	990.61	1006.95	1.41	ppb
Thallium	205-1	1017.99	1018.51	969.93	1002.14	2.78	ppb
Tin	118-1	974.38	982.12	1023.12	993.21	2.64	ppb
Titanium	47-1	9830.76	9962.75	10241.38	10011.63	2.09	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1017.22	1020.73	978.62	1005.52	2.32	ppb
Vanadium	51-2	998.75	1012.77	1008.56	1006.69	0.71	ppb
Yttrium	89-1				98		%
Yttrium	89-2				96		%
Zinc	66-2	10065.85	9989.75	9864.23	9973.28	1.02	ppb
Zirconium	90-1	985.92	1019.98	991.41	999.10	1.83	ppb
Zirconium	91-1	986.72	1008.40	1000.44	998.52	1.10	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	101442.06	99304.92	99147.88	99964.95	1.28	ppb
Antimony	121-1	0.79	0.83	0.82	0.81	2.16	ppb
Arsenic	75-2	1.10	1.21	1.14	1.15	4.57	ppb
Barium	135-1	2.91	3.07	2.98	2.98	2.62	ppb
Barium	137-1	3.10	3.13	3.15	3.13	0.75	ppb
Beryllium	9-1	0.18	0.15	0.19	0.17	11.22	ppb
Bismuth	209-1				97		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.67	0.83	0.91	0.80	15.59	ppb
Cadmium	106-1	0.18	0.12	1.33	0.55	125.42	ppb
Cadmium	111-1	0.52	0.52	0.58	0.54	6.23	ppb
Calcium	43-1	499371.02	495545.01	501049.02	498655.02	0.57	ppb
Calcium	44-1	501010.15	495448.76	500563.49	499007.47	0.62	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4.49	4.39	4.52	4.47	1.50	ppb
Cobalt	59-2	1.87	1.80	1.91	1.86	3.03	ppb
Copper	63-2	1.71	1.66	1.69	1.69	1.47	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				106		%
Indium	115-1				102		%
Indium	115-2				103		%
Iron	56-2	250551.40	242658.59	254810.98	249340.33	2.47	ppb
Iron	57-2	250051.05	246534.64	251619.68	249401.79	1.04	ppb
Iron	54-2	248983.18	246380.88	251658.62	249007.56	1.06	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.45	1.50	1.50	1.48	1.92	ppb
Lead	207-1	1.45	1.51	1.50	1.48	2.26	ppb
Lead	208-1	1.44	1.52	1.50	1.49	2.67	ppb
Lithium	6-1				107		%
Magnesium	24-2	511671.74	484278.10	502095.93	499348.59	2.78	ppb
Manganese	55-2	4.94	4.80	4.86	4.87	1.41	ppb
Molybdenum	94-1	3.50	3.28	3.56	3.45	4.23	ppb
Molybdenum	95-1	2.19	2.11	2.16	2.15	1.88	ppb
Molybdenum	96-1	3.58	3.52	3.60	3.57	1.17	ppb
Molybdenum	97-1	2.29	2.19	2.17	2.22	2.96	ppb
Molybdenum	98-1	2.15	2.19	2.19	2.18	1.02	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	3.64	3.65	3.93	3.74	4.44	ppb
Phosphorus	31-2	-25.81	-20.91	-30.13	-25.62		ppb
Potassium	39-2	250207.07	245796.43	253039.76	249681.09	1.46	ppb
Rhodium	103-1				98		%
Rhodium	103-2				98		%
Scandium	45-1				104		%
Scandium	45-2				106		%
Selenium	82-1	1.74	0.84	-0.16	0.81	117.69	ppb
Selenium	77-2	1.58	2.45	0.91	1.65	46.86	ppb
Selenium	78-2	1.71	3.28	3.34	2.78	33.42	ppb
Silicon	28-1	3.98	4.37	5.30	4.55	14.96	ppb
Silver	107-1	0.40	0.21	0.21	0.27	41.11	ppb
Silver	109-1	0.36	0.24	0.20	0.27	31.28	ppb
Sodium	23-2	500484.59	497225.58	498513.52	498741.23	0.33	ppb
Strontium	86-1	12.27	12.77	12.63	12.55	2.03	ppb
Strontium	88-1	12.21	12.40	12.63	12.41	1.70	ppb
Sulfur	34-1	-491.84	-557.71	-478.77	-509.44		ppb
Terbium	159-1				107		%
Terbium	159-2				106		%
Thallium	203-1	1.34	1.37	1.32	1.34	1.86	ppb
Thallium	205-1	1.32	1.36	1.35	1.34	1.39	ppb
Tin	118-1	0.09	0.04	0.05	0.06	47.66	ppb
Titanium	47-1	1.26	1.29	1.28	1.28	1.14	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.09	0.09	0.09	0.09	2.39	ppb
Vanadium	51-2	0.46	0.49	0.44	0.47	5.56	ppb
Yttrium	89-1				98		%
Yttrium	89-2				100		%
Zinc	66-2	5.16	4.58	4.75	4.83	6.17	ppb
Zirconium	90-1	0.85	0.87	0.90	0.87	3.33	ppb
Zirconium	91-1	0.88	0.86	0.87	0.87	1.12	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1511.84	1484.44	1500.70	1499.00	0.92	ppb
Antimony	121-1	401.10	401.21	401.80	401.37	0.09	ppb
Arsenic	75-2	389.12	386.06	384.20	386.46	0.64	ppb
Barium	135-1	1569.19	1570.55	1606.82	1582.18	1.35	ppb
Barium	137-1	1614.54	1578.62	1628.45	1607.21	1.60	ppb
Beryllium	9-1	39.04	38.65	39.47	39.05	1.05	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	184.45	183.56	181.84	183.29	0.72	ppb
Cadmium	106-1	184.01	189.44	182.75	185.40	1.92	ppb
Cadmium	111-1	200.60	204.77	203.19	202.85	1.04	ppb
Calcium	43-1	4097.57	4231.95	4150.18	4159.90	1.63	ppb
Calcium	44-1	4067.51	4178.10	4159.10	4134.90	1.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	158.98	155.71	157.42	157.37	1.04	ppb
Cobalt	59-2	397.60	397.46	386.44	393.83	1.63	ppb
Copper	63-2	184.09	182.16	181.93	182.72	0.65	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				105		%
Indium	115-2				104		%
Iron	56-2	824.16	805.12	814.73	814.67	1.17	ppb
Iron	57-2	861.84	846.88	849.32	852.68	0.94	ppb
Iron	54-2	965.13	945.68	941.85	950.89	1.31	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	365.79	358.69	368.19	364.22	1.36	ppb
Lead	207-1	358.49	350.77	358.01	355.76	1.22	ppb
Lead	208-1	371.95	366.35	369.39	369.23	0.76	ppb
Lithium	6-1				99		%
Magnesium	24-2	3591.01	3558.60	3699.90	3616.50	2.05	ppb
Manganese	55-2	367.36	362.12	361.23	363.57	0.91	ppb
Molybdenum	94-1	0.26	0.27	0.29	0.27	4.82	ppb
Molybdenum	95-1	0.26	0.21	0.28	0.25	14.14	ppb
Molybdenum	96-1	0.32	0.26	0.32	0.30	11.18	ppb
Molybdenum	97-1	0.28	0.27	0.27	0.27	0.35	ppb
Molybdenum	98-1	0.28	0.27	0.26	0.27	3.15	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	398.89	398.86	397.42	398.39	0.21	ppb
Phosphorus	31-2	-23.97	-32.15	-19.96	-25.36		ppb
Potassium	39-2	3998.78	3945.52	4026.07	3990.13	1.03	ppb
Rhodium	103-1				105		%
Rhodium	103-2				104		%
Scandium	45-1				104		%
Scandium	45-2				104		%
Selenium	82-1	392.61	396.59	403.90	397.70	1.44	ppb
Selenium	77-2	406.96	394.73	392.51	398.07	1.96	ppb
Selenium	78-2	396.08	394.96	392.02	394.35	0.53	ppb
Silicon	28-1	1.22	2.43	1.95	1.87	32.87	ppb
Silver	107-1	202.44	204.00	203.32	203.25	0.38	ppb
Silver	109-1	199.91	203.42	201.79	201.70	0.87	ppb
Sodium	23-2	4020.05	4061.32	4028.71	4036.69	0.54	ppb
Strontium	86-1	0.10	0.11	0.08	0.09	17.82	ppb
Strontium	88-1	0.07	0.07	0.08	0.08	5.72	ppb
Sulfur	34-1	-137.64	17.01	0.92	-39.91		ppb
Terbium	159-1				102		%
Terbium	159-2				103		%
Thallium	203-1	373.70	371.52	377.71	374.31	0.84	ppb
Thallium	205-1	371.64	361.86	379.80	371.10	2.42	ppb
Tin	118-1	0.36	0.42	0.34	0.37	11.16	ppb
Titanium	47-1	0.20	0.17	0.21	0.19	10.06	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	17.72	ppb
Vanadium	51-2	388.89	385.79	384.92	386.53	0.54	ppb
Yttrium	89-1				98		%
Yttrium	89-2				98		%
Zinc	66-2	368.40	367.61	360.65	365.55	1.17	ppb
Zirconium	90-1	0.10	0.02	0.02	0.04	106.21	ppb
Zirconium	91-1	0.07	0.09	0.07	0.08	12.24	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	19.60	20.10	18.43	19.38	4.44	ppb
Antimony	121-1	2.10	2.12	2.15	2.12	1.13	ppb
Arsenic	75-2	0.92	1.14	0.96	1.01	11.81	ppb
Barium	135-1	9.80	10.07	10.01	9.96	1.45	ppb
Barium	137-1	9.96	10.30	9.82	10.03	2.46	ppb
Beryllium	9-1	1.10	1.13	1.06	1.09	2.95	ppb
Bismuth	209-1				100		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.75	0.91	0.82	0.83	10.20	ppb
Cadmium	106-1	1.62	0.87	0.03	0.84	94.21	ppb
Cadmium	111-1	0.99	0.98	0.91	0.96	4.36	ppb
Calcium	43-1	527.34	541.80	530.98	533.38	1.41	ppb
Calcium	44-1	549.00	556.15	558.15	554.43	0.87	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.96	1.99	1.98	1.98	0.89	ppb
Cobalt	59-2	1.07	1.03	1.04	1.05	2.16	ppb
Copper	63-2	2.08	1.86	1.91	1.95	5.79	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				101		%
Indium	115-1				104		%
Indium	115-2				102		%
Iron	56-2	56.28	53.13	53.31	54.24	3.26	ppb
Iron	57-2	60.33	54.55	54.02	56.30	6.22	ppb
Iron	54-2	61.57	52.46	55.31	56.45	8.25	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.79	0.86	0.85	0.83	4.66	ppb
Lead	207-1	0.80	0.80	0.91	0.84	7.17	ppb
Lead	208-1	0.82	0.83	0.86	0.84	2.73	ppb
Lithium	6-1				99		%
Magnesium	24-2	522.02	485.57	489.21	498.93	4.02	ppb
Manganese	55-2	1.07	1.04	1.02	1.04	2.50	ppb
Molybdenum	94-1	5.99	5.80	6.19	5.99	3.20	ppb
Molybdenum	95-1	4.94	4.89	5.03	4.95	1.44	ppb
Molybdenum	96-1	5.08	5.04	5.05	5.06	0.45	ppb
Molybdenum	97-1	5.04	4.92	5.00	4.99	1.26	ppb
Molybdenum	98-1	4.99	4.97	5.02	4.99	0.50	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.10	0.93	0.93	0.99	9.74	ppb
Phosphorus	31-2	234.75	212.58	229.84	225.73	5.16	ppb
Potassium	39-2	583.32	535.60	531.92	550.28	5.21	ppb
Rhodium	103-1				105		%
Rhodium	103-2				103		%
Scandium	45-1				104		%
Scandium	45-2				103		%
Selenium	82-1	4.40	5.18	5.24	4.94	9.52	ppb
Selenium	77-2	5.22	6.45	6.67	6.11	12.76	ppb
Selenium	78-2	6.83	5.90	7.29	6.67	10.62	ppb
Silicon	28-1	0.25	1.04	1.88	1.06	76.77	ppb
Silver	107-1	1.13	1.12	1.13	1.13	0.61	ppb
Silver	109-1	1.12	1.14	1.12	1.13	1.26	ppb
Sodium	23-2	547.44	510.95	505.23	521.21	4.39	ppb
Strontium	86-1	0.99	1.02	1.02	1.01	1.52	ppb
Strontium	88-1	0.97	0.95	0.94	0.95	1.25	ppb
Sulfur	34-1	-548.13	-349.29	-287.89	-395.10		ppb
Terbium	159-1				102		%
Terbium	159-2				103		%
Thallium	203-1	1.06	1.02	1.07	1.05	2.80	ppb
Thallium	205-1	1.02	0.98	1.05	1.02	3.16	ppb
Tin	118-1	5.09	5.07	5.18	5.11	1.14	ppb
Titanium	47-1	4.93	5.08	5.16	5.05	2.33	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.90	0.90	0.92	0.91	1.36	ppb
Vanadium	51-2	5.22	4.99	4.89	5.03	3.35	ppb
Yttrium	89-1				98		%
Yttrium	89-2				98		%
Zinc	66-2	4.93	3.71	4.18	4.27	14.43	ppb
Zirconium	90-1	0.99	1.01	0.98	0.99	2.00	ppb
Zirconium	91-1	1.09	1.01	1.01	1.04	4.42	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-1.98	-2.36	-1.86	-2.07		ppb
Antimony	121-1	0.02	0.02	0.02	0.02	18.04	ppb
Arsenic	75-2	-0.02	-0.03	-0.01	-0.02		ppb
Barium	135-1	-0.02	-0.01	-0.01	-0.01		ppb
Barium	137-1	0.01	0.00	0.00	0.00	80.08	ppb
Beryllium	9-1	0.00	0.00	0.00	0.00		ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	0.07	-0.05	0.02	387.71	ppb
Cadmium	106-1	-0.54	-0.90	-1.11	-0.85		ppb
Cadmium	111-1	-0.03	-0.06	-0.07	-0.06		ppb
Calcium	43-1	-8.64	-5.14	-6.74	-6.84		ppb
Calcium	44-1	15.72	14.84	15.62	15.39	3.12	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.10	0.02	0.00	0.04	140.53	ppb
Cobalt	59-2	-0.01	-0.01	-0.02	-0.01		ppb
Copper	63-2	-0.01	-0.03	-0.03	-0.02		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				105		%
Indium	115-2				106		%
Iron	56-2	-0.59	-1.25	-1.28	-1.04		ppb
Iron	57-2	-2.09	0.19	-0.29	-0.73		ppb
Iron	54-2	-1.92	-0.91	-1.37	-1.40		ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.01	0.02	0.02	0.02	21.71	ppb
Lead	207-1	0.00	0.02	0.00	0.01	115.39	ppb
Lead	208-1	0.01	0.01	0.02	0.01	26.79	ppb
Lithium	6-1				100		%
Magnesium	24-2	-6.57	-6.28	-6.48	-6.44		ppb
Manganese	55-2	-0.07	-0.04	-0.02	-0.04		ppb
Molybdenum	94-1	0.00	0.01	-0.02	0.00		ppb
Molybdenum	95-1	-0.01	0.01	-0.01	-0.01		ppb
Molybdenum	96-1	0.00	0.02	0.01	0.01	95.08	ppb
Molybdenum	97-1	0.03	0.02	0.02	0.02	16.94	ppb
Molybdenum	98-1	0.00	-0.01	0.00	-0.01		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.08	-0.06	-0.07	-0.07		ppb
Phosphorus	31-2	-32.26	-30.80	-39.25	-34.10		ppb
Potassium	39-2	-0.76	10.09	9.98	6.44	96.81	ppb
Rhodium	103-1				105		%
Rhodium	103-2				105		%
Scandium	45-1				106		%
Scandium	45-2				104		%
Selenium	82-1	-0.42	-0.42	-0.55	-0.46		ppb
Selenium	77-2	0.00	0.00	0.00	0.00	N/A	ppb
Selenium	78-2	1.42	0.48	1.36	1.09	48.36	ppb
Silicon	28-1	-1.16	-1.03	-0.61	-0.94		ppb
Silver	107-1	0.01	0.01	0.00	0.01	58.90	ppb
Silver	109-1	0.02	0.01	0.01	0.01	29.04	ppb
Sodium	23-2	2.94	6.35	6.35	5.21	37.74	ppb
Strontium	86-1	0.01	0.04	0.04	0.03	56.26	ppb
Strontium	88-1	0.00	0.01	0.01	0.00	196.13	ppb
Sulfur	34-1	-704.86	-690.09	-658.91	-684.62		ppb
Terbium	159-1				104		%
Terbium	159-2				105		%
Thallium	203-1	0.02	0.00	0.01	0.01	173.77	ppb
Thallium	205-1	0.02	0.02	0.02	0.02	20.74	ppb
Tin	118-1	-0.04	-0.02	-0.03	-0.03		ppb
Titanium	47-1	-0.02	-0.02	0.03	0.00		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.01	0.01	0.00	0.00	49.59	ppb
Yttrium	89-1				98		%
Yttrium	89-2				100		%
Zinc	66-2	-0.94	-1.11	-0.92	-0.99		ppb
Zirconium	90-1	0.01	0.01	0.00	0.01	46.57	ppb
Zirconium	91-1	-0.01	-0.01	0.02	0.00	1043.20	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	89189.51	87381.04	90775.56	89115.37	1.91	ppb
Antimony	121-1	1.13	1.10	1.10	1.11	1.27	ppb
Arsenic	75-2	0.26	0.34	0.26	0.29	16.88	ppb
Barium	135-1	1.28	1.27	1.28	1.27	0.67	ppb
Barium	137-1	1.36	1.32	1.36	1.35	1.82	ppb
Beryllium	9-1	0.32	0.26	0.27	0.28	11.87	ppb
Bismuth	209-1				100		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	8.18	7.08	7.52	7.59	7.28	ppb
Cadmium	106-1	-0.90	-1.23	-1.35	-1.16		ppb
Cadmium	111-1	0.17	0.23	0.12	0.17	31.72	ppb
Calcium	43-1	104226.73	100798.82	103199.93	102741.83	1.71	ppb
Calcium	44-1	103778.99	100621.36	102486.50	102295.61	1.55	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	19.77	20.16	20.11	20.01	1.05	ppb
Cobalt	59-2	1.33	1.25	1.26	1.28	3.44	ppb
Copper	63-2	7.35	7.36	7.52	7.41	1.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				107		%
Indium	115-1				107		%
Indium	115-2				106		%
Iron	56-2	101978.70	103526.79	102146.79	102550.76	0.83	ppb
Iron	57-2	100341.36	102827.53	101802.85	101657.25	1.23	ppb
Iron	54-2	102785.03	102791.09	103209.57	102928.56	0.24	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4.26	4.51	4.42	4.40	2.96	ppb
Lead	207-1	3.85	4.16	4.03	4.01	3.80	ppb
Lead	208-1	3.97	4.22	4.11	4.10	3.06	ppb
Lithium	6-1				97		%
Magnesium	24-2	90996.10	92048.98	90425.65	91156.91	0.90	ppb
Manganese	55-2	7.63	7.57	7.36	7.52	1.94	ppb
Molybdenum	94-1	1596.05	1554.18	1590.05	1580.09	1.43	ppb
Molybdenum	95-1	1968.91	1880.90	1963.30	1937.71	2.54	ppb
Molybdenum	96-1	1943.55	1874.09	1918.07	1911.90	1.84	ppb
Molybdenum	97-1	1995.30	1978.91	1930.57	1968.26	1.71	ppb
Molybdenum	98-1	2003.67	1934.77	1930.33	1956.26	2.10	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	5.11	5.45	5.31	5.29	3.22	ppb
Phosphorus	31-2	98294.33	98857.61	97783.90	98311.95	0.55	ppb
Potassium	39-2	101497.54	101445.12	101631.61	101524.76	0.09	ppb
Rhodium	103-1				107		%
Rhodium	103-2				105		%
Scandium	45-1				103		%
Scandium	45-2				105		%
Selenium	82-1	0.00	0.06	-0.70	-0.21		ppb
Selenium	77-2	0.22	0.00	0.22	0.15	86.60	ppb
Selenium	78-2	0.46	1.21	3.07	1.58	85.11	ppb
Silicon	28-1	1.83	1.51	2.59	1.98	28.00	ppb
Silver	107-1	0.05	0.04	0.04	0.05	16.05	ppb
Silver	109-1	0.07	0.04	0.05	0.05	28.51	ppb
Sodium	23-2	100320.76	99500.67	100743.16	100188.20	0.63	ppb
Strontium	86-1	32.58	31.67	31.62	31.96	1.68	ppb
Strontium	88-1	32.69	31.78	31.52	31.99	1.92	ppb
Sulfur	34-1	102589.70	102049.44	101785.49	102141.54	0.40	ppb
Terbium	159-1				108		%
Terbium	159-2				108		%
Thallium	203-1	0.02	0.04	0.05	0.04	36.35	ppb
Thallium	205-1	0.02	0.04	0.06	0.04	40.30	ppb
Tin	118-1	0.32	0.57	0.73	0.54	39.01	ppb
Titanium	47-1	2000.24	2000.22	1959.99	1986.82	1.17	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.02	0.01	15.58	ppb
Vanadium	51-2	0.23	0.23	0.20	0.22	8.19	ppb
Yttrium	89-1				101		%
Yttrium	89-2				101		%
Zinc	66-2	8.99	9.34	9.19	9.17	1.89	ppb
Zirconium	90-1	0.02	0.02	0.02	0.02	8.99	ppb
Zirconium	91-1	0.01	0.02	0.01	0.01	45.04	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	91071.28	88369.14	89739.20	89726.54	1.51	ppb
Antimony	121-1	20.74	20.44	21.45	20.88	2.48	ppb
Arsenic	75-2	19.48	20.47	20.65	20.20	3.12	ppb
Barium	135-1	21.92	20.70	21.63	21.41	2.98	ppb
Barium	137-1	21.63	21.80	21.70	21.71	0.40	ppb
Beryllium	9-1	22.10	21.67	20.39	21.39	4.15	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	26.55	25.14	23.70	25.13	5.66	ppb
Cadmium	106-1	16.87	15.77	16.04	16.23	3.52	ppb
Cadmium	111-1	20.35	20.00	20.82	20.39	2.02	ppb
Calcium	43-1	103045.28	106429.96	96857.63	102110.96	4.75	ppb
Calcium	44-1	103357.56	104297.27	99771.27	102475.36	2.33	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	41.08	39.77	40.90	40.58	1.75	ppb
Cobalt	59-2	22.06	21.61	22.10	21.92	1.24	ppb
Copper	63-2	26.67	26.28	26.49	26.48	0.73	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				108		%
Holmium	165-2				109		%
Indium	115-1				109		%
Indium	115-2				107		%
Iron	56-2	104620.99	103866.98	102516.50	103668.15	1.03	ppb
Iron	57-2	103511.11	99985.34	102468.35	101988.27	1.78	ppb
Iron	54-2	104225.19	101228.69	102550.74	102668.20	1.46	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	24.85	24.27	23.93	24.35	1.92	ppb
Lead	207-1	24.16	24.04	22.51	23.57	3.91	ppb
Lead	208-1	23.81	23.59	23.50	23.63	0.67	ppb
Lithium	6-1				89		%
Magnesium	24-2	94173.71	90445.09	90749.49	91789.43	2.26	ppb
Manganese	55-2	28.65	28.32	28.36	28.44	0.64	ppb
Molybdenum	94-1	1573.37	1599.46	1649.83	1607.55	2.42	ppb
Molybdenum	95-1	1936.85	1932.31	1988.08	1952.41	1.59	ppb
Molybdenum	96-1	1897.18	1885.34	1928.09	1903.53	1.16	ppb
Molybdenum	97-1	1914.75	1995.83	1972.32	1960.96	2.13	ppb
Molybdenum	98-1	1959.45	1982.05	1986.15	1975.88	0.73	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	27.50	25.93	26.26	26.56	3.12	ppb
Phosphorus	31-2	99045.43	97187.51	97593.08	97942.01	1.00	ppb
Potassium	39-2	103859.25	101616.23	103485.67	102987.05	1.17	ppb
Rhodium	103-1				107		%
Rhodium	103-2				107		%
Scandium	45-1				105		%
Scandium	45-2				107		%
Selenium	82-1	19.31	18.94	20.57	19.60	4.35	ppb
Selenium	77-2	17.76	19.97	17.82	18.51	6.80	ppb
Selenium	78-2	21.56	19.58	22.18	21.10	6.42	ppb
Silicon	28-1	17.29	17.17	13.43	15.96	13.74	ppb
Silver	107-1	19.73	19.28	20.69	19.90	3.61	ppb
Silver	109-1	19.60	19.50	20.22	19.77	1.97	ppb
Sodium	23-2	103837.36	98456.00	99623.52	100638.96	2.81	ppb
Strontium	86-1	32.18	32.81	33.92	32.97	2.67	ppb
Strontium	88-1	32.09	32.50	32.95	32.51	1.33	ppb
Sulfur	34-1	101093.00	99753.64	92611.63	97819.42	4.66	ppb
Terbium	159-1				111		%
Terbium	159-2				109		%
Thallium	203-1	21.08	21.20	19.40	20.56	4.90	ppb
Thallium	205-1	20.23	20.41	20.56	20.40	0.80	ppb
Tin	118-1	0.47	0.63	0.72	0.61	20.35	ppb
Titanium	47-1	1992.34	2013.64	1943.65	1983.21	1.81	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	3.26	ppb
Vanadium	51-2	20.61	20.18	20.24	20.34	1.15	ppb
Yttrium	89-1				103		%
Yttrium	89-2				103		%
Zinc	66-2	26.67	26.79	27.90	27.12	2.49	ppb
Zirconium	90-1	0.09	0.10	0.09	0.09	2.71	ppb
Zirconium	91-1	0.10	0.11	0.09	0.10	10.40	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	46998.26	47028.06	47269.04	47098.45	0.32	ppb
Antimony	121-1	483.17	491.76	495.34	490.09	1.28	ppb
Arsenic	75-2	470.12	471.44	480.97	474.18	1.25	ppb
Barium	135-1	2431.62	2468.44	2443.41	2447.82	0.77	ppb
Barium	137-1	2432.74	2499.08	2441.74	2457.85	1.46	ppb
Beryllium	9-1	509.23	507.58	515.62	510.81	0.83	ppb
Bismuth	209-1				102		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	472.13	482.75	485.48	480.12	1.47	ppb
Cadmium	106-1	481.46	471.51	498.92	483.96	2.87	ppb
Cadmium	111-1	473.96	478.18	485.42	479.18	1.21	ppb
Calcium	43-1	253555.71	250080.75	250433.42	251356.63	0.76	ppb
Calcium	44-1	248162.16	249194.34	251240.69	249532.40	0.63	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	472.88	492.53	502.12	489.18	3.05	ppb
Cobalt	59-2	471.16	487.84	485.75	481.58	1.89	ppb
Copper	63-2	4747.71	4770.22	4795.18	4771.04	0.50	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				110		%
Holmium	165-2				109		%
Indium	115-1				109		%
Indium	115-2				108		%
Iron	56-2	125428.15	126925.27	126698.55	126350.66	0.64	ppb
Iron	57-2	123726.60	126235.08	130545.90	126835.86	2.72	ppb
Iron	54-2	125395.37	129734.16	128765.79	127965.11	1.78	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2455.71	2554.16	2498.21	2502.69	1.97	ppb
Lead	207-1	2457.75	2485.51	2482.42	2475.23	0.61	ppb
Lead	208-1	2443.21	2483.32	2459.34	2461.96	0.82	ppb
Lithium	6-1				90		%
Magnesium	24-2	236621.45	241577.66	239391.55	239196.88	1.04	ppb
Manganese	55-2	4758.51	4937.38	4959.31	4885.07	2.25	ppb
Molybdenum	94-1	4836.37	4948.56	4903.47	4896.14	1.15	ppb
Molybdenum	95-1	4915.70	4930.26	4847.36	4897.78	0.90	ppb
Molybdenum	96-1	4906.52	4882.91	4758.47	4849.30	1.64	ppb
Molybdenum	97-1	4791.14	4906.52	4824.44	4840.70	1.23	ppb
Molybdenum	98-1	4764.07	4875.06	4865.15	4834.76	1.27	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	474.05	481.26	481.11	478.81	0.86	ppb
Phosphorus	31-2	9470.80	9592.58	9549.87	9537.75	0.65	ppb
Potassium	39-2	124773.33	124455.88	123461.73	124230.31	0.55	ppb
Rhodium	103-1				106		%
Rhodium	103-2				106		%
Scandium	45-1				107		%
Scandium	45-2				110		%
Selenium	82-1	459.77	473.40	464.34	465.84	1.49	ppb
Selenium	77-2	481.82	458.06	480.07	473.32	2.80	ppb
Selenium	78-2	465.18	460.75	469.18	465.04	0.91	ppb
Silicon	28-1	486.20	487.47	484.76	486.15	0.28	ppb
Silver	107-1	477.04	487.24	490.08	484.79	1.41	ppb
Silver	109-1	479.58	485.02	487.13	483.91	0.81	ppb
Sodium	23-2	240106.66	249743.77	241538.51	243796.31	2.13	ppb
Strontium	86-1	482.40	488.97	486.20	485.86	0.68	ppb
Strontium	88-1	492.17	495.75	494.00	493.97	0.36	ppb
Sulfur	34-1	8261.47	8160.88	8082.21	8168.19	1.10	ppb
Terbium	159-1				112		%
Terbium	159-2				111		%
Thallium	203-1	502.21	511.69	518.82	510.91	1.63	ppb
Thallium	205-1	495.44	511.30	501.56	502.76	1.59	ppb
Tin	118-1	491.66	496.06	500.41	496.04	0.88	ppb
Titanium	47-1	4779.85	4881.86	4893.49	4851.73	1.29	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	476.74	492.75	491.63	487.04	1.84	ppb
Vanadium	51-2	482.05	493.95	492.90	489.63	1.35	ppb
Yttrium	89-1				105		%
Yttrium	89-2				106		%
Zinc	66-2	4588.26	4700.66	4730.43	4673.12	1.60	ppb
Zirconium	90-1	484.18	483.22	481.32	482.91	0.30	ppb
Zirconium	91-1	475.08	486.22	477.73	479.68	1.21	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-1.23	-1.37	-1.95	-1.52		ppb
Antimony	121-1	0.10	0.11	0.11	0.11	2.70	ppb
Arsenic	75-2	0.01	0.01	0.00	0.01	148.92	ppb
Barium	135-1	0.00	0.02	0.00	0.00	269.53	ppb
Barium	137-1	0.01	0.03	0.01	0.02	79.85	ppb
Beryllium	9-1	-0.01	0.00	0.00	0.00		ppb
Bismuth	209-1				100		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	0.07	0.28	0.13	102.52	ppb
Cadmium	106-1	-2.02	-0.64	-2.22	-1.63		ppb
Cadmium	111-1	-0.12	-0.03	-0.14	-0.10		ppb
Calcium	43-1	-7.40	-4.09	-9.07	-6.85		ppb
Calcium	44-1	31.15	27.95	30.48	29.86	5.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.03	0.00	-0.01	-0.01		ppb
Cobalt	59-2	-0.02	-0.01	-0.01	-0.01		ppb
Copper	63-2	0.13	0.20	0.17	0.16	21.17	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				102		%
Indium	115-1				106		%
Indium	115-2				105		%
Iron	56-2	0.68	0.20	4.98	1.95	134.81	ppb
Iron	57-2	-0.20	-3.22	3.69	0.09	3704.16	ppb
Iron	54-2	2.26	0.17	0.98	1.14	92.61	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.03	0.03	0.03	0.03	8.07	ppb
Lead	207-1	0.04	0.01	0.02	0.03	67.69	ppb
Lead	208-1	0.04	0.03	0.03	0.03	15.08	ppb
Lithium	6-1				88		%
Magnesium	24-2	-4.25	-4.73	-4.54	-4.51		ppb
Manganese	55-2	-0.02	0.02	0.00	0.00	4896.07	ppb
Molybdenum	94-1	0.20	0.18	0.10	0.16	32.60	ppb
Molybdenum	95-1	0.16	0.14	0.17	0.16	8.03	ppb
Molybdenum	96-1	0.15	0.19	0.15	0.16	13.05	ppb
Molybdenum	97-1	0.19	0.17	0.16	0.17	9.36	ppb
Molybdenum	98-1	0.15	0.16	0.13	0.15	8.93	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.03	-0.06	-0.06	-0.05		ppb
Phosphorus	31-2	-39.68	-43.48	-33.72	-38.96		ppb
Potassium	39-2	29.47	23.05	31.96	28.16	16.32	ppb
Rhodium	103-1				106		%
Rhodium	103-2				105		%
Scandium	45-1				103		%
Scandium	45-2				104		%
Selenium	82-1	0.33	-0.23	-0.71	-0.20		ppb
Selenium	77-2	0.22	0.22	0.22	0.22	0.41	ppb
Selenium	78-2	2.24	0.76	2.80	1.93	54.54	ppb
Silicon	28-1	-0.40	-0.72	0.32	-0.26		ppb
Silver	107-1	0.05	0.04	0.03	0.04	25.10	ppb
Silver	109-1	0.05	0.04	0.03	0.04	33.98	ppb
Sodium	23-2	48.90	43.76	47.10	46.59	5.60	ppb
Strontium	86-1	0.00	0.01	0.05	0.02	109.04	ppb
Strontium	88-1	0.00	0.01	0.00	0.01	118.86	ppb
Sulfur	34-1	-970.99	-944.24	-708.86	-874.69		ppb
Terbium	159-1				103		%
Terbium	159-2				103		%
Thallium	203-1	0.07	0.06	0.07	0.06	7.85	ppb
Thallium	205-1	0.08	0.06	0.06	0.07	14.46	ppb
Tin	118-1	0.12	0.12	0.12	0.12	2.56	ppb
Titanium	47-1	0.01	0.05	0.03	0.03	78.78	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	28.22	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00	23.71	ppb
Yttrium	89-1				99		%
Yttrium	89-2				100		%
Zinc	66-2	-0.48	-0.72	-0.43	-0.54		ppb
Zirconium	90-1	0.02	0.02	0.03	0.02	8.17	ppb
Zirconium	91-1	0.00	0.01	0.03	0.02	96.51	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18.48	17.09	18.53	18.03	4.55	ppb
Antimony	121-1	2.15	2.16	2.15	2.15	0.34	ppb
Arsenic	75-2	0.95	1.07	1.09	1.04	7.59	ppb
Barium	135-1	10.59	10.63	10.62	10.61	0.18	ppb
Barium	137-1	10.65	10.54	10.87	10.69	1.57	ppb
Beryllium	9-1	1.03	1.24	1.23	1.17	10.46	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.31	1.26	1.09	1.22	9.51	ppb
Cadmium	106-1	-1.13	-0.49	-0.82	-0.81		ppb
Cadmium	111-1	0.99	1.13	1.03	1.05	6.86	ppb
Calcium	43-1	588.23	560.29	582.72	577.08	2.56	ppb
Calcium	44-1	586.10	587.60	593.75	589.15	0.69	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.23	2.14	2.22	2.19	2.27	ppb
Cobalt	59-2	1.17	1.13	1.15	1.15	1.78	ppb
Copper	63-2	2.16	2.24	2.15	2.19	2.30	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				105		%
Indium	115-2				107		%
Iron	56-2	58.49	57.12	61.98	59.20	4.24	ppb
Iron	57-2	62.16	60.73	58.77	60.55	2.81	ppb
Iron	54-2	60.68	58.17	59.02	59.29	2.15	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.98	1.08	0.97	1.01	5.84	ppb
Lead	207-1	1.01	1.03	0.91	0.98	6.63	ppb
Lead	208-1	1.00	1.04	0.98	1.01	2.80	ppb
Lithium	6-1				90		%
Magnesium	24-2	529.29	518.08	518.10	521.82	1.24	ppb
Manganese	55-2	1.08	1.06	1.03	1.06	1.99	ppb
Molybdenum	94-1	6.09	6.44	6.55	6.36	3.78	ppb
Molybdenum	95-1	5.21	5.47	5.40	5.36	2.51	ppb
Molybdenum	96-1	5.43	5.63	5.47	5.51	1.88	ppb
Molybdenum	97-1	5.31	5.52	5.56	5.46	2.54	ppb
Molybdenum	98-1	5.33	5.62	5.50	5.48	2.62	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.13	1.23	1.11	1.16	5.73	ppb
Phosphorus	31-2	-8.54	-6.34	-0.45	-5.11		ppb
Potassium	39-2	596.61	590.88	586.40	591.30	0.87	ppb
Rhodium	103-1				105		%
Rhodium	103-2				105		%
Scandium	45-1				102		%
Scandium	45-2				103		%
Selenium	82-1	5.62	5.65	5.41	5.56	2.34	ppb
Selenium	77-2	7.58	7.04	6.96	7.19	4.72	ppb
Selenium	78-2	8.42	6.22	6.56	7.07	16.77	ppb
Silicon	28-1	-0.12	0.29	1.18	0.45	147.11	ppb
Silver	107-1	1.08	1.16	1.16	1.13	4.34	ppb
Silver	109-1	1.11	1.10	1.10	1.10	0.56	ppb
Sodium	23-2	568.77	557.57	558.97	561.77	1.09	ppb
Strontium	86-1	1.13	1.12	1.10	1.11	1.30	ppb
Strontium	88-1	1.06	1.08	1.07	1.07	0.94	ppb
Sulfur	34-1	-520.99	-425.93	-321.43	-422.78		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	1.01	1.01	1.02	1.01	0.39	ppb
Thallium	205-1	1.03	1.07	1.06	1.05	1.80	ppb
Tin	118-1	5.59	5.51	5.51	5.54	0.79	ppb
Titanium	47-1	5.09	5.29	5.43	5.27	3.25	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.97	0.99	0.95	0.97	1.98	ppb
Vanadium	51-2	5.47	5.67	5.45	5.53	2.21	ppb
Yttrium	89-1				98		%
Yttrium	89-2				100		%
Zinc	66-2	4.73	4.94	4.42	4.70	5.55	ppb
Zirconium	90-1	0.97	1.08	1.03	1.03	5.05	ppb
Zirconium	91-1	1.08	1.05	1.04	1.05	1.84	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-2.26	-2.87	-2.01	-2.38		ppb
Antimony	121-1	0.00	0.01	0.01	0.01	40.55	ppb
Arsenic	75-2	-0.03	-0.02	-0.03	-0.02		ppb
Barium	135-1	-0.01	-0.01	-0.03	-0.02		ppb
Barium	137-1	0.00	-0.02	0.00	-0.01		ppb
Beryllium	9-1	0.00	0.00	-0.01	0.00	1888.87	ppb
Bismuth	209-1				103		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	0.04	0.07	0.05	39.84	ppb
Cadmium	106-1	-1.10	-0.93	-2.13	-1.39		ppb
Cadmium	111-1	-0.06	-0.07	-0.14	-0.09		ppb
Calcium	43-1	-10.75	-4.70	-13.70	-9.72		ppb
Calcium	44-1	0.96	4.72	1.51	2.40	84.58	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.10	-0.05	-0.04	-0.06		ppb
Cobalt	59-2	-0.02	-0.02	-0.01	-0.02		ppb
Copper	63-2	0.37	0.80	0.35	0.50	50.07	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				100		%
Indium	115-1				105		%
Indium	115-2				103		%
Iron	56-2	-0.99	-0.93	5.07	1.05	331.19	ppb
Iron	57-2	-1.83	-1.85	-3.11	-2.26		ppb
Iron	54-2	-1.39	-0.12	-0.25	-0.59		ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.03	0.04	0.02	0.03	38.64	ppb
Lead	207-1	0.03	0.05	0.02	0.03	52.80	ppb
Lead	208-1	0.04	0.04	0.02	0.03	38.34	ppb
Lithium	6-1				93		%
Magnesium	24-2	-6.02	-6.23	-6.30	-6.18		ppb
Manganese	55-2	0.01	-0.04	-0.02	-0.01		ppb
Molybdenum	94-1	0.03	-0.02	0.05	0.02	168.92	ppb
Molybdenum	95-1	0.01	0.01	-0.01	0.00	494.89	ppb
Molybdenum	96-1	0.00	0.01	0.02	0.01	92.31	ppb
Molybdenum	97-1	0.03	0.02	0.02	0.02	49.82	ppb
Molybdenum	98-1	0.01	0.00	0.01	0.01	84.06	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.03	0.08	0.03	0.03	200.47	ppb
Phosphorus	31-2	-31.88	-27.76	-25.49	-28.38		ppb
Potassium	39-2	11.51	15.77	12.47	13.25	16.88	ppb
Rhodium	103-1				105		%
Rhodium	103-2				103		%
Scandium	45-1				103		%
Scandium	45-2				102		%
Selenium	82-1	-0.97	-0.32	-0.46	-0.58		ppb
Selenium	77-2	0.00	0.00	0.00	0.00	N/A	ppb
Selenium	78-2	0.16	3.05	1.35	1.52	95.16	ppb
Silicon	28-1	-1.98	-1.13	-1.07	-1.39		ppb
Silver	107-1	0.00	0.00	0.01	0.00	114.25	ppb
Silver	109-1	0.00	0.01	0.01	0.01	49.56	ppb
Sodium	23-2	14.61	16.26	15.21	15.36	5.43	ppb
Strontium	86-1	0.02	0.01	-0.03	0.00	1249.85	ppb
Strontium	88-1	0.00	0.00	0.01	0.00	43.38	ppb
Sulfur	34-1	-471.10	-305.89	-274.01	-350.33		ppb
Terbium	159-1				103		%
Terbium	159-2				102		%
Thallium	203-1	0.01	0.01	0.01	0.01	47.29	ppb
Thallium	205-1	0.02	0.01	0.01	0.01	35.69	ppb
Tin	118-1	0.04	-0.04	-0.03	-0.01		ppb
Titanium	47-1	-0.03	-0.04	-0.06	-0.04		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.00	0.01	-0.01	0.00	527.67	ppb
Yttrium	89-1				98		%
Yttrium	89-2				99		%
Zinc	66-2	0.13	-0.28	-0.11	-0.09		ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	7.65	ppb
Zirconium	91-1	0.00	0.01	-0.01	0.00		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9686.17	9752.41	9820.85	9753.14	0.69	ppb
Antimony	121-1	484.21	499.32	503.12	495.55	2.02	ppb
Arsenic	75-2	481.46	515.38	495.15	497.33	3.43	ppb
Barium	135-1	2422.47	2520.20	2522.63	2488.43	2.30	ppb
Barium	137-1	2428.36	2519.22	2560.00	2502.53	2.69	ppb
Beryllium	9-1	500.98	504.10	513.27	506.12	1.26	ppb
Bismuth	209-1				103		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	489.23	512.67	510.17	504.02	2.55	ppb
Cadmium	106-1	491.42	509.48	518.02	506.31	2.68	ppb
Cadmium	111-1	490.56	503.49	512.70	502.25	2.21	ppb
Calcium	43-1	52722.33	52985.00	52941.87	52883.07	0.27	ppb
Calcium	44-1	51121.25	51213.78	50775.94	51036.99	0.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	486.66	511.41	503.69	500.58	2.53	ppb
Cobalt	59-2	484.16	492.41	497.93	491.50	1.41	ppb
Copper	63-2	5047.13	4987.80	5051.47	5028.80	0.71	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				106		%
Indium	115-1				108		%
Indium	115-2				106		%
Iron	56-2	25717.79	26191.77	26385.60	26098.38	1.32	ppb
Iron	57-2	25847.29	26008.87	26077.88	25978.01	0.46	ppb
Iron	54-2	26303.19	27063.16	26682.69	26683.01	1.42	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2452.38	2471.99	2498.60	2474.33	0.94	ppb
Lead	207-1	2474.54	2472.41	2500.79	2482.58	0.64	ppb
Lead	208-1	2490.97	2464.22	2503.65	2486.28	0.81	ppb
Lithium	6-1				98		%
Magnesium	24-2	48328.89	50713.05	50432.83	49824.92	2.62	ppb
Manganese	55-2	4892.00	4968.62	5010.90	4957.17	1.22	ppb
Molybdenum	94-1	4919.77	4987.51	5024.49	4977.26	1.07	ppb
Molybdenum	95-1	4941.30	4958.52	5081.12	4993.64	1.53	ppb
Molybdenum	96-1	4957.74	4991.09	5188.18	5045.67	2.47	ppb
Molybdenum	97-1	4948.55	5028.24	5190.28	5055.69	2.44	ppb
Molybdenum	98-1	4892.42	4951.98	5097.04	4980.48	2.11	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	492.20	491.39	491.84	491.81	0.08	ppb
Phosphorus	31-2	9263.43	9678.26	9779.75	9573.81	2.86	ppb
Potassium	39-2	25425.20	25172.19	25040.52	25212.63	0.78	ppb
Rhodium	103-1				107		%
Rhodium	103-2				105		%
Scandium	45-1				105		%
Scandium	45-2				105		%
Selenium	82-1	494.62	501.25	516.29	504.05	2.20	ppb
Selenium	77-2	500.24	502.50	499.86	500.87	0.29	ppb
Selenium	78-2	496.92	510.11	486.16	497.73	2.41	ppb
Silicon	28-1	504.97	504.77	499.55	503.10	0.61	ppb
Silver	107-1	491.72	514.50	519.16	508.46	2.89	ppb
Silver	109-1	499.26	499.68	520.26	506.40	2.37	ppb
Sodium	23-2	50171.46	51216.74	50853.30	50747.17	1.05	ppb
Strontium	86-1	490.45	493.04	501.74	495.08	1.19	ppb
Strontium	88-1	497.22	498.89	505.64	500.58	0.89	ppb
Sulfur	34-1	9305.08	9256.24	9056.22	9205.84	1.43	ppb
Terbium	159-1				109		%
Terbium	159-2				107		%
Thallium	203-1	504.18	500.96	502.34	502.49	0.32	ppb
Thallium	205-1	498.15	499.09	502.70	499.98	0.48	ppb
Tin	118-1	481.36	496.83	514.70	497.63	3.35	ppb
Titanium	47-1	4918.00	4942.36	4811.90	4890.75	1.42	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	477.36	489.70	492.28	486.45	1.64	ppb
Vanadium	51-2	490.32	505.57	502.48	499.46	1.61	ppb
Yttrium	89-1				100		%
Yttrium	89-2				100		%
Zinc	66-2	5024.57	5035.41	5077.41	5045.80	0.55	ppb
Zirconium	90-1	500.38	494.16	496.59	497.04	0.63	ppb
Zirconium	91-1	488.69	486.93	503.70	493.11	1.87	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-2.36	-1.91	-1.76	-2.01		ppb
Antimony	121-1	0.02	0.03	0.03	0.02	32.73	ppb
Arsenic	75-2	-0.03	-0.01	-0.01	-0.02		ppb
Barium	135-1	-0.01	0.01	-0.04	-0.01		ppb
Barium	137-1	0.01	0.01	-0.01	0.00	269.92	ppb
Beryllium	9-1	0.00	0.00	0.02	0.01	217.54	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.01	0.07	0.28	0.12	121.66	ppb
Cadmium	106-1	-0.69	0.39	-0.45	-0.25		ppb
Cadmium	111-1	-0.03	0.03	-0.04	-0.02		ppb
Calcium	43-1	-10.23	-5.48	-11.44	-9.05		ppb
Calcium	44-1	25.14	30.36	30.59	28.69	10.74	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.43	0.01	-0.11	0.11	261.94	ppb
Cobalt	59-2	-0.01	-0.02	-0.01	-0.02		ppb
Copper	63-2	0.35	0.36	0.34	0.35	3.59	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				105		%
Indium	115-2				104		%
Iron	56-2	-1.24	-1.18	-1.45	-1.29		ppb
Iron	57-2	-3.91	-3.56	-0.96	-2.81		ppb
Iron	54-2	-1.01	-0.83	-0.88	-0.91		ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.04	0.06	0.02	0.04	42.41	ppb
Lead	207-1	0.01	0.05	0.03	0.03	57.40	ppb
Lead	208-1	0.03	0.05	0.03	0.04	26.98	ppb
Lithium	6-1				94		%
Magnesium	24-2	-6.15	-6.20	-6.29	-6.21		ppb
Manganese	55-2	-0.06	-0.04	-0.04	-0.05		ppb
Molybdenum	94-1	0.04	0.03	0.04	0.04	17.05	ppb
Molybdenum	95-1	0.04	0.06	0.02	0.04	45.92	ppb
Molybdenum	96-1	0.04	0.04	0.04	0.04	9.37	ppb
Molybdenum	97-1	0.04	0.07	0.06	0.06	29.91	ppb
Molybdenum	98-1	0.05	0.03	0.03	0.04	28.52	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.09	0.00	0.09	0.06	80.41	ppb
Phosphorus	31-2	-38.12	-40.61	-30.89	-36.54		ppb
Potassium	39-2	7.17	11.18	13.87	10.74	31.41	ppb
Rhodium	103-1				104		%
Rhodium	103-2				105		%
Scandium	45-1				102		%
Scandium	45-2				103		%
Selenium	82-1	-0.54	-1.07	-0.96	-0.86		ppb
Selenium	77-2	0.00	0.23	0.23	0.15	86.60	ppb
Selenium	78-2	1.63	4.14	2.47	2.75	46.60	ppb
Silicon	28-1	-1.68	-0.51	-0.55	-0.91		ppb
Silver	107-1	0.01	0.01	0.00	0.01	110.01	ppb
Silver	109-1	0.01	0.01	0.01	0.01	12.03	ppb
Sodium	23-2	16.98	13.26	13.13	14.46	15.15	ppb
Strontium	86-1	0.01	-0.03	0.04	0.01	402.66	ppb
Strontium	88-1	0.00	0.00	0.00	0.00		ppb
Sulfur	34-1	-745.78	-556.10	-569.71	-623.87		ppb
Terbium	159-1				103		%
Terbium	159-2				103		%
Thallium	203-1	0.01	-0.01	0.00	0.00	388.97	ppb
Thallium	205-1	0.02	0.02	0.01	0.02	48.20	ppb
Tin	118-1	0.15	0.12	0.09	0.12	25.10	ppb
Titanium	47-1	0.01	0.02	0.02	0.02	28.97	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.00	-0.01	0.00	0.00		ppb
Yttrium	89-1				99		%
Yttrium	89-2				99		%
Zinc	66-2	-0.20	-0.49	0.27	-0.14		ppb
Zirconium	90-1	0.00	0.00	0.01	0.00	23.76	ppb
Zirconium	91-1	0.01	-0.01	0.00	0.00	499.32	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9815.30	10445.38	9974.33	10078.34	3.25	ppb
Antimony	121-1	514.04	515.09	505.54	511.55	1.02	ppb
Arsenic	75-2	496.17	517.13	490.15	501.15	2.83	ppb
Barium	135-1	2599.68	2546.51	2561.77	2569.32	1.07	ppb
Barium	137-1	2623.39	2530.34	2491.81	2548.51	2.65	ppb
Beryllium	9-1	497.70	505.08	510.33	504.37	1.26	ppb
Bismuth	209-1				103		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	509.48	517.96	506.48	511.31	1.16	ppb
Cadmium	106-1	524.31	517.93	510.01	517.41	1.38	ppb
Cadmium	111-1	521.86	521.73	510.21	517.93	1.29	ppb
Calcium	43-1	53370.89	54713.92	55016.48	54367.10	1.61	ppb
Calcium	44-1	51821.98	52528.84	53593.85	52648.22	1.69	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	496.59	520.15	503.35	506.70	2.39	ppb
Cobalt	59-2	490.48	518.02	502.09	503.53	2.75	ppb
Copper	63-2	5082.20	5101.92	5159.26	5114.46	0.78	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				107		%
Indium	115-1				107		%
Indium	115-2				106		%
Iron	56-2	26090.62	27829.82	26242.58	26721.00	3.60	ppb
Iron	57-2	26211.93	27639.69	26078.15	26643.26	3.25	ppb
Iron	54-2	26303.60	27908.62	26846.20	27019.47	3.02	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2483.27	2594.24	2625.26	2567.59	2.91	ppb
Lead	207-1	2494.65	2563.39	2596.82	2551.62	2.04	ppb
Lead	208-1	2489.89	2526.47	2580.05	2532.14	1.79	ppb
Lithium	6-1				100		%
Magnesium	24-2	51253.33	53354.09	50524.62	51710.68	2.84	ppb
Manganese	55-2	4969.98	5145.78	4997.62	5037.79	1.88	ppb
Molybdenum	94-1	4877.85	5047.78	5059.22	4994.95	2.03	ppb
Molybdenum	95-1	4956.20	4957.78	5063.63	4992.54	1.23	ppb
Molybdenum	96-1	4962.60	4990.09	5074.00	5008.89	1.16	ppb
Molybdenum	97-1	4904.89	5030.78	5082.36	5006.01	1.82	ppb
Molybdenum	98-1	5019.58	5072.40	5119.32	5070.43	0.98	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	497.09	512.57	506.13	505.27	1.54	ppb
Phosphorus	31-2	9617.73	10145.01	9750.10	9837.61	2.79	ppb
Potassium	39-2	25020.45	26151.86	26204.60	25792.30	2.59	ppb
Rhodium	103-1				105		%
Rhodium	103-2				104		%
Scandium	45-1				104		%
Scandium	45-2				105		%
Selenium	82-1	491.71	503.78	509.03	501.51	1.77	ppb
Selenium	77-2	484.53	512.84	504.19	500.52	2.90	ppb
Selenium	78-2	519.77	508.72	508.81	512.43	1.24	ppb
Silicon	28-1	512.27	510.70	518.07	513.68	0.76	ppb
Silver	107-1	533.73	541.33	508.08	527.71	3.30	ppb
Silver	109-1	531.35	528.15	502.68	520.73	3.02	ppb
Sodium	23-2	53171.82	53721.57	51744.38	52879.26	1.93	ppb
Strontium	86-1	491.61	495.81	510.46	499.29	1.98	ppb
Strontium	88-1	491.84	503.07	507.40	500.77	1.60	ppb
Sulfur	34-1	9294.88	9436.39	9668.23	9466.50	1.99	ppb
Terbium	159-1				107		%
Terbium	159-2				107		%
Thallium	203-1	503.12	519.72	517.79	513.54	1.77	ppb
Thallium	205-1	486.33	507.48	513.59	502.47	2.85	ppb
Tin	118-1	521.86	514.49	509.77	515.37	1.18	ppb
Titanium	47-1	4961.18	5049.18	5157.00	5055.79	1.94	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	473.32	491.28	496.61	487.07	2.51	ppb
Vanadium	51-2	491.66	509.20	496.25	499.04	1.82	ppb
Yttrium	89-1				101		%
Yttrium	89-2				100		%
Zinc	66-2	5097.41	5126.00	5109.89	5111.10	0.28	ppb
Zirconium	90-1	485.35	493.26	503.45	494.02	1.84	ppb
Zirconium	91-1	485.18	493.54	499.07	492.60	1.42	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	58442.07	58072.61	58937.04	58483.91	0.74	ppb
Antimony	121-1	1.42	1.58	1.42	1.47	6.33	ppb
Arsenic	75-2	39.80	40.84	42.36	41.00	3.15	ppb
Barium	135-1	495.81	527.25	501.19	508.08	3.31	ppb
Barium	137-1	511.52	531.74	515.86	519.71	2.05	ppb
Beryllium	9-1	3.38	3.37	3.34	3.36	0.63	ppb
Bismuth	209-1				107		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	5.14	6.01	4.64	5.27	13.21	ppb
Cadmium	106-1	6.03	8.03	8.69	7.58	18.24	ppb
Cadmium	111-1	4.05	4.51	4.26	4.27	5.40	ppb
Calcium	43-1	18177.69	19170.02	18061.39	18469.70	3.30	ppb
Calcium	44-1	17426.34	18435.21	17681.48	17847.68	2.94	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	134.84	135.48	136.32	135.55	0.55	ppb
Cobalt	59-2	45.95	45.48	46.04	45.82	0.66	ppb
Copper	63-2	291.56	292.40	295.29	293.08	0.67	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				112		%
Holmium	165-2				110		%
Indium	115-1				111		%
Indium	115-2				110		%
Iron	56-2	120322.37	116221.89	120059.48	118867.91	1.93	ppb
Iron	57-2	118930.23	117727.74	118140.54	118266.17	0.52	ppb
Iron	54-2	120454.43	119832.82	120458.71	120248.66	0.30	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1902.38	1943.90	1888.68	1911.65	1.50	ppb
Lead	207-1	1876.56	1934.69	1829.66	1880.30	2.80	ppb
Lead	208-1	1865.98	1914.18	1851.87	1877.34	1.74	ppb
Lithium	6-1				101		%
Magnesium	24-2	18406.77	18598.88	18836.65	18614.10	1.16	ppb
Manganese	55-2	2629.42	2561.37	2616.96	2602.58	1.39	ppb
Molybdenum	94-1	39.60	39.74	40.85	40.06	1.71	ppb
Molybdenum	95-1	5.76	5.63	5.93	5.78	2.65	ppb
Molybdenum	96-1	10.04	9.45	10.37	9.95	4.68	ppb
Molybdenum	97-1	5.70	5.94	6.02	5.89	2.88	ppb
Molybdenum	98-1	5.87	5.55	5.95	5.79	3.62	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	102.89	101.06	101.10	101.68	1.03	ppb
Phosphorus	31-2	7926.54	7878.81	7914.05	7906.47	0.31	ppb
Potassium	39-2	6622.56	6665.51	6773.20	6687.09	1.16	ppb
Rhodium	103-1				110		%
Rhodium	103-2				109		%
Scandium	45-1				115		%
Scandium	45-2				114		%
Selenium	82-1	4.31	3.55	3.59	3.82	11.18	ppb
Selenium	77-2	27.76	28.61	28.63	28.33	1.76	ppb
Selenium	78-2	9.19	9.91	11.17	10.09	9.91	ppb
Silicon	28-1	198.77	195.74	209.44	201.32	3.57	ppb
Silver	107-1	0.95	1.07	0.98	1.00	6.12	ppb
Silver	109-1	1.02	1.05	1.00	1.03	2.33	ppb
Sodium	23-2	2955.14	2993.58	3081.98	3010.23	2.16	ppb
Strontium	86-1	92.14	89.49	95.22	92.28	3.11	ppb
Strontium	88-1	92.79	90.98	93.81	92.53	1.55	ppb
Sulfur	34-1	910.93	1211.44	926.55	1016.31	16.65	ppb
Terbium	159-1				112		%
Terbium	159-2				111		%
Thallium	203-1	0.68	0.61	0.62	0.64	5.60	ppb
Thallium	205-1	0.65	0.64	0.61	0.63	2.61	ppb
Tin	118-1	7.06	7.20	7.09	7.12	1.03	ppb
Titanium	47-1	949.30	965.95	954.19	956.48	0.89	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	5.80	5.72	5.71	5.75	0.82	ppb
Vanadium	51-2	166.53	166.04	166.73	166.43	0.21	ppb
Yttrium	89-1				117		%
Yttrium	89-2				118		%
Zinc	66-2	946.74	944.86	950.61	947.40	0.31	ppb
Zirconium	90-1	19.14	18.86	20.03	19.34	3.15	ppb
Zirconium	91-1	19.99	19.20	20.69	19.96	3.72	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	58828.24	58995.01	60189.98	59337.74	1.25	ppb
Antimony	121-1	1.68	1.60	1.65	1.64	2.53	ppb
Arsenic	75-2	42.43	41.72	41.77	41.97	0.93	ppb
Barium	135-1	545.41	537.90	551.12	544.81	1.22	ppb
Barium	137-1	552.44	561.62	562.75	558.94	1.01	ppb
Beryllium	9-1	3.41	3.65	3.27	3.44	5.64	ppb
Bismuth	209-1				104		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	5.85	5.97	5.49	5.77	4.30	ppb
Cadmium	106-1	8.11	8.04	7.56	7.90	3.81	ppb
Cadmium	111-1	4.37	4.29	4.39	4.35	1.14	ppb
Calcium	43-1	17332.29	17867.90	17028.92	17409.70	2.44	ppb
Calcium	44-1	16426.94	17171.19	16943.31	16847.15	2.26	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	120.48	122.62	122.10	121.73	0.92	ppb
Cobalt	59-2	41.19	41.81	42.53	41.84	1.60	ppb
Copper	63-2	288.57	291.55	291.23	290.45	0.56	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				110		%
Holmium	165-2				111		%
Indium	115-1				110		%
Indium	115-2				113		%
Iron	56-2	104072.16	106467.46	106545.82	105695.15	1.33	ppb
Iron	57-2	100901.85	104510.64	102528.59	102647.03	1.76	ppb
Iron	54-2	103998.66	104024.99	105079.05	104367.57	0.59	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2011.76	2203.07	1968.27	2061.04	6.06	ppb
Lead	207-1	2004.27	2127.00	1974.76	2035.34	3.97	ppb
Lead	208-1	1980.82	2125.14	1973.60	2026.52	4.22	ppb
Lithium	6-1				100		%
Magnesium	24-2	17658.61	17670.73	17712.85	17680.73	0.16	ppb
Manganese	55-2	2500.29	2557.39	2490.65	2516.11	1.43	ppb
Molybdenum	94-1	41.27	40.81	41.05	41.04	0.56	ppb
Molybdenum	95-1	4.61	4.52	4.50	4.54	1.25	ppb
Molybdenum	96-1	8.76	8.90	9.17	8.94	2.34	ppb
Molybdenum	97-1	4.65	4.56	4.49	4.57	1.76	ppb
Molybdenum	98-1	4.54	4.50	4.63	4.56	1.49	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	96.78	98.13	97.03	97.31	0.73	ppb
Phosphorus	31-2	7898.77	7909.44	8055.32	7954.51	1.10	ppb
Potassium	39-2	6934.28	6854.73	6842.00	6877.00	0.73	ppb
Rhodium	103-1				112		%
Rhodium	103-2				111		%
Scandium	45-1				115		%
Scandium	45-2				116		%
Selenium	82-1	3.97	4.42	3.92	4.10	6.77	ppb
Selenium	77-2	25.66	30.57	28.28	28.17	8.71	ppb
Selenium	78-2	10.87	8.96	9.27	9.70	10.54	ppb
Silicon	28-1	759.70	792.05	812.53	788.09	3.38	ppb
Silver	107-1	1.02	1.03	1.00	1.02	1.59	ppb
Silver	109-1	1.04	1.08	1.04	1.05	2.14	ppb
Sodium	23-2	2770.27	2789.90	2801.33	2787.17	0.56	ppb
Strontium	86-1	88.31	90.01	91.13	89.82	1.58	ppb
Strontium	88-1	89.77	88.89	89.71	89.46	0.55	ppb
Sulfur	34-1	595.85	753.57	308.73	552.72	40.80	ppb
Terbium	159-1				114		%
Terbium	159-2				112		%
Thallium	203-1	0.63	0.65	0.63	0.64	2.29	ppb
Thallium	205-1	0.59	0.65	0.57	0.60	6.68	ppb
Tin	118-1	4.69	4.69	4.86	4.75	2.12	ppb
Titanium	47-1	1062.63	1104.58	1057.40	1074.87	2.41	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	5.72	6.22	5.60	5.85	5.64	ppb
Vanadium	51-2	159.84	161.98	161.66	161.16	0.72	ppb
Yttrium	89-1				118		%
Yttrium	89-2				120		%
Zinc	66-2	987.32	992.24	998.94	992.83	0.59	ppb
Zirconium	90-1	20.07	20.03	20.69	20.26	1.82	ppb
Zirconium	91-1	20.09	20.81	21.38	20.76	3.10	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	12313.02	11926.70	12169.80	12136.51	1.61	ppb
Antimony	121-1	0.29	0.29	0.30	0.29	2.62	ppb
Arsenic	75-2	9.02	9.65	8.95	9.21	4.19	ppb
Barium	135-1	106.73	103.92	103.63	104.76	1.63	ppb
Barium	137-1	107.63	105.79	105.52	106.31	1.08	ppb
Beryllium	9-1	0.74	0.73	0.68	0.71	4.05	ppb
Bismuth	209-1				99		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.59	1.47	1.24	1.43	12.53	ppb
Cadmium	106-1	0.39	-0.34	0.52	0.19	247.60	ppb
Cadmium	111-1	0.79	0.73	0.78	0.77	4.45	ppb
Calcium	43-1	3881.92	4051.01	4048.45	3993.79	2.43	ppb
Calcium	44-1	3855.34	3932.56	3993.27	3927.06	1.76	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	27.82	27.78	28.45	28.02	1.34	ppb
Cobalt	59-2	9.63	9.22	9.52	9.46	2.24	ppb
Copper	63-2	61.71	60.90	61.06	61.22	0.70	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				106		%
Indium	115-2				107		%
Iron	56-2	25156.49	24498.33	24755.06	24803.29	1.34	ppb
Iron	57-2	24402.18	23969.98	24176.09	24182.75	0.89	ppb
Iron	54-2	24665.29	24605.19	24476.27	24582.25	0.39	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	365.76	369.44	359.63	364.94	1.36	ppb
Lead	207-1	355.51	364.09	357.04	358.88	1.27	ppb
Lead	208-1	365.50	380.04	366.87	370.80	2.17	ppb
Lithium	6-1				93		%
Magnesium	24-2	3915.62	3825.87	3832.15	3857.88	1.30	ppb
Manganese	55-2	541.03	539.23	539.94	540.07	0.17	ppb
Molybdenum	94-1	8.94	8.95	9.05	8.98	0.72	ppb
Molybdenum	95-1	1.27	1.21	1.31	1.26	3.89	ppb
Molybdenum	96-1	2.25	2.32	2.22	2.27	2.17	ppb
Molybdenum	97-1	1.30	1.35	1.25	1.30	3.97	ppb
Molybdenum	98-1	1.32	1.24	1.24	1.27	3.48	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	21.25	21.10	21.58	21.31	1.16	ppb
Phosphorus	31-2	1591.19	1626.83	1583.35	1600.45	1.45	ppb
Potassium	39-2	1438.11	1411.81	1466.40	1438.77	1.90	ppb
Rhodium	103-1				107		%
Rhodium	103-2				106		%
Scandium	45-1				106		%
Scandium	45-2				107		%
Selenium	82-1	0.08	0.78	-0.48	0.13	499.63	ppb
Selenium	77-2	6.89	5.40	6.28	6.19	12.13	ppb
Selenium	78-2	2.10	3.82	3.34	3.09	28.82	ppb
Silicon	28-1	49.15	49.36	50.07	49.53	0.97	ppb
Silver	107-1	0.20	0.21	0.21	0.21	1.01	ppb
Silver	109-1	0.23	0.23	0.21	0.22	5.58	ppb
Sodium	23-2	617.77	607.83	621.65	615.75	1.16	ppb
Strontium	86-1	21.25	20.85	21.15	21.08	0.98	ppb
Strontium	88-1	20.52	20.25	20.40	20.39	0.67	ppb
Sulfur	34-1	-1909.70	-1744.68	-1563.55	-1739.31		ppb
Terbium	159-1				105		%
Terbium	159-2				105		%
Thallium	203-1	0.09	0.09	0.09	0.09	2.02	ppb
Thallium	205-1	0.11	0.12	0.11	0.11	6.52	ppb
Tin	118-1	1.60	1.49	1.54	1.54	3.63	ppb
Titanium	47-1	202.53	205.87	207.79	205.40	1.30	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.13	1.18	1.11	1.14	3.24	ppb
Vanadium	51-2	33.86	34.02	35.10	34.33	1.97	ppb
Yttrium	89-1				101		%
Yttrium	89-2				104		%
Zinc	66-2	203.84	193.66	201.76	199.75	2.69	ppb
Zirconium	90-1	4.38	4.35	4.35	4.36	0.43	ppb
Zirconium	91-1	4.49	4.33	4.39	4.40	1.85	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	69570.26	69537.61	68664.84	69257.57	0.74	ppb
Antimony	121-1	463.38	468.67	471.87	467.97	0.92	ppb
Arsenic	75-2	468.96	475.22	473.65	472.61	0.69	ppb
Barium	135-1	2578.80	2689.60	2729.24	2665.88	2.92	ppb
Barium	137-1	2669.75	2695.41	2700.43	2688.53	0.61	ppb
Beryllium	9-1	478.69	477.79	483.72	480.07	0.67	ppb
Bismuth	209-1				102		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	462.83	474.34	469.26	468.81	1.23	ppb
Cadmium	106-1	480.13	483.62	484.20	482.65	0.46	ppb
Cadmium	111-1	466.39	482.12	481.56	476.69	1.87	ppb
Calcium	43-1	63804.64	63977.39	65091.58	64291.20	1.09	ppb
Calcium	44-1	61331.92	61029.49	62774.72	61712.04	1.51	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	588.66	574.17	576.00	579.61	1.36	ppb
Cobalt	59-2	501.55	502.83	493.33	499.24	1.03	ppb
Copper	63-2	4452.48	4564.74	4436.94	4484.72	1.55	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				110		%
Holmium	165-2				109		%
Indium	115-1				110		%
Indium	115-2				109		%
Iron	56-2	126528.95	126137.76	122214.79	124960.50	1.91	ppb
Iron	57-2	123565.70	124750.64	119657.03	122657.79	2.17	ppb
Iron	54-2	126388.81	126154.73	124402.35	125648.63	0.86	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3850.88	3996.89	4140.76	3996.17	3.63	ppb
Lead	207-1	3896.62	3922.43	4116.30	3978.45	3.02	ppb
Lead	208-1	3886.95	3884.08	4081.92	3950.98	2.87	ppb
Lithium	6-1				98		%
Magnesium	24-2	65901.11	65337.51	65410.46	65549.70	0.47	ppb
Manganese	55-2	6601.27	6613.87	6427.33	6547.49	1.59	ppb
Molybdenum	94-1	1925.61	2025.43	2006.87	1985.97	2.67	ppb
Molybdenum	95-1	1456.76	1491.24	1541.91	1496.63	2.86	ppb
Molybdenum	96-1	1513.60	1552.29	1556.98	1540.96	1.55	ppb
Molybdenum	97-1	1448.14	1531.66	1577.75	1519.18	4.32	ppb
Molybdenum	98-1	1443.09	1514.75	1519.60	1492.48	2.87	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	564.01	562.49	553.38	559.96	1.03	ppb
Phosphorus	31-2	8170.82	8208.52	8184.53	8187.96	0.23	ppb
Potassium	39-2	27978.68	27912.57	28677.23	28189.49	1.50	ppb
Rhodium	103-1				111		%
Rhodium	103-2				109		%
Scandium	45-1				118		%
Scandium	45-2				114		%
Selenium	82-1	410.89	424.45	428.04	421.13	2.15	ppb
Selenium	77-2	426.68	451.79	456.12	444.86	3.57	ppb
Selenium	78-2	417.87	426.96	438.62	427.82	2.43	ppb
Silicon	28-1	449.27	446.34	440.14	445.25	1.05	ppb
Silver	107-1	85.03	89.10	89.85	87.99	2.95	ppb
Silver	109-1	84.67	85.91	85.92	85.50	0.83	ppb
Sodium	23-2	47232.59	47009.41	47629.91	47290.64	0.66	ppb
Strontium	86-1	550.21	557.38	575.89	561.16	2.36	ppb
Strontium	88-1	555.67	564.72	581.08	567.16	2.27	ppb
Sulfur	34-1	281.17	202.73	308.50	264.13	20.79	ppb
Terbium	159-1				111		%
Terbium	159-2				110		%
Thallium	203-1	465.18	474.04	489.82	476.35	2.62	ppb
Thallium	205-1	457.98	470.86	486.42	471.75	3.02	ppb
Tin	118-1	419.89	435.04	430.54	428.49	1.82	ppb
Titanium	47-1	985.19	986.42	1018.56	996.73	1.90	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	457.39	450.37	475.13	460.96	2.77	ppb
Vanadium	51-2	622.44	620.42	621.91	621.59	0.17	ppb
Yttrium	89-1				119		%
Yttrium	89-2				118		%
Zinc	66-2	5129.28	5167.53	5078.28	5125.03	0.87	ppb
Zirconium	90-1	420.27	434.01	437.26	430.51	2.09	ppb
Zirconium	91-1	414.19	428.84	437.46	426.83	2.76	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	70601.80	71232.23	70928.72	70920.92	0.44	ppb
Antimony	121-1	460.01	466.71	465.52	464.08	0.77	ppb
Arsenic	75-2	452.16	477.55	474.82	468.17	2.98	ppb
Barium	135-1	2652.93	2768.44	2684.81	2702.06	2.21	ppb
Barium	137-1	2659.73	2805.21	2750.37	2738.44	2.68	ppb
Beryllium	9-1	461.40	467.64	489.24	472.76	3.09	ppb
Bismuth	209-1				103		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	457.97	479.20	456.85	464.67	2.71	ppb
Cadmium	106-1	470.60	482.82	475.26	476.23	1.30	ppb
Cadmium	111-1	460.82	483.82	469.19	471.27	2.47	ppb
Calcium	43-1	64045.71	66593.58	66531.48	65723.59	2.21	ppb
Calcium	44-1	63329.02	64977.75	64902.03	64402.93	1.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	557.97	570.94	574.07	567.66	1.50	ppb
Cobalt	59-2	481.39	499.93	490.75	490.69	1.89	ppb
Copper	63-2	4424.38	4421.39	4429.78	4425.18	0.10	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				111		%
Holmium	165-2				110		%
Indium	115-1				110		%
Indium	115-2				110		%
Iron	56-2	133464.24	134418.60	135249.79	134377.54	0.66	ppb
Iron	57-2	132013.22	133164.45	133633.24	132936.97	0.63	ppb
Iron	54-2	134623.91	135511.05	134269.69	134801.55	0.47	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4205.77	4474.63	4332.56	4337.66	3.10	ppb
Lead	207-1	4206.64	4387.03	4374.69	4322.78	2.33	ppb
Lead	208-1	4238.87	4372.47	4348.11	4319.82	1.65	ppb
Lithium	6-1				96		%
Magnesium	24-2	66832.02	69560.63	68667.71	68353.45	2.04	ppb
Manganese	55-2	6644.64	6649.41	6761.75	6685.27	0.99	ppb
Molybdenum	94-1	1908.52	1935.53	2010.75	1951.60	2.71	ppb
Molybdenum	95-1	1445.49	1477.13	1515.28	1479.30	2.36	ppb
Molybdenum	96-1	1500.28	1503.55	1520.11	1507.98	0.70	ppb
Molybdenum	97-1	1458.15	1460.35	1494.78	1471.09	1.40	ppb
Molybdenum	98-1	1429.03	1446.03	1504.53	1459.86	2.71	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	539.31	550.77	550.57	546.88	1.20	ppb
Phosphorus	31-2	8499.60	8632.25	8636.21	8589.35	0.91	ppb
Potassium	39-2	27391.56	28181.89	28223.41	27932.29	1.68	ppb
Rhodium	103-1				110		%
Rhodium	103-2				109		%
Scandium	45-1				117		%
Scandium	45-2				116		%
Selenium	82-1	402.72	411.38	421.89	412.00	2.33	ppb
Selenium	77-2	416.73	429.57	441.16	429.15	2.85	ppb
Selenium	78-2	408.60	409.46	431.46	416.51	3.11	ppb
Silicon	28-1	295.57	386.61	270.43	317.54	19.25	ppb
Silver	107-1	83.94	88.58	85.72	86.08	2.72	ppb
Silver	109-1	83.99	86.94	84.13	85.02	1.95	ppb
Sodium	23-2	47286.76	47558.40	47837.92	47561.03	0.58	ppb
Strontium	86-1	556.58	554.44	576.54	562.52	2.17	ppb
Strontium	88-1	562.02	562.96	573.39	566.12	1.11	ppb
Sulfur	34-1	521.29	820.04	677.77	673.03	22.20	ppb
Terbium	159-1				111		%
Terbium	159-2				113		%
Thallium	203-1	452.29	483.74	479.73	471.92	3.63	ppb
Thallium	205-1	449.34	467.29	462.42	459.68	2.02	ppb
Tin	118-1	429.28	437.93	435.26	434.16	1.02	ppb
Titanium	47-1	1114.11	1148.63	1169.67	1144.13	2.45	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	414.58	463.98	441.54	440.04	5.62	ppb
Vanadium	51-2	623.88	625.88	627.78	625.84	0.31	ppb
Yttrium	89-1				121		%
Yttrium	89-2				121		%
Zinc	66-2	5415.31	5382.78	5455.73	5417.94	0.67	ppb
Zirconium	90-1	408.68	410.15	422.40	413.74	1.82	ppb
Zirconium	91-1	405.45	411.94	421.00	412.80	1.89	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	48219.87	47623.79	47461.73	47768.46	0.84	ppb
Antimony	121-1	488.72	500.82	497.74	495.76	1.27	ppb
Arsenic	75-2	496.14	475.08	481.25	484.15	2.24	ppb
Barium	135-1	2482.00	2535.58	2533.76	2517.11	1.21	ppb
Barium	137-1	2449.38	2555.64	2562.16	2522.39	2.51	ppb
Beryllium	9-1	503.41	513.99	513.80	510.40	1.19	ppb
Bismuth	209-1				100		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	479.76	503.30	493.22	492.09	2.40	ppb
Cadmium	106-1	484.76	491.68	494.05	490.16	0.98	ppb
Cadmium	111-1	478.67	499.81	497.59	492.02	2.36	ppb
Calcium	43-1	255724.31	259320.51	258022.02	257688.94	0.71	ppb
Calcium	44-1	251811.60	258359.04	250075.27	253415.31	1.72	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	503.93	495.83	493.60	497.79	1.09	ppb
Cobalt	59-2	490.26	503.60	496.73	496.86	1.34	ppb
Copper	63-2	4970.56	4954.09	4944.64	4956.43	0.26	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				108		%
Holmium	165-2				107		%
Indium	115-1				107		%
Indium	115-2				108		%
Iron	56-2	131104.39	130490.11	130512.55	130702.35	0.27	ppb
Iron	57-2	131172.58	130014.06	128604.38	129930.34	0.99	ppb
Iron	54-2	130043.12	131244.59	129838.22	130375.31	0.58	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2469.94	2528.32	2489.67	2495.98	1.19	ppb
Lead	207-1	2431.72	2503.90	2472.68	2469.44	1.47	ppb
Lead	208-1	2438.61	2508.01	2487.69	2478.10	1.44	ppb
Lithium	6-1				93		%
Magnesium	24-2	242028.08	242725.99	246745.38	243833.15	1.04	ppb
Manganese	55-2	5052.60	4976.44	4956.54	4995.19	1.01	ppb
Molybdenum	94-1	4892.25	5135.15	5042.91	5023.44	2.44	ppb
Molybdenum	95-1	4866.15	5109.07	5007.19	4994.14	2.44	ppb
Molybdenum	96-1	4950.20	5076.41	4978.48	5001.70	1.32	ppb
Molybdenum	97-1	4915.63	5196.66	4959.34	5023.88	3.01	ppb
Molybdenum	98-1	4945.72	5174.94	5019.88	5046.85	2.32	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	491.03	486.92	482.18	486.71	0.91	ppb
Phosphorus	31-2	9823.25	9730.42	9730.27	9761.31	0.55	ppb
Potassium	39-2	127817.07	127689.03	127928.98	127811.69	0.09	ppb
Rhodium	103-1				107		%
Rhodium	103-2				106		%
Scandium	45-1				108		%
Scandium	45-2				108		%
Selenium	82-1	468.04	482.11	473.73	474.63	1.49	ppb
Selenium	77-2	496.80	465.51	471.88	478.06	3.46	ppb
Selenium	78-2	477.13	468.77	469.08	471.66	1.00	ppb
Silicon	28-1	491.06	500.60	507.98	499.88	1.70	ppb
Silver	107-1	497.50	501.98	507.43	502.30	0.99	ppb
Silver	109-1	498.80	500.19	496.87	498.62	0.33	ppb
Sodium	23-2	246051.18	252178.32	250161.79	249463.76	1.25	ppb
Strontium	86-1	488.27	505.84	500.01	498.04	1.80	ppb
Strontium	88-1	487.04	513.09	506.82	502.32	2.71	ppb
Sulfur	34-1	8067.56	8214.31	8273.10	8184.99	1.29	ppb
Terbium	159-1				110		%
Terbium	159-2				109		%
Thallium	203-1	503.37	531.11	513.15	515.88	2.73	ppb
Thallium	205-1	499.11	520.70	515.92	511.91	2.22	ppb
Tin	118-1	498.41	513.22	508.87	506.83	1.50	ppb
Titanium	47-1	4913.92	4918.30	4913.83	4915.35	0.05	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	485.89	491.56	491.27	489.57	0.65	ppb
Vanadium	51-2	506.93	501.84	495.59	501.45	1.13	ppb
Yttrium	89-1				104		%
Yttrium	89-2				105		%
Zinc	66-2	4873.28	4821.03	4800.84	4831.72	0.77	ppb
Zirconium	90-1	485.06	503.94	502.33	497.11	2.11	ppb
Zirconium	91-1	484.34	501.93	498.33	494.87	1.88	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-1.94	-2.30	-1.27	-1.84		ppb
Antimony	121-1	0.12	0.13	0.14	0.13	9.71	ppb
Arsenic	75-2	0.01	-0.02	-0.01	-0.01		ppb
Barium	135-1	0.01	0.01	-0.02	0.00		ppb
Barium	137-1	0.03	0.04	0.03	0.04	20.12	ppb
Beryllium	9-1	0.02	0.00	0.01	0.01	61.29	ppb
Bismuth	209-1				98		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	0.01	0.07	0.04	82.14	ppb
Cadmium	106-1	-1.15	-0.92	-0.55	-0.87		ppb
Cadmium	111-1	-0.04	-0.03	-0.01	-0.02		ppb
Calcium	43-1	-9.50	-7.56	-11.28	-9.45		ppb
Calcium	44-1	25.87	26.21	23.82	25.30	5.13	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.03	-0.06	-0.02	-0.04		ppb
Cobalt	59-2	-0.02	-0.02	-0.01	-0.02		ppb
Copper	63-2	0.15	0.16	0.13	0.15	10.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				102		%
Indium	115-1				106		%
Indium	115-2				107		%
Iron	56-2	2.38	0.52	1.38	1.43	65.27	ppb
Iron	57-2	1.44	28.90	-0.71	9.88	167.12	ppb
Iron	54-2	0.31	0.75	0.10	0.39	85.82	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.06	0.07	0.05	0.06	16.94	ppb
Lead	207-1	0.03	0.03	0.05	0.04	37.19	ppb
Lead	208-1	0.04	0.05	0.05	0.05	10.15	ppb
Lithium	6-1				90		%
Magnesium	24-2	-4.79	-4.69	-4.65	-4.71		ppb
Manganese	55-2	0.00	0.01	0.01	0.01	41.82	ppb
Molybdenum	94-1	0.21	0.21	0.17	0.20	11.48	ppb
Molybdenum	95-1	0.18	0.16	0.14	0.16	11.87	ppb
Molybdenum	96-1	0.20	0.18	0.13	0.17	20.67	ppb
Molybdenum	97-1	0.16	0.15	0.15	0.15	2.77	ppb
Molybdenum	98-1	0.21	0.16	0.15	0.17	18.63	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.01	-0.02	-0.05	-0.02		ppb
Phosphorus	31-2	-43.73	-34.49	-38.83	-39.02		ppb
Potassium	39-2	38.71	28.59	28.85	32.05	17.99	ppb
Rhodium	103-1				106		%
Rhodium	103-2				106		%
Scandium	45-1				105		%
Scandium	45-2				104		%
Selenium	82-1	-0.62	-0.41	-0.07	-0.37		ppb
Selenium	77-2	0.00	0.00	0.22	0.07	173.21	ppb
Selenium	78-2	3.16	0.22	1.95	1.78	83.34	ppb
Silicon	28-1	-0.29	-0.28	-0.16	-0.24		ppb
Silver	107-1	0.04	0.03	0.03	0.04	18.56	ppb
Silver	109-1	0.04	0.03	0.03	0.04	16.27	ppb
Sodium	23-2	58.19	51.61	52.01	53.94	6.84	ppb
Strontium	86-1	-0.01	0.01	0.07	0.02	179.23	ppb
Strontium	88-1	0.01	0.01	0.01	0.01	17.53	ppb
Sulfur	34-1	-1034.21	-848.18	-972.05	-951.48		ppb
Terbium	159-1				104		%
Terbium	159-2				104		%
Thallium	203-1	0.10	0.09	0.08	0.09	9.76	ppb
Thallium	205-1	0.11	0.10	0.10	0.10	8.10	ppb
Tin	118-1	-0.01	-0.02	0.03	0.00		ppb
Titanium	47-1	0.05	0.04	0.03	0.04	26.82	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.00	0.01	0.01	20.50	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00	65.52	ppb
Yttrium	89-1				100		%
Yttrium	89-2				102		%
Zinc	66-2	-0.34	-0.53	-1.08	-0.65		ppb
Zirconium	90-1	0.03	0.03	0.04	0.03	16.69	ppb
Zirconium	91-1	0.02	0.00	0.02	0.02	65.58	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	68909.81	71055.86	68658.09	69541.25	1.89	ppb
Antimony	121-1	471.92	464.12	470.26	468.77	0.88	ppb
Arsenic	75-2	454.46	458.68	456.69	456.61	0.46	ppb
Barium	135-1	2738.56	2759.53	2745.42	2747.84	0.39	ppb
Barium	137-1	2730.02	2758.23	2712.95	2733.73	0.84	ppb
Beryllium	9-1	476.53	494.75	476.20	482.49	2.20	ppb
Bismuth	209-1				104		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	468.99	465.24	450.07	461.43	2.17	ppb
Cadmium	106-1	489.59	491.43	472.08	484.37	2.21	ppb
Cadmium	111-1	483.38	471.53	473.50	476.14	1.33	ppb
Calcium	43-1	65078.66	66658.10	65621.30	65786.02	1.22	ppb
Calcium	44-1	62995.03	63532.00	63231.00	63252.68	0.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	582.03	583.83	571.71	579.19	1.13	ppb
Cobalt	59-2	509.48	518.00	493.39	506.96	2.47	ppb
Copper	63-2	4325.94	4428.72	4368.03	4374.23	1.18	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				110		%
Holmium	165-2				111		%
Indium	115-1				109		%
Indium	115-2				111		%
Iron	56-2	124900.79	127121.70	123647.92	125223.47	1.40	ppb
Iron	57-2	123563.93	126626.44	120795.89	123662.09	2.36	ppb
Iron	54-2	126329.61	127760.86	125740.15	126610.21	0.82	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3987.44	3927.26	4092.02	4002.24	2.08	ppb
Lead	207-1	3967.40	3883.72	4056.74	3969.29	2.18	ppb
Lead	208-1	3959.89	3890.64	3998.72	3949.75	1.39	ppb
Lithium	6-1				97		%
Magnesium	24-2	64838.23	66139.72	63868.62	64948.86	1.75	ppb
Manganese	55-2	6568.58	6606.46	6442.87	6539.30	1.31	ppb
Molybdenum	94-1	2032.82	2042.46	2005.83	2027.04	0.94	ppb
Molybdenum	95-1	1511.23	1521.70	1504.77	1512.57	0.56	ppb
Molybdenum	96-1	1566.40	1575.13	1555.76	1565.76	0.62	ppb
Molybdenum	97-1	1499.86	1531.27	1488.92	1506.68	1.46	ppb
Molybdenum	98-1	1491.37	1505.07	1492.97	1496.47	0.50	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	558.29	567.18	547.78	557.75	1.74	ppb
Phosphorus	31-2	8333.84	8214.94	8035.76	8194.85	1.83	ppb
Potassium	39-2	28002.59	29334.01	27865.65	28400.75	2.86	ppb
Rhodium	103-1				110		%
Rhodium	103-2				109		%
Scandium	45-1				115		%
Scandium	45-2				116		%
Selenium	82-1	420.86	421.89	418.17	420.31	0.46	ppb
Selenium	77-2	439.57	454.32	446.19	446.69	1.65	ppb
Selenium	78-2	428.49	435.52	416.14	426.72	2.30	ppb
Silicon	28-1	451.77	486.15	1211.19	716.37	59.87	ppb
Silver	107-1	87.39	87.19	86.11	86.89	0.79	ppb
Silver	109-1	86.09	86.25	83.91	85.42	1.53	ppb
Sodium	23-2	47156.71	47623.12	46594.33	47124.72	1.09	ppb
Strontium	86-1	564.84	573.63	567.46	568.64	0.79	ppb
Strontium	88-1	577.86	578.20	569.87	575.31	0.82	ppb
Sulfur	34-1	877.02	990.08	755.56	874.22	13.42	ppb
Terbium	159-1				110		%
Terbium	159-2				114		%
Thallium	203-1	473.70	468.40	480.45	474.18	1.27	ppb
Thallium	205-1	472.72	459.41	473.22	468.45	1.67	ppb
Tin	118-1	435.09	424.56	423.69	427.78	1.48	ppb
Titanium	47-1	994.24	1035.46	1013.41	1014.37	2.03	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	464.60	443.51	446.71	451.61	2.52	ppb
Vanadium	51-2	614.69	632.61	605.99	617.76	2.20	ppb
Yttrium	89-1				117		%
Yttrium	89-2				120		%
Zinc	66-2	5070.57	5170.55	4885.31	5042.14	2.87	ppb
Zirconium	90-1	430.19	435.74	425.57	430.50	1.18	ppb
Zirconium	91-1	428.63	436.64	429.35	431.54	1.03	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1932.54	1923.16	1930.23	1928.64	0.25	ppb
Antimony	121-1	0.48	0.55	0.59	0.54	10.04	ppb
Arsenic	75-2	1.56	1.66	1.66	1.63	3.71	ppb
Barium	135-1	29.22	31.45	30.88	30.52	3.80	ppb
Barium	137-1	29.85	31.18	30.41	30.48	2.18	ppb
Beryllium	9-1	0.19	0.15	0.18	0.18	11.93	ppb
Bismuth	209-1				102		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.62	0.52	0.67	0.60	12.88	ppb
Cadmium	106-1	-1.01	0.09	0.39	-0.18		ppb
Cadmium	111-1	0.27	0.33	0.35	0.32	13.92	ppb
Calcium	43-1	70731.87	72110.97	72948.23	71930.36	1.56	ppb
Calcium	44-1	70381.58	70565.63	72107.01	71018.07	1.33	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	12.95	12.92	13.05	12.97	0.55	ppb
Cobalt	59-2	2.92	2.94	2.82	2.89	2.19	ppb
Copper	63-2	13.21	13.49	13.61	13.44	1.51	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				106		%
Indium	115-1				111		%
Indium	115-2				111		%
Iron	56-2	6052.47	5951.63	5929.89	5978.00	1.09	ppb
Iron	57-2	6530.90	6461.54	6632.97	6541.80	1.32	ppb
Iron	54-2	6191.41	6135.26	6269.82	6198.83	1.09	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	58.95	59.31	58.34	58.87	0.83	ppb
Lead	207-1	58.24	55.33	54.55	56.04	3.47	ppb
Lead	208-1	57.28	56.69	56.86	56.94	0.54	ppb
Lithium	6-1				92		%
Magnesium	24-2	1829.36	1812.86	1850.51	1830.91	1.03	ppb
Manganese	55-2	168.28	166.09	169.71	168.03	1.09	ppb
Molybdenum	94-1	4.67	4.55	4.59	4.61	1.30	ppb
Molybdenum	95-1	0.87	0.81	0.82	0.84	3.95	ppb
Molybdenum	96-1	1.25	1.28	1.23	1.25	2.02	ppb
Molybdenum	97-1	0.87	0.81	0.86	0.85	3.87	ppb
Molybdenum	98-1	0.90	0.82	0.84	0.85	4.73	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	7.71	7.52	7.49	7.57	1.60	ppb
Phosphorus	31-2	487.59	486.38	473.88	482.62	1.57	ppb
Potassium	39-2	727.26	720.30	745.56	731.04	1.78	ppb
Rhodium	103-1				111		%
Rhodium	103-2				110		%
Scandium	45-1				110		%
Scandium	45-2				108		%
Selenium	82-1	0.15	0.16	-0.32	0.00		ppb
Selenium	77-2	1.48	1.25	1.27	1.34	9.67	ppb
Selenium	78-2	3.27	1.13	1.78	2.06	53.44	ppb
Silicon	28-1	179.22	179.33	176.93	178.49	0.76	ppb
Silver	107-1	0.10	0.12	0.10	0.10	10.03	ppb
Silver	109-1	0.11	0.09	0.11	0.10	9.05	ppb
Sodium	23-2	2023.13	2011.29	2016.01	2016.81	0.30	ppb
Strontium	86-1	42.97	40.99	41.33	41.76	2.54	ppb
Strontium	88-1	40.97	41.13	40.90	41.00	0.28	ppb
Sulfur	34-1	-853.45	-517.67	-418.65	-596.59		ppb
Terbium	159-1				107		%
Terbium	159-2				106		%
Thallium	203-1	0.19	0.14	0.12	0.15	22.44	ppb
Thallium	205-1	0.19	0.15	0.14	0.16	17.25	ppb
Tin	118-1	33.58	35.20	34.72	34.50	2.42	ppb
Titanium	47-1	74.97	75.46	76.65	75.69	1.14	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.26	0.26	0.27	0.26	2.14	ppb
Vanadium	51-2	6.70	6.73	6.87	6.77	1.36	ppb
Yttrium	89-1				106		%
Yttrium	89-2				106		%
Zinc	66-2	78.29	76.57	78.93	77.93	1.57	ppb
Zirconium	90-1	2.20	2.13	2.11	2.14	2.14	ppb
Zirconium	91-1	2.21	2.21	2.09	2.17	3.35	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18233.95	18223.66	18716.34	18391.31	1.53	ppb
Antimony	121-1	0.84	0.83	0.82	0.83	1.29	ppb
Arsenic	75-2	7.61	8.09	7.98	7.89	3.18	ppb
Barium	135-1	475.46	479.26	478.70	477.80	0.43	ppb
Barium	137-1	482.17	482.41	484.91	483.16	0.31	ppb
Beryllium	9-1	7.72	8.15	8.17	8.01	3.19	ppb
Bismuth	209-1				108		%
Bismuth	209-2				112		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	13.71	16.03	14.78	14.84	7.84	ppb
Cadmium	106-1	16.46	16.58	15.53	16.19	3.55	ppb
Cadmium	111-1	14.49	14.90	14.55	14.64	1.50	ppb
Calcium	43-1	98463.50	96864.40	99987.38	98438.43	1.59	ppb
Calcium	44-1	96414.19	97142.18	97864.62	97140.33	0.75	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	63.91	64.00	64.41	64.11	0.41	ppb
Cobalt	59-2	63.21	64.86	63.61	63.89	1.35	ppb
Copper	63-2	75.64	77.70	76.49	76.61	1.35	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				114		%
Holmium	165-2				116		%
Indium	115-1				118		%
Indium	115-2				119		%
Iron	56-2	39822.87	40846.07	39967.64	40212.19	1.38	ppb
Iron	57-2	39221.90	40887.73	39872.87	39994.16	2.10	ppb
Iron	54-2	39545.97	40024.17	40278.26	39949.47	0.93	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	136.76	137.66	146.36	140.26	3.78	ppb
Lead	207-1	133.04	131.29	138.87	134.40	2.95	ppb
Lead	208-1	134.09	134.56	140.39	136.34	2.57	ppb
Lithium	6-1				98		%
Magnesium	24-2	45160.09	46404.79	45222.05	45595.65	1.54	ppb
Manganese	55-2	1184.80	1220.42	1207.66	1204.29	1.50	ppb
Molybdenum	94-1	35.72	35.37	35.82	35.64	0.67	ppb
Molybdenum	95-1	21.80	22.01	22.50	22.10	1.62	ppb
Molybdenum	96-1	23.52	24.07	23.94	23.84	1.21	ppb
Molybdenum	97-1	22.27	22.27	22.88	22.47	1.55	ppb
Molybdenum	98-1	21.83	22.17	22.62	22.21	1.79	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	57.46	57.77	57.19	57.47	0.50	ppb
Phosphorus	31-2	17017.72	16958.73	17099.08	17025.18	0.41	ppb
Potassium	39-2	25599.54	25808.51	26218.35	25875.47	1.22	ppb
Rhodium	103-1				116		%
Rhodium	103-2				115		%
Scandium	45-1				119		%
Scandium	45-2				119		%
Selenium	82-1	3.47	2.99	3.92	3.46	13.53	ppb
Selenium	77-2	21.03	20.19	21.27	20.83	2.74	ppb
Selenium	78-2	7.73	8.14	7.71	7.86	3.09	ppb
Silicon	28-1	15693.13	15742.25	16146.94	15860.77	1.57	ppb
Silver	107-1	0.12	0.13	0.14	0.13	5.30	ppb
Silver	109-1	0.12	0.13	0.12	0.12	2.22	ppb
Sodium	23-2	121780.06	124157.65	121563.42	122500.38	1.17	ppb
Strontium	86-1	321.39	313.66	315.75	316.93	1.26	ppb
Strontium	88-1	319.77	309.33	323.75	317.62	2.35	ppb
Sulfur	34-1	32012.43	31725.44	33161.03	32299.64	2.35	ppb
Terbium	159-1				117		%
Terbium	159-2				118		%
Thallium	203-1	0.64	0.67	0.64	0.65	2.95	ppb
Thallium	205-1	0.65	0.64	0.69	0.66	3.71	ppb
Tin	118-1	3.96	3.18	3.23	3.46	12.63	ppb
Titanium	47-1	172.86	174.57	178.71	175.38	1.71	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	6.50	6.65	6.64	6.60	1.25	ppb
Vanadium	51-2	378.82	385.22	386.39	383.48	1.06	ppb
Yttrium	89-1				127		%
Yttrium	89-2				130		%
Zinc	66-2	400.30	412.54	411.30	408.05	1.65	ppb
Zirconium	90-1	9.34	9.24	9.62	9.40	2.06	ppb
Zirconium	91-1	9.79	9.62	9.67	9.70	0.91	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	31.31	32.01	32.03	31.78	1.28	ppb
Antimony	121-1	0.22	0.20	0.24	0.22	10.75	ppb
Arsenic	75-2	0.70	0.55	0.65	0.63	11.96	ppb
Barium	135-1	444.47	439.87	442.71	442.35	0.52	ppb
Barium	137-1	449.54	447.20	446.70	447.81	0.34	ppb
Beryllium	9-1	0.05	0.06	0.08	0.07	19.37	ppb
Bismuth	209-1				110		%
Bismuth	209-2				115		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.14	0.11	0.25	0.16	45.12	ppb
Cadmium	106-1	-0.22	-1.42	-0.85	-0.83		ppb
Cadmium	111-1	0.01	-0.08	-0.05	-0.04		ppb
Calcium	43-1	30437.21	30431.37	30777.50	30548.69	0.65	ppb
Calcium	44-1	29174.84	29755.24	29737.66	29555.91	1.12	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.17	0.17	0.25	0.20	23.14	ppb
Cobalt	59-2	8.17	8.15	8.37	8.23	1.52	ppb
Copper	63-2	0.19	0.19	0.23	0.20	12.17	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				117		%
Holmium	165-2				116		%
Indium	115-1				122		%
Indium	115-2				121		%
Iron	56-2	7325.65	7227.43	7373.60	7308.89	1.02	ppb
Iron	57-2	7560.07	7448.50	7761.89	7590.15	2.09	ppb
Iron	54-2	7427.97	7450.37	7618.28	7498.87	1.39	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.63	0.63	0.61	0.62	1.59	ppb
Lead	207-1	0.58	0.61	0.60	0.60	3.24	ppb
Lead	208-1	0.61	0.64	0.59	0.61	3.81	ppb
Lithium	6-1				97		%
Magnesium	24-2	9224.57	9189.50	9380.85	9264.97	1.10	ppb
Manganese	55-2	615.95	605.65	631.31	617.64	2.09	ppb
Molybdenum	94-1	11.06	11.07	11.13	11.09	0.35	ppb
Molybdenum	95-1	13.26	13.39	13.44	13.37	0.69	ppb
Molybdenum	96-1	13.14	13.12	13.32	13.19	0.83	ppb
Molybdenum	97-1	13.57	13.61	13.67	13.62	0.36	ppb
Molybdenum	98-1	13.36	13.45	13.70	13.51	1.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	11.27	11.32	11.31	11.30	0.24	ppb
Phosphorus	31-2	226.98	229.39	245.03	233.80	4.19	ppb
Potassium	39-2	7990.90	7844.34	8009.64	7948.29	1.14	ppb
Rhodium	103-1				121		%
Rhodium	103-2				120		%
Scandium	45-1				119		%
Scandium	45-2				118		%
Selenium	82-1	0.36	-0.35	0.10	0.04	934.91	ppb
Selenium	77-2	0.20	0.00	0.20	0.13	86.61	ppb
Selenium	78-2	-0.33	0.73	1.66	0.68	145.44	ppb
Silicon	28-1	3839.49	3977.49	4005.72	3940.90	2.26	ppb
Silver	107-1	0.00	0.00	0.00	0.00	123.40	ppb
Silver	109-1	0.01	0.01	0.01	0.01	20.64	ppb
Sodium	23-2	6206.36	6091.17	6226.42	6174.65	1.18	ppb
Strontium	86-1	285.91	287.41	291.85	288.39	1.07	ppb
Strontium	88-1	292.73	292.92	293.85	293.17	0.21	ppb
Sulfur	34-1	8230.66	8446.84	8665.63	8447.71	2.57	ppb
Terbium	159-1				115		%
Terbium	159-2				119		%
Thallium	203-1	0.22	0.22	0.20	0.21	6.32	ppb
Thallium	205-1	0.21	0.22	0.19	0.21	6.64	ppb
Tin	118-1	1.03	1.02	1.01	1.02	0.94	ppb
Titanium	47-1	0.74	0.81	0.96	0.84	13.20	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	3.31	ppb
Vanadium	51-2	0.31	0.30	0.27	0.29	6.33	ppb
Yttrium	89-1				115		%
Yttrium	89-2				115		%
Zinc	66-2	3.08	2.75	3.42	3.08	10.82	ppb
Zirconium	90-1	0.02	0.02	0.02	0.02	16.29	ppb
Zirconium	91-1	0.02	0.04	0.01	0.03	57.13	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	30.15	31.18	29.07	30.13	3.50	ppb
Antimony	121-1	0.21	0.21	0.19	0.20	5.34	ppb
Arsenic	75-2	0.67	0.61	0.63	0.64	4.58	ppb
Barium	135-1	433.88	429.21	429.39	430.83	0.61	ppb
Barium	137-1	442.43	432.86	433.40	436.23	1.23	ppb
Beryllium	9-1	0.03	0.05	0.05	0.04	24.91	ppb
Bismuth	209-1				110		%
Bismuth	209-2				114		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.08	0.02	0.13	0.08	69.98	ppb
Cadmium	106-1	-0.31	-1.24	-2.42	-1.32		ppb
Cadmium	111-1	-0.02	-0.08	-0.17	-0.09		ppb
Calcium	43-1	28794.97	30064.22	29844.50	29567.90	2.29	ppb
Calcium	44-1	28443.13	29367.44	28899.36	28903.31	1.60	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.28	0.30	0.34	0.31	10.22	ppb
Cobalt	59-2	8.14	8.26	8.13	8.18	0.87	ppb
Copper	63-2	0.16	0.17	0.16	0.16	3.84	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				116		%
Holmium	165-2				117		%
Indium	115-1				120		%
Indium	115-2				121		%
Iron	56-2	7221.71	7415.27	7360.54	7332.51	1.36	ppb
Iron	57-2	7495.69	7604.51	7572.25	7557.48	0.74	ppb
Iron	54-2	7448.35	7593.67	7486.02	7509.35	1.00	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.04	0.01	0.04	0.03	52.12	ppb
Lead	207-1	0.01	0.02	0.04	0.02	52.63	ppb
Lead	208-1	0.02	0.02	0.04	0.02	54.53	ppb
Lithium	6-1				97		%
Magnesium	24-2	8995.10	9433.88	9160.65	9196.55	2.41	ppb
Manganese	55-2	598.84	625.04	627.16	617.02	2.56	ppb
Molybdenum	94-1	14.22	14.30	14.34	14.29	0.44	ppb
Molybdenum	95-1	16.96	17.23	17.19	17.13	0.83	ppb
Molybdenum	96-1	16.69	17.08	16.72	16.83	1.28	ppb
Molybdenum	97-1	17.52	17.48	17.48	17.49	0.14	ppb
Molybdenum	98-1	16.92	17.12	17.38	17.14	1.34	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	10.77	11.35	11.11	11.08	2.66	ppb
Phosphorus	31-2	236.61	247.56	229.17	237.78	3.89	ppb
Potassium	39-2	7950.82	8043.52	8101.19	8031.84	0.94	ppb
Rhodium	103-1				120		%
Rhodium	103-2				119		%
Scandium	45-1				120		%
Scandium	45-2				117		%
Selenium	82-1	0.05	-0.39	-0.12	-0.15		ppb
Selenium	77-2	0.40	0.20	0.00	0.20	100.42	ppb
Selenium	78-2	1.93	1.22	1.84	1.66	23.35	ppb
Silicon	28-1	3781.62	3947.72	3941.38	3890.24	2.42	ppb
Silver	107-1	0.00	0.00	0.00	0.00		ppb
Silver	109-1	0.00	0.00	0.00	0.00	182.43	ppb
Sodium	23-2	6124.63	6217.86	6002.69	6115.06	1.76	ppb
Strontium	86-1	282.42	282.82	285.12	283.46	0.51	ppb
Strontium	88-1	280.89	288.02	285.97	284.96	1.29	ppb
Sulfur	34-1	7843.08	8560.54	8798.80	8400.81	5.92	ppb
Terbium	159-1				116		%
Terbium	159-2				117		%
Thallium	203-1	0.16	0.19	0.18	0.18	7.08	ppb
Thallium	205-1	0.17	0.20	0.19	0.19	7.97	ppb
Tin	118-1	0.95	0.91	0.93	0.93	2.09	ppb
Titanium	47-1	0.79	0.85	0.76	0.80	6.18	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	14.19	ppb
Vanadium	51-2	0.28	0.31	0.27	0.29	7.65	ppb
Yttrium	89-1				116		%
Yttrium	89-2				113		%
Zinc	66-2	3.44	3.56	3.21	3.40	5.16	ppb
Zirconium	90-1	0.01	0.01	0.01	0.01	3.48	ppb
Zirconium	91-1	0.01	0.00	0.01	0.01	93.64	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	3.93	4.88	4.38	4.40	10.87	ppb
Antimony	121-1	0.05	0.05	0.05	0.05	9.02	ppb
Arsenic	75-2	0.09	0.14	0.13	0.12	23.98	ppb
Barium	135-1	90.88	88.42	92.22	90.51	2.13	ppb
Barium	137-1	90.81	91.26	90.82	90.96	0.28	ppb
Beryllium	9-1	0.01	0.01	0.01	0.01	25.75	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.07	0.03	0.04	0.05	39.24	ppb
Cadmium	106-1	-0.43	-1.36	-2.52	-1.44		ppb
Cadmium	111-1	-0.02	-0.08	-0.17	-0.09		ppb
Calcium	43-1	6312.73	6394.15	6111.45	6272.78	2.32	ppb
Calcium	44-1	6125.42	6150.85	6129.17	6135.15	0.22	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.15	0.18	0.20	0.17	15.02	ppb
Cobalt	59-2	1.65	1.69	1.63	1.65	1.79	ppb
Copper	63-2	0.07	0.09	0.10	0.09	17.19	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				106		%
Indium	115-1				110		%
Indium	115-2				112		%
Iron	56-2	1512.42	1563.26	1514.59	1530.09	1.88	ppb
Iron	57-2	1559.28	1622.95	1585.07	1589.10	2.02	ppb
Iron	54-2	1553.40	1622.87	1563.35	1579.88	2.38	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.03	0.05	0.06	0.05	26.70	ppb
Lead	207-1	0.04	0.07	0.04	0.05	32.43	ppb
Lead	208-1	0.04	0.07	0.05	0.05	20.48	ppb
Lithium	6-1				88		%
Magnesium	24-2	1833.19	1904.62	1836.49	1858.10	2.17	ppb
Manganese	55-2	123.79	129.58	124.22	125.86	2.56	ppb
Molybdenum	94-1	2.09	2.07	2.08	2.08	0.50	ppb
Molybdenum	95-1	2.42	2.48	2.58	2.49	3.42	ppb
Molybdenum	96-1	2.40	2.56	2.48	2.48	3.26	ppb
Molybdenum	97-1	2.55	2.61	2.66	2.61	2.27	ppb
Molybdenum	98-1	2.51	2.47	2.60	2.53	2.44	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.18	2.56	2.29	2.34	8.27	ppb
Phosphorus	31-2	5.15	12.33	19.84	12.44	59.04	ppb
Potassium	39-2	1643.83	1745.81	1673.45	1687.70	3.11	ppb
Rhodium	103-1				109		%
Rhodium	103-2				111		%
Scandium	45-1				106		%
Scandium	45-2				108		%
Selenium	82-1	-0.63	-0.80	-1.45	-0.96		ppb
Selenium	77-2	0.00	0.22	0.21	0.14	86.64	ppb
Selenium	78-2	2.01	2.58	3.20	2.59	22.93	ppb
Silicon	28-1	802.69	843.44	817.98	821.37	2.51	ppb
Silver	107-1	0.00	0.00	0.00	0.00		ppb
Silver	109-1	0.00	0.00	0.00	0.00	42.85	ppb
Sodium	23-2	1235.09	1281.81	1233.83	1250.25	2.19	ppb
Strontium	86-1	58.70	60.86	60.52	60.03	1.94	ppb
Strontium	88-1	59.85	59.45	58.77	59.36	0.92	ppb
Sulfur	34-1	-324.86	105.31	-299.48	-173.01		ppb
Terbium	159-1				105		%
Terbium	159-2				109		%
Thallium	203-1	0.03	0.03	0.02	0.03	20.99	ppb
Thallium	205-1	0.03	0.02	0.02	0.03	21.41	ppb
Tin	118-1	0.11	0.10	0.13	0.11	11.66	ppb
Titanium	47-1	0.13	0.13	0.24	0.17	35.70	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	325.05	ppb
Vanadium	51-2	0.05	0.07	0.05	0.06	12.33	ppb
Yttrium	89-1				103		%
Yttrium	89-2				105		%
Zinc	66-2	0.11	0.29	-0.04	0.12	133.66	ppb
Zirconium	90-1	0.01	0.00	0.01	0.01	76.93	ppb
Zirconium	91-1	0.00	0.01	0.00	0.00	208.10	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8465.05	8258.20	8585.44	8436.23	1.96	ppb
Antimony	121-1	467.00	458.93	466.68	464.20	0.98	ppb
Arsenic	75-2	486.97	479.02	499.92	488.64	2.16	ppb
Barium	135-1	2545.48	2562.94	2628.50	2578.97	1.70	ppb
Barium	137-1	2541.26	2563.34	2639.81	2581.47	2.00	ppb
Beryllium	9-1	480.23	486.22	487.43	484.63	0.80	ppb
Bismuth	209-1				112		%
Bismuth	209-2				114		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	451.03	451.49	461.19	454.57	1.26	ppb
Cadmium	106-1	472.03	478.43	481.35	477.27	1.00	ppb
Cadmium	111-1	467.81	471.10	479.14	472.68	1.23	ppb
Calcium	43-1	72916.87	73888.12	75609.58	74138.19	1.84	ppb
Calcium	44-1	72986.01	73902.24	73615.08	73501.11	0.64	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	450.61	458.53	458.77	455.97	1.02	ppb
Cobalt	59-2	463.99	472.54	491.88	476.14	3.00	ppb
Copper	63-2	4256.00	4336.56	4444.78	4345.78	2.18	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				117		%
Holmium	165-2				118		%
Indium	115-1				119		%
Indium	115-2				119		%
Iron	56-2	28317.11	29065.98	29323.46	28902.18	1.81	ppb
Iron	57-2	28484.96	28834.04	29438.65	28919.22	1.67	ppb
Iron	54-2	28948.23	28680.88	29419.69	29016.26	1.29	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2330.54	2381.61	2279.67	2330.61	2.19	ppb
Lead	207-1	2338.78	2404.48	2279.48	2340.91	2.67	ppb
Lead	208-1	2307.98	2361.66	2289.52	2319.72	1.62	ppb
Lithium	6-1				99		%
Magnesium	24-2	56129.63	53857.04	55170.43	55052.37	2.07	ppb
Manganese	55-2	4896.77	4849.27	4953.22	4899.75	1.06	ppb
Molybdenum	94-1	2188.81	2196.98	2209.53	2198.44	0.47	ppb
Molybdenum	95-1	1713.95	1697.23	1718.50	1709.89	0.66	ppb
Molybdenum	96-1	1795.28	1765.43	1762.05	1774.26	1.03	ppb
Molybdenum	97-1	1723.62	1710.47	1721.40	1718.50	0.41	ppb
Molybdenum	98-1	1664.43	1661.74	1713.78	1679.98	1.74	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	460.08	480.55	494.30	478.31	3.60	ppb
Phosphorus	31-2	215.44	227.35	239.27	227.35	5.24	ppb
Potassium	39-2	28716.06	29782.86	30663.80	29720.91	3.28	ppb
Rhodium	103-1				119		%
Rhodium	103-2				118		%
Scandium	45-1				119		%
Scandium	45-2				117		%
Selenium	82-1	466.69	473.18	486.20	475.36	2.09	ppb
Selenium	77-2	445.46	462.21	484.53	464.07	4.22	ppb
Selenium	78-2	464.93	470.71	485.26	473.63	2.21	ppb
Silicon	28-1	3811.12	3807.76	3923.70	3847.53	1.72	ppb
Silver	107-1	85.15	83.91	86.94	85.33	1.79	ppb
Silver	109-1	82.74	83.67	85.60	84.00	1.73	ppb
Sodium	23-2	52762.55	51621.48	52335.89	52239.97	1.10	ppb
Strontium	86-1	801.46	833.84	826.68	820.66	2.07	ppb
Strontium	88-1	804.66	813.70	829.87	816.08	1.57	ppb
Sulfur	34-1	8444.26	8552.63	9381.73	8792.88	5.83	ppb
Terbium	159-1				120		%
Terbium	159-2				118		%
Thallium	203-1	457.87	484.47	462.41	468.25	3.04	ppb
Thallium	205-1	464.42	467.06	453.75	461.74	1.53	ppb
Tin	118-1	432.11	437.88	441.40	437.13	1.07	ppb
Titanium	47-1	2.38	2.61	2.55	2.51	4.76	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	424.43	436.49	437.06	432.66	1.65	ppb
Vanadium	51-2	466.87	464.53	468.60	466.67	0.44	ppb
Yttrium	89-1				114		%
Yttrium	89-2				114		%
Zinc	66-2	4532.46	4472.54	4558.72	4521.24	0.98	ppb
Zirconium	90-1	453.19	453.11	459.03	455.11	0.75	ppb
Zirconium	91-1	441.49	449.04	454.17	448.23	1.42	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8453.04	8689.38	8280.00	8474.14	2.43	ppb
Antimony	121-1	481.44	475.73	489.84	482.33	1.47	ppb
Arsenic	75-2	460.74	479.74	459.72	466.73	2.42	ppb
Barium	135-1	2631.82	2678.47	2645.49	2651.92	0.90	ppb
Barium	137-1	2612.67	2651.45	2675.28	2646.47	1.19	ppb
Beryllium	9-1	491.67	481.51	494.02	489.06	1.36	ppb
Bismuth	209-1				110		%
Bismuth	209-2				113		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	471.42	458.13	483.47	471.00	2.69	ppb
Cadmium	106-1	483.12	482.41	498.24	487.92	1.83	ppb
Cadmium	111-1	477.21	484.88	489.71	483.93	1.30	ppb
Calcium	43-1	73958.31	74839.40	74860.22	74552.64	0.69	ppb
Calcium	44-1	74508.90	73430.22	74555.69	74164.94	0.86	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	453.06	455.70	456.20	454.98	0.37	ppb
Cobalt	59-2	483.70	487.45	472.81	481.32	1.58	ppb
Copper	63-2	4392.75	4352.33	4368.39	4371.16	0.47	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				116		%
Holmium	165-2				117		%
Indium	115-1				120		%
Indium	115-2				118		%
Iron	56-2	28853.06	29367.85	28139.26	28786.72	2.14	ppb
Iron	57-2	29092.36	29391.90	28500.52	28994.93	1.56	ppb
Iron	54-2	29201.50	29862.73	28999.48	29354.57	1.54	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2356.04	2397.98	2441.75	2398.59	1.79	ppb
Lead	207-1	2313.24	2377.26	2468.81	2386.43	3.28	ppb
Lead	208-1	2336.34	2353.97	2422.01	2370.77	1.91	ppb
Lithium	6-1				97		%
Magnesium	24-2	54727.98	56477.47	53745.65	54983.70	2.52	ppb
Manganese	55-2	4967.67	5013.95	4884.11	4955.24	1.33	ppb
Molybdenum	94-1	2176.02	2164.79	2239.43	2193.42	1.83	ppb
Molybdenum	95-1	1682.40	1661.61	1748.22	1697.41	2.66	ppb
Molybdenum	96-1	1722.83	1723.76	1777.96	1741.51	1.81	ppb
Molybdenum	97-1	1698.23	1649.33	1707.40	1684.98	1.85	ppb
Molybdenum	98-1	1692.05	1653.91	1664.95	1670.30	1.18	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	487.84	476.86	482.79	482.50	1.14	ppb
Phosphorus	31-2	231.07	200.27	212.47	214.60	7.23	ppb
Potassium	39-2	30387.46	29281.22	29957.41	29875.36	1.87	ppb
Rhodium	103-1				118		%
Rhodium	103-2				116		%
Scandium	45-1				117		%
Scandium	45-2				116		%
Selenium	82-1	463.83	463.62	478.18	468.54	1.78	ppb
Selenium	77-2	480.17	476.02	467.03	474.41	1.42	ppb
Selenium	78-2	467.13	470.81	457.28	465.07	1.50	ppb
Silicon	28-1	3715.53	3752.30	3767.33	3745.06	0.71	ppb
Silver	107-1	87.30	86.95	88.52	87.59	0.94	ppb
Silver	109-1	85.87	85.29	87.10	86.09	1.07	ppb
Sodium	23-2	51823.65	53636.69	50341.35	51933.90	3.18	ppb
Strontium	86-1	799.31	808.90	837.62	815.27	2.45	ppb
Strontium	88-1	785.72	800.07	811.52	799.10	1.62	ppb
Sulfur	34-1	8791.22	8826.65	9186.44	8934.77	2.45	ppb
Terbium	159-1				117		%
Terbium	159-2				118		%
Thallium	203-1	463.40	478.26	472.35	471.34	1.59	ppb
Thallium	205-1	464.82	479.24	486.90	476.99	2.35	ppb
Tin	118-1	444.07	442.42	451.02	445.84	1.02	ppb
Titanium	47-1	2.46	2.51	2.48	2.48	1.19	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	439.40	448.14	444.14	443.89	0.98	ppb
Vanadium	51-2	475.30	482.75	469.86	475.97	1.36	ppb
Yttrium	89-1				115		%
Yttrium	89-2				114		%
Zinc	66-2	4609.54	4537.85	4492.92	4546.77	1.29	ppb
Zirconium	90-1	441.50	442.70	449.78	444.66	1.01	ppb
Zirconium	91-1	435.40	438.01	447.36	440.26	1.43	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8231.53	8357.83	8088.08	8225.81	1.64	ppb
Antimony	121-1	463.19	470.09	484.58	472.62	2.31	ppb
Arsenic	75-2	491.09	477.86	495.78	488.24	1.90	ppb
Barium	135-1	2572.93	2583.10	2643.94	2599.99	1.48	ppb
Barium	137-1	2566.57	2633.43	2621.50	2607.17	1.37	ppb
Beryllium	9-1	470.87	477.20	484.15	477.41	1.39	ppb
Bismuth	209-1				112		%
Bismuth	209-2				113		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	456.41	471.71	467.54	465.22	1.70	ppb
Cadmium	106-1	471.91	481.26	483.09	478.76	1.25	ppb
Cadmium	111-1	468.21	479.15	484.47	477.28	1.74	ppb
Calcium	43-1	73313.35	71460.09	73999.11	72924.18	1.80	ppb
Calcium	44-1	72279.95	71662.13	74039.42	72660.50	1.70	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	448.78	440.96	450.63	446.79	1.15	ppb
Cobalt	59-2	464.11	461.88	468.98	464.99	0.78	ppb
Copper	63-2	4367.33	4312.16	4315.55	4331.68	0.71	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				118		%
Holmium	165-2				119		%
Indium	115-1				120		%
Indium	115-2				119		%
Iron	56-2	28352.00	28030.68	28161.36	28181.34	0.57	ppb
Iron	57-2	28257.29	27946.41	28220.58	28141.43	0.60	ppb
Iron	54-2	28509.54	28241.87	28212.41	28321.27	0.58	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2333.74	2198.95	2303.99	2278.89	3.11	ppb
Lead	207-1	2360.51	2236.57	2317.15	2304.74	2.73	ppb
Lead	208-1	2345.17	2233.46	2318.80	2299.14	2.54	ppb
Lithium	6-1				97		%
Magnesium	24-2	53665.83	53491.88	53603.63	53587.11	0.16	ppb
Manganese	55-2	4838.61	4780.77	4807.05	4808.81	0.60	ppb
Molybdenum	94-1	2169.88	2207.21	2169.73	2182.27	0.99	ppb
Molybdenum	95-1	1675.13	1681.12	1685.45	1680.57	0.31	ppb
Molybdenum	96-1	1706.33	1761.82	1753.11	1740.42	1.71	ppb
Molybdenum	97-1	1657.16	1687.97	1713.08	1686.07	1.66	ppb
Molybdenum	98-1	1668.58	1659.60	1674.56	1667.58	0.45	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	467.04	463.12	469.59	466.58	0.70	ppb
Phosphorus	31-2	210.22	209.33	217.92	212.49	2.22	ppb
Potassium	39-2	28722.57	28184.64	28651.60	28519.60	1.02	ppb
Rhodium	103-1				120		%
Rhodium	103-2				119		%
Scandium	45-1				118		%
Scandium	45-2				117		%
Selenium	82-1	462.42	474.97	472.04	469.81	1.40	ppb
Selenium	77-2	455.53	459.49	467.84	460.96	1.36	ppb
Selenium	78-2	483.61	469.87	477.94	477.14	1.45	ppb
Silicon	28-1	3655.16	3647.74	3731.78	3678.22	1.26	ppb
Silver	107-1	84.67	85.72	86.80	85.73	1.24	ppb
Silver	109-1	83.91	84.87	85.31	84.70	0.85	ppb
Sodium	23-2	51314.99	50706.37	49769.14	50596.83	1.54	ppb
Strontium	86-1	796.50	809.10	825.09	810.23	1.77	ppb
Strontium	88-1	804.04	820.54	813.48	812.69	1.02	ppb
Sulfur	34-1	8588.15	8440.41	8836.71	8621.76	2.32	ppb
Terbium	159-1				118		%
Terbium	159-2				119		%
Thallium	203-1	465.22	441.95	463.86	457.01	2.86	ppb
Thallium	205-1	466.65	436.84	464.21	455.90	3.63	ppb
Tin	118-1	419.65	429.76	447.43	432.28	3.25	ppb
Titanium	47-1	2.31	2.39	2.39	2.36	2.02	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	450.10	428.31	431.46	436.62	2.70	ppb
Vanadium	51-2	467.07	453.45	461.49	460.67	1.49	ppb
Yttrium	89-1				114		%
Yttrium	89-2				114		%
Zinc	66-2	4481.07	4428.09	4408.06	4439.08	0.85	ppb
Zirconium	90-1	441.45	442.49	448.75	444.23	0.89	ppb
Zirconium	91-1	437.20	446.32	439.73	441.08	1.07	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11.78	13.69	11.32	12.26	10.23	ppb
Antimony	121-1	0.22	0.24	0.22	0.23	4.29	ppb
Arsenic	75-2	0.55	0.62	0.45	0.54	15.99	ppb
Barium	135-1	430.46	428.70	422.37	427.17	1.00	ppb
Barium	137-1	432.93	434.42	424.51	430.62	1.24	ppb
Beryllium	9-1	0.04	0.08	0.07	0.06	31.46	ppb
Bismuth	209-1				112		%
Bismuth	209-2				115		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.08	0.10	0.08	0.09	19.10	ppb
Cadmium	106-1	0.76	-1.09	-1.47	-0.60		ppb
Cadmium	111-1	0.10	-0.03	-0.05	0.01	1078.78	ppb
Calcium	43-1	32617.62	32451.27	33105.39	32724.76	1.04	ppb
Calcium	44-1	32036.44	31527.74	32011.12	31858.43	0.90	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.24	-0.22	-0.21	-0.22		ppb
Cobalt	59-2	8.17	8.12	8.11	8.13	0.44	ppb
Copper	63-2	0.28	0.31	0.29	0.29	4.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				116		%
Holmium	165-2				116		%
Indium	115-1				122		%
Indium	115-2				121		%
Iron	56-2	6723.34	6864.50	6851.09	6812.97	1.14	ppb
Iron	57-2	7056.17	6961.26	7046.31	7021.25	0.74	ppb
Iron	54-2	6928.24	6908.49	7018.19	6951.64	0.84	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.03	0.04	0.04	0.04	17.20	ppb
Lead	207-1	0.03	0.02	0.03	0.03	27.71	ppb
Lead	208-1	0.03	0.03	0.04	0.03	10.09	ppb
Lithium	6-1				97		%
Magnesium	24-2	10055.05	9719.16	9658.98	9811.06	2.18	ppb
Manganese	55-2	638.85	645.51	642.05	642.14	0.52	ppb
Molybdenum	94-1	10.73	10.75	10.83	10.77	0.48	ppb
Molybdenum	95-1	12.63	12.66	12.77	12.68	0.59	ppb
Molybdenum	96-1	12.31	12.49	12.32	12.37	0.80	ppb
Molybdenum	97-1	13.28	12.82	13.05	13.05	1.74	ppb
Molybdenum	98-1	12.77	12.62	12.73	12.71	0.60	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	10.37	10.03	10.43	10.28	2.08	ppb
Phosphorus	31-2	203.19	173.31	199.96	192.15	8.53	ppb
Potassium	39-2	8239.77	8144.91	8192.10	8192.26	0.58	ppb
Rhodium	103-1				120		%
Rhodium	103-2				120		%
Scandium	45-1				116		%
Scandium	45-2				116		%
Selenium	82-1	-0.17	-0.56	0.31	-0.14		ppb
Selenium	77-2	0.20	0.00	0.00	0.07	173.21	ppb
Selenium	78-2	1.30	0.32	-0.37	0.42	201.36	ppb
Silicon	28-1	4031.02	3978.31	4105.50	4038.28	1.58	ppb
Silver	107-1	0.04	0.03	0.03	0.03	19.92	ppb
Silver	109-1	0.03	0.04	0.04	0.04	4.75	ppb
Sodium	23-2	6554.42	6492.20	6508.00	6518.21	0.50	ppb
Strontium	86-1	305.61	306.51	308.91	307.01	0.56	ppb
Strontium	88-1	307.10	308.86	306.07	307.34	0.46	ppb
Sulfur	34-1	11901.55	11941.37	12128.05	11990.32	1.01	ppb
Terbium	159-1				118		%
Terbium	159-2				118		%
Thallium	203-1	0.44	0.43	0.35	0.40	11.73	ppb
Thallium	205-1	0.43	0.39	0.36	0.39	8.04	ppb
Tin	118-1	-0.12	-0.13	-0.12	-0.12		ppb
Titanium	47-1	0.52	0.50	0.62	0.54	12.11	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.01	0.02	0.02	11.06	ppb
Vanadium	51-2	0.16	0.17	0.19	0.17	9.47	ppb
Yttrium	89-1				115		%
Yttrium	89-2				113		%
Zinc	66-2	1.42	1.51	1.20	1.38	11.75	ppb
Zirconium	90-1	0.17	0.15	0.16	0.16	6.38	ppb
Zirconium	91-1	0.15	0.15	0.13	0.14	7.77	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47712.38	49399.13	49080.31	48730.61	1.84	ppb
Antimony	121-1	507.61	497.31	503.37	502.76	1.03	ppb
Arsenic	75-2	479.22	488.20	471.09	479.50	1.78	ppb
Barium	135-1	2504.88	2515.55	2572.82	2531.08	1.44	ppb
Barium	137-1	2530.65	2436.63	2531.14	2499.47	2.18	ppb
Beryllium	9-1	505.66	518.64	505.88	510.06	1.46	ppb
Bismuth	209-1				97		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	490.40	487.02	494.77	490.73	0.79	ppb
Cadmium	106-1	498.00	498.25	507.82	501.36	1.12	ppb
Cadmium	111-1	494.73	486.76	501.11	494.20	1.45	ppb
Calcium	43-1	251828.93	258245.10	248624.64	252899.55	1.94	ppb
Calcium	44-1	247013.86	259872.82	251949.75	252945.48	2.56	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	494.89	497.61	491.42	494.64	0.63	ppb
Cobalt	59-2	492.66	488.24	492.41	491.10	0.51	ppb
Copper	63-2	4882.79	4874.53	4912.64	4889.99	0.41	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				107		%
Indium	115-1				108		%
Indium	115-2				108		%
Iron	56-2	126248.97	129620.24	130623.76	128830.99	1.78	ppb
Iron	57-2	127059.89	125671.06	128569.66	127100.20	1.14	ppb
Iron	54-2	127680.48	130367.80	130209.35	129419.21	1.17	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2473.84	2532.46	2593.55	2533.28	2.36	ppb
Lead	207-1	2480.76	2500.33	2527.69	2502.93	0.94	ppb
Lead	208-1	2478.52	2507.70	2515.24	2500.48	0.78	ppb
Lithium	6-1				91		%
Magnesium	24-2	242864.53	244387.89	242571.69	243274.70	0.40	ppb
Manganese	55-2	4784.52	5011.11	4991.53	4929.05	2.55	ppb
Molybdenum	94-1	4905.43	4876.01	4931.58	4904.34	0.57	ppb
Molybdenum	95-1	4906.83	4904.83	4960.55	4924.07	0.64	ppb
Molybdenum	96-1	4910.86	4860.02	4959.78	4910.22	1.02	ppb
Molybdenum	97-1	4899.38	4870.59	5056.88	4942.28	2.03	ppb
Molybdenum	98-1	4852.19	4895.02	5040.48	4929.23	2.00	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	475.95	482.22	483.15	480.44	0.81	ppb
Phosphorus	31-2	9634.16	9740.36	9689.18	9687.90	0.55	ppb
Potassium	39-2	126033.50	128922.97	129746.27	128234.25	1.52	ppb
Rhodium	103-1				106		%
Rhodium	103-2				106		%
Scandium	45-1				108		%
Scandium	45-2				108		%
Selenium	82-1	464.22	469.07	482.40	471.90	2.00	ppb
Selenium	77-2	463.62	482.35	464.74	470.23	2.23	ppb
Selenium	78-2	475.84	490.49	465.35	477.23	2.65	ppb
Silicon	28-1	488.92	505.66	483.59	492.72	2.34	ppb
Silver	107-1	509.55	498.30	515.90	507.92	1.76	ppb
Silver	109-1	510.31	494.79	515.00	506.70	2.09	ppb
Sodium	23-2	243248.51	246364.43	247509.95	245707.63	0.90	ppb
Strontium	86-1	488.96	491.71	497.83	492.84	0.92	ppb
Strontium	88-1	494.89	496.63	500.28	497.27	0.55	ppb
Sulfur	34-1	8104.60	8603.02	8250.03	8319.22	3.08	ppb
Terbium	159-1				109		%
Terbium	159-2				110		%
Thallium	203-1	507.36	511.49	529.85	516.23	2.32	ppb
Thallium	205-1	499.46	512.32	521.75	511.18	2.19	ppb
Tin	118-1	511.17	503.60	515.90	510.22	1.22	ppb
Titanium	47-1	4771.90	4983.42	4860.88	4872.07	2.18	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	489.93	495.53	494.78	493.41	0.62	ppb
Vanadium	51-2	493.41	503.45	498.43	498.43	1.01	ppb
Yttrium	89-1				105		%
Yttrium	89-2				105		%
Zinc	66-2	4737.45	4804.63	4795.56	4779.22	0.76	ppb
Zirconium	90-1	485.00	495.32	490.98	490.43	1.06	ppb
Zirconium	91-1	481.59	479.79	495.62	485.67	1.78	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-1.90	-2.18	-0.66	-1.58		ppb
Antimony	121-1	0.14	0.15	0.14	0.14	4.89	ppb
Arsenic	75-2	0.00	-0.01	-0.02	-0.01		ppb
Barium	135-1	0.00	-0.02	-0.03	-0.02		ppb
Barium	137-1	0.02	0.03	0.02	0.02	22.99	ppb
Beryllium	9-1	0.02	0.01	0.00	0.01	81.84	ppb
Bismuth	209-1				99		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.19	0.01	0.13	0.11	84.10	ppb
Cadmium	106-1	-1.45	0.11	-1.54	-0.96		ppb
Cadmium	111-1	-0.07	0.05	-0.07	-0.03		ppb
Calcium	43-1	-10.19	-3.75	-15.48	-9.81		ppb
Calcium	44-1	26.80	26.07	26.05	26.31	1.62	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.01	0.03	-0.04	0.00		ppb
Cobalt	59-2	-0.02	-0.01	-0.01	-0.02		ppb
Copper	63-2	0.17	0.18	0.18	0.18	2.61	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				100		%
Indium	115-1				107		%
Indium	115-2				106		%
Iron	56-2	0.11	0.45	0.04	0.20	107.31	ppb
Iron	57-2	0.32	3.66	0.64	1.54	119.44	ppb
Iron	54-2	0.38	1.29	1.60	1.09	58.10	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.04	0.00	0.02	0.02	76.83	ppb
Lead	207-1	0.02	0.02	0.03	0.02	35.98	ppb
Lead	208-1	0.02	0.02	0.02	0.02	8.29	ppb
Lithium	6-1				91		%
Magnesium	24-2	-4.37	-3.99	-4.39	-4.25		ppb
Manganese	55-2	0.01	0.02	0.06	0.03	85.86	ppb
Molybdenum	94-1	0.23	0.17	0.18	0.19	15.33	ppb
Molybdenum	95-1	0.20	0.15	0.12	0.16	23.75	ppb
Molybdenum	96-1	0.17	0.17	0.17	0.17	3.00	ppb
Molybdenum	97-1	0.20	0.20	0.13	0.18	23.74	ppb
Molybdenum	98-1	0.18	0.16	0.16	0.17	6.15	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.01	-0.07	-0.10	-0.05		ppb
Phosphorus	31-2	-46.94	-46.60	-54.17	-49.24		ppb
Potassium	39-2	35.00	34.14	34.99	34.71	1.42	ppb
Rhodium	103-1				106		%
Rhodium	103-2				105		%
Scandium	45-1				102		%
Scandium	45-2				103		%
Selenium	82-1	0.19	-1.08	-0.57	-0.48		ppb
Selenium	77-2	0.00	0.23	0.22	0.15	86.62	ppb
Selenium	78-2	0.46	1.68	0.43	0.86	83.34	ppb
Silicon	28-1	-0.33	-0.05	0.78	0.13	434.28	ppb
Silver	107-1	0.06	0.06	0.07	0.06	6.89	ppb
Silver	109-1	0.07	0.06	0.07	0.06	7.12	ppb
Sodium	23-2	54.74	53.37	50.65	52.92	3.94	ppb
Strontium	86-1	0.03	0.01	0.01	0.02	74.96	ppb
Strontium	88-1	0.01	0.01	0.01	0.01	27.80	ppb
Sulfur	34-1	-1111.68	-988.85	-737.66	-946.06		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	0.07	0.10	0.08	0.08	17.15	ppb
Thallium	205-1	0.08	0.09	0.09	0.09	4.77	ppb
Tin	118-1	-0.02	-0.04	-0.01	-0.02		ppb
Titanium	47-1	0.03	0.00	-0.01	0.01	408.06	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	28.17	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00	25.71	ppb
Yttrium	89-1				99		%
Yttrium	89-2				101		%
Zinc	66-2	-0.37	-0.71	-0.38	-0.49		ppb
Zirconium	90-1	0.03	0.03	0.03	0.03	2.52	ppb
Zirconium	91-1	0.02	0.03	0.02	0.02	22.93	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8206.61	8228.41	8515.06	8316.69	2.07	ppb
Antimony	121-1	475.31	468.89	478.32	474.18	1.02	ppb
Arsenic	75-2	488.53	485.11	501.06	491.56	1.71	ppb
Barium	135-1	2545.09	2540.87	2678.30	2588.09	3.02	ppb
Barium	137-1	2569.77	2582.12	2643.56	2598.48	1.52	ppb
Beryllium	9-1	468.32	471.74	481.09	473.72	1.40	ppb
Bismuth	209-1				114		%
Bismuth	209-2				116		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	455.53	459.87	472.60	462.67	1.92	ppb
Cadmium	106-1	465.23	473.31	482.39	473.64	1.81	ppb
Cadmium	111-1	469.68	469.04	480.34	473.02	1.34	ppb
Calcium	43-1	72575.00	73800.45	71556.81	72644.08	1.55	ppb
Calcium	44-1	73409.27	73560.17	70075.37	72348.27	2.72	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	464.19	454.04	459.80	459.34	1.11	ppb
Cobalt	59-2	480.58	470.62	473.75	474.98	1.07	ppb
Copper	63-2	4246.09	4212.97	4281.58	4246.88	0.81	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				119		%
Holmium	165-2				117		%
Indium	115-1				121		%
Indium	115-2				119		%
Iron	56-2	27324.31	26482.84	27082.96	26963.37	1.61	ppb
Iron	57-2	26612.17	27297.09	27141.25	27016.84	1.33	ppb
Iron	54-2	27210.93	26473.00	27987.89	27223.94	2.78	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2333.15	2348.98	2256.01	2312.71	2.15	ppb
Lead	207-1	2331.17	2364.87	2249.60	2315.21	2.56	ppb
Lead	208-1	2302.85	2348.84	2256.49	2302.72	2.01	ppb
Lithium	6-1				105		%
Magnesium	24-2	55821.72	54897.19	55736.76	55485.22	0.92	ppb
Manganese	55-2	4841.92	4708.52	4879.60	4810.01	1.87	ppb
Molybdenum	94-1	2185.15	2268.97	2197.14	2217.09	2.04	ppb
Molybdenum	95-1	1721.93	1720.36	1686.46	1709.58	1.17	ppb
Molybdenum	96-1	1735.04	1770.46	1750.01	1751.84	1.01	ppb
Molybdenum	97-1	1740.86	1740.79	1685.79	1722.48	1.84	ppb
Molybdenum	98-1	1713.20	1725.29	1715.56	1718.01	0.37	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	468.73	459.64	461.90	463.42	1.02	ppb
Phosphorus	31-2	168.38	163.59	160.19	164.05	2.51	ppb
Potassium	39-2	29116.20	28583.54	29494.40	29064.71	1.57	ppb
Rhodium	103-1				121		%
Rhodium	103-2				119		%
Scandium	45-1				121		%
Scandium	45-2				118		%
Selenium	82-1	465.24	477.89	461.88	468.33	1.80	ppb
Selenium	77-2	465.57	438.90	456.88	453.78	3.00	ppb
Selenium	78-2	478.11	469.71	468.11	471.98	1.14	ppb
Silicon	28-1	3687.49	3708.35	3626.60	3674.15	1.16	ppb
Silver	107-1	85.31	83.82	86.61	85.25	1.64	ppb
Silver	109-1	83.86	84.31	86.53	84.90	1.68	ppb
Sodium	23-2	53657.94	51677.56	54046.00	53127.17	2.39	ppb
Strontium	86-1	820.14	824.67	793.00	812.60	2.11	ppb
Strontium	88-1	819.54	813.07	810.10	814.24	0.59	ppb
Sulfur	34-1	8744.39	8753.72	8625.65	8707.92	0.82	ppb
Terbium	159-1				120		%
Terbium	159-2				119		%
Thallium	203-1	458.67	480.56	451.51	463.58	3.26	ppb
Thallium	205-1	458.53	464.28	453.22	458.67	1.21	ppb
Tin	118-1	431.87	430.60	439.69	434.05	1.14	ppb
Titanium	47-1	2.06	1.95	2.04	2.02	2.89	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	430.20	433.68	434.63	432.84	0.54	ppb
Vanadium	51-2	469.62	466.57	477.97	471.39	1.25	ppb
Yttrium	89-1				114		%
Yttrium	89-2				113		%
Zinc	66-2	4440.03	4389.32	4453.85	4427.73	0.77	ppb
Zirconium	90-1	456.54	455.64	444.10	452.10	1.53	ppb
Zirconium	91-1	443.30	453.84	442.89	446.68	1.39	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8780.39	8291.59	7892.02	8321.34	5.35	ppb
Antimony	121-1	463.38	462.52	473.73	466.55	1.34	ppb
Arsenic	75-2	498.66	504.98	478.96	494.20	2.75	ppb
Barium	135-1	2452.35	2570.46	2546.10	2522.97	2.47	ppb
Barium	137-1	2477.58	2488.75	2587.79	2518.04	2.41	ppb
Beryllium	9-1	464.02	477.80	466.29	469.37	1.57	ppb
Bismuth	209-1				115		%
Bismuth	209-2				115		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	428.81	435.35	461.24	441.80	3.88	ppb
Cadmium	106-1	456.52	462.72	476.45	465.23	2.19	ppb
Cadmium	111-1	449.85	463.08	474.48	462.47	2.67	ppb
Calcium	43-1	72034.68	70871.67	74585.29	72497.21	2.62	ppb
Calcium	44-1	71697.41	72099.40	73948.86	72581.89	1.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	473.09	457.13	442.08	457.43	3.39	ppb
Cobalt	59-2	476.28	468.31	450.40	465.00	2.85	ppb
Copper	63-2	4344.44	4312.19	4240.24	4298.96	1.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				119		%
Holmium	165-2				118		%
Indium	115-1				120		%
Indium	115-2				120		%
Iron	56-2	27482.67	26860.49	25499.18	26614.11	3.81	ppb
Iron	57-2	27204.63	27043.20	25774.50	26674.11	2.94	ppb
Iron	54-2	28613.03	27022.25	26023.87	27219.71	4.80	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2254.76	2248.80	2268.17	2257.24	0.44	ppb
Lead	207-1	2222.64	2304.00	2278.89	2268.51	1.84	ppb
Lead	208-1	2253.69	2274.68	2270.75	2266.37	0.49	ppb
Lithium	6-1				102		%
Magnesium	24-2	56477.43	55822.10	55097.47	55799.00	1.24	ppb
Manganese	55-2	5161.33	4883.13	4743.71	4929.39	4.31	ppb
Molybdenum	94-1	2179.86	2228.40	2155.95	2188.07	1.69	ppb
Molybdenum	95-1	1660.15	1685.57	1679.04	1674.92	0.79	ppb
Molybdenum	96-1	1694.97	1744.56	1702.89	1714.14	1.55	ppb
Molybdenum	97-1	1688.31	1684.24	1681.90	1684.82	0.19	ppb
Molybdenum	98-1	1688.13	1673.74	1676.68	1679.52	0.45	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	471.32	461.05	452.18	461.52	2.08	ppb
Phosphorus	31-2	120.39	107.16	119.74	115.76	6.44	ppb
Potassium	39-2	29768.02	28476.39	27930.90	28725.10	3.28	ppb
Rhodium	103-1				119		%
Rhodium	103-2				118		%
Scandium	45-1				119		%
Scandium	45-2				118		%
Selenium	82-1	463.42	474.08	466.14	467.88	1.18	ppb
Selenium	77-2	474.54	481.68	433.95	463.39	5.56	ppb
Selenium	78-2	461.07	477.97	463.56	467.53	1.95	ppb
Silicon	28-1	3770.59	3641.75	3782.48	3731.61	2.09	ppb
Silver	107-1	80.61	83.21	84.52	82.78	2.40	ppb
Silver	109-1	80.69	82.67	83.77	82.38	1.89	ppb
Sodium	23-2	54187.96	53427.44	52347.97	53321.12	1.73	ppb
Strontium	86-1	812.16	820.20	826.39	819.59	0.87	ppb
Strontium	88-1	813.99	825.07	792.36	810.47	2.05	ppb
Sulfur	34-1	8891.58	8766.14	9104.25	8920.66	1.92	ppb
Terbium	159-1				121		%
Terbium	159-2				121		%
Thallium	203-1	450.76	463.26	452.74	455.59	1.48	ppb
Thallium	205-1	455.36	460.54	447.09	454.33	1.49	ppb
Tin	118-1	411.40	421.18	424.83	419.13	1.66	ppb
Titanium	47-1	2.14	1.91	1.94	1.99	6.26	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	420.58	429.55	425.94	425.36	1.06	ppb
Vanadium	51-2	484.81	468.36	463.23	472.13	2.39	ppb
Yttrium	89-1				114		%
Yttrium	89-2				113		%
Zinc	66-2	4500.19	4428.66	4368.62	4432.49	1.49	ppb
Zirconium	90-1	428.87	448.53	441.76	439.72	2.27	ppb
Zirconium	91-1	442.08	446.45	437.48	442.00	1.01	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9.57	9.76	9.59	9.64	1.06	ppb
Antimony	121-1	0.20	0.23	0.21	0.22	6.05	ppb
Arsenic	75-2	14.64	14.83	15.59	15.02	3.35	ppb
Barium	135-1	65.78	66.03	66.52	66.11	0.57	ppb
Barium	137-1	66.38	66.83	65.24	66.15	1.24	ppb
Beryllium	9-1	0.01	0.03	0.02	0.02	54.50	ppb
Bismuth	209-1				114		%
Bismuth	209-2				121		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.14	0.05	0.13	0.11	44.07	ppb
Cadmium	106-1	0.27	-0.71	0.01	-0.14		ppb
Cadmium	111-1	0.07	0.03	0.06	0.05	43.91	ppb
Calcium	43-1	4436.56	4588.31	4461.72	4495.53	1.81	ppb
Calcium	44-1	4553.76	4761.43	4722.95	4679.38	2.36	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.30	1.19	1.18	1.22	5.32	ppb
Cobalt	59-2	1.07	1.12	1.08	1.09	2.15	ppb
Copper	63-2	2.90	2.80	2.77	2.82	2.41	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				118		%
Holmium	165-2				119		%
Indium	115-1				122		%
Indium	115-2				122		%
Iron	56-2	21269.55	20777.57	20762.77	20936.63	1.38	ppb
Iron	57-2	20872.96	20466.74	20079.87	20473.19	1.94	ppb
Iron	54-2	21106.39	20889.82	20582.83	20859.68	1.26	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	7.09	7.11	7.17	7.12	0.63	ppb
Lead	207-1	6.82	7.06	6.72	6.87	2.56	ppb
Lead	208-1	6.78	7.02	6.80	6.87	1.96	ppb
Lithium	6-1				98		%
Magnesium	24-2	1538.83	1525.92	1537.11	1533.96	0.46	ppb
Manganese	55-2	840.74	840.07	815.39	832.07	1.74	ppb
Molybdenum	94-1	1.56	0.90	0.87	1.11	35.06	ppb
Molybdenum	95-1	0.32	0.31	0.29	0.31	5.05	ppb
Molybdenum	96-1	0.58	0.53	0.46	0.52	11.67	ppb
Molybdenum	97-1	0.40	0.32	0.26	0.32	21.55	ppb
Molybdenum	98-1	0.31	0.31	0.27	0.30	8.08	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.69	2.62	2.79	2.70	3.13	ppb
Phosphorus	31-2	159.20	143.49	154.28	152.32	5.27	ppb
Potassium	39-2	513.39	510.84	517.52	513.92	0.66	ppb
Rhodium	103-1				119		%
Rhodium	103-2				123		%
Scandium	45-1				115		%
Scandium	45-2				119		%
Selenium	82-1	31.22	35.07	36.37	34.22	7.82	ppb
Selenium	77-2	39.49	40.78	51.10	43.79	14.53	ppb
Selenium	78-2	42.49	45.29	43.37	43.71	3.27	ppb
Silicon	28-1	192.67	224.90	215.18	210.92	7.84	ppb
Silver	107-1	0.07	0.09	0.09	0.08	10.53	ppb
Silver	109-1	0.09	0.10	0.10	0.10	3.13	ppb
Sodium	23-2	12686.37	12721.31	12554.85	12654.18	0.69	ppb
Strontium	86-1	18.69	19.64	18.52	18.95	3.18	ppb
Strontium	88-1	18.36	18.64	17.90	18.30	2.05	ppb
Sulfur	34-1	-906.84	-775.91	-1188.63	-957.12		ppb
Terbium	159-1				116		%
Terbium	159-2				121		%
Thallium	203-1	0.56	0.70	0.61	0.62	11.94	ppb
Thallium	205-1	0.60	0.67	0.61	0.63	5.80	ppb
Tin	118-1	0.69	0.62	0.62	0.64	6.53	ppb
Titanium	47-1	0.07	0.10	0.10	0.09	17.60	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.00	0.00	0.00	56.79	ppb
Vanadium	51-2	0.04	0.02	0.03	0.03	22.10	ppb
Yttrium	89-1				110		%
Yttrium	89-2				115		%
Zinc	66-2	43.04	43.24	42.53	42.94	0.85	ppb
Zirconium	90-1	0.68	0.38	0.35	0.47	38.11	ppb
Zirconium	91-1	0.73	0.36	0.36	0.48	44.11	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	828.22	820.15	815.68	821.35	0.77	ppb
Antimony	121-1	0.66	0.67	0.66	0.66	0.72	ppb
Arsenic	75-2	1.89	1.60	1.71	1.74	8.39	ppb
Barium	135-1	28.88	30.02	30.08	29.66	2.28	ppb
Barium	137-1	28.79	30.12	30.10	29.67	2.57	ppb
Beryllium	9-1	0.03	0.03	0.02	0.03	33.39	ppb
Bismuth	209-1				114		%
Bismuth	209-2				114		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.38	0.29	0.35	0.34	13.54	ppb
Cadmium	106-1	0.07	0.47	1.14	0.56	95.94	ppb
Cadmium	111-1	0.19	0.24	0.25	0.23	13.32	ppb
Calcium	43-1	4070.17	4219.59	4223.26	4171.01	2.09	ppb
Calcium	44-1	4040.80	4079.09	4184.22	4101.37	1.81	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	39.04	38.99	38.67	38.90	0.52	ppb
Cobalt	59-2	2.25	2.21	2.22	2.23	0.86	ppb
Copper	63-2	49.12	49.09	48.25	48.82	1.01	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				118		%
Holmium	165-2				117		%
Indium	115-1				121		%
Indium	115-2				119		%
Iron	56-2	279772.04	274621.52	272334.17	275575.91	1.38	ppb
Iron	57-2	269938.76	269753.66	270853.07	270181.83	0.22	ppb
Iron	54-2	279102.46	271619.07	275205.21	275308.91	1.36	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	15.06	15.46	15.56	15.36	1.74	ppb
Lead	207-1	14.31	14.18	14.73	14.41	2.00	ppb
Lead	208-1	14.55	14.98	15.00	14.84	1.72	ppb
Lithium	6-1				103		%
Magnesium	24-2	485.26	477.65	479.78	480.90	0.82	ppb
Manganese	55-2	679.48	668.45	652.40	666.78	2.04	ppb
Molybdenum	94-1	6.00	6.00	6.27	6.09	2.52	ppb
Molybdenum	95-1	5.74	5.70	5.75	5.73	0.40	ppb
Molybdenum	96-1	7.29	7.28	7.43	7.33	1.19	ppb
Molybdenum	97-1	5.89	5.82	5.78	5.83	0.99	ppb
Molybdenum	98-1	5.67	5.70	5.73	5.70	0.48	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	6.03	6.05	5.91	5.99	1.28	ppb
Phosphorus	31-2	-21.64	-18.41	-15.67	-18.57		ppb
Potassium	39-2	6331.95	6214.21	6310.04	6285.40	1.00	ppb
Rhodium	103-1				119		%
Rhodium	103-2				116		%
Scandium	45-1				117		%
Scandium	45-2				116		%
Selenium	82-1	10.42	10.05	8.69	9.72	9.37	ppb
Selenium	77-2	7.27	8.71	8.86	8.28	10.64	ppb
Selenium	78-2	7.81	8.36	7.74	7.97	4.27	ppb
Silicon	28-1	246.77	257.39	256.94	253.70	2.37	ppb
Silver	107-1	0.09	0.09	0.11	0.09	12.35	ppb
Silver	109-1	0.09	0.10	0.07	0.09	16.52	ppb
Sodium	23-2	216115.03	215628.47	214259.29	215334.26	0.45	ppb
Strontium	86-1	17.80	17.62	18.35	17.92	2.10	ppb
Strontium	88-1	17.50	18.13	17.92	17.85	1.79	ppb
Sulfur	34-1	345.40	643.13	746.53	578.35	36.01	ppb
Terbium	159-1				119		%
Terbium	159-2				119		%
Thallium	203-1	0.11	0.09	0.08	0.10	13.73	ppb
Thallium	205-1	0.11	0.11	0.10	0.11	1.85	ppb
Tin	118-1	5.37	5.60	5.74	5.57	3.34	ppb
Titanium	47-1	11.50	11.71	11.89	11.70	1.69	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	16.08	ppb
Vanadium	51-2	2.17	2.20	2.26	2.21	2.10	ppb
Yttrium	89-1				113		%
Yttrium	89-2				114		%
Zinc	66-2	455.07	458.21	456.19	456.49	0.35	ppb
Zirconium	90-1	0.60	0.61	0.61	0.61	1.35	ppb
Zirconium	91-1	0.61	0.60	0.65	0.62	4.33	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	333.35	349.83	349.41	344.20	2.73	ppb
Antimony	121-1	7.47	8.13	7.43	7.68	5.10	ppb
Arsenic	75-2	26.90	29.27	28.04	28.07	4.21	ppb
Barium	135-1	37.64	40.07	36.33	38.01	5.00	ppb
Barium	137-1	38.00	39.89	36.49	38.13	4.47	ppb
Beryllium	9-1	0.01	0.03	0.01	0.02	65.54	ppb
Bismuth	209-1				81		%
Bismuth	209-2				86		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	164.25	168.81	150.20	161.09	6.02	ppb
Cadmium	106-1	-0.76	0.26	-0.82	-0.44		ppb
Cadmium	111-1	5.56	5.98	5.16	5.57	7.29	ppb
Calcium	43-1	2074.84	2197.58	2065.44	2112.62	3.49	ppb
Calcium	44-1	2586.76	2773.93	2597.32	2652.67	3.96	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	6.30	6.55	6.78	6.54	3.69	ppb
Cobalt	59-2	0.30	0.32	0.29	0.30	5.62	ppb
Copper	63-2	196.04	208.90	210.07	205.01	3.80	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				87		%
Holmium	165-2				90		%
Indium	115-1				89		%
Indium	115-2				91		%
Iron	56-2	398.11	417.55	415.90	410.52	2.63	ppb
Iron	57-2	394.01	425.34	430.80	416.71	4.76	ppb
Iron	54-2	405.11	427.25	432.41	421.59	3.44	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	18.96	19.69	18.42	19.02	3.34	ppb
Lead	207-1	18.96	19.55	18.66	19.06	2.39	ppb
Lead	208-1	18.76	19.48	18.48	18.91	2.73	ppb
Lithium	6-1				78		%
Magnesium	24-2	653.95	702.32	710.68	688.98	4.44	ppb
Manganese	55-2	91.60	97.59	99.11	96.10	4.13	ppb
Molybdenum	94-1	37425.33	39561.10	36962.29	37982.91	3.65	ppb
Molybdenum	95-1	45854.35	47834.11	44576.97	46088.48	3.56	ppb
Molybdenum	96-1	45227.47	47318.99	44618.40	45721.62	3.10	ppb
Molybdenum	97-1	46725.49	47903.77	44812.00	46480.42	3.36	ppb
Molybdenum	98-1	46401.61	48327.27	44225.93	46318.27	4.43	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.35	1.59	1.31	1.41	10.62	ppb
Phosphorus	31-2	4675.07	4986.31	5098.83	4920.07	4.46	ppb
Potassium	39-2	923164.42	970708.80	979442.40	957771.87	3.16	ppb
Rhodium	103-1				87		%
Rhodium	103-2				90		%
Scandium	45-1				86		%
Scandium	45-2				91		%
Selenium	82-1	47.33	55.24	53.76	52.11	8.07	ppb
Selenium	77-2	39.90	53.34	44.24	45.83	14.97	ppb
Selenium	78-2	45.52	48.46	47.01	46.99	3.13	ppb
Silicon	28-1	1783.22	2157.18	2204.87	2048.42	11.27	ppb
Silver	107-1	0.50	0.49	0.48	0.49	2.78	ppb
Silver	109-1	0.50	0.55	0.49	0.51	5.76	ppb
Sodium	23-2	385261.84	405879.89	399284.71	396808.81	2.65	ppb
Strontium	86-1	71.58	74.66	70.54	72.26	2.96	ppb
Strontium	88-1	71.32	74.49	69.71	71.84	3.38	ppb
Sulfur	34-1	6792.26	7169.41	6241.28	6734.32	6.93	ppb
Terbium	159-1				88		%
Terbium	159-2				91		%
Thallium	203-1	0.30	0.36	0.39	0.35	13.26	ppb
Thallium	205-1	0.32	0.37	0.39	0.36	10.58	ppb
Tin	118-1	5.80	6.11	5.57	5.83	4.69	ppb
Titanium	47-1	4.61	4.71	4.79	4.70	1.93	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.03	0.02	0.02	9.85	ppb
Vanadium	51-2	1.84	1.97	1.97	1.92	3.91	ppb
Yttrium	89-1				84		%
Yttrium	89-2				88		%
Zinc	66-2	773.92	819.22	822.94	805.36	3.39	ppb
Zirconium	90-1	0.26	0.27	0.24	0.26	5.15	ppb
Zirconium	91-1	0.27	0.27	0.25	0.26	3.19	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	10166.17	10431.14	10835.91	10477.74	3.22	ppb
Antimony	121-1	0.95	0.91	0.97	0.94	3.06	ppb
Arsenic	75-2	23.75	23.57	24.07	23.79	1.06	ppb
Barium	135-1	30.04	29.78	29.33	29.71	1.21	ppb
Barium	137-1	30.28	30.45	29.59	30.11	1.51	ppb
Beryllium	9-1	0.02	0.02	0.01	0.02	37.59	ppb
Bismuth	209-1				84		%
Bismuth	209-2				90		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	34.83	37.13	32.05	34.67	7.34	ppb
Cadmium	106-1	0.28	-0.02	0.49	0.25	100.57	ppb
Cadmium	111-1	1.81	1.47	1.79	1.69	11.32	ppb
Calcium	43-1	8155.62	8285.48	8013.96	8151.69	1.67	ppb
Calcium	44-1	8442.58	8730.97	8515.23	8562.93	1.75	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	5.30	5.68	5.74	5.58	4.33	ppb
Cobalt	59-2	0.13	0.15	0.18	0.15	14.97	ppb
Copper	63-2	291.68	298.23	309.56	299.82	3.02	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				89		%
Holmium	165-2				94		%
Indium	115-1				90		%
Indium	115-2				95		%
Iron	56-2	469.44	473.17	493.26	478.62	2.68	ppb
Iron	57-2	497.41	513.15	538.94	516.50	4.06	ppb
Iron	54-2	478.70	486.16	507.90	490.92	3.09	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.89	1.96	1.96	1.94	2.17	ppb
Lead	207-1	1.88	1.97	1.97	1.94	2.56	ppb
Lead	208-1	1.88	1.98	1.95	1.93	2.68	ppb
Lithium	6-1				80		%
Magnesium	24-2	623.89	638.77	669.80	644.15	3.64	ppb
Manganese	55-2	444.64	450.45	464.90	453.33	2.30	ppb
Molybdenum	94-1	8024.40	8189.58	8312.45	8175.48	1.77	ppb
Molybdenum	95-1	9840.27	10085.90	10079.46	10001.88	1.40	ppb
Molybdenum	96-1	9672.01	9907.12	9664.18	9747.77	1.42	ppb
Molybdenum	97-1	10021.57	10222.13	9868.27	10037.32	1.77	ppb
Molybdenum	98-1	9709.18	10189.66	9797.16	9898.67	2.58	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.68	1.75	1.61	1.68	4.13	ppb
Phosphorus	31-2	224.96	251.26	228.18	234.80	6.11	ppb
Potassium	39-2	742586.13	746898.28	775845.37	755109.93	2.40	ppb
Rhodium	103-1				88		%
Rhodium	103-2				93		%
Scandium	45-1				89		%
Scandium	45-2				95		%
Selenium	82-1	53.79	61.79	62.97	59.51	8.39	ppb
Selenium	77-2	39.73	57.55	53.63	50.31	18.61	ppb
Selenium	78-2	54.58	53.34	57.41	55.11	3.78	ppb
Silicon	28-1	1029.34	1230.88	1256.43	1172.22	10.61	ppb
Silver	107-1	0.05	0.05	0.04	0.04	17.04	ppb
Silver	109-1	0.05	0.06	0.05	0.06	4.47	ppb
Sodium	23-2	306983.91	313015.39	317108.35	312369.21	1.63	ppb
Strontium	86-1	29.15	29.45	29.53	29.38	0.69	ppb
Strontium	88-1	28.64	28.99	29.15	28.93	0.90	ppb
Sulfur	34-1	11236.16	11600.27	10644.20	11160.21	4.32	ppb
Terbium	159-1				89		%
Terbium	159-2				94		%
Thallium	203-1	0.45	0.46	0.47	0.46	2.47	ppb
Thallium	205-1	0.47	0.45	0.45	0.46	2.18	ppb
Tin	118-1	5.61	5.62	5.34	5.52	2.86	ppb
Titanium	47-1	2.00	2.22	2.06	2.09	5.43	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.06	0.06	0.06	0.06	7.14	ppb
Vanadium	51-2	4.26	4.46	4.36	4.36	2.35	ppb
Yttrium	89-1				85		%
Yttrium	89-2				91		%
Zinc	66-2	1798.35	1832.34	1894.85	1841.85	2.66	ppb
Zirconium	90-1	0.34	0.37	0.37	0.36	4.15	ppb
Zirconium	91-1	0.34	0.34	0.36	0.35	3.22	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	74.95	76.33	74.99	75.42	1.04	ppb
Antimony	121-1	2.22	2.20	2.13	2.18	2.27	ppb
Arsenic	75-2	16.15	16.04	16.80	16.33	2.51	ppb
Barium	135-1	21.47	20.03	19.88	20.46	4.28	ppb
Barium	137-1	21.38	21.33	20.27	20.99	2.98	ppb
Beryllium	9-1	0.02	0.02	0.00	0.02	71.73	ppb
Bismuth	209-1				113		%
Bismuth	209-2				114		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.35	0.33	0.56	0.41	30.10	ppb
Cadmium	106-1	-1.18	-0.94	-1.58	-1.23		ppb
Cadmium	111-1	0.01	-0.01	-0.07	-0.02		ppb
Calcium	43-1	9273.34	9803.65	9688.65	9588.55	2.91	ppb
Calcium	44-1	9747.33	9737.62	9497.33	9660.76	1.47	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.03	1.09	1.09	1.07	3.18	ppb
Cobalt	59-2	0.84	0.82	0.81	0.83	2.24	ppb
Copper	63-2	2.33	2.22	2.35	2.30	3.11	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				117		%
Holmium	165-2				116		%
Indium	115-1				119		%
Indium	115-2				119		%
Iron	56-2	48912.26	50651.03	48978.42	49513.90	1.99	ppb
Iron	57-2	47490.86	48980.49	48513.63	48328.32	1.58	ppb
Iron	54-2	49264.25	49574.11	50086.98	49641.78	0.84	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.97	0.96	0.91	0.94	3.12	ppb
Lead	207-1	0.85	0.91	0.89	0.88	3.56	ppb
Lead	208-1	0.87	0.93	0.88	0.90	3.64	ppb
Lithium	6-1				106		%
Magnesium	24-2	1374.60	1406.46	1384.28	1388.44	1.18	ppb
Manganese	55-2	239.74	245.68	240.00	241.81	1.39	ppb
Molybdenum	94-1	3.05	2.85	2.66	2.85	6.94	ppb
Molybdenum	95-1	3.30	3.03	2.82	3.05	7.82	ppb
Molybdenum	96-1	3.55	3.39	3.11	3.35	6.60	ppb
Molybdenum	97-1	3.36	3.14	2.91	3.14	7.16	ppb
Molybdenum	98-1	3.35	2.97	2.85	3.06	8.59	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	3.56	3.52	3.49	3.52	0.93	ppb
Phosphorus	31-2	24196.11	24667.17	24448.58	24437.29	0.96	ppb
Potassium	39-2	232727.73	236659.78	229772.13	233053.21	1.48	ppb
Rhodium	103-1				120		%
Rhodium	103-2				118		%
Scandium	45-1				119		%
Scandium	45-2				117		%
Selenium	82-1	28.10	30.98	30.22	29.77	5.02	ppb
Selenium	77-2	34.40	30.64	33.01	32.68	5.82	ppb
Selenium	78-2	32.48	32.47	32.54	32.50	0.13	ppb
Silicon	28-1	349.95	362.19	354.31	355.49	1.74	ppb
Silver	107-1	0.04	0.05	0.04	0.04	12.45	ppb
Silver	109-1	0.04	0.04	0.04	0.04	2.85	ppb
Sodium	23-2	11379.67	11501.08	11367.38	11416.05	0.65	ppb
Strontium	86-1	26.22	26.74	25.83	26.26	1.76	ppb
Strontium	88-1	25.44	26.36	25.49	25.76	2.01	ppb
Sulfur	34-1	2946.71	3951.80	3609.40	3502.64	14.59	ppb
Terbium	159-1				120		%
Terbium	159-2				118		%
Thallium	203-1	0.05	0.06	0.08	0.06	20.48	ppb
Thallium	205-1	0.06	0.08	0.07	0.07	14.56	ppb
Tin	118-1	0.30	0.25	0.26	0.27	8.76	ppb
Titanium	47-1	4.15	3.97	4.22	4.11	3.06	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.00	0.01	33.22	ppb
Vanadium	51-2	0.15	0.14	0.15	0.15	5.54	ppb
Yttrium	89-1				113		%
Yttrium	89-2				114		%
Zinc	66-2	448.76	460.61	452.08	453.82	1.35	ppb
Zirconium	90-1	0.14	0.14	0.13	0.14	5.39	ppb
Zirconium	91-1	0.13	0.15	0.14	0.14	8.53	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	64.62	62.01	64.72	63.78	2.41	ppb
Antimony	121-1	0.36	0.34	0.34	0.34	3.44	ppb
Arsenic	75-2	8.82	8.57	9.28	8.89	4.03	ppb
Barium	135-1	8.44	8.35	8.43	8.40	0.61	ppb
Barium	137-1	8.64	8.62	8.63	8.63	0.14	ppb
Beryllium	9-1	0.00	0.01	0.00	0.00	167.53	ppb
Bismuth	209-1				115		%
Bismuth	209-2				117		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	-0.03	0.13	-0.03	0.02	422.26	ppb
Cadmium	106-1	-1.37	-1.62	-0.95	-1.31		ppb
Cadmium	111-1	-0.07	-0.10	-0.05	-0.08		ppb
Calcium	43-1	1689.45	1685.11	1686.35	1686.97	0.13	ppb
Calcium	44-1	1788.91	1778.39	1807.29	1791.53	0.82	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.90	1.83	1.91	1.88	2.42	ppb
Cobalt	59-2	0.16	0.12	0.15	0.14	15.31	ppb
Copper	63-2	2.84	2.62	2.78	2.75	4.23	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				116		%
Holmium	165-2				117		%
Indium	115-1				121		%
Indium	115-2				121		%
Iron	56-2	30824.38	30340.32	31338.25	30834.31	1.62	ppb
Iron	57-2	30556.30	29811.06	30833.23	30400.20	1.74	ppb
Iron	54-2	30926.98	31132.04	31349.56	31136.19	0.68	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.48	0.50	0.49	0.49	1.88	ppb
Lead	207-1	0.47	0.48	0.50	0.48	3.11	ppb
Lead	208-1	0.47	0.49	0.48	0.48	1.58	ppb
Lithium	6-1				105		%
Magnesium	24-2	210.48	208.25	218.32	212.35	2.49	ppb
Manganese	55-2	902.35	889.43	923.35	905.04	1.89	ppb
Molybdenum	94-1	0.56	0.58	0.59	0.58	2.40	ppb
Molybdenum	95-1	0.51	0.50	0.52	0.51	1.95	ppb
Molybdenum	96-1	0.69	0.73	0.76	0.73	5.32	ppb
Molybdenum	97-1	0.58	0.57	0.54	0.56	4.32	ppb
Molybdenum	98-1	0.54	0.52	0.49	0.52	4.25	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.63	0.68	0.62	0.64	5.44	ppb
Phosphorus	31-2	2459.14	2403.65	2470.72	2444.50	1.47	ppb
Potassium	39-2	25477.14	24717.84	25925.88	25373.62	2.41	ppb
Rhodium	103-1				120		%
Rhodium	103-2				121		%
Scandium	45-1				118		%
Scandium	45-2				118		%
Selenium	82-1	19.75	21.31	21.02	20.70	4.01	ppb
Selenium	77-2	24.39	26.48	24.28	25.05	4.96	ppb
Selenium	78-2	25.34	23.33	22.45	23.70	6.24	ppb
Silicon	28-1	160.10	165.58	174.73	166.80	4.43	ppb
Silver	107-1	0.05	0.07	0.09	0.07	26.94	ppb
Silver	109-1	0.06	0.08	0.09	0.08	18.79	ppb
Sodium	23-2	2109.82	2119.58	2125.94	2118.45	0.38	ppb
Strontium	86-1	2.42	2.53	2.47	2.47	2.17	ppb
Strontium	88-1	2.41	2.52	2.45	2.46	2.36	ppb
Sulfur	34-1	-1679.83	-1652.73	-1573.91	-1635.49		ppb
Terbium	159-1				119		%
Terbium	159-2				119		%
Thallium	203-1	0.02	0.02	0.03	0.03	29.42	ppb
Thallium	205-1	0.03	0.05	0.04	0.04	20.93	ppb
Tin	118-1	0.44	0.44	0.45	0.44	0.95	ppb
Titanium	47-1	0.85	0.86	0.72	0.81	9.39	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.07	0.07	0.08	0.07	4.89	ppb
Yttrium	89-1				113		%
Yttrium	89-2				115		%
Zinc	66-2	30.26	27.84	31.05	29.71	5.63	ppb
Zirconium	90-1	0.06	0.07	0.07	0.07	7.73	ppb
Zirconium	91-1	0.07	0.08	0.09	0.08	17.19	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	3.50	2.56	2.37	2.81	21.49	ppb
Antimony	121-1	0.06	0.05	0.05	0.05	7.55	ppb
Arsenic	75-2	1.87	1.73	1.66	1.75	6.09	ppb
Barium	135-1	182.39	185.51	187.23	185.04	1.33	ppb
Barium	137-1	184.21	187.99	191.21	187.80	1.87	ppb
Beryllium	9-1	0.01	0.02	0.02	0.01	46.86	ppb
Bismuth	209-1				120		%
Bismuth	209-2				120		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.16	0.02	-0.01	0.06	159.83	ppb
Cadmium	106-1	-0.92	-0.02	-1.03	-0.66		ppb
Cadmium	111-1	-0.06	0.00	-0.06	-0.04		ppb
Calcium	43-1	62217.63	61721.65	64290.35	62743.21	2.17	ppb
Calcium	44-1	61903.48	62537.82	63150.11	62530.47	1.00	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.13	-0.10	-0.16	-0.13		ppb
Cobalt	59-2	1.75	1.75	1.68	1.73	2.30	ppb
Copper	63-2	0.05	0.05	0.04	0.05	9.33	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				123		%
Holmium	165-2				120		%
Indium	115-1				128		%
Indium	115-2				126		%
Iron	56-2	21868.63	21936.54	21482.62	21762.60	1.13	ppb
Iron	57-2	22118.14	22021.35	21683.73	21941.07	1.04	ppb
Iron	54-2	21999.06	22124.84	22001.45	22041.78	0.33	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	14.38	13.99	14.06	14.14	1.49	ppb
Lead	207-1	13.27	12.78	12.87	12.98	2.03	ppb
Lead	208-1	13.69	13.30	13.33	13.44	1.59	ppb
Lithium	6-1				110		%
Magnesium	24-2	21262.05	21503.90	21141.69	21302.55	0.87	ppb
Manganese	55-2	1183.67	1205.52	1208.38	1199.19	1.13	ppb
Molybdenum	94-1	0.53	0.53	0.52	0.53	1.11	ppb
Molybdenum	95-1	0.60	0.66	0.62	0.63	4.40	ppb
Molybdenum	96-1	0.72	0.70	0.73	0.72	1.97	ppb
Molybdenum	97-1	0.57	0.60	0.57	0.58	2.93	ppb
Molybdenum	98-1	0.60	0.61	0.59	0.60	1.97	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.98	0.93	0.93	0.95	3.39	ppb
Phosphorus	31-2	142.34	135.32	112.78	130.14	11.87	ppb
Potassium	39-2	9808.47	9902.90	9607.72	9773.03	1.54	ppb
Rhodium	103-1				128		%
Rhodium	103-2				123		%
Scandium	45-1				128		%
Scandium	45-2				123		%
Selenium	82-1	9.65	9.25	7.58	8.82	12.46	ppb
Selenium	77-2	9.56	7.74	8.08	8.46	11.47	ppb
Selenium	78-2	6.38	7.54	8.60	7.51	14.84	ppb
Silicon	28-1	3900.75	3904.34	3877.61	3894.23	0.37	ppb
Silver	107-1	0.00	0.00	0.00	0.00	224.92	ppb
Silver	109-1	0.00	0.00	0.00	0.00	47.87	ppb
Sodium	23-2	11609.31	11666.76	11507.54	11594.54	0.70	ppb
Strontium	86-1	719.30	732.45	718.44	723.40	1.09	ppb
Strontium	88-1	701.67	707.40	700.26	703.11	0.54	ppb
Sulfur	34-1	55970.86	56846.11	56755.28	56524.08	0.85	ppb
Terbium	159-1				125		%
Terbium	159-2				124		%
Thallium	203-1	-0.04	-0.04	-0.04	-0.04		ppb
Thallium	205-1	-0.02	-0.03	-0.03	-0.03		ppb
Tin	118-1	1.68	1.78	1.74	1.73	2.84	ppb
Titanium	47-1	0.47	0.46	0.46	0.47	1.47	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.01	0.02	7.06	ppb
Vanadium	51-2	0.08	0.08	0.07	0.08	6.83	ppb
Yttrium	89-1				122		%
Yttrium	89-2				121		%
Zinc	66-2	-1.32	-1.29	-1.09	-1.23		ppb
Zirconium	90-1	0.01	0.01	0.01	0.01	28.12	ppb
Zirconium	91-1	0.00	0.01	0.01	0.01	86.33	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4.50	4.43	4.71	4.55	3.29	ppb
Antimony	121-1	0.03	0.03	0.02	0.03	15.79	ppb
Arsenic	75-2	0.93	1.03	1.06	1.01	7.05	ppb
Barium	135-1	161.04	161.90	158.47	160.47	1.11	ppb
Barium	137-1	163.74	163.75	161.80	163.10	0.69	ppb
Beryllium	9-1	0.00	0.02	0.00	0.01	178.70	ppb
Bismuth	209-1				114		%
Bismuth	209-2				116		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.09	-0.03	-0.01	0.02	368.82	ppb
Cadmium	106-1	-0.18	-1.37	-1.10	-0.88		ppb
Cadmium	111-1	-0.01	-0.08	-0.08	-0.06		ppb
Calcium	43-1	61713.44	61569.70	63194.29	62159.14	1.45	ppb
Calcium	44-1	60324.54	60357.74	60985.56	60555.94	0.62	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.22	-0.24	-0.18	-0.21		ppb
Cobalt	59-2	1.60	1.65	1.65	1.63	1.63	ppb
Copper	63-2	-0.03	-0.01	-0.03	-0.02		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				120		%
Holmium	165-2				118		%
Indium	115-1				124		%
Indium	115-2				123		%
Iron	56-2	21293.36	21491.99	21334.83	21373.39	0.49	ppb
Iron	57-2	21310.99	21496.69	21037.14	21281.61	1.09	ppb
Iron	54-2	21269.27	20860.44	21115.67	21081.80	0.98	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.01	0.00	0.01	0.00		ppb
Lead	207-1	0.00	0.01	0.00	0.00	116.36	ppb
Lead	208-1	-0.01	-0.01	0.00	-0.01		ppb
Lithium	6-1				105		%
Magnesium	24-2	20813.31	20761.38	20917.99	20830.89	0.38	ppb
Manganese	55-2	1192.12	1190.40	1172.39	1184.97	0.92	ppb
Molybdenum	94-1	0.58	0.59	0.61	0.59	2.26	ppb
Molybdenum	95-1	0.67	0.70	0.72	0.70	4.14	ppb
Molybdenum	96-1	0.79	0.79	0.81	0.79	1.24	ppb
Molybdenum	97-1	0.69	0.68	0.80	0.72	8.75	ppb
Molybdenum	98-1	0.71	0.70	0.70	0.70	0.87	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.68	0.58	0.61	0.62	8.09	ppb
Phosphorus	31-2	149.66	152.76	126.58	143.00	10.00	ppb
Potassium	39-2	9678.77	9742.63	9701.93	9707.78	0.33	ppb
Rhodium	103-1				123		%
Rhodium	103-2				120		%
Scandium	45-1				123		%
Scandium	45-2				120		%
Selenium	82-1	6.87	6.93	6.34	6.71	4.80	ppb
Selenium	77-2	6.74	5.61	6.06	6.14	9.28	ppb
Selenium	78-2	7.07	7.14	5.38	6.53	15.30	ppb
Silicon	28-1	3726.25	3824.77	3983.08	3844.70	3.37	ppb
Silver	107-1	0.00	0.00	0.00	0.00		ppb
Silver	109-1	0.00	0.00	0.00	0.00	431.99	ppb
Sodium	23-2	11117.97	11045.74	11335.36	11166.36	1.35	ppb
Strontium	86-1	693.22	701.72	697.85	697.60	0.61	ppb
Strontium	88-1	685.35	709.92	689.28	694.85	1.90	ppb
Sulfur	34-1	55751.21	57811.22	58424.15	57328.86	2.44	ppb
Terbium	159-1				122		%
Terbium	159-2				119		%
Thallium	203-1	-0.05	-0.05	-0.05	-0.05		ppb
Thallium	205-1	-0.04	-0.04	-0.04	-0.04		ppb
Tin	118-1	0.04	0.05	0.02	0.04	36.29	ppb
Titanium	47-1	0.42	0.33	0.51	0.42	21.47	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	14.56	ppb
Vanadium	51-2	0.07	0.08	0.08	0.08	5.10	ppb
Yttrium	89-1				118		%
Yttrium	89-2				116		%
Zinc	66-2	-1.89	-1.89	-2.03	-1.94		ppb
Zirconium	90-1	0.00	0.01	0.00	0.00	93.85	ppb
Zirconium	91-1	0.01	0.01	-0.01	0.00	573.67	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47405.10	47493.08	47987.88	47628.69	0.66	ppb
Antimony	121-1	486.42	497.34	499.27	494.34	1.40	ppb
Arsenic	75-2	475.14	472.43	476.96	474.84	0.48	ppb
Barium	135-1	2409.93	2520.75	2453.73	2461.47	2.27	ppb
Barium	137-1	2471.60	2475.34	2458.81	2468.58	0.35	ppb
Beryllium	9-1	497.00	513.68	504.13	504.93	1.66	ppb
Bismuth	209-1				105		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	473.34	491.64	480.05	481.68	1.92	ppb
Cadmium	106-1	480.09	490.74	492.33	487.72	1.37	ppb
Cadmium	111-1	476.44	486.92	479.74	481.03	1.11	ppb
Calcium	43-1	245093.49	254137.41	253052.19	250761.03	1.97	ppb
Calcium	44-1	249336.80	256360.24	250244.84	251980.63	1.52	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	483.79	495.51	470.91	483.41	2.55	ppb
Cobalt	59-2	486.54	478.84	481.01	482.13	0.82	ppb
Copper	63-2	4743.98	4686.44	4627.99	4686.14	1.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				112		%
Holmium	165-2				113		%
Indium	115-1				113		%
Indium	115-2				114		%
Iron	56-2	127448.09	126515.68	125516.17	126493.31	0.76	ppb
Iron	57-2	124520.64	127101.98	125636.29	125752.97	1.03	ppb
Iron	54-2	129399.63	123859.18	127237.58	126832.13	2.20	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2426.86	2452.09	2435.59	2438.18	0.53	ppb
Lead	207-1	2429.42	2434.14	2419.38	2427.64	0.31	ppb
Lead	208-1	2424.10	2461.46	2406.27	2430.61	1.16	ppb
Lithium	6-1				99		%
Magnesium	24-2	244350.51	240831.09	235584.16	240255.25	1.84	ppb
Manganese	55-2	4828.45	4856.72	4851.16	4845.44	0.31	ppb
Molybdenum	94-1	4941.06	4859.87	4879.30	4893.41	0.87	ppb
Molybdenum	95-1	4852.16	4813.89	4735.87	4800.64	1.23	ppb
Molybdenum	96-1	4934.85	4946.45	4780.57	4887.29	1.89	ppb
Molybdenum	97-1	4940.45	4887.16	4803.66	4877.09	1.41	ppb
Molybdenum	98-1	4973.61	4892.24	4881.18	4915.68	1.03	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	471.85	470.15	468.99	470.33	0.31	ppb
Phosphorus	31-2	9737.51	9652.58	9632.40	9674.16	0.58	ppb
Potassium	39-2	123730.36	124464.85	124194.95	124130.05	0.30	ppb
Rhodium	103-1				111		%
Rhodium	103-2				110		%
Scandium	45-1				115		%
Scandium	45-2				115		%
Selenium	82-1	482.01	476.64	472.11	476.92	1.04	ppb
Selenium	77-2	463.32	490.83	484.90	479.68	3.02	ppb
Selenium	78-2	466.42	478.16	475.02	473.20	1.28	ppb
Silicon	28-1	499.56	527.44	524.60	517.20	2.97	ppb
Silver	107-1	483.70	503.61	491.09	492.80	2.04	ppb
Silver	109-1	484.65	493.38	487.71	488.58	0.91	ppb
Sodium	23-2	244824.54	239807.39	246882.57	243838.17	1.49	ppb
Strontium	86-1	487.14	480.48	476.69	481.44	1.10	ppb
Strontium	88-1	491.19	492.44	486.37	490.00	0.65	ppb
Sulfur	34-1	7719.29	7916.77	7704.75	7780.27	1.52	ppb
Terbium	159-1				115		%
Terbium	159-2				115		%
Thallium	203-1	489.18	488.28	510.25	495.90	2.51	ppb
Thallium	205-1	500.48	487.45	501.36	496.43	1.57	ppb
Tin	118-1	492.83	499.37	501.99	498.06	0.95	ppb
Titanium	47-1	4744.84	4875.04	4815.14	4811.67	1.35	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	469.06	483.55	465.46	472.69	2.03	ppb
Vanadium	51-2	475.30	491.88	472.25	479.81	2.20	ppb
Yttrium	89-1				110		%
Yttrium	89-2				111		%
Zinc	66-2	4676.56	4731.30	4736.94	4714.94	0.71	ppb
Zirconium	90-1	486.85	484.59	477.17	482.87	1.05	ppb
Zirconium	91-1	485.48	476.94	475.84	479.42	1.10	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-2.48	-1.59	-3.06	-2.38		ppb
Antimony	121-1	0.14	0.12	0.12	0.13	7.53	ppb
Arsenic	75-2	0.25	0.18	0.17	0.20	22.23	ppb
Barium	135-1	-0.02	-0.02	-0.03	-0.02		ppb
Barium	137-1	-0.01	0.01	0.01	0.00	430.37	ppb
Beryllium	9-1	0.00	-0.01	0.00	0.00		ppb
Bismuth	209-1				105		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.06	0.12	0.09	0.09	34.48	ppb
Cadmium	106-1	-2.78	-3.67	-3.64	-3.36		ppb
Cadmium	111-1	-0.19	-0.24	-0.25	-0.23		ppb
Calcium	43-1	-12.81	-12.29	-11.59	-12.23		ppb
Calcium	44-1	6.47	3.48	7.80	5.92	37.39	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.06	0.06	0.05	0.06	14.84	ppb
Cobalt	59-2	-0.02	-0.02	-0.02	-0.02		ppb
Copper	63-2	0.20	0.21	0.23	0.21	5.11	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				108		%
Holmium	165-2				104		%
Indium	115-1				110		%
Indium	115-2				108		%
Iron	56-2	-0.86	-0.74	-0.72	-0.78		ppb
Iron	57-2	-2.68	-0.64	-0.82	-1.38		ppb
Iron	54-2	-1.03	0.36	-0.79	-0.49		ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.01	0.00	0.00	0.01	49.66	ppb
Lead	207-1	-0.03	-0.02	-0.02	-0.02		ppb
Lead	208-1	-0.01	-0.01	-0.01	-0.01		ppb
Lithium	6-1				89		%
Magnesium	24-2	-6.38	-5.87	-6.06	-6.10		ppb
Manganese	55-2	-0.01	-0.01	-0.07	-0.03		ppb
Molybdenum	94-1	-0.01	0.03	0.04	0.02	123.49	ppb
Molybdenum	95-1	0.01	0.02	0.03	0.02	61.41	ppb
Molybdenum	96-1	0.03	0.03	0.05	0.04	24.63	ppb
Molybdenum	97-1	0.03	0.03	0.06	0.04	37.00	ppb
Molybdenum	98-1	0.02	0.01	0.03	0.02	43.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.02	0.04	0.00	0.02	98.23	ppb
Phosphorus	31-2	-56.44	-58.56	-59.33	-58.11		ppb
Potassium	39-2	13.98	19.40	17.45	16.94	16.19	ppb
Rhodium	103-1				108		%
Rhodium	103-2				108		%
Scandium	45-1				101		%
Scandium	45-2				103		%
Selenium	82-1	1.27	2.27	3.03	2.19	40.37	ppb
Selenium	77-2	0.88	2.50	2.39	1.92	47.21	ppb
Selenium	78-2	2.76	2.73	2.79	2.76	1.19	ppb
Silicon	28-1	0.09	0.43	0.77	0.43	79.13	ppb
Silver	107-1	0.05	0.06	0.07	0.06	17.18	ppb
Silver	109-1	0.07	0.07	0.08	0.07	13.07	ppb
Sodium	23-2	30.48	29.78	30.06	30.11	1.18	ppb
Strontium	86-1	-0.02	0.04	0.02	0.01	260.99	ppb
Strontium	88-1	0.00	0.00	0.00	0.00	98.33	ppb
Sulfur	34-1	-1246.16	-1150.39	-1037.27	-1144.60		ppb
Terbium	159-1				110		%
Terbium	159-2				107		%
Thallium	203-1	-0.06	-0.06	-0.06	-0.06		ppb
Thallium	205-1	-0.05	-0.05	-0.06	-0.05		ppb
Tin	118-1	0.08	0.20	0.14	0.14	44.43	ppb
Titanium	47-1	0.02	-0.02	0.00	0.00	574.31	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.00	0.01	0.00	0.01	84.75	ppb
Yttrium	89-1				101		%
Yttrium	89-2				101		%
Zinc	66-2	-0.42	-0.31	-0.21	-0.32		ppb
Zirconium	90-1	-0.01	0.01	0.00	0.00		ppb
Zirconium	91-1	-0.01	-0.01	0.00	-0.01		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.15	3.71	7.16	4.01	75.20	ppb
Antimony	121-1	0.08	0.08	0.08	0.08	1.97	ppb
Arsenic	75-2	0.52	0.36	0.47	0.45	17.70	ppb
Barium	135-1	183.37	175.98	176.52	178.62	2.31	ppb
Barium	137-1	185.99	180.81	179.51	182.10	1.88	ppb
Beryllium	9-1	0.00	0.00	0.01	0.00	359.18	ppb
Bismuth	209-1				119		%
Bismuth	209-2				118		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.09	-0.01	0.04	0.04	120.43	ppb
Cadmium	106-1	-3.77	-2.71	-1.89	-2.79		ppb
Cadmium	111-1	-0.26	-0.18	-0.13	-0.19		ppb
Calcium	43-1	69223.98	67148.46	67896.37	68089.60	1.54	ppb
Calcium	44-1	67733.89	65122.24	65899.80	66251.98	2.02	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.39	0.46	0.43	0.43	8.44	ppb
Cobalt	59-2	1.85	1.87	1.83	1.85	1.02	ppb
Copper	63-2	0.13	0.15	0.17	0.15	14.80	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				120		%
Holmium	165-2				120		%
Indium	115-1				125		%
Indium	115-2				122		%
Iron	56-2	22865.82	22947.42	22662.71	22825.32	0.64	ppb
Iron	57-2	23041.45	22937.53	23252.41	23077.13	0.70	ppb
Iron	54-2	23045.69	23031.92	22681.66	22919.76	0.90	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.02	0.00	-0.01	-0.01		ppb
Lead	207-1	-0.03	-0.01	-0.03	-0.02		ppb
Lead	208-1	-0.02	-0.02	-0.02	-0.02		ppb
Lithium	6-1				101		%
Magnesium	24-2	22226.59	22101.72	22345.21	22224.51	0.55	ppb
Manganese	55-2	1245.37	1257.30	1252.11	1251.59	0.48	ppb
Molybdenum	94-1	1.30	1.24	1.33	1.29	3.38	ppb
Molybdenum	95-1	1.55	1.48	1.50	1.51	2.32	ppb
Molybdenum	96-1	1.71	1.53	1.61	1.62	5.48	ppb
Molybdenum	97-1	1.56	1.47	1.51	1.51	3.10	ppb
Molybdenum	98-1	1.58	1.51	1.54	1.55	2.34	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.60	0.68	0.63	0.64	6.28	ppb
Phosphorus	31-2	144.73	151.91	133.66	143.43	6.41	ppb
Potassium	39-2	10081.42	9971.93	9802.37	9951.91	1.41	ppb
Rhodium	103-1				120		%
Rhodium	103-2				123		%
Scandium	45-1				116		%
Scandium	45-2				119		%
Selenium	82-1	1.85	1.61	1.50	1.65	10.80	ppb
Selenium	77-2	0.95	1.95	2.90	1.94	50.35	ppb
Selenium	78-2	2.30	1.53	1.60	1.81	23.46	ppb
Silicon	28-1	4217.47	4018.66	4022.99	4086.37	2.78	ppb
Silver	107-1	0.02	0.02	0.01	0.01	50.54	ppb
Silver	109-1	0.03	0.02	0.03	0.02	23.23	ppb
Sodium	23-2	11868.12	11911.31	12216.66	11998.70	1.58	ppb
Strontium	86-1	789.08	752.54	746.07	762.56	3.04	ppb
Strontium	88-1	777.08	755.89	754.28	762.42	1.67	ppb
Sulfur	34-1	65288.34	61823.95	63584.61	63565.63	2.73	ppb
Terbium	159-1				124		%
Terbium	159-2				121		%
Thallium	203-1	-0.07	-0.07	-0.08	-0.07		ppb
Thallium	205-1	-0.06	-0.06	-0.06	-0.06		ppb
Tin	118-1	0.06	0.03	0.05	0.05	41.00	ppb
Titanium	47-1	0.50	0.43	0.55	0.50	12.22	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	5.85	ppb
Vanadium	51-2	0.10	0.07	0.09	0.09	13.53	ppb
Yttrium	89-1				114		%
Yttrium	89-2				116		%
Zinc	66-2	-2.00	-1.47	-1.70	-1.73		ppb
Zirconium	90-1	0.02	0.01	0.02	0.02	30.88	ppb
Zirconium	91-1	0.01	0.02	0.02	0.02	16.99	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	6.66	6.95	6.62	6.74	2.63	ppb
Antimony	121-1	0.09	0.10	0.10	0.10	6.53	ppb
Arsenic	75-2	0.39	0.42	0.50	0.44	12.19	ppb
Barium	135-1	168.52	170.15	167.24	168.64	0.87	ppb
Barium	137-1	169.53	171.26	169.24	170.01	0.64	ppb
Beryllium	9-1	0.00	0.01	0.01	0.01	84.98	ppb
Bismuth	209-1				116		%
Bismuth	209-2				116		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	-0.03	-0.03	-0.01		ppb
Cadmium	106-1	-0.29	-2.06	-1.73	-1.36		ppb
Cadmium	111-1	-0.01	-0.14	-0.10	-0.08		ppb
Calcium	43-1	62641.17	63514.99	64200.75	63452.30	1.23	ppb
Calcium	44-1	61860.76	61357.94	61402.81	61540.50	0.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.20	-0.18	-0.12	-0.17		ppb
Cobalt	59-2	1.67	1.71	1.68	1.69	1.02	ppb
Copper	63-2	0.16	0.16	0.11	0.14	19.67	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				120		%
Holmium	165-2				120		%
Indium	115-1				125		%
Indium	115-2				120		%
Iron	56-2	21640.16	21642.32	21653.17	21645.22	0.03	ppb
Iron	57-2	21582.24	22201.95	22213.59	21999.26	1.64	ppb
Iron	54-2	21298.13	21506.61	21779.53	21528.09	1.12	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.01	0.03	0.00	0.01	101.11	ppb
Lead	207-1	0.00	0.02	0.03	0.01	99.92	ppb
Lead	208-1	0.02	0.02	0.01	0.02	19.56	ppb
Lithium	6-1				101		%
Magnesium	24-2	20718.50	21305.17	21100.35	21041.34	1.42	ppb
Manganese	55-2	1176.46	1193.35	1198.43	1189.41	0.97	ppb
Molybdenum	94-1	1.56	1.48	1.48	1.50	3.07	ppb
Molybdenum	95-1	1.79	1.83	1.83	1.82	1.16	ppb
Molybdenum	96-1	1.88	1.88	1.86	1.88	0.56	ppb
Molybdenum	97-1	1.83	1.77	1.80	1.80	1.48	ppb
Molybdenum	98-1	1.77	1.86	1.72	1.78	4.13	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.66	0.73	0.62	0.67	8.18	ppb
Phosphorus	31-2	159.63	163.81	141.10	154.85	7.81	ppb
Potassium	39-2	9818.87	9883.63	10155.99	9952.83	1.80	ppb
Rhodium	103-1				123		%
Rhodium	103-2				121		%
Scandium	45-1				118		%
Scandium	45-2				116		%
Selenium	82-1	1.03	1.78	1.33	1.38	27.57	ppb
Selenium	77-2	1.96	2.71	1.40	2.03	32.48	ppb
Selenium	78-2	3.47	3.73	0.29	2.50	76.65	ppb
Silicon	28-1	3786.18	3853.57	3979.33	3873.03	2.53	ppb
Silver	107-1	0.01	0.01	0.02	0.01	24.17	ppb
Silver	109-1	0.02	0.02	0.01	0.02	28.98	ppb
Sodium	23-2	12300.39	12072.41	12446.23	12273.01	1.54	ppb
Strontium	86-1	715.92	697.13	704.61	705.89	1.34	ppb
Strontium	88-1	697.16	706.79	707.97	703.97	0.84	ppb
Sulfur	34-1	56208.13	59165.07	59488.61	58287.27	3.10	ppb
Terbium	159-1				123		%
Terbium	159-2				119		%
Thallium	203-1	-0.07	-0.07	-0.07	-0.07		ppb
Thallium	205-1	-0.06	-0.06	-0.06	-0.06		ppb
Tin	118-1	0.78	0.83	0.78	0.80	4.17	ppb
Titanium	47-1	0.50	0.51	0.54	0.52	4.23	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.03	0.03	0.03	0.03	2.13	ppb
Vanadium	51-2	0.09	0.09	0.08	0.09	11.16	ppb
Yttrium	89-1				116		%
Yttrium	89-2				114		%
Zinc	66-2	-0.56	-0.45	-0.68	-0.56		ppb
Zirconium	90-1	0.01	0.01	0.01	0.01	33.13	ppb
Zirconium	91-1	0.00	0.02	0.00	0.01	179.81	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.34	1.28	0.78	1.13	27.30	ppb
Antimony	121-1	0.04	0.03	0.04	0.03	11.35	ppb
Arsenic	75-2	0.01	0.06	0.04	0.03	73.57	ppb
Barium	135-1	-0.04	-0.01	-0.04	-0.03		ppb
Barium	137-1	-0.01	0.00	-0.01	0.00		ppb
Beryllium	9-1	0.01	0.00	0.01	0.01	115.93	ppb
Bismuth	209-1				120		%
Bismuth	209-2				119		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	-0.03	0.02	0.01	409.35	ppb
Cadmium	106-1	-2.39	-2.68	-1.21	-2.09		ppb
Cadmium	111-1	-0.16	-0.18	-0.08	-0.14		ppb
Calcium	43-1	-8.42	-10.02	-12.42	-10.29		ppb
Calcium	44-1	0.98	0.78	4.13	1.96	95.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.22	-0.26	-0.24	-0.24		ppb
Cobalt	59-2	-0.02	-0.02	-0.02	-0.02		ppb
Copper	63-2	0.15	0.15	0.10	0.13	19.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				120		%
Holmium	165-2				118		%
Indium	115-1				127		%
Indium	115-2				122		%
Iron	56-2	28.33	27.12	27.15	27.53	2.50	ppb
Iron	57-2	23.18	25.36	27.87	25.47	9.21	ppb
Iron	54-2	27.92	26.63	24.66	26.40	6.22	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.00	0.00	0.00	0.00	8939.92	ppb
Lead	207-1	0.00	0.00	0.01	0.00	299.52	ppb
Lead	208-1	0.01	0.00	0.01	0.01	49.60	ppb
Lithium	6-1				102		%
Magnesium	24-2	-5.60	-5.65	-5.45	-5.56		ppb
Manganese	55-2	16.45	15.61	15.80	15.95	2.75	ppb
Molybdenum	94-1	0.92	1.03	0.99	0.98	5.71	ppb
Molybdenum	95-1	1.24	1.21	1.17	1.21	2.91	ppb
Molybdenum	96-1	1.22	1.31	1.25	1.26	3.58	ppb
Molybdenum	97-1	1.29	1.23	1.26	1.26	2.44	ppb
Molybdenum	98-1	1.22	1.22	1.25	1.23	1.64	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.04	0.00	-0.10	-0.05		ppb
Phosphorus	31-2	-43.32	-53.78	-39.98	-45.69		ppb
Potassium	39-2	190.19	189.58	184.04	187.93	1.80	ppb
Rhodium	103-1				125		%
Rhodium	103-2				121		%
Scandium	45-1				119		%
Scandium	45-2				118		%
Selenium	82-1	1.28	0.90	0.91	1.03	21.11	ppb
Selenium	77-2	0.80	2.58	0.58	1.32	82.87	ppb
Selenium	78-2	-0.17	2.12	0.65	0.86	134.55	ppb
Silicon	28-1	-5.17	-4.54	-3.96	-4.56		ppb
Silver	107-1	0.02	0.03	0.04	0.03	30.12	ppb
Silver	109-1	0.03	0.03	0.04	0.03	30.65	ppb
Sodium	23-2	38.35	37.73	35.62	37.23	3.84	ppb
Strontium	86-1	-0.03	0.00	0.02	0.00		ppb
Strontium	88-1	0.04	0.04	0.04	0.04	5.35	ppb
Sulfur	34-1	-2056.79	-1924.28	-1771.25	-1917.44		ppb
Terbium	159-1				123		%
Terbium	159-2				120		%
Thallium	203-1	-0.07	-0.07	-0.08	-0.07		ppb
Thallium	205-1	-0.07	-0.07	-0.07	-0.07		ppb
Tin	118-1	0.66	0.68	0.81	0.72	11.19	ppb
Titanium	47-1	-0.03	0.04	-0.04	-0.01		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	154.04	ppb
Vanadium	51-2	0.01	0.00	0.00	0.00	61.98	ppb
Yttrium	89-1				116		%
Yttrium	89-2				113		%
Zinc	66-2	-0.60	-0.61	-0.51	-0.58		ppb
Zirconium	90-1	0.00	0.01	0.01	0.00	82.29	ppb
Zirconium	91-1	0.00	-0.01	0.00	0.00		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2.02	1.17	0.79	1.32	47.40	ppb
Antimony	121-1	0.03	0.03	0.02	0.03	17.72	ppb
Arsenic	75-2	0.02	0.03	0.02	0.02	30.86	ppb
Barium	135-1	-0.01	-0.02	-0.05	-0.03		ppb
Barium	137-1	-0.01	0.00	-0.01	-0.01		ppb
Beryllium	9-1	0.03	-0.01	0.01	0.01	149.08	ppb
Bismuth	209-1				116		%
Bismuth	209-2				116		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.18	0.07	0.02	0.09	86.11	ppb
Cadmium	106-1	0.28	-1.69	-0.75	-0.72		ppb
Cadmium	111-1	0.02	-0.12	-0.05	-0.05		ppb
Calcium	43-1	-12.99	-15.05	-8.83	-12.29		ppb
Calcium	44-1	0.60	1.77	-0.42	0.65	167.88	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.15	-0.14	-0.22	-0.17		ppb
Cobalt	59-2	-0.02	-0.01	-0.02	-0.02		ppb
Copper	63-2	0.00	0.00	0.04	0.01	165.80	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				117		%
Holmium	165-2				114		%
Indium	115-1				122		%
Indium	115-2				118		%
Iron	56-2	3.78	3.40	3.35	3.51	6.68	ppb
Iron	57-2	1.42	-0.40	1.39	0.80	129.47	ppb
Iron	54-2	2.87	1.32	2.10	2.10	36.97	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.02	0.01	-0.03	-0.01		ppb
Lead	207-1	-0.04	-0.02	-0.03	-0.03		ppb
Lead	208-1	-0.02	-0.01	-0.02	-0.02		ppb
Lithium	6-1				100		%
Magnesium	24-2	-3.51	-2.80	-3.19	-3.17		ppb
Manganese	55-2	7.23	7.09	7.30	7.21	1.51	ppb
Molybdenum	94-1	0.28	0.28	0.29	0.28	3.49	ppb
Molybdenum	95-1	0.31	0.31	0.31	0.31	1.23	ppb
Molybdenum	96-1	0.32	0.35	0.33	0.33	3.76	ppb
Molybdenum	97-1	0.32	0.33	0.30	0.32	4.25	ppb
Molybdenum	98-1	0.33	0.33	0.34	0.33	2.04	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.06	-0.10	-0.14	-0.10		ppb
Phosphorus	31-2	-53.52	-64.71	-57.07	-58.43		ppb
Potassium	39-2	26.47	24.99	27.88	26.44	5.47	ppb
Rhodium	103-1				121		%
Rhodium	103-2				118		%
Scandium	45-1				117		%
Scandium	45-2				115		%
Selenium	82-1	1.30	0.41	0.54	0.75	64.15	ppb
Selenium	77-2	0.58	1.00	1.22	0.94	34.89	ppb
Selenium	78-2	0.57	1.81	0.96	1.11	57.04	ppb
Silicon	28-1	-3.75	-3.64	-2.45	-3.28		ppb
Silver	107-1	0.02	0.01	0.01	0.02	13.20	ppb
Silver	109-1	0.02	0.02	0.02	0.02	17.18	ppb
Sodium	23-2	15.18	14.26	14.17	14.54	3.84	ppb
Strontium	86-1	-0.03	-0.01	-0.06	-0.03		ppb
Strontium	88-1	0.02	0.01	0.01	0.01	23.58	ppb
Sulfur	34-1	-1774.55	-1741.71	-1654.86	-1723.71		ppb
Terbium	159-1				119		%
Terbium	159-2				116		%
Thallium	203-1	-0.08	-0.08	-0.08	-0.08		ppb
Thallium	205-1	-0.07	-0.07	-0.07	-0.07		ppb
Tin	118-1	0.37	0.39	0.35	0.37	5.55	ppb
Titanium	47-1	-0.05	-0.01	-0.01	-0.02		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	5042.85	ppb
Vanadium	51-2	0.01	0.00	0.01	0.01	70.52	ppb
Yttrium	89-1				113		%
Yttrium	89-2				112		%
Zinc	66-2	-0.46	-0.74	-0.87	-0.69		ppb
Zirconium	90-1	0.00	0.00	0.00	0.00		ppb
Zirconium	91-1	0.00	-0.01	-0.01	-0.01		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	3705.48	3776.67	3701.25	3727.80	1.14	ppb
Antimony	121-1	0.18	0.18	0.18	0.18	2.19	ppb
Arsenic	75-2	1.61	1.75	1.67	1.68	4.24	ppb
Barium	135-1	93.92	94.61	95.37	94.63	0.77	ppb
Barium	137-1	98.12	95.59	100.56	98.09	2.54	ppb
Beryllium	9-1	1.49	1.61	1.75	1.61	7.91	ppb
Bismuth	209-1				107		%
Bismuth	209-2				111		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.79	2.69	2.59	2.69	3.64	ppb
Cadmium	106-1	1.02	2.29	2.19	1.83	38.27	ppb
Cadmium	111-1	2.80	2.93	2.83	2.85	2.39	ppb
Calcium	43-1	20979.92	20305.97	20717.26	20667.71	1.64	ppb
Calcium	44-1	20095.34	20105.95	19959.70	20053.67	0.41	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	13.39	13.50	13.19	13.36	1.17	ppb
Cobalt	59-2	13.36	13.40	13.14	13.30	1.05	ppb
Copper	63-2	15.89	16.03	16.22	16.05	1.04	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				112		%
Holmium	165-2				112		%
Indium	115-1				113		%
Indium	115-2				115		%
Iron	56-2	8232.13	8272.12	8347.42	8283.89	0.71	ppb
Iron	57-2	8438.38	8572.30	8566.63	8525.77	0.89	ppb
Iron	54-2	8436.08	8518.46	8203.36	8385.97	1.95	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	27.17	28.99	27.66	27.94	3.38	ppb
Lead	207-1	26.63	28.30	27.91	27.61	3.15	ppb
Lead	208-1	27.02	27.97	27.49	27.49	1.74	ppb
Lithium	6-1				92		%
Magnesium	24-2	9788.88	9684.83	9502.96	9658.89	1.50	ppb
Manganese	55-2	241.66	247.21	245.94	244.94	1.19	ppb
Molybdenum	94-1	6.99	6.63	7.11	6.91	3.59	ppb
Molybdenum	95-1	3.52	3.57	3.72	3.60	2.90	ppb
Molybdenum	96-1	3.86	4.10	4.27	4.08	5.07	ppb
Molybdenum	97-1	3.60	3.66	3.65	3.64	0.90	ppb
Molybdenum	98-1	3.60	3.69	3.60	3.63	1.48	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	11.69	12.22	12.02	11.98	2.26	ppb
Phosphorus	31-2	3398.55	3424.15	3463.65	3428.78	0.96	ppb
Potassium	39-2	5200.56	5349.46	5271.68	5273.90	1.41	ppb
Rhodium	103-1				115		%
Rhodium	103-2				113		%
Scandium	45-1				109		%
Scandium	45-2				111		%
Selenium	82-1	1.24	2.16	1.71	1.70	27.16	ppb
Selenium	77-2	4.96	7.10	6.82	6.29	18.52	ppb
Selenium	78-2	3.30	4.51	3.15	3.65	20.32	ppb
Silicon	28-1	3153.75	3243.59	3285.56	3227.63	2.09	ppb
Silver	107-1	0.03	0.02	0.03	0.03	17.41	ppb
Silver	109-1	0.02	0.03	0.03	0.03	13.66	ppb
Sodium	23-2	25746.21	25752.10	25394.59	25630.96	0.80	ppb
Strontium	86-1	70.49	72.33	74.00	72.27	2.43	ppb
Strontium	88-1	70.39	69.78	71.97	70.71	1.60	ppb
Sulfur	34-1	5706.91	5387.53	5445.07	5513.17	3.09	ppb
Terbium	159-1				113		%
Terbium	159-2				114		%
Thallium	203-1	0.03	0.06	0.04	0.04	25.50	ppb
Thallium	205-1	0.06	0.07	0.07	0.06	4.17	ppb
Tin	118-1	0.64	0.66	0.66	0.65	1.61	ppb
Titanium	47-1	35.09	36.19	36.21	35.83	1.78	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.30	1.36	1.36	1.34	2.43	ppb
Vanadium	51-2	77.78	79.25	78.38	78.47	0.94	ppb
Yttrium	89-1				108		%
Yttrium	89-2				111		%
Zinc	66-2	84.98	86.59	84.53	85.37	1.27	ppb
Zirconium	90-1	2.11	2.13	2.08	2.10	1.17	ppb
Zirconium	91-1	2.18	2.07	2.15	2.13	2.75	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	422.77	422.88	419.57	421.74	0.45	ppb
Antimony	121-1	0.35	0.36	0.32	0.34	5.94	ppb
Arsenic	75-2	0.37	0.50	0.46	0.44	15.50	ppb
Barium	135-1	14.57	15.34	15.15	15.02	2.70	ppb
Barium	137-1	14.41	15.59	15.19	15.06	3.96	ppb
Beryllium	9-1	0.00	0.01	0.01	0.01	37.49	ppb
Bismuth	209-1				108		%
Bismuth	209-2				113		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.29	0.22	0.22	0.24	15.41	ppb
Cadmium	106-1	-0.71	0.89	-1.33	-0.38		ppb
Cadmium	111-1	0.02	0.15	-0.03	0.05	182.44	ppb
Calcium	43-1	2061.42	2096.66	2121.11	2093.06	1.43	ppb
Calcium	44-1	2082.75	2139.62	2092.16	2104.84	1.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	20.01	20.08	19.75	19.95	0.87	ppb
Cobalt	59-2	1.17	1.10	1.14	1.14	2.96	ppb
Copper	63-2	25.29	25.45	24.80	25.18	1.33	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				114		%
Holmium	165-2				114		%
Indium	115-1				116		%
Indium	115-2				115		%
Iron	56-2	144422.04	143607.74	141927.78	143319.19	0.89	ppb
Iron	57-2	141757.56	141001.33	137302.81	140020.56	1.70	ppb
Iron	54-2	146079.25	142278.00	140510.82	142956.02	1.99	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	7.70	7.79	7.90	7.80	1.29	ppb
Lead	207-1	7.23	7.47	7.49	7.40	1.98	ppb
Lead	208-1	7.45	7.49	7.64	7.53	1.33	ppb
Lithium	6-1				94		%
Magnesium	24-2	241.30	241.33	240.19	240.94	0.27	ppb
Manganese	55-2	335.10	337.78	329.15	334.01	1.32	ppb
Molybdenum	94-1	3.04	3.07	2.83	2.98	4.47	ppb
Molybdenum	95-1	2.88	3.01	2.86	2.92	2.83	ppb
Molybdenum	96-1	3.72	3.76	3.65	3.71	1.50	ppb
Molybdenum	97-1	2.87	2.99	2.98	2.95	2.21	ppb
Molybdenum	98-1	2.82	2.93	2.88	2.88	1.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	3.11	2.89	2.95	2.98	3.76	ppb
Phosphorus	31-2	-37.75	-40.24	-44.86	-40.95		ppb
Potassium	39-2	3333.34	3213.34	3156.99	3234.56	2.78	ppb
Rhodium	103-1				116		%
Rhodium	103-2				114		%
Scandium	45-1				112		%
Scandium	45-2				111		%
Selenium	82-1	0.35	-0.02	1.08	0.47	119.48	ppb
Selenium	77-2	1.63	1.63	1.66	1.64	1.10	ppb
Selenium	78-2	2.09	2.58	2.24	2.31	10.89	ppb
Silicon	28-1	127.53	132.17	130.42	130.04	1.80	ppb
Silver	107-1	0.04	0.04	0.04	0.04	2.35	ppb
Silver	109-1	0.05	0.04	0.04	0.04	3.06	ppb
Sodium	23-2	109145.16	109511.73	106681.65	108446.18	1.42	ppb
Strontium	86-1	8.87	9.19	8.96	9.01	1.86	ppb
Strontium	88-1	8.86	9.09	8.96	8.97	1.24	ppb
Sulfur	34-1	-752.52	-331.51	-220.24	-434.76		ppb
Terbium	159-1				115		%
Terbium	159-2				115		%
Thallium	203-1	-0.07	-0.08	-0.07	-0.07		ppb
Thallium	205-1	-0.07	-0.07	-0.07	-0.07		ppb
Tin	118-1	2.64	2.90	2.76	2.76	4.67	ppb
Titanium	47-1	5.80	6.10	6.00	5.97	2.48	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	14.12	ppb
Vanadium	51-2	1.17	1.18	1.13	1.16	2.17	ppb
Yttrium	89-1				110		%
Yttrium	89-2				110		%
Zinc	66-2	236.51	236.10	232.91	235.17	0.84	ppb
Zirconium	90-1	0.30	0.30	0.30	0.30	0.54	ppb
Zirconium	91-1	0.28	0.32	0.32	0.31	7.21	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	84.21	82.62	81.83	82.89	1.46	ppb
Antimony	121-1	1.79	1.93	2.07	1.93	7.29	ppb
Arsenic	75-2	4.30	4.29	4.56	4.38	3.48	ppb
Barium	135-1	7.92	7.97	7.77	7.89	1.35	ppb
Barium	137-1	7.69	7.89	7.86	7.81	1.33	ppb
Beryllium	9-1	0.02	0.01	0.00	0.01	85.58	ppb
Bismuth	209-1				75		%
Bismuth	209-2				77		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	31.85	33.89	33.10	32.95	3.11	ppb
Cadmium	106-1	0.01	-1.45	-2.42	-1.29		ppb
Cadmium	111-1	1.10	1.08	0.83	1.00	15.30	ppb
Calcium	43-1	411.17	433.26	456.63	433.69	5.24	ppb
Calcium	44-1	573.70	583.02	606.32	587.68	2.86	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.48	1.63	1.63	1.58	5.27	ppb
Cobalt	59-2	0.07	0.06	0.07	0.07	12.88	ppb
Copper	63-2	51.79	51.55	52.69	52.01	1.15	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				77		%
Holmium	165-2				78		%
Indium	115-1				79		%
Indium	115-2				79		%
Iron	56-2	100.56	102.45	103.62	102.21	1.51	ppb
Iron	57-2	110.37	113.46	112.32	112.05	1.40	ppb
Iron	54-2	107.71	108.21	106.39	107.44	0.88	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3.85	3.99	4.06	3.97	2.71	ppb
Lead	207-1	3.84	3.86	3.90	3.87	0.75	ppb
Lead	208-1	3.84	3.98	3.97	3.93	2.00	ppb
Lithium	6-1				69		%
Magnesium	24-2	159.89	163.41	165.72	163.01	1.80	ppb
Manganese	55-2	23.81	23.62	24.15	23.86	1.13	ppb
Molybdenum	94-1	7751.35	7929.44	8010.74	7897.18	1.68	ppb
Molybdenum	95-1	9412.10	9600.86	9956.24	9656.40	2.86	ppb
Molybdenum	96-1	9152.84	9420.98	9745.09	9439.64	3.14	ppb
Molybdenum	97-1	9463.45	9676.87	9949.14	9696.49	2.51	ppb
Molybdenum	98-1	9322.05	9698.87	9821.30	9614.07	2.71	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.45	0.45	0.67	0.52	24.08	ppb
Phosphorus	31-2	960.67	986.90	1013.65	987.07	2.68	ppb
Potassium	39-2	236005.00	228212.50	235077.52	233098.34	1.83	ppb
Rhodium	103-1				76		%
Rhodium	103-2				78		%
Scandium	45-1				73		%
Scandium	45-2				75		%
Selenium	82-1	9.49	11.16	11.54	10.73	10.16	ppb
Selenium	77-2	12.62	12.24	9.76	11.54	13.44	ppb
Selenium	78-2	17.60	17.99	17.18	17.59	2.32	ppb
Silicon	28-1	338.10	364.08	396.92	366.36	8.04	ppb
Silver	107-1	0.11	0.11	0.11	0.11	1.32	ppb
Silver	109-1	0.10	0.12	0.13	0.11	10.85	ppb
Sodium	23-2	97633.46	97670.80	99542.96	98282.40	1.11	ppb
Strontium	86-1	15.08	15.48	15.72	15.43	2.09	ppb
Strontium	88-1	14.73	15.09	15.42	15.08	2.29	ppb
Sulfur	34-1	2766.47	2223.29	2303.87	2431.21	12.06	ppb
Terbium	159-1				78		%
Terbium	159-2				79		%
Thallium	203-1	-0.03	-0.04	-0.03	-0.04		ppb
Thallium	205-1	-0.03	-0.03	-0.02	-0.03		ppb
Tin	118-1	1.26	1.32	1.34	1.31	3.10	ppb
Titanium	47-1	0.75	0.88	0.83	0.82	7.57	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.00	0.00	12.61	ppb
Vanadium	51-2	0.47	0.47	0.44	0.46	3.71	ppb
Yttrium	89-1				72		%
Yttrium	89-2				73		%
Zinc	66-2	203.72	202.50	206.84	204.35	1.10	ppb
Zirconium	90-1	0.07	0.04	0.05	0.06	26.94	ppb
Zirconium	91-1	0.07	0.07	0.04	0.06	25.38	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2073.15	2076.20	2048.19	2065.85	0.74	ppb
Antimony	121-1	0.54	0.65	0.75	0.65	16.72	ppb
Arsenic	75-2	4.73	4.91	4.55	4.73	3.82	ppb
Barium	135-1	5.71	6.44	5.81	5.99	6.65	ppb
Barium	137-1	5.58	5.98	5.82	5.79	3.49	ppb
Beryllium	9-1	0.01	0.00	0.01	0.01	83.40	ppb
Bismuth	209-1				78		%
Bismuth	209-2				78		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	6.14	6.16	7.34	6.55	10.49	ppb
Cadmium	106-1	-1.82	-2.15	-2.33	-2.10		ppb
Cadmium	111-1	0.19	0.20	0.19	0.20	2.22	ppb
Calcium	43-1	1590.58	1621.02	1636.22	1615.94	1.44	ppb
Calcium	44-1	1719.47	1753.84	1742.09	1738.47	1.00	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.30	1.29	1.28	1.29	0.75	ppb
Cobalt	59-2	0.01	0.01	0.02	0.01	55.86	ppb
Copper	63-2	62.74	62.82	63.09	62.88	0.30	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				78		%
Holmium	165-2				78		%
Indium	115-1				80		%
Indium	115-2				79		%
Iron	56-2	99.60	100.48	99.72	99.94	0.48	ppb
Iron	57-2	116.57	116.96	122.23	118.59	2.66	ppb
Iron	54-2	101.75	104.67	108.26	104.89	3.11	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.38	0.37	0.37	0.37	2.38	ppb
Lead	207-1	0.35	0.34	0.41	0.37	11.51	ppb
Lead	208-1	0.37	0.36	0.38	0.37	3.53	ppb
Lithium	6-1				70		%
Magnesium	24-2	127.44	123.73	122.88	124.68	1.95	ppb
Manganese	55-2	92.72	93.15	93.90	93.26	0.64	ppb
Molybdenum	94-1	1563.07	1645.96	1624.83	1611.29	2.67	ppb
Molybdenum	95-1	1892.22	1951.34	1910.37	1917.98	1.58	ppb
Molybdenum	96-1	1904.42	1943.12	1849.65	1899.07	2.47	ppb
Molybdenum	97-1	1932.57	1989.89	1967.24	1963.23	1.47	ppb
Molybdenum	98-1	1937.25	1982.43	1921.47	1947.05	1.63	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.38	0.51	0.44	0.44	14.88	ppb
Phosphorus	31-2	4.34	14.93	9.20	9.49	55.86	ppb
Potassium	39-2	153786.51	153981.99	153681.98	153816.83	0.10	ppb
Rhodium	103-1				79		%
Rhodium	103-2				79		%
Scandium	45-1				76		%
Scandium	45-2				76		%
Selenium	82-1	9.20	12.31	12.78	11.43	17.01	ppb
Selenium	77-2	11.04	13.27	8.62	10.98	21.18	ppb
Selenium	78-2	16.78	17.19	17.04	17.01	1.21	ppb
Silicon	28-1	184.65	208.00	214.50	202.38	7.75	ppb
Silver	107-1	0.02	0.01	0.02	0.02	18.87	ppb
Silver	109-1	0.02	0.02	0.02	0.02	15.27	ppb
Sodium	23-2	63441.59	63665.96	63248.13	63451.89	0.33	ppb
Strontium	86-1	6.02	5.75	5.99	5.92	2.51	ppb
Strontium	88-1	5.69	5.68	5.61	5.66	0.71	ppb
Sulfur	34-1	2340.15	2246.83	2053.00	2213.33	6.62	ppb
Terbium	159-1				80		%
Terbium	159-2				79		%
Thallium	203-1	0.00	-0.02	0.00	-0.01		ppb
Thallium	205-1	0.00	0.01	-0.01	0.00	735.00	ppb
Tin	118-1	1.00	1.18	1.14	1.11	8.41	ppb
Titanium	47-1	0.43	0.43	0.38	0.42	6.69	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	15.48	ppb
Vanadium	51-2	0.88	0.89	0.94	0.90	3.30	ppb
Yttrium	89-1				74		%
Yttrium	89-2				76		%
Zinc	66-2	386.08	388.21	383.05	385.78	0.67	ppb
Zirconium	90-1	0.08	0.08	0.07	0.08	3.29	ppb
Zirconium	91-1	0.08	0.09	0.05	0.07	25.20	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	48633.45	46745.29	48060.83	47813.19	2.02	ppb
Antimony	121-1	476.34	522.13	512.82	503.76	4.80	ppb
Arsenic	75-2	463.25	467.76	475.86	468.96	1.36	ppb
Barium	135-1	2365.08	2594.93	2703.98	2554.66	6.77	ppb
Barium	137-1	2398.58	2584.37	2650.75	2544.57	5.14	ppb
Beryllium	9-1	486.21	538.86	517.13	514.07	5.15	ppb
Bismuth	209-1				106		%
Bismuth	209-2				110		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	456.53	506.72	512.73	491.99	6.27	ppb
Cadmium	106-1	463.08	503.50	524.49	497.02	6.28	ppb
Cadmium	111-1	462.90	501.09	509.91	491.30	5.09	ppb
Calcium	43-1	248067.96	270846.52	264759.34	261224.60	4.51	ppb
Calcium	44-1	246192.01	265747.38	266997.77	259645.72	4.49	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	475.96	491.61	491.93	486.50	1.88	ppb
Cobalt	59-2	474.64	485.25	488.50	482.79	1.50	ppb
Copper	63-2	4726.98	4630.78	4861.10	4739.62	2.44	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				114		%
Holmium	165-2				117		%
Indium	115-1				113		%
Indium	115-2				115		%
Iron	56-2	126781.93	125844.40	126765.76	126464.03	0.42	ppb
Iron	57-2	125079.02	126232.97	128092.01	126468.00	1.20	ppb
Iron	54-2	125809.90	127533.75	128835.42	127393.02	1.19	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2473.08	2725.35	2622.61	2607.01	4.87	ppb
Lead	207-1	2474.70	2685.20	2563.89	2574.59	4.10	ppb
Lead	208-1	2470.63	2642.52	2578.61	2563.92	3.39	ppb
Lithium	6-1				98		%
Magnesium	24-2	241056.06	236132.97	243829.20	240339.41	1.62	ppb
Manganese	55-2	4795.77	4851.44	4873.64	4840.28	0.83	ppb
Molybdenum	94-1	4913.30	5414.35	5171.16	5166.27	4.85	ppb
Molybdenum	95-1	4867.31	5273.31	5171.34	5103.99	4.14	ppb
Molybdenum	96-1	4797.39	5261.00	5090.61	5049.67	4.64	ppb
Molybdenum	97-1	4843.67	5375.39	5079.46	5099.51	5.22	ppb
Molybdenum	98-1	4839.73	5322.95	5081.82	5081.50	4.75	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	471.67	466.21	473.52	470.47	0.81	ppb
Phosphorus	31-2	9464.06	9386.74	9638.52	9496.44	1.36	ppb
Potassium	39-2	123283.53	121983.39	126573.07	123946.67	1.91	ppb
Rhodium	103-1				111		%
Rhodium	103-2				113		%
Scandium	45-1				113		%
Scandium	45-2				116		%
Selenium	82-1	465.95	515.15	496.81	492.64	5.05	ppb
Selenium	77-2	454.81	474.31	480.28	469.80	2.84	ppb
Selenium	78-2	450.33	459.87	468.70	459.63	2.00	ppb
Silicon	28-1	497.83	572.45	541.03	537.10	6.98	ppb
Silver	107-1	479.38	519.50	528.43	509.11	5.13	ppb
Silver	109-1	467.03	500.07	526.83	497.98	6.02	ppb
Sodium	23-2	241477.64	240621.45	238497.88	240198.99	0.64	ppb
Strontium	86-1	476.76	526.51	499.39	500.89	4.97	ppb
Strontium	88-1	485.93	537.99	505.48	509.80	5.16	ppb
Sulfur	34-1	6977.24	8523.99	8189.29	7896.84	10.31	ppb
Terbium	159-1				116		%
Terbium	159-2				118		%
Thallium	203-1	515.60	548.02	524.36	529.33	3.17	ppb
Thallium	205-1	498.95	543.13	524.41	522.16	4.25	ppb
Tin	118-1	485.82	517.47	515.43	506.24	3.50	ppb
Titanium	47-1	4706.70	5224.58	5040.93	4990.74	5.26	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	479.51	516.21	508.44	501.39	3.86	ppb
Vanadium	51-2	483.50	480.80	485.93	483.41	0.53	ppb
Yttrium	89-1				108		%
Yttrium	89-2				113		%
Zinc	66-2	4648.97	4670.04	4726.97	4681.99	0.86	ppb
Zirconium	90-1	482.27	520.38	510.58	504.41	3.92	ppb
Zirconium	91-1	474.44	517.47	499.81	497.24	4.35	ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-2.28	-2.84	-1.97	-2.36		ppb
Antimony	121-1	0.04	0.04	0.04	0.04	1.51	ppb
Arsenic	75-2	0.01	0.02	0.05	0.03	77.22	ppb
Barium	135-1	-0.01	-0.01	-0.01	-0.01		ppb
Barium	137-1	-0.02	0.02	-0.01	0.00		ppb
Beryllium	9-1	0.00	0.00	0.01	0.01	55.78	ppb
Bismuth	209-1				108		%
Bismuth	209-2				108		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.12	-0.06	0.15	0.07	155.57	ppb
Cadmium	106-1	-2.35	-1.90	-3.45	-2.57		ppb
Cadmium	111-1	-0.16	-0.12	-0.25	-0.17		ppb
Calcium	43-1	-12.65	-7.11	-9.77	-9.84		ppb
Calcium	44-1	26.21	27.06	29.83	27.70	6.84	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.04	-0.04	-0.05	-0.05		ppb
Cobalt	59-2	-0.02	-0.02	-0.02	-0.02		ppb
Copper	63-2	-0.01	0.00	-0.05	-0.02		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				107		%
Indium	115-1				112		%
Indium	115-2				111		%
Iron	56-2	-0.85	-0.89	-0.81	-0.85		ppb
Iron	57-2	-2.40	-2.71	-0.68	-1.93		ppb
Iron	54-2	26.26	-0.62	0.46	8.70	174.92	ppb
Krypton	83-1						cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.02	-0.02	-0.02	-0.02		ppb
Lead	207-1	-0.03	-0.04	-0.04	-0.04		ppb
Lead	208-1	-0.03	-0.03	-0.03	-0.03		ppb
Lithium	6-1				95		%
Magnesium	24-2	-5.60	-6.29	-5.73	-5.87		ppb
Manganese	55-2	-0.04	-0.03	0.01	-0.02		ppb
Molybdenum	94-1	0.03	0.03	0.04	0.03	25.18	ppb
Molybdenum	95-1	0.02	0.02	0.02	0.02	12.27	ppb
Molybdenum	96-1	0.06	0.03	0.06	0.05	40.63	ppb
Molybdenum	97-1	0.08	0.06	0.05	0.06	19.79	ppb
Molybdenum	98-1	0.02	0.04	0.02	0.03	48.77	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.01	0.03	0.05	0.03	64.75	ppb
Phosphorus	31-2	-70.79	-66.75	-53.15	-63.56		ppb
Potassium	39-2	24.54	16.96	23.94	21.81	19.33	ppb
Rhodium	103-1				111		%
Rhodium	103-2				111		%
Scandium	45-1				109		%
Scandium	45-2				106		%
Selenium	82-1	0.34	1.30	0.50	0.71	72.42	ppb
Selenium	77-2	1.51	2.59	1.75	1.95	29.15	ppb
Selenium	78-2	2.81	2.12	3.24	2.72	20.85	ppb
Silicon	28-1	-1.59	-0.70	-0.41	-0.90		ppb
Silver	107-1	0.04	0.03	0.03	0.04	17.50	ppb
Silver	109-1	0.04	0.03	0.03	0.03	24.10	ppb
Sodium	23-2	29.31	27.49	29.78	28.86	4.19	ppb
Strontium	86-1	-0.01	-0.02	0.03	0.00		ppb
Strontium	88-1	0.00	0.00	0.00	0.00	422.96	ppb
Sulfur	34-1	-1418.93	-1221.80	-1169.40	-1270.04		ppb
Terbium	159-1				108		%
Terbium	159-2				109		%
Thallium	203-1	-0.07	-0.08	-0.07	-0.07		ppb
Thallium	205-1	-0.06	-0.07	-0.07	-0.06		ppb
Tin	118-1	-0.04	-0.05	-0.06	-0.05		ppb
Titanium	47-1	-0.04	0.01	-0.03	-0.02		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.00	0.01	0.00	0.00	280.57	ppb
Yttrium	89-1				105		%
Yttrium	89-2				104		%
Zinc	66-2	-0.28	-0.48	-0.61	-0.46		ppb
Zirconium	90-1	0.00	0.00	0.01	0.01	92.55	ppb
Zirconium	91-1	-0.01	-0.02	0.00	-0.01		ppb

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1870	1977	1860	1902	3.40	cps
Antimony	121-1	53	43	37	44	18.87	cps
Arsenic	75-2	13	17	13	14	13.35	cps
Barium	135-1	137	147	153	146	5.76	cps
Barium	137-1	163	220	237	207	18.60	cps
Beryllium	9-1	3	3	7	5	43.02	cps
Bismuth	209-1	4685697	4609567	4686321	4660528	0.95	cps
Bismuth	209-2	4058298	4174512	4230726	4154512	2.12	cps
Bromine	81-1	9650	9446	9386	9494	1.45	cps
Bromine	81-2	250	187	267	234	18.00	cps
Cadmium	108-1	13	0	13	9	86.60	cps
Cadmium	106-1	2930	2737	2907	2858	3.69	cps
Cadmium	111-1	2055	1926	2056	2012	3.72	cps
Calcium	43-1	643	647	643	644	0.30	cps
Calcium	44-1	27936	28370	29272	28526	2.39	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2537	2770	2860	2723	6.13	cps
Cobalt	59-2	163	223	163	183	18.89	cps
Copper	63-2	1537	1467	1540	1515	2.73	cps
Dysprosium	156-1	3	10	17	10	66.70	cps
Dysprosium	156-2	0	7	0	2	173.21	cps
Erbium	164-1	43	50	73	56	28.35	cps
Erbium	164-2	23	30	27	27	12.51	cps
Gadolinium	160-1	60	57	20	46	48.72	cps
Gadolinium	160-2	417	473	353	414	14.49	cps
Holmium	165-1	7233970	7258752	7344554	7279092	0.80	cps
Holmium	165-2	5525391	5575146	5507415	5535984	0.63	cps
Indium	115-1	6277346	6157988	6122314	6185883	1.31	cps
Indium	115-2	2548402	2525083	2550891	2541458	0.56	cps
Iron	56-2	28497	27285	26734	27505	3.28	cps
Iron	57-2	1933	2020	1890	1948	3.40	cps
Iron	54-2	3110	3234	3067	3137	2.76	cps
Krypton	83-1	480	483	483	482	0.40	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	733	743	713	730	2.09	cps
Lead	207-1	607	637	660	634	4.21	cps
Lead	208-1	2923	2897	2833	2885	1.60	cps
Lithium	6-1	764906	742010	737455	748124	1.97	cps
Magnesium	24-2	3014	2960	3150	3041	3.22	cps
Manganese	55-2	1157	1198	1157	1170	2.03	cps
Molybdenum	94-1	300	307	280	296	4.70	cps
Molybdenum	95-1	310	283	297	297	4.49	cps
Molybdenum	96-1	300	310	333	314	5.44	cps
Molybdenum	97-1	180	200	213	198	8.48	cps
Molybdenum	98-1	413	473	407	431	8.52	cps
Neodymium	150-1	7	0	10	6	91.64	cps
Neodymium	150-2	13	10	10	11	17.30	cps
Nickel	60-2	593	677	593	621	7.75	cps
Phosphorus	31-2	947	970	937	951	1.80	cps
Potassium	39-2	78219	78444	78343	78336	0.14	cps
Rhodium	103-1	6291654	6089343	6145891	6175629	1.69	cps
Rhodium	103-2	4002330	3980855	3966977	3983387	0.45	cps
Scandium	45-1	3809159	3734871	3766403	3770144	0.99	cps
Scandium	45-2	423442	422871	414261	420191	1.22	cps
Selenium	82-1	208	150	199	186	16.79	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	447	467	437	450	3.39	cps
Silicon	28-1	545780	545401	553688	548290	0.85	cps
Silver	107-1	143	130	123	132	7.70	cps
Silver	109-1	107	67	90	88	22.89	cps
Sodium	23-2	39006	39267	38662	38978	0.78	cps
Strontium	86-1	827	823	860	837	2.42	cps
Strontium	88-1	717	773	800	763	5.58	cps
Sulfur	34-1	165515	165946	166196	165886	0.21	cps
Terbium	159-1	7722755	7549724	7181996	7484825	3.69	cps
Terbium	159-2	5431942	5548641	5563174	5514586	1.30	cps
Thallium	203-1	520	493	460	491	6.12	cps
Thallium	205-1	993	1060	1123	1059	6.14	cps
Tin	118-1	2140	2027	1943	2037	4.85	cps
Titanium	47-1	83	113	117	104	17.58	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:34:51 DataFile Name : 002CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	37	20	43	33	36.05	cps
Vanadium	51-2	7	27	37	23	65.46	cps
Yttrium	89-1	11582816	11424621	11305372	11437603	1.22	cps
Yttrium	89-2	3409172	3339090	3397418	3381893	1.11	cps
Zinc	66-2	4537	4617	4634	4596	1.12	cps
Zirconium	90-1	363	417	413	398	7.51	cps
Zirconium	91-1	90	117	63	90	29.63	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5371	5151	5144	5222	2.47	cps
Antimony	121-1	11681	11338	11658	11559	1.66	cps
Arsenic	75-2	453	417	420	430	4.72	cps
Barium	135-1	13967	13820	13944	13910	0.57	cps
Barium	137-1	24298	24134	23830	24087	0.98	cps
Beryllium	9-1	888	943	827	886	6.53	cps
Bismuth	209-1	4599785	4675228	4610566	4628526	0.88	cps
Bismuth	209-2	4193556	4110349	4053012	4118972	1.72	cps
Bromine	81-1	9767	9533	9416	9572	1.86	cps
Bromine	81-2	267	220	190	226	17.13	cps
Cadmium	108-1	133	127	97	119	16.43	cps
Cadmium	106-1	3060	2994	3187	3080	3.19	cps
Cadmium	111-1	3842	3707	3700	3750	2.14	cps
Calcium	43-1	9290	9366	9313	9323	0.42	cps
Calcium	44-1	168013	168614	169631	168753	0.48	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	9163	9560	9216	9313	2.31	cps
Cobalt	59-2	5548	5554	5494	5532	0.59	cps
Copper	63-2	8969	8896	9073	8979	0.99	cps
Dysprosium	156-1	7	17	0	8	107.84	cps
Dysprosium	156-2	3	10	10	8	49.52	cps
Erbium	164-1	70	73	60	68	10.24	cps
Erbium	164-2	23	30	47	33	36.07	cps
Gadolinium	160-1	10	47	57	38	65.04	cps
Gadolinium	160-2	390	420	407	406	3.71	cps
Holmium	165-1	7392078	7381365	7249640	7341028	1.08	cps
Holmium	165-2	5644497	5313010	5389034	5448847	3.19	cps
Indium	115-1	6196739	6247201	6241731	6228557	0.44	cps
Indium	115-2	2543654	2481618	2568470	2531247	1.77	cps
Iron	56-2	182704	183445	184114	183421	0.38	cps
Iron	57-2	6362	5638	5885	5961	6.17	cps
Iron	54-2	11945	11868	11785	11866	0.67	cps
Krypton	83-1	457	480	397	444	9.67	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	4914	5134	4864	4971	2.89	cps
Lead	207-1	4167	4224	4527	4306	4.50	cps
Lead	208-1	19759	19632	19532	19641	0.58	cps
Lithium	6-1	783982	770892	765757	773544	1.21	cps
Magnesium	24-2	159946	156270	157246	157821	1.21	cps
Manganese	55-2	3881	3795	3727	3801	2.04	cps
Molybdenum	94-1	13019	12772	12636	12809	1.52	cps
Molybdenum	95-1	16373	15785	15729	15962	2.23	cps
Molybdenum	96-1	17250	17414	17007	17224	1.19	cps
Molybdenum	97-1	9573	10154	9793	9840	2.98	cps
Molybdenum	98-1	25552	25553	25803	25636	0.56	cps
Neodymium	150-1	10	3	7	7	50.03	cps
Neodymium	150-2	7	0	0	2	173.21	cps
Nickel	60-2	1893	1763	1807	1821	3.63	cps
Phosphorus	31-2	1230	1293	1240	1255	2.71	cps
Potassium	39-2	273725	272642	269214	271860	0.87	cps
Rhodium	103-1	6226001	6209316	6087834	6174384	1.22	cps
Rhodium	103-2	3967486	3944976	3976679	3963047	0.41	cps
Scandium	45-1	3802461	3761988	3805137	3789862	0.64	cps
Scandium	45-2	429017	415569	417248	420611	1.74	cps
Selenium	82-1	805	704	854	788	9.71	cps
Selenium	77-2	93	97	47	79	35.43	cps
Selenium	78-2	727	757	793	759	4.40	cps
Silicon	28-1	555583	558788	560086	558152	0.42	cps
Silver	107-1	8256	8016	8089	8120	1.52	cps
Silver	109-1	7989	7692	7946	7876	2.04	cps
Sodium	23-2	280596	278573	276043	278404	0.82	cps
Strontium	86-1	2697	2767	2844	2769	2.65	cps
Strontium	88-1	17634	17798	17487	17640	0.88	cps
Sulfur	34-1	168716	169363	168683	168921	0.23	cps
Terbium	159-1	7743090	7628231	7537182	7636168	1.35	cps
Terbium	159-2	5572056	5555588	5430834	5519493	1.40	cps
Thallium	203-1	5865	5555	5555	5658	3.16	cps
Thallium	205-1	13177	13307	13357	13280	0.70	cps
Tin	118-1	26965	26137	25539	26214	2.73	cps
Titanium	47-1	4147	4241	4127	4172	1.45	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S2 Instrumnet Name : P7
Client Sample ID : S2 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:48:25 DataFile Name : 006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16954	16380	16711	16682	1.73	cps
Vanadium	51-2	14757	14254	14427	14479	1.77	cps
Yttrium	89-1	10876401	10793652	10573135	10747729	1.46	cps
Yttrium	89-2	3151405	3220004	3215080	3195496	1.20	cps
Zinc	66-2	7729	7392	7499	7540	2.28	cps
Zirconium	90-1	10607	10434	10717	10586	1.35	cps
Zirconium	91-1	2144	2207	2347	2232	4.66	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	176880	175516	174787	175728	0.60	cps
Antimony	121-1	273092	276360	271607	273686	0.89	cps
Arsenic	75-2	19690	19773	19603	19689	0.43	cps
Barium	135-1	320231	326345	318746	321774	1.25	cps
Barium	137-1	564981	562673	562983	563546	0.22	cps
Beryllium	9-1	40616	40862	40921	40800	0.40	cps
Bismuth	209-1	4740245	4754834	4793607	4762895	0.58	cps
Bismuth	209-2	4259280	4246272	4121671	4209074	1.80	cps
Bromine	81-1	9380	9670	9640	9563	1.67	cps
Bromine	81-2	153	230	197	193	19.88	cps
Cadmium	108-1	5504	5621	5371	5499	2.28	cps
Cadmium	106-1	10814	10651	10777	10747	0.80	cps
Cadmium	111-1	79404	81202	78905	79837	1.51	cps
Calcium	43-1	82922	82517	81380	82273	0.97	cps
Calcium	44-1	1269607	1280046	1301350	1283668	1.26	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	156922	160341	157909	158391	1.11	cps
Cobalt	59-2	238819	233763	238852	237145	1.23	cps
Copper	63-2	1858267	1822247	1841611	1840709	0.98	cps
Dysprosium	156-1	13	20	3	12	68.66	cps
Dysprosium	156-2	30	17	23	23	28.56	cps
Erbium	164-1	53	77	50	60	24.22	cps
Erbium	164-2	63	50	53	56	12.49	cps
Gadolinium	160-1	47	53	63	54	15.40	cps
Gadolinium	160-2	400	377	400	392	3.43	cps
Holmium	165-1	7591325	7503463	7440975	7511921	1.01	cps
Holmium	165-2	5664170	5611194	5501563	5592309	1.48	cps
Indium	115-1	6287432	6346071	6334147	6322550	0.49	cps
Indium	115-2	2574243	2599731	2600811	2591595	0.58	cps
Iron	56-2	7329805	7309349	7305746	7314967	0.18	cps
Iron	57-2	191877	189307	191050	190745	0.69	cps
Iron	54-2	435013	431333	431561	432636	0.48	cps
Krypton	83-1	430	510	410	450	11.76	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1016312	1012253	1003940	1010835	0.62	cps
Lead	207-1	872818	879244	869494	873852	0.57	cps
Lead	208-1	4083079	4171264	4077415	4110586	1.28	cps
Lithium	6-1	781934	791585	774511	782677	1.09	cps
Magnesium	24-2	1520685	1533045	1502533	1518755	1.01	cps
Manganese	55-2	1337539	1317820	1324997	1326785	0.75	cps
Molybdenum	94-1	1004661	1010683	997370	1004238	0.66	cps
Molybdenum	95-1	1454665	1473744	1442233	1456880	1.09	cps
Molybdenum	96-1	1539416	1590300	1548055	1559257	1.75	cps
Molybdenum	97-1	893057	901796	895646	896833	0.50	cps
Molybdenum	98-1	2284080	2299823	2244079	2275994	1.26	cps
Neodymium	150-1	43	10	23	26	65.65	cps
Neodymium	150-2	10	10	3	8	49.52	cps
Nickel	60-2	62011	62038	63146	62398	1.04	cps
Phosphorus	31-2	10574	10250	10257	10360	1.78	cps
Potassium	39-2	991907	986231	995152	991097	0.46	cps
Rhodium	103-1	6216386	6303585	6309621	6276531	0.83	cps
Rhodium	103-2	4091512	3985399	3984575	4020495	1.53	cps
Scandium	45-1	3869420	3894168	3847403	3870330	0.60	cps
Scandium	45-2	432354	431054	424606	429338	0.97	cps
Selenium	82-1	5842	5947	6111	5967	2.28	cps
Selenium	77-2	703	653	713	690	4.66	cps
Selenium	78-2	2860	2997	2720	2859	4.84	cps
Silicon	28-1	1448618	1486477	1469060	1468052	1.29	cps
Silver	107-1	376285	382132	378382	378933	0.78	cps
Silver	109-1	359427	369897	369472	366265	1.62	cps
Sodium	23-2	2364393	2362110	2364467	2363657	0.06	cps
Strontium	86-1	93431	95451	95159	94680	1.15	cps
Strontium	88-1	815713	825397	814850	818653	0.72	cps
Sulfur	34-1	192152	193367	192426	192648	0.33	cps
Terbium	159-1	7819494	7653261	7700748	7724501	1.11	cps
Terbium	159-2	5653454	5708095	5590946	5650832	1.04	cps
Thallium	203-1	248436	248911	247651	248333	0.26	cps
Thallium	205-1	588361	589673	584328	587454	0.47	cps
Tin	118-1	231780	232907	231705	232131	0.29	cps
Titanium	47-1	380641	382689	380306	381212	0.34	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S3 Instrumnet Name : P7
Client Sample ID : S3 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:51:42 DataFile Name : 007CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	779409	783748	777667	780275	0.40	cps
Vanadium	51-2	131848	130836	130550	131078	0.52	cps
Yttrium	89-1	10820420	10822229	10635278	10759309	1.00	cps
Yttrium	89-2	3166598	3260009	3221128	3215912	1.46	cps
Zinc	66-2	350072	344851	349453	348126	0.82	cps
Zirconium	90-1	488572	497719	492451	492914	0.93	cps
Zirconium	91-1	110067	110113	109797	109992	0.16	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	441518	439710	447513	442914	0.92	cps
Antimony	121-1	697089	698513	694883	696828	0.26	cps
Arsenic	75-2	48870	49191	48322	48795	0.90	cps
Barium	135-1	822849	834122	820488	825820	0.88	cps
Barium	137-1	1431108	1440488	1442438	1438012	0.42	cps
Beryllium	9-1	101173	101473	101769	101472	0.29	cps
Bismuth	209-1	4847156	4756145	4752878	4785393	1.12	cps
Bismuth	209-2	4198425	4165043	4179382	4180950	0.40	cps
Bromine	81-1	9700	9380	9396	9492	1.90	cps
Bromine	81-2	160	200	227	196	17.16	cps
Cadmium	108-1	14167	14250	14407	14275	0.85	cps
Cadmium	106-1	22905	22204	22641	22583	1.57	cps
Cadmium	111-1	200076	200327	201641	200682	0.42	cps
Calcium	43-1	208623	209092	208695	208803	0.12	cps
Calcium	44-1	3233964	3267207	3217092	3239421	0.79	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	397410	397555	396924	397296	0.08	cps
Cobalt	59-2	602683	599265	598025	599991	0.40	cps
Copper	63-2	4567405	4741245	4584289	4630980	2.07	cps
Dysprosium	156-1	27	30	30	29	6.65	cps
Dysprosium	156-2	43	63	77	61	27.46	cps
Erbium	164-1	97	57	80	78	25.83	cps
Erbium	164-2	87	40	60	62	37.63	cps
Gadolinium	160-1	70	67	77	71	7.16	cps
Gadolinium	160-2	487	367	440	431	14.03	cps
Holmium	165-1	7413258	7356145	7562559	7443987	1.43	cps
Holmium	165-2	5670918	5631820	5523294	5608678	1.36	cps
Indium	115-1	6213886	6224247	6332101	6256744	1.05	cps
Indium	115-2	2590633	2581723	2539600	2570652	1.06	cps
Iron	56-2	18473682	18605033	18390173	18489629	0.59	cps
Iron	57-2	478845	484896	473477	479073	1.19	cps
Iron	54-2	1087448	1073444	1084990	1081960	0.69	cps
Krypton	83-1	480	473	547	500	8.11	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	2671348	2715939	2699464	2695584	0.84	cps
Lead	207-1	2379117	2324387	2365507	2356337	1.21	cps
Lead	208-1	10869033	10856618	10776734	10834128	0.46	cps
Lithium	6-1	787587	783824	801197	790869	1.16	cps
Magnesium	24-2	3926321	3840032	3911513	3892622	1.19	cps
Manganese	55-2	3234458	3375927	3357670	3322685	2.32	cps
Molybdenum	94-1	2522107	2548797	2614452	2561785	1.86	cps
Molybdenum	95-1	3771962	3748983	3842586	3787844	1.29	cps
Molybdenum	96-1	4113438	4074443	4078185	4088689	0.53	cps
Molybdenum	97-1	2267367	2310875	2247082	2275108	1.43	cps
Molybdenum	98-1	5928072	5952378	5944917	5941789	0.21	cps
Neodymium	150-1	33	30	33	32	5.97	cps
Neodymium	150-2	10	20	10	13	43.30	cps
Nickel	60-2	156522	156525	158749	157265	0.82	cps
Phosphorus	31-2	25712	25061	24934	25235	1.65	cps
Potassium	39-2	2320418	2338301	2297027	2318582	0.89	cps
Rhodium	103-1	6227475	6220509	6207421	6218468	0.16	cps
Rhodium	103-2	4062597	3992873	3949537	4001669	1.43	cps
Scandium	45-1	3861810	3813798	3841083	3838897	0.63	cps
Scandium	45-2	429972	422381	427816	426723	0.92	cps
Selenium	82-1	14697	14899	14619	14738	0.98	cps
Selenium	77-2	1887	1903	1720	1837	5.52	cps
Selenium	78-2	6662	6742	6605	6669	1.03	cps
Silicon	28-1	2573423	2622751	2569970	2588714	1.14	cps
Silver	107-1	971896	968392	965230	968506	0.34	cps
Silver	109-1	934323	935808	925087	931739	0.62	cps
Sodium	23-2	5938004	5868321	6064389	5956905	1.67	cps
Strontium	86-1	241148	242641	239986	241258	0.55	cps
Strontium	88-1	2079546	2114036	2093637	2095740	0.83	cps
Sulfur	34-1	230120	228591	227296	228669	0.62	cps
Terbium	159-1	7859658	7545518	7701687	7702288	2.04	cps
Terbium	159-2	5656909	5524531	5580843	5587428	1.19	cps
Thallium	203-1	632325	631284	626109	629906	0.53	cps
Thallium	205-1	1490170	1509872	1487737	1495926	0.81	cps
Tin	118-1	595592	592892	585334	591272	0.90	cps
Titanium	47-1	974649	970690	963519	969619	0.58	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S4 Instrumnet Name : P7
Client Sample ID : S4 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:54:53 DataFile Name : 008CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	2108231	2094159	2100143	2100844	0.34	cps
Vanadium	51-2	331796	332578	329367	331247	0.51	cps
Yttrium	89-1	10913619	10822914	11031482	10922672	0.96	cps
Yttrium	89-2	3197507	3242060	3114171	3184579	2.04	cps
Zinc	66-2	877764	866280	860689	868244	1.00	cps
Zirconium	90-1	1253419	1252241	1259152	1254937	0.29	cps
Zirconium	91-1	281804	280574	277082	279820	0.88	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	875620	869979	860043	868547	0.91	cps
Antimony	121-1	1371258	1376642	1379811	1375903	0.31	cps
Arsenic	75-2	100142	98233	97877	98750	1.23	cps
Barium	135-1	1674279	1646073	1705621	1675324	1.78	cps
Barium	137-1	2886829	2925153	2951611	2921198	1.12	cps
Beryllium	9-1	199425	202831	199772	200676	0.93	cps
Bismuth	209-1	4747196	4733023	4779934	4753384	0.51	cps
Bismuth	209-2	4142629	4160756	4119881	4141088	0.49	cps
Bromine	81-1	9096	9196	9433	9242	1.87	cps
Bromine	81-2	237	190	277	234	18.50	cps
Cadmium	108-1	27690	28368	27650	27902	1.45	cps
Cadmium	106-1	41429	41787	41259	41492	0.65	cps
Cadmium	111-1	389238	393961	390806	391335	0.61	cps
Calcium	43-1	405566	407597	405585	406249	0.29	cps
Calcium	44-1	6267334	6380990	6344356	6330893	0.92	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	779326	769872	783038	777412	0.87	cps
Cobalt	59-2	1181518	1177032	1182420	1180323	0.24	cps
Copper	63-2	9270178	9159988	9118763	9182976	0.85	cps
Dysprosium	156-1	43	37	23	34	29.57	cps
Dysprosium	156-2	123	97	100	107	13.62	cps
Erbium	164-1	110	93	110	104	9.21	cps
Erbium	164-2	73	63	37	58	32.80	cps
Gadolinium	160-1	63	73	47	61	22.04	cps
Gadolinium	160-2	377	413	353	381	7.94	cps
Holmium	165-1	7533259	7426945	7482876	7481026	0.71	cps
Holmium	165-2	5671204	5558961	5547714	5592627	1.22	cps
Indium	115-1	6365156	6261082	6283722	6303320	0.87	cps
Indium	115-2	2586667	2554689	2576262	2572540	0.63	cps
Iron	56-2	36587105	36823374	36237294	36549258	0.81	cps
Iron	57-2	953987	954654	943879	950840	0.63	cps
Iron	54-2	2068770	2035454	2036853	2047026	0.92	cps
Krypton	83-1	480	497	450	476	4.97	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	5200922	5281700	5377476	5286699	1.67	cps
Lead	207-1	4645904	4659252	4669872	4658342	0.26	cps
Lead	208-1	21140861	21426560	21594265	21387228	1.07	cps
Lithium	6-1	795201	782103	779576	785627	1.07	cps
Magnesium	24-2	7710046	7756854	7446208	7637703	2.19	cps
Manganese	55-2	6646393	6639912	6536679	6607661	0.93	cps
Molybdenum	94-1	4899746	5040215	5061599	5000520	1.76	cps
Molybdenum	95-1	7382875	7438224	7427145	7416081	0.39	cps
Molybdenum	96-1	8050086	8077297	8025632	8051005	0.32	cps
Molybdenum	97-1	4480246	4544072	4524066	4516128	0.72	cps
Molybdenum	98-1	11806383	11812677	11660169	11759743	0.73	cps
Neodymium	150-1	60	70	43	58	23.32	cps
Neodymium	150-2	23	23	7	18	54.11	cps
Nickel	60-2	308108	309231	303060	306800	1.07	cps
Phosphorus	31-2	49748	49333	47410	48830	2.55	cps
Potassium	39-2	4544318	4494906	4452566	4497263	1.02	cps
Rhodium	103-1	6119398	6232475	6275180	6209018	1.30	cps
Rhodium	103-2	4026753	3983066	3943432	3984417	1.05	cps
Scandium	45-1	3771167	3770537	3780823	3774176	0.15	cps
Scandium	45-2	426983	421918	420419	423106	0.81	cps
Selenium	82-1	28427	28877	28602	28635	0.79	cps
Selenium	77-2	3657	3714	3334	3568	5.75	cps
Selenium	78-2	12409	12792	12602	12601	1.52	cps
Silicon	28-1	4290238	4355284	4319808	4321777	0.75	cps
Silver	107-1	1895013	1950508	1908010	1917844	1.51	cps
Silver	109-1	1842956	1832527	1855674	1843719	0.63	cps
Sodium	23-2	11897241	11652247	11463554	11671014	1.86	cps
Strontium	86-1	470678	479808	475881	475456	0.96	cps
Strontium	88-1	4102410	4231863	4071913	4135395	2.05	cps
Sulfur	34-1	282017	281400	281590	281669	0.11	cps
Terbium	159-1	7842195	7597118	7828450	7755921	1.78	cps
Terbium	159-2	5612464	5687441	5449592	5583166	2.18	cps
Thallium	203-1	1235874	1242517	1242779	1240390	0.32	cps
Thallium	205-1	3148028	3023995	3162403	3111475	2.45	cps
Tin	118-1	1152251	1171477	1157399	1160376	0.86	cps
Titanium	47-1	1888524	1943779	1945011	1925771	1.68	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S5 Instrumnet Name : P7
Client Sample ID : S5 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 12:57:53 DataFile Name : 009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4146462	4157153	4243151	4182256	1.27	cps
Vanadium	51-2	664404	659085	657659	660383	0.54	cps
Yttrium	89-1	10892197	10661035	10699328	10750854	1.15	cps
Yttrium	89-2	3248956	3171147	3167250	3195784	1.44	cps
Zinc	66-2	1759668	1695717	1700804	1718730	2.07	cps
Zirconium	90-1	2506401	2487855	2513247	2502501	0.52	cps
Zirconium	91-1	549497	556760	558139	554799	0.84	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1739591	1732278	1751942	1741270	0.57	cps
Antimony	121-1	2808215	2868783	2799180	2825393	1.34	cps
Arsenic	75-2	194051	194209	193831	194030	0.10	cps
Barium	135-1	3396133	3468175	3405751	3423353	1.14	cps
Barium	137-1	5772129	5854378	5774358	5800288	0.81	cps
Beryllium	9-1	397714	400348	393641	397234	0.85	cps
Bismuth	209-1	5046556	4577329	4848967	4824284	4.88	cps
Bismuth	209-2	4325566	4298531	4194783	4272960	1.62	cps
Bromine	81-1	8926	9356	9360	9214	2.71	cps
Bromine	81-2	163	200	190	184	10.27	cps
Cadmium	108-1	55703	57005	54827	55845	1.96	cps
Cadmium	106-1	78652	80643	79540	79611	1.25	cps
Cadmium	111-1	774106	785430	781381	780306	0.74	cps
Calcium	43-1	806307	821176	812279	813254	0.92	cps
Calcium	44-1	12428494	12713054	12501519	12547689	1.18	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1531112	1526588	1556788	1538163	1.06	cps
Cobalt	59-2	2367261	2342758	2374337	2361452	0.70	cps
Copper	63-2	18447590	18233100	18003540	18228077	1.22	cps
Dysprosium	156-1	47	53	80	60	29.40	cps
Dysprosium	156-2	203	203	187	198	4.87	cps
Erbium	164-1	167	90	130	129	29.75	cps
Erbium	164-2	100	80	93	91	11.18	cps
Gadolinium	160-1	120	73	110	101	24.30	cps
Gadolinium	160-2	403	427	343	391	10.99	cps
Holmium	165-1	7749853	7182729	7539073	7490552	3.83	cps
Holmium	165-2	5895976	5785811	5640151	5773979	2.22	cps
Indium	115-1	6337967	6028526	6239901	6202131	2.55	cps
Indium	115-2	2655551	2643163	2591673	2630129	1.29	cps
Iron	56-2	72396334	72117661	72749549	72421181	0.44	cps
Iron	57-2	1823075	1782937	1838401	1814804	1.58	cps
Iron	54-2	4047860	4022180	4073402	4047814	0.63	cps
Krypton	83-1	483	510	443	479	7.01	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10938337	10804083	10655332	10799251	1.31	cps
Lead	207-1	9374122	9554934	9401197	9443417	1.03	cps
Lead	208-1	43285638	43580712	43154708	43340353	0.50	cps
Lithium	6-1	793704	759806	769638	774382	2.25	cps
Magnesium	24-2	15051656	14922306	15432728	15135563	1.75	cps
Manganese	55-2	13261174	13174276	13103590	13179680	0.60	cps
Molybdenum	94-1	10002759	10064644	10067158	10044854	0.36	cps
Molybdenum	95-1	14494628	14943205	14985328	14807720	1.84	cps
Molybdenum	96-1	15546940	16323474	16057714	15976043	2.47	cps
Molybdenum	97-1	8955208	9147359	9105772	9069446	1.11	cps
Molybdenum	98-1	23590793	23825454	23331104	23582450	1.05	cps
Neodymium	150-1	103	97	140	113	20.59	cps
Neodymium	150-2	50	20	20	30	57.74	cps
Nickel	60-2	618453	603361	600097	607304	1.61	cps
Phosphorus	31-2	96355	93766	95533	95218	1.39	cps
Potassium	39-2	8897438	8799126	8725473	8807346	0.98	cps
Rhodium	103-1	6370677	5917718	6246178	6178191	3.79	cps
Rhodium	103-2	4090757	4133041	3994365	4072721	1.75	cps
Scandium	45-1	3852210	3683272	3776899	3770794	2.24	cps
Scandium	45-2	439622	438821	430635	436360	1.14	cps
Selenium	82-1	56858	57475	57160	57164	0.54	cps
Selenium	77-2	7352	6882	7008	7081	3.44	cps
Selenium	78-2	24337	24444	24230	24337	0.44	cps
Silicon	28-1	7596054	7886784	7966270	7816369	2.49	cps
Silver	107-1	3805942	3844210	3882368	3844173	0.99	cps
Silver	109-1	3695076	3686834	3714923	3698944	0.39	cps
Sodium	23-2	23411419	22977620	23178342	23189127	0.94	cps
Strontium	86-1	947562	959925	945067	950851	0.84	cps
Strontium	88-1	8238169	8394914	8329828	8320970	0.95	cps
Sulfur	34-1	388692	395037	394950	392893	0.93	cps
Terbium	159-1	7849593	7503270	7596592	7649818	2.34	cps
Terbium	159-2	5761197	5715681	5631447	5702775	1.15	cps
Thallium	203-1	2657422	2662470	2609184	2643026	1.11	cps
Thallium	205-1	6415352	6278223	6365249	6352941	1.09	cps
Tin	118-1	2330124	2401000	2352363	2361162	1.54	cps
Titanium	47-1	3815740	3882131	3786950	3828274	1.28	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S6 Instrumnet Name : P7
Client Sample ID : S6 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:00:51 DataFile Name : 010CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8412147	8556245	8578942	8515778	1.06	cps
Vanadium	51-2	1310170	1318498	1333934	1320867	0.91	cps
Yttrium	89-1	11256286	10420972	10707610	10794956	3.93	cps
Yttrium	89-2	3441776	3278169	3220069	3313338	3.47	cps
Zinc	66-2	3345335	3400405	3405411	3383717	0.99	cps
Zirconium	90-1	5028809	5092386	5044285	5055160	0.66	cps
Zirconium	91-1	1094733	1122549	1120223	1112502	1.39	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3519300	3500602	3530970	3516957	0.44	cps
Antimony	121-1	5875188	5825586	5827571	5842782	0.48	cps
Arsenic	75-2	397855	390051	391857	393254	1.04	cps
Barium	135-1	6982209	6944622	6967727	6964853	0.27	cps
Barium	137-1	11895014	11923481	11945848	11921448	0.21	cps
Beryllium	9-1	783053	780359	771410	778274	0.78	cps
Bismuth	209-1	4849737	4786593	4902558	4846296	1.20	cps
Bismuth	209-2	4246751	4147374	4246980	4213702	1.36	cps
Bromine	81-1	9086	9126	9430	9214	2.04	cps
Bromine	81-2	130	100	130	120	14.43	cps
Cadmium	108-1	111986	114301	111079	112455	1.48	cps
Cadmium	106-1	157953	156785	156448	157062	0.50	cps
Cadmium	111-1	1565657	1562902	1564894	1564484	0.09	cps
Calcium	43-1	1576384	1594630	1575904	1582306	0.67	cps
Calcium	44-1	25618786	25816085	25629580	25688150	0.43	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3065085	3178969	3111402	3118485	1.84	cps
Cobalt	59-2	4754336	4711978	4765197	4743837	0.59	cps
Copper	63-2	36752200	36593745	35858279	36401408	1.31	cps
Dysprosium	156-1	163	107	173	148	24.33	cps
Dysprosium	156-2	390	417	390	399	3.86	cps
Erbium	164-1	200	193	203	199	2.56	cps
Erbium	164-2	183	183	173	180	3.21	cps
Gadolinium	160-1	107	170	123	133	24.62	cps
Gadolinium	160-2	390	390	560	447	21.97	cps
Holmium	165-1	7820443	7727797	7691368	7746536	0.86	cps
Holmium	165-2	5857221	5655554	5786879	5766551	1.78	cps
Indium	115-1	6421766	6416259	6317867	6385297	0.92	cps
Indium	115-2	2551571	2501178	2498503	2517084	1.19	cps
Iron	56-2	146897894	148834341	143291821	146341352	1.92	cps
Iron	57-2	3623976	3630754	3635553	3630094	0.16	cps
Iron	54-2	8162368	8341379	8261176	8254974	1.09	cps
Krypton	83-1	557	493	410	487	15.12	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	22196086	22114327	22031315	22113909	0.37	cps
Lead	207-1	19053662	19383992	19211575	19216410	0.86	cps
Lead	208-1	88108311	88366843	88059047	88178067	0.19	cps
Lithium	6-1	799164	779799	778640	785868	1.47	cps
Magnesium	24-2	30717146	30285662	30419398	30474069	0.72	cps
Manganese	55-2	26334221	26705112	26721843	26587058	0.82	cps
Molybdenum	94-1	20894246	20806848	20480548	20727214	1.05	cps
Molybdenum	95-1	30719483	30823950	29647986	30397140	2.14	cps
Molybdenum	96-1	33047397	33305968	32204842	32852736	1.75	cps
Molybdenum	97-1	18765120	18488249	18262518	18505296	1.36	cps
Molybdenum	98-1	48245028	48542191	47803813	48197010	0.77	cps
Neodymium	150-1	227	223	207	219	4.90	cps
Neodymium	150-2	73	87	77	79	8.80	cps
Nickel	60-2	1214436	1215942	1204214	1211531	0.53	cps
Phosphorus	31-2	193385	191684	191858	192309	0.49	cps
Potassium	39-2	17995023	17504691	17740996	17746903	1.38	cps
Rhodium	103-1	6358343	6212632	6280127	6283701	1.16	cps
Rhodium	103-2	4034964	4014647	3937765	3995792	1.28	cps
Scandium	45-1	3892192	3829171	3804759	3842041	1.17	cps
Scandium	45-2	429479	427116	425444	427346	0.47	cps
Selenium	82-1	114107	114928	114364	114466	0.37	cps
Selenium	77-2	14180	14247	14591	14339	1.54	cps
Selenium	78-2	47717	47911	47400	47676	0.54	cps
Silicon	28-1	14193727	14580408	14259201	14344445	1.44	cps
Silver	107-1	7657503	7688587	7644496	7663528	0.30	cps
Silver	109-1	7425153	7439253	7374481	7412962	0.46	cps
Sodium	23-2	47346789	46517839	46100524	46655051	1.36	cps
Strontium	86-1	1947208	1952560	1937191	1945653	0.40	cps
Strontium	88-1	17181089	17277092	16699663	17052615	1.81	cps
Sulfur	34-1	629206	618925	619729	622620	0.92	cps
Terbium	159-1	8162212	8126892	7826047	8038384	2.30	cps
Terbium	159-2	5871850	5727417	5712264	5770510	1.53	cps
Thallium	203-1	5451357	5386723	5380851	5406310	0.72	cps
Thallium	205-1	13045786	12882547	12565288	12831207	1.90	cps
Tin	118-1	4779500	4796628	4811802	4795977	0.34	cps
Titanium	47-1	7803423	7780144	7946746	7843438	1.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S7 Instrumnet Name : P7
Client Sample ID : S7 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:03:39 DataFile Name : 011CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	17712627	17542337	17226160	17493708	1.41	cps
Vanadium	51-2	2669858	2692422	2670752	2677677	0.48	cps
Yttrium	89-1	11329482	11116432	11169136	11205017	0.99	cps
Yttrium	89-2	3309599	3245011	3185639	3246750	1.91	cps
Zinc	66-2	6720863	6633381	6524445	6626229	1.49	cps
Zirconium	90-1	10419536	10576807	10329298	10441880	1.20	cps
Zirconium	91-1	2296064	2302380	2295035	2297826	0.17	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	18292051	18187302	17863275	18114209	1.23	cps
Antimony	121-1	4768	4854	4697	4773	1.64	cps
Arsenic	75-2	460	510	473	481	5.38	cps
Barium	135-1	4297	4414	4204	4305	2.44	cps
Barium	137-1	7772	7669	7539	7660	1.53	cps
Beryllium	9-1	151	124	151	142	10.90	cps
Bismuth	209-1	4656477	4480699	4484882	4540686	2.21	cps
Bismuth	209-2	4018131	3965331	3989889	3991117	0.66	cps
Bromine	81-1	9243	9523	9773	9513	2.79	cps
Bromine	81-2	207	227	220	218	4.68	cps
Cadmium	108-1	87	103	110	100	12.02	cps
Cadmium	106-1	3147	3064	3184	3131	1.96	cps
Cadmium	111-1	3032	2966	2992	2997	1.12	cps
Calcium	43-1	7801748	7695967	7654984	7717566	0.98	cps
Calcium	44-1	127010825	124854821	124093511	125319719	1.21	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	17320	17277	17410	17336	0.39	cps
Cobalt	59-2	9366	9173	9573	9371	2.14	cps
Copper	63-2	8052	7992	7986	8010	0.46	cps
Dysprosium	156-1	293	287	273	284	3.58	cps
Dysprosium	156-2	267	260	273	267	2.50	cps
Erbium	164-1	607	623	700	643	7.74	cps
Erbium	164-2	430	533	463	476	11.09	cps
Gadolinium	160-1	373	327	413	371	11.69	cps
Gadolinium	160-2	670	567	543	593	11.36	cps
Holmium	165-1	7865508	7770651	7669445	7768535	1.26	cps
Holmium	165-2	5879430	5908882	5824509	5870940	0.73	cps
Indium	115-1	6334966	6254594	6286860	6292140	0.64	cps
Indium	115-2	2626502	2616873	2574939	2606105	1.05	cps
Iron	56-2	721827909	710043883	733477669	721783154	1.62	cps
Iron	57-2	18027999	18053009	18125723	18068911	0.28	cps
Iron	54-2	39454824	39654236	39844991	39651350	0.49	cps
Krypton	83-1	420	533	553	502	14.32	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	6872	6825	6805	6834	0.50	cps
Lead	207-1	5961	5965	5915	5947	0.47	cps
Lead	208-1	27251	27451	27201	27301	0.49	cps
Lithium	6-1	827057	792682	776410	798716	3.24	cps
Magnesium	24-2	157415881	151322864	154339181	154359309	1.97	cps
Manganese	55-2	14805	14654	14588	14682	0.75	cps
Molybdenum	94-1	7675	7102	7522	7433	4.00	cps
Molybdenum	95-1	7082	6732	6742	6852	2.91	cps
Molybdenum	96-1	12299	11908	11928	12045	1.83	cps
Molybdenum	97-1	4524	4267	4147	4313	4.46	cps
Molybdenum	98-1	10994	11031	10784	10936	1.22	cps
Neodymium	150-1	127	140	137	134	5.16	cps
Neodymium	150-2	70	60	103	78	29.17	cps
Nickel	60-2	5224	5314	5584	5374	3.49	cps
Phosphorus	31-2	910	973	867	917	5.85	cps
Potassium	39-2	89286229	89088122	90219699	89531350	0.67	cps
Rhodium	103-1	6056779	6130679	5904612	6030690	1.91	cps
Rhodium	103-2	3896060	3915904	3884375	3898780	0.41	cps
Scandium	45-1	3940232	3916815	3853163	3903403	1.15	cps
Scandium	45-2	443916	450871	443540	446109	0.93	cps
Selenium	82-1	389	279	160	276	41.39	cps
Selenium	77-2	23	37	13	24	47.91	cps
Selenium	78-2	530	617	607	584	8.11	cps
Silicon	28-1	630561	632407	635339	632769	0.38	cps
Silver	107-1	3300	1730	1703	2245	40.74	cps
Silver	109-1	2867	1870	1577	2105	32.13	cps
Sodium	23-2	230385177	232471597	229283750	230713508	0.70	cps
Strontium	86-1	25145	25739	24971	25285	1.59	cps
Strontium	88-1	212739	212875	212566	212727	0.07	cps
Sulfur	34-1	163036	160539	159731	161102	1.07	cps
Terbium	159-1	8157574	7967006	7801449	7975343	2.23	cps
Terbium	159-2	5915168	5802769	5849467	5855802	0.96	cps
Thallium	203-1	7409	7272	7032	7238	2.64	cps
Thallium	205-1	17338	17098	17041	17159	0.92	cps
Tin	118-1	2680	2357	2354	2464	7.62	cps
Titanium	47-1	1123	1140	1110	1124	1.34	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : S8 Instrumnet Name : P7
Client Sample ID : S8 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:06:29 DataFile Name : 012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	1607	1480	1493	1527	4.56	cps
Vanadium	51-2	1297	1413	1250	1320	6.37	cps
Yttrium	89-1	11370889	11203061	10983847	11185932	1.74	cps
Yttrium	89-2	3355131	3408080	3336644	3366618	1.10	cps
Zinc	66-2	8412	8139	8122	8225	1.98	cps
Zirconium	90-1	9376	9506	9653	9512	1.46	cps
Zirconium	91-1	2154	2077	2057	2096	2.44	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	267263	267297	268528	267696	0.27	cps
Antimony	121-1	2248468	2239020	2238000	2241829	0.26	cps
Arsenic	75-2	153423	156040	155451	154971	0.89	cps
Barium	135-1	2105717	2098093	2142404	2115405	1.12	cps
Barium	137-1	3705360	3606694	3713356	3675136	1.62	cps
Beryllium	9-1	29128	28840	29144	29038	0.59	cps
Bismuth	209-1	4683720	4758476	4650212	4697469	1.18	cps
Bismuth	209-2	4238052	4299280	4266148	4267827	0.72	cps
Bromine	81-1	9923	10727	10264	10305	3.92	cps
Bromine	81-2	223	243	183	217	14.10	cps
Cadmium	108-1	20074	19887	19663	19875	1.03	cps
Cadmium	106-1	30385	31057	30004	30482	1.75	cps
Cadmium	111-1	305557	310466	307492	307838	0.80	cps
Calcium	43-1	65461	65950	64681	65364	0.98	cps
Calcium	44-1	1073449	1075152	1070270	1072957	0.23	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	500544	499357	501664	500522	0.23	cps
Cobalt	59-2	1898907	1933251	1868019	1900059	1.72	cps
Copper	63-2	677331	682610	677539	679160	0.44	cps
Dysprosium	156-1	37	17	10	21	65.74	cps
Dysprosium	156-2	110	160	123	131	19.75	cps
Erbium	164-1	47	47	70	54	24.74	cps
Erbium	164-2	27	43	50	40	30.04	cps
Gadolinium	160-1	60	27	57	48	38.42	cps
Gadolinium	160-2	403	467	413	428	7.96	cps
Holmium	165-1	7512195	7462745	7285745	7420228	1.60	cps
Holmium	165-2	5781674	5714515	5675189	5723793	0.94	cps
Indium	115-1	6516964	6500667	6422868	6480167	0.78	cps
Indium	115-2	2624440	2641242	2643850	2636511	0.40	cps
Iron	56-2	2339140	2327931	2340781	2335951	0.30	cps
Iron	57-2	62473	62557	62342	62457	0.17	cps
Iron	54-2	152067	151819	150279	151388	0.64	cps
Krypton	83-1	433	503	543	493	11.29	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1556930	1551102	1555930	1554654	0.20	cps
Lead	207-1	1328565	1320712	1317311	1322196	0.44	cps
Lead	208-1	6325662	6329842	6237185	6297563	0.83	cps
Lithium	6-1	746698	746728	738972	744133	0.60	cps
Magnesium	24-2	1078378	1088392	1124470	1097080	2.21	cps
Manganese	55-2	983460	987346	978820	983209	0.43	cps
Molybdenum	94-1	850	853	877	860	1.69	cps
Molybdenum	95-1	1103	940	1123	1056	9.53	cps
Molybdenum	96-1	1397	1193	1360	1317	8.23	cps
Molybdenum	97-1	717	710	693	707	1.70	cps
Molybdenum	98-1	1790	1747	1660	1732	3.82	cps
Neodymium	150-1	57	60	83	67	21.79	cps
Neodymium	150-2	13	3	17	11	62.48	cps
Nickel	60-2	487984	496956	492087	492342	0.91	cps
Phosphorus	31-2	903	840	953	899	6.32	cps
Potassium	39-2	1468123	1476393	1495533	1480017	0.95	cps
Rhodium	103-1	6493685	6665784	6365080	6508183	2.32	cps
Rhodium	103-2	4128654	4139479	4165288	4144474	0.45	cps
Scandium	45-1	3987487	3891001	3890538	3923009	1.42	cps
Scandium	45-2	432061	440033	437306	436467	0.93	cps
Selenium	82-1	46215	46405	46219	46280	0.23	cps
Selenium	77-2	5851	5818	5791	5820	0.52	cps
Selenium	78-2	19323	19753	19630	19569	1.13	cps
Silicon	28-1	597723	600648	593610	597327	0.59	cps
Silver	107-1	1504637	1509393	1501481	1505170	0.26	cps
Silver	109-1	1435096	1453763	1439322	1442727	0.68	cps
Sodium	23-2	1840865	1893661	1867141	1867222	1.41	cps
Strontium	86-1	1027	1043	957	1009	4.56	cps
Strontium	88-1	2047	1997	2094	2046	2.36	cps
Sulfur	34-1	173350	172717	172326	172798	0.30	cps
Terbium	159-1	7695378	7660903	7646126	7667469	0.33	cps
Terbium	159-2	5643165	5778171	5626978	5682771	1.46	cps
Thallium	203-1	1939575	1959040	1946393	1948336	0.51	cps
Thallium	205-1	4600335	4550734	4667642	4606237	1.27	cps
Tin	118-1	3737	3997	3627	3787	5.02	cps
Titanium	47-1	270	243	273	262	6.27	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICV01 Instrumnet Name : P7
Client Sample ID : ICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:22:05 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	67	67	57	63	9.12	cps
Vanadium	51-2	1045844	1056648	1047736	1050076	0.55	cps
Yttrium	89-1	11388447	11321119	11072416	11260661	1.48	cps
Yttrium	89-2	3270931	3353077	3356582	3326863	1.46	cps
Zinc	66-2	252009	256122	249801	252644	1.27	cps
Zirconium	90-1	1452	563	577	864	58.96	cps
Zirconium	91-1	257	293	250	267	8.75	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5188	5584	5251	5341	3.99	cps
Antimony	121-1	11762	11718	11912	11797	0.86	cps
Arsenic	75-2	367	483	397	416	14.58	cps
Barium	135-1	13233	13410	13386	13343	0.72	cps
Barium	137-1	22959	23406	22421	22929	2.15	cps
Beryllium	9-1	826	832	789	816	2.81	cps
Bismuth	209-1	4714834	4652751	4639994	4669193	0.86	cps
Bismuth	209-2	4142899	4357949	4282277	4261042	2.56	cps
Bromine	81-1	10140	9953	10410	10168	2.26	cps
Bromine	81-2	237	250	240	242	2.86	cps
Cadmium	108-1	90	107	97	98	8.58	cps
Cadmium	106-1	3167	3014	2904	3028	4.37	cps
Cadmium	111-1	3554	3482	3402	3479	2.19	cps
Calcium	43-1	9056	9073	8786	8972	1.79	cps
Calcium	44-1	171701	169564	167794	169686	1.15	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	8579	9196	9113	8963	3.74	cps
Cobalt	59-2	5111	5204	5218	5178	1.12	cps
Copper	63-2	8833	8549	8683	8688	1.63	cps
Dysprosium	156-1	3	7	0	3	100.05	cps
Dysprosium	156-2	3	23	7	11	96.43	cps
Erbium	164-1	57	33	20	37	50.62	cps
Erbium	164-2	43	53	47	48	10.66	cps
Gadolinium	160-1	57	50	40	49	17.16	cps
Gadolinium	160-2	363	390	357	370	4.77	cps
Holmium	165-1	7633283	7366324	7251680	7417096	2.64	cps
Holmium	165-2	5520321	5661025	5632201	5604516	1.33	cps
Indium	115-1	6578003	6404758	6408677	6463813	1.53	cps
Indium	115-2	2524764	2597425	2649823	2590671	2.42	cps
Iron	56-2	178886	180523	179856	179755	0.46	cps
Iron	57-2	5995	5938	5861	5931	1.13	cps
Iron	54-2	12228	11525	11895	11883	2.96	cps
Krypton	83-1	560	473	450	494	11.72	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	4121	4377	4301	4266	3.09	cps
Lead	207-1	3637	3594	3964	3732	5.42	cps
Lead	208-1	16904	16871	17374	17050	1.65	cps
Lithium	6-1	750352	734937	739509	741600	1.07	cps
Magnesium	24-2	153271	151261	151380	151971	0.74	cps
Manganese	55-2	3903	4048	3968	3973	1.83	cps
Molybdenum	94-1	12889	12399	12919	12736	2.29	cps
Molybdenum	95-1	15538	15278	15371	15396	0.86	cps
Molybdenum	96-1	17280	16997	16680	16986	1.77	cps
Molybdenum	97-1	9680	9376	9336	9464	1.99	cps
Molybdenum	98-1	24865	24594	24304	24587	1.14	cps
Neodymium	150-1	17	0	7	8	107.84	cps
Neodymium	150-2	10	0	7	6	91.64	cps
Nickel	60-2	1903	1810	1797	1837	3.16	cps
Phosphorus	31-2	3257	3234	3380	3290	2.40	cps
Potassium	39-2	272025	271321	268268	270538	0.74	cps
Rhodium	103-1	6476473	6517213	6510300	6501329	0.34	cps
Rhodium	103-2	4053386	4152112	4046070	4083856	1.45	cps
Scandium	45-1	4004136	3912316	3860044	3925499	1.86	cps
Scandium	45-2	415354	440029	437172	430852	3.13	cps
Selenium	82-1	698	784	774	752	6.21	cps
Selenium	77-2	73	97	97	89	15.16	cps
Selenium	78-2	743	747	790	760	3.43	cps
Silicon	28-1	586081	583971	587992	586015	0.34	cps
Silver	107-1	8519	8302	8403	8408	1.29	cps
Silver	109-1	8092	8146	7999	8079	0.92	cps
Sodium	23-2	274277	273920	269549	272582	0.97	cps
Strontium	86-1	2790	2817	2764	2790	0.96	cps
Strontium	88-1	17510	17120	16606	17079	2.66	cps
Sulfur	34-1	164346	165181	164377	164635	0.29	cps
Terbium	159-1	7657580	7551039	7583204	7597274	0.72	cps
Terbium	159-2	5539102	5658806	5791634	5663181	2.23	cps
Thallium	203-1	6018	5728	6005	5917	2.77	cps
Thallium	205-1	13737	13136	13887	13587	2.92	cps
Tin	118-1	25503	25035	25646	25395	1.26	cps
Titanium	47-1	4134	4161	4167	4154	0.42	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : LLICV01 Instrumnet Name : P7
Client Sample ID : LLICV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:40:20 DataFile Name : 016LLIC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	15216	15095	15362	15224	0.88	cps
Vanadium	51-2	13516	13696	13329	13514	1.36	cps
Yttrium	89-1	11345426	11256710	11017609	11206582	1.51	cps
Yttrium	89-2	3195176	3407216	3299394	3300595	3.21	cps
Zinc	66-2	7725	7349	7622	7565	2.57	cps
Zirconium	90-1	10917	11048	10404	10790	3.16	cps
Zirconium	91-1	2630	2424	2377	2477	5.44	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1673	1540	1643	1619	4.32	cps
Antimony	121-1	157	143	183	161	12.64	cps
Arsenic	75-2	7	3	10	7	50.03	cps
Barium	135-1	127	133	137	132	3.85	cps
Barium	137-1	233	223	220	226	3.08	cps
Beryllium	9-1	3	3	3	3	0.52	cps
Bismuth	209-1	4821349	4826469	4671524	4773114	1.84	cps
Bismuth	209-2	4336659	4264941	4253869	4285156	1.05	cps
Bromine	81-1	10350	10080	10144	10191	1.39	cps
Bromine	81-2	247	243	267	252	5.00	cps
Cadmium	108-1	13	17	3	11	62.48	cps
Cadmium	106-1	2894	2857	2810	2854	1.46	cps
Cadmium	111-1	2043	2012	1992	2015	1.28	cps
Calcium	43-1	547	603	573	574	4.94	cps
Calcium	44-1	34309	34148	34075	34177	0.35	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3241	2847	2824	2971	7.90	cps
Cobalt	59-2	130	143	107	127	14.65	cps
Copper	63-2	1600	1463	1457	1507	5.37	cps
Dysprosium	156-1	7	7	7	7	0.00	cps
Dysprosium	156-2	0	10	0	3	173.21	cps
Erbium	164-1	40	70	47	52	30.16	cps
Erbium	164-2	27	37	27	30	19.24	cps
Gadolinium	160-1	50	37	53	47	18.89	cps
Gadolinium	160-2	453	417	417	429	4.94	cps
Holmium	165-1	7409349	7427536	7531696	7456194	0.89	cps
Holmium	165-2	5917346	5560086	5614301	5697244	3.38	cps
Indium	115-1	6515037	6555763	6388797	6486532	1.34	cps
Indium	115-2	2703414	2691847	2702263	2699175	0.24	cps
Iron	56-2	27686	24734	24897	25772	6.44	cps
Iron	57-2	1930	2014	2000	1981	2.26	cps
Iron	54-2	3047	3080	3040	3056	0.70	cps
Krypton	83-1	493	483	480	486	1.43	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	820	857	820	832	2.54	cps
Lead	207-1	667	743	650	687	7.25	cps
Lead	208-1	3180	3200	3197	3192	0.34	cps
Lithium	6-1	760301	747536	738002	748613	1.49	cps
Magnesium	24-2	1207	1247	1200	1218	2.07	cps
Manganese	55-2	1046	1103	1152	1100	4.83	cps
Molybdenum	94-1	300	297	260	286	7.77	cps
Molybdenum	95-1	270	303	253	276	9.24	cps
Molybdenum	96-1	323	367	330	340	6.86	cps
Molybdenum	97-1	253	230	240	241	4.86	cps
Molybdenum	98-1	410	377	410	399	4.82	cps
Neodymium	150-1	17	17	3	12	63.01	cps
Neodymium	150-2	7	7	3	6	34.70	cps
Nickel	60-2	563	570	563	566	0.68	cps
Phosphorus	31-2	857	837	763	819	6.00	cps
Potassium	39-2	83497	83947	84761	84068	0.76	cps
Rhodium	103-1	6413522	6597341	6402118	6470994	1.69	cps
Rhodium	103-2	4250452	4098442	4201501	4183465	1.85	cps
Scandium	45-1	3999596	4007247	3975253	3994032	0.42	cps
Scandium	45-2	449300	431485	435875	438887	2.11	cps
Selenium	82-1	136	131	119	129	7.03	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	523	470	513	502	5.65	cps
Silicon	28-1	564617	567643	569283	567181	0.42	cps
Silver	107-1	217	220	157	198	18.02	cps
Silver	109-1	227	193	167	196	15.37	cps
Sodium	23-2	43050	42866	43304	43073	0.51	cps
Strontium	86-1	857	887	897	880	2.37	cps
Strontium	88-1	697	870	847	804	11.69	cps
Sulfur	34-1	160449	161106	160553	160703	0.22	cps
Terbium	159-1	7784475	7832957	7789056	7802163	0.34	cps
Terbium	159-2	5988624	5735604	5612652	5778960	3.32	cps
Thallium	203-1	623	483	523	543	13.27	cps
Thallium	205-1	1413	1327	1277	1339	5.17	cps
Tin	118-1	1953	2060	1980	1998	2.78	cps
Titanium	47-1	97	97	137	110	20.99	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICB01 Instrumnet Name : P7
Client Sample ID : ICB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:43:39 DataFile Name : 017CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	13	13	3	10	57.75	cps
Vanadium	51-2	43	40	30	38	18.36	cps
Yttrium	89-1	11428387	11055820	11291923	11258710	1.67	cps
Yttrium	89-2	3401803	3353957	3355261	3370341	0.81	cps
Zinc	66-2	4261	3974	4147	4127	3.50	cps
Zirconium	90-1	457	460	420	446	4.98	cps
Zirconium	91-1	70	70	137	92	41.74	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	16122159	15624884	16315903	16020982	2.22	cps
Antimony	121-1	6562	6548	6732	6614	1.55	cps
Arsenic	75-2	120	157	123	133	15.21	cps
Barium	135-1	1920	1943	2020	1961	2.67	cps
Barium	137-1	3447	3414	3610	3490	3.02	cps
Beryllium	9-1	244	187	191	207	15.31	cps
Bismuth	209-1	4803535	4591086	4655998	4683539	2.32	cps
Bismuth	209-2	4250039	4231829	4190350	4224073	0.72	cps
Bromine	81-1	10847	11348	11341	11179	2.57	cps
Bromine	81-2	353	313	297	321	9.07	cps
Cadmium	108-1	927	820	893	880	6.20	cps
Cadmium	106-1	2894	2900	2960	2918	1.26	cps
Cadmium	111-1	2394	2546	2437	2459	3.20	cps
Calcium	43-1	1621195	1580891	1565489	1589192	1.81	cps
Calcium	44-1	26208317	25622380	25241508	25690735	1.90	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	66643	67166	67343	67051	0.54	cps
Cobalt	59-2	6742	6291	6338	6457	3.83	cps
Copper	63-2	29406	29112	29850	29456	1.26	cps
Dysprosium	156-1	30	33	30	31	6.18	cps
Dysprosium	156-2	27	7	23	19	56.72	cps
Erbium	164-1	63	70	107	80	29.17	cps
Erbium	164-2	40	67	47	51	27.15	cps
Gadolinium	160-1	80	77	50	69	23.87	cps
Gadolinium	160-2	347	400	380	376	7.17	cps
Holmium	165-1	8057701	7635265	7731686	7808217	2.84	cps
Holmium	165-2	5933454	5919691	5877047	5910064	0.50	cps
Indium	115-1	6736028	6546526	6574108	6618887	1.55	cps
Indium	115-2	2693254	2686238	2680071	2686521	0.25	cps
Iron	56-2	294530509	295775356	293345302	294550389	0.41	cps
Iron	57-2	7253227	7352699	7317170	7307699	0.69	cps
Iron	54-2	16329406	16154183	16303985	16262524	0.58	cps
Krypton	83-1	457	437	503	466	7.35	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	19321	19541	19411	19424	0.57	cps
Lead	207-1	15289	15716	15466	15490	1.39	cps
Lead	208-1	72225	73206	72281	72571	0.76	cps
Lithium	6-1	751837	711908	702452	722065	3.63	cps
Magnesium	24-2	28065929	28084331	27732069	27960776	0.71	cps
Manganese	55-2	22265	21862	21379	21835	2.03	cps
Molybdenum	94-1	3410035	3381255	3387997	3393096	0.44	cps
Molybdenum	95-1	6179227	6010885	6144906	6111673	1.46	cps
Molybdenum	96-1	6591765	6472302	6487680	6517249	1.00	cps
Molybdenum	97-1	3815954	3853751	3682131	3783945	2.38	cps
Molybdenum	98-1	9976999	9809931	9585771	9790901	2.00	cps
Neodymium	150-1	17	20	63	33	78.09	cps
Neodymium	150-2	17	17	17	17	0.00	cps
Nickel	60-2	7092	7439	7295	7275	2.40	cps
Phosphorus	31-2	973406	968415	962874	968232	0.54	cps
Potassium	39-2	36357027	35946249	36198788	36167355	0.57	cps
Rhodium	103-1	6567444	6665784	6531354	6588194	1.06	cps
Rhodium	103-2	4170850	4207923	4149691	4176154	0.71	cps
Scandium	45-1	3921636	3954134	3824547	3900105	1.73	cps
Scandium	45-2	444999	440198	442479	442559	0.54	cps
Selenium	82-1	187	198	104	163	31.45	cps
Selenium	77-2	3	0	3	2	86.60	cps
Selenium	78-2	477	517	610	534	12.80	cps
Silicon	28-1	596630	597016	592540	595395	0.42	cps
Silver	107-1	550	450	497	499	10.03	cps
Silver	109-1	580	387	447	471	21.01	cps
Sodium	23-2	46325776	45451699	46257218	46011564	1.06	cps
Strontium	86-1	66285	65649	64183	65372	1.65	cps
Strontium	88-1	576119	570356	554033	566836	2.02	cps
Sulfur	34-1	2554966	2563494	2473507	2530656	1.96	cps
Terbium	159-1	7932644	8094370	8313141	8113385	2.35	cps
Terbium	159-2	6004471	5928875	5853035	5928794	1.28	cps
Thallium	203-1	630	667	747	681	8.76	cps
Thallium	205-1	1383	1580	1740	1568	11.40	cps
Tin	118-1	3664	4958	5925	4849	23.39	cps
Titanium	47-1	1599844	1613079	1528836	1580586	2.87	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSA01 Instrumnet Name : P7
Client Sample ID : ICSA01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:46:58 DataFile Name : 018ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	260	243	313	272	13.43	cps
Vanadium	51-2	653	650	567	623	7.88	cps
Yttrium	89-1	11528073	11738679	11496774	11587842	1.14	cps
Yttrium	89-2	3409498	3428669	3430413	3422860	0.34	cps
Zinc	66-2	11084	11201	11158	11148	0.53	cps
Zirconium	90-1	643	673	617	644	4.40	cps
Zirconium	91-1	113	143	117	124	13.21	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	16590129	16241004	16313124	16381419	1.13	cps
Antimony	121-1	124353	126317	126757	125809	1.02	cps
Arsenic	75-2	8239	8543	8596	8459	2.28	cps
Barium	135-1	31605	30773	30740	31039	1.58	cps
Barium	137-1	53296	55371	52701	53789	2.61	cps
Beryllium	9-1	14875	14414	13574	14288	4.62	cps
Bismuth	209-1	4606778	4651149	4796167	4684698	2.11	cps
Bismuth	209-2	4284795	4298365	4329395	4304185	0.53	cps
Bromine	81-1	12265	12268	11818	12117	2.14	cps
Bromine	81-2	367	427	380	391	8.06	cps
Cadmium	108-1	3097	3024	2727	2949	6.64	cps
Cadmium	106-1	5835	5835	5621	5763	2.14	cps
Cadmium	111-1	35134	35620	35371	35375	0.69	cps
Calcium	43-1	1606709	1632131	1555500	1598113	2.44	cps
Calcium	44-1	26165250	25968025	26014855	26049377	0.40	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	136402	133318	135530	135083	1.18	cps
Cobalt	59-2	109520	108272	109530	109107	0.66	cps
Copper	63-2	103237	102659	102318	102738	0.45	cps
Dysprosium	156-1	53	43	47	48	10.66	cps
Dysprosium	156-2	20	27	40	29	35.25	cps
Erbium	164-1	90	127	93	103	19.62	cps
Erbium	164-2	70	97	77	81	17.11	cps
Gadolinium	160-1	87	80	120	96	22.43	cps
Gadolinium	160-2	467	467	473	469	0.82	cps
Holmium	165-1	7945153	7895799	7789542	7876831	1.01	cps
Holmium	165-2	6015635	5976498	6081497	6024544	0.88	cps
Indium	115-1	6883977	6712545	6687094	6761205	1.58	cps
Indium	115-2	2673480	2730024	2731934	2711813	1.22	cps
Iron	56-2	304507915	305000009	297755776	302421233	1.34	cps
Iron	57-2	7540402	7348359	7448766	7445843	1.29	cps
Iron	54-2	16686759	16351130	16384203	16474031	1.12	cps
Krypton	83-1	523	533	493	517	4.03	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	104718	103267	104996	104327	0.89	cps
Lead	207-1	88649	89065	86025	87913	1.87	cps
Lead	208-1	400948	401105	412082	404712	1.58	cps
Lithium	6-1	673473	665636	666030	668380	0.66	cps
Magnesium	24-2	29271479	28362483	28147951	28593971	2.09	cps
Manganese	55-2	80761	80542	79800	80368	0.63	cps
Molybdenum	94-1	3502885	3522050	3496993	3507309	0.37	cps
Molybdenum	95-1	6334129	6250228	6189946	6258101	1.16	cps
Molybdenum	96-1	6704975	6590304	6487499	6594259	1.65	cps
Molybdenum	97-1	3815839	3933940	3742100	3830626	2.53	cps
Molybdenum	98-1	10166993	10171841	9811394	10050076	2.06	cps
Neodymium	150-1	37	50	57	48	21.31	cps
Neodymium	150-2	30	37	33	33	10.01	cps
Nickel	60-2	35539	33845	33888	34424	2.81	cps
Phosphorus	31-2	988453	978553	971928	979645	0.85	cps
Potassium	39-2	37490008	37008125	37276906	37258346	0.65	cps
Rhodium	103-1	6744386	6728034	6261828	6578082	4.17	cps
Rhodium	103-2	4243258	4242030	4248873	4244720	0.09	cps
Scandium	45-1	3931137	3866393	4048869	3948799	2.34	cps
Scandium	45-2	448455	452440	447512	449469	0.58	cps
Selenium	82-1	2583	2509	2607	2566	1.99	cps
Selenium	77-2	273	303	270	282	6.50	cps
Selenium	78-2	1567	1447	1573	1529	4.66	cps
Silicon	28-1	821058	805744	788320	805041	2.03	cps
Silver	107-1	156910	158062	162176	159049	1.74	cps
Silver	109-1	150528	154373	153064	152655	1.28	cps
Sodium	23-2	48320613	46225724	46263908	46936748	2.55	cps
Strontium	86-1	68241	68800	68435	68492	0.41	cps
Strontium	88-1	589367	590406	576224	585332	1.35	cps
Sulfur	34-1	2526331	2454072	2398736	2459713	2.60	cps
Terbium	159-1	8227293	8480688	8110087	8272689	2.29	cps
Terbium	159-2	6028055	6086564	5953173	6022597	1.11	cps
Thallium	203-1	108094	109709	103582	107129	2.96	cps
Thallium	205-1	247286	251932	261590	253603	2.88	cps
Tin	118-1	4581	5511	5705	5265	11.41	cps
Titanium	47-1	1597385	1587868	1605022	1596758	0.54	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : ICSAB01 Instrumnet Name : P7
Client Sample ID : ICSAB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:50:08 DataFile Name : 019ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	297	317	313	309	3.47	cps
Vanadium	51-2	57560	56860	56395	56938	1.03	cps
Yttrium	89-1	12012666	11881386	11436701	11776918	2.56	cps
Yttrium	89-2	3502338	3456187	3447465	3468663	0.85	cps
Zinc	66-2	23489	23782	24290	23854	1.70	cps
Zirconium	90-1	1480	1487	1377	1448	4.26	cps
Zirconium	91-1	347	367	303	339	9.55	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9029469	8772865	8817164	8873166	1.55	cps
Antimony	121-1	2961339	3006387	2987843	2985189	0.76	cps
Arsenic	75-2	207244	205710	203178	205377	1.00	cps
Barium	135-1	3567485	3612369	3528037	3569297	1.18	cps
Barium	137-1	6104110	6254716	6029692	6129506	1.87	cps
Beryllium	9-1	345439	344547	346538	345508	0.29	cps
Bismuth	209-1	4833046	4692294	4686092	4737144	1.75	cps
Bismuth	209-2	4292184	4137634	4157301	4195706	2.01	cps
Bromine	81-1	10924	11061	10611	10865	2.12	cps
Bromine	81-2	203	183	220	202	9.08	cps
Cadmium	108-1	56162	57279	56835	56759	0.99	cps
Cadmium	106-1	81726	79902	83234	81621	2.04	cps
Cadmium	111-1	786234	791210	792448	789964	0.42	cps
Calcium	43-1	4019116	4027721	4060343	4035727	0.54	cps
Calcium	44-1	63838972	65134117	66107402	65026831	1.75	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1623571	1641823	1673599	1646331	1.54	cps
Cobalt	59-2	2462937	2476085	2465290	2468104	0.28	cps
Copper	63-2	19078234	18612142	18708131	18799502	1.31	cps
Dysprosium	156-1	177	200	187	188	6.23	cps
Dysprosium	156-2	310	320	400	343	14.37	cps
Erbium	164-1	310	360	367	346	8.96	cps
Erbium	164-2	363	287	327	326	11.78	cps
Gadolinium	160-1	190	217	207	204	6.59	cps
Gadolinium	160-2	533	503	580	539	7.17	cps
Holmium	165-1	8047159	8059760	7995689	8034203	0.42	cps
Holmium	165-2	6094365	5940488	6131881	6055578	1.67	cps
Indium	115-1	6798732	6743982	6682312	6741675	0.86	cps
Indium	115-2	2763737	2701943	2757286	2740988	1.24	cps
Iron	56-2	384972834	378256821	377553714	380261123	1.08	cps
Iron	57-2	9504151	9415260	9736001	9551804	1.73	cps
Iron	54-2	21170365	21266831	21106577	21181257	0.38	cps
Krypton	83-1	497	590	463	517	12.71	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10781293	10886887	10634360	10767513	1.18	cps
Lead	207-1	9395026	9224436	9200803	9273422	1.14	cps
Lead	208-1	42858745	42293686	41829923	42327451	1.22	cps
Lithium	6-1	678927	679363	672638	676976	0.56	cps
Magnesium	24-2	77553662	76879152	76177932	76870249	0.89	cps
Manganese	55-2	13927747	14031675	14092965	14017463	0.60	cps
Molybdenum	94-1	10843952	10901151	11066349	10937151	1.06	cps
Molybdenum	95-1	16190566	15954091	16069969	16071542	0.74	cps
Molybdenum	96-1	17464141	17075684	17048047	17195957	1.35	cps
Molybdenum	97-1	9616135	9675218	9746348	9679234	0.67	cps
Molybdenum	98-1	24895554	25029359	25590169	25171694	1.46	cps
Neodymium	150-1	193	193	127	171	22.49	cps
Neodymium	150-2	63	67	67	66	2.94	cps
Nickel	60-2	634628	625575	625335	628513	0.84	cps
Phosphorus	31-2	100794	99110	98667	99523	1.13	cps
Potassium	39-2	47477951	45982408	45612463	46357607	2.13	cps
Rhodium	103-1	6602675	6559953	6496475	6553034	0.82	cps
Rhodium	103-2	4339598	4151042	4188290	4226310	2.36	cps
Scandium	45-1	3997367	4061579	4088711	4049219	1.16	cps
Scandium	45-2	472913	459182	459149	463748	1.71	cps
Selenium	82-1	57462	58124	58411	57999	0.84	cps
Selenium	77-2	7745	7289	7395	7476	3.20	cps
Selenium	78-2	25288	24798	24437	24841	1.72	cps
Silicon	28-1	7709656	7852401	7864213	7808757	1.10	cps
Silver	107-1	3876288	3949196	3919219	3914901	0.94	cps
Silver	109-1	3764006	3797122	3762731	3774620	0.52	cps
Sodium	23-2	117769542	118938158	115023502	117243734	1.71	cps
Strontium	86-1	1017887	1013672	1032606	1021388	0.97	cps
Strontium	88-1	9092738	8998328	9186225	9092430	1.03	cps
Sulfur	34-1	372503	376068	376676	375082	0.60	cps
Terbium	159-1	8413636	8392399	8280432	8362155	0.86	cps
Terbium	159-2	6191809	6095765	6007739	6098438	1.51	cps
Thallium	203-1	2689521	2660471	2693976	2681323	0.68	cps
Thallium	205-1	6327865	6340249	6211208	6293108	1.13	cps
Tin	118-1	2487103	2503008	2491230	2493780	0.33	cps
Titanium	47-1	3896710	4043802	4080512	4007008	2.43	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV01 Instrumnet Name : P7
Client Sample ID : CCV01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:53:13 DataFile Name : 020CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8272756	8301683	8271852	8282097	0.20	cps
Vanadium	51-2	1418947	1411766	1408653	1413122	0.37	cps
Yttrium	89-1	12098659	11886750	12177838	12054416	1.25	cps
Yttrium	89-2	3657142	3619941	3504490	3593857	2.21	cps
Zinc	66-2	3376165	3358324	3379313	3371267	0.34	cps
Zirconium	90-1	5464654	5358296	5467917	5430289	1.15	cps
Zirconium	91-1	1180597	1187115	1194956	1187556	0.61	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1750	1757	1600	1702	5.20	cps
Antimony	121-1	643	650	650	648	0.59	cps
Arsenic	75-2	17	20	13	17	20.01	cps
Barium	135-1	150	167	147	154	6.94	cps
Barium	137-1	240	283	233	252	10.76	cps
Beryllium	9-1	0	4	7	4	91.64	cps
Bismuth	209-1	4643927	4703116	4609067	4652037	1.02	cps
Bismuth	209-2	4260945	4286772	4222900	4256873	0.75	cps
Bromine	81-1	11204	10978	10934	11039	1.31	cps
Bromine	81-2	233	207	223	221	6.09	cps
Cadmium	108-1	13	17	40	23	62.27	cps
Cadmium	106-1	2687	2784	2604	2691	3.35	cps
Cadmium	111-1	1922	1987	1851	1920	3.53	cps
Calcium	43-1	557	610	517	561	8.35	cps
Calcium	44-1	37620	36922	36464	37002	1.57	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2734	2860	2757	2784	2.42	cps
Cobalt	59-2	117	137	120	124	8.61	cps
Copper	63-2	2044	2344	2150	2179	6.98	cps
Dysprosium	156-1	7	3	3	4	43.40	cps
Dysprosium	156-2	3	3	3	3	0.00	cps
Erbium	164-1	27	40	67	44	45.82	cps
Erbium	164-2	57	53	27	46	36.09	cps
Gadolinium	160-1	47	50	33	43	20.36	cps
Gadolinium	160-2	400	413	390	401	2.92	cps
Holmium	165-1	7505160	7517799	7462364	7495108	0.39	cps
Holmium	165-2	5640303	5742691	5574917	5652637	1.50	cps
Indium	115-1	6742864	6561763	6411338	6571989	2.53	cps
Indium	115-2	2708335	2666815	2637906	2671019	1.33	cps
Iron	56-2	30344	29546	41919	33936	20.41	cps
Iron	57-2	2000	1820	2244	2021	10.52	cps
Iron	54-2	3594	3330	3350	3425	4.29	cps
Krypton	83-1	410	470	520	467	11.80	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	857	850	850	852	0.45	cps
Lead	207-1	793	677	710	727	8.27	cps
Lead	208-1	3514	3457	3324	3431	2.84	cps
Lithium	6-1	669626	656719	655907	660750	1.16	cps
Magnesium	24-2	1867	1753	1753	1791	3.65	cps
Manganese	55-2	1152	1286	1202	1214	5.57	cps
Molybdenum	94-1	730	663	503	632	18.43	cps
Molybdenum	95-1	807	727	797	777	5.61	cps
Molybdenum	96-1	817	930	820	856	7.54	cps
Molybdenum	97-1	560	520	487	522	7.03	cps
Molybdenum	98-1	1203	1180	1067	1150	6.36	cps
Neodymium	150-1	13	7	3	8	65.47	cps
Neodymium	150-2	10	3	3	6	69.34	cps
Nickel	60-2	607	577	560	581	4.07	cps
Phosphorus	31-2	757	733	803	764	4.66	cps
Potassium	39-2	91288	90701	90916	90968	0.33	cps
Rhodium	103-1	6547022	6636519	6498651	6560730	1.07	cps
Rhodium	103-2	4188588	4180658	4167622	4178956	0.25	cps
Scandium	45-1	3930563	3942418	3827181	3900054	1.63	cps
Scandium	45-2	434470	442528	428613	435204	1.61	cps
Selenium	82-1	226	156	100	160	39.35	cps
Selenium	77-2	3	3	3	3	0.00	cps
Selenium	78-2	560	490	587	546	9.15	cps
Silicon	28-1	565914	562999	561137	563350	0.43	cps
Silver	107-1	487	400	340	409	18.03	cps
Silver	109-1	467	363	270	367	26.83	cps
Sodium	23-2	62332	61129	60692	61385	1.38	cps
Strontium	86-1	847	850	913	870	4.32	cps
Strontium	88-1	760	963	827	850	12.19	cps
Sulfur	34-1	151488	152569	153442	152500	0.64	cps
Terbium	159-1	7830722	7526197	7674930	7677283	1.98	cps
Terbium	159-2	5763694	5720469	5623298	5702487	1.26	cps
Thallium	203-1	840	800	820	820	2.44	cps
Thallium	205-1	1987	1793	1793	1858	6.01	cps
Tin	118-1	2694	2607	2631	2644	1.69	cps
Titanium	47-1	113	147	127	129	13.02	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB01 Instrumnet Name : P7
Client Sample ID : CCB01 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 13:56:14 DataFile Name : 021CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	100	87	70	86	17.57	cps
Vanadium	51-2	33	33	37	34	5.60	cps
Yttrium	89-1	11503653	11245201	11238468	11329107	1.33	cps
Yttrium	89-2	3377139	3398657	3372965	3382920	0.41	cps
Zinc	66-2	4431	4347	4404	4394	0.97	cps
Zirconium	90-1	637	630	660	642	2.45	cps
Zirconium	91-1	97	117	163	126	27.25	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5191	4958	5258	5135	3.07	cps
Antimony	121-1	12255	12219	11992	12155	1.17	cps
Arsenic	75-2	403	447	460	437	6.79	cps
Barium	135-1	14518	14484	14297	14433	0.82	cps
Barium	137-1	24939	24541	24999	24826	1.00	cps
Beryllium	9-1	703	826	828	786	9.09	cps
Bismuth	209-1	4762909	4648318	4682908	4698045	1.25	cps
Bismuth	209-2	4294785	4298864	4230809	4274819	0.89	cps
Bromine	81-1	10794	10554	11168	10839	2.85	cps
Bromine	81-2	197	253	233	228	12.62	cps
Cadmium	108-1	153	147	127	142	9.76	cps
Cadmium	106-1	2804	2884	2800	2829	1.67	cps
Cadmium	111-1	3615	3806	3605	3675	3.08	cps
Calcium	43-1	9737	9280	9410	9475	2.48	cps
Calcium	44-1	176082	175962	173532	175192	0.82	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	9760	9500	9843	9701	1.85	cps
Cobalt	59-2	5771	5588	5745	5701	1.74	cps
Copper	63-2	9486	9797	9536	9606	1.73	cps
Dysprosium	156-1	13	13	13	13	0.00	cps
Dysprosium	156-2	7	0	7	4	86.60	cps
Erbium	164-1	53	33	70	52	35.16	cps
Erbium	164-2	33	53	60	49	28.39	cps
Gadolinium	160-1	37	77	30	48	52.83	cps
Gadolinium	160-2	367	423	323	371	13.51	cps
Holmium	165-1	7436896	7500490	7255130	7397505	1.72	cps
Holmium	165-2	5669660	5760724	5717146	5715843	0.80	cps
Indium	115-1	6676223	6411197	6387751	6491724	2.47	cps
Indium	115-2	2696679	2734083	2722779	2717847	0.71	cps
Iron	56-2	191988	188562	203996	194849	4.16	cps
Iron	57-2	6355	6268	6185	6269	1.36	cps
Iron	54-2	12565	12205	12445	12405	1.48	cps
Krypton	83-1	467	477	517	487	5.44	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	4978	5281	4878	5045	4.16	cps
Lead	207-1	4467	4407	4001	4292	5.92	cps
Lead	208-1	20303	20363	19559	20075	2.23	cps
Lithium	6-1	682173	661332	670712	671406	1.55	cps
Magnesium	24-2	161376	158359	159770	159835	0.94	cps
Manganese	55-2	4073	4042	4005	4040	0.83	cps
Molybdenum	94-1	13236	13470	13593	13433	1.35	cps
Molybdenum	95-1	16553	16726	16403	16561	0.98	cps
Molybdenum	96-1	18632	18589	17958	18393	2.05	cps
Molybdenum	97-1	10287	10310	10307	10301	0.12	cps
Molybdenum	98-1	26835	27215	26431	26827	1.46	cps
Neodymium	150-1	7	3	0	3	100.05	cps
Neodymium	150-2	0	7	3	3	100.05	cps
Nickel	60-2	2013	2144	2010	2056	3.70	cps
Phosphorus	31-2	1050	1073	1140	1088	4.29	cps
Potassium	39-2	287154	285780	286738	286557	0.25	cps
Rhodium	103-1	6544057	6493475	6415522	6484351	1.00	cps
Rhodium	103-2	4125286	4230889	4142848	4166341	1.36	cps
Scandium	45-1	3889796	3878907	3792272	3853658	1.39	cps
Scandium	45-2	431425	432346	436172	433314	0.58	cps
Selenium	82-1	849	821	788	820	3.73	cps
Selenium	77-2	113	103	103	107	5.42	cps
Selenium	78-2	870	747	773	797	8.15	cps
Silicon	28-1	564051	568281	567968	566767	0.42	cps
Silver	107-1	8222	8833	8683	8579	3.71	cps
Silver	109-1	8136	7999	7946	8027	1.22	cps
Sodium	23-2	294426	290035	293237	292566	0.78	cps
Strontium	86-1	3087	2964	2900	2984	3.18	cps
Strontium	88-1	19303	18929	18682	18971	1.65	cps
Sulfur	34-1	160277	162011	160738	161009	0.56	cps
Terbium	159-1	7781621	7736668	7643816	7720702	0.91	cps
Terbium	159-2	5688736	5790057	5792366	5757053	1.03	cps
Thallium	203-1	5828	5708	5771	5769	1.04	cps
Thallium	205-1	14084	14188	14198	14156	0.44	cps
Tin	118-1	28234	27717	27393	27781	1.53	cps
Titanium	47-1	4144	4291	4304	4246	2.09	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CRI Instrumnet Name : P7
Client Sample ID : CRI Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:03:24 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16664	16477	15966	16369	2.21	cps
Vanadium	51-2	14717	15291	14827	14945	2.04	cps
Yttrium	89-1	11462213	11041409	10961922	11155181	2.41	cps
Yttrium	89-2	3400434	3337671	3379910	3372672	0.95	cps
Zinc	66-2	7889	8049	7769	7902	1.78	cps
Zirconium	90-1	10791	11458	10917	11055	3.20	cps
Zirconium	91-1	2624	2460	2427	2504	4.20	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1563	1430	1593	1529	5.69	cps
Antimony	121-1	67	93	93	84	18.23	cps
Arsenic	75-2	3	7	3	4	43.40	cps
Barium	135-1	137	140	107	128	14.37	cps
Barium	137-1	223	163	200	196	15.47	cps
Beryllium	9-1	7	7	0	4	85.44	cps
Bismuth	209-1	4712669	4919249	4775917	4802612	2.20	cps
Bismuth	209-2	4322611	4021419	4159694	4167908	3.62	cps
Bromine	81-1	11074	10974	10554	10867	2.54	cps
Bromine	81-2	230	257	260	249	6.61	cps
Cadmium	108-1	13	13	17	14	13.35	cps
Cadmium	106-1	2797	2827	2597	2740	4.56	cps
Cadmium	111-1	1992	1983	1833	1936	4.61	cps
Calcium	43-1	503	590	450	514	13.74	cps
Calcium	44-1	29926	30487	29576	29996	1.53	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2497	2604	2664	2588	3.26	cps
Cobalt	59-2	113	113	117	114	1.68	cps
Copper	63-2	2920	4409	2820	3383	26.29	cps
Dysprosium	156-1	13	17	7	12	41.65	cps
Dysprosium	156-2	10	3	0	4	114.60	cps
Erbium	164-1	43	83	80	69	32.22	cps
Erbium	164-2	37	67	37	47	37.11	cps
Gadolinium	160-1	47	67	57	57	17.65	cps
Gadolinium	160-2	453	420	453	442	4.35	cps
Holmium	165-1	7459199	7613934	7231930	7435021	2.58	cps
Holmium	165-2	5678506	5472282	5539422	5563403	1.89	cps
Indium	115-1	6449106	6689028	6405105	6514413	2.35	cps
Indium	115-2	2664101	2584811	2616243	2621718	1.52	cps
Iron	56-2	25579	25278	42240	31032	31.28	cps
Iron	57-2	1880	1843	1773	1832	2.96	cps
Iron	54-2	3020	3157	3167	3115	2.63	cps
Krypton	83-1	607	490	500	532	12.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	877	957	827	887	7.40	cps
Lead	207-1	740	857	713	770	9.90	cps
Lead	208-1	3607	3694	3264	3521	6.46	cps
Lithium	6-1	701815	701742	688253	697270	1.12	cps
Magnesium	24-2	1330	1243	1233	1269	4.19	cps
Manganese	55-2	1239	1086	1156	1160	6.62	cps
Molybdenum	94-1	353	257	393	334	21.01	cps
Molybdenum	95-1	320	317	257	298	11.97	cps
Molybdenum	96-1	313	343	360	339	6.98	cps
Molybdenum	97-1	260	227	220	236	9.10	cps
Molybdenum	98-1	497	440	447	461	6.72	cps
Neodymium	150-1	3	13	10	9	57.30	cps
Neodymium	150-2	7	7	3	6	34.70	cps
Nickel	60-2	603	730	673	669	9.49	cps
Phosphorus	31-2	830	853	883	856	3.13	cps
Potassium	39-2	84811	84677	84362	84617	0.27	cps
Rhodium	103-1	6501335	6600078	6312305	6471239	2.26	cps
Rhodium	103-2	4152463	4035227	4162955	4116882	1.72	cps
Scandium	45-1	3922659	3873596	3858755	3885003	0.86	cps
Scandium	45-2	433396	425273	429409	429359	0.95	cps
Selenium	82-1	70	148	126	115	35.22	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	460	577	520	519	11.24	cps
Silicon	28-1	542093	547332	546133	545186	0.50	cps
Silver	107-1	130	170	177	159	15.89	cps
Silver	109-1	117	140	163	140	16.67	cps
Sodium	23-2	46771	46620	46607	46666	0.20	cps
Strontium	86-1	867	867	747	827	8.38	cps
Strontium	88-1	797	837	840	824	2.92	cps
Sulfur	34-1	162790	164542	164639	163990	0.63	cps
Terbium	159-1	7750019	7762168	7615801	7709329	1.05	cps
Terbium	159-2	5750024	5508721	5608746	5622497	2.16	cps
Thallium	203-1	573	580	530	561	4.84	cps
Thallium	205-1	1333	1273	1233	1280	3.93	cps
Tin	118-1	2297	1930	1954	2060	9.97	cps
Titanium	47-1	87	73	63	74	15.73	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BL Instrumnet Name : P7
Client Sample ID : PB169251BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:10:08 DataFile Name : 024CCBD.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7	7	20	11	69.25	cps
Vanadium	51-2	23	50	10	28	73.33	cps
Yttrium	89-1	11264454	11441363	10982527	11229448	2.06	cps
Yttrium	89-2	3394745	3248548	3402146	3348480	2.59	cps
Zinc	66-2	4827	4467	4627	4641	3.89	cps
Zirconium	90-1	430	440	417	429	2.73	cps
Zirconium	91-1	80	110	63	84	28.01	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1763034	1753712	1757158	1757968	0.27	cps
Antimony	121-1	2943013	2954198	2908348	2935187	0.81	cps
Arsenic	75-2	201609	204523	202243	202791	0.76	cps
Barium	135-1	3524500	3569254	3490653	3528136	1.12	cps
Barium	137-1	6042423	6101954	6058306	6067561	0.51	cps
Beryllium	9-1	366960	370868	369957	369262	0.55	cps
Bismuth	209-1	4817749	4811880	4777145	4802258	0.46	cps
Bismuth	209-2	4284920	4322923	4368522	4325455	0.97	cps
Bromine	81-1	10507	10507	10160	10392	1.93	cps
Bromine	81-2	233	153	190	192	20.83	cps
Cadmium	108-1	57711	58869	57236	57939	1.45	cps
Cadmium	106-1	82658	83304	82704	82889	0.44	cps
Cadmium	111-1	806930	806140	801983	805017	0.33	cps
Calcium	43-1	834450	834308	833553	834104	0.06	cps
Calcium	44-1	13146539	13102814	12989860	13079738	0.62	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1581445	1641714	1608876	1610678	1.87	cps
Cobalt	59-2	2395582	2407036	2421846	2408155	0.55	cps
Copper	63-2	19196770	18742787	18886901	18942153	1.22	cps
Dysprosium	156-1	80	47	67	64	26.03	cps
Dysprosium	156-2	167	200	193	187	9.45	cps
Erbium	164-1	133	117	143	131	10.28	cps
Erbium	164-2	100	97	87	94	7.35	cps
Gadolinium	160-1	100	73	107	93	18.90	cps
Gadolinium	160-2	367	410	430	402	8.05	cps
Holmium	165-1	7885734	7910687	7578667	7791696	2.37	cps
Holmium	165-2	5937419	5786936	5823689	5849348	1.34	cps
Indium	115-1	6698627	6645129	6628769	6657508	0.55	cps
Indium	115-2	2722340	2645637	2682700	2683559	1.43	cps
Iron	56-2	74737014	75197794	75374524	75103111	0.44	cps
Iron	57-2	1880945	1869911	1865473	1872110	0.43	cps
Iron	54-2	4205908	4275248	4194072	4225076	1.04	cps
Krypton	83-1	583	493	463	513	12.17	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10732601	10805232	10842686	10793506	0.52	cps
Lead	207-1	9429279	9409676	9448965	9429307	0.21	cps
Lead	208-1	43558122	43037999	43410823	43335648	0.62	cps
Lithium	6-1	733093	736323	721388	730268	1.08	cps
Magnesium	24-2	14995517	15545733	15382394	15307881	1.85	cps
Manganese	55-2	13552739	13599338	13646313	13599463	0.34	cps
Molybdenum	94-1	10577054	10692693	10435199	10568315	1.22	cps
Molybdenum	95-1	15605214	15615786	15501613	15574204	0.41	cps
Molybdenum	96-1	16920378	16986535	17105275	17004063	0.55	cps
Molybdenum	97-1	9523391	9649682	9649255	9607442	0.76	cps
Molybdenum	98-1	24514335	24743342	24671920	24643199	0.48	cps
Neodymium	150-1	150	130	117	132	12.69	cps
Neodymium	150-2	53	33	47	44	22.92	cps
Nickel	60-2	623672	615154	612630	617152	0.94	cps
Phosphorus	31-2	93340	96295	96805	95480	1.96	cps
Potassium	39-2	9223728	9022852	8931113	9059231	1.65	cps
Rhodium	103-1	6779511	6544934	6471736	6598727	2.44	cps
Rhodium	103-2	4278032	4061773	4189658	4176488	2.60	cps
Scandium	45-1	3988791	3968358	3967994	3975048	0.30	cps
Scandium	45-2	447624	442236	440020	443293	0.88	cps
Selenium	82-1	59260	59884	59746	59630	0.55	cps
Selenium	77-2	7639	7272	7445	7452	2.46	cps
Selenium	78-2	25629	24921	24467	25006	2.34	cps
Silicon	28-1	7967630	7924007	7847255	7912964	0.77	cps
Silver	107-1	3962352	4035757	3978750	3992286	0.97	cps
Silver	109-1	3885882	3785831	3851195	3840969	1.32	cps
Sodium	23-2	23325410	23523931	23240270	23363204	0.62	cps
Strontium	86-1	992281	994719	980611	989204	0.76	cps
Strontium	88-1	8807932	8812947	8652813	8757897	1.04	cps
Sulfur	34-1	396342	393165	388431	392646	1.01	cps
Terbium	159-1	8343654	8121954	7935423	8133677	2.51	cps
Terbium	159-2	5826221	5892992	5976682	5898632	1.28	cps
Thallium	203-1	2691507	2671091	2659097	2673898	0.61	cps
Thallium	205-1	6342324	6346620	6346306	6345083	0.04	cps
Tin	118-1	2414794	2426097	2455574	2432155	0.87	cps
Titanium	47-1	4000731	3999952	3894018	3964901	1.55	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169251BS Instrumnet Name : P7
Client Sample ID : PB169251BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:13:27 DataFile Name : 025LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8257419	8460419	8443622	8387154	1.34	cps
Vanadium	51-2	1366117	1391646	1376218	1377994	0.93	cps
Yttrium	89-1	11600846	11568388	11206729	11458654	1.91	cps
Yttrium	89-2	3473940	3292189	3388501	3384877	2.69	cps
Zinc	66-2	3499035	3464369	3475724	3479710	0.51	cps
Zirconium	90-1	5415011	5332776	5191511	5313099	2.13	cps
Zirconium	91-1	1164447	1157000	1159442	1160296	0.33	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1540	1627	1660	1609	3.85	cps
Antimony	121-1	137	230	193	187	25.19	cps
Arsenic	75-2	3	10	10	8	49.52	cps
Barium	135-1	140	167	100	136	24.75	cps
Barium	137-1	237	237	197	223	10.34	cps
Beryllium	9-1	3	3	17	8	98.17	cps
Bismuth	209-1	4786760	4608464	4696020	4697081	1.90	cps
Bismuth	209-2	4340925	4207916	4255180	4268007	1.58	cps
Bromine	81-1	10587	10731	10347	10555	1.84	cps
Bromine	81-2	257	243	223	241	6.96	cps
Cadmium	108-1	10	17	40	22	70.88	cps
Cadmium	106-1	2844	3037	2887	2923	3.47	cps
Cadmium	111-1	2024	2138	2013	2058	3.36	cps
Calcium	43-1	510	570	480	520	8.81	cps
Calcium	44-1	35922	36364	36437	36241	0.77	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4150	2830	2474	3151	28.02	cps
Cobalt	59-2	117	100	123	113	10.61	cps
Copper	63-2	2850	2907	2824	2860	1.49	cps
Dysprosium	156-1	17	10	10	12	31.50	cps
Dysprosium	156-2	3	3	3	3	0.00	cps
Erbium	164-1	37	33	47	39	17.85	cps
Erbium	164-2	47	50	27	41	30.69	cps
Gadolinium	160-1	30	50	40	40	25.00	cps
Gadolinium	160-2	397	393	457	416	8.58	cps
Holmium	165-1	7520411	7516865	7465032	7500769	0.41	cps
Holmium	165-2	5654657	5684491	5744368	5694506	0.80	cps
Indium	115-1	6500525	6431272	6502888	6478228	0.63	cps
Indium	115-2	2622872	2626316	2665170	2638119	0.89	cps
Iron	56-2	24777	25074	24393	24748	1.38	cps
Iron	57-2	1727	1760	1950	1812	6.65	cps
Iron	54-2	3067	3110	3114	3097	0.84	cps
Krypton	83-1	483	487	490	487	0.68	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	937	957	830	908	7.50	cps
Lead	207-1	707	810	747	754	6.91	cps
Lead	208-1	3510	3704	3490	3568	3.30	cps
Lithium	6-1	716592	701357	696175	704708	1.51	cps
Magnesium	24-2	1287	1277	1253	1272	1.34	cps
Manganese	55-2	1039	1099	1103	1080	3.30	cps
Molybdenum	94-1	377	357	373	369	2.91	cps
Molybdenum	95-1	433	493	360	429	15.57	cps
Molybdenum	96-1	453	433	447	444	2.29	cps
Molybdenum	97-1	270	333	310	304	10.52	cps
Molybdenum	98-1	663	563	583	603	8.77	cps
Neodymium	150-1	3	13	13	10	57.75	cps
Neodymium	150-2	0	0	10	3	173.21	cps
Nickel	60-2	750	647	750	716	8.34	cps
Phosphorus	31-2	767	747	843	786	6.50	cps
Potassium	39-2	82958	84791	86021	84590	1.82	cps
Rhodium	103-1	6417803	6431944	6440226	6429991	0.18	cps
Rhodium	103-2	4171218	4137585	4189834	4166212	0.64	cps
Scandium	45-1	3911392	3819796	3821478	3850888	1.36	cps
Scandium	45-2	431618	433878	435364	433620	0.44	cps
Selenium	82-1	122	59	71	84	39.81	cps
Selenium	77-2	0	3	3	2	86.60	cps
Selenium	78-2	523	643	567	578	10.52	cps
Silicon	28-1	544835	548335	548111	547094	0.36	cps
Silver	107-1	237	200	127	188	29.83	cps
Silver	109-1	177	160	167	168	5.00	cps
Sodium	23-2	47640	46212	46313	46722	1.71	cps
Strontium	86-1	857	777	897	843	7.24	cps
Strontium	88-1	683	797	763	748	7.79	cps
Sulfur	34-1	155963	156599	156360	156308	0.21	cps
Terbium	159-1	7708240	7792173	7729965	7743459	0.56	cps
Terbium	159-2	5698820	5635071	5672363	5668751	0.56	cps
Thallium	203-1	583	437	517	512	14.34	cps
Thallium	205-1	1373	1317	1163	1285	8.46	cps
Tin	118-1	2804	2694	2527	2675	5.21	cps
Titanium	47-1	117	120	120	119	1.62	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BL Instrumnet Name : P7
Client Sample ID : PB169194BL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:19:44 DataFile Name : 027CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	13	30	17	20	44.10	cps
Vanadium	51-2	33	10	27	23	51.50	cps
Yttrium	89-1	11378079	11379089	11130223	11295797	1.27	cps
Yttrium	89-2	3336905	3324271	3349898	3337025	0.38	cps
Zinc	66-2	4587	4414	4948	4650	5.85	cps
Zirconium	90-1	443	437	450	443	1.50	cps
Zirconium	91-1	107	73	97	92	18.55	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1780750	1831880	1783808	1798813	1.59	cps
Antimony	121-1	2975784	2971612	2969572	2972323	0.11	cps
Arsenic	75-2	204674	205598	199729	203334	1.55	cps
Barium	135-1	3602530	3516724	3602106	3573787	1.38	cps
Barium	137-1	6217424	5976278	5992271	6061991	2.22	cps
Beryllium	9-1	375632	375023	378410	376355	0.48	cps
Bismuth	209-1	4899020	4842848	4707591	4816486	2.04	cps
Bismuth	209-2	4334827	4253140	4266231	4284733	1.02	cps
Bromine	81-1	10667	10444	11455	10855	4.89	cps
Bromine	81-2	397	413	490	433	11.49	cps
Cadmium	108-1	57243	57996	57741	57660	0.66	cps
Cadmium	106-1	83794	82527	82789	83036	0.81	cps
Cadmium	111-1	817470	814461	811012	814314	0.40	cps
Calcium	43-1	835794	849212	845250	843419	0.82	cps
Calcium	44-1	13185726	13246674	13377674	13270025	0.74	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1608455	1628583	1607108	1614715	0.74	cps
Cobalt	59-2	2419016	2469799	2441001	2443272	1.04	cps
Copper	63-2	19267807	18698913	19281485	19082735	1.74	cps
Dysprosium	156-1	67	90	77	78	15.05	cps
Dysprosium	156-2	257	203	183	214	17.68	cps
Erbium	164-1	120	120	120	120	0.00	cps
Erbium	164-2	110	87	53	83	34.18	cps
Gadolinium	160-1	80	80	90	83	6.93	cps
Gadolinium	160-2	350	440	417	402	11.61	cps
Holmium	165-1	7850791	7818613	7761862	7810422	0.58	cps
Holmium	165-2	5990109	5765795	5931967	5895957	1.97	cps
Indium	115-1	6622451	6655598	6556293	6611447	0.76	cps
Indium	115-2	2722962	2640893	2682002	2681952	1.53	cps
Iron	56-2	75575606	77929117	74933414	76146046	2.07	cps
Iron	57-2	1901302	1938043	1864678	1901341	1.93	cps
Iron	54-2	4192414	4300015	4217911	4236780	1.33	cps
Krypton	83-1	547	467	507	507	7.89	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11051074	11412517	11226417	11230003	1.61	cps
Lead	207-1	9666254	9818711	9668941	9717969	0.90	cps
Lead	208-1	44273811	44409056	44084174	44255680	0.37	cps
Lithium	6-1	755363	743123	742116	746867	0.99	cps
Magnesium	24-2	15851447	15951947	15403657	15735684	1.86	cps
Manganese	55-2	13724368	13736960	13604211	13688513	0.54	cps
Molybdenum	94-1	10569578	10763099	10624737	10652471	0.94	cps
Molybdenum	95-1	15775621	15528640	15620883	15641714	0.80	cps
Molybdenum	96-1	17070433	16890853	16915734	16959006	0.57	cps
Molybdenum	97-1	9513774	9602103	9554197	9556691	0.46	cps
Molybdenum	98-1	25349658	25207340	25056656	25204551	0.58	cps
Neodymium	150-1	103	127	200	143	35.19	cps
Neodymium	150-2	40	40	40	40	0.00	cps
Nickel	60-2	627833	625825	630134	627931	0.34	cps
Phosphorus	31-2	96553	98395	96473	97140	1.12	cps
Potassium	39-2	9048955	9139777	9338412	9175715	1.61	cps
Rhodium	103-1	6550917	6528424	6396714	6492018	1.28	cps
Rhodium	103-2	4191623	4089617	4150200	4143813	1.24	cps
Scandium	45-1	3946707	3911714	3872068	3910163	0.96	cps
Scandium	45-2	446180	431333	439828	439114	1.70	cps
Selenium	82-1	59377	59858	59567	59600	0.41	cps
Selenium	77-2	7289	7435	7492	7405	1.42	cps
Selenium	78-2	26387	24901	25525	25605	2.91	cps
Silicon	28-1	7989214	7895983	7920541	7935246	0.61	cps
Silver	107-1	4096375	4140466	3956769	4064537	2.36	cps
Silver	109-1	3939030	3901886	3781267	3874061	2.13	cps
Sodium	23-2	24638144	24064140	23636460	24112914	2.08	cps
Strontium	86-1	1002465	994877	1008783	1002042	0.69	cps
Strontium	88-1	8781310	8838446	8780035	8799930	0.38	cps
Sulfur	34-1	391922	391724	393067	392238	0.18	cps
Terbium	159-1	7947046	7919725	8063708	7976826	0.96	cps
Terbium	159-2	5972419	5943991	5824419	5913610	1.33	cps
Thallium	203-1	2731171	2788929	2700963	2740354	1.63	cps
Thallium	205-1	6296391	6494772	6389410	6393524	1.55	cps
Tin	118-1	2493325	2449725	2471381	2471477	0.88	cps
Titanium	47-1	3993279	4028076	4072393	4031249	0.98	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : PB169194BS Instrumnet Name : P7
Client Sample ID : PB169194BS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:23:36 DataFile Name : 028LCS6.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8325613	8542490	8393890	8420665	1.32	cps
Vanadium	51-2	1365415	1367090	1358545	1363683	0.33	cps
Yttrium	89-1	11692267	11505544	11331944	11509918	1.57	cps
Yttrium	89-2	3422165	3298335	3380480	3366993	1.87	cps
Zinc	66-2	3538243	3439665	3496403	3491437	1.42	cps
Zirconium	90-1	5293771	5294125	5321992	5303296	0.31	cps
Zirconium	91-1	1165191	1166351	1161615	1164386	0.21	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11367530	11327152	11351801	11348828	0.18	cps
Antimony	121-1	8940	9400	8853	9064	3.24	cps
Arsenic	75-2	19413	19586	20034	19678	1.63	cps
Barium	135-1	740984	746643	746185	744604	0.42	cps
Barium	137-1	1307362	1287789	1313486	1302879	1.03	cps
Beryllium	9-1	2615	2543	2498	2552	2.31	cps
Bismuth	209-1	4931799	5012653	4953839	4966097	0.84	cps
Bismuth	209-2	4416993	4420219	4378595	4405269	0.53	cps
Bromine	81-1	30392	38156	42020	36856	16.07	cps
Bromine	81-2	8382	8666	8969	8673	3.38	cps
Cadmium	108-1	633	700	570	634	10.25	cps
Cadmium	106-1	4277	4367	4701	4449	5.02	cps
Cadmium	111-1	9132	9389	9440	9320	1.77	cps
Calcium	43-1	317723	322175	315017	318305	1.14	cps
Calcium	44-1	4963298	5047372	5024665	5011779	0.87	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	470943	474483	471428	472285	0.41	cps
Cobalt	59-2	243365	241541	241484	242130	0.44	cps
Copper	63-2	1187839	1194583	1191247	1191223	0.28	cps
Dysprosium	156-1	98842	97459	98133	98145	0.70	cps
Dysprosium	156-2	93099	91886	89510	91498	2.00	cps
Erbium	164-1	95356	97344	95504	96068	1.15	cps
Erbium	164-2	73476	73885	72169	73177	1.22	cps
Gadolinium	160-1	101421	103802	102341	102521	1.17	cps
Gadolinium	160-2	81982	81124	80513	81207	0.91	cps
Holmium	165-1	8267146	8047483	8142112	8152247	1.35	cps
Holmium	165-2	6111255	6168787	6028959	6103000	1.15	cps
Indium	115-1	6848891	6835823	6889074	6857930	0.40	cps
Indium	115-2	2820578	2835973	2734482	2797011	1.96	cps
Iron	56-2	373906674	362171555	369445114	368507781	1.61	cps
Iron	57-2	9249683	9181692	9098521	9176632	0.83	cps
Iron	54-2	20589837	20540668	20389421	20506642	0.51	cps
Krypton	83-1	603	627	647	626	3.47	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8522837	8851609	8499274	8624574	2.28	cps
Lead	207-1	7320111	7670547	7169081	7386580	3.48	cps
Lead	208-1	33402450	34826978	33298170	33842532	2.52	cps
Lithium	6-1	773976	753275	747892	758381	1.82	cps
Magnesium	24-2	6111295	6192249	6192848	6165464	0.76	cps
Manganese	55-2	7792615	7612148	7679951	7694905	1.18	cps
Molybdenum	94-1	98452	101845	98720	99672	1.89	cps
Molybdenum	95-1	21326	21473	21329	21376	0.39	cps
Molybdenum	96-1	39861	38685	39991	39512	1.82	cps
Molybdenum	97-1	12866	13810	13213	13296	3.59	cps
Molybdenum	98-1	34391	33569	33876	33945	1.22	cps
Neodymium	150-1	153653	160266	154654	156191	2.28	cps
Neodymium	150-2	103366	101744	102050	102387	0.84	cps
Nickel	60-2	140009	137916	136243	138056	1.37	cps
Phosphorus	31-2	85615	85344	84647	85202	0.59	cps
Potassium	39-2	2635924	2659827	2667534	2654428	0.62	cps
Rhodium	103-1	6929012	6624606	6915169	6822929	2.52	cps
Rhodium	103-2	4332053	4409094	4227538	4322895	2.11	cps
Scandium	45-1	4398168	4229453	4388722	4338781	2.18	cps
Scandium	45-2	478808	480142	474131	477694	0.66	cps
Selenium	82-1	810	728	692	743	8.18	cps
Selenium	77-2	493	500	493	496	0.78	cps
Selenium	78-2	1080	1103	1160	1115	3.69	cps
Silicon	28-1	3846082	3651488	4009532	3835701	4.67	cps
Silver	107-1	8032	8516	8189	8246	2.99	cps
Silver	109-1	8289	8049	8082	8140	1.60	cps
Sodium	23-2	1511403	1534748	1558985	1535045	1.55	cps
Strontium	86-1	215609	215931	216588	216043	0.23	cps
Strontium	88-1	1894840	1915537	1862336	1890904	1.42	cps
Sulfur	34-1	218501	217642	218437	218193	0.22	cps
Terbium	159-1	8569015	8119794	8536644	8408484	2.98	cps
Terbium	159-2	6197743	6108890	6011813	6106149	1.52	cps
Thallium	203-1	4221	3914	3947	4027	4.18	cps
Thallium	205-1	9537	9583	9163	9428	2.44	cps
Tin	118-1	38676	37318	38659	38218	2.04	cps
Titanium	47-1	851604	833290	854144	846346	1.34	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01 Instrumnet Name : P7
Client Sample ID : 0804-SOIL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:26:19 DataFile Name : 029SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	102731	103033	101660	102475	0.70	cps
Vanadium	51-2	496316	496226	492066	494869	0.49	cps
Yttrium	89-1	13368651	13781988	12996084	13382241	2.94	cps
Yttrium	89-2	4043418	3975569	3920338	3979775	1.55	cps
Zinc	66-2	709478	710055	705396	708310	0.36	cps
Zirconium	90-1	239121	242909	243208	241746	0.94	cps
Zirconium	91-1	55000	54461	55321	54927	0.79	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11750250	11582381	11833906	11722179	1.09	cps
Antimony	121-1	10437	9927	10297	10220	2.58	cps
Arsenic	75-2	20718	20231	20541	20497	1.20	cps
Barium	135-1	808998	796784	816974	807585	1.26	cps
Barium	137-1	1401385	1422746	1426684	1416938	0.96	cps
Beryllium	9-1	2573	2626	2497	2565	2.52	cps
Bismuth	209-1	4914629	4656186	4975254	4848690	3.49	cps
Bismuth	209-2	4427815	4365421	4467546	4420261	1.16	cps
Bromine	81-1	38788	46209	51369	45455	13.91	cps
Bromine	81-2	10140	10177	10317	10211	0.91	cps
Cadmium	108-1	713	727	670	703	4.21	cps
Cadmium	106-1	4587	4571	4494	4551	1.09	cps
Cadmium	111-1	9597	9456	9619	9557	0.92	cps
Calcium	43-1	301822	301172	299433	300809	0.41	cps
Calcium	44-1	4662651	4716521	4854883	4744685	2.09	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	432440	432571	431345	432119	0.16	cps
Cobalt	59-2	224064	223528	227722	225105	1.01	cps
Copper	63-2	1207288	1198907	1199303	1201833	0.39	cps
Dysprosium	156-1	92386	92372	96221	93660	2.37	cps
Dysprosium	156-2	89369	89849	88709	89309	0.64	cps
Erbium	164-1	95303	93524	93753	94194	1.03	cps
Erbium	164-2	73214	73928	74488	73877	0.86	cps
Gadolinium	160-1	99085	99635	100954	99892	0.96	cps
Gadolinium	160-2	78392	78774	79924	79030	1.01	cps
Holmium	165-1	7977522	7876488	8186472	8013494	1.97	cps
Holmium	165-2	6166728	6127325	6072919	6122324	0.77	cps
Indium	115-1	6966833	6615589	6910077	6830833	2.76	cps
Indium	115-2	2923365	2795038	2873940	2864114	2.26	cps
Iron	56-2	332107275	333949868	334675382	333577508	0.40	cps
Iron	57-2	8058860	8204503	8060501	8107955	1.03	cps
Iron	54-2	18255387	17948266	18156122	18119925	0.86	cps
Krypton	83-1	627	553	653	611	8.47	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8981455	9318269	8895706	9065144	2.46	cps
Lead	207-1	7790997	7833246	7770996	7798413	0.41	cps
Lead	208-1	35334564	35915414	35640055	35630011	0.82	cps
Lithium	6-1	754678	718249	763063	745330	3.20	cps
Magnesium	24-2	6020659	5921930	5944559	5962383	0.87	cps
Manganese	55-2	7609198	7650077	7461172	7573482	1.31	cps
Molybdenum	94-1	104279	102651	101218	102716	1.49	cps
Molybdenum	95-1	17390	16983	16590	16988	2.36	cps
Molybdenum	96-1	35400	35771	36148	35773	1.05	cps
Molybdenum	97-1	10711	10474	10100	10428	2.95	cps
Molybdenum	98-1	27175	26778	27022	26992	0.74	cps
Neodymium	150-1	154313	152977	152837	153376	0.53	cps
Neodymium	150-2	103054	101788	102625	102489	0.63	cps
Nickel	60-2	135289	134814	133510	134538	0.68	cps
Phosphorus	31-2	87613	86232	87925	87257	1.03	cps
Potassium	39-2	2829875	2750689	2749690	2776751	1.66	cps
Rhodium	103-1	6985910	6871903	6928450	6928754	0.82	cps
Rhodium	103-2	4463588	4471308	4297946	4410947	2.22	cps
Scandium	45-1	4381337	4241169	4423897	4348801	2.20	cps
Scandium	45-2	491680	483286	483980	486315	0.96	cps
Selenium	82-1	776	836	750	787	5.55	cps
Selenium	77-2	457	540	507	501	8.37	cps
Selenium	78-2	1180	1060	1093	1111	5.58	cps
Silicon	28-1	12845195	12937386	13827070	13203217	4.11	cps
Silver	107-1	8549	8569	8332	8484	1.55	cps
Silver	109-1	8332	8639	8346	8439	2.06	cps
Sodium	23-2	1457795	1442743	1450549	1450362	0.52	cps
Strontium	86-1	210115	213153	211572	211613	0.72	cps
Strontium	88-1	1863582	1836775	1817385	1839247	1.26	cps
Sulfur	34-1	209494	206751	204011	206752	1.33	cps
Terbium	159-1	8505004	8493475	8499897	8499459	0.07	cps
Terbium	159-2	6203207	6173502	6114655	6163788	0.73	cps
Thallium	203-1	3964	3864	3974	3934	1.55	cps
Thallium	205-1	8790	9040	8639	8823	2.29	cps
Tin	118-1	26264	26221	27133	26539	1.94	cps
Titanium	47-1	949606	955503	954108	953072	0.32	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01DUP Instrumnet Name : P7
Client Sample ID : 0804-SOILDUP Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:29:26 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	100884	104071	100145	101700	2.05	cps
Vanadium	51-2	489176	487263	487022	487821	0.24	cps
Yttrium	89-1	13589054	13526633	13261775	13459154	1.29	cps
Yttrium	89-2	4048145	4019258	4076462	4047955	0.71	cps
Zinc	66-2	759547	750274	756392	755404	0.62	cps
Zirconium	90-1	254925	253183	256381	254829	0.63	cps
Zirconium	91-1	56190	57907	58336	57478	1.98	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2256938	2208703	2201690	2222444	1.35	cps
Antimony	121-1	1660	1710	1767	1712	3.12	cps
Arsenic	75-2	3847	4104	3794	3915	4.23	cps
Barium	135-1	144342	142989	142057	143129	0.80	cps
Barium	137-1	248902	248881	247336	248373	0.36	cps
Beryllium	9-1	513	511	479	501	3.80	cps
Bismuth	209-1	4628651	4508753	4645322	4594242	1.62	cps
Bismuth	209-2	4288167	4300571	4294473	4294404	0.14	cps
Bromine	81-1	22921	23345	23419	23228	1.16	cps
Bromine	81-2	2854	2960	2877	2897	1.94	cps
Cadmium	108-1	183	173	147	168	11.30	cps
Cadmium	106-1	3020	2960	3080	3020	1.99	cps
Cadmium	111-1	3289	3248	3317	3285	1.06	cps
Calcium	43-1	63442	63917	63620	63660	0.38	cps
Calcium	44-1	1041871	1025874	1037054	1034933	0.79	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	93827	94662	94632	94374	0.50	cps
Cobalt	59-2	48207	46629	46997	47278	1.75	cps
Copper	63-2	238023	237327	232478	235943	1.28	cps
Dysprosium	156-1	18376	17995	17885	18085	1.42	cps
Dysprosium	156-2	17741	17765	17925	17810	0.56	cps
Erbium	164-1	17274	17665	17878	17606	1.74	cps
Erbium	164-2	14137	13403	13877	13806	2.70	cps
Gadolinium	160-1	18789	19300	19330	19140	1.59	cps
Gadolinium	160-2	15949	16083	15996	16009	0.42	cps
Holmium	165-1	7590486	7499765	7388361	7492871	1.35	cps
Holmium	165-2	5843248	5740708	5663065	5749007	1.57	cps
Indium	115-1	6577469	6444330	6587232	6536344	1.22	cps
Indium	115-2	2739668	2696761	2720000	2718810	0.79	cps
Iron	56-2	73639171	72451794	71521842	72537602	1.46	cps
Iron	57-2	1788837	1775286	1749227	1771117	1.14	cps
Iron	54-2	3972951	4004092	3891242	3956095	1.47	cps
Krypton	83-1	570	513	630	571	10.22	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1538506	1513718	1518165	1523463	0.87	cps
Lead	207-1	1302027	1298902	1312372	1304434	0.54	cps
Lead	208-1	6142812	6221663	6188110	6184195	0.64	cps
Lithium	6-1	691707	699159	697326	696064	0.56	cps
Magnesium	24-2	1226793	1211094	1185090	1207659	1.74	cps
Manganese	55-2	1510904	1521397	1488250	1506850	1.12	cps
Molybdenum	94-1	19463	19567	19757	19595	0.76	cps
Molybdenum	95-1	4294	4141	4444	4293	3.53	cps
Molybdenum	96-1	7989	8249	7909	8049	2.21	cps
Molybdenum	97-1	2700	2807	2604	2704	3.76	cps
Molybdenum	98-1	7018	6652	6662	6777	3.08	cps
Neodymium	150-1	28659	28957	29718	29111	1.88	cps
Neodymium	150-2	19867	19530	19293	19564	1.47	cps
Nickel	60-2	27756	27849	27812	27806	0.17	cps
Phosphorus	31-2	17130	17667	16829	17209	2.47	cps
Potassium	39-2	604828	601433	607070	604444	0.47	cps
Rhodium	103-1	6525477	6592516	6658134	6592042	1.01	cps
Rhodium	103-2	4290437	4189062	4173412	4217637	1.51	cps
Scandium	45-1	4076869	3937738	3921860	3978822	2.14	cps
Scandium	45-2	450885	455528	445019	450477	1.17	cps
Selenium	82-1	198	282	131	203	37.25	cps
Selenium	77-2	107	83	97	96	12.25	cps
Selenium	78-2	577	663	637	626	7.10	cps
Silicon	28-1	1327859	1285548	1290584	1301330	1.78	cps
Silver	107-1	1657	1707	1710	1691	1.77	cps
Silver	109-1	1763	1750	1613	1709	4.86	cps
Sodium	23-2	330615	329323	328103	329347	0.38	cps
Strontium	86-1	43701	43060	43591	43451	0.79	cps
Strontium	88-1	363332	359885	362090	361769	0.48	cps
Sulfur	34-1	134475	133732	137398	135202	1.43	cps
Terbium	159-1	7747860	7882322	7853011	7827731	0.90	cps
Terbium	159-2	5809886	5744504	5789644	5781344	0.58	cps
Thallium	203-1	957	927	973	952	2.48	cps
Thallium	205-1	2344	2454	2410	2402	2.31	cps
Tin	118-1	9547	9183	9403	9378	1.95	cps
Titanium	47-1	168502	165435	166304	166747	0.95	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01LX5 Instrumnet Name : P7
Client Sample ID : 0804-SOILL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 14:32:33 DataFile Name : 031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18753	19144	18546	18815	1.61	cps
Vanadium	51-2	95038	96483	97248	96257	1.17	cps
Yttrium	89-1	11573751	11617278	11597748	11596259	0.19	cps
Yttrium	89-2	3523319	3513668	3501702	3512896	0.31	cps
Zinc	66-2	147718	142033	144358	144703	1.98	cps
Zirconium	90-1	47718	47564	47450	47577	0.28	cps
Zirconium	91-1	10764	10420	10541	10575	1.65	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13683584	13420948	13405763	13503432	1.16	cps
Antimony	121-1	2848745	2819727	2823280	2830584	0.56	cps
Arsenic	75-2	231641	226016	225221	227626	1.54	cps
Barium	135-1	3794997	3873561	3908845	3859135	1.51	cps
Barium	137-1	6719303	6639062	6614538	6657634	0.82	cps
Beryllium	9-1	352279	351066	346859	350068	0.81	cps
Bismuth	209-1	4824117	4857063	4649109	4776763	2.34	cps
Bismuth	209-2	4485922	4301049	4384131	4390368	2.11	cps
Bromine	81-1	39824	47587	53487	46966	14.59	cps
Bromine	81-2	10571	10591	11428	10863	4.50	cps
Cadmium	108-1	55225	55389	54492	55035	0.87	cps
Cadmium	106-1	81759	80573	80218	80850	1.00	cps
Cadmium	111-1	776088	785072	779812	780324	0.58	cps
Calcium	43-1	1122507	1137656	1131866	1130676	0.68	cps
Calcium	44-1	17527643	17629061	17731343	17629349	0.58	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2068501	1979849	2009122	2019157	2.24	cps
Cobalt	59-2	2684270	2640727	2620799	2648599	1.23	cps
Copper	63-2	18318408	18428423	18119636	18288823	0.86	cps
Dysprosium	156-1	97252	97074	98067	97464	0.54	cps
Dysprosium	156-2	93714	92140	91849	92568	1.08	cps
Erbium	164-1	97569	100600	100151	99440	1.65	cps
Erbium	164-2	77521	77343	76415	77093	0.77	cps
Gadolinium	160-1	102963	104615	103628	103735	0.80	cps
Gadolinium	160-2	83772	82713	82378	82954	0.88	cps
Holmium	165-1	7917131	8086505	8060160	8021265	1.14	cps
Holmium	165-2	6214042	5985134	5874416	6024531	2.88	cps
Indium	115-1	6976644	6775379	6695483	6815835	2.13	cps
Indium	115-2	2828228	2771969	2743933	2781377	1.54	cps
Iron	56-2	397606874	388952301	381214128	389257768	2.11	cps
Iron	57-2	9718014	9627390	9341166	9562190	2.06	cps
Iron	54-2	21846613	21397662	21344462	21529579	1.28	cps
Krypton	83-1	743	703	777	741	4.95	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	16874803	17634222	17486766	17331930	2.32	cps
Lead	207-1	14867395	15068061	15135769	15023742	0.93	cps
Lead	208-1	68057037	68471312	68877818	68468722	0.60	cps
Lithium	6-1	736542	735377	717670	729863	1.45	cps
Magnesium	24-2	22116675	21516789	21789880	21807781	1.38	cps
Manganese	55-2	19781335	19447817	19117882	19449011	1.71	cps
Molybdenum	94-1	4956130	5166807	4955395	5026111	2.42	cps
Molybdenum	95-1	5507712	5588072	5592792	5562859	0.86	cps
Molybdenum	96-1	6184304	6286157	6103065	6191175	1.48	cps
Molybdenum	97-1	3336428	3497545	3487337	3440437	2.62	cps
Molybdenum	98-1	8656544	9005799	8745090	8802478	2.06	cps
Neodymium	150-1	161970	161864	160872	161569	0.37	cps
Neodymium	150-2	107681	105565	105411	106219	1.19	cps
Nickel	60-2	772927	756408	752777	760704	1.41	cps
Phosphorus	31-2	89205	87932	88693	88610	0.72	cps
Potassium	39-2	10970037	10739307	11158631	10955992	1.92	cps
Rhodium	103-1	6721665	6902571	6885992	6836742	1.46	cps
Rhodium	103-2	4446517	4289700	4280787	4339001	2.15	cps
Scandium	45-1	4434394	4482112	4383019	4433175	1.12	cps
Scandium	45-2	484184	475114	480607	479968	0.95	cps
Selenium	82-1	58971	60369	58926	59422	1.38	cps
Selenium	77-2	7685	7836	7909	7810	1.46	cps
Selenium	78-2	25509	25085	25749	25448	1.32	cps
Silicon	28-1	7951911	7989228	7712922	7884687	1.90	cps
Silver	107-1	693135	710854	712836	705609	1.54	cps
Silver	109-1	666690	661944	658349	662328	0.63	cps
Sodium	23-2	23755264	23200320	23777767	23577784	1.39	cps
Strontium	86-1	1332487	1337870	1337967	1336108	0.23	cps
Strontium	88-1	11783632	11869235	11821711	11824859	0.36	cps
Sulfur	34-1	203771	203883	202120	203258	0.49	cps
Terbium	159-1	8439416	8259270	8213454	8304047	1.44	cps
Terbium	159-2	6331022	5953717	5946169	6076969	3.62	cps
Thallium	203-1	2486637	2551281	2523349	2520423	1.29	cps
Thallium	205-1	5838775	6043883	5976282	5952980	1.76	cps
Tin	118-1	2130890	2160552	2126398	2139280	0.87	cps
Titanium	47-1	891071	901784	910578	901144	1.08	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MS Instrumnet Name : P7
Client Sample ID : 0804-SOILMS Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:35:48 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7922411	7853990	7931107	7902503	0.53	cps
Vanadium	51-2	1875846	1834738	1860423	1857002	1.12	cps
Yttrium	89-1	13887543	13764398	13323282	13658408	2.17	cps
Yttrium	89-2	4097736	3945606	3944709	3996017	2.20	cps
Zinc	66-2	3863588	3819443	3796974	3826668	0.89	cps
Zirconium	90-1	5444694	5572764	5434634	5484031	1.40	cps
Zirconium	91-1	1181498	1212434	1197164	1197032	1.29	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	14104326	13951573	13947487	14001129	0.64	cps
Antimony	121-1	2803570	2778796	2811385	2797917	0.61	cps
Arsenic	75-2	230956	232464	228014	230478	0.98	cps
Barium	135-1	3870300	3945714	3881282	3899098	1.04	cps
Barium	137-1	6636132	6837794	6800041	6757989	1.59	cps
Beryllium	9-1	339575	340156	337629	339120	0.39	cps
Bismuth	209-1	4904993	4706967	4790756	4800905	2.07	cps
Bismuth	209-2	4414662	4363590	4344172	4374141	0.83	cps
Bromine	81-1	41977	49791	52014	47927	11.00	cps
Bromine	81-2	10641	10714	10801	10718	0.75	cps
Cadmium	108-1	54171	55376	53549	54365	1.71	cps
Cadmium	106-1	79507	79610	79533	79550	0.07	cps
Cadmium	111-1	760205	779640	766953	768932	1.28	cps
Calcium	43-1	1152206	1142939	1146149	1147098	0.41	cps
Calcium	44-1	18506216	18114268	18160952	18260479	1.17	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1991590	1997874	2016811	2002092	0.66	cps
Cobalt	59-2	2616800	2664358	2625851	2635670	0.96	cps
Copper	63-2	18488432	18114095	18220782	18274436	1.06	cps
Dysprosium	156-1	98282	101384	98755	99474	1.68	cps
Dysprosium	156-2	95274	94033	94204	94504	0.71	cps
Erbium	164-1	101171	102763	104112	102682	1.43	cps
Erbium	164-2	79022	80746	78835	79534	1.32	cps
Gadolinium	160-1	104148	104034	106327	104836	1.23	cps
Gadolinium	160-2	84651	83927	84795	84458	0.55	cps
Holmium	165-1	7973176	8324278	7861162	8052872	3.00	cps
Holmium	165-2	6208552	6070269	6072766	6117196	1.29	cps
Indium	115-1	6881056	6934802	6643079	6819646	2.28	cps
Indium	115-2	2775431	2777968	2822478	2791959	0.95	cps
Iron	56-2	425979247	420623314	424910327	423837629	0.67	cps
Iron	57-2	10545130	10428760	10507163	10493685	0.57	cps
Iron	54-2	23634946	23324770	23203218	23387645	0.95	cps
Krypton	83-1	663	680	677	673	1.31	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	18738895	19131859	18854182	18908312	1.07	cps
Lead	207-1	16319261	16331938	16575924	16409041	0.88	cps
Lead	208-1	75462869	74698624	75604725	75255406	0.65	cps
Lithium	6-1	736586	728001	690680	718422	3.40	cps
Magnesium	24-2	22780961	23246577	23039647	23022395	1.01	cps
Manganese	55-2	20223711	19841900	20257478	20107696	1.15	cps
Molybdenum	94-1	4926844	5012888	5037358	4992363	1.16	cps
Molybdenum	95-1	5481455	5619738	5576304	5559166	1.27	cps
Molybdenum	96-1	6148247	6181760	6045409	6125139	1.16	cps
Molybdenum	97-1	3369548	3385648	3352118	3369105	0.50	cps
Molybdenum	98-1	8597825	8728541	8784564	8703643	1.10	cps
Neodymium	150-1	160908	160110	161329	160782	0.39	cps
Neodymium	150-2	107517	105535	106107	106386	0.96	cps
Nickel	60-2	750709	751631	754354	752231	0.25	cps
Phosphorus	31-2	94199	93776	94192	94055	0.26	cps
Potassium	39-2	10910273	11002646	11062627	10991848	0.70	cps
Rhodium	103-1	6835497	6850042	6695014	6793517	1.26	cps
Rhodium	103-2	4379916	4309403	4384443	4357921	0.97	cps
Scandium	45-1	4534594	4326136	4342336	4401022	2.63	cps
Scandium	45-2	491781	482150	484071	486001	1.05	cps
Selenium	82-1	57975	59410	58929	58771	1.24	cps
Selenium	77-2	7762	7625	7725	7704	0.92	cps
Selenium	78-2	25806	24644	25589	25346	2.44	cps
Silicon	28-1	5575322	6763443	4938573	5759113	16.08	cps
Silver	107-1	678342	699362	686477	688061	1.54	cps
Silver	109-1	655580	662939	650747	656422	0.94	cps
Sodium	23-2	24155615	23818353	24053540	24009169	0.72	cps
Strontium	86-1	1351942	1351137	1358995	1354025	0.32	cps
Strontium	88-1	11953965	12013080	11835316	11934120	0.76	cps
Sulfur	34-1	214820	212595	209734	212383	1.20	cps
Terbium	159-1	8366374	8173534	8290477	8276795	1.17	cps
Terbium	159-2	6327042	6172281	6199827	6233050	1.32	cps
Thallium	203-1	2458305	2523060	2546689	2509352	1.82	cps
Thallium	205-1	5824673	5812707	5854534	5830638	0.37	cps
Tin	118-1	2159628	2152329	2169853	2160603	0.41	cps
Titanium	47-1	1030424	1013510	1035941	1026625	1.14	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01MSD Instrumnet Name : P7
Client Sample ID : 0804-SOILMSD Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:38:34 DataFile Name : 033SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7301303	7841346	7595029	7579226	3.57	cps
Vanadium	51-2	1909683	1878286	1891500	1893156	0.83	cps
Yttrium	89-1	13929097	13974564	13517504	13807055	1.82	cps
Yttrium	89-2	4237477	4038380	3983834	4086564	3.27	cps
Zinc	66-2	4142740	4037245	4108195	4096060	1.31	cps
Zirconium	90-1	5310453	5346838	5326412	5327901	0.34	cps
Zirconium	91-1	1160022	1182419	1168911	1170451	0.96	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8874374	8837891	8776651	8829639	0.56	cps
Antimony	121-1	2965802	2982742	2940572	2963039	0.72	cps
Arsenic	75-2	207733	207526	205891	207050	0.49	cps
Barium	135-1	3605515	3614856	3583257	3601209	0.45	cps
Barium	137-1	6085280	6231206	6196936	6171140	1.24	cps
Beryllium	9-1	351828	356473	353305	353869	0.67	cps
Bismuth	209-1	4792234	4621700	4597433	4670456	2.27	cps
Bismuth	209-2	4209081	4240070	4132784	4193979	1.32	cps
Bromine	81-1	18715	17657	17597	17990	3.50	cps
Bromine	81-2	1343	1327	1170	1280	7.47	cps
Cadmium	108-1	56507	58176	56553	57079	1.67	cps
Cadmium	106-1	81454	81035	80757	81082	0.43	cps
Cadmium	111-1	786206	805572	795553	795777	1.22	cps
Calcium	43-1	4171676	4198138	4150981	4173598	0.57	cps
Calcium	44-1	66666101	67878177	65291647	66611975	1.94	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1657204	1644219	1631057	1644160	0.80	cps
Cobalt	59-2	2454972	2542830	2499240	2499014	1.76	cps
Copper	63-2	19133381	19229213	19124634	19162410	0.30	cps
Dysprosium	156-1	190	203	167	187	9.94	cps
Dysprosium	156-2	273	327	280	293	9.91	cps
Erbium	164-1	353	423	420	399	9.90	cps
Erbium	164-2	297	277	230	268	12.78	cps
Gadolinium	160-1	147	217	230	198	22.63	cps
Gadolinium	160-2	603	480	563	549	11.46	cps
Holmium	165-1	8050286	7914214	7584196	7849565	3.05	cps
Holmium	165-2	5919042	5973658	5837415	5910038	1.16	cps
Indium	115-1	6593236	6693117	6567175	6617843	1.00	cps
Indium	115-2	2788813	2740221	2722089	2750374	1.25	cps
Iron	56-2	385464781	386862781	385559901	385962488	0.20	cps
Iron	57-2	9652093	9646757	9508414	9602422	0.85	cps
Iron	54-2	21031273	21402804	21098560	21177545	0.93	cps
Krypton	83-1	493	520	480	498	4.09	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10752195	10614641	10397520	10588119	1.69	cps
Lead	207-1	9217048	9152888	8991322	9120419	1.28	cps
Lead	208-1	42416729	42071402	41511609	41999913	1.09	cps
Lithium	6-1	699475	694116	688206	693932	0.81	cps
Magnesium	24-2	75988424	76843946	77839912	76890761	1.21	cps
Manganese	55-2	14166305	14069319	13963467	14066364	0.72	cps
Molybdenum	94-1	10935611	11194036	11071198	11066948	1.17	cps
Molybdenum	95-1	15978217	16360055	16147919	16162064	1.18	cps
Molybdenum	96-1	17565599	17566945	17350701	17494415	0.71	cps
Molybdenum	97-1	9835761	10140334	9746118	9907404	2.09	cps
Molybdenum	98-1	25765604	26291507	25685239	25914117	1.27	cps
Neodymium	150-1	197	183	117	166	25.89	cps
Neodymium	150-2	77	67	43	62	27.50	cps
Nickel	60-2	629691	629633	621312	626879	0.77	cps
Phosphorus	31-2	100102	99996	99640	99913	0.24	cps
Potassium	39-2	46587969	46930053	46851676	46789899	0.38	cps
Rhodium	103-1	6569988	6634045	6551917	6585317	0.66	cps
Rhodium	103-2	4118654	4241381	4265768	4208601	1.87	cps
Scandium	45-1	4113922	4082614	4057067	4084534	0.70	cps
Scandium	45-2	453018	456801	455184	455001	0.42	cps
Selenium	82-1	58313	58572	57966	58284	0.52	cps
Selenium	77-2	7585	7415	7362	7454	1.57	cps
Selenium	78-2	24624	25248	24744	24872	1.33	cps
Silicon	28-1	8007679	8089657	8148905	8082080	0.88	cps
Silver	107-1	4002721	3963641	3974493	3980285	0.51	cps
Silver	109-1	3876254	3814826	3759091	3816724	1.54	cps
Sodium	23-2	115607028	119474282	118099795	117727035	1.67	cps
Strontium	86-1	1027099	1037651	1033007	1032586	0.51	cps
Strontium	88-1	8970364	9215866	9168041	9118090	1.43	cps
Sulfur	34-1	378642	379307	378345	378765	0.13	cps
Terbium	159-1	8330716	8175825	8110162	8205568	1.38	cps
Terbium	159-2	6070623	5991609	5923449	5995227	1.23	cps
Thallium	203-1	2672951	2719900	2614134	2668995	1.99	cps
Thallium	205-1	6320929	6359612	6268248	6316263	0.73	cps
Tin	118-1	2496373	2522680	2481260	2500105	0.84	cps
Titanium	47-1	4122817	4095090	4065763	4094557	0.70	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV02 Instrumnet Name : P7
Client Sample ID : CCV02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:41:26 DataFile Name : 034CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8360399	8157050	8109354	8208934	1.62	cps
Vanadium	51-2	1429404	1426862	1404122	1420130	0.98	cps
Yttrium	89-1	12061573	11762611	11846321	11890168	1.30	cps
Yttrium	89-2	3473532	3623902	3549255	3548896	2.12	cps
Zinc	66-2	3434727	3426331	3399923	3420327	0.53	cps
Zirconium	90-1	5457776	5529661	5551181	5512873	0.89	cps
Zirconium	91-1	1199912	1212669	1212554	1208378	0.61	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1607	1573	1753	1645	5.82	cps
Antimony	121-1	723	783	857	788	8.48	cps
Arsenic	75-2	17	7	10	11	45.82	cps
Barium	135-1	163	160	123	149	14.91	cps
Barium	137-1	287	313	290	297	4.90	cps
Beryllium	9-1	17	7	14	13	40.84	cps
Bismuth	209-1	4557953	4590669	4580097	4576240	0.36	cps
Bismuth	209-2	4162670	4239737	4243317	4215242	1.08	cps
Bromine	81-1	13506	13610	13403	13506	0.77	cps
Bromine	81-2	767	627	637	677	11.54	cps
Cadmium	108-1	13	10	17	13	25.01	cps
Cadmium	106-1	2820	2810	2870	2834	1.13	cps
Cadmium	111-1	2052	2037	2064	2051	0.67	cps
Calcium	43-1	527	553	500	527	5.06	cps
Calcium	44-1	36474	36317	36046	36279	0.60	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2680	2650	2784	2705	2.59	cps
Cobalt	59-2	113	100	130	114	13.13	cps
Copper	63-2	2100	2177	2060	2112	2.81	cps
Dysprosium	156-1	3	7	0	3	100.05	cps
Dysprosium	156-2	13	10	10	11	17.30	cps
Erbium	164-1	67	40	67	58	26.65	cps
Erbium	164-2	37	50	33	40	22.05	cps
Gadolinium	160-1	53	47	33	44	22.92	cps
Gadolinium	160-2	350	423	410	394	9.90	cps
Holmium	165-1	7547023	7497649	7546050	7530241	0.37	cps
Holmium	165-2	5641085	5642629	5641390	5641701	0.01	cps
Indium	115-1	6614732	6711223	6296604	6540853	3.32	cps
Indium	115-2	2728214	2744197	2680881	2717764	1.21	cps
Iron	56-2	34776	30146	32538	32486	7.13	cps
Iron	57-2	2094	4086	1977	2719	43.60	cps
Iron	54-2	3257	3387	3280	3308	2.10	cps
Krypton	83-1	510	490	480	493	3.10	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	967	1030	940	979	4.72	cps
Lead	207-1	743	717	813	758	6.59	cps
Lead	208-1	3550	3714	3720	3661	2.63	cps
Lithium	6-1	666529	677057	668547	670711	0.83	cps
Magnesium	24-2	1687	1747	1757	1730	2.19	cps
Manganese	55-2	1209	1245	1233	1229	1.50	cps
Molybdenum	94-1	750	740	653	714	7.44	cps
Molybdenum	95-1	853	787	723	788	8.25	cps
Molybdenum	96-1	1007	920	757	894	14.19	cps
Molybdenum	97-1	500	480	480	487	2.37	cps
Molybdenum	98-1	1493	1223	1173	1297	13.28	cps
Neodymium	150-1	7	20	7	11	69.25	cps
Neodymium	150-2	3	0	0	1	173.21	cps
Nickel	60-2	647	627	590	621	4.63	cps
Phosphorus	31-2	710	813	770	764	6.79	cps
Potassium	39-2	93515	91721	91674	92303	1.14	cps
Rhodium	103-1	6603008	6567251	6506861	6559040	0.74	cps
Rhodium	103-2	4230520	4205835	4226590	4220982	0.31	cps
Scandium	45-1	3951023	3924677	3961435	3945712	0.48	cps
Scandium	45-2	429890	437998	437344	435077	1.04	cps
Selenium	82-1	113	137	176	142	22.06	cps
Selenium	77-2	0	0	3	1	173.21	cps
Selenium	78-2	610	473	557	547	12.60	cps
Silicon	28-1	570402	566782	573798	570327	0.62	cps
Silver	107-1	467	367	383	406	13.21	cps
Silver	109-1	403	330	320	351	12.96	cps
Sodium	23-2	65813	64069	64153	64678	1.52	cps
Strontium	86-1	820	863	963	882	8.33	cps
Strontium	88-1	907	910	860	892	3.13	cps
Sulfur	34-1	150798	154114	152653	152522	1.09	cps
Terbium	159-1	7835623	7712014	7724952	7757530	0.88	cps
Terbium	159-2	5725783	5697430	5762586	5728600	0.57	cps
Thallium	203-1	977	933	893	934	4.46	cps
Thallium	205-1	2387	2217	2240	2281	4.04	cps
Tin	118-1	2064	2017	2240	2107	5.59	cps
Titanium	47-1	150	140	133	141	5.94	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB02 Instrumnet Name : P7
Client Sample ID : CCB02 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:44:12 DataFile Name : 035CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	143	107	123	124	14.75	cps
Vanadium	51-2	37	27	37	33	17.32	cps
Yttrium	89-1	11547424	11439847	11351539	11446270	0.86	cps
Yttrium	89-2	3404102	3474103	3445102	3441102	1.02	cps
Zinc	66-2	4477	4431	4051	4320	5.42	cps
Zirconium	90-1	727	683	787	732	7.09	cps
Zirconium	91-1	150	100	143	131	20.71	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13685082	13862661	13687724	13745156	0.74	cps
Antimony	121-1	2784789	2815684	2868000	2822824	1.49	cps
Arsenic	75-2	226405	223121	223296	224274	0.82	cps
Barium	135-1	3868347	4007437	4008009	3961264	2.03	cps
Barium	137-1	6595192	6850488	6773638	6739773	1.94	cps
Beryllium	9-1	345912	355120	343270	348101	1.79	cps
Bismuth	209-1	4756187	4980415	4860185	4865596	2.31	cps
Bismuth	209-2	4490127	4442862	4448544	4460511	0.58	cps
Bromine	81-1	35687	44995	50232	43638	16.88	cps
Bromine	81-2	10287	10594	11064	10648	3.68	cps
Cadmium	108-1	53713	54780	53274	53922	1.44	cps
Cadmium	106-1	79962	82506	79802	80757	1.88	cps
Cadmium	111-1	772004	774282	781604	775963	0.65	cps
Calcium	43-1	1122871	1129781	1132388	1128347	0.44	cps
Calcium	44-1	17655550	17491117	17724076	17623581	0.68	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2065057	2034945	2036318	2045440	0.83	cps
Cobalt	59-2	2753155	2749880	2676486	2726507	1.59	cps
Copper	63-2	17970382	18073253	18215223	18086286	0.68	cps
Dysprosium	156-1	98832	99554	101688	100025	1.48	cps
Dysprosium	156-2	94687	93838	92866	93797	0.97	cps
Erbium	164-1	100631	102299	101007	101312	0.86	cps
Erbium	164-2	78231	79287	76988	78169	1.47	cps
Gadolinium	160-1	103685	106593	107046	105775	1.72	cps
Gadolinium	160-2	84936	86744	83635	85105	1.83	cps
Holmium	165-1	7966008	8045234	8098687	8036643	0.83	cps
Holmium	165-2	6223326	5991882	6155099	6123436	1.94	cps
Indium	115-1	6723776	6777187	6806466	6769143	0.62	cps
Indium	115-2	2824380	2810498	2800005	2811628	0.43	cps
Iron	56-2	396295447	396237367	393834701	395455839	0.36	cps
Iron	57-2	9812073	9878101	9629338	9773171	1.32	cps
Iron	54-2	22048049	21905061	22029884	21994331	0.35	cps
Krypton	83-1	627	743	720	697	8.86	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	17227158	17767076	18065517	17686584	2.40	cps
Lead	207-1	14924273	15298258	15594000	15272177	2.20	cps
Lead	208-1	68357873	70328677	70537456	69741335	1.72	cps
Lithium	6-1	726501	718380	721446	722109	0.57	cps
Magnesium	24-2	21970952	22017150	21725971	21904691	0.71	cps
Manganese	55-2	19874166	19636699	19569083	19693316	0.81	cps
Molybdenum	94-1	5020821	5054514	5040900	5038745	0.34	cps
Molybdenum	95-1	5482980	5531789	5555128	5523299	0.67	cps
Molybdenum	96-1	6141633	6187980	6206723	6178779	0.54	cps
Molybdenum	97-1	3316061	3392147	3349525	3352578	1.14	cps
Molybdenum	98-1	8584941	8680784	8744597	8670107	0.93	cps
Neodymium	150-1	161621	165399	164099	163706	1.17	cps
Neodymium	150-2	109112	109468	109143	109241	0.18	cps
Nickel	60-2	772514	770974	760916	768135	0.82	cps
Phosphorus	31-2	91841	88955	88944	89913	1.86	cps
Potassium	39-2	11085761	11404047	11074572	11188126	1.67	cps
Rhodium	103-1	6740842	6872710	6792652	6802068	0.98	cps
Rhodium	103-2	4373390	4319000	4384899	4359096	0.81	cps
Scandium	45-1	4349049	4272196	4349665	4323637	1.03	cps
Scandium	45-2	488877	480267	490764	486636	1.15	cps
Selenium	82-1	57957	58212	58596	58255	0.55	cps
Selenium	77-2	7986	8059	7956	8000	0.66	cps
Selenium	78-2	26367	26160	25148	25892	2.52	cps
Silicon	28-1	7838736	8238856	19954991	12010861	57.30	cps
Silver	107-1	683774	701376	696373	693841	1.31	cps
Silver	109-1	650596	670154	655422	658724	1.55	cps
Sodium	23-2	23947071	23757523	23753345	23819313	0.46	cps
Strontium	86-1	1312675	1335682	1341832	1330063	1.16	cps
Strontium	88-1	11759459	11789436	11799753	11782883	0.18	cps
Sulfur	34-1	215188	214244	212091	213841	0.74	cps
Terbium	159-1	8100679	8328181	8372195	8267018	1.76	cps
Terbium	159-2	6255407	6250131	6277320	6260953	0.23	cps
Thallium	203-1	2496532	2584956	2587443	2556310	2.03	cps
Thallium	205-1	5941727	6046670	6078144	6022180	1.19	cps
Tin	118-1	2119335	2126171	2133009	2126172	0.32	cps
Titanium	47-1	881949	902273	899078	894433	1.22	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-01A Instrumnet Name : P7
Client Sample ID : 0804-SOILA Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:50:45 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7933882	7930778	7795294	7886651	1.00	cps
Vanadium	51-2	1870460	1891090	1851084	1870878	1.07	cps
Yttrium	89-1	13326852	13352952	13560166	13413323	0.95	cps
Yttrium	89-2	4132883	4035527	4056280	4074897	1.26	cps
Zinc	66-2	3856449	3863124	3730085	3816553	1.96	cps
Zirconium	90-1	5348123	5427819	5383335	5386426	0.74	cps
Zirconium	91-1	1173298	1197560	1195848	1188902	1.14	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	359960	358745	353670	357458	0.93	cps
Antimony	121-1	2974	3210	3414	3199	6.88	cps
Arsenic	75-2	687	743	733	721	4.19	cps
Barium	135-1	42604	43242	42433	42760	1.00	cps
Barium	137-1	74388	73268	71419	73025	2.05	cps
Beryllium	9-1	137	110	130	126	11.15	cps
Bismuth	209-1	4706592	4769049	4716707	4730783	0.71	cps
Bismuth	209-2	4376597	4384714	4252911	4338074	1.70	cps
Bromine	81-1	17978	17507	17183	17556	2.28	cps
Bromine	81-2	1413	1393	1433	1413	1.42	cps
Cadmium	108-1	83	67	83	78	12.37	cps
Cadmium	106-1	3020	3017	3060	3033	0.80	cps
Cadmium	111-1	2683	2629	2661	2657	1.02	cps
Calcium	43-1	1179852	1191994	1190208	1187351	0.55	cps
Calcium	44-1	19067473	18944835	19107241	19039850	0.44	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	45736	45699	45331	45589	0.49	cps
Cobalt	59-2	14904	15048	14187	14713	3.13	cps
Copper	63-2	52826	53981	53469	53425	1.08	cps
Dysprosium	156-1	3184	3420	3104	3236	5.09	cps
Dysprosium	156-2	3380	3277	3147	3268	3.58	cps
Erbium	164-1	3450	3447	3517	3472	1.14	cps
Erbium	164-2	2757	2454	2670	2627	5.95	cps
Gadolinium	160-1	3457	3317	3417	3397	2.12	cps
Gadolinium	160-2	3284	3077	3004	3121	4.65	cps
Holmium	165-1	7632139	7831881	7646322	7703447	1.45	cps
Holmium	165-2	5877847	5931662	5853294	5887601	0.68	cps
Indium	115-1	6890527	6916715	6763636	6856959	1.19	cps
Indium	115-2	2774607	2826990	2839072	2813556	1.22	cps
Iron	56-2	17939323	17666669	17290066	17632019	1.85	cps
Iron	57-2	485699	481264	485216	484060	0.50	cps
Iron	54-2	1011063	1003386	1007136	1007195	0.38	cps
Krypton	83-1	497	463	513	491	5.18	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	252762	257690	250701	253717	1.42	cps
Lead	207-1	217441	209346	204148	210312	3.19	cps
Lead	208-1	981438	984087	976336	980621	0.40	cps
Lithium	6-1	695021	696598	683815	691811	1.01	cps
Magnesium	24-2	581367	576995	578467	578943	0.38	cps
Manganese	55-2	476104	470627	472332	473021	0.59	cps
Molybdenum	94-1	10694	10591	10647	10644	0.49	cps
Molybdenum	95-1	3164	3004	3040	3069	2.73	cps
Molybdenum	96-1	4724	4918	4731	4791	2.29	cps
Molybdenum	97-1	1937	1843	1943	1908	2.93	cps
Molybdenum	98-1	5111	4784	4861	4919	3.47	cps
Neodymium	150-1	5351	5398	5461	5403	1.02	cps
Neodymium	150-2	3754	3557	3527	3613	3.41	cps
Nickel	60-2	10617	10390	10157	10388	2.22	cps
Phosphorus	31-2	6138	6135	5901	6058	2.24	cps
Potassium	39-2	351334	349294	352197	350942	0.42	cps
Rhodium	103-1	7034247	6825373	6670135	6843252	2.67	cps
Rhodium	103-2	4399040	4350493	4380530	4376688	0.56	cps
Scandium	45-1	4204732	4166811	4112843	4161462	1.11	cps
Scandium	45-2	455965	456630	448532	453709	0.99	cps
Selenium	82-1	213	218	157	196	17.31	cps
Selenium	77-2	23	20	20	21	9.11	cps
Selenium	78-2	647	543	570	587	9.14	cps
Silicon	28-1	3375436	3346572	3267033	3329680	1.69	cps
Silver	107-1	913	1013	900	942	6.57	cps
Silver	109-1	980	787	873	880	11.00	cps
Sodium	23-2	998703	994553	979111	990789	1.04	cps
Strontium	86-1	90678	87837	88364	88959	1.70	cps
Strontium	88-1	750985	765281	759429	758565	0.95	cps
Sulfur	34-1	164980	171774	171959	169571	2.35	cps
Terbium	159-1	8329928	7856879	7852166	8012991	3.43	cps
Terbium	159-2	5906493	5864639	5834596	5868576	0.62	cps
Thallium	203-1	1470	1217	1147	1278	13.31	cps
Thallium	205-1	3430	2947	2797	3058	10.83	cps
Tin	118-1	170274	168273	165884	168144	1.31	cps
Titanium	47-1	64407	64235	64404	64349	0.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-02 Instrumnet Name : P7
Client Sample ID : 0804-D Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:53:34 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4494	4434	4574	4501	1.56	cps
Vanadium	51-2	19032	19152	19206	19130	0.46	cps
Yttrium	89-1	11993479	12173975	12147598	12105018	0.81	cps
Yttrium	89-2	3576936	3634009	3584679	3598541	0.86	cps
Zinc	66-2	60449	59314	59910	59891	0.95	cps
Zirconium	90-1	24988	24571	24330	24630	1.35	cps
Zirconium	91-1	5541	5631	5301	5491	3.11	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3750247	3672584	3824833	3749221	2.03	cps
Antimony	121-1	5438	5274	5208	5307	2.23	cps
Arsenic	75-2	4087	4304	4237	4210	2.64	cps
Barium	135-1	732888	722780	724709	726792	0.74	cps
Barium	137-1	1271070	1244219	1255461	1256917	1.07	cps
Beryllium	9-1	5743	6023	5857	5874	2.40	cps
Bismuth	209-1	5169950	5101313	4875773	5049012	3.05	cps
Bismuth	209-2	4697372	4606257	4601533	4635054	1.17	cps
Bromine	81-1	34334	42211	46898	41147	15.43	cps
Bromine	81-2	9163	9360	9093	9205	1.50	cps
Cadmium	108-1	1723	1970	1823	1839	6.75	cps
Cadmium	106-1	6198	6085	5931	6071	2.21	cps
Cadmium	111-1	27556	27652	27163	27457	0.94	cps
Calcium	43-1	1779598	1728929	1734170	1747566	1.59	cps
Calcium	44-1	28293598	28152335	27559290	28001741	1.39	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	237669	233190	237966	236275	1.13	cps
Cobalt	59-2	353769	355723	353738	354410	0.32	cps
Copper	63-2	327063	329155	328623	328280	0.33	cps
Dysprosium	156-1	78981	78509	78488	78659	0.35	cps
Dysprosium	156-2	74426	74999	74510	74645	0.41	cps
Erbium	164-1	90954	91881	91068	91301	0.55	cps
Erbium	164-2	71257	71783	71744	71595	0.41	cps
Gadolinium	160-1	83370	84279	84728	84126	0.82	cps
Gadolinium	160-2	68885	69039	68523	68816	0.39	cps
Holmium	165-1	8446567	8233519	8287048	8322378	1.33	cps
Holmium	165-2	6471298	6390680	6414509	6425495	0.64	cps
Indium	115-1	7474709	7260801	7176248	7303919	2.11	cps
Indium	115-2	3042558	2989042	3020220	3017274	0.89	cps
Iron	56-2	130821461	131477328	130458941	130919244	0.39	cps
Iron	57-2	3225752	3294878	3258322	3259650	1.06	cps
Iron	54-2	7147311	7087889	7233127	7156109	1.02	cps
Krypton	83-1	653	780	670	701	9.82	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	643041	638647	648943	643544	0.80	cps
Lead	207-1	544680	530373	536158	537070	1.34	cps
Lead	208-1	2519152	2494393	2487297	2500281	0.67	cps
Lithium	6-1	744139	739285	716623	733349	2.00	cps
Magnesium	24-2	15842390	15950814	15762868	15852024	0.60	cps
Manganese	55-2	3712052	3746559	3759504	3739372	0.66	cps
Molybdenum	94-1	96499	95691	96858	96349	0.62	cps
Molybdenum	95-1	86548	87520	89392	87820	1.65	cps
Molybdenum	96-1	100859	103363	102762	102328	1.28	cps
Molybdenum	97-1	53896	53983	55405	54428	1.56	cps
Molybdenum	98-1	137444	139786	142532	139921	1.82	cps
Neodymium	150-1	117776	116265	115594	116545	0.96	cps
Neodymium	150-2	79372	78270	78963	78868	0.71	cps
Nickel	60-2	82975	81738	82070	82261	0.78	cps
Phosphorus	31-2	192755	188220	192434	191137	1.32	cps
Potassium	39-2	10499100	10370703	10681998	10517267	1.49	cps
Rhodium	103-1	7251050	7162799	7080109	7164653	1.19	cps
Rhodium	103-2	4626143	4516223	4556226	4566197	1.22	cps
Scandium	45-1	4556666	4499979	4372704	4476450	2.10	cps
Scandium	45-2	506078	495878	502848	501601	1.04	cps
Selenium	82-1	754	683	823	753	9.27	cps
Selenium	77-2	410	390	410	403	2.86	cps
Selenium	78-2	1090	1107	1077	1091	1.38	cps
Silicon	28-1	262933589	260473196	259596496	261001094	0.66	cps
Silver	107-1	1213	1273	1303	1263	3.63	cps
Silver	109-1	1130	1133	1093	1119	1.98	cps
Sodium	23-2	63943809	63877297	63422744	63747950	0.44	cps
Strontium	86-1	814298	795851	800672	803607	1.19	cps
Strontium	88-1	7090792	6868977	7185185	7048318	2.30	cps
Sulfur	34-1	1065195	1044299	1051918	1053804	1.00	cps
Terbium	159-1	8838165	8647109	8680344	8721873	1.17	cps
Terbium	159-2	6540799	6441976	6534715	6505830	0.85	cps
Thallium	203-1	4187	4337	3991	4172	4.17	cps
Thallium	205-1	10047	9814	9967	9943	1.19	cps
Tin	118-1	23422	18869	19196	20496	12.39	cps
Titanium	47-1	160759	160334	159486	160193	0.40	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03 Instrumnet Name : P7
Client Sample ID : 0804-E Dilution Factor : 1
Date & Time Acquired : 2025-08-15 14:56:49 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	120769	121880	116280	119643	2.48	cps
Vanadium	51-2	1193282	1188999	1209372	1197218	0.90	cps
Yttrium	89-1	14521169	14541563	14533249	14531994	0.07	cps
Yttrium	89-2	4435367	4395410	4385075	4405284	0.60	cps
Zinc	66-2	320258	323239	326811	323436	1.01	cps
Zirconium	90-1	126990	125902	130867	127920	2.04	cps
Zirconium	91-1	29326	28852	28992	29057	0.84	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8569	8823	8606	8666	1.58	cps
Antimony	121-1	1457	1273	1560	1430	10.15	cps
Arsenic	75-2	343	280	313	312	10.15	cps
Barium	135-1	666101	658394	659400	661298	0.63	cps
Barium	137-1	1152127	1144718	1137866	1144904	0.62	cps
Beryllium	9-1	44	51	61	52	16.18	cps
Bismuth	209-1	5150512	5093675	5164019	5136069	0.73	cps
Bismuth	209-2	4832151	4787946	4745948	4788682	0.90	cps
Bromine	81-1	38277	45734	50586	44866	13.82	cps
Bromine	81-2	9603	9340	9583	9509	1.54	cps
Cadmium	108-1	27	23	40	30	29.40	cps
Cadmium	106-1	3247	3044	3124	3138	3.27	cps
Cadmium	111-1	2331	2168	2209	2236	3.80	cps
Calcium	43-1	549424	539367	538146	542312	1.14	cps
Calcium	44-1	8566544	8577999	8457494	8534012	0.78	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3841	3874	4064	3926	3.07	cps
Cobalt	59-2	45078	45546	45636	45420	0.66	cps
Copper	63-2	2587	2627	2737	2650	2.93	cps
Dysprosium	156-1	33	43	47	41	16.89	cps
Dysprosium	156-2	53	47	33	44	22.92	cps
Erbium	164-1	87	90	73	83	10.59	cps
Erbium	164-2	83	43	63	63	31.58	cps
Gadolinium	160-1	87	103	100	97	9.12	cps
Gadolinium	160-2	503	470	533	502	6.31	cps
Holmium	165-1	8482844	8590322	8470301	8514489	0.77	cps
Holmium	165-2	6428101	6425470	6461289	6438287	0.31	cps
Indium	115-1	7681794	7459598	7441306	7527566	1.78	cps
Indium	115-2	3101256	3089261	3071660	3087393	0.48	cps
Iron	56-2	23654658	23642482	23516437	23604526	0.32	cps
Iron	57-2	612330	611201	620883	614804	0.86	cps
Iron	54-2	1321108	1342372	1338198	1333893	0.84	cps
Krypton	83-1	453	520	420	464	10.96	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	3751	3704	3677	3711	1.00	cps
Lead	207-1	3047	3164	3167	3126	2.19	cps
Lead	208-1	14580	14917	14253	14583	2.28	cps
Lithium	6-1	733241	728576	712457	724758	1.50	cps
Magnesium	24-2	3180091	3209357	3194132	3194527	0.46	cps
Manganese	55-2	1895429	1888052	1918768	1900750	0.84	cps
Molybdenum	94-1	27773	27513	27092	27459	1.25	cps
Molybdenum	95-1	48667	48627	47788	48361	1.03	cps
Molybdenum	96-1	52096	51460	51166	51574	0.92	cps
Molybdenum	97-1	30368	30134	29640	30048	1.24	cps
Molybdenum	98-1	77749	77437	77253	77480	0.32	cps
Neodymium	150-1	53	73	50	59	21.43	cps
Neodymium	150-2	37	17	20	24	43.83	cps
Nickel	60-2	16566	16856	16416	16613	1.35	cps
Phosphorus	31-2	3811	3887	3961	3886	1.93	cps
Potassium	39-2	3281472	3265046	3248498	3265005	0.50	cps
Rhodium	103-1	7624145	7502838	7357583	7494856	1.78	cps
Rhodium	103-2	4791206	4822261	4742536	4785334	0.84	cps
Scandium	45-1	4546520	4464146	4404026	4471564	1.60	cps
Scandium	45-2	496883	503367	490772	497007	1.27	cps
Selenium	82-1	266	168	224	220	22.58	cps
Selenium	77-2	3	0	3	2	86.60	cps
Selenium	78-2	497	570	600	556	9.57	cps
Silicon	28-1	64685731	65773232	65343334	65267432	0.84	cps
Silver	107-1	150	190	163	168	12.14	cps
Silver	109-1	160	167	143	157	7.67	cps
Sodium	23-2	3243342	3225552	3213660	3227518	0.46	cps
Strontium	86-1	667755	664152	660557	664155	0.54	cps
Strontium	88-1	5982921	5923303	5820287	5908837	1.39	cps
Sulfur	34-1	422847	420898	420933	421559	0.26	cps
Terbium	159-1	8592524	8581915	8539986	8571475	0.32	cps
Terbium	159-2	6543427	6583948	6519956	6549111	0.49	cps
Thallium	203-1	1777	1787	1667	1743	3.82	cps
Thallium	205-1	4024	4064	3764	3951	4.12	cps
Tin	118-1	7652	7569	7512	7578	0.93	cps
Titanium	47-1	817	857	983	886	9.83	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:00:01 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	217	213	207	212	2.40	cps
Vanadium	51-2	987	963	863	938	6.99	cps
Yttrium	89-1	13383811	13241919	12970197	13198642	1.59	cps
Yttrium	89-2	3873567	3963717	3815972	3884419	1.92	cps
Zinc	66-2	7812	7659	7976	7815	2.03	cps
Zirconium	90-1	770	723	663	719	7.44	cps
Zirconium	91-1	173	223	137	178	24.47	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8292	8386	8066	8248	2.00	cps
Antimony	121-1	1373	1377	1250	1333	5.41	cps
Arsenic	75-2	323	300	307	310	3.88	cps
Barium	135-1	653185	654066	646887	651380	0.60	cps
Barium	137-1	1139067	1128081	1116636	1127928	0.99	cps
Beryllium	9-1	27	37	41	35	19.89	cps
Bismuth	209-1	5239733	5010655	5075839	5108742	2.31	cps
Bismuth	209-2	4704594	4770214	4793482	4756097	0.97	cps
Bromine	81-1	44805	51436	54524	50255	9.88	cps
Bromine	81-2	10884	11134	10914	10978	1.24	cps
Cadmium	108-1	20	13	27	20	33.35	cps
Cadmium	106-1	3247	3130	2897	3091	5.77	cps
Cadmium	111-1	2296	2215	2043	2184	5.92	cps
Calcium	43-1	534927	534113	523360	530800	1.22	cps
Calcium	44-1	8595262	8486414	8243716	8441797	2.13	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4211	4231	4427	4290	2.79	cps
Cobalt	59-2	44690	44743	44609	44681	0.15	cps
Copper	63-2	2450	2454	2434	2446	0.44	cps
Dysprosium	156-1	50	43	33	42	19.87	cps
Dysprosium	156-2	63	70	63	66	5.87	cps
Erbium	164-1	87	90	97	91	5.59	cps
Erbium	164-2	67	73	83	74	11.26	cps
Gadolinium	160-1	77	90	67	78	15.05	cps
Gadolinium	160-2	380	447	483	437	12.00	cps
Holmium	165-1	8640343	8559668	8230984	8476998	2.56	cps
Holmium	165-2	6445562	6467084	6487863	6466837	0.33	cps
Indium	115-1	7564725	7434296	7350976	7449999	1.45	cps
Indium	115-2	3074091	3085954	3042537	3067527	0.73	cps
Iron	56-2	23202366	23504966	23617059	23441464	0.92	cps
Iron	57-2	604085	604631	609439	606052	0.49	cps
Iron	54-2	1318070	1325754	1322995	1322273	0.29	cps
Krypton	83-1	460	523	447	477	8.59	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1000	843	997	947	9.46	cps
Lead	207-1	763	777	843	794	5.40	cps
Lead	208-1	3544	3420	3870	3611	6.44	cps
Lithium	6-1	742544	719561	712574	724893	2.16	cps
Magnesium	24-2	3085499	3192583	3138137	3138740	1.71	cps
Manganese	55-2	1833546	1888145	1917703	1879798	2.27	cps
Molybdenum	94-1	35727	35600	35062	35463	1.00	cps
Molybdenum	95-1	62363	62737	61479	62193	1.04	cps
Molybdenum	96-1	66320	67211	64633	66054	1.98	cps
Molybdenum	97-1	39266	38808	38116	38730	1.49	cps
Molybdenum	98-1	98666	98862	98584	98704	0.15	cps
Neodymium	150-1	47	53	70	57	21.21	cps
Neodymium	150-2	23	33	33	30	19.25	cps
Nickel	60-2	15782	16379	16239	16133	1.94	cps
Phosphorus	31-2	3897	3964	3811	3891	1.98	cps
Potassium	39-2	3249079	3241966	3304484	3265176	1.05	cps
Rhodium	103-1	7462923	7307317	7422991	7397744	1.09	cps
Rhodium	103-2	4661320	4760222	4788469	4736670	1.41	cps
Scandium	45-1	4678630	4474572	4416722	4523308	3.04	cps
Scandium	45-2	494387	487779	493745	491971	0.74	cps
Selenium	82-1	225	163	197	195	15.95	cps
Selenium	77-2	7	3	0	3	100.05	cps
Selenium	78-2	617	583	610	603	2.92	cps
Silicon	28-1	65572242	65438226	64489481	65166650	0.91	cps
Silver	107-1	160	120	160	147	15.75	cps
Silver	109-1	130	90	120	113	18.37	cps
Sodium	23-2	3185162	3189740	3118605	3164502	1.26	cps
Strontium	86-1	661870	656644	650178	656231	0.89	cps
Strontium	88-1	5760568	5851741	5706486	5772932	1.27	cps
Sulfur	34-1	424401	424893	425628	424974	0.15	cps
Terbium	159-1	8631580	8737295	8637795	8668890	0.68	cps
Terbium	159-2	6392461	6463194	6495021	6450225	0.81	cps
Thallium	203-1	1500	1570	1550	1540	2.34	cps
Thallium	205-1	3551	3774	3731	3685	3.21	cps
Tin	118-1	7269	7155	7152	7192	0.92	cps
Titanium	47-1	887	903	803	864	6.20	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03DUP Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:03:09 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	217	213	260	230	11.32	cps
Vanadium	51-2	893	983	867	914	6.68	cps
Yttrium	89-1	13429278	13304290	13067292	13266953	1.39	cps
Yttrium	89-2	3823862	3866753	3814277	3834964	0.73	cps
Zinc	66-2	8049	8032	7865	7982	1.27	cps
Zirconium	90-1	610	607	587	601	2.10	cps
Zirconium	91-1	123	107	140	123	13.51	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2810	2884	2874	2856	1.39	cps
Antimony	121-1	303	360	330	331	8.56	cps
Arsenic	75-2	53	73	73	67	17.32	cps
Barium	135-1	123659	122214	125434	123769	1.30	cps
Barium	137-1	211291	215683	211219	212731	1.20	cps
Beryllium	9-1	10	10	13	11	17.31	cps
Bismuth	209-1	4805886	4601471	4799955	4735771	2.46	cps
Bismuth	209-2	4489606	4266460	4373725	4376597	2.55	cps
Bromine	81-1	25596	25679	26043	25773	0.92	cps
Bromine	81-2	3504	3220	3497	3407	4.75	cps
Cadmium	108-1	17	13	13	14	13.35	cps
Cadmium	106-1	2914	2817	2597	2776	5.85	cps
Cadmium	111-1	2068	1999	1835	1967	6.08	cps
Calcium	43-1	102105	98679	98830	99871	1.94	cps
Calcium	44-1	1627168	1558999	1627457	1604541	2.46	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3460	3440	3607	3503	2.60	cps
Cobalt	59-2	8549	8442	8396	8462	0.93	cps
Copper	63-2	1930	1917	2030	1959	3.16	cps
Dysprosium	156-1	0	10	10	7	86.60	cps
Dysprosium	156-2	7	13	23	14	58.06	cps
Erbium	164-1	70	87	60	72	18.66	cps
Erbium	164-2	33	63	47	48	31.46	cps
Gadolinium	160-1	37	57	33	42	29.89	cps
Gadolinium	160-2	470	403	490	454	9.99	cps
Holmium	165-1	7510174	7429328	7550282	7496595	0.82	cps
Holmium	165-2	6057135	5820811	5802834	5893593	2.41	cps
Indium	115-1	6834303	6614640	6996744	6815229	2.81	cps
Indium	115-2	2887091	2798609	2815823	2833841	1.66	cps
Iron	56-2	4534785	4521951	4509515	4522083	0.28	cps
Iron	57-2	118346	118776	119427	118849	0.46	cps
Iron	54-2	257908	259849	257724	258494	0.46	cps
Krypton	83-1	517	500	623	547	12.24	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	903	937	1013	951	5.93	cps
Lead	207-1	810	877	807	831	4.75	cps
Lead	208-1	3744	3947	3870	3854	2.67	cps
Lithium	6-1	676278	633601	664492	658124	3.35	cps
Magnesium	24-2	586412	587780	583357	585850	0.39	cps
Manganese	55-2	352877	356372	351640	353630	0.69	cps
Molybdenum	94-1	4901	4798	4807	4835	1.18	cps
Molybdenum	95-1	8102	8206	8533	8280	2.71	cps
Molybdenum	96-1	8703	9156	8869	8909	2.58	cps
Molybdenum	97-1	5211	5281	5371	5288	1.52	cps
Molybdenum	98-1	13320	12959	13553	13277	2.25	cps
Neodymium	150-1	23	27	13	21	32.88	cps
Neodymium	150-2	3	10	3	6	69.34	cps
Nickel	60-2	3510	3857	3620	3663	4.84	cps
Phosphorus	31-2	1257	1283	1397	1312	5.67	cps
Potassium	39-2	691496	703529	697506	697510	0.86	cps
Rhodium	103-1	6793266	6798828	6601026	6731040	1.67	cps
Rhodium	103-2	4444692	4377776	4413743	4412070	0.76	cps
Scandium	45-1	4051952	3866448	4050242	3989547	2.67	cps
Scandium	45-2	458965	442876	455758	452533	1.88	cps
Selenium	82-1	115	94	15	75	70.78	cps
Selenium	77-2	0	3	3	2	86.60	cps
Selenium	78-2	587	593	647	609	5.40	cps
Silicon	28-1	12518388	12523161	12740259	12593936	1.01	cps
Silver	107-1	117	133	113	121	8.85	cps
Silver	109-1	100	103	110	104	4.88	cps
Sodium	23-2	630289	629646	625290	628408	0.43	cps
Strontium	86-1	122090	125040	124190	123773	1.23	cps
Strontium	88-1	1083892	1063766	1050209	1065955	1.59	cps
Sulfur	34-1	171663	173648	172199	172503	0.60	cps
Terbium	159-1	7794351	7917228	7790990	7834190	0.92	cps
Terbium	159-2	6189274	5884712	5963894	6012627	2.63	cps
Thallium	203-1	667	617	610	631	4.91	cps
Thallium	205-1	1517	1330	1390	1412	6.75	cps
Tin	118-1	2620	2637	2710	2656	1.80	cps
Titanium	47-1	223	210	307	247	21.24	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03LX5 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:06:24 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	50	20	47	39	42.28	cps
Vanadium	51-2	180	207	177	188	8.76	cps
Yttrium	89-1	11851342	11709939	11694304	11751862	0.74	cps
Yttrium	89-2	3613273	3476320	3598536	3562710	2.11	cps
Zinc	66-2	5101	5044	4961	5035	1.40	cps
Zirconium	90-1	500	417	517	478	11.21	cps
Zirconium	91-1	83	123	100	102	19.66	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1751234	1646661	1676701	1691532	3.18	cps
Antimony	121-1	3102713	3024530	2996168	3041137	1.81	cps
Arsenic	75-2	233195	225245	224524	227655	2.11	cps
Barium	135-1	4048267	4043256	4039552	4043692	0.11	cps
Barium	137-1	6912083	6916045	6938376	6922168	0.20	cps
Beryllium	9-1	362337	359136	357258	359577	0.71	cps
Bismuth	209-1	5256778	5122000	5251346	5210041	1.46	cps
Bismuth	209-2	4935399	4609983	4617808	4721063	3.93	cps
Bromine	81-1	57207	67926	75227	66787	13.57	cps
Bromine	81-2	16026	15872	16356	16084	1.54	cps
Cadmium	108-1	58160	57751	57467	57793	0.60	cps
Cadmium	106-1	86926	87348	85591	86621	1.06	cps
Cadmium	111-1	841266	840342	832580	838062	0.57	cps
Calcium	43-1	1321021	1339297	1312631	1324317	1.03	cps
Calcium	44-1	21474613	21754971	20755812	21328465	2.42	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1664282	1632200	1599546	1632009	1.98	cps
Cobalt	59-2	2608951	2560817	2610925	2593564	1.09	cps
Copper	63-2	18396085	18065865	18136746	18199565	0.96	cps
Dysprosium	156-1	103	93	90	96	7.26	cps
Dysprosium	156-2	267	220	267	251	10.73	cps
Erbium	164-1	137	120	147	134	10.02	cps
Erbium	164-2	77	117	97	97	20.69	cps
Gadolinium	160-1	140	137	130	136	3.76	cps
Gadolinium	160-2	547	460	543	517	9.50	cps
Holmium	165-1	8615789	8508770	8500134	8541565	0.75	cps
Holmium	165-2	6687396	6522996	6336560	6515651	2.69	cps
Indium	115-1	7501289	7359403	7257624	7372772	1.66	cps
Indium	115-2	3124112	3058819	2890937	3024623	3.98	cps
Iron	56-2	93512469	92510625	91414915	92479336	1.13	cps
Iron	57-2	2355410	2297953	2297953	2317105	1.43	cps
Iron	54-2	5259879	5022724	5046304	5109636	2.56	cps
Krypton	83-1	520	473	467	487	5.97	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11128845	11081113	10874688	11028215	1.22	cps
Lead	207-1	9724118	9740929	9467770	9644273	1.59	cps
Lead	208-1	44036306	43905076	43639082	43860155	0.46	cps
Lithium	6-1	755131	739240	733548	742639	1.51	cps
Magnesium	24-2	19791004	18302551	18364114	18819223	4.47	cps
Manganese	55-2	15416444	14714358	14721381	14950728	2.70	cps
Molybdenum	94-1	5409447	5342943	5252084	5334825	1.48	cps
Molybdenum	95-1	6222316	6063235	6000526	6095359	1.88	cps
Molybdenum	96-1	7043423	6815723	6648993	6836047	2.90	cps
Molybdenum	97-1	3813148	3723637	3662764	3733183	2.03	cps
Molybdenum	98-1	9587086	9418752	9494267	9500035	0.89	cps
Neodymium	150-1	150	113	130	131	14.00	cps
Neodymium	150-2	63	43	60	56	19.29	cps
Nickel	60-2	662545	666941	671930	667139	0.70	cps
Phosphorus	31-2	3771	3764	3814	3783	0.72	cps
Potassium	39-2	11826341	11818364	11915624	11853443	0.46	cps
Rhodium	103-1	7541753	7278561	7284719	7368344	2.04	cps
Rhodium	103-2	4874983	4726255	4542821	4714686	3.53	cps
Scandium	45-1	4566840	4569195	4376320	4504118	2.46	cps
Scandium	45-2	508683	490272	480213	493056	2.93	cps
Selenium	82-1	64285	64135	64406	64275	0.21	cps
Selenium	77-2	7779	7926	7936	7880	1.11	cps
Selenium	78-2	27456	27289	26855	27200	1.14	cps
Silicon	28-1	64499649	64476552	63615864	64197355	0.78	cps
Silver	107-1	750153	733253	740105	741170	1.15	cps
Silver	109-1	704040	706209	703805	704685	0.19	cps
Sodium	23-2	27873685	26284843	26101217	26753248	3.64	cps
Strontium	86-1	1863318	1907597	1848519	1873145	1.64	cps
Strontium	88-1	16384662	16304176	16252593	16313810	0.41	cps
Sulfur	34-1	430511	433664	436834	433670	0.73	cps
Terbium	159-1	9120477	9047115	8813398	8993663	1.78	cps
Terbium	159-2	6741544	6556891	6289243	6529226	3.48	cps
Thallium	203-1	2667093	2749635	2690748	2702492	1.57	cps
Thallium	205-1	6451814	6322144	6297124	6357028	1.31	cps
Tin	118-1	2369801	2382090	2339179	2363690	0.93	cps
Titanium	47-1	2344	2560	2397	2434	4.64	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-04 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:09:39 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8010773	8027262	8240643	8092893	1.58	cps
Vanadium	51-2	1478208	1417556	1400655	1432140	2.85	cps
Yttrium	89-1	13335215	13122295	12825843	13094451	1.95	cps
Yttrium	89-2	3972683	3900922	3725936	3866514	3.28	cps
Zinc	66-2	3587430	3411949	3406237	3468539	2.97	cps
Zirconium	90-1	5637691	5546697	5492168	5558852	1.32	cps
Zirconium	91-1	1209266	1210314	1196465	1205348	0.64	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1667511	1713360	1650420	1677097	1.94	cps
Antimony	121-1	3117195	3066791	3060889	3081625	1.00	cps
Arsenic	75-2	215124	219962	216190	217092	1.17	cps
Barium	135-1	4079055	4133243	3957137	4056479	2.22	cps
Barium	137-1	6925460	6997557	6843911	6922309	1.11	cps
Beryllium	9-1	358655	351577	350468	353567	1.26	cps
Bismuth	209-1	5253240	5137213	5006264	5132239	2.41	cps
Bismuth	209-2	4700640	4708715	4702409	4703921	0.09	cps
Bromine	81-1	59855	71965	78632	70150	13.57	cps
Bromine	81-2	16459	17280	16870	16870	2.43	cps
Cadmium	108-1	59241	57320	58635	58399	1.68	cps
Cadmium	106-1	86624	86124	86114	86287	0.34	cps
Cadmium	111-1	836275	845971	828174	836806	1.06	cps
Calcium	43-1	1310389	1309482	1281118	1300330	1.28	cps
Calcium	44-1	21439505	20866537	20721190	21009077	1.81	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1595572	1604177	1623336	1607695	0.88	cps
Cobalt	59-2	2593400	2612390	2561393	2589061	1.00	cps
Copper	63-2	18105034	17931021	18192015	18076023	0.74	cps
Dysprosium	156-1	107	77	130	104	25.60	cps
Dysprosium	156-2	330	287	250	289	13.86	cps
Erbium	164-1	133	120	127	127	5.27	cps
Erbium	164-2	73	93	113	93	21.43	cps
Gadolinium	160-1	137	117	113	122	10.32	cps
Gadolinium	160-2	573	560	497	543	7.54	cps
Holmium	165-1	8552996	8402170	8443651	8466272	0.92	cps
Holmium	165-2	6520461	6487253	6370854	6459523	1.22	cps
Indium	115-1	7507333	7421309	7282046	7403563	1.54	cps
Indium	115-2	3026636	3028255	2952394	3002428	1.44	cps
Iron	56-2	90855219	92437295	89530482	90940999	1.60	cps
Iron	57-2	2293828	2316461	2270597	2293629	1.00	cps
Iron	54-2	5059372	5171705	5076685	5102587	1.19	cps
Krypton	83-1	533	430	473	479	10.84	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11243024	11190404	11104217	11179215	0.63	cps
Lead	207-1	9611454	9659265	9775539	9682086	0.87	cps
Lead	208-1	44547304	43892221	44009686	44149737	0.79	cps
Lithium	6-1	730071	730772	710014	723619	1.63	cps
Magnesium	24-2	18400408	18980618	18258254	18546427	2.06	cps
Manganese	55-2	14913084	15045746	14814846	14924559	0.78	cps
Molybdenum	94-1	5393580	5326848	5357373	5359267	0.62	cps
Molybdenum	95-1	6125664	6006116	6143546	6091775	1.23	cps
Molybdenum	96-1	6778927	6733433	6752143	6754834	0.34	cps
Molybdenum	97-1	3767949	3632944	3656331	3685741	1.96	cps
Molybdenum	98-1	9774664	9485080	9283057	9514267	2.60	cps
Neodymium	150-1	137	160	133	143	10.13	cps
Neodymium	150-2	77	57	60	64	16.63	cps
Nickel	60-2	669841	654511	669806	664719	1.33	cps
Phosphorus	31-2	3764	3430	3600	3598	4.63	cps
Potassium	39-2	11928027	11492317	11882894	11767746	2.04	cps
Rhodium	103-1	7503593	7202258	7135481	7280444	2.69	cps
Rhodium	103-2	4647811	4630722	4641670	4640068	0.19	cps
Scandium	45-1	4466332	4410724	4313986	4397014	1.75	cps
Scandium	45-2	485051	484850	490099	486666	0.61	cps
Selenium	82-1	64079	63586	63754	63806	0.39	cps
Selenium	77-2	8176	7959	8009	8048	1.41	cps
Selenium	78-2	26895	26614	26527	26679	0.72	cps
Silicon	28-1	61514254	61343222	60235592	61031023	1.14	cps
Silver	107-1	749508	743250	733510	742090	1.09	cps
Silver	109-1	712078	704207	697046	704443	1.07	cps
Sodium	23-2	26106593	27007142	25625098	26246278	2.67	cps
Strontium	86-1	1863752	1872425	1884998	1873725	0.57	cps
Strontium	88-1	16045741	16220435	15995432	16087203	0.73	cps
Sulfur	34-1	430208	425777	425625	427204	0.61	cps
Terbium	159-1	8888374	8849581	8578160	8772038	1.93	cps
Terbium	159-2	6425452	6581639	6495133	6500741	1.20	cps
Thallium	203-1	2697453	2722503	2620283	2680080	1.99	cps
Thallium	205-1	6453071	6506297	6441777	6467048	0.53	cps
Tin	118-1	2373368	2354207	2326342	2351306	1.01	cps
Titanium	47-1	2360	2384	2300	2348	1.83	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-05 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:12:27 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8287847	8265869	7983409	8179042	2.08	cps
Vanadium	51-2	1434996	1456887	1433323	1441735	0.91	cps
Yttrium	89-1	13374193	13277245	12908315	13186584	1.86	cps
Yttrium	89-2	3873502	3803729	3901297	3859509	1.30	cps
Zinc	66-2	3478855	3423407	3426263	3442842	0.91	cps
Zirconium	90-1	5508335	5483216	5416055	5469202	0.87	cps
Zirconium	91-1	1196068	1194522	1186122	1192238	0.45	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1655104	1674128	1605157	1644797	2.17	cps
Antimony	121-1	3027317	3021561	3061290	3036722	0.71	cps
Arsenic	75-2	226910	226028	225082	226007	0.40	cps
Barium	135-1	4025365	3974383	3998262	3999337	0.64	cps
Barium	137-1	6867366	6929624	6779994	6858995	1.10	cps
Beryllium	9-1	348275	347816	344017	346702	0.67	cps
Bismuth	209-1	5125579	5405854	5137588	5223007	3.03	cps
Bismuth	209-2	4673106	4772400	4674375	4706627	1.21	cps
Bromine	81-1	59901	68957	76011	68290	11.83	cps
Bromine	81-2	16389	16146	16706	16414	1.71	cps
Cadmium	108-1	57895	58846	57326	58023	1.32	cps
Cadmium	106-1	85491	85675	84515	85227	0.73	cps
Cadmium	111-1	828292	833558	828332	830061	0.36	cps
Calcium	43-1	1291653	1268750	1287918	1282773	0.96	cps
Calcium	44-1	20682003	20663971	20927742	20757905	0.71	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1610921	1576940	1596531	1594797	1.07	cps
Cobalt	59-2	2536232	2514526	2529518	2526758	0.44	cps
Copper	63-2	18346478	18046420	17893315	18095405	1.27	cps
Dysprosium	156-1	120	97	97	104	12.90	cps
Dysprosium	156-2	263	310	273	282	8.71	cps
Erbium	164-1	93	117	103	104	11.21	cps
Erbium	164-2	130	117	87	111	19.97	cps
Gadolinium	160-1	130	143	103	126	16.22	cps
Gadolinium	160-2	470	430	490	463	6.59	cps
Holmium	165-1	8750222	8508293	8463343	8573953	1.80	cps
Holmium	165-2	6553669	6722147	6562647	6612821	1.43	cps
Indium	115-1	7616858	7378408	7350761	7448675	1.96	cps
Indium	115-2	3023483	3085663	2963717	3024287	2.02	cps
Iron	56-2	90995102	89624942	89208465	89942836	1.04	cps
Iron	57-2	2270904	2237476	2238473	2248951	0.85	cps
Iron	54-2	5034575	4968535	4917374	4973495	1.18	cps
Krypton	83-1	623	563	427	538	18.74	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10865977	10798317	10752633	10805643	0.53	cps
Lead	207-1	9569500	9562888	9415763	9516050	0.91	cps
Lead	208-1	43629096	43823062	43239523	43563894	0.68	cps
Lithium	6-1	740252	729475	711148	726958	2.02	cps
Magnesium	24-2	18390401	18261676	18130284	18260787	0.71	cps
Manganese	55-2	14805042	14572932	14517282	14631752	1.04	cps
Molybdenum	94-1	5316008	5301137	5182478	5266541	1.39	cps
Molybdenum	95-1	6028472	5931082	5913676	5957743	1.04	cps
Molybdenum	96-1	6636192	6717268	6647319	6666926	0.66	cps
Molybdenum	97-1	3634208	3629005	3662732	3641981	0.50	cps
Molybdenum	98-1	9527365	9289795	9322022	9379727	1.37	cps
Neodymium	150-1	153	140	167	153	8.70	cps
Neodymium	150-2	83	33	83	67	43.30	cps
Nickel	60-2	653641	645723	648666	649344	0.62	cps
Phosphorus	31-2	3607	3584	3644	3612	0.84	cps
Potassium	39-2	11496407	11240274	11319137	11351940	1.16	cps
Rhodium	103-1	7450940	7347688	7363478	7387369	0.75	cps
Rhodium	103-2	4739483	4831051	4674655	4748396	1.65	cps
Scandium	45-1	4441177	4475497	4387321	4434665	1.00	cps
Scandium	45-2	494380	492516	487953	491616	0.67	cps
Selenium	82-1	63145	63578	62839	63187	0.59	cps
Selenium	77-2	7675	7926	7745	7782	1.66	cps
Selenium	78-2	27536	27402	26744	27227	1.56	cps
Silicon	28-1	60184324	60527722	60687424	60466490	0.43	cps
Silver	107-1	733833	730640	727160	730544	0.46	cps
Silver	109-1	702355	698674	690243	697090	0.89	cps
Sodium	23-2	26347996	25937860	25223418	25836425	2.20	cps
Strontium	86-1	1835667	1828043	1853900	1839204	0.72	cps
Strontium	88-1	16229561	16236992	16008819	16158457	0.80	cps
Sulfur	34-1	422447	421798	423779	422675	0.24	cps
Terbium	159-1	8972120	8823613	8672383	8822705	1.70	cps
Terbium	159-2	6587065	6736267	6380575	6567969	2.72	cps
Thallium	203-1	2642251	2647374	2640710	2643445	0.13	cps
Thallium	205-1	6320950	6240839	6302761	6288183	0.67	cps
Tin	118-1	2264088	2280199	2333171	2292486	1.58	cps
Titanium	47-1	2214	2307	2260	2260	2.07	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-03A Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025/ Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:15:13 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8283365	8313319	7958821	8185168	2.40	cps
Vanadium	51-2	1437266	1390101	1401638	1409668	1.74	cps
Yttrium	89-1	13219145	12959275	12888035	13022151	1.34	cps
Yttrium	89-2	3833220	3924005	3766381	3841202	2.06	cps
Zinc	66-2	3447097	3393558	3346935	3395863	1.48	cps
Zirconium	90-1	5443762	5349361	5395179	5396101	0.87	cps
Zirconium	91-1	1187091	1188026	1164052	1179723	1.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4531	4964	4407	4634	6.31	cps
Antimony	121-1	1467	1573	1493	1511	3.67	cps
Arsenic	75-2	270	303	223	266	15.13	cps
Barium	135-1	663985	655633	650522	656713	1.03	cps
Barium	137-1	1142062	1136226	1118151	1132146	1.10	cps
Beryllium	9-1	34	61	54	50	27.60	cps
Bismuth	209-1	5264458	5206413	5205102	5225324	0.65	cps
Bismuth	209-2	4831423	4784554	4719953	4778643	1.17	cps
Bromine	81-1	51664	60721	64738	59041	11.34	cps
Bromine	81-2	13029	13390	13303	13240	1.42	cps
Cadmium	108-1	20	23	20	21	9.11	cps
Cadmium	106-1	3510	3167	3124	3267	6.49	cps
Cadmium	111-1	2553	2314	2289	2386	6.11	cps
Calcium	43-1	575196	567019	562967	568394	1.10	cps
Calcium	44-1	9187277	8959032	8852657	8999655	1.90	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2300	2400	2400	2367	2.44	cps
Cobalt	59-2	44162	44359	43486	44002	1.04	cps
Copper	63-2	2917	3057	2950	2975	2.46	cps
Dysprosium	156-1	30	23	23	26	15.07	cps
Dysprosium	156-2	57	37	63	52	26.57	cps
Erbium	164-1	73	73	70	72	2.66	cps
Erbium	164-2	47	33	27	36	28.64	cps
Gadolinium	160-1	50	63	53	56	12.49	cps
Gadolinium	160-2	500	490	423	471	8.85	cps
Holmium	165-1	8513822	8429716	8378207	8440581	0.81	cps
Holmium	165-2	6491218	6507327	6346530	6448358	1.37	cps
Indium	115-1	7551232	7550213	7511141	7537529	0.30	cps
Indium	115-2	3102488	3068997	3036492	3069326	1.08	cps
Iron	56-2	21259609	21955911	21504922	21573481	1.64	cps
Iron	57-2	559747	558625	554890	557754	0.46	cps
Iron	54-2	1206773	1217237	1213475	1212495	0.44	cps
Krypton	83-1	460	503	407	457	10.60	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	970	1007	1017	998	2.46	cps
Lead	207-1	823	790	850	821	3.66	cps
Lead	208-1	3857	3864	3944	3888	1.24	cps
Lithium	6-1	737126	719190	721726	726014	1.34	cps
Magnesium	24-2	3393777	3318406	3236450	3316211	2.37	cps
Manganese	55-2	1924829	1967348	1920342	1937506	1.34	cps
Molybdenum	94-1	26818	26380	26237	26478	1.14	cps
Molybdenum	95-1	46113	45383	45210	45569	1.05	cps
Molybdenum	96-1	48597	48393	47149	48047	1.63	cps
Molybdenum	97-1	29557	28040	28164	28587	2.95	cps
Molybdenum	98-1	73923	71769	71457	72383	1.86	cps
Neodymium	150-1	27	37	23	29	24.03	cps
Neodymium	150-2	3	17	13	11	62.48	cps
Nickel	60-2	14981	14687	14954	14874	1.09	cps
Phosphorus	31-2	3474	3187	3414	3358	4.50	cps
Potassium	39-2	3310274	3311003	3267637	3296305	0.75	cps
Rhodium	103-1	7566755	7418324	7338266	7441115	1.56	cps
Rhodium	103-2	4772047	4818752	4719833	4770211	1.04	cps
Scandium	45-1	4442018	4401293	4283603	4375638	1.88	cps
Scandium	45-2	486520	492137	482970	487209	0.95	cps
Selenium	82-1	193	136	251	194	29.76	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	580	530	490	533	8.46	cps
Silicon	28-1	66319284	64860382	65124367	65434678	1.19	cps
Silver	107-1	490	387	403	427	13.00	cps
Silver	109-1	383	407	403	398	3.17	cps
Sodium	23-2	3351267	3358214	3303574	3337685	0.89	cps
Strontium	86-1	709942	699227	695859	701676	1.05	cps
Strontium	88-1	6243403	6166339	6034034	6147926	1.72	cps
Sulfur	34-1	509642	506007	497210	504287	1.27	cps
Terbium	159-1	8844061	8768652	8830616	8814443	0.46	cps
Terbium	159-2	6544779	6545775	6451195	6513916	0.83	cps
Thallium	203-1	3097	3007	2564	2889	9.88	cps
Thallium	205-1	7119	6515	6178	6604	7.22	cps
Tin	118-1	1783	1697	1760	1747	2.57	cps
Titanium	47-1	593	567	660	607	7.92	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-09 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:18:02 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	380	317	320	339	10.52	cps
Vanadium	51-2	503	547	597	549	8.51	cps
Yttrium	89-1	13313044	13073992	12909880	13098972	1.55	cps
Yttrium	89-2	3816282	3847435	3836252	3833323	0.41	cps
Zinc	66-2	6398	6538	6181	6373	2.82	cps
Zirconium	90-1	2520	2237	2374	2377	5.96	cps
Zirconium	91-1	510	507	447	488	7.31	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8912314	9108563	9051285	9024054	1.12	cps
Antimony	121-1	3019830	2984254	2934507	2979530	1.44	cps
Arsenic	75-2	205274	204613	203624	204504	0.41	cps
Barium	135-1	3567154	3613437	3590369	3590320	0.64	cps
Barium	137-1	6163494	5986006	6040931	6063477	1.50	cps
Beryllium	9-1	344939	350553	350673	348722	0.94	cps
Bismuth	209-1	4561782	4542011	4504612	4536135	0.64	cps
Bismuth	209-2	4140818	4141297	4170787	4150967	0.41	cps
Bromine	81-1	21422	20318	19960	20567	3.71	cps
Bromine	81-2	1833	1627	1517	1659	9.69	cps
Cadmium	108-1	56624	56721	55981	56442	0.71	cps
Cadmium	106-1	81947	82698	81826	82157	0.58	cps
Cadmium	111-1	796521	790526	790574	792541	0.44	cps
Calcium	43-1	4122736	4114172	4036866	4091258	1.16	cps
Calcium	44-1	65628861	67188731	66390001	66402531	1.17	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1651857	1639563	1619471	1636964	1.00	cps
Cobalt	59-2	2503879	2449476	2470786	2474714	1.11	cps
Copper	63-2	19076614	18799259	18949349	18941741	0.73	cps
Dysprosium	156-1	180	167	140	162	12.56	cps
Dysprosium	156-2	253	347	307	302	15.49	cps
Erbium	164-1	440	370	427	412	9.02	cps
Erbium	164-2	313	363	310	329	9.08	cps
Gadolinium	160-1	203	240	217	220	8.43	cps
Gadolinium	160-2	460	583	610	551	14.52	cps
Holmium	165-1	7834454	7828507	7746745	7803236	0.63	cps
Holmium	165-2	5982027	5900990	5924399	5935805	0.70	cps
Indium	115-1	6773772	6630009	6576445	6660076	1.53	cps
Indium	115-2	2731078	2739675	2726357	2732370	0.25	cps
Iron	56-2	376741301	381822808	384842234	381135448	1.07	cps
Iron	57-2	9489354	9264864	9480077	9411432	1.35	cps
Iron	54-2	20958037	21123676	21101495	21061070	0.43	cps
Krypton	83-1	533	520	587	547	6.45	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10251297	10448678	10612635	10437537	1.73	cps
Lead	207-1	8950752	8982234	9005762	8979583	0.31	cps
Lead	208-1	41037739	41340882	41123853	41167491	0.38	cps
Lithium	6-1	682731	676474	693770	684325	1.28	cps
Magnesium	24-2	77391339	76874516	76315847	76860567	0.70	cps
Manganese	55-2	13615358	14076583	14023893	13905278	1.82	cps
Molybdenum	94-1	11057877	10981057	10769188	10936041	1.37	cps
Molybdenum	95-1	16248193	16226051	15912379	16128874	1.16	cps
Molybdenum	96-1	17573526	17374974	17193539	17380680	1.09	cps
Molybdenum	97-1	9886227	9818742	9884915	9863295	0.39	cps
Molybdenum	98-1	25492287	25692735	25653392	25612805	0.41	cps
Neodymium	150-1	210	200	177	196	8.75	cps
Neodymium	150-2	70	73	57	67	13.22	cps
Nickel	60-2	619497	619566	620862	619975	0.12	cps
Phosphorus	31-2	99667	99455	98955	99359	0.37	cps
Potassium	39-2	46626059	47079264	47387201	47030842	0.81	cps
Rhodium	103-1	6541531	6551987	6527863	6540460	0.18	cps
Rhodium	103-2	4178236	4194343	4257645	4210075	1.00	cps
Scandium	45-1	4128538	4017618	4094638	4080265	1.39	cps
Scandium	45-2	459792	453875	453950	455873	0.74	cps
Selenium	82-1	58328	58879	58709	58639	0.48	cps
Selenium	77-2	7242	7372	7325	7313	0.90	cps
Selenium	78-2	25125	25325	24804	25085	1.05	cps
Silicon	28-1	8003733	8035489	7858054	7965759	1.19	cps
Silver	107-1	4018972	3964323	3987391	3990228	0.69	cps
Silver	109-1	3887693	3802186	3844672	3844850	1.11	cps
Sodium	23-2	115999852	115973238	116531542	116168210	0.27	cps
Strontium	86-1	1037271	1042091	1023045	1034136	0.96	cps
Strontium	88-1	9192096	9215531	9001566	9136397	1.28	cps
Sulfur	34-1	380893	382512	381290	381565	0.22	cps
Terbium	159-1	8166812	8237676	8002889	8135792	1.48	cps
Terbium	159-2	6050663	6030402	6069909	6050325	0.33	cps
Thallium	203-1	2564602	2574243	2644688	2594511	1.69	cps
Thallium	205-1	6021168	6149454	6211083	6127235	1.58	cps
Tin	118-1	2509855	2494188	2482191	2495412	0.56	cps
Titanium	47-1	4017889	4083246	4059203	4053446	0.82	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV03 Instrumnet Name : P7
Client Sample ID : CCV03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:21:12 DataFile Name : 046CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8024445	8081044	8002447	8035978	0.50	cps
Vanadium	51-2	1412094	1422283	1408326	1414234	0.51	cps
Yttrium	89-1	12163671	12152049	11783281	12033000	1.80	cps
Yttrium	89-2	3553608	3476972	3585837	3538806	1.58	cps
Zinc	66-2	3389068	3392818	3386986	3389624	0.09	cps
Zirconium	90-1	5503225	5614998	5396883	5505035	1.98	cps
Zirconium	91-1	1203216	1197554	1199536	1200102	0.24	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1623	1567	1830	1673	8.28	cps
Antimony	121-1	857	897	817	857	4.67	cps
Arsenic	75-2	17	10	7	11	45.82	cps
Barium	135-1	153	123	113	130	16.02	cps
Barium	137-1	257	273	260	263	3.35	cps
Beryllium	9-1	20	10	7	12	55.53	cps
Bismuth	209-1	4616393	4572022	4592272	4593562	0.48	cps
Bismuth	209-2	4165459	4240965	4102898	4169774	1.66	cps
Bromine	81-1	15145	14414	14451	14670	2.81	cps
Bromine	81-2	917	840	797	851	7.14	cps
Cadmium	108-1	30	10	23	21	48.24	cps
Cadmium	106-1	2780	2947	2684	2804	4.75	cps
Cadmium	111-1	2002	2133	1949	2028	4.66	cps
Calcium	43-1	513	603	413	510	18.64	cps
Calcium	44-1	36531	35732	34867	35710	2.33	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2780	2897	2657	2778	4.32	cps
Cobalt	59-2	100	127	117	114	11.77	cps
Copper	63-2	2194	2217	2197	2202	0.57	cps
Dysprosium	156-1	0	3	3	2	86.60	cps
Dysprosium	156-2	7	7	3	6	34.70	cps
Erbium	164-1	47	90	47	61	40.93	cps
Erbium	164-2	43	27	37	36	23.58	cps
Gadolinium	160-1	57	60	43	53	16.54	cps
Gadolinium	160-2	453	373	370	399	11.83	cps
Holmium	165-1	7464594	7308506	7253052	7342051	1.49	cps
Holmium	165-2	5661215	5489362	5460177	5536918	1.96	cps
Indium	115-1	6678882	6574172	6614132	6622396	0.80	cps
Indium	115-2	2726559	2705620	2637258	2689812	1.74	cps
Iron	56-2	28621	29432	28250	28768	2.10	cps
Iron	57-2	2027	2250	2037	2105	6.00	cps
Iron	54-2	3287	3410	3454	3384	2.56	cps
Krypton	83-1	410	533	520	488	13.88	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	870	730	823	808	8.83	cps
Lead	207-1	693	690	743	709	4.21	cps
Lead	208-1	3264	3174	3237	3225	1.43	cps
Lithium	6-1	687776	672296	681138	680403	1.14	cps
Magnesium	24-2	1820	1927	1803	1850	3.62	cps
Manganese	55-2	1235	1246	1349	1277	4.92	cps
Molybdenum	94-1	783	653	660	699	10.48	cps
Molybdenum	95-1	910	767	657	778	16.33	cps
Molybdenum	96-1	877	887	880	881	0.58	cps
Molybdenum	97-1	583	577	433	531	15.96	cps
Molybdenum	98-1	1313	1207	1183	1235	5.61	cps
Neodymium	150-1	10	3	0	4	114.60	cps
Neodymium	150-2	20	7	3	10	88.20	cps
Nickel	60-2	657	557	517	577	12.51	cps
Phosphorus	31-2	683	683	610	659	6.43	cps
Potassium	39-2	92777	92033	92194	92334	0.42	cps
Rhodium	103-1	6634238	6460929	6453981	6516383	1.57	cps
Rhodium	103-2	4205466	4185360	4167341	4186056	0.46	cps
Scandium	45-1	3931866	3865467	3772260	3856531	2.08	cps
Scandium	45-2	432410	430325	429703	430813	0.33	cps
Selenium	82-1	210	58	116	128	59.90	cps
Selenium	77-2	0	3	3	2	86.60	cps
Selenium	78-2	490	527	477	498	5.20	cps
Silicon	28-1	567128	561450	559375	562651	0.71	cps
Silver	107-1	593	560	617	590	4.83	cps
Silver	109-1	583	513	567	554	6.59	cps
Sodium	23-2	64655	63731	62426	63604	1.76	cps
Strontium	86-1	900	847	830	859	4.26	cps
Strontium	88-1	893	910	957	920	3.57	cps
Sulfur	34-1	148264	148571	150597	149144	0.85	cps
Terbium	159-1	7850157	7668470	7615688	7711438	1.60	cps
Terbium	159-2	5868808	5707213	5690614	5755545	1.71	cps
Thallium	203-1	857	987	883	909	7.56	cps
Thallium	205-1	2054	2134	2097	2095	1.91	cps
Tin	118-1	2044	1907	2004	1985	3.54	cps
Titanium	47-1	130	110	93	111	16.52	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB03 Instrumnet Name : P7
Client Sample ID : CCB03 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:24:00 DataFile Name : 047CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	87	70	100	86	17.57	cps
Vanadium	51-2	33	30	30	31	6.18	cps
Yttrium	89-1	11516319	11261323	11174189	11317277	1.57	cps
Yttrium	89-2	3503235	3341682	3421904	3422273	2.36	cps
Zinc	66-2	4481	4231	4451	4387	3.11	cps
Zirconium	90-1	747	730	710	729	2.52	cps
Zirconium	91-1	133	150	130	138	7.78	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1653429	1668161	1709933	1677175	1.75	cps
Antimony	121-1	3134281	3107267	3048539	3096696	1.42	cps
Arsenic	75-2	228247	224051	229730	227343	1.30	cps
Barium	135-1	4017350	4030561	4086149	4044687	0.90	cps
Barium	137-1	6937289	7005185	6897675	6946716	0.78	cps
Beryllium	9-1	372014	372589	370066	371556	0.36	cps
Bismuth	209-1	5371348	5272054	5338299	5327234	0.95	cps
Bismuth	209-2	4851319	4846095	4729048	4808821	1.44	cps
Bromine	81-1	51557	62636	69228	61140	14.61	cps
Bromine	81-2	14514	14794	15151	14820	2.16	cps
Cadmium	108-1	58300	59147	58461	58636	0.77	cps
Cadmium	106-1	85081	86928	85145	85718	1.22	cps
Cadmium	111-1	838290	841306	828573	836056	0.80	cps
Calcium	43-1	1321646	1341213	1293699	1318852	1.81	cps
Calcium	44-1	21710894	21711152	20576696	21332914	3.07	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1669504	1643244	1648396	1653715	0.84	cps
Cobalt	59-2	2631512	2593071	2585748	2603443	0.94	cps
Copper	63-2	17873226	17844493	17964213	17893977	0.35	cps
Dysprosium	156-1	70	80	97	82	16.39	cps
Dysprosium	156-2	227	217	187	210	9.91	cps
Erbium	164-1	110	113	130	118	9.10	cps
Erbium	164-2	87	70	90	82	13.03	cps
Gadolinium	160-1	137	120	127	128	6.57	cps
Gadolinium	160-2	497	520	523	513	2.83	cps
Holmium	165-1	8663714	8648197	8607993	8639968	0.33	cps
Holmium	165-2	6529421	6492343	6333034	6451599	1.62	cps
Indium	115-1	7498089	7633670	7405867	7512542	1.53	cps
Indium	115-2	3036915	3024039	2982457	3014471	0.94	cps
Iron	56-2	87874829	85701220	86816825	86797625	1.25	cps
Iron	57-2	2143141	2211960	2178624	2177908	1.58	cps
Iron	54-2	4815112	4713870	4936451	4821811	2.31	cps
Krypton	83-1	487	563	383	478	18.91	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11384150	11249479	10940050	11191226	2.03	cps
Lead	207-1	9903733	9861149	9498392	9754425	2.28	cps
Lead	208-1	44896045	44946062	43721571	44521226	1.56	cps
Lithium	6-1	795012	790484	769860	785119	1.71	cps
Magnesium	24-2	19167628	18967879	19076472	19070660	0.52	cps
Manganese	55-2	14845051	14526215	14912126	14761131	1.40	cps
Molybdenum	94-1	5267020	5435761	5345522	5349435	1.58	cps
Molybdenum	95-1	6096883	6054269	6027242	6059465	0.58	cps
Molybdenum	96-1	6638966	6733253	6758954	6710391	0.94	cps
Molybdenum	97-1	3756149	3733139	3671414	3720234	1.18	cps
Molybdenum	98-1	9624260	9633150	9727761	9661724	0.59	cps
Neodymium	150-1	150	87	123	120	26.50	cps
Neodymium	150-2	67	60	57	61	8.33	cps
Nickel	60-2	657328	648618	645669	650538	0.93	cps
Phosphorus	31-2	3154	3120	3054	3109	1.64	cps
Potassium	39-2	11676201	11535831	11788361	11666798	1.08	cps
Rhodium	103-1	7516471	7517243	7400042	7477919	0.90	cps
Rhodium	103-2	4759958	4810435	4613545	4727980	2.16	cps
Scandium	45-1	4590510	4581163	4557339	4576337	0.37	cps
Scandium	45-2	495377	498469	493773	495873	0.48	cps
Selenium	82-1	62502	63806	62634	62981	1.14	cps
Selenium	77-2	7932	7392	7639	7654	3.53	cps
Selenium	78-2	27532	26748	26464	26915	2.06	cps
Silicon	28-1	62752416	62975152	61281222	62336264	1.48	cps
Silver	107-1	745986	736580	731996	738188	0.97	cps
Silver	109-1	708265	715521	706313	710033	0.68	cps
Sodium	23-2	27604475	26753294	27713721	27357163	1.92	cps
Strontium	86-1	1859637	1858517	1814960	1844371	1.38	cps
Strontium	88-1	16275440	16048776	16238602	16187606	0.75	cps
Sulfur	34-1	440897	440252	434508	438552	0.80	cps
Terbium	159-1	9052204	9097062	8749293	8966186	2.11	cps
Terbium	159-2	6508897	6658193	6479755	6548948	1.46	cps
Thallium	203-1	2729975	2807378	2670862	2736071	2.50	cps
Thallium	205-1	6508812	6468620	6393852	6457095	0.90	cps
Tin	118-1	2350741	2355449	2313213	2339801	0.99	cps
Titanium	47-1	2060	1953	2027	2013	2.71	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-10 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:27:19 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8296708	8209209	8330548	8278821	0.76	cps
Vanadium	51-2	1448034	1447610	1468993	1454879	0.84	cps
Yttrium	89-1	13005834	12926670	13127658	13020054	0.78	cps
Yttrium	89-2	3876012	3831557	3803648	3837072	0.95	cps
Zinc	66-2	3422460	3404551	3421979	3416330	0.30	cps
Zirconium	90-1	5539079	5494450	5438603	5490711	0.92	cps
Zirconium	91-1	1184219	1205017	1194217	1194484	0.87	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1729527	1668689	1631899	1676705	2.94	cps
Antimony	121-1	3158436	3049322	3039139	3082299	2.15	cps
Arsenic	75-2	229827	226989	227365	228061	0.68	cps
Barium	135-1	4001247	4056595	3909951	3989264	1.86	cps
Barium	137-1	6913521	6717257	6796466	6809081	1.45	cps
Beryllium	9-1	357667	359265	352592	356508	0.98	cps
Bismuth	209-1	5475013	5318382	5312013	5368469	1.72	cps
Bismuth	209-2	4606944	4804533	4930155	4780544	3.41	cps
Bromine	81-1	56542	67922	73730	66065	13.23	cps
Bromine	81-2	15635	15682	15655	15657	0.15	cps
Cadmium	108-1	56727	55707	57430	56621	1.53	cps
Cadmium	106-1	86366	84625	84689	85227	1.16	cps
Cadmium	111-1	830036	826388	823874	826766	0.37	cps
Calcium	43-1	1290135	1272352	1303977	1288821	1.23	cps
Calcium	44-1	20855001	21021865	20996593	20957820	0.43	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1663608	1642340	1631910	1645953	0.98	cps
Cobalt	59-2	2549939	2561508	2531081	2547509	0.60	cps
Copper	63-2	17880334	18131431	18317416	18109727	1.21	cps
Dysprosium	156-1	137	130	117	128	7.97	cps
Dysprosium	156-2	317	330	370	339	8.19	cps
Erbium	164-1	120	93	100	104	13.29	cps
Erbium	164-2	90	113	97	100	12.02	cps
Gadolinium	160-1	173	157	123	151	16.85	cps
Gadolinium	160-2	473	447	507	476	6.32	cps
Holmium	165-1	8725707	8657175	8627189	8670024	0.58	cps
Holmium	165-2	6306098	6532642	6806158	6548299	3.82	cps
Indium	115-1	7602056	7374919	7265188	7414054	2.32	cps
Indium	115-2	2967407	3063742	3087105	3039418	2.09	cps
Iron	56-2	86417822	86288645	84161392	85622620	1.48	cps
Iron	57-2	2142068	2175419	2130276	2149254	1.09	cps
Iron	54-2	4950401	4776486	4726177	4817688	2.44	cps
Krypton	83-1	543	533	490	522	5.43	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11214009	10864389	10944854	11007751	1.66	cps
Lead	207-1	9624929	9691797	9574683	9630470	0.61	cps
Lead	208-1	44785710	43909544	43781319	44158858	1.24	cps
Lithium	6-1	771441	752549	756791	760260	1.30	cps
Magnesium	24-2	18961346	19146664	19415916	19174642	1.19	cps
Manganese	55-2	15472266	14954931	14925984	15117727	2.03	cps
Molybdenum	94-1	5337752	5335467	5219076	5297432	1.28	cps
Molybdenum	95-1	5971549	5928378	5970680	5956869	0.41	cps
Molybdenum	96-1	6588663	6630907	6544039	6587870	0.66	cps
Molybdenum	97-1	3700643	3609770	3644613	3651675	1.26	cps
Molybdenum	98-1	9634104	9339935	9459739	9477926	1.56	cps
Neodymium	150-1	193	237	220	217	10.09	cps
Neodymium	150-2	120	80	73	91	27.70	cps
Nickel	60-2	646251	645865	650797	647638	0.42	cps
Phosphorus	31-2	2567	2477	2687	2577	4.09	cps
Potassium	39-2	11670019	11409082	11498931	11526011	1.15	cps
Rhodium	103-1	7525682	7247945	7331722	7368450	1.93	cps
Rhodium	103-2	4536470	4748028	4822682	4702393	3.16	cps
Scandium	45-1	4514652	4525443	4407137	4482411	1.46	cps
Scandium	45-2	484358	494832	508389	495860	2.43	cps
Selenium	82-1	63248	63262	62893	63134	0.33	cps
Selenium	77-2	7976	7896	7512	7794	3.18	cps
Selenium	78-2	26210	26481	27138	26610	1.79	cps
Silicon	28-1	63091361	61103811	61781162	61992111	1.63	cps
Silver	107-1	728619	727403	719011	725011	0.72	cps
Silver	109-1	704427	698059	688267	696918	1.17	cps
Sodium	23-2	27256567	27455835	27639066	27450489	0.70	cps
Strontium	86-1	1870810	1847371	1881887	1866689	0.94	cps
Strontium	88-1	16422037	16276061	15803543	16167214	2.00	cps
Sulfur	34-1	437544	435230	432672	435149	0.56	cps
Terbium	159-1	9356861	9050420	8806694	9071325	3.04	cps
Terbium	159-2	6708365	6657648	6640298	6668770	0.53	cps
Thallium	203-1	2734668	2730109	2664927	2709901	1.44	cps
Thallium	205-1	6588629	6472853	6276358	6445946	2.45	cps
Tin	118-1	2314802	2292160	2249760	2285574	1.44	cps
Titanium	47-1	2094	1887	1863	1948	6.50	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2842-11 Instrumnet Name : P7
Client Sample ID : MW-17B-55.5-08122025I Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:30:07 DataFile Name : 049SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8267740	8202547	8123724	8198004	0.88	cps
Vanadium	51-2	1461612	1442536	1465822	1456656	0.85	cps
Yttrium	89-1	13212445	12919172	13061961	13064526	1.12	cps
Yttrium	89-2	3823536	3729034	3938172	3830247	2.73	cps
Zinc	66-2	3391597	3409953	3455957	3419169	0.97	cps
Zirconium	90-1	5285977	5405609	5382853	5358146	1.19	cps
Zirconium	91-1	1199743	1184691	1173723	1186052	1.10	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4197	4261	4217	4225	0.77	cps
Antimony	121-1	1320	1493	1427	1413	6.19	cps
Arsenic	75-2	6998	6965	7269	7077	2.35	cps
Barium	135-1	98536	99666	102479	100227	2.03	cps
Barium	137-1	170012	172462	171859	171444	0.74	cps
Beryllium	9-1	11	24	17	17	38.66	cps
Bismuth	209-1	5207641	5204103	5467271	5293005	2.85	cps
Bismuth	209-2	5013611	5085745	4941185	5013514	1.44	cps
Bromine	81-1	68978	80692	81892	77187	9.24	cps
Bromine	81-2	16172	15972	16126	16090	0.65	cps
Cadmium	108-1	27	17	27	23	24.74	cps
Cadmium	106-1	3324	3184	3374	3294	2.99	cps
Cadmium	111-1	2435	2374	2480	2429	2.19	cps
Calcium	43-1	75859	77945	80452	78086	2.94	cps
Calcium	44-1	1283425	1332216	1402548	1339396	4.47	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	7922	7582	7525	7677	2.80	cps
Cobalt	59-2	6135	6418	6208	6254	2.35	cps
Copper	63-2	14087	13753	13583	13808	1.86	cps
Dysprosium	156-1	17	13	13	14	13.35	cps
Dysprosium	156-2	13	13	17	14	13.35	cps
Erbium	164-1	60	50	83	64	26.54	cps
Erbium	164-2	90	40	70	67	37.75	cps
Gadolinium	160-1	43	57	53	51	13.58	cps
Gadolinium	160-2	487	403	487	459	10.49	cps
Holmium	165-1	8539538	8480481	8708798	8576272	1.38	cps
Holmium	165-2	6632227	6439024	6616939	6562730	1.64	cps
Indium	115-1	7399555	7559402	7631191	7530049	1.57	cps
Indium	115-2	3128870	3114663	3068411	3103981	1.02	cps
Iron	56-2	68881724	67693271	67491221	68022072	1.10	cps
Iron	57-2	1693040	1670105	1634853	1666000	1.76	cps
Iron	54-2	3761475	3745273	3681896	3729548	1.13	cps
Krypton	83-1	463	547	587	532	11.82	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	34337	34437	36482	35085	3.45	cps
Lead	207-1	28777	29773	29806	29452	1.98	cps
Lead	208-1	131296	135828	138336	135153	2.64	cps
Lithium	6-1	722677	714800	766272	734583	3.77	cps
Magnesium	24-2	535549	534273	536943	535589	0.25	cps
Manganese	55-2	2596575	2610108	2527689	2578124	1.71	cps
Molybdenum	94-1	3921	2387	2417	2908	30.16	cps
Molybdenum	95-1	1413	1353	1357	1375	2.45	cps
Molybdenum	96-1	2464	2267	2094	2275	8.14	cps
Molybdenum	97-1	1030	870	770	890	14.74	cps
Molybdenum	98-1	2157	2134	1997	2096	4.12	cps
Neodymium	150-1	10	7	10	9	21.63	cps
Neodymium	150-2	3	10	3	6	69.34	cps
Nickel	60-2	4527	4464	4691	4561	2.57	cps
Phosphorus	31-2	3074	2917	3030	3007	2.69	cps
Potassium	39-2	298661	299423	301429	299838	0.48	cps
Rhodium	103-1	7360934	7411499	7201451	7324628	1.50	cps
Rhodium	103-2	4900213	4931163	4884738	4905371	0.48	cps
Scandium	45-1	4271229	4244897	4504520	4340215	3.29	cps
Scandium	45-2	498794	501789	500648	500410	0.30	cps
Selenium	82-1	4192	4659	5028	4626	9.06	cps
Selenium	77-2	687	697	867	750	13.49	cps
Selenium	78-2	2977	3084	2954	3005	2.31	cps
Silicon	28-1	3639476	4118895	4210219	3989530	7.68	cps
Silver	107-1	767	880	933	860	9.90	cps
Silver	109-1	847	873	920	880	4.22	cps
Sodium	23-2	6606879	6664731	6563168	6611593	0.77	cps
Strontium	86-1	41348	43173	42485	42335	2.18	cps
Strontium	88-1	348862	352406	352672	351314	0.61	cps
Sulfur	34-1	166240	168504	167807	167517	0.69	cps
Terbium	159-1	8576526	8641401	8820308	8679411	1.45	cps
Terbium	159-2	6676932	6749693	6633369	6686665	0.88	cps
Thallium	203-1	3754	4604	4294	4217	10.20	cps
Thallium	205-1	9473	10371	10034	9959	4.55	cps
Tin	118-1	5905	5588	5671	5721	2.87	cps
Titanium	47-1	180	203	213	199	8.60	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2849-01DLX5 Instrumnet Name : P7
Client Sample ID : BUR-25-0071DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:32:53 DataFile Name : 050SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	167	110	80	119	37.02	cps
Vanadium	51-2	140	100	120	120	16.67	cps
Yttrium	89-1	12415081	12351177	12871929	12546062	2.26	cps
Yttrium	89-2	3956169	3887065	3858944	3900726	1.28	cps
Zinc	66-2	38807	39198	38556	38854	0.83	cps
Zirconium	90-1	8249	4854	4641	5915	34.23	cps
Zirconium	91-1	1960	1013	1053	1342	39.89	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	167074	165180	163827	165360	0.99	cps
Antimony	121-1	4451	4354	4281	4362	1.96	cps
Arsenic	75-2	897	757	810	821	8.61	cps
Barium	135-1	46081	46319	46222	46207	0.26	cps
Barium	137-1	78512	79424	79038	78992	0.58	cps
Beryllium	9-1	28	31	18	26	27.70	cps
Bismuth	209-1	5361671	5283626	5226663	5290653	1.28	cps
Bismuth	209-2	4741848	4699329	4769923	4737033	0.75	cps
Bromine	81-1	38664	45761	49536	44654	12.36	cps
Bromine	81-2	9643	9460	9660	9588	1.16	cps
Cadmium	108-1	60	47	53	53	12.50	cps
Cadmium	106-1	3497	3450	3550	3499	1.43	cps
Cadmium	111-1	2803	2793	2798	2798	0.18	cps
Calcium	43-1	73145	74197	73397	73580	0.75	cps
Calcium	44-1	1199744	1185147	1200705	1195199	0.73	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	141810	141400	139845	141018	0.73	cps
Cobalt	59-2	12379	12152	12198	12243	0.98	cps
Copper	63-2	206281	205808	201744	204611	1.22	cps
Dysprosium	156-1	407	310	377	364	13.58	cps
Dysprosium	156-2	320	353	390	354	9.88	cps
Erbium	164-1	447	413	467	442	6.09	cps
Erbium	164-2	270	250	350	290	18.25	cps
Gadolinium	160-1	397	403	387	396	2.12	cps
Gadolinium	160-2	733	750	687	723	4.54	cps
Holmium	165-1	8687104	8465516	8575529	8576050	1.29	cps
Holmium	165-2	6564801	6453569	6455361	6491244	0.98	cps
Indium	115-1	7637796	7349308	7477332	7488145	1.93	cps
Indium	115-2	3047619	2998214	2999394	3015075	0.93	cps
Iron	56-2	889814707	871911854	862209454	874645338	1.60	cps
Iron	57-2	21485265	21433114	21459718	21459366	0.12	cps
Iron	54-2	48825881	47434029	47924599	48061503	1.47	cps
Krypton	83-1	523	460	570	518	10.66	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	74165	75013	74698	74625	0.57	cps
Lead	207-1	61392	59983	61610	60995	1.45	cps
Lead	208-1	286464	290587	287790	288280	0.73	cps
Lithium	6-1	772769	778880	769690	773779	0.60	cps
Magnesium	24-2	168359	165485	165738	166527	0.96	cps
Manganese	55-2	2062109	2025130	1970961	2019400	2.27	cps
Molybdenum	94-1	14854	14718	15115	14895	1.35	cps
Molybdenum	95-1	20758	20404	20248	20470	1.28	cps
Molybdenum	96-1	28381	28024	28197	28201	0.63	cps
Molybdenum	97-1	12993	12702	12422	12706	2.25	cps
Molybdenum	98-1	32500	32289	31975	32255	0.82	cps
Neodymium	150-1	590	590	587	589	0.33	cps
Neodymium	150-2	370	417	357	381	8.27	cps
Nickel	60-2	9076	9093	8869	9013	1.38	cps
Phosphorus	31-2	1050	1083	1110	1081	2.78	cps
Potassium	39-2	2583555	2532793	2563188	2559845	1.00	cps
Rhodium	103-1	7522752	7231470	7328617	7360946	2.01	cps
Rhodium	103-2	4586298	4643758	4593333	4607796	0.68	cps
Scandium	45-1	4485265	4390264	4339197	4404908	1.68	cps
Scandium	45-2	490073	489219	487838	489043	0.23	cps
Selenium	82-1	1614	1548	1346	1503	9.28	cps
Selenium	77-2	123	147	150	140	10.38	cps
Selenium	78-2	953	977	947	959	1.64	cps
Silicon	28-1	4711835	4783122	4720261	4738406	0.82	cps
Silver	107-1	960	880	1067	969	9.67	cps
Silver	109-1	883	960	713	852	14.81	cps
Sodium	23-2	109852812	109414572	108413255	109226879	0.68	cps
Strontium	86-1	41488	40633	41639	41253	1.32	cps
Strontium	88-1	349965	358628	349292	352628	1.48	cps
Sulfur	34-1	207814	211149	211349	210104	0.95	cps
Terbium	159-1	9115276	8814618	8779318	8903071	2.07	cps
Terbium	159-2	6690902	6465860	6468545	6541769	1.97	cps
Thallium	203-1	1213	1100	1033	1116	8.16	cps
Thallium	205-1	2740	2690	2620	2684	2.25	cps
Tin	118-1	31889	32073	32654	32205	1.24	cps
Titanium	47-1	10640	10607	10644	10630	0.19	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01 Instrumnet Name : P7
Client Sample ID : #452 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:36:09 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	360	433	327	373	14.62	cps
Vanadium	51-2	6635	6738	6885	6753	1.86	cps
Yttrium	89-1	13065221	12925513	12733851	12908195	1.29	cps
Yttrium	89-2	3860949	3830497	3849603	3847017	0.40	cps
Zinc	66-2	351830	353604	351077	352170	0.37	cps
Zirconium	90-1	7739	7775	7735	7750	0.29	cps
Zirconium	91-1	1737	1693	1800	1743	3.08	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	55895	55269	54483	55216	1.28	cps
Antimony	121-1	38565	37255	34826	36882	5.14	cps
Arsenic	75-2	10244	10294	9747	10095	3.00	cps
Barium	135-1	46586	44025	40812	43807	6.60	cps
Barium	137-1	80396	74923	70085	75135	6.87	cps
Beryllium	9-1	8	18	11	12	42.20	cps
Bismuth	209-1	4053116	3680685	3638062	3790621	6.02	cps
Bismuth	209-2	3668968	3481473	3506385	3552275	2.87	cps
Bromine	81-1	3745846	4583200	4907125	4412057	13.58	cps
Bromine	81-2	1055769	1061577	1053263	1056869	0.40	cps
Cadmium	108-1	16439	15001	13647	15029	9.29	cps
Cadmium	106-1	2600	2434	2354	2462	5.11	cps
Cadmium	111-1	9642	9074	8251	8989	7.78	cps
Calcium	43-1	29222	27892	26303	27806	5.26	cps
Calcium	44-1	605368	583978	549467	579604	4.87	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	20818	20334	20691	20614	1.22	cps
Cobalt	59-2	1477	1493	1340	1437	5.86	cps
Copper	63-2	666830	670447	665380	667552	0.39	cps
Dysprosium	156-1	37	53	37	42	22.78	cps
Dysprosium	156-2	53	47	37	46	18.41	cps
Erbium	164-1	93	80	80	84	9.11	cps
Erbium	164-2	73	77	53	68	18.62	cps
Gadolinium	160-1	87	83	83	84	2.28	cps
Gadolinium	160-2	413	413	340	389	10.89	cps
Holmium	165-1	6878064	5970684	6095756	6314835	7.79	cps
Holmium	165-2	5182372	4829630	4905234	4972412	3.73	cps
Indium	115-1	5880089	5217332	5327791	5475071	6.49	cps
Indium	115-2	2419174	2279422	2216111	2304902	4.51	cps
Iron	56-2	1058214	1046181	1028501	1044298	1.43	cps
Iron	57-2	27412	27786	27752	27650	0.75	cps
Iron	54-2	60746	60304	60200	60416	0.48	cps
Krypton	83-1	430	473	557	487	13.23	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	70441	66402	61455	66099	6.81	cps
Lead	207-1	61341	57417	54177	57645	6.22	cps
Lead	208-1	278417	262507	246271	262398	6.13	cps
Lithium	6-1	621758	556391	570153	582768	5.91	cps
Magnesium	24-2	183937	186210	185931	185360	0.67	cps
Manganese	55-2	227568	228709	229231	228503	0.37	cps
Molybdenum	94-1	71557594	68059766	63806057	67807806	5.72	cps
Molybdenum	95-1	128790145	120884845	113038292	120904427	6.51	cps
Molybdenum	96-1	137278318	129231135	122271878	129593777	5.79	cps
Molybdenum	97-1	79972484	73771486	69245891	74329953	7.24	cps
Molybdenum	98-1	206777417	193773517	177934457	192828464	7.49	cps
Neodymium	150-1	70	87	60	72	18.66	cps
Neodymium	150-2	63	43	60	56	19.29	cps
Nickel	60-2	2110	2247	1927	2095	7.67	cps
Phosphorus	31-2	42559	42770	43141	42823	0.69	cps
Potassium	39-2	296251136	293957129	292723242	294310502	0.61	cps
Rhodium	103-1	5676246	5176091	5177758	5343365	5.40	cps
Rhodium	103-2	3748735	3533703	3423433	3568624	4.64	cps
Scandium	45-1	3479984	3139725	3146353	3255354	5.98	cps
Scandium	45-2	399475	376972	372043	382830	3.82	cps
Selenium	82-1	5195	5430	5306	5310	2.22	cps
Selenium	77-2	553	683	560	599	12.22	cps
Selenium	78-2	2514	2447	2357	2439	3.22	cps
Silicon	28-1	23266271	25297750	25901500	24821840	5.56	cps
Silver	107-1	3550	3114	3060	3242	8.30	cps
Silver	109-1	3374	3280	3034	3229	5.44	cps
Sodium	23-2	159599918	158667688	154049191	157438932	1.89	cps
Strontium	86-1	129440	121447	115179	122022	5.86	cps
Strontium	88-1	1124263	1056513	992161	1057645	6.25	cps
Sulfur	34-1	294027	272287	255577	273964	7.04	cps
Terbium	159-1	7076957	6283460	6423612	6594676	6.42	cps
Terbium	159-2	5248491	4940003	4800096	4996197	4.59	cps
Thallium	203-1	1783	1853	1974	1870	5.14	cps
Thallium	205-1	4304	4457	4557	4440	2.87	cps
Tin	118-1	26595	24775	23222	24864	6.79	cps
Titanium	47-1	3367	3100	3160	3209	4.36	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DLX5 Instrumnet Name : P7
Client Sample ID : #456DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:39:23 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	360	383	327	357	7.98	cps
Vanadium	51-2	4587	4634	4574	4599	0.69	cps
Yttrium	89-1	10317405	9283304	9315012	9638574	6.10	cps
Yttrium	89-2	3155122	2914618	2879911	2983217	5.02	cps
Zinc	66-2	484675	483899	479721	482765	0.55	cps
Zirconium	90-1	2877	2647	2434	2652	8.36	cps
Zirconium	91-1	653	583	560	599	8.11	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1720743	1677697	1673750	1690730	1.54	cps
Antimony	121-1	4561	4484	4784	4610	3.38	cps
Arsenic	75-2	9136	8623	8626	8795	3.36	cps
Barium	135-1	34442	34849	34468	34586	0.66	cps
Barium	137-1	59340	60903	59451	59898	1.46	cps
Beryllium	9-1	16	16	9	14	27.10	cps
Bismuth	209-1	3882853	3902083	3902271	3895736	0.29	cps
Bismuth	209-2	3917068	3724495	3514876	3718813	5.41	cps
Bromine	81-1	4663601	6121854	6676762	5820739	17.86	cps
Bromine	81-2	1537568	1540526	1567148	1548414	1.05	cps
Cadmium	108-1	3234	3517	3050	3267	7.20	cps
Cadmium	106-1	2540	2554	2630	2575	1.89	cps
Cadmium	111-1	4095	3733	4177	4002	5.90	cps
Calcium	43-1	108636	109385	110432	109484	0.82	cps
Calcium	44-1	1840235	1885474	1919703	1881804	2.12	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	18685	18859	18278	18607	1.60	cps
Cobalt	59-2	800	823	920	848	7.51	cps
Copper	63-2	1033143	1003748	1000583	1012491	1.77	cps
Dysprosium	156-1	13	30	30	24	39.37	cps
Dysprosium	156-2	10	33	23	22	52.68	cps
Erbium	164-1	53	70	63	62	13.48	cps
Erbium	164-2	57	40	30	42	31.91	cps
Gadolinium	160-1	53	67	60	60	11.12	cps
Gadolinium	160-2	460	373	363	399	13.33	cps
Holmium	165-1	6424841	6559178	6481172	6488397	1.04	cps
Holmium	165-2	5589977	5167255	4841373	5199535	7.22	cps
Indium	115-1	5551445	5623729	5483536	5552904	1.26	cps
Indium	115-2	2543906	2398800	2286894	2409867	5.35	cps
Iron	56-2	1295439	1240555	1240997	1258997	2.51	cps
Iron	57-2	35555	34797	35011	35121	1.11	cps
Iron	54-2	74236	71596	71710	72514	2.06	cps
Krypton	83-1	440	447	470	452	3.48	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7269	7549	7576	7464	2.28	cps
Lead	207-1	6302	6598	6598	6499	2.64	cps
Lead	208-1	28836	30394	30027	29752	2.74	cps
Lithium	6-1	603407	586701	600535	596881	1.50	cps
Magnesium	24-2	183001	177976	179101	180026	1.46	cps
Manganese	55-2	1146624	1103799	1094084	1114836	2.51	cps
Molybdenum	94-1	14390569	14644489	14939464	14658174	1.87	cps
Molybdenum	95-1	25922799	26493320	26610476	26342198	1.40	cps
Molybdenum	96-1	27535250	28123340	27572620	27743737	1.19	cps
Molybdenum	97-1	16087748	16362435	15876017	16108733	1.51	cps
Molybdenum	98-1	40581233	42466774	41037719	41361909	2.38	cps
Neodymium	150-1	13	37	13	21	63.83	cps
Neodymium	150-2	23	10	13	16	44.60	cps
Nickel	60-2	2594	2544	2294	2477	6.49	cps
Phosphorus	31-2	3174	3247	2924	3115	5.44	cps
Potassium	39-2	248342403	237355803	236793263	240830490	2.70	cps
Rhodium	103-1	5466303	5428458	5387701	5427487	0.72	cps
Rhodium	103-2	3915833	3666396	3488525	3690252	5.82	cps
Scandium	45-1	3342087	3312643	3457072	3370601	2.26	cps
Scandium	45-2	416281	395569	379912	397254	4.59	cps
Selenium	82-1	5516	6294	6444	6085	8.19	cps
Selenium	77-2	557	767	700	674	15.91	cps
Selenium	78-2	2960	2760	2880	2867	3.51	cps
Silicon	28-1	13103516	15436736	16433690	14991314	11.40	cps
Silver	107-1	413	447	353	404	11.69	cps
Silver	109-1	407	443	420	423	4.38	cps
Sodium	23-2	132530098	128409541	124939131	128626257	2.95	cps
Strontium	86-1	49861	50219	50616	50232	0.75	cps
Strontium	88-1	423845	427825	432287	427985	0.99	cps
Sulfur	34-1	370283	374160	370910	371784	0.56	cps
Terbium	159-1	6551802	6685832	6715312	6650982	1.31	cps
Terbium	159-2	5505398	5136242	4933901	5191847	5.58	cps
Thallium	203-1	2344	2410	2454	2402	2.31	cps
Thallium	205-1	5658	5481	5578	5572	1.59	cps
Tin	118-1	23846	24404	23376	23875	2.16	cps
Titanium	47-1	1453	1590	1550	1531	4.59	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DLX5 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:42:38 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	800	890	920	870	7.18	cps
Vanadium	51-2	11051	11004	10320	10792	3.79	cps
Yttrium	89-1	9676981	9649105	9697962	9674683	0.25	cps
Yttrium	89-2	3187367	3030770	2969393	3062510	3.67	cps
Zinc	66-2	1167577	1130374	1122529	1140160	2.11	cps
Zirconium	90-1	3427	3644	3674	3582	3.76	cps
Zirconium	91-1	760	757	800	772	3.12	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	17463	17370	17230	17354	0.68	cps
Antimony	121-1	14147	14474	14297	14306	1.14	cps
Arsenic	75-2	7485	7532	7739	7585	1.78	cps
Barium	135-1	32744	31618	32036	32133	1.77	cps
Barium	137-1	55729	57506	55802	56346	1.78	cps
Beryllium	9-1	24	20	7	17	52.01	cps
Bismuth	209-1	5334469	5160169	5368747	5287795	2.12	cps
Bismuth	209-2	4792629	4682825	4757061	4744172	1.18	cps
Bromine	81-1	796619	754981	707411	753004	5.93	cps
Bromine	81-2	136683	133779	128325	132929	3.19	cps
Cadmium	108-1	53	53	83	63	27.35	cps
Cadmium	106-1	3127	3277	3230	3212	2.39	cps
Cadmium	111-1	2352	2400	2353	2368	1.16	cps
Calcium	43-1	170808	170400	173073	171427	0.84	cps
Calcium	44-1	2934370	2766979	2774260	2825203	3.35	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	6975	7028	7098	7034	0.88	cps
Cobalt	59-2	4881	4657	4614	4717	3.03	cps
Copper	63-2	11695	10981	11618	11431	3.43	cps
Dysprosium	156-1	10	17	10	12	31.50	cps
Dysprosium	156-2	10	23	17	17	39.99	cps
Erbium	164-1	103	53	57	71	39.32	cps
Erbium	164-2	40	40	57	46	21.13	cps
Gadolinium	160-1	73	63	67	68	7.51	cps
Gadolinium	160-2	423	500	370	431	15.16	cps
Holmium	165-1	8581571	8297476	8703784	8527610	2.44	cps
Holmium	165-2	6528658	6402918	6344776	6425451	1.46	cps
Indium	115-1	7519878	7223794	7405403	7383025	2.02	cps
Indium	115-2	3016126	2985885	3040931	3014314	0.91	cps
Iron	56-2	158545928	160726684	156575434	158616015	1.31	cps
Iron	57-2	3853618	3890803	3882385	3875602	0.50	cps
Iron	54-2	8784929	8654173	8808662	8749255	0.95	cps
Krypton	83-1	413	507	507	476	11.33	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	5515	5298	5284	5366	2.41	cps
Lead	207-1	4311	4424	4501	4412	2.17	cps
Lead	208-1	20199	20676	20516	20464	1.19	cps
Lithium	6-1	812090	781796	793765	795884	1.92	cps
Magnesium	24-2	479337	480047	476044	478476	0.45	cps
Manganese	55-2	742299	744654	732870	739941	0.84	cps
Molybdenum	94-1	7682	7062	6778	7174	6.44	cps
Molybdenum	95-1	12005	10854	10397	11085	7.47	cps
Molybdenum	96-1	13930	13059	12349	13113	6.04	cps
Molybdenum	97-1	7472	6848	6552	6957	6.75	cps
Molybdenum	98-1	19313	16823	16616	17584	8.53	cps
Neodymium	150-1	27	10	23	20	44.10	cps
Neodymium	150-2	7	7	20	11	69.25	cps
Nickel	60-2	5765	5588	5594	5649	1.77	cps
Phosphorus	31-2	269882	269322	268931	269378	0.18	cps
Potassium	39-2	93431159	93009275	90976279	92472238	1.42	cps
Rhodium	103-1	7402657	7363163	7550069	7438630	1.32	cps
Rhodium	103-2	4746256	4615423	4723412	4695030	1.49	cps
Scandium	45-1	4624326	4364828	4485685	4491613	2.89	cps
Scandium	45-2	499377	488871	492505	493584	1.08	cps
Selenium	82-1	3969	4271	4284	4175	4.27	cps
Selenium	77-2	580	523	553	552	5.13	cps
Selenium	78-2	2327	2357	2317	2334	0.89	cps
Silicon	28-1	6607966	6433063	6481605	6507545	1.39	cps
Silver	107-1	487	563	487	512	8.64	cps
Silver	109-1	450	447	460	452	1.53	cps
Sodium	23-2	5938077	5874683	5850080	5887613	0.77	cps
Strontium	86-1	60293	60307	59851	60150	0.43	cps
Strontium	88-1	505427	513646	510172	509748	0.81	cps
Sulfur	34-1	285457	295405	294494	291785	1.88	cps
Terbium	159-1	8701881	9002745	9187943	8964190	2.74	cps
Terbium	159-2	6581057	6512896	6506362	6533438	0.63	cps
Thallium	203-1	873	897	1033	934	9.25	cps
Thallium	205-1	2040	2247	2200	2162	5.01	cps
Tin	118-1	3921	3807	3957	3895	2.01	cps
Titanium	47-1	4044	3657	3981	3894	5.33	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-01DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2005DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:45:45 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	217	160	130	169	26.06	cps
Vanadium	51-2	500	450	500	483	5.97	cps
Yttrium	89-1	12990509	12740388	13086936	12939278	1.38	cps
Yttrium	89-2	3836187	3886217	3813396	3845267	0.97	cps
Zinc	66-2	353616	355180	351294	353363	0.55	cps
Zirconium	90-1	2114	2144	2023	2094	2.98	cps
Zirconium	91-1	440	493	480	471	5.89	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	15214	15068	14927	15070	0.95	cps
Antimony	121-1	2394	2207	2257	2286	4.23	cps
Arsenic	75-2	4251	4101	4244	4198	2.02	cps
Barium	135-1	13383	13006	13253	13214	1.45	cps
Barium	137-1	23379	22905	23156	23147	1.02	cps
Beryllium	9-1	7	13	3	8	65.13	cps
Bismuth	209-1	5374886	5320317	5372576	5355926	0.58	cps
Bismuth	209-2	4849758	4891632	4790069	4843820	1.05	cps
Bromine	81-1	153677	147636	136465	145926	5.98	cps
Bromine	81-2	24874	24273	22808	23985	4.43	cps
Cadmium	108-1	7	27	7	13	86.58	cps
Cadmium	106-1	3194	3094	3237	3175	2.32	cps
Cadmium	111-1	2291	2194	2305	2263	2.68	cps
Calcium	43-1	30391	30835	30214	30480	1.05	cps
Calcium	44-1	542575	548839	545588	545667	0.57	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	10023	10000	9837	9953	1.02	cps
Cobalt	59-2	1097	887	1010	998	10.58	cps
Copper	63-2	13723	13109	13199	13344	2.48	cps
Dysprosium	156-1	13	3	13	10	57.75	cps
Dysprosium	156-2	3	17	23	14	70.51	cps
Erbium	164-1	80	60	93	78	21.57	cps
Erbium	164-2	53	63	43	53	18.75	cps
Gadolinium	160-1	33	53	40	42	24.12	cps
Gadolinium	160-2	530	443	503	492	9.02	cps
Holmium	165-1	8587291	8404267	8321495	8437684	1.61	cps
Holmium	165-2	6607045	6421181	6417063	6481763	1.67	cps
Indium	115-1	7575625	7508305	7375388	7486439	1.36	cps
Indium	115-2	3123066	3077704	3023970	3074913	1.61	cps
Iron	56-2	98926735	99880422	98555022	99120726	0.69	cps
Iron	57-2	2455464	2457304	2427924	2446897	0.67	cps
Iron	54-2	5461132	5638835	5424500	5508156	2.08	cps
Krypton	83-1	463	490	463	472	3.26	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	3187	3240	3250	3226	1.06	cps
Lead	207-1	2717	2730	2840	2763	2.45	cps
Lead	208-1	12559	12719	12749	12676	0.81	cps
Lithium	6-1	783164	778794	788139	783366	0.60	cps
Magnesium	24-2	75692	76858	76804	76452	0.86	cps
Manganese	55-2	2762094	2792641	2769552	2774762	0.57	cps
Molybdenum	94-1	1707	1710	1737	1718	0.96	cps
Molybdenum	95-1	2174	2080	2147	2134	2.25	cps
Molybdenum	96-1	3004	3120	3230	3118	3.64	cps
Molybdenum	97-1	1490	1447	1363	1433	4.49	cps
Molybdenum	98-1	3527	3387	3217	3377	4.60	cps
Neodymium	150-1	7	23	17	16	53.90	cps
Neodymium	150-2	17	13	13	14	13.35	cps
Nickel	60-2	1610	1730	1563	1635	5.26	cps
Phosphorus	31-2	28320	28423	27875	28206	1.03	cps
Potassium	39-2	10207754	10161297	10177291	10182114	0.23	cps
Rhodium	103-1	7468398	7442220	7340179	7416932	0.91	cps
Rhodium	103-2	4884353	4830331	4740554	4818413	1.51	cps
Scandium	45-1	4424248	4500176	4406352	4443592	1.12	cps
Scandium	45-2	494377	507104	484446	495309	2.29	cps
Selenium	82-1	2879	3026	2986	2964	2.57	cps
Selenium	77-2	427	460	403	430	6.62	cps
Selenium	78-2	2000	1870	1740	1870	6.95	cps
Silicon	28-1	3241396	3387488	3464693	3364526	3.37	cps
Silver	107-1	587	743	897	742	20.88	cps
Silver	109-1	633	743	870	749	15.82	cps
Sodium	23-2	1127269	1161422	1112721	1133804	2.20	cps
Strontium	86-1	6495	6602	6461	6519	1.12	cps
Strontium	88-1	49109	50359	48925	49464	1.58	cps
Sulfur	34-1	151953	155283	154101	153779	1.10	cps
Terbium	159-1	8978316	8819932	8898045	8898764	0.89	cps
Terbium	159-2	6617466	6599120	6477652	6564746	1.16	cps
Thallium	203-1	690	693	773	719	6.56	cps
Thallium	205-1	1700	1907	1747	1785	6.07	cps
Tin	118-1	4808	4724	4804	4779	0.99	cps
Titanium	47-1	893	913	773	860	8.80	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2867-02DLX5 Instrumnet Name : P7
Client Sample ID : Tre-2006DL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 15:49:06 DataFile Name : 055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	20	20	13	18	21.66	cps
Vanadium	51-2	253	253	263	257	2.25	cps
Yttrium	89-1	13111323	12843367	12838053	12930915	1.21	cps
Yttrium	89-2	3980671	3951849	3779863	3904128	2.78	cps
Zinc	66-2	28648	27479	28668	28265	2.41	cps
Zirconium	90-1	1203	1297	1270	1257	3.83	cps
Zirconium	91-1	283	310	350	314	10.67	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3080	2874	2847	2934	4.35	cps
Antimony	121-1	453	383	410	416	8.50	cps
Arsenic	75-2	963	853	823	880	8.38	cps
Barium	135-1	307114	300970	299354	302479	1.35	cps
Barium	137-1	530434	521549	522800	524927	0.92	cps
Beryllium	9-1	11	18	21	17	30.38	cps
Bismuth	209-1	5554202	5594057	5562069	5570110	0.38	cps
Bismuth	209-2	5084371	4934109	4875752	4964744	2.17	cps
Bromine	81-1	88879	93273	96368	92840	4.05	cps
Bromine	81-2	18969	18455	18148	18524	2.24	cps
Cadmium	108-1	33	13	10	19	66.81	cps
Cadmium	106-1	3517	3550	3320	3463	3.59	cps
Cadmium	111-1	2475	2506	2361	2447	3.13	cps
Calcium	43-1	1212111	1182009	1196498	1196873	1.26	cps
Calcium	44-1	19589747	19453600	19090559	19377969	1.33	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2887	2980	2757	2875	3.90	cps
Cobalt	59-2	10270	10220	9880	10123	2.10	cps
Copper	63-2	2100	2070	2064	2078	0.94	cps
Dysprosium	156-1	67	33	37	46	40.30	cps
Dysprosium	156-2	57	57	43	52	14.75	cps
Erbium	164-1	110	90	90	97	11.95	cps
Erbium	164-2	57	83	93	78	24.37	cps
Gadolinium	160-1	87	87	83	86	2.25	cps
Gadolinium	160-2	507	490	440	479	7.24	cps
Holmium	165-1	9268335	8915860	8761984	8982060	2.89	cps
Holmium	165-2	6782424	6547569	6571588	6633860	1.95	cps
Indium	115-1	8015076	7881247	7903624	7933315	0.90	cps
Indium	115-2	3204125	3219504	3155824	3193151	1.04	cps
Iron	56-2	73525442	73520594	72323194	73123077	0.95	cps
Iron	57-2	1862406	1848407	1828277	1846363	0.93	cps
Iron	54-2	4070116	4080431	4075913	4075487	0.13	cps
Krypton	83-1	400	460	543	468	15.39	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	73434	71947	71913	72431	1.20	cps
Lead	207-1	59068	57308	57368	57915	1.73	cps
Lead	208-1	279332	273529	272565	275142	1.33	cps
Lithium	6-1	840565	837877	796932	825125	2.96	cps
Magnesium	24-2	7634193	7696591	7600984	7643923	0.64	cps
Manganese	55-2	3794717	3852523	3879012	3842084	1.12	cps
Molybdenum	94-1	1763	1720	1710	1731	1.64	cps
Molybdenum	95-1	2687	2844	2700	2744	3.17	cps
Molybdenum	96-1	3397	3244	3354	3332	2.37	cps
Molybdenum	97-1	1587	1630	1567	1595	2.03	cps
Molybdenum	98-1	4184	4194	4064	4147	1.75	cps
Neodymium	150-1	77	87	50	71	26.66	cps
Neodymium	150-2	47	70	50	56	22.71	cps
Nickel	60-2	2207	2120	2127	2151	2.24	cps
Phosphorus	31-2	2997	2907	2660	2855	6.11	cps
Potassium	39-2	4175796	4201751	4097689	4158412	1.30	cps
Rhodium	103-1	7954219	7903619	7887091	7914976	0.44	cps
Rhodium	103-2	5080628	4797347	4877299	4918425	2.97	cps
Scandium	45-1	4910421	4826926	4691004	4809451	2.30	cps
Scandium	45-2	517841	516205	518523	517523	0.23	cps
Selenium	82-1	1631	1548	1312	1497	11.04	cps
Selenium	77-2	177	137	143	152	14.08	cps
Selenium	78-2	950	977	1043	990	4.86	cps
Silicon	28-1	70966327	69823179	67397286	69395597	2.63	cps
Silver	107-1	197	143	197	179	17.21	cps
Silver	109-1	140	123	147	137	8.79	cps
Sodium	23-2	6280922	6291825	6234475	6269074	0.49	cps
Strontium	86-1	1769809	1773369	1744539	1762572	0.89	cps
Strontium	88-1	15119693	15000021	14891687	15003800	0.76	cps
Sulfur	34-1	1844224	1837871	1783596	1821897	1.83	cps
Terbium	159-1	9651171	9298821	9164284	9371425	2.68	cps
Terbium	159-2	6913166	6833946	6840217	6862443	0.64	cps
Thallium	203-1	367	370	350	362	2.96	cps
Thallium	205-1	980	807	820	869	11.10	cps
Tin	118-1	12392	12489	12055	12312	1.85	cps
Titanium	47-1	610	587	573	590	3.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-01 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:52:19 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	357	350	317	341	6.28	cps
Vanadium	51-2	287	283	257	276	5.97	cps
Yttrium	89-1	14111809	13886614	13927027	13975150	0.86	cps
Yttrium	89-2	4202476	4018198	4035495	4085390	2.49	cps
Zinc	66-2	4601	4614	4797	4671	2.35	cps
Zirconium	90-1	567	617	593	592	4.23	cps
Zirconium	91-1	113	150	133	132	13.88	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3247	3174	3207	3209	1.14	cps
Antimony	121-1	240	277	220	246	11.70	cps
Arsenic	75-2	457	503	513	491	6.16	cps
Barium	135-1	259894	256341	252802	256346	1.38	cps
Barium	137-1	451886	443387	441395	445556	1.25	cps
Beryllium	9-1	5	18	5	9	83.87	cps
Bismuth	209-1	5489019	5268287	5200336	5319214	2.84	cps
Bismuth	209-2	4871257	4851215	4782930	4835134	0.96	cps
Bromine	81-1	73586	81802	84192	79860	6.97	cps
Bromine	81-2	17070	17510	16960	17180	1.70	cps
Cadmium	108-1	23	7	10	13	66.12	cps
Cadmium	106-1	3504	3227	3300	3344	4.29	cps
Cadmium	111-1	2463	2301	2320	2361	3.75	cps
Calcium	43-1	1154477	1140925	1136582	1143995	0.82	cps
Calcium	44-1	18331733	18168773	17817713	18106073	1.45	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2507	2377	2570	2485	3.97	cps
Cobalt	59-2	9280	9350	9310	9313	0.38	cps
Copper	63-2	1720	1747	1680	1716	1.96	cps
Dysprosium	156-1	43	37	30	37	18.18	cps
Dysprosium	156-2	50	60	70	60	16.67	cps
Erbium	164-1	90	70	67	76	16.70	cps
Erbium	164-2	90	47	87	74	32.39	cps
Gadolinium	160-1	73	67	90	77	15.68	cps
Gadolinium	160-2	513	490	477	493	3.76	cps
Holmium	165-1	8839608	8693071	8661007	8731229	1.09	cps
Holmium	165-2	6575038	6602317	6453283	6543546	1.21	cps
Indium	115-1	7787752	7683585	7584072	7685136	1.33	cps
Indium	115-2	3144460	3149404	3106849	3133571	0.74	cps
Iron	56-2	70641456	70019386	68962116	69874319	1.22	cps
Iron	57-2	1770699	1754025	1703106	1742610	2.02	cps
Iron	54-2	3882966	3739994	3755998	3792986	2.07	cps
Krypton	83-1	430	453	467	450	4.12	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	820	807	853	827	2.91	cps
Lead	207-1	750	757	717	741	2.89	cps
Lead	208-1	3230	3127	3214	3190	1.74	cps
Lithium	6-1	797909	784010	779211	787043	1.23	cps
Magnesium	24-2	7373907	7223399	7220744	7272684	1.21	cps
Manganese	55-2	3771064	3697983	3613462	3694170	2.13	cps
Molybdenum	94-1	1837	1793	1843	1825	1.49	cps
Molybdenum	95-1	2840	2880	2967	2896	2.24	cps
Molybdenum	96-1	3547	3477	3520	3515	1.01	cps
Molybdenum	97-1	1803	1743	1987	1845	6.87	cps
Molybdenum	98-1	4717	4524	4551	4597	2.28	cps
Neodymium	150-1	107	80	50	79	35.94	cps
Neodymium	150-2	40	53	50	48	14.52	cps
Nickel	60-2	1737	1567	1593	1632	5.60	cps
Phosphorus	31-2	3040	3020	2707	2923	6.40	cps
Potassium	39-2	4067118	4019798	3971978	4019632	1.18	cps
Rhodium	103-1	7707711	7607635	7408236	7574527	2.01	cps
Rhodium	103-2	4855473	4813593	4635968	4768345	2.44	cps
Scandium	45-1	4715122	4670642	4533333	4639699	2.04	cps
Scandium	45-2	510965	501787	497849	503533	1.34	cps
Selenium	82-1	1191	1169	1085	1148	4.86	cps
Selenium	77-2	117	97	103	106	9.65	cps
Selenium	78-2	930	930	820	893	7.11	cps
Silicon	28-1	65126232	66199556	66885527	66070438	1.34	cps
Silver	107-1	143	123	203	157	26.57	cps
Silver	109-1	127	83	130	113	22.97	cps
Sodium	23-2	5937222	5792998	5897027	5875749	1.27	cps
Strontium	86-1	1656197	1632771	1619363	1636110	1.14	cps
Strontium	88-1	14339689	14466525	14007679	14271297	1.66	cps
Sulfur	34-1	1764745	1805046	1768427	1779406	1.25	cps
Terbium	159-1	9249174	9074436	9142559	9155390	0.96	cps
Terbium	159-2	6622535	6602838	6532988	6586120	0.71	cps
Thallium	203-1	277	273	260	270	3.27	cps
Thallium	205-1	710	650	633	664	6.07	cps
Tin	118-1	2754	2734	2610	2699	2.88	cps
Titanium	47-1	533	443	597	524	14.69	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-02 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:55:29 DataFile Name : 057SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	413	383	480	426	11.63	cps
Vanadium	51-2	260	273	277	270	3.27	cps
Yttrium	89-1	13702434	13345159	13308839	13452144	1.62	cps
Yttrium	89-2	3937015	3920028	3876876	3911306	0.79	cps
Zinc	66-2	4091	4017	3874	3994	2.76	cps
Zirconium	90-1	517	557	470	514	8.43	cps
Zirconium	91-1	130	137	70	112	32.72	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9346306	9334254	9399743	9360101	0.37	cps
Antimony	121-1	3068995	3078304	3122496	3089931	0.93	cps
Arsenic	75-2	215478	213768	213601	214283	0.48	cps
Barium	135-1	3639711	3734766	3673484	3682654	1.31	cps
Barium	137-1	6384078	6272331	6295564	6317324	0.93	cps
Beryllium	9-1	371197	376326	376017	374513	0.77	cps
Bismuth	209-1	4918000	4936981	4880747	4911909	0.58	cps
Bismuth	209-2	4427316	4447462	4369563	4414780	0.92	cps
Bromine	81-1	38052	37875	35931	37286	3.16	cps
Bromine	81-2	5158	4964	4964	5029	2.22	cps
Cadmium	108-1	57962	59060	58270	58431	0.97	cps
Cadmium	106-1	83901	84062	85205	84389	0.84	cps
Cadmium	111-1	813590	815646	812061	813766	0.22	cps
Calcium	43-1	4226991	4325869	4352048	4301636	1.53	cps
Calcium	44-1	69787161	70817767	69845751	70150226	0.83	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1704503	1740233	1648416	1697717	2.73	cps
Cobalt	59-2	2609987	2560636	2563598	2578074	1.07	cps
Copper	63-2	19562866	19264915	18960657	19262813	1.56	cps
Dysprosium	156-1	170	230	153	184	21.86	cps
Dysprosium	156-2	400	283	360	348	17.05	cps
Erbium	164-1	350	403	360	371	7.64	cps
Erbium	164-2	290	350	273	304	13.24	cps
Gadolinium	160-1	240	193	157	197	21.24	cps
Gadolinium	160-2	607	577	523	569	7.42	cps
Holmium	165-1	8192305	8254431	8104063	8183600	0.92	cps
Holmium	165-2	6306879	6215606	6173343	6231942	1.10	cps
Indium	115-1	6944018	7014270	7013397	6990561	0.58	cps
Indium	115-2	2892413	2888134	2897372	2892640	0.16	cps
Iron	56-2	401423914	397236421	392772247	397144194	1.09	cps
Iron	57-2	9815815	9987801	9839429	9881015	0.94	cps
Iron	54-2	22418836	21391730	21901167	21903911	2.34	cps
Krypton	83-1	487	503	513	501	2.69	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10841892	10996889	10798506	10879096	0.96	cps
Lead	207-1	9449973	9504877	9339639	9431496	0.89	cps
Lead	208-1	43270976	44107401	42627430	43335269	1.71	cps
Lithium	6-1	747502	733226	746501	742410	1.07	cps
Magnesium	24-2	82185700	80747740	78722784	80552075	2.16	cps
Manganese	55-2	14502842	14541956	14476408	14507069	0.23	cps
Molybdenum	94-1	11382823	11349489	11474874	11402395	0.57	cps
Molybdenum	95-1	16420043	16514187	16360596	16431609	0.47	cps
Molybdenum	96-1	18047193	18337981	17847442	18077538	1.36	cps
Molybdenum	97-1	10188044	10216519	10112451	10172338	0.53	cps
Molybdenum	98-1	26704101	26627846	26754140	26695362	0.24	cps
Neodymium	150-1	173	143	207	174	18.16	cps
Neodymium	150-2	73	80	83	79	6.45	cps
Nickel	60-2	648251	643894	640146	644097	0.63	cps
Phosphorus	31-2	106312	105065	104495	105291	0.88	cps
Potassium	39-2	48315749	48449453	48181906	48315703	0.28	cps
Rhodium	103-1	6882658	6831883	6883395	6865978	0.43	cps
Rhodium	103-2	4421585	4351651	4393426	4388887	0.80	cps
Scandium	45-1	4349245	4292614	4337109	4326323	0.69	cps
Scandium	45-2	485307	483784	482156	483749	0.33	cps
Selenium	82-1	61886	62039	61883	61936	0.14	cps
Selenium	77-2	7662	8099	7919	7893	2.78	cps
Selenium	78-2	26083	26667	26223	26325	1.16	cps
Silicon	28-1	8601353	8928320	8975622	8835099	2.31	cps
Silver	107-1	4046056	4132617	4071999	4083557	1.09	cps
Silver	109-1	3915746	3910608	3906062	3910805	0.12	cps
Sodium	23-2	123229878	120326562	123458455	122338298	1.43	cps
Strontium	86-1	1056101	1055985	1055010	1055699	0.06	cps
Strontium	88-1	9323713	9475879	9424748	9408113	0.82	cps
Sulfur	34-1	391336	391258	389871	390821	0.21	cps
Terbium	159-1	8661210	8496724	8585539	8581158	0.96	cps
Terbium	159-2	6396779	6413134	6220557	6343490	1.68	cps
Thallium	203-1	2665823	2671154	2759547	2698841	1.95	cps
Thallium	205-1	6504683	6359696	6466692	6443690	1.17	cps
Tin	118-1	2566352	2551013	2591185	2569517	0.79	cps
Titanium	47-1	4208675	4267855	4259113	4245214	0.75	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV04 Instrumnet Name : P7
Client Sample ID : CCV04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 15:58:35 DataFile Name : 058CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8282646	8571376	8156804	8336942	2.55	cps
Vanadium	51-2	1435747	1481175	1417276	1444732	2.28	cps
Yttrium	89-1	12430813	12601479	12689949	12574080	1.05	cps
Yttrium	89-2	3762241	3753845	3715275	3743787	0.67	cps
Zinc	66-2	3531224	3561280	3553521	3548675	0.44	cps
Zirconium	90-1	5645597	5696575	5648635	5663602	0.50	cps
Zirconium	91-1	1239558	1234480	1240267	1238102	0.25	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1533	1673	1400	1536	8.90	cps
Antimony	121-1	887	757	783	809	8.49	cps
Arsenic	75-2	120	87	87	98	19.68	cps
Barium	135-1	130	137	110	126	11.05	cps
Barium	137-1	217	240	237	231	5.46	cps
Beryllium	9-1	3	0	3	2	84.32	cps
Bismuth	209-1	4962850	4883286	4895815	4913984	0.87	cps
Bismuth	209-2	4409105	4409979	4396785	4405290	0.17	cps
Bromine	81-1	18672	18145	18191	18336	1.59	cps
Bromine	81-2	1737	1867	1907	1837	4.84	cps
Cadmium	108-1	17	23	20	20	16.65	cps
Cadmium	106-1	2740	2540	2524	2601	4.64	cps
Cadmium	111-1	1927	1812	1783	1840	4.14	cps
Calcium	43-1	457	463	473	464	1.81	cps
Calcium	44-1	30337	29532	30557	30142	1.79	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3027	2997	2917	2980	1.91	cps
Cobalt	59-2	100	77	90	89	13.17	cps
Copper	63-2	2324	2337	2354	2338	0.64	cps
Dysprosium	156-1	7	10	10	9	21.63	cps
Dysprosium	156-2	7	3	3	4	43.40	cps
Erbium	164-1	50	70	67	62	17.22	cps
Erbium	164-2	50	17	30	32	52.06	cps
Gadolinium	160-1	43	20	20	28	48.49	cps
Gadolinium	160-2	377	357	307	347	10.40	cps
Holmium	165-1	8012827	7802333	7754523	7856561	1.75	cps
Holmium	165-2	5812976	5759714	5769474	5780721	0.49	cps
Indium	115-1	6827808	6834903	6810711	6824474	0.18	cps
Indium	115-2	2750005	2772586	2727755	2750115	0.82	cps
Iron	56-2	26079	26150	25916	26048	0.46	cps
Iron	57-2	1830	1953	1920	1901	3.36	cps
Iron	54-2	3094	3274	3064	3144	3.61	cps
Krypton	83-1	573	477	447	499	13.27	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	820	783	787	797	2.54	cps
Lead	207-1	573	590	600	588	2.29	cps
Lead	208-1	2850	2820	2893	2855	1.29	cps
Lithium	6-1	667511	674688	656113	666104	1.41	cps
Magnesium	24-2	1230	1367	1297	1298	5.27	cps
Manganese	55-2	1189	1172	1009	1123	8.86	cps
Molybdenum	94-1	287	367	373	342	14.09	cps
Molybdenum	95-1	330	357	390	359	8.38	cps
Molybdenum	96-1	437	427	477	447	5.92	cps
Molybdenum	97-1	270	257	307	278	9.32	cps
Molybdenum	98-1	537	500	583	540	7.73	cps
Neodymium	150-1	13	13	10	12	15.73	cps
Neodymium	150-2	3	3	7	4	43.40	cps
Nickel	60-2	673	690	630	664	4.66	cps
Phosphorus	31-2	597	570	557	574	3.55	cps
Potassium	39-2	86142	87097	85505	86248	0.93	cps
Rhodium	103-1	6765229	6714541	6568409	6682727	1.53	cps
Rhodium	103-2	4263698	4338492	4319737	4307309	0.90	cps
Scandium	45-1	3799980	3790535	3787143	3792553	0.18	cps
Scandium	45-2	435774	431121	426547	431147	1.07	cps
Selenium	82-1	346	454	541	447	21.80	cps
Selenium	77-2	13	37	37	29	46.64	cps
Selenium	78-2	600	577	607	594	2.65	cps
Silicon	28-1	553905	557334	561530	557590	0.68	cps
Silver	107-1	587	603	727	639	11.97	cps
Silver	109-1	630	610	733	658	10.06	cps
Sodium	23-2	54198	53305	52864	53456	1.27	cps
Strontium	86-1	817	920	877	871	5.96	cps
Strontium	88-1	880	783	780	814	6.97	cps
Sulfur	34-1	140266	142066	144475	142269	1.48	cps
Terbium	159-1	8344686	8166587	8094182	8201818	1.57	cps
Terbium	159-2	5920538	5833695	5909441	5887891	0.80	cps
Thallium	203-1	180	180	183	181	1.06	cps
Thallium	205-1	450	470	363	428	13.25	cps
Tin	118-1	2657	3204	2874	2911	9.46	cps
Titanium	47-1	123	93	107	108	13.95	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB04 Instrumnet Name : P7
Client Sample ID : CCB04 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:10:16 DataFile Name : 061CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	23	20	13	19	26.96	cps
Vanadium	51-2	33	60	33	42	36.47	cps
Yttrium	89-1	11827769	11476151	11406623	11570181	1.95	cps
Yttrium	89-2	3460377	3336074	3490405	3428952	2.39	cps
Zinc	66-2	4481	4507	4527	4505	0.52	cps
Zirconium	90-1	350	460	383	398	14.18	cps
Zirconium	91-1	77	77	83	79	4.87	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2500	2998	3703	3067	19.71	cps
Antimony	121-1	603	600	587	597	1.48	cps
Arsenic	75-2	267	187	237	230	17.57	cps
Barium	135-1	296967	285571	287631	290057	2.09	cps
Barium	137-1	515095	501754	500196	505682	1.62	cps
Beryllium	9-1	1	4	14	7	106.17	cps
Bismuth	209-1	5524150	5507417	5584962	5538843	0.74	cps
Bismuth	209-2	4891673	4985244	4863348	4913422	1.30	cps
Bromine	81-1	52708	60945	64189	59281	9.98	cps
Bromine	81-2	12796	12906	12859	12853	0.43	cps
Cadmium	108-1	23	10	17	17	39.99	cps
Cadmium	106-1	2870	3067	3227	3055	5.85	cps
Cadmium	111-1	2026	2167	2284	2159	5.98	cps
Calcium	43-1	1180604	1184655	1179088	1181449	0.24	cps
Calcium	44-1	18762814	18661707	18588480	18671001	0.47	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4651	4871	4791	4771	2.33	cps
Cobalt	59-2	10464	10480	10304	10416	0.94	cps
Copper	63-2	2340	2410	2514	2421	3.60	cps
Dysprosium	156-1	40	53	43	46	15.23	cps
Dysprosium	156-2	73	73	77	74	2.59	cps
Erbium	164-1	120	97	70	96	26.18	cps
Erbium	164-2	77	87	87	83	6.93	cps
Gadolinium	160-1	103	63	73	80	26.02	cps
Gadolinium	160-2	500	477	457	478	4.54	cps
Holmium	165-1	8646843	8644403	8868431	8719893	1.48	cps
Holmium	165-2	6628415	6722777	6649575	6666922	0.74	cps
Indium	115-1	7610219	7787098	7829519	7742278	1.50	cps
Indium	115-2	3124398	3125510	3086000	3111969	0.72	cps
Iron	56-2	74310847	74001512	73297589	73869983	0.70	cps
Iron	57-2	1875302	1852476	1883368	1870382	0.86	cps
Iron	54-2	4121278	4087094	4036769	4081714	1.04	cps
Krypton	83-1	497	467	470	478	3.44	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	757	863	827	816	6.64	cps
Lead	207-1	640	727	617	661	8.77	cps
Lead	208-1	2937	3107	3150	3065	3.68	cps
Lithium	6-1	756566	757282	746346	753398	0.81	cps
Magnesium	24-2	7714002	7611603	7717932	7681179	0.78	cps
Manganese	55-2	3859213	3866147	3861470	3862277	0.09	cps
Molybdenum	94-1	3417	3377	3540	3445	2.47	cps
Molybdenum	95-1	5738	5658	5651	5682	0.85	cps
Molybdenum	96-1	6812	6318	6548	6559	3.76	cps
Molybdenum	97-1	3547	3444	3484	3492	1.49	cps
Molybdenum	98-1	9266	9120	9183	9190	0.80	cps
Neodymium	150-1	90	97	67	84	18.65	cps
Neodymium	150-2	23	40	47	37	32.79	cps
Nickel	60-2	1593	1690	1617	1633	3.09	cps
Phosphorus	31-2	2924	2980	2787	2897	3.43	cps
Potassium	39-2	4146157	4070547	4014615	4077106	1.62	cps
Rhodium	103-1	7366759	7421272	7532805	7440279	1.14	cps
Rhodium	103-2	4929162	4876001	4883931	4896365	0.59	cps
Scandium	45-1	4299005	4447007	4377385	4374465	1.69	cps
Scandium	45-2	500559	496703	498156	498473	0.39	cps
Selenium	82-1	453	433	413	433	4.62	cps
Selenium	77-2	17	33	50	33	50.00	cps
Selenium	78-2	663	603	613	627	5.13	cps
Silicon	28-1	67123881	66192127	65225342	66180450	1.43	cps
Silver	107-1	310	317	213	280	20.65	cps
Silver	109-1	327	270	370	322	15.56	cps
Sodium	23-2	6205610	6180055	6355842	6247169	1.52	cps
Strontium	86-1	1765602	1731938	1694968	1730836	2.04	cps
Strontium	88-1	15228379	15235961	15007921	15157420	0.85	cps
Sulfur	34-1	1851678	1824239	1841296	1839071	0.75	cps
Terbium	159-1	9282260	9300717	9339248	9307408	0.31	cps
Terbium	159-2	6598087	6809122	6646288	6684499	1.65	cps
Thallium	203-1	153	167	113	144	19.22	cps
Thallium	205-1	357	373	427	386	9.48	cps
Tin	118-1	2877	2674	2834	2795	3.83	cps
Titanium	47-1	560	517	617	564	8.89	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-05 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:13:37 DataFile Name : 062SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	320	327	300	316	4.40	cps
Vanadium	51-2	327	253	297	292	12.62	cps
Yttrium	89-1	12833978	13200349	13030449	13021592	1.41	cps
Yttrium	89-2	3979987	3881425	3919262	3926891	1.27	cps
Zinc	66-2	3917	4297	4131	4115	4.63	cps
Zirconium	90-1	683	587	660	643	7.84	cps
Zirconium	91-1	133	150	147	143	6.15	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3570	3594	3500	3555	1.37	cps
Antimony	121-1	677	753	690	707	5.80	cps
Arsenic	75-2	200	217	243	220	9.94	cps
Barium	135-1	272249	272308	265646	270068	1.42	cps
Barium	137-1	468371	468694	459698	465588	1.10	cps
Beryllium	9-1	5	12	8	9	38.94	cps
Bismuth	209-1	5460819	5356155	5359048	5392008	1.11	cps
Bismuth	209-2	4855835	4860871	4782431	4833046	0.91	cps
Bromine	81-1	50045	58479	63141	57222	11.60	cps
Bromine	81-2	12585	12242	12148	12325	1.87	cps
Cadmium	108-1	17	7	7	10	57.72	cps
Cadmium	106-1	3487	3140	3174	3267	5.86	cps
Cadmium	111-1	2473	2220	2264	2319	5.82	cps
Calcium	43-1	1138361	1112715	1110657	1120577	1.38	cps
Calcium	44-1	18260925	17461305	17255551	17659260	3.01	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2474	2544	2717	2578	4.86	cps
Cobalt	59-2	9343	9430	9213	9329	1.17	cps
Copper	63-2	2454	2437	2210	2367	5.74	cps
Dysprosium	156-1	50	50	47	49	3.93	cps
Dysprosium	156-2	77	60	60	66	14.68	cps
Erbium	164-1	90	87	80	86	5.95	cps
Erbium	164-2	73	50	67	63	18.98	cps
Gadolinium	160-1	77	87	93	86	9.80	cps
Gadolinium	160-2	413	483	417	438	9.02	cps
Holmium	165-1	9061768	8603589	8512830	8726063	3.37	cps
Holmium	165-2	6626127	6659354	6594616	6626699	0.49	cps
Indium	115-1	7883482	7651498	7631155	7722045	1.82	cps
Indium	115-2	3036185	3069819	3029800	3045268	0.71	cps
Iron	56-2	69364577	68716337	68185011	68755308	0.86	cps
Iron	57-2	1732588	1765440	1751828	1749952	0.94	cps
Iron	54-2	3756789	3757702	3774026	3762839	0.26	cps
Krypton	83-1	457	443	477	459	3.66	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	910	967	843	907	6.81	cps
Lead	207-1	737	807	840	794	6.64	cps
Lead	208-1	3717	3687	3567	3657	2.17	cps
Lithium	6-1	769959	756338	737962	754753	2.13	cps
Magnesium	24-2	7092202	7224050	7095755	7137335	1.05	cps
Manganese	55-2	3595741	3612876	3598401	3602340	0.26	cps
Molybdenum	94-1	4191	3977	3944	4037	3.32	cps
Molybdenum	95-1	6852	6958	6892	6901	0.78	cps
Molybdenum	96-1	7742	7699	7585	7675	1.05	cps
Molybdenum	97-1	4271	4131	4164	4188	1.75	cps
Molybdenum	98-1	10674	11171	10284	10710	4.15	cps
Neodymium	150-1	70	80	70	73	7.87	cps
Neodymium	150-2	77	50	57	61	22.71	cps
Nickel	60-2	1653	1730	1567	1650	4.95	cps
Phosphorus	31-2	3047	3064	2794	2968	5.10	cps
Potassium	39-2	3985178	3972977	4046374	4001510	0.98	cps
Rhodium	103-1	7730836	7549806	7562614	7614419	1.33	cps
Rhodium	103-2	4788802	4801154	4812857	4800938	0.25	cps
Scandium	45-1	4580490	4415741	4360526	4452252	2.57	cps
Scandium	45-2	493692	489029	485005	489242	0.89	cps
Selenium	82-1	357	458	394	403	12.69	cps
Selenium	77-2	33	47	23	34	34.00	cps
Selenium	78-2	710	733	520	654	17.88	cps
Silicon	28-1	64273424	63053071	64275957	63867484	1.10	cps
Silver	107-1	293	260	320	291	10.33	cps
Silver	109-1	287	243	203	244	17.05	cps
Sodium	23-2	6341739	6166257	6303486	6270494	1.47	cps
Strontium	86-1	1662809	1610526	1618532	1630622	1.73	cps
Strontium	88-1	14181014	14299866	14242293	14241058	0.42	cps
Sulfur	34-1	1726743	1741920	1728489	1732384	0.48	cps
Terbium	159-1	9259089	9172226	9103503	9178273	0.85	cps
Terbium	159-2	6584136	6636167	6470798	6563700	1.29	cps
Thallium	203-1	170	147	130	149	13.49	cps
Thallium	205-1	363	370	397	377	4.68	cps
Tin	118-1	6835	7095	6735	6888	2.70	cps
Titanium	47-1	590	583	600	591	1.42	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-06 Instrumnet Name : P7
Client Sample ID : MW-19B-71.5-081325-FL Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:16:45 DataFile Name : 063SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	637	630	607	624	2.52	cps
Vanadium	51-2	313	310	257	293	10.84	cps
Yttrium	89-1	13321326	13250004	13174526	13248619	0.55	cps
Yttrium	89-2	3864177	3912839	3785699	3854238	1.66	cps
Zinc	66-2	4974	5011	4791	4925	2.39	cps
Zirconium	90-1	623	630	543	599	8.05	cps
Zirconium	91-1	103	170	103	126	30.66	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2504	2497	2400	2467	2.34	cps
Antimony	121-1	300	257	290	282	8.04	cps
Arsenic	75-2	20	43	33	32	36.33	cps
Barium	135-1	117	167	117	133	21.65	cps
Barium	137-1	243	263	223	243	8.22	cps
Beryllium	9-1	10	4	13	9	55.68	cps
Bismuth	209-1	5621737	5536908	5582965	5580537	0.76	cps
Bismuth	209-2	4924182	4892860	5013361	4943467	1.26	cps
Bromine	81-1	23402	22554	21993	22650	3.13	cps
Bromine	81-2	2320	2244	2160	2241	3.57	cps
Cadmium	108-1	17	7	13	12	41.65	cps
Cadmium	106-1	3187	3084	3167	3146	1.74	cps
Cadmium	111-1	2253	2187	2244	2228	1.60	cps
Calcium	43-1	623	590	537	583	7.49	cps
Calcium	44-1	34573	34245	34539	34452	0.52	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2427	2277	2364	2356	3.20	cps
Cobalt	59-2	110	100	83	98	13.78	cps
Copper	63-2	2400	2424	2227	2350	4.57	cps
Dysprosium	156-1	10	7	10	9	21.63	cps
Dysprosium	156-2	7	13	7	9	43.25	cps
Erbium	164-1	57	93	57	69	30.73	cps
Erbium	164-2	53	33	53	47	24.75	cps
Gadolinium	160-1	53	57	40	50	17.64	cps
Gadolinium	160-2	460	413	423	432	5.68	cps
Holmium	165-1	9057651	8543388	8710514	8770518	2.99	cps
Holmium	165-2	6497013	6544080	6577593	6539562	0.62	cps
Indium	115-1	7975427	7742723	7760988	7826379	1.65	cps
Indium	115-2	3087529	3079390	3140099	3102339	1.06	cps
Iron	56-2	123076	119468	119760	120768	1.66	cps
Iron	57-2	4147	4331	4541	4340	4.54	cps
Iron	54-2	8606	8396	8062	8355	3.28	cps
Krypton	83-1	403	477	453	444	8.43	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	897	863	863	874	2.20	cps
Lead	207-1	763	740	810	771	4.62	cps
Lead	208-1	3677	3490	3640	3602	2.75	cps
Lithium	6-1	777135	765239	756917	766430	1.33	cps
Magnesium	24-2	1660	1647	1717	1675	2.22	cps
Manganese	55-2	51641	49185	49841	50222	2.53	cps
Molybdenum	94-1	2647	2844	2727	2739	3.61	cps
Molybdenum	95-1	4911	4684	4481	4692	4.59	cps
Molybdenum	96-1	5218	5424	5148	5263	2.73	cps
Molybdenum	97-1	3120	2904	2944	2989	3.86	cps
Molybdenum	98-1	7649	7409	7539	7532	1.60	cps
Neodymium	150-1	7	3	10	7	50.03	cps
Neodymium	150-2	3	3	7	4	43.40	cps
Nickel	60-2	670	727	590	662	10.37	cps
Phosphorus	31-2	820	707	860	796	10.00	cps
Potassium	39-2	167469	167567	165673	166903	0.64	cps
Rhodium	103-1	7902075	7656550	7557509	7705378	2.30	cps
Rhodium	103-2	4861632	4808190	4829121	4832981	0.56	cps
Scandium	45-1	4531161	4495551	4409729	4478814	1.39	cps
Scandium	45-2	493729	494736	495654	494706	0.19	cps
Selenium	82-1	398	336	333	356	10.32	cps
Selenium	77-2	13	43	10	22	82.62	cps
Selenium	78-2	493	627	557	559	11.93	cps
Silicon	28-1	573159	578948	577279	576462	0.52	cps
Silver	107-1	377	400	493	423	14.58	cps
Silver	109-1	340	330	450	373	17.83	cps
Sodium	23-2	65431	65247	64287	64988	0.95	cps
Strontium	86-1	920	973	1003	966	4.37	cps
Strontium	88-1	1743	1643	1710	1699	3.00	cps
Sulfur	34-1	145515	147898	149069	147494	1.23	cps
Terbium	159-1	9487868	9336431	8827875	9217391	3.75	cps
Terbium	159-2	6689400	6508408	6721246	6639685	1.73	cps
Thallium	203-1	137	183	100	140	29.83	cps
Thallium	205-1	310	257	247	271	12.56	cps
Tin	118-1	6358	6361	6698	6473	3.02	cps
Titanium	47-1	93	163	83	113	38.47	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-07 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:19:54 DataFile Name : 064SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	80	67	27	58	48.03	cps
Vanadium	51-2	47	33	37	39	17.85	cps
Yttrium	89-1	13513788	13155616	13008702	13226035	1.96	cps
Yttrium	89-2	3778526	3820683	3905715	3834975	1.69	cps
Zinc	66-2	4941	4941	5028	4970	1.01	cps
Zirconium	90-1	473	547	537	519	7.66	cps
Zirconium	91-1	113	87	93	98	14.19	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2597	2434	2337	2456	5.35	cps
Antimony	121-1	250	227	187	221	14.49	cps
Arsenic	75-2	27	30	23	27	12.51	cps
Barium	135-1	157	147	97	133	24.11	cps
Barium	137-1	210	253	213	226	10.69	cps
Beryllium	9-1	27	0	13	13	99.00	cps
Bismuth	209-1	5427749	5342648	5514535	5428311	1.58	cps
Bismuth	209-2	4860268	4804512	4832255	4832345	0.58	cps
Bromine	81-1	18245	18388	17841	18158	1.56	cps
Bromine	81-2	1827	1707	1470	1668	10.88	cps
Cadmium	108-1	33	20	13	22	45.83	cps
Cadmium	106-1	3480	3107	3234	3274	5.80	cps
Cadmium	111-1	2446	2184	2281	2304	5.75	cps
Calcium	43-1	527	490	593	537	9.76	cps
Calcium	44-1	33537	33824	32909	33424	1.40	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2630	2677	2340	2549	7.16	cps
Cobalt	59-2	83	133	97	104	24.79	cps
Copper	63-2	1763	1747	1890	1800	4.36	cps
Dysprosium	156-1	7	10	23	13	66.12	cps
Dysprosium	156-2	3	7	3	4	43.40	cps
Erbium	164-1	87	70	53	70	23.81	cps
Erbium	164-2	37	43	30	37	18.18	cps
Gadolinium	160-1	43	53	27	41	32.76	cps
Gadolinium	160-2	483	413	410	436	9.51	cps
Holmium	165-1	8644937	8648559	8348908	8547468	2.01	cps
Holmium	165-2	6399163	6335454	6275215	6336611	0.98	cps
Indium	115-1	7673058	7428603	7578453	7560038	1.63	cps
Indium	115-2	3071583	3030347	2931507	3011146	2.39	cps
Iron	56-2	43710	42584	42029	42774	2.00	cps
Iron	57-2	2364	2224	2344	2310	3.28	cps
Iron	54-2	4124	3861	3961	3982	3.34	cps
Krypton	83-1	393	507	487	462	13.09	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	750	883	737	790	10.27	cps
Lead	207-1	580	663	620	621	6.71	cps
Lead	208-1	2943	3047	2930	2973	2.15	cps
Lithium	6-1	761017	739724	741433	747391	1.58	cps
Magnesium	24-2	2337	2577	2424	2446	4.97	cps
Manganese	55-2	23103	22703	23136	22981	1.05	cps
Molybdenum	94-1	1023	1000	1033	1019	1.68	cps
Molybdenum	95-1	1457	1427	1413	1432	1.55	cps
Molybdenum	96-1	1607	1687	1583	1626	3.33	cps
Molybdenum	97-1	927	933	867	909	4.04	cps
Molybdenum	98-1	2337	2364	2374	2358	0.80	cps
Neodymium	150-1	3	13	3	7	86.65	cps
Neodymium	150-2	0	0	3	1	173.21	cps
Nickel	60-2	640	587	517	581	10.64	cps
Phosphorus	31-2	697	577	653	642	9.46	cps
Potassium	39-2	100902	100462	100680	100681	0.22	cps
Rhodium	103-1	7561263	7438676	7459783	7486574	0.88	cps
Rhodium	103-2	4786223	4677918	4601580	4688574	1.98	cps
Scandium	45-1	4409295	4403591	4364408	4392431	0.56	cps
Scandium	45-2	485795	486461	482098	484785	0.48	cps
Selenium	82-1	387	266	280	311	21.25	cps
Selenium	77-2	10	17	20	16	32.73	cps
Selenium	78-2	553	603	547	568	5.45	cps
Silicon	28-1	580580	581617	595472	585890	1.42	cps
Silver	107-1	313	280	277	290	6.99	cps
Silver	109-1	243	290	290	274	9.82	cps
Sodium	23-2	52713	52322	51807	52281	0.87	cps
Strontium	86-1	890	933	810	878	7.13	cps
Strontium	88-1	1177	1057	1120	1118	5.37	cps
Sulfur	34-1	148968	149631	150543	149714	0.53	cps
Terbium	159-1	8978297	8889595	8794621	8887504	1.03	cps
Terbium	159-2	6523149	6504634	6224012	6417265	2.61	cps
Thallium	203-1	73	100	80	84	16.44	cps
Thallium	205-1	263	273	270	269	1.89	cps
Tin	118-1	4414	4521	4261	4399	2.97	cps
Titanium	47-1	73	113	113	100	23.10	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2869-08 Instrumnet Name : P7
Client Sample ID : EB01-081325 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:23:11 DataFile Name : 065SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	40	53	23	39	38.65	cps
Vanadium	51-2	53	33	67	51	32.83	cps
Yttrium	89-1	13054658	13018353	12838885	12970632	0.89	cps
Yttrium	89-2	3913491	3782944	3716465	3804300	2.63	cps
Zinc	66-2	4964	4764	4621	4783	3.61	cps
Zirconium	90-1	483	467	400	450	9.80	cps
Zirconium	91-1	103	80	70	84	20.25	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	708460	705257	701877	705198	0.47	cps
Antimony	121-1	1173	1163	1130	1156	1.96	cps
Arsenic	75-2	760	807	753	773	3.76	cps
Barium	135-1	137155	141863	139205	139408	1.69	cps
Barium	137-1	245002	245085	250989	247025	1.39	cps
Beryllium	9-1	1041	1121	1194	1119	6.85	cps
Bismuth	209-1	5016129	4853462	5025203	4964931	1.95	cps
Bismuth	209-2	4669463	4552125	4662845	4628145	1.42	cps
Bromine	81-1	21346	22514	23813	22557	5.47	cps
Bromine	81-2	3037	3114	3024	3058	1.59	cps
Cadmium	108-1	340	337	317	331	3.81	cps
Cadmium	106-1	3364	3664	3550	3526	4.30	cps
Cadmium	111-1	6861	7260	6894	7005	3.16	cps
Calcium	43-1	340721	334256	336122	337033	0.99	cps
Calcium	44-1	5314898	5389570	5273874	5326114	1.10	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	48575	47829	47525	47976	1.13	cps
Cobalt	59-2	69494	68101	67806	68467	1.32	cps
Copper	63-2	65068	64090	65844	65001	1.35	cps
Dysprosium	156-1	14798	15429	15102	15109	2.09	cps
Dysprosium	156-2	14341	14671	14381	14464	1.25	cps
Erbium	164-1	18055	17511	17801	17789	1.53	cps
Erbium	164-2	13050	13727	13957	13578	3.47	cps
Gadolinium	160-1	16667	16400	16396	16488	0.94	cps
Gadolinium	160-2	13570	13690	14027	13762	1.72	cps
Holmium	165-1	8090585	8322258	8012693	8141845	1.98	cps
Holmium	165-2	6313265	6087769	6213928	6204987	1.82	cps
Indium	115-1	6880297	7084795	7005110	6990067	1.47	cps
Indium	115-2	2899579	2877038	2964242	2913620	1.55	cps
Iron	56-2	25103428	24639223	25246803	24996485	1.27	cps
Iron	57-2	645304	640283	649733	645106	0.73	cps
Iron	54-2	1416704	1397273	1366475	1393484	1.82	cps
Krypton	83-1	563	490	490	514	8.23	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	124582	128576	127047	126735	1.59	cps
Lead	207-1	106347	109278	111596	109074	2.41	cps
Lead	208-1	494964	495750	504532	498415	1.07	cps
Lithium	6-1	695979	694793	682112	690961	1.11	cps
Magnesium	24-2	3187235	3080138	3068986	3112120	2.10	cps
Manganese	55-2	703197	702605	709764	705189	0.56	cps
Molybdenum	94-1	16419	15542	16253	16071	2.90	cps
Molybdenum	95-1	12235	12362	12572	12390	1.37	cps
Molybdenum	96-1	14454	15281	15535	15090	3.75	cps
Molybdenum	97-1	7639	7735	7539	7638	1.29	cps
Molybdenum	98-1	19770	20211	19273	19751	2.38	cps
Neodymium	150-1	23059	22528	22762	22783	1.17	cps
Neodymium	150-2	14691	14941	14918	14850	0.93	cps
Nickel	60-2	16206	16523	16516	16415	1.10	cps
Phosphorus	31-2	36684	36093	37058	36612	1.33	cps
Potassium	39-2	2047711	2054959	2057586	2053418	0.25	cps
Rhodium	103-1	6945522	7126481	7147447	7073150	1.57	cps
Rhodium	103-2	4500222	4557489	4480098	4512603	0.89	cps
Scandium	45-1	4087842	4143099	4083693	4104878	0.81	cps
Scandium	45-2	469321	458418	465491	464410	1.19	cps
Selenium	82-1	360	477	409	415	14.10	cps
Selenium	77-2	83	117	110	103	17.07	cps
Selenium	78-2	693	743	657	698	6.23	cps
Silicon	28-1	47878583	49890916	49804248	49191249	2.31	cps
Silver	107-1	383	343	413	380	9.24	cps
Silver	109-1	287	343	343	324	10.08	cps
Sodium	23-2	12571171	12281929	12298876	12383992	1.31	cps
Strontium	86-1	153576	157000	156881	155819	1.25	cps
Strontium	88-1	1336592	1320256	1330297	1329049	0.62	cps
Sulfur	34-1	319125	315607	312472	315734	1.05	cps
Terbium	159-1	8365416	8589370	8361511	8438766	1.55	cps
Terbium	159-2	6360408	6295872	6207094	6287791	1.22	cps
Thallium	203-1	717	817	777	770	6.54	cps
Thallium	205-1	1950	1940	2020	1970	2.21	cps
Tin	118-1	5498	5741	5584	5608	2.20	cps
Titanium	47-1	29369	30691	30267	30109	2.24	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2840-03DLX5 Instrumnet Name : P7
Client Sample ID : 0804-EDL Dilution Factor : 5
Date & Time Acquired : 2025-08-15 16:26:31 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	23511	23718	24590	23940	2.39	cps
Vanadium	51-2	227244	226145	227119	226836	0.27	cps
Yttrium	89-1	12428889	12384108	12098529	12303842	1.46	cps
Yttrium	89-2	3825036	3739190	3667917	3744048	2.10	cps
Zinc	66-2	67097	66682	66220	66666	0.66	cps
Zirconium	90-1	24848	25025	23893	24589	2.48	cps
Zirconium	91-1	5668	5358	5434	5487	2.95	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	82293	82233	82096	82207	0.12	cps
Antimony	121-1	2304	2274	2047	2208	6.36	cps
Arsenic	75-2	180	240	217	212	14.25	cps
Barium	135-1	22815	22755	22708	22759	0.24	cps
Barium	137-1	38562	39485	38870	38973	1.21	cps
Beryllium	9-1	7	11	10	9	20.47	cps
Bismuth	209-1	5097025	5050469	4969635	5039043	1.28	cps
Bismuth	209-2	4718788	4700078	4651107	4689991	0.75	cps
Bromine	81-1	28264	32055	34160	31493	9.49	cps
Bromine	81-2	5618	5591	5614	5608	0.26	cps
Cadmium	108-1	47	37	37	40	14.43	cps
Cadmium	106-1	3284	3374	3040	3233	5.34	cps
Cadmium	111-1	2443	2523	2253	2406	5.77	cps
Calcium	43-1	35726	35335	36033	35698	0.98	cps
Calcium	44-1	606014	604747	596910	602557	0.82	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	70748	70899	70216	70621	0.51	cps
Cobalt	59-2	6245	5891	6098	6078	2.92	cps
Copper	63-2	102005	102546	100589	101713	0.99	cps
Dysprosium	156-1	143	207	150	167	20.88	cps
Dysprosium	156-2	163	167	123	151	15.96	cps
Erbium	164-1	227	240	207	224	7.47	cps
Erbium	164-2	160	203	127	163	23.54	cps
Gadolinium	160-1	200	243	227	223	9.79	cps
Gadolinium	160-2	623	513	603	580	10.10	cps
Holmium	165-1	8345133	8213293	8292958	8283795	0.80	cps
Holmium	165-2	6332023	6320738	6203157	6285306	1.14	cps
Indium	115-1	7225385	7249283	7041749	7172139	1.58	cps
Indium	115-2	2914462	2942134	2936524	2931040	0.50	cps
Iron	56-2	437669220	434767714	432265700	434900878	0.62	cps
Iron	57-2	10751419	10683412	10465792	10633541	1.40	cps
Iron	54-2	24350460	23693259	23539713	23861144	1.80	cps
Krypton	83-1	567	603	460	543	13.70	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	36456	36549	36453	36486	0.15	cps
Lead	207-1	29830	30511	30133	30158	1.13	cps
Lead	208-1	140894	140502	140846	140747	0.15	cps
Lithium	6-1	705515	706492	706889	706299	0.10	cps
Magnesium	24-2	81465	81394	81512	81457	0.07	cps
Manganese	55-2	969637	976418	957206	967754	1.01	cps
Molybdenum	94-1	7495	7392	6858	7249	4.72	cps
Molybdenum	95-1	10307	10524	10047	10293	2.32	cps
Molybdenum	96-1	14277	14084	13740	14034	1.94	cps
Molybdenum	97-1	6281	6385	6391	6353	0.97	cps
Molybdenum	98-1	15972	16219	15979	16057	0.88	cps
Neodymium	150-1	263	320	290	291	9.74	cps
Neodymium	150-2	230	187	240	219	12.95	cps
Nickel	60-2	4794	4504	4607	4635	3.17	cps
Phosphorus	31-2	833	807	763	801	4.41	cps
Potassium	39-2	1337104	1290811	1277336	1301750	2.41	cps
Rhodium	103-1	7166957	7193801	7093391	7151383	0.73	cps
Rhodium	103-2	4569630	4560384	4514100	4548038	0.65	cps
Scandium	45-1	4281669	4165059	4199379	4215369	1.42	cps
Scandium	45-2	466942	466477	469282	467567	0.32	cps
Selenium	82-1	252	199	341	264	27.20	cps
Selenium	77-2	27	27	27	27	0.00	cps
Selenium	78-2	610	637	607	618	2.66	cps
Silicon	28-1	2625471	2624778	2619445	2623231	0.13	cps
Silver	107-1	527	483	500	503	4.34	cps
Silver	109-1	483	437	450	457	5.26	cps
Sodium	23-2	52882184	53006739	51948516	52612480	1.10	cps
Strontium	86-1	20581	20831	20404	20606	1.04	cps
Strontium	88-1	172919	173327	171532	172593	0.55	cps
Sulfur	34-1	170557	176291	180509	175785	2.84	cps
Terbium	159-1	8913836	8443116	8531292	8629415	2.90	cps
Terbium	159-2	6404703	6296829	6303063	6334865	0.96	cps
Thallium	203-1	123	110	123	119	6.48	cps
Thallium	205-1	277	247	233	252	8.80	cps
Tin	118-1	16546	16980	16463	16663	1.67	cps
Titanium	47-1	5188	5294	5251	5244	1.02	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-01DLX2 Instrumnet Name : P7
Client Sample ID : #452DL Dilution Factor : 2
Date & Time Acquired : 2025-08-15 16:29:43 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	180	207	163	183	11.92	cps
Vanadium	51-2	3437	3440	3330	3403	1.84	cps
Yttrium	89-1	12713979	12433812	12481920	12543237	1.19	cps
Yttrium	89-2	3728985	3728952	3658870	3705602	1.09	cps
Zinc	66-2	176676	176208	174939	175941	0.51	cps
Zirconium	90-1	4047	3921	3961	3976	1.63	cps
Zirconium	91-1	833	903	923	887	5.33	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	12198	12032	11718	11983	2.04	cps
Antimony	121-1	7765	8166	8739	8224	5.95	cps
Arsenic	75-2	1287	1293	1377	1319	3.80	cps
Barium	135-1	8329	8162	7949	8147	2.34	cps
Barium	137-1	13803	13780	13710	13765	0.35	cps
Beryllium	9-1	14	10	4	9	55.93	cps
Bismuth	209-1	3491130	3524221	3501827	3505726	0.48	cps
Bismuth	209-2	3214726	3171395	3200782	3195634	0.69	cps
Bromine	81-1	563022	718389	769551	683654	15.73	cps
Bromine	81-2	175307	178404	180812	178174	1.55	cps
Cadmium	108-1	2684	2780	2714	2726	1.82	cps
Cadmium	106-1	2274	2050	1940	2088	8.14	cps
Cadmium	111-1	2890	2783	2496	2723	7.49	cps
Calcium	43-1	4958	5298	5361	5205	4.17	cps
Calcium	44-1	122403	126415	125998	124939	1.77	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	5431	5771	5681	5628	3.13	cps
Cobalt	59-2	390	333	363	362	7.83	cps
Copper	63-2	139737	139514	140023	139758	0.18	cps
Dysprosium	156-1	3	27	30	20	72.66	cps
Dysprosium	156-2	7	13	3	8	65.47	cps
Erbium	164-1	47	57	50	51	9.96	cps
Erbium	164-2	43	47	40	43	7.70	cps
Gadolinium	160-1	33	40	40	38	10.19	cps
Gadolinium	160-2	343	307	303	318	6.98	cps
Holmium	165-1	5670346	5681308	5486141	5612598	1.95	cps
Holmium	165-2	4316130	4323088	4306789	4315336	0.19	cps
Indium	115-1	4836880	4951042	4827319	4871747	1.41	cps
Indium	115-2	1990393	2018646	1998203	2002414	0.73	cps
Iron	56-2	226157	230713	228966	228612	1.01	cps
Iron	57-2	7105	7285	7098	7163	1.48	cps
Iron	54-2	14460	14560	14100	14374	1.68	cps
Krypton	83-1	483	533	510	509	4.92	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	12759	13327	13470	13185	2.85	cps
Lead	207-1	11075	11248	11268	11197	0.95	cps
Lead	208-1	50761	53035	52584	52126	2.31	cps
Lithium	6-1	515034	514940	519340	516438	0.49	cps
Magnesium	24-2	37182	38064	37883	37710	1.24	cps
Manganese	55-2	47285	47067	47248	47200	0.25	cps
Molybdenum	94-1	12024307	12173051	12121958	12106439	0.62	cps
Molybdenum	95-1	21447594	21650911	22131187	21743231	1.61	cps
Molybdenum	96-1	22539567	22959370	23409535	22969490	1.89	cps
Molybdenum	97-1	13140940	13297986	13476582	13305170	1.26	cps
Molybdenum	98-1	33703194	34702049	34637463	34347569	1.63	cps
Neodymium	150-1	7	17	20	14	48.02	cps
Neodymium	150-2	10	10	10	10	0.00	cps
Nickel	60-2	863	870	1043	926	11.03	cps
Phosphorus	31-2	7552	7759	7805	7705	1.75	cps
Potassium	39-2	59758267	57960167	58641284	58786573	1.54	cps
Rhodium	103-1	4716008	4772644	4683164	4723939	0.96	cps
Rhodium	103-2	3105623	3110921	3110588	3109044	0.10	cps
Scandium	45-1	2752504	2804916	2705711	2754377	1.80	cps
Scandium	45-2	314969	315913	310300	313727	0.96	cps
Selenium	82-1	954	1087	1102	1048	7.79	cps
Selenium	77-2	137	133	107	126	13.10	cps
Selenium	78-2	960	980	953	964	1.44	cps
Silicon	28-1	3813565	4153425	4332393	4099794	6.43	cps
Silver	107-1	727	717	700	714	1.89	cps
Silver	109-1	633	690	750	691	8.44	cps
Sodium	23-2	31911729	32019540	32052985	31994751	0.23	cps
Strontium	86-1	22614	22945	22968	22842	0.87	cps
Strontium	88-1	188832	191420	192825	191026	1.06	cps
Sulfur	34-1	166974	161136	156727	161612	3.18	cps
Terbium	159-1	5944272	5789681	5784273	5839409	1.56	cps
Terbium	159-2	4359925	4407826	4284179	4350643	1.43	cps
Thallium	203-1	247	207	233	229	8.90	cps
Thallium	205-1	553	550	573	559	2.26	cps
Tin	118-1	6128	6171	6235	6178	0.87	cps
Titanium	47-1	500	580	533	538	7.47	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2856-03DL2X25 Instrumnet Name : P7
Client Sample ID : #456DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:32:59 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	90	90	77	86	9.00	cps
Vanadium	51-2	930	940	863	911	4.57	cps
Yttrium	89-1	8370604	8283829	8165346	8273259	1.25	cps
Yttrium	89-2	2463966	2478409	2485778	2476051	0.45	cps
Zinc	66-2	103130	102842	103106	103026	0.16	cps
Zirconium	90-1	850	617	677	714	16.96	cps
Zirconium	91-1	183	187	137	169	16.55	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	269727	269648	266184	268520	0.75	cps
Antimony	121-1	2377	2844	3330	2850	16.73	cps
Arsenic	75-2	1460	1533	1410	1468	4.23	cps
Barium	135-1	6075	6778	6198	6350	5.92	cps
Barium	137-1	10124	10744	10581	10483	3.07	cps
Beryllium	9-1	10	4	7	7	46.00	cps
Bismuth	209-1	3635024	3671674	3612193	3639630	0.82	cps
Bismuth	209-2	3244466	3267775	3182467	3231569	1.36	cps
Bromine	81-1	657190	877923	963367	832827	18.97	cps
Bromine	81-2	219428	221931	223831	221730	1.00	cps
Cadmium	108-1	527	523	630	560	10.83	cps
Cadmium	106-1	2077	2020	2024	2040	1.56	cps
Cadmium	111-1	1838	1831	1845	1838	0.39	cps
Calcium	43-1	18465	18728	19022	18738	1.49	cps
Calcium	44-1	336874	341708	341767	340116	0.83	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	5058	5044	5011	5038	0.48	cps
Cobalt	59-2	177	160	207	181	13.06	cps
Copper	63-2	170982	170895	171738	171205	0.27	cps
Dysprosium	156-1	7	3	0	3	100.05	cps
Dysprosium	156-2	3	7	7	6	34.70	cps
Erbium	164-1	53	40	43	46	15.23	cps
Erbium	164-2	17	57	23	32	66.51	cps
Gadolinium	160-1	30	53	30	38	35.66	cps
Gadolinium	160-2	367	290	310	322	12.34	cps
Holmium	165-1	5708072	5690191	5742710	5713658	0.47	cps
Holmium	165-2	4345277	4272380	4308104	4308587	0.85	cps
Indium	115-1	4901225	4931611	4929808	4920881	0.35	cps
Indium	115-2	2000905	2026143	2012338	2013129	0.63	cps
Iron	56-2	226802	228221	226788	227270	0.36	cps
Iron	57-2	7509	7515	7792	7605	2.13	cps
Iron	54-2	13950	14257	14674	14294	2.54	cps
Krypton	83-1	480	470	523	491	5.77	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1837	1823	1767	1809	2.05	cps
Lead	207-1	1490	1477	1673	1547	7.11	cps
Lead	208-1	7077	7017	7241	7112	1.63	cps
Lithium	6-1	521933	522850	523049	522611	0.11	cps
Magnesium	24-2	30447	29575	29405	29809	1.88	cps
Manganese	55-2	183705	184242	185812	184586	0.59	cps
Molybdenum	94-1	2421192	2563847	2558803	2514614	3.22	cps
Molybdenum	95-1	4305456	4464800	4419203	4396486	1.87	cps
Molybdenum	96-1	4682820	4804711	4623966	4703832	1.96	cps
Molybdenum	97-1	2679587	2774498	2773136	2742407	1.98	cps
Molybdenum	98-1	6993587	7196712	7052240	7080846	1.48	cps
Neodymium	150-1	7	7	3	6	34.70	cps
Neodymium	150-2	0	3	0	1	173.21	cps
Nickel	60-2	813	930	863	869	6.74	cps
Phosphorus	31-2	867	940	900	902	4.07	cps
Potassium	39-2	39411494	39392303	39338639	39380812	0.10	cps
Rhodium	103-1	4830770	4884230	4856070	4857023	0.55	cps
Rhodium	103-2	3138607	3186119	3160065	3161597	0.75	cps
Scandium	45-1	2850756	2838564	2856992	2848771	0.33	cps
Scandium	45-2	318616	318057	318243	318306	0.09	cps
Selenium	82-1	927	1202	1256	1128	15.60	cps
Selenium	77-2	123	150	97	123	21.62	cps
Selenium	78-2	960	987	973	973	1.37	cps
Silicon	28-1	2345285	2578289	2663147	2528907	6.51	cps
Silver	107-1	227	187	217	210	9.91	cps
Silver	109-1	193	190	163	182	9.02	cps
Sodium	23-2	20986487	21023665	20898086	20969413	0.31	cps
Strontium	86-1	9380	9036	9493	9303	2.56	cps
Strontium	88-1	73118	73456	73399	73324	0.25	cps
Sulfur	34-1	165741	163464	161248	163484	1.37	cps
Terbium	159-1	5984568	5930415	6003739	5972907	0.64	cps
Terbium	159-2	4395320	4470560	4280424	4382101	2.19	cps
Thallium	203-1	380	317	390	362	10.98	cps
Thallium	205-1	837	950	737	841	12.69	cps
Tin	118-1	5228	5815	5735	5592	5.69	cps
Titanium	47-1	330	330	303	321	4.80	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : Q2865-03DL2X25 Instrumnet Name : P7
Client Sample ID : LAW-25-0109DL2 Dilution Factor : 25
Date & Time Acquired : 2025-08-15 16:36:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	133	173	147	151	13.48	cps
Vanadium	51-2	1767	1787	1877	1810	3.24	cps
Yttrium	89-1	8357856	8404610	8497173	8419880	0.84	cps
Yttrium	89-2	2541221	2571445	2551084	2554583	0.60	cps
Zinc	66-2	194592	195302	192863	194252	0.65	cps
Zirconium	90-1	893	903	877	891	1.51	cps
Zirconium	91-1	203	213	157	191	15.83	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9664572	9263277	9358534	9428794	2.22	cps
Antimony	121-1	3169420	3240189	3113460	3174357	2.00	cps
Arsenic	75-2	219697	215148	214252	216366	1.35	cps
Barium	135-1	3766973	3854750	3929684	3850469	2.12	cps
Barium	137-1	6533707	6565767	6588430	6562635	0.42	cps
Beryllium	9-1	370238	382331	375439	376003	1.61	cps
Bismuth	209-1	5148535	4813025	4853962	4938507	3.71	cps
Bismuth	209-2	4618661	4488316	4548171	4551716	1.43	cps
Bromine	81-1	108609	98140	90660	99136	9.09	cps
Bromine	81-2	13863	13553	12632	13349	4.80	cps
Cadmium	108-1	58956	61031	60415	60134	1.77	cps
Cadmium	106-1	85470	86389	87908	86589	1.42	cps
Cadmium	111-1	833701	841518	837738	837652	0.47	cps
Calcium	43-1	4392423	4401258	4379010	4390897	0.26	cps
Calcium	44-1	70745797	70082044	71666142	70831328	1.12	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1690279	1740820	1711715	1714271	1.48	cps
Cobalt	59-2	2566363	2616356	2588134	2590284	0.97	cps
Copper	63-2	19647529	19193511	19798211	19546417	1.61	cps
Dysprosium	156-1	190	170	233	198	16.37	cps
Dysprosium	156-2	367	307	350	341	9.08	cps
Erbium	164-1	403	383	493	427	13.73	cps
Erbium	164-2	293	273	283	283	3.53	cps
Gadolinium	160-1	223	260	250	244	7.75	cps
Gadolinium	160-2	557	673	567	599	10.80	cps
Holmium	165-1	8727423	8070130	8100917	8299490	4.47	cps
Holmium	165-2	6561618	6482812	6354232	6466220	1.62	cps
Indium	115-1	7232162	6702967	7091223	7008784	3.91	cps
Indium	115-2	2983499	2860932	2961688	2935373	2.23	cps
Iron	56-2	402496941	398395714	394343847	398412167	1.02	cps
Iron	57-2	9938118	10001540	9972574	9970744	0.32	cps
Iron	54-2	21970097	22208363	22045431	22074631	0.55	cps
Krypton	83-1	503	493	537	511	4.44	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	11566302	11915441	11563774	11681839	1.73	cps
Lead	207-1	10077325	10221890	9843141	10047452	1.90	cps
Lead	208-1	46168817	46162772	45429558	45920382	0.93	cps
Lithium	6-1	762105	710106	726611	732941	3.63	cps
Magnesium	24-2	81721522	79827160	80997529	80848737	1.18	cps
Manganese	55-2	14519072	14646240	14457770	14541027	0.66	cps
Molybdenum	94-1	11687722	11997961	11883202	11856295	1.32	cps
Molybdenum	95-1	17008050	17165394	17456530	17209991	1.32	cps
Molybdenum	96-1	18116238	18507008	18570441	18397896	1.34	cps
Molybdenum	97-1	10313957	10662663	10448601	10475074	1.68	cps
Molybdenum	98-1	26832055	27490975	27217086	27180039	1.22	cps
Neodymium	150-1	213	210	163	196	14.29	cps
Neodymium	150-2	53	80	80	71	21.65	cps
Nickel	60-2	653143	643770	642512	646475	0.90	cps
Phosphorus	31-2	104183	103052	103945	103726	0.58	cps
Potassium	39-2	48523888	47877909	48813333	48405043	0.99	cps
Rhodium	103-1	7091777	6691768	6753842	6845796	3.14	cps
Rhodium	103-2	4588842	4442113	4436411	4489122	1.92	cps
Scandium	45-1	4465281	4098030	4171043	4244785	4.58	cps
Scandium	45-2	489161	487783	479313	485419	1.10	cps
Selenium	82-1	61780	63608	63620	63003	1.68	cps
Selenium	77-2	7866	7956	7886	7902	0.60	cps
Selenium	78-2	26354	26090	26020	26154	0.67	cps
Silicon	28-1	8802546	9200160	8883338	8962015	2.34	cps
Silver	107-1	4228832	4274195	4253411	4252146	0.53	cps
Silver	109-1	3979372	3973995	4095889	4016419	1.71	cps
Sodium	23-2	122511068	121733068	118563988	120936042	1.73	cps
Strontium	86-1	1067303	1097898	1079947	1081716	1.42	cps
Strontium	88-1	9524526	9822903	9571127	9639519	1.67	cps
Sulfur	34-1	382164	388251	386905	385773	0.83	cps
Terbium	159-1	9134053	8519012	8334359	8662475	4.83	cps
Terbium	159-2	6680406	6497649	6355376	6511144	2.50	cps
Thallium	203-1	2941457	2922654	2820267	2894793	2.25	cps
Thallium	205-1	6788687	6908151	6726806	6807881	1.35	cps
Tin	118-1	2668011	2650339	2582694	2633681	1.71	cps
Titanium	47-1	4286231	4366521	4288101	4313618	1.06	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCV05 Instrumnet Name : P7
Client Sample ID : CCV05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:39:21 DataFile Name : 070CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8863992	8920570	8861114	8881892	0.38	cps
Vanadium	51-2	1472115	1459757	1449722	1460531	0.77	cps
Yttrium	89-1	12835901	11957273	12399839	12397671	3.54	cps
Yttrium	89-2	3934373	3815777	3735212	3828454	2.62	cps
Zinc	66-2	3538295	3544299	3525151	3535915	0.28	cps
Zirconium	90-1	5774696	5804508	5906028	5828411	1.18	cps
Zirconium	91-1	1250858	1270909	1272956	1264907	0.97	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1600	1510	1637	1582	4.12	cps
Antimony	121-1	267	257	263	262	1.94	cps
Arsenic	75-2	20	23	37	27	33.08	cps
Barium	135-1	140	140	137	139	1.38	cps
Barium	137-1	183	263	207	218	18.89	cps
Beryllium	9-1	7	7	10	8	24.40	cps
Bismuth	209-1	5005473	5112509	5035442	5051142	1.09	cps
Bismuth	209-2	4590191	4425755	4384818	4466921	2.43	cps
Bromine	81-1	17327	17250	17213	17264	0.34	cps
Bromine	81-2	1437	1350	1450	1412	3.84	cps
Cadmium	108-1	23	3	27	18	71.00	cps
Cadmium	106-1	2744	2770	2544	2686	4.61	cps
Cadmium	111-1	1931	1977	1782	1897	5.38	cps
Calcium	43-1	503	583	533	540	7.48	cps
Calcium	44-1	38556	38034	38265	38285	0.68	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2740	2767	2674	2727	1.76	cps
Cobalt	59-2	120	80	87	96	22.43	cps
Copper	63-2	1560	1597	1403	1520	6.76	cps
Dysprosium	156-1	7	3	10	7	50.03	cps
Dysprosium	156-2	3	3	10	6	69.34	cps
Erbium	164-1	67	53	43	54	21.51	cps
Erbium	164-2	53	40	47	47	14.28	cps
Gadolinium	160-1	63	40	40	48	28.19	cps
Gadolinium	160-2	477	487	507	490	3.12	cps
Holmium	165-1	7799741	7881616	7722288	7801215	1.02	cps
Holmium	165-2	6034107	5838234	5904726	5925689	1.68	cps
Indium	115-1	6978419	7038941	6797993	6938451	1.81	cps
Indium	115-2	2830106	2780289	2839571	2816656	1.13	cps
Iron	56-2	26637	26704	26447	26596	0.50	cps
Iron	57-2	1887	1877	1987	1917	3.17	cps
Iron	54-2	7482	3240	3350	4691	51.54	cps
Krypton	83-1	530	487	527	514	4.69	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	687	690	700	692	1.00	cps
Lead	207-1	557	550	533	547	2.20	cps
Lead	208-1	2613	2663	2613	2630	1.10	cps
Lithium	6-1	713981	707344	700634	707320	0.94	cps
Magnesium	24-2	1493	1290	1437	1407	7.46	cps
Manganese	55-2	1119	1165	1252	1179	5.72	cps
Molybdenum	94-1	380	383	393	386	1.80	cps
Molybdenum	95-1	387	380	357	374	4.21	cps
Molybdenum	96-1	560	437	527	508	12.56	cps
Molybdenum	97-1	363	330	303	332	9.05	cps
Molybdenum	98-1	537	680	550	589	13.45	cps
Neodymium	150-1	7	17	3	9	78.08	cps
Neodymium	150-2	0	17	13	10	88.20	cps
Nickel	60-2	670	703	713	696	3.26	cps
Phosphorus	31-2	467	510	633	537	16.11	cps
Potassium	39-2	91606	89504	90313	90474	1.17	cps
Rhodium	103-1	6900712	6939819	6689259	6843263	1.97	cps
Rhodium	103-2	4402988	4455781	4400777	4419849	0.70	cps
Scandium	45-1	4166825	4086314	4033832	4095657	1.64	cps
Scandium	45-2	444349	447376	439112	443613	0.94	cps
Selenium	82-1	238	365	247	284	25.01	cps
Selenium	77-2	23	40	27	30	29.40	cps
Selenium	78-2	613	577	627	606	4.28	cps
Silicon	28-1	581777	583807	580566	582050	0.28	cps
Silver	107-1	487	393	397	426	12.44	cps
Silver	109-1	430	313	320	354	18.48	cps
Sodium	23-2	54723	54252	54292	54422	0.48	cps
Strontium	86-1	857	867	897	873	2.38	cps
Strontium	88-1	870	857	720	816	10.18	cps
Sulfur	34-1	149546	151425	150731	150567	0.63	cps
Terbium	159-1	8148823	8020802	8063876	8077834	0.81	cps
Terbium	159-2	6037538	5965509	6003889	6002312	0.60	cps
Thallium	203-1	140	80	130	117	27.55	cps
Thallium	205-1	303	267	273	281	6.95	cps
Tin	118-1	2014	1950	1913	1959	2.58	cps
Titanium	47-1	83	120	83	96	22.16	cps

LB Number : LB136861 Operator : Jaswal
Lab Sample ID : CCB05 Instrumnet Name : P7
Client Sample ID : CCB05 Dilution Factor : 1
Date & Time Acquired : 2025-08-15 16:50:13 DataFile Name : 071CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	37	30	33	33	10.01	cps
Vanadium	51-2	17	40	30	29	40.51	cps
Yttrium	89-1	12084998	12314498	11550929	11983475	3.27	cps
Yttrium	89-2	3522569	3517042	3474347	3504653	0.75	cps
Zinc	66-2	4667	4557	4391	4539	3.07	cps
Zirconium	90-1	450	460	530	480	9.08	cps
Zirconium	91-1	80	57	97	78	25.83	cps