



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB27	Mercury	0.076	+/-0.2	U	0.20	CV	08/21/2025	09:01	LB136901

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB84	Mercury	0.076	+/-0.2	U	0.20	CV	08/21/2025	09:05	LB136901
CCB85	Mercury	0.076	+/-0.2	U	0.20	CV	08/21/2025	09:38	LB136901
CCB86	Mercury	0.076	+/-0.2	U	0.20	CV	08/21/2025	10:14	LB136901

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	13:43	LB136861
	Antimony	0.11	+/-1	U	2.00	P	08/15/2025	13:43	LB136861
	Arsenic	0.089	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	13:43	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	13:43	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	13:43	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	13:43	LB136861
	Iron	7.81	+/-25	U	50.0	P	08/15/2025	13:43	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Magnesium	19.5	+/-250	U	500	P	08/15/2025	13:43	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	13:43	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	13:43	LB136861
	Silver	0.060	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	13:43	LB136861
	Thallium	0.060	+/-0.5	U	1.00	P	08/15/2025	13:43	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	13:43	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	13:43	LB136861

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	13:56	LB136861
	Antimony	0.11	+/-1	J	2.00	P	08/15/2025	13:56	LB136861
	Arsenic	0.089	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	13:56	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	13:56	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	13:56	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	13:56	LB136861
	Iron	7.81	+/-25	U	50.0	P	08/15/2025	13:56	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Magnesium	19.5	+/-250	U	500	P	08/15/2025	13:56	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	13:56	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	13:56	LB136861
	Silver	0.060	+/-0.5	U	1.00	P	08/15/2025	13:56	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	13:56	LB136861
	Thallium	0.070	+/-0.5	J	1.00	P	08/15/2025	13:56	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	13:56	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	13:56	LB136861
CCB02	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	14:44	LB136861
	Antimony	0.13	+/-1	J	2.00	P	08/15/2025	14:44	LB136861
	Arsenic	0.089	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	14:44	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	14:44	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	14:44	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	14:44	LB136861
	Iron	9.88	+/-25	J	50.0	P	08/15/2025	14:44	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Magnesium	19.5	+/-250	U	500	P	08/15/2025	14:44	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	14:44	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	14:44	LB136861

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.060	+/-0.5	U	1.00	P	08/15/2025	14:44	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	14:44	LB136861
	Thallium	0.10	+/-0.5	J	1.00	P	08/15/2025	14:44	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	14:44	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	14:44	LB136861
CCB03	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	15:24	LB136861
	Antimony	0.14	+/-1	J	2.00	P	08/15/2025	15:24	LB136861
	Arsenic	0.089	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	15:24	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	15:24	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	15:24	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	15:24	LB136861
	Iron	7.81	+/-25	U	50.0	P	08/15/2025	15:24	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Magnesium	19.5	+/-250	U	500	P	08/15/2025	15:24	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	15:24	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	15:24	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	15:24	LB136861
	Silver	0.060	+/-0.5	J	1.00	P	08/15/2025	15:24	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	15:24	LB136861
	Thallium	0.090	+/-0.5	J	1.00	P	08/15/2025	15:24	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	15:24	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	15:24	LB136861
CCB04	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	16:10	LB136861
	Antimony	0.13	+/-1	J	2.00	P	08/15/2025	16:10	LB136861
	Arsenic	0.20	+/-0.5	J	1.00	P	08/15/2025	16:10	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	16:10	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	16:10	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	16:10	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	16:10	LB136861
	Iron	7.81	+/-25	U	50.0	P	08/15/2025	16:10	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	19.5	+/-250	U	500	P	08/15/2025	16:10	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	16:10	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	16:10	LB136861
	Silver	0.060	+/-0.5	J	1.00	P	08/15/2025	16:10	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	16:10	LB136861
	Thallium	0.060	+/-0.5	U	1.00	P	08/15/2025	16:10	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	16:10	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	16:10	LB136861
CCB05	Aluminum	1.94	+/-10	U	20.0	P	08/15/2025	16:50	LB136861
	Antimony	0.11	+/-1	U	2.00	P	08/15/2025	16:50	LB136861
	Arsenic	0.089	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Barium	0.21	+/-5	U	10.0	P	08/15/2025	16:50	LB136861
	Beryllium	0.32	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Cadmium	0.34	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Calcium	45.7	+/-250	U	500	P	08/15/2025	16:50	LB136861
	Chromium	0.21	+/-1	U	2.00	P	08/15/2025	16:50	LB136861
	Cobalt	0.070	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Copper	0.30	+/-1	U	2.00	P	08/15/2025	16:50	LB136861
	Iron	7.81	+/-25	U	50.0	P	08/15/2025	16:50	LB136861
	Lead	0.21	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Magnesium	19.5	+/-250	U	500	P	08/15/2025	16:50	LB136861
	Manganese	0.43	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Nickel	0.27	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Potassium	36.4	+/-250	U	500	P	08/15/2025	16:50	LB136861
	Selenium	2.90	+/-2.5	U	5.00	P	08/15/2025	16:50	LB136861
	Silver	0.060	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Sodium	128	+/-250	U	500	P	08/15/2025	16:50	LB136861
	Thallium	0.060	+/-0.5	U	1.00	P	08/15/2025	16:50	LB136861
	Vanadium	0.077	+/-2.5	U	5.00	P	08/15/2025	16:50	LB136861
	Zinc	1.25	+/-2.5	U	5.00	P	08/15/2025	16:50	LB136861

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	1.94	+/-10	U	20.0	P	08/21/2025	12:29	LB136916
	Antimony	0.11	+/-1	U	2.00	P	08/21/2025	12:29	LB136916
	Arsenic	0.089	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Barium	0.21	+/-5	U	10.0	P	08/21/2025	12:29	LB136916
	Beryllium	0.32	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Cadmium	0.34	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Calcium	45.7	+/-250	U	500	P	08/21/2025	12:29	LB136916
	Chromium	0.21	+/-1	U	2.00	P	08/21/2025	12:29	LB136916
	Cobalt	0.070	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Copper	0.30	+/-1	U	2.00	P	08/21/2025	12:29	LB136916
	Iron	7.81	+/-25	U	50.0	P	08/21/2025	12:29	LB136916
	Lead	0.21	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Magnesium	19.5	+/-250	U	500	P	08/21/2025	12:29	LB136916
	Manganese	0.43	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Nickel	0.27	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Potassium	36.4	+/-250	U	500	P	08/21/2025	12:29	LB136916
	Selenium	2.90	+/-2.5	U	5.00	P	08/21/2025	12:29	LB136916
	Silver	0.060	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Sodium	128	+/-250	U	500	P	08/21/2025	12:29	LB136916
	Thallium	0.060	+/-0.5	U	1.00	P	08/21/2025	12:29	LB136916
	Vanadium	0.077	+/-2.5	U	5.00	P	08/21/2025	12:29	LB136916
	Zinc	1.25	+/-2.5	U	5.00	P	08/21/2025	12:29	LB136916

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	1.94	+/-10	U	20.0	P	08/21/2025	12:44	LB136916
	Antimony	0.11	+/-1	U	2.00	P	08/21/2025	12:44	LB136916
	Arsenic	0.089	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Barium	0.21	+/-5	U	10.0	P	08/21/2025	12:44	LB136916
	Beryllium	0.32	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Cadmium	0.34	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Calcium	45.7	+/-250	U	500	P	08/21/2025	12:44	LB136916
	Chromium	0.21	+/-1	U	2.00	P	08/21/2025	12:44	LB136916
	Cobalt	0.070	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Copper	0.30	+/-1	U	2.00	P	08/21/2025	12:44	LB136916
	Iron	7.81	+/-25	U	50.0	P	08/21/2025	12:44	LB136916
	Lead	0.21	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Magnesium	19.5	+/-250	U	500	P	08/21/2025	12:44	LB136916
	Manganese	0.43	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Nickel	0.27	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Potassium	36.4	+/-250	U	500	P	08/21/2025	12:44	LB136916
	Selenium	2.90	+/-2.5	U	5.00	P	08/21/2025	12:44	LB136916
	Silver	0.060	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Sodium	128	+/-250	U	500	P	08/21/2025	12:44	LB136916
	Thallium	0.060	+/-0.5	U	1.00	P	08/21/2025	12:44	LB136916
	Vanadium	0.077	+/-2.5	U	5.00	P	08/21/2025	12:44	LB136916
	Zinc	1.25	+/-2.5	U	5.00	P	08/21/2025	12:44	LB136916
CCB02	Aluminum	1.94	+/-10	U	20.0	P	08/21/2025	13:43	LB136916
	Antimony	0.15	+/-1	J	2.00	P	08/21/2025	13:43	LB136916
	Arsenic	0.089	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Barium	0.21	+/-5	U	10.0	P	08/21/2025	13:43	LB136916
	Beryllium	0.32	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Cadmium	0.34	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Calcium	45.7	+/-250	U	500	P	08/21/2025	13:43	LB136916
	Chromium	0.21	+/-1	U	2.00	P	08/21/2025	13:43	LB136916
	Cobalt	0.070	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Copper	0.30	+/-1	U	2.00	P	08/21/2025	13:43	LB136916
	Iron	7.81	+/-25	U	50.0	P	08/21/2025	13:43	LB136916
	Lead	0.21	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Magnesium	19.5	+/-250	U	500	P	08/21/2025	13:43	LB136916
	Manganese	0.43	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Nickel	0.27	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Potassium	36.4	+/-250	U	500	P	08/21/2025	13:43	LB136916
	Selenium	2.90	+/-2.5	U	5.00	P	08/21/2025	13:43	LB136916

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.060	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Sodium	128	+/-250	U	500	P	08/21/2025	13:43	LB136916
	Thallium	0.060	+/-0.5	U	1.00	P	08/21/2025	13:43	LB136916
	Vanadium	0.077	+/-2.5	U	5.00	P	08/21/2025	13:43	LB136916
	Zinc	1.25	+/-2.5	U	5.00	P	08/21/2025	13:43	LB136916
CCB03	Aluminum	1.94	+/-10	U	20.0	P	08/21/2025	14:34	LB136916
	Antimony	0.15	+/-1	J	2.00	P	08/21/2025	14:34	LB136916
	Arsenic	0.089	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Barium	0.21	+/-5	U	10.0	P	08/21/2025	14:34	LB136916
	Beryllium	0.32	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Cadmium	0.34	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Calcium	45.7	+/-250	U	500	P	08/21/2025	14:34	LB136916
	Chromium	0.21	+/-1	U	2.00	P	08/21/2025	14:34	LB136916
	Cobalt	0.070	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Copper	0.30	+/-1	U	2.00	P	08/21/2025	14:34	LB136916
	Iron	7.81	+/-25	U	50.0	P	08/21/2025	14:34	LB136916
	Lead	0.21	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Magnesium	19.5	+/-250	U	500	P	08/21/2025	14:34	LB136916
	Manganese	0.43	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Nickel	0.27	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Potassium	36.4	+/-250	U	500	P	08/21/2025	14:34	LB136916
	Selenium	2.90	+/-2.5	U	5.00	P	08/21/2025	14:34	LB136916
	Silver	0.060	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Sodium	128	+/-250	U	500	P	08/21/2025	14:34	LB136916
	Thallium	0.060	+/-0.5	U	1.00	P	08/21/2025	14:34	LB136916
	Vanadium	0.077	+/-2.5	U	5.00	P	08/21/2025	14:34	LB136916
	Zinc	1.25	+/-2.5	U	5.00	P	08/21/2025	14:34	LB136916
CCB04	Aluminum	1.94	+/-10	U	20.0	P	08/21/2025	15:18	LB136916
	Antimony	0.11	+/-1	U	2.00	P	08/21/2025	15:18	LB136916
	Arsenic	0.089	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Barium	0.21	+/-5	U	10.0	P	08/21/2025	15:18	LB136916
	Beryllium	0.32	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Cadmium	0.34	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Calcium	45.7	+/-250	U	500	P	08/21/2025	15:18	LB136916
	Chromium	0.21	+/-1	U	2.00	P	08/21/2025	15:18	LB136916
	Cobalt	0.070	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Copper	0.30	+/-1	U	2.00	P	08/21/2025	15:18	LB136916
	Iron	7.81	+/-25	U	50.0	P	08/21/2025	15:18	LB136916
	Lead	0.21	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	19.5	+/-250	U	500	P	08/21/2025	15:18	LB136916
	Manganese	0.43	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Nickel	0.27	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Potassium	36.4	+/-250	U	500	P	08/21/2025	15:18	LB136916
	Selenium	2.90	+/-2.5	U	5.00	P	08/21/2025	15:18	LB136916
	Silver	0.060	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Sodium	128	+/-250	U	500	P	08/21/2025	15:18	LB136916
	Thallium	0.060	+/-0.5	U	1.00	P	08/21/2025	15:18	LB136916
	Vanadium	0.077	+/-2.5	U	5.00	P	08/21/2025	15:18	LB136916
	Zinc	1.25	+/-2.5	U	5.00	P	08/21/2025	15:18	LB136916