

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169250BS							
Aluminum	ug/L	10000	9810		98	80 - 120	P
Antimony	ug/L	500	502		100	80 - 120	P
Arsenic	ug/L	500	486		97	80 - 120	P
Barium	ug/L	2500	2530		101	80 - 120	P
Beryllium	ug/L	500	493		99	80 - 120	P
Cadmium	ug/L	500	511		102	80 - 120	P
Calcium	ug/L	50000	50400		101	80 - 120	P
Chromium	ug/L	500	485		97	80 - 120	P
Cobalt	ug/L	500	488		98	80 - 120	P
Copper	ug/L	5000	5000		100	80 - 120	P
Iron	ug/L	25000	25700		103	80 - 120	P
Lead	ug/L	2500	2420		97	80 - 120	P
Magnesium	ug/L	50000	48900		98	80 - 120	P
Manganese	ug/L	5000	4890		98	80 - 120	P
Nickel	ug/L	500	488		98	80 - 120	P
Potassium	ug/L	25000	24600		98	80 - 120	P
Selenium	ug/L	500	505		101	80 - 120	P
Silver	ug/L	500	517		103	80 - 120	P
Sodium	ug/L	50000	50200		100	80 - 120	P
Thallium	ug/L	500	488		98	80 - 120	P
Vanadium	ug/L	500	485		97	80 - 120	P
Zinc	ug/L	5000	4950		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2869

Contract: JACO05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169327BS Mercury	ug/L	4.0	4.18		104	80 - 120	CV