



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082125\
Data File : VX047454.D
Acq On : 21 Aug 2025 07:58
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Title : SW846 8260
Last Update : Mon Aug 18 04:18:28 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	21493	PASS
75	95	30	60	51.8	57413	PASS
95	95	100	100	100.0	110837	PASS
96	95	5	9	6.5	7187	PASS
173	174	0.00	2	1.0	873	PASS
174	95	50	100	77.9	86360	PASS
175	174	5	9	7.1	6128	PASS
176	174	95	101	95.4	82403	PASS
177	176	5	9	6.7	5524	PASS