

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q2869
 Lab File ID: VX047400.D Date Analyzed: 08/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:37
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	167955	12.018				
UPPER LIMIT	335910	12.518				
LOWER LIMIT	83977.5	11.518				
EPA SAMPLE NO.						
MW-19B-71.5-081325	210685	12.02				
EB01-081325	238761	12.02				
EB02-081325	212656	12.02				
TB01-081325	219893	12.02				
VX0819WBL01	212865	12.02				
VX0819WBS01	189862	12.02				
VX0819WBSD01	189413	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.