

## Initial and Continuing Calibration Verification

**Client:** JACOBS Engineering Group, Inc.

**SDG No.:** Q2878

**Project:** Former Schlumberger STC PTC Site D3868221

**RunNo.:** LB136833

| Analyte                | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|------------------------|-------|--------|------------|------------|------------------------|---------------|
| <b>Sample ID: ICV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.3   | 10         | 103        | 90-110                 | 08/06/2025    |
| Chloride               | mg/L  | 3.2    | 3          | 107        | 90-110                 | 08/06/2025    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 08/06/2025    |
| Nitrite                | mg/L  | 3.1    | 3          | 103        | 90-110                 | 08/06/2025    |
| Nitrate                | mg/L  | 2.6    | 2.5        | 104        | 90-110                 | 08/06/2025    |
| Sulfate                | mg/L  | 15.4   | 15         | 103        | 90-110                 | 08/06/2025    |
| Orthophosphate as P    | mg/L  | 5      | 5          | 100        | 90-110                 | 08/06/2025    |
| <b>Sample ID: CCV1</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.2   | 10         | 102        | 90-110                 | 08/15/2025    |
| Chloride               | mg/L  | 3.1    | 3          | 103        | 90-110                 | 08/15/2025    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 08/15/2025    |
| Nitrite                | mg/L  | 3      | 3          | 100        | 90-110                 | 08/15/2025    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 08/15/2025    |
| Sulfate                | mg/L  | 15.2   | 15         | 101        | 90-110                 | 08/15/2025    |
| Orthophosphate as P    | mg/L  | 5      | 5          | 100        | 90-110                 | 08/15/2025    |
| <b>Sample ID: CCV2</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.2   | 10         | 102        | 90-110                 | 08/15/2025    |
| Chloride               | mg/L  | 3.1    | 3          | 103        | 90-110                 | 08/15/2025    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 08/15/2025    |
| Nitrite                | mg/L  | 3      | 3          | 100        | 90-110                 | 08/15/2025    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 08/15/2025    |
| Sulfate                | mg/L  | 15.2   | 15         | 101        | 90-110                 | 08/15/2025    |
| Orthophosphate as P    | mg/L  | 4.6    | 5          | 92         | 90-110                 | 08/15/2025    |
| <b>Sample ID: CCV3</b> |       |        |            |            |                        |               |
| Bromide                | mg/L  | 10.2   | 10         | 102        | 90-110                 | 08/15/2025    |
| Chloride               | mg/L  | 3.1    | 3          | 103        | 90-110                 | 08/15/2025    |
| Fluoride               | mg/L  | 2      | 2          | 100        | 90-110                 | 08/15/2025    |
| Nitrite                | mg/L  | 3      | 3          | 100        | 90-110                 | 08/15/2025    |
| Nitrate                | mg/L  | 2.5    | 2.5        | 100        | 90-110                 | 08/15/2025    |
| Sulfate                | mg/L  | 15.1   | 15         | 101        | 90-110                 | 08/15/2025    |
| Orthophosphate as P    | mg/L  | 5.2    | 5          | 104        | 90-110                 | 08/15/2025    |