

Laboratory Control Sample Summary

Client:	JACOBS Engineering Group, Inc.	SDG No.:	Q2878
Project:	Former Schlumberger STC PTC Site D3868221	Run No.:	LB136833

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136833BSW							
Bromide	mg/L	10	10.2		102	1	90-110	08/15/2025
Chloride	mg/L	3	3.10		103	1	90-110	08/15/2025
Fluoride	mg/L	2	2.00		100	1	90-110	08/15/2025
Nitrite	mg/L	3	3.00		100	1	90-110	08/15/2025
Nitrate	mg/L	2.5	2.50		100	1	90-110	08/15/2025
Sulfate	mg/L	15	15.3		102	1	90-110	08/15/2025
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	08/15/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB136835

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136835BS							
TDS	mg/L	100	95.0		95	1	90-110	08/14/2025

Laboratory Control Sample Summary

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Project: Former Schlumberger STC PTC Site D3868221

Run No.: LB136847

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136847BSW							
Alkalinity	mg/L	50	53.4		107	1	80-120	08/15/2025