

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136901

Review By	jaswal	Review On	8/21/2025 9:25:37 PM
Supervise By	mohan	Supervise On	8/26/2025 9:41:32 AM

STD. NAME	STD REF.#
ICAL Standard	MP86848,MP86849,MP86850,MP86851,MP86852,MP86853
ICV Standard	MP86854
CCV Standard	MP86856
ICSA Standard	
CRI Standard	MP86858
LCS Standard	
Chk Standard	MP86855,MP86857,,MP868,MP86859,MP86872

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/21/25 08:41		MOHAN	OK
2	S0.2	S0.2	CAL2	08/21/25 08:44		MOHAN	OK
3	S2.5	S2.5	CAL3	08/21/25 08:46		MOHAN	OK
4	S5	S5	CAL4	08/21/25 08:48		MOHAN	OK
5	S7.5	S7.5	CAL5	08/21/25 08:51		MOHAN	OK
6	S10	S10	CAL6	08/21/25 08:56		MOHAN	OK
7	ICV27	ICV27	ICV	08/21/25 08:59		MOHAN	OK
8	ICB27	ICB27	ICB	08/21/25 09:01		MOHAN	OK
9	CCV84	CCV84	CCV	08/21/25 09:03		MOHAN	OK
10	CCB84	CCB84	CCB	08/21/25 09:05		MOHAN	OK
11	CRA	CRA	CRDL	08/21/25 09:08		MOHAN	OK
12	HighStd	HighStd	HIGH STD	08/21/25 09:10		MOHAN	OK
13	ChkStd	ChkStd	SAM	08/21/25 09:12		MOHAN	OK
14	PB169327BL	PB169327BL	MB	08/21/25 09:14		MOHAN	OK
15	PB169327BS	PB169327BS	LCS	08/21/25 09:19		MOHAN	OK
16	Q2869-01	MW-19B-71.5-081325	SAM	08/21/25 09:22		MOHAN	OK
17	Q2869-01DUP	MW-19B-71.5-081325	DUP	08/21/25 09:24		MOHAN	OK
18	Q2869-01MS	MW-19B-71.5-081325	MS	08/21/25 09:26		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136901

Review By	jaswal	Review On	8/21/2025 9:25:37 PM
Supervise By	mohan	Supervise On	8/26/2025 9:41:32 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86848,MP86849,MP86850,MP86851,MP86852,MP86853		
ICV Standard	MP86854		
CCV Standard	MP86856		
ICSA Standard			
CRI Standard	MP86858		
LCS Standard			
Chk Standard	MP86855,MP86857,,MP868,MP86859,MP86872		

19	Q2869-01MSD	MW-19B-71.5-081325	MSD	08/21/25 09:31		MOHAN	OK
20	Q2869-05	MW-19B-71.5-081325	SAM	08/21/25 09:33		MOHAN	OK
21	CCV85	CCV85	CCV	08/21/25 09:36		MOHAN	OK
22	CCB85	CCB85	CCB	08/21/25 09:38		MOHAN	OK
23	Q2869-07	EB01-081325	SAM	08/21/25 09:40		MOHAN	OK
24	Q2878-01	MW-18B-57.5-081325	SAM	08/21/25 09:43		MOHAN	OK
25	Q2890-01	60296	SAM	08/21/25 09:45		MOHAN	OK
26	Q2890-02	60295	SAM	08/21/25 09:47		MOHAN	OK
27	Q2890-03	60272	SAM	08/21/25 09:49		MOHAN	OK
28	Q2906-01	3409	SAM	08/21/25 09:52		MOHAN	OK
29	Q2909-03	4066	SAM	08/21/25 09:54		MOHAN	OK
30	Q2909-04	4067	SAM	08/21/25 09:56		MOHAN	OK
31	Q2869-01L	MW-19B-71.5-081325	SD	08/21/25 10:02		MOHAN	OK
32	Q2869-01A	MW-19B-71.5-081325	PS	08/21/25 10:07		MOHAN	OK
33	CCV86	CCV86	CCV	08/21/25 10:12		MOHAN	OK
34	CCB86	CCB86	CCB	08/21/25 10:14		MOHAN	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136916

Review By	Janvi	Review On	8/25/2025 10:29:28 AM
Supervise By	jaswal	Supervise On	8/25/2025 3:50:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86739,MP86789,MP86788,MP86745,MP86744,MP86743,MP86742,MP86741,MP86740		
ICV Standard	MP86863,MP86788		
CCV Standard	MP86794		
ICSA Standard	MP86746,MP86747		
CRI Standard			
LCS Standard			
Chk Standard	MP86750,MP86748		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	08/21/25 10:59		Jaswal	OK
2	S0	S0	CAL1	08/21/25 11:33		Jaswal	OK
3	S2	S2	CAL3	08/21/25 11:39		Jaswal	OK
4	S3	S3	CAL4	08/21/25 11:43		Jaswal	OK
5	S4	S4	CAL5	08/21/25 11:46		Jaswal	OK
6	S5	S5	CAL6	08/21/25 11:49		Jaswal	OK
7	S6	S6	CAL7	08/21/25 11:52		Jaswal	OK
8	S7	S7	CAL8	08/21/25 11:54		Jaswal	OK
9	S8	S8	CAL9	08/21/25 11:57		Jaswal	OK
10	ICV01	ICV01	ICV	08/21/25 12:15		Jaswal	OK
11	LLICV01	LLICV01	LLICV	08/21/25 12:26		Jaswal	OK
12	ICB01	ICB01	ICB	08/21/25 12:29		Jaswal	OK
13	ICSA01	ICSA01	ICSA	08/21/25 12:33		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	08/21/25 12:36		Jaswal	OK
15	CCV01	CCV01	CCV	08/21/25 12:39		Jaswal	OK
16	CCB01	CCB01	CCB	08/21/25 12:44		Jaswal	OK
17	CRI	CRI	CRDL	08/21/25 12:48		Jaswal	OK
18	Q2126-01	LOD-MDL-SOIL-03-Q	SAM	08/21/25 12:58		Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136916

Review By	Janvi	Review On	8/25/2025 10:29:28 AM
Supervise By	jaswal	Supervise On	8/25/2025 3:50:25 PM

STD. NAME	STD REF.#
ICAL Standard	MP86739,MP86789,MP86788,MP86745,MP86744,MP86743,MP86742,MP86741,MP86740
ICV Standard	MP86863,MP86788
CCV Standard	MP86794
ICSA Standard	MP86746,MP86747
CRI Standard	
LCS Standard	
Chk Standard	MP86750,MP86748

19	Q2126-16	LLOQ-SOIL-QT2-202	SAM	08/21/25 13:14		Jaswal	Not Ok
20	Q2126-02	LOQ-SOIL-02-QT2-20	LOQ	08/21/25 13:20		Jaswal	Not Ok
21	Q2126-01RE	LOD-MDL-SOIL-03-Q	SAM	08/21/25 13:27		Jaswal	Not Ok
22	Q2126-07	LOD-MDL-WATER-01	SAM	08/21/25 13:30		Jaswal	Not Ok
23	Q2126-08	LOQ-WATER-02-QT2	LOQ	08/21/25 13:34		Jaswal	Not Ok
24	Q2126-08RE	LOQ-WATER-02-QT2	SAM	08/21/25 13:37		Jaswal	Not Ok
25	CCV02	CCV02	CCV	08/21/25 13:40		Jaswal	OK
26	CCB02	CCB02	CCB	08/21/25 13:43		Jaswal	OK
27	PB169250BL	PB169250BL	MB	08/21/25 13:46		Jaswal	OK
28	PB169250BS	PB169250BS	LCS	08/21/25 13:50		Jaswal	OK
29	PB169323BL	PB169323BL	MB	08/21/25 13:52		Jaswal	OK
30	PB169323BS	PB169323BS	LCS	08/21/25 13:56		Jaswal	OK
31	Q2878-01	MW-18B-57.5-081325	SAM	08/21/25 13:58		Jaswal	OK
32	Q2878-01DUP	MW-18B-57.5-081325	DUP	08/21/25 14:02		Jaswal	OK
33	Q2878-01L	MW-18B-57.5-081325	SD	08/21/25 14:05		Jaswal	OK
34	Q2878-01MS	MW-18B-57.5-081325	MS	08/21/25 14:08		Jaswal	OK
35	Q2878-01MSD	MW-18B-57.5-081325	MSD	08/21/25 14:11		Jaswal	OK
36	Q2878-01A	MW-18B-57.5-081325	PS	08/21/25 14:14	0.1 ML OF MP86790 AND 0.2ML OF EACH MP86791,MP86792 AND MP86793 IN 10ML SAMPLE.	Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136916

Review By	Janvi	Review On	8/25/2025 10:29:28 AM
Supervise By	jaswal	Supervise On	8/25/2025 3:50:25 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86739,MP86789,MP86788,MP86745,MP86744,MP86743,MP86742,MP86741,MP86740		
ICV Standard	MP86863,MP86788		
CCV Standard	MP86794		
ICSA Standard	MP86746,MP86747		
CRI Standard			
LCS Standard			
Chk Standard	MP86750,MP86748		

37	CCV03	CCV03	CCV	08/21/25 14:32		Jaswal	OK
38	CCB03	CCB03	CCB	08/21/25 14:34		Jaswal	OK
39	Q2878-02	MW-18B-57.5-081325	SAM	08/21/25 14:39		Jaswal	OK
40	Q2890-01DL	60296DL	SAM	08/21/25 14:42		Jaswal	OK
41	Q2890-02DL	60295DL	SAM	08/21/25 14:45	Fe,Mn high	Jaswal	Dilution
42	Q2890-03DL	60272DL	SAM	08/21/25 14:48	K high	Jaswal	Dilution
43	Q2906-01DL	3409DL	SAM	08/21/25 14:51		Jaswal	OK
44	Q2890-02DL2	60295DL2	SAM	08/21/25 15:06	25X for Fe, Mn	Jaswal	Confirms
45	Q2890-03DL2	60272DL2	SAM	08/21/25 15:09	25X for K	Jaswal	Confirms
46	CCV04	CCV04	CCV	08/21/25 15:12		Jaswal	OK
47	CCB04	CCB04	CCB	08/21/25 15:18		Jaswal	OK