



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Contract: JACO05

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.23	0.2	115	70 - 130	CV	08/21/2025	09:08	LB136901
CRI	Aluminum	20.2	20.0	101	70 - 130	P	08/21/2025	12:48	LB136916
	Antimony	2.16	2.0	108	70 - 130	P	08/21/2025	12:48	LB136916
	Arsenic	1.05	1.0	105	70 - 130	P	08/21/2025	12:48	LB136916
	Barium	10.7	10.0	107	70 - 130	P	08/21/2025	12:48	LB136916
	Beryllium	1.18	1.0	118	70 - 130	P	08/21/2025	12:48	LB136916
	Cadmium	0.98	1.0	98	70 - 130	P	08/21/2025	12:48	LB136916
	Calcium	555	500	111	70 - 130	P	08/21/2025	12:48	LB136916
	Chromium	2.13	2.0	106	70 - 130	P	08/21/2025	12:48	LB136916
	Cobalt	1.16	1.0	116	50 - 150	P	08/21/2025	12:48	LB136916
	Copper	2.11	2.0	106	70 - 130	P	08/21/2025	12:48	LB136916
	Iron	59.0	50.0	118	70 - 130	P	08/21/2025	12:48	LB136916
	Lead	1.03	1.0	103	70 - 130	P	08/21/2025	12:48	LB136916
	Magnesium	524	500	105	70 - 130	P	08/21/2025	12:48	LB136916
	Manganese	1.01	1.0	101	50 - 150	P	08/21/2025	12:48	LB136916
	Nickel	1.19	1.0	119	70 - 130	P	08/21/2025	12:48	LB136916
	Potassium	561	500	112	70 - 130	P	08/21/2025	12:48	LB136916
	Selenium	5.74	5.0	115	70 - 130	P	08/21/2025	12:48	LB136916
	Silver	1.10	1.0	110	70 - 130	P	08/21/2025	12:48	LB136916
	Sodium	558	500	112	70 - 130	P	08/21/2025	12:48	LB136916
	Thallium	1.00	1.0	100	70 - 130	P	08/21/2025	12:48	LB136916
	Vanadium	5.41	5.0	108	70 - 130	P	08/21/2025	12:48	LB136916
	Zinc	5.49	5.0	110	50 - 150	P	08/21/2025	12:48	LB136916