

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Contract: JACO05

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: MW-18B-57.5-081325A

Sample ID: Q2878-01 **Spiked ID:** Q2878-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Calcium	ug/L	75 - 125	127000		89300		50000	76		P
Potassium	ug/L	75 - 125	41000		20700		25000	81		P
Silver	ug/L	75 - 125	77.9		1.00	U	500	16	N	P