

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Contract: JACO05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169323BS							
Aluminum	ug/L	10000	10100		101	80 - 120	P
Antimony	ug/L	500	507		101	80 - 120	P
Arsenic	ug/L	500	500		100	80 - 120	P
Barium	ug/L	2500	2540		102	80 - 120	P
Beryllium	ug/L	500	494		99	80 - 120	P
Cadmium	ug/L	500	508		102	80 - 120	P
Calcium	ug/L	50000	50300		101	80 - 120	P
Chromium	ug/L	500	491		98	80 - 120	P
Cobalt	ug/L	500	500		100	80 - 120	P
Copper	ug/L	5000	4980		100	80 - 120	P
Iron	ug/L	25000	25800		103	80 - 120	P
Lead	ug/L	2500	2440		98	80 - 120	P
Magnesium	ug/L	50000	50600		101	80 - 120	P
Manganese	ug/L	5000	4930		99	80 - 120	P
Nickel	ug/L	500	495		99	80 - 120	P
Potassium	ug/L	25000	25600		102	80 - 120	P
Selenium	ug/L	500	503		101	80 - 120	P
Silver	ug/L	500	514		103	80 - 120	P
Sodium	ug/L	50000	51700		103	80 - 120	P
Thallium	ug/L	500	487		97	80 - 120	P
Vanadium	ug/L	500	493		99	80 - 120	P
Zinc	ug/L	5000	5020		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q2878

Contract: JACO05

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169327BS Mercury	ug/L	4.0	4.18		104	80 - 120	CV