

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2878

Client ID : SSTDCCC0.4

Date Analyzed: 08/19/2025

Lab File ID: BN037600.D

Time Analyzed: 10:39

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2628	7.717	6457	10.50	3216	14.34
UPPER LIMIT	5256	8.217	12914	10.998	6432	14.844
LOWER LIMIT	1314	7.217	3228.5	9.998	1608	13.844
EPA SAMPLE NO.						
01 MW-18B-57.5-081325	2177	7.72	5462	10.50	2861	14.35
02 MW-11A-37.5-081425	2250	7.72	5479	10.50	2729	14.35
03 MW-17B-55.5-08122025MS	2023	7.72	4765	10.50	2380	14.35
04 MW-17B-55.5-08122025MSD	2022	7.72	4839	10.50	2406	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.