

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135148
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158
CCC	VP135152
Internal Standard/PEM	
ICV/I.BLK	VP135159
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047347.D	15 Aug 2025 08:28		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX047348.D	15 Aug 2025 09:18	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX047349.D	15 Aug 2025 09:40	LR-58,81	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX047350.D	15 Aug 2025 10:01		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX047351.D	15 Aug 2025 10:22		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX047352.D	15 Aug 2025 10:43		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX047353.D	15 Aug 2025 11:05		JC/MD	Ok,M
8	IBLK	IBLK	VX047354.D	15 Aug 2025 11:26		JC/MD	Ok
9	VSTDIC001	VSTDIC001	VX047355.D	15 Aug 2025 12:08	Not used	JC/MD	Not Ok
10	VSTDIC001	VSTDIC001	VX047356.D	15 Aug 2025 12:30		JC/MD	Ok,M
11	VSTDICV050	ICVVX081525	VX047357.D	15 Aug 2025 13:11		JC/MD	Ok,M
12	VX0815WBL01	VX0815WBL01	VX047358.D	15 Aug 2025 13:39		JC/MD	Ok
13	VX0815MBL01	VX0815MBL01	VX047359.D	15 Aug 2025 14:00		JC/MD	Ok
14	VX0815WBS01	VX0815WBS01	VX047360.D	15 Aug 2025 14:21		JC/MD	Ok,M
15	VX0815WBSD01	VX0815WBSD01	VX047361.D	15 Aug 2025 14:49		JC/MD	Ok,M
16	Q2880-01	STORAGE-BLANK-SO	VX047362.D	15 Aug 2025 15:10	vial A pH<2	JC/MD	Ok
17	Q2880-02	STORAGE-BLANK-WA	VX047363.D	15 Aug 2025 15:32	vial A pH<2	JC/MD	Ok
18	Q2880-03	STORAGE-BLANK-WA	VX047364.D	15 Aug 2025 15:53	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135148		
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158		
CCC	VP135152		
Internal Standard/PEM			
ICV/I.BLK	VP135159		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2880-04	STORAGE-BLANK-SAM	VX047365.D	15 Aug 2025 16:15	vial A pH<2	JC/MD	Ok
20	Q2886-01	BP-TT182D-TB-202508	VX047366.D	15 Aug 2025 16:36	vial A pH<2	JC/MD	Ok
21	Q2886-02	BP-TT182D-GW-202508	VX047367.D	15 Aug 2025 16:58	vial A pH<2 Surrogate Fail	JC/MD	ReRun
22	Q2837-01	TW-WTS-13	VX047368.D	15 Aug 2025 17:19	vial B pH<2	JC/MD	Ok,M
23	VSTDCCC050	VSTDCCC050EC	VX047369.D	15 Aug 2025 17:41		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX082025

Review By	John Carlone	Review On	8/21/2025 8:55:57 AM
Supervise By	Maresh Dadoda	Supervise On	8/22/2025 1:13:48 AM
SubDirectory	VX082025	HP Acquire Method	MSVOA_X HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135191
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135192,VP135193

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047426.D	20 Aug 2025 07:50		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047427.D	20 Aug 2025 09:25	pH#Lot#V12668	JC/MD	Ok,M
3	VX0820MBL01	VX0820MBL01	VX047428.D	20 Aug 2025 09:46		JC/MD	Ok
4	VX0820WBL01	VX0820WBL01	VX047429.D	20 Aug 2025 10:07		JC/MD	Ok
5	VX0820WBS01	VX0820WBS01	VX047430.D	20 Aug 2025 10:35		JC/MD	Ok,M
6	VX0820WBSD01	VX0820WBSD01	VX047431.D	20 Aug 2025 11:06		JC/MD	Ok,M
7	Q2815-21RE	TW-11M-ERE	VX047432.D	20 Aug 2025 11:28	vial B pH<2 Surrogate Fail	JC/MD	Confirms
8	Q2815-22	TW-11M-S	VX047433.D	20 Aug 2025 11:49	vial B pH<2	JC/MD	Ok
9	Q2815-26	FB	VX047434.D	20 Aug 2025 12:11	vial B pH<2 FB	JC/MD	Ok
10	Q2878-01	MW-18B-57.5-081325	VX047435.D	20 Aug 2025 12:32	Run @ lower dilution	JC/MD	Not Ok
11	Q2878-05	MW-11A-37.5-081425	VX047436.D	20 Aug 2025 12:54	Run @ lower dilution	JC/MD	Not Ok
12	Q2869-03	MW-01-6.5-081325	VX047437.D	20 Aug 2025 13:15	Need straight run	JC/MD	Not Ok
13	Q2869-04	MW-11A-13.5-081325	VX047438.D	20 Aug 2025 13:37	Run @ 10x	JC/MD	Not Ok
14	Q2869-05	MW-19B-71.5-081325	VX047439.D	20 Aug 2025 13:59	Run @ 40x	JC/MD	Not Ok
15	IBLK	IBLK	VX047440.D	20 Aug 2025 14:20		JC/MD	Ok
16	Q2878-06	TB01-081425	VX047441.D	20 Aug 2025 14:42	vial A pH<2 TB	JC/MD	Ok
17	Q2900-07	MW-11B-081825	VX047442.D	20 Aug 2025 15:04	vial A pH<2 Surrogate Fail	JC/MD	ReRun
18	Q2900-01	MN-9C-081825	VX047443.D	20 Aug 2025 15:25	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX082025

Review By	John Carlone	Review On	8/21/2025 8:55:57 AM
Supervise By	Maresh Dadoda	Supervise On	8/22/2025 1:13:48 AM
SubDirectory	VX082025	HP Acquire Method	MSVOA_X HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135191		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135192,VP135193		

19	Q2900-02	MN-12C-081825	VX047444.D	20 Aug 2025 15:47	vial A pH<2 Surrogate Fail;Need 50X	JC/MD	Dilution
20	Q2900-03	DUP-01-081825	VX047445.D	20 Aug 2025 16:09	vial A pH<2 Surrogate Fail;Need 50X	JC/MD	Dilution
21	Q2900-04	MW-17B-081825	VX047446.D	20 Aug 2025 16:30	vial A pH<2 E flag of previous sample	JC/MD	ReRun
22	Q2900-05	MW-11C-081825	VX047447.D	20 Aug 2025 16:52	vial A pH<2 Need 50X	JC/MD	Dilution
23	Q2900-06	MW-17C-081825	VX047448.D	20 Aug 2025 17:14	vial A pH<2 Surrogate Fail	JC/MD	ReRun
24	Q2900-08	TB-081825	VX047449.D	20 Aug 2025 17:35	vial A pH<2 TB,Surrogate Fail	JC/MD	ReRun
25	Q2878-01	MW-18B-57.5-081325	VX047450.D	20 Aug 2025 17:57	vial A pH<2 Need 10X	JC/MD	Dilution
26	Q2878-01DL	MW-18B-57.5-081325	VX047451.D	20 Aug 2025 18:18	vial B pH<2	JC/MD	Ok,M
27	Q2878-05	MW-11A-37.5-081425	VX047452.D	20 Aug 2025 18:40	vial A pH<2	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX047453.D	20 Aug 2025 19:01		JC/MD	Ok,M

M : Manual Integration