

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q2878
 Lab File ID: VX047427.D Date Analyzed: 08/20/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:25
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	222102	5.56	369648	6.76	338607	10.05
UPPER LIMIT	444204	6.056	739296	7.263	677214	10.549
LOWER LIMIT	111051	5.056	184824	6.263	169304	9.549
EPA SAMPLE NO.						
MW-18B-57.5-081325	205051	5.57	415601	6.77	411051	10.06
MW-18B-57.5-081325DL	215725	5.57	432331	6.77	427580	10.06
MW-11A-37.5-081425	233413	5.57	475099	6.78	465473	10.06
TB01-081425	298809	5.57	610463	6.78	600442	10.06
VX0820WBL01	280589	5.57	574542	6.77	560929	10.06
VX0820WBS01	244648	5.56	407608	6.76	368354	10.06
VX0820WBSD01	227988	5.56	384011	6.76	353527	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.