



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	103	70 - 130	CV	08/19/2025	14:16	LB136879
CRI	Antimony	2.19	2.0	110	70 - 130	P	08/20/2025	12:15	LB136893
	Arsenic	1.04	1.0	104	70 - 130	P	08/20/2025	12:15	LB136893
	Barium	10.6	10.0	106	70 - 130	P	08/20/2025	12:15	LB136893
	Beryllium	1.19	1.0	119	70 - 130	P	08/20/2025	12:15	LB136893
	Cadmium	0.71	1.0	71	70 - 130	P	08/20/2025	12:15	LB136893
	Chromium	2.15	2.0	108	70 - 130	P	08/20/2025	12:15	LB136893
	Copper	2.09	2.0	104	70 - 130	P	08/20/2025	12:15	LB136893
	Lead	0.92	1.0	92	70 - 130	P	08/20/2025	12:15	LB136893
	Nickel	1.02	1.0	102	70 - 130	P	08/20/2025	12:15	LB136893
	Selenium	4.95	5.0	99	70 - 130	P	08/20/2025	12:15	LB136893
	Silver	1.19	1.0	119	70 - 130	P	08/20/2025	12:15	LB136893
	Thallium	1.03	1.0	103	70 - 130	P	08/20/2025	12:15	LB136893
	Vanadium	5.41	5.0	108	70 - 130	P	08/20/2025	12:15	LB136893
	Zinc	4.29	5.0	86	50 - 150	P	08/20/2025	12:15	LB136893