



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB23	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/19/2025	14:09	LB136879

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB71	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/19/2025	14:14	LB136879
CCB72	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/19/2025	14:47	LB136879
CCB73	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/19/2025	15:16	LB136879
CCB74	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	08/19/2025	15:39	LB136879

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	08/20/2025	11:56	LB136893
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/20/2025	11:56	LB136893
	Barium	0.21	+/-5	U	1.25	10.0	P	08/20/2025	11:56	LB136893
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/20/2025	11:56	LB136893
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/20/2025	11:56	LB136893
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/20/2025	11:56	LB136893
	Copper	0.30	+/-1	U	1.50	2.00	P	08/20/2025	11:56	LB136893
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/20/2025	11:56	LB136893
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/20/2025	11:56	LB136893
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/20/2025	11:56	LB136893
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	11:56	LB136893
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	11:56	LB136893
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/20/2025	11:56	LB136893
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/20/2025	11:56	LB136893

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.11	+/-1	J	0.25	2.00	P	08/20/2025	12:09	LB136893
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/20/2025	12:09	LB136893
	Barium	0.21	+/-5	U	1.25	10.0	P	08/20/2025	12:09	LB136893
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/20/2025	12:09	LB136893
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/20/2025	12:09	LB136893
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/20/2025	12:09	LB136893
	Copper	0.30	+/-1	U	1.50	2.00	P	08/20/2025	12:09	LB136893
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/20/2025	12:09	LB136893
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/20/2025	12:09	LB136893
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/20/2025	12:09	LB136893
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	12:09	LB136893
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	12:09	LB136893
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/20/2025	12:09	LB136893
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/20/2025	12:09	LB136893
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	08/20/2025	12:52	LB136893
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/20/2025	12:52	LB136893
	Barium	0.21	+/-5	U	1.25	10.0	P	08/20/2025	12:52	LB136893
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/20/2025	12:52	LB136893
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/20/2025	12:52	LB136893
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/20/2025	12:52	LB136893
	Copper	0.30	+/-1	U	1.50	2.00	P	08/20/2025	12:52	LB136893
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/20/2025	12:52	LB136893
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/20/2025	12:52	LB136893
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/20/2025	12:52	LB136893
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	12:52	LB136893
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	12:52	LB136893
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/20/2025	12:52	LB136893
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/20/2025	12:52	LB136893
CCB03	Antimony	0.14	+/-1	J	0.25	2.00	P	08/20/2025	13:30	LB136893
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	08/20/2025	13:30	LB136893
	Barium	0.21	+/-5	U	1.25	10.0	P	08/20/2025	13:30	LB136893
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	08/20/2025	13:30	LB136893
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	08/20/2025	13:30	LB136893
	Chromium	0.21	+/-1	U	0.75	2.00	P	08/20/2025	13:30	LB136893
	Copper	0.30	+/-1	U	1.50	2.00	P	08/20/2025	13:30	LB136893
	Lead	0.21	+/-0.5	U	0.75	1.00	P	08/20/2025	13:30	LB136893
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	08/20/2025	13:30	LB136893
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	08/20/2025	13:30	LB136893
	Silver	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	13:30	LB136893

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	U	0.50	1.00	P	08/20/2025	13:30	LB136893
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	08/20/2025	13:30	LB136893
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	08/20/2025	13:30	LB136893