

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2879
Contract: NOBI03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2879-18 **Client ID:** OU4-TS-GRILLO-TSCP11-081425DUP
Percent Solids for Sample: NA **Duplicate ID** Q2879-18DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	0.75	JD	0.80	JD	6		P
Arsenic	ug/L	20	31.9	D	32.5	D	2		P
Barium	ug/L	20	132	D	129	D	2		P
Beryllium	ug/L	20	12.2	D	12.9	D	6		P
Cadmium	ug/L	20	9.70	D	9.40	D	3		P
Chromium	ug/L	20	64.5	D	65.3	D	1		P
Copper	ug/L	20	248	D	252	D	2		P
Lead	ug/L	20	280	D	288	D	3		P
Nickel	ug/L	20	100	D	103	D	3		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	25.0	UD	25.0	UD			P
Silver	ug/L	20	5.00	UD	5.00	UD			P
Thallium	ug/L	20	0.65	JD	0.65	JD	0		P
Vanadium	ug/L	20	302	D	307	D	2		P
Zinc	ug/L	20	1750	D	1780	D	2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Nobis Group</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2879</u>
Contract:	<u>NOBI03</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2879-18MS</u>	Client ID:	<u>OU4-TS-GRILLO-TSCP11-081425MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2879-18MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	474	D	484	D	2		P
Arsenic	ug/L	20	693	D	711	D	3		P
Barium	ug/L	20	2260	D	2310	D	2		P
Beryllium	ug/L	20	476	D	483	D	1		P
Cadmium	ug/L	20	488	D	495	D	1		P
Chromium	ug/L	20	518	D	527	D	2		P
Copper	ug/L	20	4730	D	4820	D	2		P
Lead	ug/L	20	2560	D	2660	D	4		P
Nickel	ug/L	20	569	D	580	D	2		P
Mercury	ug/L	20	4.03		4.00		1		CV
Selenium	ug/L	20	442	D	455	D	3		P
Silver	ug/L	20	79.4	D	80.4	D	1		P
Thallium	ug/L	20	424	D	434	D	2		P
Vanadium	ug/L	20	749	D	765	D	2		P
Zinc	ug/L	20	6210	D	6340	D	2		P