

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169293BS Mercury	ug/L	4.0	4.15		104	82 - 119	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2879

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169300BS							
Antimony	ug/L	500	515		103	85 - 117	P
Arsenic	ug/L	500	501		100	84 - 116	P
Barium	ug/L	2500	2540		102	86 - 114	P
Beryllium	ug/L	500	515		103	83 - 121	P
Cadmium	ug/L	500	517		103	87 - 115	P
Chromium	ug/L	500	514		103	85 - 116	P
Copper	ug/L	5000	5140		103	85 - 118	P
Lead	ug/L	2500	2480		99	88 - 115	P
Nickel	ug/L	500	513		103	85 - 117	P
Selenium	ug/L	500	499		100	80 - 120	P
Silver	ug/L	500	516		103	85 - 116	P
Thallium	ug/L	500	502		100	82 - 116	P
Vanadium	ug/L	500	514		103	86 - 115	P
Zinc	ug/L	5000	5230		105	83 - 119	P